**Dieses Dokument ist eine Zweitveröffentlichung (Verlagsversion) /
This is a self-archiving document (published version):**

Joanne E. Adaway, Mirko Peitzsch, Brian G. Keevil

**A novel method for the measurement of plasma metanephrines using
online solid phase extraction-liquid chromatography tandem mass
spectrometry**

Erstveröffentlichung in / First published in:

Annals of Clinical Biochemistry. 2015, 52(3), S. 361 – 369 [Zugriff am: 13.08.2019]. SAGE
journals. ISSN 1758-1001.

DOI: <https://doi.org/10.1177/0004563214546560>

Diese Version ist verfügbar / This version is available on:

<https://nbn-resolving.org/urn:nbn:de:bsz:14-qucosa2-354293>

„Dieser Beitrag ist mit Zustimmung des Rechteinhabers aufgrund einer (DFGgeförderten) Allianz- bzw. Nationallizenz frei zugänglich.“

This publication is openly accessible with the permission of the copyright owner. The permission is granted within a nationwide license, supported by the German Research Foundation (abbr. in German DFG).

www.nationallizenzen.de/

A novel method for the measurement of plasma metanephrines using online solid phase extraction-liquid chromatography tandem mass spectrometry

Joanne E Adaway^{1,2}, Mirko Peitzsch³ and Brian G Keevil^{1,2}

Abstract

Background: Measurement of plasma metanephrine, normetanephrine and 3-methoxytyramine is useful in the diagnosis of pheochromocytomas, but many assays require a large volume of plasma due to poor assay sensitivity, and often require lengthy sample preparation. Our aim was to develop a method for measurement of plasma metanephrines using a small sample volume with minimal hands-on preparation.

Methods: Samples were deproteinised using 10 K spin filters prior to online solid phase extraction using a Waters Acquity UPLC Online SPE Manager (Waters, Manchester, UK) coupled to a Waters Xevo TQ-S mass spectrometer (Waters, Manchester, UK). The assay was validated and results compared to a previously published method.

Results: We achieved a limit of quantification of 37.5 pmol/L for metanephrine and 3-methoxytyramine and 75 pmol/L for normetanephrine using only 150 µL of sample. The assay was linear up to 30,000 pmol/L for all analytes and in a method comparison study results showed good agreement with a previously published LC-MS/MS assay.

Conclusions: We have developed a simple method for measurement of plasma metanephrine, normetanephrine and 3-methoxytyramine using only 150 µL of sample. There is minimal hands-on sample preparation required and the assay is suitable for routine use in a clinical laboratory.

Keywords

Metanephrine, normetanephrine, 3-methoxytyramine, pheochromocytoma, LC-MS/MS, online solid phase extraction

Accepted: 16th July 2014

Introduction

Pheochromocytomas and paragangliomas (PPGLs) are rare tumours of adrenal chromaffin cells or similar tissue in extra-adrenal paraganglia. Biochemical diagnosis of these tumours is important due to the non-specific nature of symptoms such as hypertension, palpitations, flushing and sweating. Analysis of plasma metanephrines is used as a first line test for the diagnosis of adrenal and extra-adrenal sympathetic PPGLs.¹

Plasma metanephrines are produced from catecholamines within chromaffin cells by the action of

¹Department of Clinical Biochemistry, University Hospital South Manchester NHS Foundation Trust, Manchester, UK

²Manchester Academic Health Science Centre, University Hospital South Manchester, The University of Manchester, Manchester, UK

³Institute of Clinical Chemistry and Laboratory Medicine, University Hospital Carl Gustav Carus, Medical Faculty Carl Gustav Carus, Technische Universität Dresden, Dresden, Germany

Corresponding author:

Joanne Adaway, Department of Clinical Biochemistry, Wythenshawe Hospital, Southmoor Road, Wythenshawe, Manchester M23 9LT, UK.
Email: Jo.adaway@uhsm.nhs.uk

catechol-O-methyl transferase (COMT). Adrenaline is metabolized to metanephrine (MN), noradrenaline to normetanephrine (NMN) and dopamine to 3-methoxytyramine (3MT). Many of the current plasma metanephrine assays only measure MN and NMN, but there is evidence that measurement of 3MT gives important information on tumour location, the presence of underlying mutations of SDHB and SDHD genes and the likelihood of metastases,² with 3MT shown to be higher in patients with metastases compared to those without. It is advantageous to offer 3MT measurement as part of plasma metanephrine analysis, but this can be challenging due to the low picomolar concentrations found in the majority of samples.

Current methods for the measurement of plasma metanephrines include immunoassay,^{3,4} HPLC with electrochemical detection⁵ and liquid chromatography tandem mass spectrometry (LC-MS/MS).^{6–10} Immunoassays do not offer the option of 3MT analysis, and HPLC with electrochemical detection suffers from interferences from common substances such as paracetamol,¹¹ so LC-MS/MS is increasingly the method of choice for metanephrine analysis.

LC-MS/MS analysis of analytes in complex matrices such as plasma often requires extensive sample preparation. Solid phase extraction is the most common sample preparation method for plasma metanephrines, with both offline^{7–10} and online⁶ methods reported in the literature; however, 3MT was not measured in two of these assays,^{7,9} and some required high sample volumes for analysis.^{6,7}

We wanted to develop an online solid phase extraction (SPE) method for the measurement of plasma metanephrines to decrease the hands-on preparation time for use in a busy clinical laboratory. The Waters Online SPE Manager (OSM), (Waters, Manchester, UK) is an online separation device designed for use with the very high back pressure encountered when running UPLC conditions. In particular, the extraction cartridges have been manufactured with thicker walls, to ensure a liquid tight junction under the higher pressures, but this reduces the amount of chromatographic packing material available for use. As a result of this fundamental design change, the preconditioning, washing and elution parameters of the OSM are uniquely different from those reported for similar systems designed for use under HPLC conditions.⁶ We often receive low volume samples for plasma metanephrine analysis and wished to develop a method with minimal sample volume requirements with sufficient sensitivity to allow the quantification of MN, NMN and 3MT. We felt that the OSM combined with a highly sensitive mass spectrometer would allow us to develop such a method.

In this report, we describe the development and validation of a semi-automated online SPE-LC-MS/MS

method for the measurement of MN, NMN and 3MT using only 150 µL of plasma.

Materials and methods

Calibrator and reagent preparation

Stock solutions were prepared by dissolving 10 mg of 3-methoxytyramine powder (Sigma-Aldrich, Poole, UK) in 0.1 mol/L hydrochloric acid. Catecholamine mix 2, consisting of 1 mg/mL of MN and NMN in methanol was used as the stock for MN and NMN (LGC standards, Middlesex, UK). Separate stocks were used for standards and quality control samples (QCs). The stocks were then diluted in phosphate buffered saline pH 7.4 (PBS; Sigma-Aldrich, Poole, UK) containing 0.1% (w/v) bovine serum albumin (BSA; Sigma, Poole, UK), to give standard concentrations of 0–30,000 pmol/L for MN and 3MT and 0–32,300 pmol/L for NMN and QC concentrations of 200, 1000 and 3000 pmol/L for MN and 3MT and 220, 1090 and 3270 pmol/L for NMN. Aliquots (200 µL) of these were stored at –80°C for up to six months. A mixed internal standard of deuterated d3-metanephrine (α -d2 β -d1), d3-normetanephrine (α -d1, β -d2) and d4-3-methoxytyramine (α -d2, β -d2) (Medical Isotopes, New Hampshire, USA) was used as the internal standard at a working concentration of 2000 pmol/L in water. This was found to be stable at 4°C for six months.

Sample preparation

Standards, QCs and patients' samples (150 µL) were pipetted directly into a 10 K modified PES centrifugal filter (VWR, Leicestershire, UK). To this, 150 µL of working internal standard was added and the filter vortexed for 10 s. The filter was then centrifuged at 9962 g for 20 min, and the filtrate transferred into a well of a 96-deep well block (Porvair Sciences, Wrexham, UK). The block was heat sealed (Thermo, Hemel Hempstead, UK), and then centrifuged at 8000 g for 5 min.

Online solid phase extraction and liquid chromatography

The plate was transferred into a Waters Acquity UPLC system coupled to an Online SPE manager (OSM) (Waters, Manchester, UK). The OSM is under full software control and allows preconditioning, loading, washing and elution of the SPE cartridges. An Oasis weak cation exchange (WCX) 1 × 10 mm and 30 µm SPE cartridge (Waters, Manchester, UK) were placed in the right clamp and preconditioned twice with 200 µL of acetonitrile containing 2% (v/v) formic acid, followed by 250 µL of 80% acetonitrile: 20%

10 mmol/L ammonium formate pH 3.2 in water followed by 250 μ L of 95% acetonitrile. The cartridge was then equilibrated with 250 μ L water, and 100 μ L of sample was loaded onto the cartridge via the Acquity autosampler with 250 μ L of water. Unbound contaminants were removed by washing the cartridge with 200 μ L of water followed by 200 μ L of 95% acetonitrile. The cartridge was then automatically transferred to the left hand clamp for elution of the analytes from the cartridge directly onto the analytical column under the initial chromatographic conditions. The analytical column was an Atlantis HILIC Silica 2.1 $\times$ 50 mm, 3 μ m column (Waters, Manchester, UK), coupled to a SecurityGuard Gemini C18 4 $\times$ 2 mm guard cartridge (Phenomenex, Macclesfield, UK). Whilst eluting the sample in the left clamp, the next sample is prepared in the right clamp, thus improving throughput.

Mobile phase A contained 100 mmol/L ammonium formate in deionized water adjusted to pH 3.2 with formic acid, and mobile phase B contained acetonitrile. Initial conditions were 5:95 (v/v) A:B. Metanephrines were eluted from the column using a gradient of 5–20% mobile phase A over 4.1 min. These conditions were held for 0.3 min before returning to initial conditions for a further 2.74 min to re-equilibrate the column. The flow rate was 0.3 mL/minute throughout and the total run time was 7.15 min.

Mass spectrometry

The eluate was injected from the OSM directly into a Xevo TQ-S tandem mass spectrometer (Waters, Manchester, UK). MassLynx software was used for system control and the MassLynx TargetLynx programme allowed data processing. The software used the height of the detected peaks, $1/\times$ weighting and linear least squares regression to produce a standard curve. The mass spectrometer was operated in electrospray positive ionization mode, the capillary was maintained at 0.5 kV and the source temperature was 150°C. The desolvation gas flow and temperature were 800 L/h and 650°C, respectively. The transitions and collision energies for the analytes are given in Table 1. The cone voltage was 40 V for all analytes with a source offset of 30 V. Transitions were monitored in multiple reaction monitoring (MRM) mode with a dwell time of 0.15 s.

Assay validation

Ion suppression

To investigate ion suppression, we infused a 2 nmol/L solution of d3-MN, d3-NMN and d4-3MT in water directly into the mass spectrometer via the Waters

Table 1. MRM Transitions and collision energies for each analyte.

Analyte	Precursor ion (m/z)	Product ion (m/z)	Collision energy (eV)
MN	180	148.1	17
d3-MN	183.1	151.1	15
NMN	166.0	134.1	15
d3-NMN	169.0	137.1	17
3MT	151.0	119.05	12
d4-3MT	155.0	123.1	14

Intellistart™ Fluidics system to give a constant background signal. Plasma samples ($n = 6$) were prepared as detailed earlier and injected simultaneously via the autosampler. Ion suppression is seen as a reduction in background signal, and this is significant if it occurs where the compound of interest elutes, and although there is no official guidance on acceptable concentrations of ion suppression, for the purpose of this study we decided that a reduction in signal of less than 10% would be acceptable. In addition, plasma samples ($n = 5$) were doubly diluted with PBS/0.1% (v/v) BSA up to a 1:32 dilution to assess the linearity of the dilutions, as the compounds of interest would dilute non-linearly in the presence of ion suppression.

Lower limit of quantification

The lower limit of quantification (LLOQ) was in accordance with FDA Guidance.¹² Fifteen replicates of low, medium and high concentrations of MN, NMN and 3MT in PBS/0.1% BSA were analysed and the LLOQ was determined as the lowest concentration of each with % coefficient of variation (CV) <20 and accuracy within 20% of expected concentrations.

Imprecision and accuracy

To determine the intra-assay imprecision of the assay, the three QC samples were prepared and analysed 10 times within one batch. To determine the inter-assay imprecision, the three samples were analysed 15 times in different batches. The imprecision was deemed acceptable if the CV was <15%, and the accuracy was deemed acceptable if the mean value was within 15% of the weighed in value.

Linearity

The linearity of the method was evaluated by analysing calibrators containing MN and 3MT up to a concentration of 30,000 pmol/L and NMN up to

32,300 pmol/L. Calibration curves ($n=10$) were produced using TargetLynx software, which plotted calibrator concentration values against the corresponding Online SPE-LC-MS/MS response (analyte peak height/stable isotope labelled analyte peak height). The assay was determined to be linear if the $1/\times$ linear regression analysis produced an $r^2 > 0.99$.

Recovery

Three different concentrations of the three analytes (500, 1000 and 3000 pmol/L) were spiked onto six different plasma samples. Total recovery was calculated from measured compared to expected concentrations, and was deemed to be acceptable if between 80 and 120%. Extraction recovery was determined using the Advanced Method Development (AMD) feature in the MassLynx software. Briefly, two cartridges were placed in series, and a sample and washes were applied to the first cartridge. Any sample lost from the first cartridge (breakthrough analyte) was captured on the second cartridge. The second cartridge was then eluted, followed by the first, and the amount of analyte eluted was quantified using TargetLynx software. Extraction recovery was calculated as the peak height of the analyte recovered from cartridge 1 divided by the peak height of breakthrough analyte plus the peak height of the analyte eluted from cartridge 1.

Method comparison

Plasma samples ($n=206$) were analysed by online SPE-LC-MS/MS and by LC-MS/MS with offline SPE.¹⁰ All samples used were part of the prospective monoamine-producing tumor study (<https://pmt-study.pressor.org>) and were taken after an overnight fast and with the patient in the fully supine position for 30 min before blood sampling. Samples were initially analysed by offline SPE within one month of collection. They were then stored at -80°C for a mean of 1.68 years, prior to shipping frozen in dry ice and storage for a maximum of 2 weeks before analysis by online SPE. Samples have been shown to be stable under these conditions.¹³ All subjects provided informed consent under protocols approved by the local Ethics committees at each participating centre.

Diagnostic sensitivity and specificity

Receiver Operating Characteristic (ROC) analysis was carried out on 206 plasma samples to ascertain the diagnostic sensitivity and specificity of both online and offline SPE assays. Of the samples, 36 were from patients with confirmed PPGLs and 170 were

from patients in whom PPGL had been excluded as a diagnosis.

Drug interference

It has previously been shown that isoproterenol and MDMA interfere with normetanephrine quantification, and isoetharine and MDA interfere with metanephrine.¹⁴ To investigate this, compounds were spiked into water plus two plasma pools at concentrations at least three times the upper limit of the appropriate reference limit (isoproterenol 47 $\mu\text{mol/L}$, MDMA 5.2 $\mu\text{mol/L}$, isoetharine 42 $\mu\text{mol/L}$ and MDA 5.6 $\mu\text{mol/L}$). The concentrations of metanephrine, normetanephrine and 3MT were calculated and compared to those in the pools spiked with an equivalent volume of water.

Statistical analysis

All statistical analyses were carried out using Analyse-it software (Analyse-it software Ltd, Leeds, UK) apart from ROC analysis which was done using the JMP Pro 10.0 software package (SAS, Cary, NC, USA).

Results

SPE and liquid chromatography

Chromatographic separation of 3MT, MN and NMN was achieved within 4 min (Figure 1) with a further 3 min required to wash and re-equilibrate the column, giving a total run time of 7.15 min. Each analyte co-eluted with the stable isotope labelled version.

Ion suppression

There was no significant suppression of the background signal at the time of analyte elution for any of the three analytes. Ion enhancement occurs after 5 min, but this does not interfere with analyte quantification. We further investigated ion suppression by serially diluting patients' samples in PBS/0.1% (v/v) BSA. The analyte:internal standard peak ratio was used to calculate the concentration of each dilution from a standard curve, and these concentrations were plotted against the expected values. Dilution linearity of five patients' samples showed an r^2 value > 0.99 for all three analytes, indicating that ion suppression does not affect this assay.

Lower limit of quantification

The lower limit of quantification for MN was 37.5 pmol/L, with a CV of 7% and a bias of +5.9%. The LLOQ for NMN and 3MT was 75 pmol/L, with

Figure 1. Chromatogram of a plasma sample containing 149 pmol/L metanephrine, 345 pmol/L normetanephrine and 110 pmol/L 3-methoxytyramine. The internal standards d3-metanephrine, d3-normetanephrine and d4-3MT are also shown on the chromatogram.

CVs of 5.8 and 5.4% and deviation from the target value of -3.4 and $+8.4\%$, respectively.

Imprecision

The intra- and inter-assay imprecision and bias was acceptable at low, medium and high concentrations for all three analytes, with CV and deviation from the target value below 10% for all analytes at all three concentrations (Table 2).

Linearity

Standard curves were produced by plotting the analyte concentration on the x -axis and the response (analyte peak height/internal standard peak height) on the y -axis. The curve was linear up to 30,000 pmol/L for MN and 3MT and up to 32,300 pmol/L for NMN, with r^2 values greater than 0.99 over multiple batches. A representative slope for MN was 0.9997 (SE 0.0002) with an intercept of 2.4604 (SE 2.0274), and the SE of the residuals was 4.1438. For NMN, a representative slope was 0.9971 with an intercept of 21.91 (SE 0.0014 and 19.09, respectively), and SE of the residuals was 39.03. A typical 3MT standard curve had a slope of 1.0053 (SE 0.0030) with an intercept of -36.95 (SE 38.88), and the SE of the residuals was 79.47.

Table 2. Intra- and inter-assay imprecision and deviation from target values.

Analyte/ concentration (pmol/L)	Intra-assay CV (%)	Deviation from target (%)	Inter-assay CV (%)	Deviation from target
Metanephrine				
200	9.5	8.3	9.3	3.5
1000	6.2	6.5	4.7	7.0
3000	3.7	5.3	3.3	6.3
Normetanephrine				
220	7.9	3.6	2.2	3.4
1090	4.1	0.1	1.6	2.6
3270	3.3	2.4	1.6	4.3
3-Methoxytyramine				
200	8.7	8.5	2.9	-3.8
1000	3.7	-1.8	2.5	-5.9
3000	2.1	-0.6	2.0	-2.1

Recovery

The mean total recovery of MN, NMN and 3MT was 98% (87–108%), 97% (86–107%) and 95% (80–109%), respectively, which were all within acceptable limits. The extraction recovery was 92% for MN, 94% for

NMN and 97% for 3MT, indicating that losses from the cartridge are minimal.

Method comparison

MN, NMN and 3MT were measured in 206 samples by online SPE-LC-MS/MS and by a previously published LC-MS/MS method¹⁰ with offline solid phase extraction. Sixteen of the 206 samples were below the LLOQ of our assay for MN and were excluded from the comparison. Bland-Altman analysis showed the online SPE method had a mean negative bias of 23% over a range of 37.5 pmol/L to >12000 pmol/L (Figure 2(a)). All the NMN results were within the measuring range, and Bland-Altman analysis showed a mean positive bias

of 0.1% (Figure 2(b)). The NMN results ranged from 113 pmol/L to greater than 40,000 pmol/L. Of the 206 samples, 163 were below the measuring range of 3MT in our assay, leaving 43 for inclusion in the comparison, with the highest 3MT concentration >22,000 pmol/L. A mean negative bias of 12.4% was shown in the Bland-Altman analysis (Figure 2(c)).

Diagnostic sensitivity and specificity

The diagnostic sensitivity and specificity was calculated for each assay using the cut-offs used in each laboratory (Table 3). The reference ranges used for online SPE were based on the work of Peaston et al.⁸ with <510 pmol/L for MN, <1180 pmol/L for NMN and

Figure 2. Comparison between offline and online SPE methods for metanephrine, normetanephrine and 3-methoxytyramine. Panel (a) shows Bland-Altman analysis for metanephrine samples, and panels (b) and (c) show Bland-Altman analysis for normetanephrine and 3-methoxytyramine, respectively. The solid line represents the mean percentage difference between the two methods; dotted lines represent the confidence limits which are 1.96% SD.

<180 pmol/L for 3MT. The reference ranges for the offline SPE method were 446 pmol/L for MN and 179 pmol/L for 3MT.¹⁰ Age-adjusted reference ranges were used for NMN, as described by Eisenhofer et al.¹⁵ ranging from 590 pmol/L for a 5-year-old to 1037 pmol/L for patients 60 years old and above. Using these cut-offs, the sensitivity of the online assay was 94.4%, whilst the offline assay had a sensitivity of 100%. The specificity of the two assays was similar at 96.5% for online SPE and 95.8% for offline SPE (Table 3). ROC analysis confirms the good diagnostic sensitivity and specificity of both assays, with an area under the curve (AUC) of 0.995 for offline SPE and 0.979 for online SPE (Figure 3).

Drug interference

None of the drugs tested showed peaks in water at the retention times of MN, NMN or 3MT. In addition, the drugs did not affect the concentration of any of the three analytes in the plasma pools, indicating that interference from these drugs is negligible in this assay.

Discussion

We have developed a method for the measurement of MN, NMN and 3MT in plasma that is suitable for routine clinical use. The combination of centrifugal filters and online SPE for sample preparation is novel and require minimal hands-on preparation time, which is important in a busy clinical laboratory. The filtration is required in this assay as the OSM is coupled to an Acquity UPLC system, and the viscosity of plasma prepared without filtration would be too high for the narrow bore tubing of the UPLC system and would quickly cause blockages. We tried other methods to remove proteins such as acetonitrile precipitation and trichloroacetic acid precipitation, but found that we did not get reproducible results using these techniques. We found that the recovery obtained from deproteinising the samples using the centrifugal filters was acceptable and consistent and therefore decided to validate the method using this sample preparation.

It has recently been reported that MN may cause ionic cross-talk in LC-MS/MS assays leading to

overestimation of 3MT.¹⁶ This is not a problem with this assay as the resolution between 3MT and MN is sufficient that cross-talk does not interfere in the quantification of 3MT (Figure 1).

The lower limits of quantification for MN, NMN and 3MT allow quantification of samples within published reference ranges,⁸ and the linearity of the assay up to 30,000 pmol/L for MN and 3MT and 32,300 pmol/L for NMN allows quantification of the majority of raised samples. If we obtain a result higher than this, the dilution linearity experiment proved that it is valid to dilute the sample with PBS/0.1% (v/v) BSA to produce a concentration within the measuring range of the assay that can then be multiplied back up.

A method comparison against a published offline SPE LC-MS/MS method¹⁰ was carried out using samples taken from patients who were enrolled in the prospective monoamine producing tumour study. The comparison between the two methods was acceptable, with a small negative bias observed with the online-SPE method for MN and 3MT and a small positive bias for

Figure 3. Receiver Operating Characteristic (ROC) curve analysis for offline SPE and online SPE. The sensitivity and specificity of each assay are plotted for different cut-off points. The area under the curve is shown for each method.

Table 3. Diagnostic sensitivity and specificity of online and offline SPE methods.

	True positive	True negative	False positive	False negative	Diagnostic sensitivity (%)	Diagnostic specificity (%)
Online SPE	34	164	6	2	94.4	96.5
Offline SPE	36	163	7	0	100	95.8

Note: The sensitivity and specificity were calculated using the cut-offs currently in use in the two laboratories.

NMN, and the diagnostic sensitivity and specificity for both the assays were shown to be very good. The largest differences were in very high results; this would not affect the diagnosis of any patient and may be due to calibration differences between the two assays. The samples were taken from patients after an overnight fast who had been in the supine position for 30 min. We currently use reference ranges derived from the work of Peaston et al.⁸ which were based on samples taken from patients in the sitting position. This reference range was used as we carried out a comparison of our previous plasma metanephrine assay against the Peaston method, and our previous method compared well with our current method, with a mean bias of -3.2% for metanephrine and -4.1% for normetanephrine. It has been established that supine rest prior to blood sampling gives significantly lower results for the metanephrines compared to samples taken in the sitting position,^{17,18} and the difference between the cut-offs used may explain the false negative results with our method. Using the age-adjusted reference ranges for NMN and the supine cut-offs for MN and 3MT,¹⁵ the sensitivity of our assay does improve to 97.2% (35 true positive, one false negative) in this supine, fasted population. This is at the expense of specificity, which falls to 91.3% (11 false positives, 115 true negatives); however, it may be argued that as plasma metanephrine analysis is used as a screening test for PPGL, the importance of not missing the diagnosis is such that the sensitivity should be maximized.

This work was carried out using samples that had been collected with ideal pre-venepuncture patient preparation, i.e. patient fully supine for 30 min. However, we are aware that the vast majority of samples that we receive are taken from patients in the sitting position and that the facilities for supine blood sampling may not be readily available in phlebotomy clinics. In such patients, therefore, the use of a fasting, supine reference range would significantly decrease the specificity of this test. Asking for a repeat sample collected under ideal conditions would increase the specificity, but as Därr et al.¹⁸ suggest that there is a problem with this approach, as for a variety of reasons the majority of such patients do not get repeat metanephrine analysis undertaken. Därr et al.¹⁸ also suggest that if supine samples cannot be taken, urine fractionated metanephrines should be used as an alternative test. However, the accurate collection of timed urine samples may also pose difficulties and we therefore suggest that the use of a 'sitting position' reference range for plasma metanephrines may represent a pragmatic alternative approach.

In conclusion, we have developed a method for plasma metanephrine analysis that is suitable for use in the clinical laboratory using a minimal amount of

sample. We plan to carry out further work on appropriate cut-offs for our population to maximize the diagnostic sensitivity and specificity of the test in our routine patients' samples.

Acknowledgements

Thanks are extended to members of Prospective Monoamine Tumor trial (Graeme Eisenhofer & Roland Därr, University Hospital Dresden at the TU Dresden, Dresden, Germany; Aleksander Prejbisz & Andrzej Januszewicz, Institute of Cardiology, Warsaw, Poland; Dirk Weismann & Martin Fassnacht, University Hospital Würzburg, Würzburg, Germany; and Felix Beuschlein, University Hospital Munich, Munich, Germany) for their contributions of patient materials.

Declaration of conflicting interests

None declared.

Funding

This research received no specific grant from any funding agency in the public, commercial, or not-for-profit sectors.

Ethical approval

University Hospital of Dresden (EK189062010), Institute of Cardiology, Warsaw (1233-08112010), Medical University Hospital of Würzburg (104/11), University Hospital of Munich (173-11).

Guarantor

JA.

Contributorship

BK conceived the study. The development and validation of the assay was performed by JA with aid from BK. MP provided the patients' samples with corresponding ethical approval and carried out some of the statistical analysis. JA wrote the first draft of the paper, all authors reviewed and edited the manuscript and approved the final version.

References

1. Pheochromocytoma and Paraganglioma: An Endocrine Society clinical practice guideline. *JCEM* 2014; 99: 1915–1942.
2. Eisenhofer G, Lenders JW, Siegert G, et al. Plasma methoxytyramine: a novel biomarker of metastatic pheochromocytoma and paraganglioma in relation to established risk factors of tumour size, location and SDHB mutation status. *Eur J Cancer* 2012; 48: 1739–1749.
3. Sarathi V, Pandit R, Patil VK, et al. Performance of plasma fractionated free metanephrines by enzyme immunoassay in the diagnosis of pheochromocytoma

- and paraganglioma in children. *Endocr Pract* 2012; 18: 694–699.
4. Christensen TT, Frystyk J and Poulsen PL. Comparison of plasma metanephrines measured by a commercial immunoassay and urinary catecholamines in the diagnosis of pheochromocytoma. *Scand J Clin Lab Invest* 2011; 71: 695–700.
 5. Lenders JW, Eisenhofer G, Armando I, et al. Determination of metanephrines in plasma by liquid chromatography with electrochemical detection. *Clin Chem* 1993; 39: 97–103.
 6. de Jong WH, Graham KS, van der Molen KC, et al. Plasma free metanephrine measurement using automated online solid phase extraction HPLC tandem mass spectrometry. *Clin Chem* 2007; 53: 1684–1693.
 7. Lagerstedt SA, O’Kane DJ and Singh RJ. Measurement of plasma free metanephrine and normetanephrine by liquid chromatography-tandem mass spectrometry for diagnosis of pheochromocytoma. *Clin Chem* 2004; 50: 603–611.
 8. Peaston RT, Graham KS, Chambers E, et al. Performance of plasma free metanephrines measured by liquid chromatography-tandem mass spectrometry in the diagnosis of pheochromocytoma. *Clin Chim Acta* 2010; 411: 546–552.
 9. Petteys BJ, Graham KS, Parnas ML, et al. Performance characteristics of an LC-MS/MS method for the determination of plasma metanephrines. *Clin Chim Acta* 2012; 413: 19–20.
 10. Peitzsch M, Prejbisz A, Kroiß M, et al. Analysis of plasma 3-methoxytyramine, normetanephrine and metanephrine by ultraperformance liquid chromatography-tandem mass spectrometry: utility for diagnosis of dopamine-producing metastatic phaeochromocytoma. *Ann Clin Biochem* 2013; 50: 147–155.
 11. Wilson SP, Kamin DL and Feldman JM. Acetaminophen administration interferes with urinary metanephrine (and catecholamine) determinations. *Clin Chem* 1985; 31: 1093–1094.
 12. *Guidance for Industry. Bioanalytical Method Validation.* US Department of Health and Human Services. Food and Drug Administration, Center for Drug Evaluation and Research (CDER) and Center for Veterinary Medicine (CVM) May 2001.
 13. Ross HA, Lenders JW and Sweep FC. A study of longer time stability of plasma free metanephrines. *Ann Clin Biochem* 2011; 48: 270–271.
 14. Petteys BJ, Graham KS, Parnás L, et al. Performance characteristics of an LC-MS/MS method for the determination of plasma metanephrines. *Clin Chim Acta* 2012; 413: 1459–1465.
 15. Eisenhofer G, Lettke P, Herberg M, et al. Reference intervals for plasma free metanephrines with an age adjustment for normetanephrine for optimized laboratory testing of phaeochromocytoma. *Ann Clin Biochem* 2013; 50: 62–69.
 16. Twentyman JM, Cradic KW, Singh RJ, et al. Ionic cross talk can lead to over-estimation of 3-methoxytyramine during quantification of metanephrines by mass spectrometry. *Clin Chem* 2012; 58: 1156–1158.
 17. Lenders JWM, Willemsen JJ, Eisenhofer G, et al. Is supine rest necessary before blood sampling for plasma metanephrines? *Clin Chem* 2007; 53: 352–354.
 18. Därr R, Deinum J, Schultzekool, et al. Biochemical diagnosis of phaeochromocytoma using plasma-free normetanephrine, metanephrine and methoxytyramine: importance of supine sampling under fasting conditions. *Clin Endocrinol* 2014; 80: 478–486.