**Development of novel polymer matrices for
MALDI MS and MALDI MS Imaging**

D I S S E R T A T I O N

zur Erlangung des akademischen Grades

**Doctor rerum naturalium
(Dr. rer. nat.)**

vorgelegt

dem Bereich Mathematik und Naturwissenschaften
der Fakultät Chemie und Lebensmittelchemie
der Technischen Universität Dresden

von

M.Sc. Kilian Martin Horatz
geboren am 16.05.1992 in Köln

Eingereicht am 25.06.2021

Verteidigt am 28.10.2021

Diese Dissertation wurde in der Zeit von 10/2017 bis 03/2021
am Leibniz Institut für Polymerforschung Dresden e.V. angefertigt.

Gutachter:

1) Prof. Dr. Brigitte Voit

2) Prof. Dr. Dominik L. Michels

3) Dr. Franziska Lissel

“Always look on the bright side of life.”
- Monty Python -

This dissertation was carried out under the scientific supervision of:
Dr. Franziska Lissel and Prof. Dr. Brigitte Voit

1st reviewer: Prof. Dr. Brigitte Voit
2nd reviewer: Prof. Dr. Dominik L. Michels
3rd reviewer: Dr. Franziska Lissel

Submitted: 25.06.2021

Defended: 28.10.2021

Abstract

Matrix assisted laser desorption/ionization mass spectrometry (MALDI MS) and the corresponding visualization technique *MALDI MS Imaging* (MSI) have emerged as important analytical tools in biochemical sciences, e.g., for drug development or to trace the metabolomic changes in cancerous tissues. Initially developed for the detection of *high molecular weight compounds* (HMWC; $M > 1000$ Da), in recent years the reliable and reproducible detection of *low molecular weight compounds* (LMWC; $M < 1000$ Da) has gained high attention, e.g., in the research fields of metabolomics and lipidomics. By using a protective matrix, the MALDI technique is capable of soft ionization of analytes to prevent their fragmentation or degradation. This matrix is responsible for the spatial separation of the analyte molecules, their protection from the strong laser shots, and their ionization. Commonly used matrices are *small organic matrices* (SOMs; $M < 500$ Da), which are utilized in HMWC analytics and recently also in LMWC analytics since they show sufficient absorption of the laser radiation, high crystallinity, and good ionization efficiency. However, their utilization can cause several drawbacks: (i) High background interferences below $m/z = 1000$ (not *MALDI silent*), which is disadvantageous specifically for LMWC analytics; (ii) low vacuum stability, which is especially problematic for standard instruments operated under *high vacuum* (HV); (iii) challenging homogeneous thin-layer coating, potentially causing inconsistent measurement conditions; and (iv) usually no suitability for dual polarity mode experiments, i.e., carrying out positive and negative mode measurements with the same matrix.

Polymeric materials are promising candidates for *MALDI silent* matrices, as the large variety of possible molecular layouts potentially allows to meet all prerequisites of a MALDI matrix: (a) Sufficient *ultra-violet* (UV) laser radiation absorption, implemented by introducing conjugated π -electron systems in the polymer backbone or side chains; (b) high ionization efficiency, enhanced by adding acidic and/or basic functional groups to the polymer's molecular structure, potentially also allowing dual polarity mode measurements; (c) *MALDI silence*, enabled by the high molar mass of the polymer chains; (d) high vacuum stability, also granted by the polymer's molar mass; and (e) homogeneous thin-films, achieved by multiple available coating methods. Yet, despite their high potential only a handful of polymeric matrices were reported in literature and so far, investigations to develop conscious design strategies are missing.

The target of this thesis is to contribute to the field of *MALDI silent* matrices by developing and investigating different polymers as macromolecular MALDI MS and MSI

matrices for LMWC analytics. Therefore, two different strategies were explored: (i) Investigating conjugated polymers, and (ii) polymerizing SOMs. For the first strategy, five conjugated polymers were tested as MALDI matrices for the detection of various LMWCs. Among these, four were found to be excellent matrices, with sufficient ionization efficiencies and rare dual polarity mode suitability and allowed LMWC detection with low background interferences (*MALDI silent*). A high crystallinity of the matrix (SOM) is reported to be crucial to ensure successful measurements, yet *conjugated polymer matrices* (CPMs) are semi-crystalline, i.e., they contain crystalline and amorphous domains. Hence, the analytes are expected to be incorporated in the crystalline domains of the CPMs, depending on their degree of crystallization. Therefore, two amorphous CPMs were synthesized and tested, showing similar matrix performances (e.g., ionization efficiencies, dual polarity mode, *MALDI silence*) as a structurally related semi-crystalline CPM. This indicates that the analytes are incorporated in the amorphous parts of the CPM. The second strategy towards *polymeric matrices* (PMs) is the polymerization of standard SOMs. As the matrix performance of the corresponding SOMs is known, the performance of the respective *polymerized SOMs* (P(SOMs)) can be validated against this benchmark. At the same time, polymerization can induce the properties needed to enable efficient LMWC analytics. Therefore, two standard SOMs were modified and polymerized, resulting in P(SOMs), which were vacuum stable and *MALDI silent*, and showed similar optical properties, analyte scopes and ionization efficiencies in benchmark tests with their respective SOMs.

For the fast and facile comparison of the matrix performances of PMs and standard matrices, the graphing software OriginPro was used to visualize, process, and evaluate the acquired mass spectra. To automatize these tasks, a script was programmed using the OriginPro-native programming languages LabTalk and OriginC: X Functions.

Dedication

I dedicate this thesis
to the insatiable curiosity,
deeply enrooted
in our human nature.
May our endless search
for knowledge and explanations
lead us to a better place.

Acknowledgements

First, I want to thank the best Ph.D. mum in the world, Dr. Franziska Lissel, enabling me to work in her research group *Functional Electronic Materials* (FEM). She always supported me with all her heart, helping me through good and bad times and gave me hope in my darkest hours. Thanks to her, I did not lose my mind during my Ph.D.

Special thanks goes to Prof. Dr. Brigitte Voit for allowing me to work at the Leibniz Institute for Polymer Research Dresden. Additionally, I appreciate that she was always very kind to me and helped me whenever I needed advice.

I also want to thank Kristina Ditte for always helping and supporting me, giving advice, and proofreading my texts.

Marco Giampà, I want to thank for a very productive cooperation resulting in two publications. I enjoyed our discussions and I really appreciate all the work he did to lead our research to success.

My dear MALDI mentor, Dr. Karin Sahre, I want to thank for teaching me everything about sample preparation, spectra measuring and evaluation.

My friends Sarah Straub and Dr. Alexander Sasinska, I want to thank for proofreading this thesis and for giving me a lot of input to improve it.

My colleagues Dr. Laura Plüschke and Dr. Martin Geisler, I want to thank for sharing their thesis with me as benchmark and helping me by answering many questions.

And last but not least, I want to thank all my family, friends, and colleagues, who supported me through the tough times and celebrated with me the good times.

Table of Content

Abstract	IV
Dedication.....	VI
Acknowledgements	VII
Table of Content	VIII
List of Figures and Tables	X
List of Symbols	XIV
List of Abbreviations	XV
Research Outputs.....	XVII
1. Motivation.....	1
2. Theory.....	4
2.1. MALDI MS and MS Imaging	4
2.2. Polymers as MALDI Matrix Systems	14
2.3. Software for MALDI MS Data Handling	18
3. Research Objectives and Aim of Thesis.....	20
4. Results & Discussion.....	21
4.1. Automation of Data Handling with OriginPro 2020b	21
4.2. Conjugated Polymers as a New Class of Dual-Mode Matrices for MALDI Mass Spectrometry and Imaging (Publication 1)	34
4.3. Amorphous Conjugated Polymers as Efficient Dual-Mode MALDI Matrices for Low-Molecular-Weight Analytes (Publication 2)	38
4.4. Polymerization as a Strategy to Improve Small Organic Matrices for Low Molecular Weight Compound Analytics with MALDI MS and MALDI MS Imaging (Publication 3).....	42
5. Conclusion & Outlook.....	47
6. List of References	50
7. Publications.....	58
7.1. Conjugated Polymers as a New Class of Dual-Mode Matrices for MALDI Mass Spectrometry and Imaging (Publication 1)	58
7.2. Amorphous Conjugated Polymers as Efficient Dual-Mode MALDI Matrices for Low-Molecular-Weight Analytes (Publication 2)	67
7.3. Polymerization as a Strategy to Improve Small Organic Matrices for Low Molecular Weight Compound Analytics with MALDI MS and MALDI MS Imaging (Publication 3).....	77

Table of Content

8. Appendix	88
8.1. LabTalk script (OriginPro 2020b).....	88
8.2. Supporting Information	152
8.3. Copyright & Permission	250
Declaration / Eidesstattliche Erklärung	252

List of Figures and Tables

Figure 1: Illustration of a MALDI MS experiment: (i) sample preparation (left), (ii) MALDI process (middle) and (iii) ion separation and detection, here by reflectron ToF - mass analyzer (right).	4
Figure 2: Illustration of the procedure to generate MALDI MSI heat maps.	7
Figure 3: Molecular structure of several commonly used SOMs: 2,5-DHB, 9AA, α -CHCA, harmine, and DCTB.	10
Figure 4: Exemplary Gaussian (red) and Flory-Schulz (black) distributions (redrawn from [1]).	14
Figure 5: Molecular structure of sol-gel poly(siloxane)/DHB. ^[2]	17
Figure 6: Molecular structure of oligomeric matrices Qthio, Sthio, 3DIO, and 4DIO. ^[4]	17
Figure 7: Molecular structure of P3OT. ^[3]	17
Figure 8: Dialogue to insert information about the measurement parameters of a 3-stacked plot.	22
Figure 9: Example of a 3-stacked graph showing three individual plots. The plots show MALDI MS spectra of a complex biological sample (red), a polymer (blue) and a single LMWC (black). In contrast to the Y-axis, the X-axis of the plots are linked to each other.	23
Figure 10: Exemplary smooth command used for smoothing MALDI MS spectra in this thesis (including descriptions of subfunctions).	24
Figure 11: Example of smoothing a MALDI MS spectrum of a polymer sample. Top: original spectrum (blue); bottom: smoothed spectrum (black). The spectra are divided into three parts showing different m/z ranges.	25
Figure 12: Manual baseline correction of a MALDI MS polymer spectrum showing the original spectrum (black), the manually defined baseline (red) and baseline corrected spectrum (blue). The baseline consists of 52 data points.	26

Figure 13: Exemplary original spectrum from Figure 14 with left peak markers (red), right peak markers (green), and the corresponding local maxima (blue).	26
Figure 14: Automatic baseline correction of a MALDI MS polymer spectrum showing the original spectrum (black), the automatically defined baseline (red) and the baseline corrected spectrum (blue). The baseline consists of 322 data points.	27
Figure 15: Exemplary manual baseline correction of a MALDI MS spectrum of a glycodendrimer, showing very broad signals.	28
Figure 16: Smoothed, baseline corrected, and peak picked MALDI MS polymer spectrum showing two signal series for the calculation of the polymeric properties.	30
Figure 17: Exemplary headlines (yellow) of a mass spectrum data set in OriginPro. The comment field of the Y-axis values is marked.	32
Figure 18: Dialogue to input information of the parts to highlight (here: two parts).	32
Figure 19: Molecular structures of a) PNDI(T2), b) P3DDT, c) P9OFI, d) PTQ1, and e) PII(T2). ^[5]	34
Figure 20: Reflectron positive mode MALDI MS spectra of the sample plate (green), the pure analyte RP (red), the pure matrix PNDI(T2) (blue), and PNDI(T2) coated on RP (black). Peaks: $m/z = 39 [K]^+$; 395 $[RP3]^+$ (see SI); 607 $[RP-H]^+$. ^[5]	35
Figure 21: Reflectron negative mode MALDI MS spectra of the sample plate (green), the pure analyte CiA (red), the pure matrix PNDI(T2) (blue), and PNDI(T2) coated on CiA (black). Peaks: $m/z = 42 [C_3H_6]^-$; 87 $[CiA1=C_3H_3O_3]^-$; 191 $[CiA-H]^+$. ^[5]	36
Figure 22: MALDI Imaging heat maps of a rat brain visualizing the spatial distribution of eight different molecules. ^[5]	37
Figure 23: Synthetic routes for the NDI(T2)/fluorene co-polymers P(TNDIT-FI(C₄C₂)) (red) and P(TNDIT-FI(C₁₀C₈)) (green). The alkyl side-chains R1 and R2 are defined in this scheme. ^[6]	38

Figure 24: GIWAXS curves of P(TNDIT-FI(C₄C₂)) (drop-coated: red; spin-coated: green) and P(TNDIT-FI(C₁₀C₈)) (drop-coated: blue; spin-coated: pink). ^[6]	39
Figure 25: Reflectron positive mode MALDI MS spectra of spin coated samples: PNDI(T2) coated on RP (red), P(TNDIT-FI(C₄C₂)) coated on RP (blue) and P(TNDIT-FI(C₁₀C₈)) coated on RP (black). Peaks: m/z = 395 [RP3] ⁺ (see SI); 397 [RP1=C ₂₃ H ₂₉ N ₂ O ₄] ⁺ ; 607 [RP-H] ⁺ . ^[6]	40
Figure 26: Reflectron negative mode MALDI MS spectra of spin-coated samples: PNDI(T2) coated on CiA (red), P(TNDIT-FI(C₄C₂)) coated on CiA (blue) and P(TNDIT-FI(C₁₀C₈)) coated on CiA (black). Peaks: m/z = 26 acetonitrile fragment [CN] ⁻ ; 42 [C ₃ H ₆] ⁻ ; 57 [C ₄ H ₈] ⁻ ; 85 [C ₆ H ₁₂] ⁻ ; 87 [CiA1=C ₃ H ₃ O ₃] ⁻ ; 111 (unknown); 129 [CiA2=C ₅ H ₅ O ₄] ⁻ ; 155 (unknown); 191 [CiA-H] ⁻ . ^[6]	40
Figure 27: Synthesis of poly(vinyl dihydroxybenzoic acid) (P(VDHB)). (a) n-BuLi, DMF, Et ₂ O, 1h, 0 °C-rt (b) KMnO ₄ , acetone/H ₂ O, 24h, rt (c) H ₂ SO ₄ , methanol, 24h, 80 °C (d) Pd(PPh ₃) ₄ , Na ₂ CO ₃ , dioxane/H ₂ O, 72h, 100 °C (e) KOH, acetone/H ₂ O, 24h, 100 °C (f) AIBN, dioxane, 7h, 90 °C (g) HBr, AcOH, 5h, 120 °C. ^[7]	42
Figure 28: Synthesis of poly(vinyl harmine) (P(VHar)). (a) NBS, DCM, 24h, rt (b) Pd(PPh ₃) ₄ , Na ₂ CO ₃ , dioxane/H ₂ O, 24h, 100 °C (c) AIBN, dioxane, 7h, 90 °C. ^[7]	43
Figure 29: Vacuum stability under HV (~10 ⁻⁷ mbar) over 24 h (average of 3 measurements, straight plots) and 72 h (single measurements, dashed plots) of DHB (yellow), P(VDHB) (red), harmine (blue) and P(VHar) (black). The balance error of 0.01 mg is visualized relative to the sample weight as error bars. ^[7]	43

Figure 30: Reflectron positive mode MALDI MS spectra of the sample plate (violet), the pure analyte MA (yellow), the pure DHB (green), DHB coated on MA (red), the pure P(VDHB) (blue), and P(VDHB) coated on MA (black). Peaks: $m/z = 23$ $[\text{Na}]^+$; 39 $[\text{K}]^+$; 57 $[\text{MA1}=\text{C}_4\text{H}_9]^+$; 88 $[\text{MA2}=\text{C}_4\text{H}_8\text{O}_2]^+$; 137 $[\text{DHB-OH}]^+$; 154 $[\text{DHB}]^{\bullet+}$; 177 $[\text{DHB}+\text{Na}]^+$; 193 $[\text{DHB}+\text{K}]^+$; 215 [unknown DHB signal] $^+$; 251 $[\text{MA}+\text{Na}]^+$; 267 $[\text{MA}+\text{K}]^+$. Laser intensity: 30%. ^[7]	44
Figure 31: Reflectron negative mode MALDI MS spectra of the sample plate (violet), the pure analyte FT (yellow), the pure harmine (green), harmine coated on FT (red), the pure P(VHar) (blue), and P(VHar) coated on FT (black). Peaks: $m/z = 35$ $[\text{Cl}]^-$; 71 $[\text{FT1}=\text{C}_3\text{H}_3\text{O}_2]^-$; 101 $[\text{FT2}=\text{C}_4\text{H}_5\text{O}_3]^-$; 119 $[\text{FT3}=\text{C}_4\text{H}_7\text{O}_4]^-$; 130 $[\text{FT4}=\text{C}_6\text{H}_{10}\text{O}_3]^-$; 179 $[\text{FT-H}]^-$; 211 $[\text{Har-H}]^-$. Laser intensity: 30%. ^[7]	44
Figure 32: Positive mode MALDI MSI measurements of a cryosection from a patient-derived breast cancer xenograft tumor using P(VDHB) as matrix. Top (from left to right): optical picture, ionic intensity heat maps of ions at m/z values 114.0, 147.0, 666.5, 762.4. Bottom (from left to right): ionic intensity heat maps of the ions at m/z values 794.4, 798.5, 874.4, 920.5. ^[7]	45
Figure 33: Negative mode MALDI MSI measurements of a cryosection from a patient-derived breast cancer xenograft tumor using P(VHar) as matrix. Top (from left to right) optical pictures, ionic intensity maps of ions at m/z values 87.0, 89.1, 124.0, 145.1. Bottom (from left to right) ionic intensity heat maps of the ions at m/z values 283.2, 303.2, 835.4, 887.5. ^[7]	45
Table 1: Polymeric properties calculated for the first (red) and second (blue) signal series.	31

List of Symbols

$\bar{M}_w$	Polydispersity index
Da	Dalton
h	Hour
kg	Kilogram
m ²	Square meter
mg	Milligram
mbar	Millibar
M_i	Molar mass of the i^{th} polymer chain
M_n	Number average molar mass
M_{RU}	Average molar mass of the repeating unit
M_w	Weight average molar mass
m/z	Mass-to-charge ratio
N_i	Number of the i^{th} polymer chain
nm	Nanometer

List of Abbreviations

2D	Two-dimensional
2,5-DHB	2,5-Dihydroxybenzoic acid
3DIO	Tetrahydro-hexaoxa-pentacene
4DIO	Tetrahydro-octaoxa-heptacene
9AA	9-Aminoacridine
α -CHCA	α -Cyano-4-hydroxycinnamic acid
ChA	Cholic acid
CiA	Citric acid
CPM	Conjugated polymer matrix
DCTB	<i>Trans</i> -2-[3-(4- <i>tert</i> -butylphenyl)-2-methyl-2-propenylidene]malono-nitrile
Dithranol	Anthracenetriol
DP	Degree of polymerization
ESI	Electrospray ionization
FAB	Fast atom bombardment
FT	Fructose
FT-ICR	Fourier transform ion cyclotron resonance
FTIR	Fourier transform infrared spectroscopy
GPC	Gel permeation chromatography
Harmine	7-Methoxy-1-methyl-9H-pyrido[3,4-b]indole
HMWC	High molecular weight compound
IR	Infrared (spectroscopy)
ITO	Indium-tin-oxide
L.	Line
Lin.	Linear (detection mode)
LMWC	Low molecular weight compound
LoD	Limit of detection
MA	Myristic acid
MAC	Mass attenuation coefficient
MALDI MS	Matrix assisted desorption/ionization mass spectrometry
MALDI MSI	Matrix assisted desorption/ionization mass spectrometry imaging
MALDI silent	MALDI measurements without or with low background interferences (matrix signals) in the low mass range ($m/z < 1000$)

List of Abbreviations

MOF	Metal organic framework
N ₂	Nitrogen
NDI(T2)	Naphthalene diimide
Nd:YAG	Neodymium : yttrium-aluminum-garnet
NMR	Nuclear magnetic resonance spectroscopy
PALDI	Polymer assisted laser desorption/ionization
P3DDT	Poly(3-dodecylthiophene-2,5-diyl)
P3OT	Poly(3-octylthiophene-2,5-diyl)
P9OFI	Poly(9,9-di- <i>n</i> -octylfluorenyl-2,7-diyl)
PII(T2)	Poly{[<i>N,N'</i> -bis(2-octyldodecyl)-isoindigo-5,5'-diyl]- <i>alt</i> -5,5'-(2,2'-bithiophene)}
PM	Polymer matrix
PNDI(T2)	Poly{[<i>N,N'</i> -bis(2-octyldodecyl)-naphthalene-1,4,5,8-bis(dicarboximide)-2,5-diyl]- <i>alt</i> -5,5'-(2,2'-bithiophene)}
P(SOM)	Polymerized small organic matrix (SOM)
PTQ1	Poly{[2,3-bis(3-octyloxyphenyl)quinoxaline-5,8-diyl]- <i>alt</i> -(thiophene-2,5-diyl)}
P(VDHB)	Poly(vinyl dihydroxybenzoic acid)
P(VHar)	Poly(vinyl harmine)
Qthio	Quaterthiophene
Refl.	Reflectron (detection mode)
RP	Reserpine
RU	Repeating unit
S/N	Signal-to-noise (ratio)
SI	Supporting information
SIMS	Secondary ion mass spectrometry
SOM	Small organic matrix
Sthio	Sexithiophene
ToF	Time-of-flight
UV	Ultra-violet

Research Outputs

First Author Publications:

1. Conjugated Polymers as a New Class of Dual-Mode Matrices for MALDI Mass Spectrometry and Imaging

K. Horatz, M. Giampà, Y. Karpov, K. Sahre, H. Bednarz, A. Kiriy, B. Voit, K. Niehaus, N. Hadjichristidis, D. L. Michels, F. Lissel, *J. Am. Chem. Soc.* **2018**, *140*, 11416–11423.

The author contributed to the design of the study, the writing process of the manuscript as well as the SI, and performed all analytical experiments (excl. MALDI MSI).

2. Amorphous Conjugated Polymers as Efficient Dual-Mode MALDI Matrices for Low-Molecular-Weight Analytes

K. Horatz[§], K. Ditte[§], T. Prenveille, K.-N. Zhang, D. Jehnichen, A. Kiriy, B. Voit, F. Lissel, *ChemPlusChem* **2019**, *84*, 1338 – 1345.

[§] These authors contributed equally to this work.

The author contributed to the design of the study, the writing process of the manuscript as well as the SI, and performed all MALDI MS experiments.

3. Polymerization as a Strategy to Improve Small Organic Matrices for Low Molecular Weight Compound Analytics with MALDI MS and MALDI MS Imaging

K. Horatz, M. Giampà, Z. Qiao, S. A. Moestue, F. Lissel, *ACS Appl. Polym. Mater.* **2021**, under Revis.

The author contributed to the design of the study, wrote the manuscript and SI, performed all optical and vacuum stability measurements and most of the MALDI MS experiments (excl. MALDI MSI).

Co-Author Publications:

1. Synthesis and Characterization of a Semiconducting and Solution-processable Ruthenium-based Polymetallayne

P.-Y. Ho, H. Komber, K. Horatz, T. Tsuda, S. Mannsfeld, E. Dmitrieva, O. Blacque, U. Kraft, H. Sirringhaus, F. Lissel, *Polym. Chem.* **2020**, 11, 472.

The author performed and evaluated all MALDI MS experiments.

2. Synthesis and Aggregation Behavior of a Glycolated Naphthalene Diimide Bithiophene Copolymer for Application in Low-Level n-Doped Organic Thermoelectrics

Y.-h. Shin, H. Komber, D. Caiola, M. Cassinelli, H. Sun, D. Stegerer, M. Schreiter, K. Horatz, F. Lissel, X. Jiao, C. R. McNeill, S. Cimò, C. Bertarelli, S. Fabiano, M. Caironi, M. Sommer, *Macromolecules* **2020**, 53, 13, 5158-5168.

The author performed and evaluated all MALDI MS experiments.

1. Motivation

Matrix assisted laser desorption/ionization mass spectrometry (MALDI MS) was developed in the 1980s for the detection of *high molecular weight compounds* (HMWC, $M > 1000$ Da), and allows the analysis of biopolymers (e.g., proteins, peptides, polysaccharides) and synthetic polymers.^[8-10] On basis of this MS technique, the visualization method MALDI MS *Imaging* (MSI) was developed in the 1990s to trace the spatial distribution and concentration of molecules in biological tissues.^[11] This information of selected molecules is especially helpful for medical and pathological research, e.g., to investigate the metabolism of drugs or the progress of a disease.^[12]

In contrast to other MS techniques, like *secondary ion mass spectrometry* (SIMS), *electrospray ionization* (ESI), *fast atom bombardment* (FAB), the analytes are incorporated in an excess of a matrix during the MALDI measurement process. This matrix protects the analyte molecules from the high energy of the laser shots, separates them from each other to avoid interactions (e.g., analyte cluster formation) and supports their ionization by absorbing and transferring the laser radiation energy.^[13-14] This indirect ionization of the analyte is also called soft ionization and prevents the analyzed compounds from fragmentation as well as degradation. Additionally, the indirect ionization promotes predominant formation of single charged ions, which is useful for the determination of the analytes' molar mass from their *mass-to-charge* (m/z) ratios during evaluation of the mass spectra.^[15-16]

Apart from HMWC analytics, there is a growing interest to use the MALDI technique also for the reliable and reproducible detection of *low molecular weight compounds* (LMWC), i.e., compounds with molar masses below 1000 Da. One reason for this trend is the developing awareness of the importance to visualize interesting compounds of lipidomics and metabolomics for medical research via MALDI MSI, as the targeted molecules in these research fields are often small.^[17-18] Standard matrices used for MALDI analytics are *small organic matrices* (SOMs; $M < 500$ Da), such as *2,5-dihydroxybenzoic acid* (2,5-DHB) or *9-aminoacridine* (9AA), which meet the fundamental prerequisites for MALDI matrices, i.e., (i) a conjugated π -electron system (e.g., benzene rings) for absorption of laser radiation in the *ultra-violet* (UV) range; (ii) crystalline morphology for sufficient analyte incorporation and (iii) Brønsted or Lewis acidic/basic functionalities to increase the ionization efficiency.^[14] Furthermore, the matrices ideally fulfill four additional prerequisites, e.g., to enhance spot-to-spot reproducibility of

MALDI MSI measurements: (iv) No background interferences in the low mass range (*MALDI silent*; $m/z < 1000$, irrelevant for HMWC analytics); (v) stability in *high vacuum* (HV; $\sim 10^{-7}$ mbar); (vi) uniform thin-layer coating and (vii) suitability in positive as well as negative polarity mode (dual polarity mode).^[19]

However, many SOMs fail regarding the additional prerequisites (iv) - (vii): Standard SOMs can produce numerous interferences in the low mass range, such as molecule ions, fragment ions and cluster ions, which can overlap with LMWC signals. This can potentially impede the evaluation of the corresponding mass spectra of LMWCs. Due to their low molar mass, most SOMs tend to sublime in HV during long-term MALDI MSI measurements, which often reduces reproducibility and reliability, as a spot measured at the end of a long experiment has a lower matrix coverage compared to a spot measured at the beginning.^[14] This spot-to-spot reproducibility is also influenced by the uniform coating of matrix layers on the tissue sample since the size and distribution of the matrix crystals affect the ionization efficiency and spatial resolution of MALDI MSI measurements.^[20] Although several promising approaches, like electrospray coating, were developed, the crystallization process of standard SOMs remains difficult to control. This process depends on various parameters, such as solvent, temperature and surface, and obtaining a uniform crystal layer often requires extensive experience.^[21] A further setback of standard SOMs is that they often contain either Brønsted acidic or basic groups to support ionization via ion transfer. Such groups increase the ionization efficiency in either positive or negative measurements, while concomitantly reducing the ionization efficiency in the other. Dual polarity mode matrices, i.e., matrices supporting the ionization in both ion modes, would be preferable for unique samples, such as biological tissue slices or when only very small amounts of sample are available (e.g., fine needle biopsies).^[22-23]

Different research efforts focus on improving standard SOMs specifically for LMWC analytics, e.g., to reduce their background interferences^[24] or to increase their vacuum stability^[25]. Another approach is to develop matrices with enhanced properties, such as *MALDI silence*, high vacuum stability, dual polarity mode suitability or a specific analyte scope. These can be organic molecules, like SOMs,^[26] porphyrins^[27], liquid crystals^[28], oligomers^[4] and polymers^[3] but also inorganic materials, such as metal oxides^[29], graphene (-oxide),^[30-31] and graphite^[32].

Especially the application of polymers as MALDI matrices for LMWC analytics is highly promising, as their chemical and physical properties strongly depends on their

molecular structure and hence can be adjusted to meet the fundamental and additional matrix requirements (e.g., conjugated π -electron system for UV laser radiation absorption or high molar mass for *MALDI silence* and vacuum stability). Thus, Soltzberg et al. (2004) and Lin et al. (2002) presented promising strategies for *polymer matrices* (PMs): Polymers with conjugated backbones (here: poly(3-octylthiophene-2,5-diyl)) and SOMs (here: 2,5-dihydroxybenzoic acid) incorporated into a polysiloxane network.^[2-3] Both works show that PMs are capable of *MALDI silent* measurements and highly promising for further investigations. This thesis aims to explore strategies to develop PMs, and to further understand their properties and potential for reliable and reproducible measurements of LMWCs with MALDI MS and MALDI MSI.

2. Theory

2.1. MALDI MS and MS Imaging

2.1.1. Procedures of MALDI MS and MS Imaging

MALDI MS was developed in the 1980s for the detection and analysis of biological and non-biological HMWCs, such as proteins, peptides, and synthetic polymers. It emerged as an important analytical tool, e.g., in biochemistry and material sciences.^[8-9, 33-36] By using a protective matrix, MALDI MS allows a mass spectrometric analysis of HMWCs without causing their fragmentation or degradation.^[37] Additionally, it does not require analyte labeling and is compatible with buffers and impurities.^[12] Moreover, its high sensitivity allows measurements in the range of attomolar ($10^{-18} \frac{\text{mol}}{\text{l}}$) analyte concentrations.^[12, 38] Generally speaking, a MALDI MS measurement procedure consists of three parts (Figure 1):^[13] (i) Sample preparation (left), (ii) MALDI process (irradiation of sample, desorption, ionization) (middle) and (iii) separation/detection (right).

Figure 1: Illustration of a MALDI MS experiment: (i) sample preparation (left), (ii) MALDI process (middle) and (iii) ion separation and detection, here by reflectron ToF - mass analyzer (right).

The sample preparation encompasses a potential pretreatment of the investigated material (e.g., the extraction of a sample for MALDI MS), as well as the application of the matrix. The homogeneous incorporation of analyte molecules into the crystal lattice of the matrix is crucial for MALDI MS and for optimal measurement results.^[39] A homogeneous matrix/analyte mixture is required for a homogeneous energy deposition and material ablation and influences ion yields. As such, a homogenous mixture is important for so-called spot-to-spot reproducibility. Furthermore, analyte molecules incorporated in an excess of matrix rarely interact with each other, preventing analyte-

analyte cluster formation. This cluster formation can lead to redundant signals, which cannot be distinguished from molecule ions of other analytes, as MALDI is not suitable to formally identify the origin of a signal. Hence, these additional analyte cluster signals can significantly impede the evaluation of the spectra.^[14] Instead of reacting with each other, the excess of matrix ensures that the analytes will rather interact with the matrix molecules, supporting the ionization process (vide infra, next passage).^[13-14] The matrix can be applied from solution, in a so-called wet preparation, or without solvent, i.e., dry. Wet methods are, e.g., dried-droplet^[8], electrospray deposition^[21, 40], or spin-coating^[41], while dry methods encompass, e.g., a mixing of solid matrix and analyte with a ball mill^[13], or the sublimation^[42] of the matrix.^[20] Regarding MALDI MS, the most commonly used method is the dried-droplet method, which can either be performed by mixing or layering the analytes and matrices. For mixing, the analyte solution is mixed with an excess of a matrix solution and is applied dropwise on the sample plate (Figure 1 (i)).^[20, 37] For the layering method, the analyte and matrix are coated subsequently without prior mixing. The order is interchangeable, some sequences are analyte-matrix (normal)^[43], matrix-analyte (reverse)^[44-45] or matrix-analyte-matrix (sandwich)^[46] (Figure 1 (ii)).^[20, 47] Both mixing and layering can be performed manually or automatically.^[48] After the sample preparation, which takes place outside the MALDI MS instrument, the sample is transferred into the instrument, and the MALDI process is then started (Figure 1 (ii)). First, the sample is irradiated by multiple laser shots of variable frequency and intensity. Commonly used MALDI lasers are N₂ (*nitrogen*, $\lambda = 337$ nm) or Nd:YAG (*neodymium : yttrium-aluminum-garnet*, $\lambda = 355$ nm) lasers, which emit UV radiation, as well as *infrared* (IR) lasers such as Cr:LiSAF (*Cr³⁺:LiSrAlF₆*, $\lambda = 2.88$ -3.4 μ m).^[10, 49] Upon absorption of the radiation, the matrix molecules are excited, and the strong laser shots on the sample lead to the desorption of matrix and analyte molecules; the matrix/analyte blend ablates, forming a MALDI plume above the sample.^[50] The ionization mechanism is not yet completely understood, but three general pathways are frequently discussed in the literature: (a) Interionic ionization, (b) lucky survivor model, and (c) ionization by electron transfer. In interionic ionization (a), the analyte is presumably ionized by ion transfer or cluster formation within the MALDI plume, e.g., due to the addition of alkali ions, such as sodium [Na]⁺ or potassium [K]⁺.^[51] A special case of ion transfer is the proton transfer (gas phase protonation model).^[52] Proton transfer is supported by Brønsted acidic/basic functional groups of the matrix, and the excited states, formed by the absorption of the laser radiation, can further

influence the corresponding pK_a and pK_b values.^[50, 53] The lucky survivor model (b) postulates that analytes are incorporated in the matrix with a respective charge and that some of these charged species survive after charge exchange in the MALDI plume.^[52] Ionization by electron transfer (c) assumes that so-called charge transfer matrices act as Lewis acids/bases to create charged analyte radicals.^[51, 54-55] Another not fully understood phenomenon of matrix-assisted ionization is the predominant production of single charged ions.^[15, 52] This means that the m/z value of the detected signal is equal to the molar mass of these ions. As MALDI MS is not suitable to formally identify origins of the detected signals, single and multi-charged ions are not distinguishable in the mass spectra. Hence, the predominant formation of single charged analyte ions is important in order to give reliable information about the molar mass of the ion from the m/z value of the signal.^[16] The molar mass ($\triangleq m/z$ value) of the ion can, e.g., be compared with databases (e.g., www.hmdb.ca) to make suggestions about the originating molecule.

In the third and last part of the MALDI MS measurement procedure, i.e., ion separation and detection, the ionized matrix and analyte molecules travel to a mass analyzer, which can be *time-of-flight* (ToF), *fourier-transform ion cyclotron resonance* (FT-ICR), quadrupole mass filter or ion trap.^[56] The commonly used ToF-mass analyzer (Figure 1 (iii)) separates the ions according to their masses and charges: For equally charged ions, the heavier ones are slower accelerated in the electromagnetic field and hence take longer to reach the detector. Then, the travel time of the ions from the source to the detector is converted into the mass of the ion in relation to its charge state (m/z ratio). Two different ToF-modes are available: linear and reflectron. In the linear mode the ions travel directly from source to detector, whereas in reflectron mode the ions are deflected by strong electromagnetic fields. This deflection lengthens the flight path and can potentially lead to the fragmentation of larger or unstable molecules, yet the longer flight path also increases the resolution of the resulting mass spectra.^[57] In these spectra, the m/z value of a signal is plotted against its intensity, which has an arbitrary unit and is proportional to the number of detected molecules with equal m/z values.

MALDI MS Imaging (MSI) was developed in the 1990s on basis of MALDI MS and allows the visualization of the spatial distribution and concentration of molecules in biological tissues.^[11] Since then, this imaging technique emerged to an important analytical tool for pathological and medical research, as there is no need for sample

labeling or high energy radiation.^[12, 38, 58] The MALDI MSI measurement procedure can also be divided into three parts, which are illustrated in Figure 2.

Figure 2: Illustration of the procedure to generate MALDI MSI heat maps.

For the sample preparation (Figure 2 (i)), a thin slice of biological tissue is prepared and pretreated using several steps (e.g., stabilizing, cryo-sectioning, washing) and mounted on an *indium-tin-oxide* (ITO) coated glass slide.^[59-60] For the matrix coating, the mixing method is not an option in this case, so the matrix is applied on the tissue using a layering method, such as spin coating^[41], sublimation^[42], automatic/manual spray^[21, 61-62] or drop^[39, 63] coating.^[12] Next, a series of MALDI MS measurements, using the aforementioned MALDI process and ion separation/detection methods, is performed in a raster-oriented grid pattern (Figure 2 (ii)), leading to a high number of individual measurements. E.g., when a 256×256 spot array is used, the experiment results in $256^2 = 65536$ individual mass spectra.^[64] Each mass spectrum reflects the chemical composition of the respective spot on the tissue slice and represents a pixel of the later generated image. The combination of all individual spectra allows to display the relative signal intensities as 2D-heat map images (Figure 2 (iii)). For every detected ion signal a corresponding heat map image can be generated, visualizing the relative concentration and spatial distribution of targeted ion within the tissue sample. The spatial resolution of these images is proportional to the grid size, i.e., the amount and density of the measured spectra, resulting in increased measurement time when aiming for a high spatial resolution.

2.1.2. Applications of MALDI MS and MALDI MS Imaging

Due to the soft ionization and protection of the analyte by the matrix, MALDI was the first MS technique capable of measuring intact HMWC, like proteins and polysaccharides.^[8, 37] In recent years also the detection of LMWCs with high sensitivity and accuracy came into focus of research, as MALDI MS requires only small amounts of sample and is a fast method, enabling, e.g., label-free high throughput screening.^[65]

However, especially the visualization of the spatial distribution of LMWCs in biological tissues via MALDI MSI is of high interest for many different kinds of biochemical investigations, like lipidomics and metabolomics.^[14] MALDI MSI can give detailed information about the distribution of metabolites of interest in different organs, such as eyes, lungs and bones, contributing to basic research on biological processes.^[17, 66] Furthermore, the imaging of small metabolites can help to understand disease progression and aid drug development as shown in context of illness or injury, e.g., cerebral ischemia- or stroke-reperfusion injury.^[18, 67] Here, not only the drug metabolism and distribution, but also drug quantification is a highly interesting topic in MALDI MSI research, which are all excellently reviewed in literature.^[68] Current research focuses also on improving the measurement procedure of MALDI MSI (chapter 2.1.1) to increase the reliability and reproducibility of the resulting heat maps. The detection sensitivity, e.g., can be improved by an optimized on-tissue pretreatment protocol or chemical derivatization, and the spatial resolution can be increased by combining salt doping and matrix sublimation.^[69-71] The spectral resolution and thus the quality of the generated heat maps can also be optimized by developing of the instrumentation.^[68]

2.1.3. MALDI Matrix Systems

During the MALDI MS process including irradiation, desorption, and ionization (chapter 2.1.1, Figure 1 (ii)), the matrix is responsible for the separation, protection, and ionization of the analyte. The separation of the analyte molecules is required to avoid analyte-analyte interactions, such as cluster formation (chapter 2.1.1). The protection of the analyte from strong laser shots is important to prevent its fragmentation and degradation. To guarantee separation and protection, the matrix is usually applied in excess compared to the analyte (ca. $\times 10^2$ - $\times 10^7$).^[13] The matrix also supports the ionization of the analytes, e.g., by ion or charge transfer (chapter 2.1.1). Therefore, MALDI matrices need to meet three fundamental prerequisites:^[57]

1. Absorption of laser radiation: A strong absorption at the wavelength of the MALDI laser is crucial for an efficient excitation of the matrix, which in turn supports the analyte's ionization.^[13] Standard MALDI lasers emit radiation either in the UV range (e.g., 355 nm) or IR range (e.g., 2.88 μm). For a sufficient absorption of the radiation, the molecular structure of the matrix requires either conjugated π -electron systems, such as benzene rings (UV), or small functional groups, such as alcohol or amino groups (IR).^[14]
 2. Crystalline morphology: A crystalline morphology of the SOM was found to be important to ensure an efficient analyte incorporation in its crystal lattice.^[72] The resulting separation of analyte molecules and increased interactions between matrix and analyte (e.g., ion / charge transfer), support the analyte ionization.^[13] Furthermore, a homogenous matrix crystal distribution is important for an enhanced spot-to-spot reproducibility.^[20]
 3. Acidic/basic functional groups: Although the MALDI ionization mechanism is not completely understood in detail, the proton transfer between matrix and analyte, as described in the gas phase protonation model (chapter 2.1.1), is assumed to be the dominant ionization mechanism.^[50] Hence, Brønsted acidic or basic functional groups, as proton donors or acceptors, are important for an enhanced ionization efficiency. For charge transfer matrices (Lewis acid/base), which ionize the analytes via electron transfer, the availability of Brønsted acidic/basic groups is not required.
- SOMs used as standard matrices fulfill all three fundamental requirements. Some well-studied proton transfer SOMs are, e.g., *2,5-dihydroxybenzoic acid* (2,5-DHB)^[39], *9-aminoacridine* (9AA)^[73], *α -cyano-4-hydroxycinnamic acid* (α -CHCA)^[74] and *7-methoxy-1-methyl-9H-pyrido[3,4-*b*]indole* (harmine)^[75], while *trans-2-[3-(4-tert-Butylphenyl)-2-*

methyl-2-propenyliden]-malonsäuredinitril (DCTB)^[76] is an example for a charge transfer SOM (Figure 3).

Figure 3: Molecular structure of several commonly used SOMs: 2,5-DHB, 9AA, α-CHCA, harmine, and DCTB.

Their molecular structures contain aromatic systems for sufficient UV absorption, acidic or basic functional groups for an enhanced ionization via proton transfer and they give highly crystalline layers.^[50] The analyte scope, ionization efficiency^[77-78], cluster formation^[79], isomers^[80-81] and best sample preparation methods^[82] for SOMs, like 2,5-DHB, were intensively studied to investigate and improve their matrix performances.

Standard SOMs often have broad analyte scopes and high ionization efficiencies, and consequently they are also commonly used for MALDI MSI analytics. Yet, for reliable and reproducible analysis of the spatial distributions of analytes via MALDI MSI, matrices should meet four additional prerequisites:

4. Low background interferences (*MALDI silent*): Background interferences from the matrix (e.g., molecule ions, fragment ions, cluster ions) in the low mass range below $m/z = 1000$ can cause overlapping signals of matrix and LMWCs, which complicates the spectra evaluation. Furthermore, matrix ions, which are normally cut off for HMWC analytics, can surcharge the detector during LMWC detection, which significantly reduces its lifetime.^[19] This prerequisite is optional for the analysis of HMWCs as their signals are observed in the high mass range and, hence do not interfere with SOM signals.
5. Stability in HV: MALDI MS measurements are mostly performed in HV ($\sim 10^{-7}$ mbar) to ensure an unhampered ion flight. It is therefore important that the matrix does not sublime in low pressure atmosphere to guarantee identical and reproducible measurement conditions for all measurement spots during long-term HV MALDI MSI

experiments. This is important for all analytics, as a changing matrix/analyte ratio reduces the spot-to-spot reproducibility in general.^[14]

6. Uniform thin-layer coating: Similarly affecting the reproducibility of MALDI MSI results, the matrix needs to be homogeneously coated on the tissue slice. Deviations in layer thickness significantly influences the ratio between matrix and analytes and, hence the accuracy of the measurements.^[62]
7. Dual polarity mode suitability: The ability for detection of positively, as well as negatively charged analyte ions by using the same matrix is a strategy to acquire the most chemical information from unique samples or when only small sample amounts are available (miniaturization procedure).^[22]

Standard SOMs often fail to meet these additional prerequisites (4. - 7.), resulting in different challenges during sample measurements and spectra evaluations, especially for LMWC analytics. When using SOMs, signals originating from molecule ions, fragment ions and cluster ions of the SOM are observed in the low molar mass range ($m/z < 1000$). The molecule ions are either (de-)protonated ($[M \pm H^+]^{+/-}$) or radicalized ($[M \pm e]^{•+/-}$) species of the matrix, whereas the cluster ions are formed by a certain number of matrix molecules ($[n \cdot M]^{+/-}$) or in combination with ubiquitous alkali metals or halogens (e.g., $[M + Na^+/K^+]^+$, $[M + Cl^-]^-$).^[51, 83] As the matrix is commonly used in excess, to separate and protect the analyte, these signals are mostly inevitable and usually of high intensity. The fragment ions are caused by high mechanical forces due to the strong laser shots.^[57] All these matrix related signals can overlap with the LMWC ion signals, and analyte and SOM signals are not distinguishable during the evaluation. A further drawback of SOMs is their tendency to sublime in HV due to their low molar mass.^[14, 84] MALDI MSI measurements, especially when aiming for high spatial resolution, can take several hours. The thin matrix layer can start to evaporate, leading to differences in matrix coating for spots measured later in the acquisition time. This variation of the measurement conditions (matrix/analyte ratio) on different parts of the specimen reduces the spot-to-spot reproducibility of the measured spectra, which can significantly reduce the comparability of MALDI MSI pictures (2D-heat maps).^[14] Apart from the vacuum stability, the reproducibility is also influenced by the homogeneous coating of the tissue slice with the matrix. Due to the necessity for crystal formation of standard SOMs for a sufficient analyte separation and ionization, the available coating methods (e.g., electrospray^[21]) must be performed very carefully and require extensive experience by the operator.^[20] The homogeneity of the crystal size and distribution

strongly affects the ionic intensities of the mass spectra and spatial resolution of the MSI heat maps.^[62] Additionally, depending on the functionalization of the SOMs, they are used for the detection of either protonated cations in positive mode (acid) or for deprotonated anions in negative mode (base) but rarely for both modes simultaneously.^[22] Dual polarity mode suitability can be important, e.g., for unique samples or when only low amounts of sample are available.^[23, 85]

2.1.3.1. Modification and Development of MALDI Matrices

As standard SOMs often fail to meet the additional matrix prerequisites for LMWC analytics via MALDI MSI (chapter 2.1.3, 4. - 7.), research is carried out to improve them or to develop new materials.^[14, 19] The improvement of standard SOMs targets their structural modification to decrease the background interferences in the low mass range, to increase their ionization efficiency and vacuum stability, to specify their analyte scope or to enhance homogeneous crystal formation.^[86]

The background interferences of SOMs can be reduced by increasing their molar mass, e.g., by binding them to silica nanoparticles or cyclodextrin, which is reported to significantly decrease detected matrix peaks.^[24, 87-89] Another strategy to achieve high mass matrices is to convert SOMs to polymeric materials, either by incorporating them into a macromolecular network or by introducing them into a polymer's sidechain. The covalent incorporation of the SOM (here: 2,5-DHB) in a sol-gel silicon dioxide polymer was reported to prevent the SOM from being detected during measurements (chapter 2.2.2 for more details).^[2] The other method, namely the introduction of SOMs in a polymer's sidechain, is an important part of this thesis and is discussed in detail in chapter 4.4. To increase the vacuum stability during long-term MALDI MSI measurements, acidic SOMs like 2,5-DHB or α -CHCA were converted into ionic liquids using different amino-group-containing molecules as counter ions.^[25, 28] Also, introducing reactive functional groups allows an enhanced detection sensitivity, e.g., boronic acids can be incorporated for predominantly detecting diols, and aldehydes to target amine containing analytes.^[90-91] The previously described modification of SOMs with cyclodextrin to reduce matrix signals is also reported to improve homogeneous crystal formation, leading to an increased spot-to-spot reproducibility due to enhanced matrix-analyte distribution.^[92] All these variations of SOMs were reported to increase the signal intensities of the analyzed LMWCs, compared to standard SOMs.

Regarding the development of new MALDI matrices for LMWC analytics via MALDI MSI, several well-known inorganic and organic materials were reported to be promising

candidates.^[19, 93] Inorganic materials, like graphene (-oxides), can enable the sensitive detection of various small molecules, like amino acids, anticancer drugs, steroids and components of medical herbs, with low background interferences.^[30-31] Additionally, when doped with phosphorus and nitrogen, graphene is suitable for dual polarity mode measurements.^[94] Another carbon based matrix are graphite dots, which were successfully used for *MALDI silent* MSI measurements of small antibiotic metabolites, and showed strong signal intensities and a good tolerance to salt contaminants.^[32, 95] Vacuum stable heavy metal oxides, like tungsten- and rhenium oxide, were applied as matrices for the detection of LMWCs, such as cocaine, quinuclidine, apigenin and glucose, without matrix signals in the low mass range.^[29] In addition to metal oxides, iron-based *metal organic frameworks* (MOFs) were also found to be capable of small molecule detection (e.g., dopamine, glyphosate and pyrene).^[96]

Organic materials as potential MALDI matrices for LMWC detection can be divided into SOMs and high molecular weight matrices. Newly developed SOMs, like *N*-phenyl-2-naphthylamine and 3,4-dimethoxycinnamic acid, are marked by low background interferences, strong UV absorption and high salt tolerance while allowing the detection of a wide range of LMWCs, e.g., fatty- and amino acids, antioxidants, phospholipids and small peptides.^[26, 97] Also, SOMs can be developed to target specific analytes: 2-phenyl-3-(*p*-aminophenyl)acrylonitrile is a derivative of α -CHCA, which is functionalized with amine group for an increased sensitivity towards aldehyde containing LMWCs, like saccharides.^[98] The low suitability of standard SOMs for dual polarity mode measurements was improved by developing matrices like 3-aminophthalhydrazide (luminol) and 2-(methylamino)benzoic acid, which contain both acidic and basic functional groups.^[22, 99] High molecular weight matrices have a molar mass outside the mass range needed for LMWC analytics to avoid background interferences and overlapping of matrix and analyte signals. An early approach was meso-tetrakis(pentafluorophenyl)porphyrin ($M = 974.57$ Da), which allowed the *MALDI silent* detection of nonylphenol ethoxylates, fatty acids and sugars with ionic intensities comparable to α -CHCA.^[27, 100-101] Additionally, porphyrin-based matrices were also used for the detection of the water-soluble vitamins B1, B2, B6, B12 and C.^[102] Metal-phthalocyanine-based matrices with different metal-centers allowed to detect amino acids, peptides and fatty acids.^[103] Also, several conjugated oligomers and polymers were found to be promising matrices for LMWC detection.^[3-4] These approaches are explicitly discussed in chapter 2.2.2.

2.2. Polymers as MALDI Matrix Systems

2.2.1. Basics of Polymers

Polymers are macromolecules which consist of multiple repeating units (monomers). The molecular structure of the monomers and the chosen polymerization method define the molecular structure of the resulting polymer, which can be linear, branched, or cross-linked.^[104] Almost all known reaction mechanisms can be used for polymerization reactions, and are grouped in two main categories based on the growing procedure of the polymer: Chain growth polymerizations (e.g., radical polymerization, ionic polymerization) and step growth polymerizations (e.g., polycondensation, polyaddition). These polymerization reactions can proceed in different media, such as solutions, emulsions, suspensions, gas-phase and in highly viscous liquid-phase-polymers.^[104] The synthesized polymer chains generally have non-uniform chain lengths, resulting in a varying distribution of their molar masses, which is often described as an Gaussian or Flory-Schulz distribution (Figure 4).^[1] Consequently, synthetic polymers do not have a distinct molar mass but rather averaged ones: *number average molar mass* (M_n) and *weight average molar mass* (M_w).^[104] The ratio between M_w and M_n , also called *poly dispersity index*

Figure 4: Exemplary Gaussian (red) and Flory-Schulz (black) distributions (redrawn from [1]).

($\bar{D}$), describes the broadness of the molar mass distribution of a polymer. Another interesting measure is the *degree of polymerization* (DP), giving the average number of repeating units per polymer chain.^[1] The molecular structure and the average molar mass strongly affect the chemical and physical properties of the polymer.^[104] Chemical properties, e.g., are acidity/basicity, polarity, reactivity towards other chemicals, solubility and flammability, whereas physical properties are, e.g., melting-/glass-/degradation-temperature, density, elasticity, viscosity, absorbance and conductivity.^[104-105]

The acidity/basicity and the reactivity towards other chemicals depend on the availability of, e.g., acidic, basic or redox active functional groups in the molecular structure of the polymer.^[106] Morphological properties, such as the degree of crystallinity, are influenced by the molar mass, melting point, cooling rate, and degree of branching or

cross-linking for a given polymer. The molar mass also contributes to the vacuum stability: Small chain oligomers are often not vacuum stable, whereas the large chain polymers retain their full mass in HV.^[104] The absorbance (UV or IR range) and conductivity depend on the molecular structures containing conjugated π -electron systems like benzene- or thiophene rings, or specific functional groups, such as alcohol-, amino- or thio-groups.^[104, 107]

Conjugated polymers, i.e., polymers with sp and sp^2 hybridized backbones, are a special polymer class. As they tend to aggregate via π - π stacking, sidechains are introduced to impart solubility. Long alkyl sidechains lead to an enhanced solubility in non-polar solvents, whereas, e.g., ethylene oxide sidechains allows conjugated polymer to dissolve in polar solvents to ensure processibility via wet coating methods (e.g., spin-, spray-, drop-coating, etc.).^[108-109]

Consequently, the properties of a polymer are tunable by varying the design of their molecular structure, which makes them potentially suitable for many different applications, e.g., as LMWC matrix for MALDI MS and MSI.

2.2.2. Pathway to Polymeric Matrices

The chemical and physical properties of polymers, described in chapter 2.2.1, can be modified by varying the molecular structure of the backbone and/or the sidechains. Hence, polymers can be designed to meet the fundamental and additional prerequisites of MALDI matrices (chapter 2.1.3), yielding PMs with a high potential for LMWC analytics with MALDI MS and MALDI MSI:

1. The laser absorption can be achieved by introducing conjugated π -electron systems into the polymer backbone or sidechains.
2. Depending on molar mass, melting point and chemical structure, polymers can potentially achieve a high degree of crystallinity to support analyte incorporation. However, most polymers contain amorphous regions and therefore do not adopt a fully crystalline morphology.^[104]
3. Acidic or basic functional groups can be introduced in the polymer backbone and/or the sidechains, to potentially enhance the ionization efficiency and the analyte scope.
4. As long as polymer backbone and sidechains do not fragment, there will be no matrix signals in the low mass range, as the polymers' molar mass is usually higher than 1000 Da.
5. Due to the high molar mass, polymers do not sublime in low pressure atmosphere.
6. Many different thin-layer coating methods for polymers are established and intensively studied, such as spin coating^[6, 110], spray coating^[5], shear coating^[111] and drop coating^[5-6]. They require a high solubility in commonly used solvents, which is influenced by the chain length, the molecular structure of the backbone, and the sidechains.
7. In addition to point (3), the backbones and sidechains of a polymer can potentially be functionalized also with acidic and basic functional groups simultaneously to introduce dual polarity mode suitability. This can be achieved, e.g., by co-polymerizing two different monomers with acidic and basic sidechains, respectively.

As these modifications can be facilely introduced to polymers, PMs have previously been used for LMWC analytics with MALDI MS. In the first approach from Lin et al. in 2002, a sol-gel matrix consisting of poly(siloxane) and the SOM 2,5-DHB was used for the detection of small proteins, peptides, and amino acids (Figure 5). The 2,5-DHB (red) was covalently incorporated in the poly(siloxane) network (black), introducing UV

absorption and acidic functional groups from the SOM into the polymer, while preventing the SOM from producing matrix signals in the low mass range.^[2] In 2004, Woldegiorgis et al. introduced the term *polymer assisted laser desorption/ionization* (PALDI) and used four conjugated oligomers with defined lengths, *quaterthiophene* (Qthio), *sexithiophene* (Sthio), *tetrahydro-hexaoxa-pentacene* (3DIO), and *tetrahydro-octaoxa-heptacene* (4DIO) (Figure 6) for the *MALDI silent* detection of several small organic analytes with various functionalities. Here, the conjugated π -electron systems give a sufficient UV absorption. Additionally, the charge

Figure 6: Molecular structure of oligomeric matrices Qthio, Sthio, 3DIO, and 4DIO.^[4]

transfer ionization mechanism was investigated and discussed for these oligomeric matrices.^[4] Woldegiorgis et al. also used the oligomeric matrices for the analysis of low molecular weight polystyrenes and polyethylene glycols. The resulting spectra showed significantly better *signal-to-noise* (S/N) ratios compared to the conventional SOMs 2,5-DHB and *anthracenetriol* (dithranol).^[112] Also in 2004, Soltzberg et al. showed that the conjugated polymer *poly(3-octylthiophene-2,5-diyl)* (P3OT) can be used for the detection of small organic compounds, such as carboxylic acids and esters in negative polarity mode with sufficient sensitivity (Figure 7). Similar to the aforementioned conjugated oligomers, this polymer has a conjugated π -electron system to allow UV absorption. To impart solubility, the thiophene-backbone is functionalized with alkyl sidechains. Interestingly, the analytes were mostly detected as $[M-H]^+$ ions, indicating an ionization via proton transfer, which is unexpected as charge transfer ionization was reported for the oligomers.^[3-4]

Figure 7: Molecular structure of P3OT.^[3]

The results and especially the publications of Lin and Soltzberg, show the potential of PMs for LMWC analytics. Still, more systematic investigations are needed, e.g., regarding the molecular structure, or to investigate the influence of intrinsic properties (e.g., morphology) on the PM's performances.

Figure 5: Molecular structure of sol-gel poly(siloxane)/DHB.^[2]

2.3. Software for MALDI MS Data Handling

To develop new matrices, a fast and facile comparison of the obtained mass spectra is desirable. This also encompasses the clear depiction of measurement parameters (e.g., laser intensity and detection / polarity mode), and sample information (e.g., sample composition, preparation method) and molecular structures. Software used for this should allow three main tasks:

1. Visualization: Plotting options for single and stacked arrangements of spectra and visualization options for the clear depiction of sample information and measurement parameters. Furthermore, the software should provide tools to prepare the graph for export as image, which will be used for documents and presentations.
2. Data processing: After plotting, most spectra need to be further processed, e.g., by smoothing or baseline correction. This can significantly simplify the evaluation but can also cause loss of information (e.g., erasing isotopic pattern due to smoothing).
3. Data evaluation: After the spectra are plotted and processed, they are evaluated, e.g., by peak picking. When characterizing polymers, the polymeric properties (M_n , M_w , $\bar{M}_n$, DP) and the *molar mass of the repeating unit* (M_{RU}) are calculated from these values.

There are three software options available: commercial software from the instrument manufacturer (e.g., flexAnalysis), open-source software (e.g., openchrom and mass-up), and graphing software (e.g., OriginPro).

The first software option is the commercial software of the used MALDI MS instrument, e.g., the software flexAnalysis3.4 from the German company Bruker Daltonik GmbH. This software has several plotting options, like single, stacked and overlapping spectra, and allows smoothing, baseline corrections, peak picking and recalibration.^[113] With the addon-software polytools1.2, the evaluation of polymer spectra regarding the polymeric properties (M_n , M_w , $\bar{M}_n$, DP) is possible.

Open-source MS software, such as openchrom and mass-up, are alternative options.^[114-115] The program openchrom was developed and published in 2010 for the facile analysis of data from various analytical instruments, such as UV, FTIR, MS and NMR.^[116] It is platform independent, as it is written in the programming language Java, and is capable of raw data import from the used MALDI instrument (Bruker Autoflex Speed). The focus of the software's main functions is the handling of mass spectrometric and UV/Vis spectroscopic data after sample separation via chromatographic

methods. The program mass-up was developed and published in 2015, specifically for processing and evaluation of MALDI MS data.^[117] This program is capable of comparing obtained results with databases, hierarchical clustering, component analysis and quality control. The software was specifically developed to allow the evaluation of MS data for bioinformatical analysis (e.g., proteomics).

The third option is a graphing software, e.g., OriginPro, a graphing and data analysis software first published in 1991 by the American company MicroCal, Inc., which was renamed to Microcal Software, Inc. in 1993 and to OriginLab Corporation in 2000.^[118-119] The software allows different functions, such as graphing and data exploration, curve and surface fitting, peak analysis, statistics, signal-, data-, and image-processing and simple and advanced mathematics.^[120] Due to this broad functionality scope, OriginPro is potentially capable of the reproducible visualization, processing and evaluation of MALDI MS data. However, performing these tasks manually for every data set is highly time consuming, as multiple functions are involved. Yet OriginPro allows the integration of scripts to repetitively execute commands and automizing the whole data handling procedure. These scripts can be written by using OriginPro's native programming languages LabTalk and Origin C: X Functions, which is based on the high-level programming language ANSI C. The integrated Code Builder of OriginPro helps programming by visualizing the text in different colors, allowing to debug, and run the whole script or individual sections.

3. Research Objectives and Aim of Thesis

The aim of this doctoral thesis is to develop PMs and explore their potential for the reliable and reproducible analysis of LMWCs with MALDI MS and MALDI MSI. Two strategies were pursued: (i) The investigation of conjugated polymers as matrices, and (ii) the polymerization of standard SOMs.

As previously described in chapter 2.2.2, conjugated oligomers and polymers are promising candidates as MALDI matrices for LMWC detection, because their molecular structure meets several prerequisites for a good matrix performance: Absorption in UV range due to their conjugated backbone, and vacuum stability and *MALDI silence* due to their high molar mass. On basis of strategy (i), the first objective of this work was to investigate the matrix performances of various conjugated homo- and heteropolymers and compare them with standard SOMs (2,5-DHB and 9AA). Conjugated polymers adopt a semi-crystalline morphology, i.e., contain crystalline and amorphous domains. For standard SOMs, a high crystallinity was reported to be necessary, as the analytes are incorporated into the crystal lattice (chapter 2.1.3). Hence, when using *conjugated polymer matrices* (CPMs) it is expected that the analyte molecules are incorporated in the crystalline domains of the CPMs. Therefore, the second objective of this work was to investigate this interdependence by comparing the matrix performance of fully amorphous CPMs with semi-crystalline CPMs.

Lin et al. had previously shown that incorporating SOMs in a polymer backbone is a feasible way to obtain matrices, which are *MALDI silent* and vacuum stable while retaining the good matrix performances of SOMs. Yet, the group polymerized DHB by using two functional hydroxy groups to directly incorporate the SOM in the polymer backbone. Another pathway would be to modify the SOM structure, thereby leaving the functional groups untouched, and using the modification to incorporate the SOM as a sidechain in a polymer. This idea forms the core of strategy (ii) of this work: Modifying and then polymerizing standard SOMs to obtain *polymerized SOMs* (P(SOMs)). The matrix performances of these P(SOMs) were investigated and compared with their corresponding SOMs.

The development of PMs requires a fast and facile comparison of mass spectra to benchmark the matrix performances of new matrices with standard matrices. Therefore, a script was programmed in the graphing software OriginPro (chapter 2.3) to automatize the visualization, processing, and evaluation of the generated data.

4. Results & Discussion

4.1. Automation of Data Handling with OriginPro 2020b

The graphing software OriginPro was used to handle and benchmark all MALDI MS data obtained over the course of this thesis. As the manual data handling with this software is time-consuming, a script containing sets of different functions was programmed to automatize the following tasks: (1) Visualization, (2) data processing and (3) data evaluation. This script was written in the programming languages LabTalk and Origin C: X Functions, using the English version of the code builder of OriginPro 2020b and contains in total 8135 *lines* (L.) of code. The full script is given in chapter 8.1.

4.1.1. Task (1): Visualization

The first task is the visualization of the measured MALDI MS spectra, starting from importing the data to OriginPro, arranging the data in data tables, and then plotting in graphs (chapter 8.1.1, L. 3 - 2673). This script allows different visualizations for different analyte classes (e.g., LMWCs, synthetic polymer, biopolymers, etc.), in line with their requirements for plotting, such as data normalization and different X-axis scales. Furthermore, the script offers options for arranging the spectra: *Multiple single spectra* in individual graphs or up to six *Stacked spectra* in one graph.

For plotting *Multiple single spectra* (L. 13 - 132), the imported data files are visualized in individual graphs by repetitively executing the following six functions for each file:

1. Import the data from the file into an individual data table (L. 17 - 28).
2. Visualize the data as a spectrum in an individual graph (L. 30 - 53).
3. Perform peak picking to find the local maximum of the signal with the highest intensity, insert this peak data point as dotted plot in the graph and edit the peak plot design to small, red circle dots (Figure 16) (L. 55 - 72).
4. Edit the design and position of the legend label and use the data file name as legend label text (L. 74 - 80).
5. Add a Y-axis scrollbar and X-axis scroller to the graph and edit their design (L. 82 - 106).
6. Set the final axis scale and execute final functions, i.e., renaming data table, hiding this table, and reactivating the graph window (L. 108 - 132).

Examples for *single spectra* are depicted in Figure 12 to Figure 16.

The plotting of spectra in the *Stacked spectra* layout allows to visualize up to six spectra stacked above each other in one single graph. This layout enables a simple and direct comparison of different spectra, e.g., to emphasize how measuring the same analyte with different matrices changes the peak pattern.

This visualization task is carried out using a variation of the six functions described for plotting *Multiple single spectra*. The differences are as follows:

Regarding function 1:

The data from all imported files are arranged in a single data table. Additionally, information about measurement parameters and sample composition can be manually logged in a dialogue shown in Figure 8 for the labels of the stacked plots.

Regarding function 2:

The spectra are visualized in individual plots stacked above each other in one graph. Each of the plots are visualized in a different color (Figure 9).

Regarding function 3:

Function 3 from *Multiple single spectra* is executed for each individual plot, except that for the design of the individual dotted peak plots the corresponding plot colors are used.

Regarding function 4:

Create labels displaying information (Figure 8) regarding laser intensity (1 label per plot), sample preparation method (1 label per graph) and polarity mode (1 label per graph).

Regarding function 5:

Create Y-axis scrollbars for each plot and X-axis scrollers for the bottom plot and edit their designs (e.g., using corresponding plot color).

Regarding function 6:

Carry out final functions, i.e., hide data tables, show graph window, and open a dialogue, giving the names of the plotted data files to finally control whether the correct data are visualized.

Figure 8: Dialogue to insert information about the measurement parameters of a 3-stacked plot.

Figure 9 shows a graph containing three stacked plots with exemplary MALDI MS spectra of a biological complex sample (red), a polymer sample (blue) and a simple LMWC analyte sample (black). The labels and plot legends are filled with placeholders ("XX") to show where the previously logged (Figure 8) sample names and measurement conditions are inserted. By default, the Y-axes and X-axes of the stacked spectra are linked to each other for better comparability, allowing a scale change of all spectra simultaneously. If this is not required, the linking can be deactivated manually for every individual spectrum. Before exporting the graph in an image format, the Y-axis scrollbars and X-axis scrollers can be hidden, using a script described in chapter 4.1.4.

Figure 9: Example of a 3-stacked graph showing three individual plots. The plots show MALDI MS spectra of a complex biological sample (red), a polymer (blue) and a single LMWC (black). In contrast to the Y-axis, the X-axis of the plots are linked to each other.

Depending on the number and size of the data files and the computer performance, the data visualization with the OriginPro script takes between 10 and 30 seconds (without user input). If needed, the final plots can be further optimized manually, e.g., by rescaling the axes via the scrollbars and scrollers, adding peaks (chapter 4.1.3) or using different data processing options, like smoothing and baseline correction (chapter 4.1.2).

4.1.2. Task (2): Data Processing

It is often necessary to further process the data, e.g., to improve the visualization or to simplify later evaluation. The script supports (i) smoothing and (ii) baseline correction (L. 2675 - 4686).

Smoothing of MALDI MS spectra can increase the S/N ratio, flatten the spectrum line, and simplify a later baseline correction. The *smooth* command of OriginPro can use six different algorithms for the smoothing process (Adjacent-Averaging, Savitzky-Golay, Percentile Filter, FFT Filter, Lowess, Loess and Binomial).^[120] Of course, when comparing spectra in a stacked plot, the same smoothing algorithm must be applied for all of them. Here, the algorithm *FFT Filter* is used as it results in more distinct signals and simplifies the evaluation. If needed, the algorithm can be applied multiple times per spectrum.

The *smooth* script (chapter 8.1.2, L. 2675 - 3079) allows *Smoothing of all spectra* simultaneously or only *Smoothing of marked spectrum* within a graph. The smoothing process itself consists of only one *smooth* command, using the *FFT Filter* algorithm. It automatically exchanges the raw input data by the smoothed output data, so that no further commands are required for this process (Figure 10).

Figure 10: Exemplary *smooth* command used for smoothing MALDI MS spectra in this thesis (including descriptions of subfunctions).

Figure 11 gives an example of the same dataset with (bottom) and without (top) applying the smoothing algorithm. As can be observed for this example, the smoothing process leads to more distinct signals and flattens the spectrum line. At the same time, it also erases information from the spectra (here: isotopic patterns) and decreases the signal intensities. In this thesis, the smoothing option is mostly applied for polymer spectra showing sufficient signal intensities, where isotopic patterns are not important for evaluation and can be neglected.

Figure 11: Example of smoothing a MALDI MS spectrum of a polymer sample. Top: original spectrum (blue); bottom: smoothed spectrum (black). The spectra are divided into three parts showing different m/z ranges.

The second option for data processing [\(ii\)](#) is a baseline correction, which is a subtraction of a predefined spectrum from a sample spectrum. E.g., the predefined spectrum can be a spectrum of a blank sample, which is then directly subtracted during the measurement of the sample spectrum, like it is often performed for UV measurements. In case of mass spectra, the predefined spectrum is a line, also called baseline. The developed *baseline* script (chapter 8.1.3, L. 3080 - 4686) can be used to define this line either manually or automatically and subtract it from the original spectrum. In the manual option, the baseline is defined by hand using the data cursor and directly plotted in the corresponding spectrum as light-green dot-line-graph (L. 3088 - 3135). After pushing the escape button when finished, the baseline corrected spectrum is calculated and plotted together with the original spectrum and the manually defined baseline in a separate plot for a direct comparison (Figure 12). Finally, either all performed changes are reset when the results are not satisfying, or the original spectrum is exchanged by the baseline corrected spectrum if the baseline correction was successful (L. 3136 - 3633).

Figure 12: Manual baseline correction of a MALDI MS polymer spectrum showing the original spectrum (black), the manually defined baseline (red) and baseline corrected spectrum (blue). The baseline consists of 52 data points.

The procedure for the automatic baseline correction is more complicated (L. 3635 - 4686). Here, the baseline is defined by the peak picking command *pkfind*. This command identifies a local maximum (Figure 13, blue dots) in a defined range and can also output the corresponding left and right peak markers (start and end point of this defined range, Figure 13, red/green dots). These peak markers are the data points of the baseline. For a correct determination, the following peak picking parameters for the *pkfind* command need to be defined:

Figure 13: Exemplary original spectrum from Figure 14 with left peak markers (red), right peak markers (green), and the corresponding local maxima (blue).

- **Local points:** This number defines the amount of plot data points, which are used to find a local maximum. The smaller the value, the more signals are picked, and the baseline is more precise.
- **Minimum height:** This number defines the minimum intensity (Y-value) of a local maximum. All maxima with lower Y-values will be neglected.

- Maximum half width: This value defines the maximum range for searching the peak markers. Especially for broad signals this value needs to be large. In general, it is recommended to set rather a high value (~100 a.u.) than a small value.

When the parameters are set, the baseline is calculated and plotted together with the original spectrum in a separate plot for a direct comparison. Three options are available, depending on the success of the baseline calculation: (a) Reset all performed changes and functions; (b) change the parameters to calculate a new baseline; and (c) go on with the calculated baseline and subtract it from the original spectrum. The time consumption of this calculation significantly depends on the *pkfind* parameters and the resulting number of data points of the baseline (up to 15 minutes, on average 1-3 minutes). The baseline corrected spectrum is plotted together with the calculated baseline and the original spectrum in a separate plot for a direct comparison (Figure 14). Finally, the original spectrum in the original plot is exchanged by the baseline corrected spectrum and the automatic baseline correction is finished.

Figure 14: Automatic baseline correction of a MALDI MS polymer spectrum showing the original spectrum (black), the automatically defined baseline (red) and the baseline corrected spectrum (blue). The baseline consists of 322 data points.

In general, the manual baseline correction is faster but less precise than the automatic one (Figure 12: manual = 52 baseline data points; Figure 14: automatic = 322 baseline

data points). However, although the automatic baseline correction is more precise, manual execution is required, e.g., when the signals are very broad, like in spectra of polypropylene imine based glycodendrimers (Figure 15). Then the successful determination of suitable *pkfind* parameters for automatic baseline correction is unlikely.

Figure 15: Exemplary manual baseline correction of a MALDI MS spectrum of a glycodendrimer, showing very broad signals.

4.1.3. **Task (3): Data Evaluation**

After visualization and data processing (chapter 4.1.1 and 4.1.2), the spectra are ready for evaluation: (i) peak picking, and, for polymer spectra with a Gaussian or Flory-Schulz signal distribution (chapter 2.2.1), (ii) calculation of the polymeric properties (M_n , M_w , $\bar{D}$, DP). Regarding (ii), also the *average molar mass of the repeating unit* (M_{RU}) and its standard error (ΔM_{RU}) can be calculated. The written script allows both peak picking and calculation of polymeric properties (L. 4688 - 7718).

The peak picking process (L. 4688 - 5938) can be performed either manually or automatically. The basic principle of the manual peak picking process is similar to the first step of manual baseline correction (chapter 4.1.2). It is used for one plot when either only a small number of signals (<15) are detected or when the detected signals are broad (Figure 15). The developed *manualpeakpicking* script (chapter 8.1.4, L. 4688 - 5147) can be utilized to manually pick relevant signals, using a data cursor. The X- and Y-coordinates (X = m/z value; Y = intensity) of the picked signal maxima are put in the corresponding spectrum data tables and sorted according to their m/z values. The spectrum plot is updated automatically, showing the X-coordinates of all picked signals. The peak picking can be simplified by previous smoothing of the spectrum (chapter 4.1.2).

The *autopeakpicking* script (chapter 8.1.5, L. 5158 - 5673) can perform automatic peak picking either in one plot or, in case of stacked spectra, in up to six ones. Therefore, parameters for the *pkfind* command need to be defined, which are described for the automatic baseline correction in chapter 4.1.2. The X- and Y-coordinates of all signals matching these parameters are added to the corresponding spectrum data table. The plot is updated automatically, visualizing all matching peaks and the automatic peak picking process is finished. The successful determination of the *pkfind* parameters is unlikely when the signals are too broad (Figure 15), thus requiring manual peak picking.

When picked signals need to be deleted, the script *deletesignals* (chapter 8.1.6, L. 5675 - 5938) can be used.

The calculation of the polymeric properties (M_n , M_w , $\bar{D}$, DP and M_{RU}) can be performed with the developed *PDI* script (chapter 8.1.7, L. 5940 - 7718). At the beginning, all equidistant peaks are picked (Figure 16, red dots) either manually or automatically with the aforementioned peak picking scripts. Then, the *PDI* script uses the peak data to

calculate M_n according to formula (1), M_w according to formula (2) and $\bar{D}$ according to formula (3).^[104] The peak data is required for these calculations, as the molar amount of the i^{th} polymer chain N_i correlates with the intensity of the corresponding mass signal. Additionally, the molar mass of the respective polymer chain i is set to be equal to the m/z -value of the corresponding mass signal.

$$M_n = \frac{\sum_i N_i M_i}{\sum_i N_i} \quad (1)$$

$$M_w = \frac{\sum_i N_i M_i^2}{\sum_i N_i M_i} \quad (2)$$

$$\bar{D} = \frac{M_w}{M_n} \quad (3)$$

i = number of polymer chain

N_i = molar amount of i^{th} polymer chain (here: intensity of mass signal)

M_i = molar mass of i^{th} polymer chain (here: m/z value of mass signal)

Afterwards, the *PDI* script calculates M_{RU} according to formula (4) and the DP by dividing M_w through M_{RU} (formula (5)).

$$M_{RU} = \frac{\sum_i (M_i - M_{i-1})}{i-1} \quad (4)$$

$$DP = \frac{M_w}{M_{RU}} \quad (5)$$

Figure 16: Smoothed, baseline corrected, and peak picked MALDI MS polymer spectrum showing two signal series for the calculation of the polymeric properties.

Depending on the amount of different end group combinations, a MALDI MS spectrum of a polymer can contain more than one signal series (Figure 16). In this case, the polymeric properties (M_n , M_w , $\bar{D}$, DP and M_{RU}) need to be calculated individually for each signal series. In Table 1 the calculated values of the 1. signal series (red dots) and the 2. signal series (blue dots) of the exemplary plot in Figure 16 are listed. The automatic calculation of more than one signal series is not implemented in the script and must be done manually.

Table 1: Polymeric properties calculated for the first (red) and second (blue) signal series.

Property	1. Signal series (red)	2. Signal series (blue)
M_n	4061.1 g/mol	4116.3 g/mol
M_w	4495.7 g/mol	4627.6 g/mol
$\bar{D}$	1.1	1.1
M_{RU}	74.2 g/mol	74.1 g/mol
DP	60.6	62.4

4.1.4. Small Scripts for graph design manipulation

When finalizing the graphs for presentation, it can be necessary to perform small design changes, e.g., for better visualization of the plots. Using scripts to automatize these operations (L. 7720 - 8135) can speed up the finalization significantly.

These small scripts include: (i) Update legend label text, (ii) show/hide scrollbars and scrollers, and (iii) insert X-axis breaks.

(i) The first script *nachbearbeitung* (chapter 8.1.8, L. 7720 - 7751) allows to update the legend labels of all plots within a graph. During the visualization in a stacked graph layout (Figure 9), the legend labels are filled with the text from the

comment field of the corresponding Y-value column of the plot (here: *absolute intensity*; marked in Figure 17). When the text of this *comment* field is changed, the legend label text is not updated automatically, so this script can be used to do this for all stacked plots simultaneously.

	A(X1)	B(Y1)
Long Name	m/z	Absolute Intensity
Units		
Comments		Matrix: XX; Analyte: XX; Additive: XX
F(x)=		
1	0	227
2	0	201
3	0	236
4	0	204

Figure 17: Exemplary headlines (yellow) of a mass spectrum data set in OriginPro. The comment field of the Y-axis values is marked.

(ii) The second script allows to *hide* (chapter 8.1.9, L. 7753 - 7819) and *show* (chapter 8.1.10, L. 7821 - 7887) the Y-axis scrollbars and X-axis scrollers. While these scrollbars and scrollers are helpful during plot scale manipulation (Figure 9), they are not wanted when depicting the graph in documents or presentations. The script allows to hide (not delete) them before depicting the graph or show them again if a further axis scale manipulation is needed.

(iii) The third script *splitspectrum* (chapter 8.1.11, L. 7889 - 8127) and *splitspectrumstop* (chapter 8.1.12, L. 8129 - 8135) can be used to highlight several specific parts of a spectrum, as demonstrated in Figure 11 (chapter 4.1.2). This can be helpful, e.g., to show the successful detection of several LMWCs with various molar masses in a complex biological sample. *Splitspectrum* divides the graph in up to four parts, and starts with opening a dialogue (Figure 18) to determine the scale parameters for each part:

Please insert the m/z of the signal you want to highlight and give the size of the area around it (the areas must not overlap) including the increment of the x-axis (recommended --> 20% of area size value).

☐ Signal 1

m/z: XX

+/-: XX

Increment: XX

☐ Signal 2

m/z: XX

+/-: XX

Increment: XX

OK Abbrechen

Figure 18: Dialogue to input information of the parts to highlight (here: two parts).

- m/z: Center of the individual part; can, e.g., be the m/z (X-) value of a signal to highlight the signal.
- +/-: The range of an individual part; when, e.g., a “+/-” value of 50 and a “m/z” value (center) of 426 are set, the X-range from $(426 - 50 =)$ 376 to $(426 + 50 =)$ 476 is depicted.
- Increment: This value gives the increment of the individual part.

Splitspectrum will automatically abort if the values would result in overlapping plot parts. If the results are not satisfying, *splitspectrumstop* can be used to reset the X-scale.

4.2. Conjugated Polymers as a New Class of Dual-Mode Matrices for MALDI Mass Spectrometry and Imaging (Publication 1)

As detailed in chapter 3, two strategies were explored to develop PMs for LMWC analytics with MALDI MS and MALDI MSI: (i) the investigation of conjugated polymers as matrices, and (ii) the polymerization of standard SOMs.

One of the most important prerequisites of a MALDI matrix is the sufficient absorption of the laser radiation to support the ionization of the analytes.^[13] As the common UV lasers emit radiation at 355 nm, the PMs require a conjugated π -electron system either in the backbone or the sidechains. To explore strategy (i), five polymers with conjugated backbones were investigated as CPMs (Figure 19): **a)** *poly{[N,N'-bis(2-octyldodecyl)-naphthalene-1,4,5,8-bis(dicarboximide)-2,5-diyl]-alt-5,5'-(2,2'-bithiophene)}* (**PNDI(T2)**), **b)** *poly(3-dodecylthiophene-2,5-diyl)* (**P3DDT**), **c)** *poly(9,9-di-n-octylfluorenyl-2,7-diyl)* (**P9OFI**), **d)** *poly{[2,3-bis(3-octyloxyphenyl)quinoxaline-5,8-diyl]-alt-(thiophene-2,5-diyl)}* (**PTQ1**), and **e)** *poly{[N,N'-bis(2-octyldodecyl)-isoindigo-5,5'-diyl]-alt-5,5'-(2,2'-bithiophene)}* (**PII(T2)**).

Figure 19: Molecular structures of a) **PNDI(T2)**, b) **P3DDT**, c) **P9OFI**, d) **PTQ1**, and e) **PII(T2)**.^[5]

UV/Vis measurements of drop-coated thin-films of the polymers were carried out to evaluate their absorption at the laser wavelength (355 nm). The drop-coating sample preparation was used to simulate the conditions during MALDI measurements. The measurements showed high relative absorptions at the laser wavelength for the CPMs **PNDI(T2)** (82%), **PTQ1** (93%), and **P9OFI** (72%). The CPMs **PII(T2)** (48%) and **P3DDT** (23%) exhibited less but still sufficient relative absorptions.

To explore the polymers as matrices for LMWC detection, MALDI measurements of selected analytes were carried out in positive and negative modes. For each mode, three analytes from different biochemical classes were used. For positive mode: *reserpine* (RP; Figure 20), an indole alkaloid used as MS standard, tetraethylammonium chloride, used to block potassium channels in histological assays, and coumaphos, an insecticide used to control varroa mite infestation in bee colonies.

Figure 20: Reflectron positive mode MALDI MS spectra of the sample plate (green), the pure analyte RP (red), the pure matrix **PNDI(T2)** (blue), and **PNDI(T2)** coated on RP (black). Peaks: m/z = 39 [K] $^+$; 395 [$RP3$] $^+$ (see SI); 607 [$RP-H$] $^+$. Laser intensity = 10%.^[5]

For negative mode: *cholic acid* (ChA), a steroid structured primary bile acid, *citric acid* (CiA; Figure 21), an important metabolite of the citric acid cycle, and phenylalanine, an essential amino acid. The spectra were plotted in 4-stacked graphs processed with OriginPro, showing the *pure sample plate* (green), the *pure analyte without matrix* (red), the *pure matrix without analyte* (blue) and the *analyte coated with matrix* (black). A comparison of the spectra of *analyte without matrix* (red) and *with matrix* (black) showed that the analytes are not observed without matrix. Furthermore, all analytes were detected with low background interferences using the CPMs **PNDI(T2)**, **P3DDT**, **PTQ1** and **P1I(T2)**: When using higher laser intensities ($\geq 50\%$), matrix related signals were observed in positive and negative mode at $m/z \leq 85$, originating from alkyl

sidechain fragments. Only **P9OFI** showed strong matrix signals in the low mass range in addition to the analyte signals and was disregarded for further measurements.

Limit of detection (LoD) measurements were carried out for RP (positive mode) and ChA (negative mode). The results showed that the CPMs enable a similar level of detection as standard SOMs in positive (DHB) and negative (9AA) mode, respectively. Similarly, a complex biological sample (*Ginkgo Biloba* extract), was measured in both polarity modes. Again, the CPMs allowed to detect a comparable peak number as the standard SOMs.

Figure 21: Reflectron negative mode MALDI MS spectra of the sample plate (green), the pure analyte CiA (red), the pure matrix **PNDI(T2)** (blue), and **PNDI(T2)** coated on CiA (black). Peaks: $m/z = 42$ [C_3H_6]⁻; 87 [$CiA1=C_3H_3O_3$]⁻; 191 [$CiA-H^+$]⁻. Laser intensity = 20%.^[5]

These results can be summarized as follows:

- Sufficient relative absorption at laser wavelength (355 nm).
- CPMs function as MALDI matrices for the detection of various LMWCs.
- Low amount of matrix related signals in the low mass range are observed (*MALDI silent*).
- They are dual polarity mode matrices, i.e., support measurements in positive and negative mode.
- Due to their high molar mass, the CPMs are fully vacuum stable.

- The matrix performances of the CPMs are competitive to standard SOMs regarding the analytes investigated.

For conjugated oligomers, an ionization via charge transfer was reported. Consequently, for the CPMs, which are not functionalized with acidic/basic groups, a similar type of ionization is expected.^[4] Yet, when further analyzing the detected signals, it is obvious that most of the analytes are ionized via ion transfer (e.g., proton transfer, cluster formation with $[\text{Na}]^+$ or $[\text{K}]^+$).^[3]

As described in chapter 2.1, MALDI MSI is capable of visualizing the spatial distribution and concentration of molecules in heat map images. These information can be used in medical and pathological research, e.g., to investigate drug metabolisms in organs and disease progress.

To access the potential of CPMs for MSI measurements, **P3DDT** was tested in MALDI MSI experiments of frozen rat brain sections. The collected data sets allowed to show the spatial distribution and relative concentration of different LMWCs (Figure 22; not identified). The results of the MSI measurements allow to add another point to the summary:

- CPMs have a high potential as matrices for MSI experiments to trace LMWCs.

Figure 22: MALDI Imaging heat maps of a rat brain visualizing the spatial distribution of eight different molecules.^[5]

The results were successfully published in the *Journal of American Chemical Society* in 2018 (chapter 7.1).^[5]

4.3. Amorphous Conjugated Polymers as Efficient Dual-Mode MALDI Matrices for Low-Molecular-Weight Analytes (Publication 2)

As shown in the previous project (chapter 4.2), CPMs are promising candidates for *MALDI silent* LMWC analytics with MALDI MS and MALDI MSI. This raises interesting questions with regards to the morphology: Polymers are only capable of adopting a semi-crystalline morphology, i.e., forming crystalline and amorphous domains. Yet, a fully crystalline structure is reported as one of the general prerequisites for MALDI matrices to ensure a sufficient incorporation of the analyte molecules into the crystal lattice (chapter 2.1.3). For CPMs, the formation of a semi-crystalline structure, represented by the degree of crystallinity, is influenced by the molecular structure, the molar mass, and physical properties of the polymer (chapter 2.2.2).

Consequently, the analyte molecules are expected to be incorporated in the crystalline domains of the CPMs. This in turn should result in a dependence of matrix performance and degree of crystallinity. To investigate this hypothesis, two completely amorphous *naphthalene diimide* (NDI(T2))/fluorene co-polymers were synthesized, **P(TNDIT-FI(C₄C₂))** and **P(TNDIT-FI(C₁₀C₈))** (Figure 23). The polymers were tested as CPMs, and their performance compared to **PNDI(T2)**, a semi-crystalline CPM containing an NDI(T2) repeating unit (chapter 4.2).

Figure 23: Synthetic routes for the NDI(T2)/fluorene co-polymers **P(TNDIT-FI(C₄C₂))** (red) and **P(TNDIT-FI(C₁₀C₈))** (green). The alkyl side-chains R₁ and R₂ are defined in this scheme.^[6]

The monomers NDI(T2) and fluorene were used as the corresponding homopolymers **PNDI(T2)** and **P9OFI** show excellent absorption behaviors and semi-crystalline morphologies. The structural relation allows to compare the matrix performances and furthermore to investigate the influence of the fluorene moiety on the MALDI stability of the co-polymers. As described before (chapter 4.2), the fluorene homopolymer **P9OFI** is not *MALDI silent*, i.e., shows strong background interferences in low mass range. To ensure a sufficient solubility in non-polar solvents and an amorphous structure, long and branched alkyl-sidechains were introduced to the monomers. The alternating co-polymers **P(TNDIT-FI(C₄C₂))** and **P(TNDIT-FI(C₁₀C₈))** were synthesized using a Suzuki-type coupling. ¹H-NMR and MALDI MS experiments confirmed the successful polymerization and molecular structures. GIWAXS and 2D-WAXS experiments were carried out to investigate the morphology and degree of crystallinity. Both showed a fully amorphous morphology for both co-polymers (Figure 24).

Figure 24: GIWAXS curves of **P(TNDIT-FI(C₄C₂))** (drop-coated: red; spin-coated: green) and **P(TNDIT-FI(C₁₀C₈))** (drop-coated: blue; spin-coated: pink).^[6]

In line with the previous project (chapter 4.2), the polymers were tested in MALDI MS measurements in positive and negative mode. Six single analytes (positive: *reserpine* (RP), tetraethylammonium chloride, coumaphos; negative: cholic acid, *citric acid* (CiA) and phenylalanine) and one complex sample (*Ginkgo Biloba* extract, in positive and negative mode) were prepared with the layered dried droplet method and then measured.

Figure 25: Reflectron positive mode MALDI MS spectra of spin coated samples: **PNDI(T2)** coated on RP (red), **P(TNDIT-FI(C₄C₂))** coated on RP (blue) and **P(TNDIT-FI(C₁₀C₈))** coated on RP (black). Peaks: $m/z = 395$ [RP3]⁺ (see SI); 397 [RP1=C₂₃H₂₉N₂O₄]⁺; 607 [RP-H]⁺. Laser intensity = 40%.^[6]

Figure 26: Reflectron negative mode MALDI MS spectra of spin-coated samples: **PNDI(T2)** coated on CiA (red), **P(TNDIT-FI(C₄C₂))** coated on CiA (blue) and **P(TNDIT-FI(C₁₀C₈))** coated on CiA (black). Peaks: $m/z = 26$ acetonitrile fragment [CN]⁻; 42 [C₃H₆]⁻; 57 [C₄H₈]⁻; 85 [C₆H₁₂]⁻; 87 [CiA1=C₃H₃O₃]⁻; 111 (unknown); 129 [CiA2=C₅H₅O₄]⁻; 155 (unknown); 191 [CiA-H]⁻. Laser intensity = 50%.^[6]

These measurements showed that both amorphous CPMs are well suited for dual polarity mode detection of LMWCs via MALDI MS. Furthermore, the absence of strong background interferences in the low mass range indicates that the NDI(T2) moiety stabilized the fluorene moiety, resulting in *MALDI silent* CPMs.

As spin-coating enables polymers to adopt a semi-crystalline morphology, spin-coated samples were prepared to allow a correlation with the GIWAXS results: RP, CiA and *Ginkgo Biloba*. Then a CPM, i.e., **P(TNDIT-FI(C₄C₂))**, **P(TNDIT-FI(C₁₀C₈))** or **PNDI(T2)**, was spin-coated on top of the analytes, resulting in a layered sample composition. The comparison of the amorphous co-polymers with the structurally related **PNDI(T2)** allowed to investigate the dependency of matrix performance and degree of crystallinity. Interestingly, the matrix performances of the amorphous co-polymers are comparable to the semi-crystalline **PNDI(T2)**, and in some cases, even higher signal intensities are detected with the amorphous CPMs.

The key findings of these investigations can be summarized as follows:

- The syntheses of two NDI(T2)/fluorene co-polymers via Suzuki-coupling were successful.
- Both co-polymers have an amorphous morphology, i.e., no crystalline domains.
- Both amorphous CPMs can be used as MALDI matrices; also, they allow to detect LMWCs in both polarity modes (dual polarity mode suitability).
- Both amorphous CPMs are *MALDI silent*, indicating that the NDI(T2) moiety stabilizes the fluorene moiety.
- The amorphous CPMs showed similar matrix performances compared to semi-crystalline parent homopolymer **PNDI(T2)**.

Hence, the matrix performance of a CPM seems to be independent from its degree of crystallinity. The results indicate that the analyte incorporation takes place in the amorphous domains of the CPMs and not as expected in the crystalline domains. Consequently, a semi-crystalline morphology or high degree of crystallinity seems not mandatory or beneficial for CPMs.

These results were published in the journal *ChemPlusChem* in 2019 (chapter 7.2).^[6]

4.4. Polymerization as a Strategy to Improve Small Organic Matrices for Low Molecular Weight Compound Analytics with MALDI MS and MALDI MS Imaging (Publication 3)

To explore the second strategy (ii) to PMs (chapter 3), the last objective of this thesis is the polymerization of SOMs to introduce them into the sidechains of a polymer. This approach potentially results in matrices combining the SOMs' broad analyte scopes and high ionization efficiencies with the PMs *MALDI silence* and vacuum stability. SOMs are standard MALDI matrices for the analysis of both HMWCs and LMWCs.^[12, 19] The advantages of SOMs (i.e., broad analyte scopes and high ionization efficiencies) are mostly caused by their excellent UV absorption due to conjugated systems in their molecular structure and their functionalization with acidic/basic groups.^[39, 50] On the other hand, PMs are intrinsically vacuum stable and potentially *MALDI silent* because of their high molar mass.

Figure 27: Synthesis of poly(vinyl dihydroxybenzoic acid) (P(VDHB)). (a) *n*-BuLi, DMF, Et₂O, 1h, 0 °C-rt (b) KMnO₄, acetone/H₂O, 24h, rt (c) H₂SO₄, methanol, 24h, 80 °C (d) Pd(PPh₃)₄, Na₂CO₃, dioxane/H₂O, 72h, 100 °C (e) KOH, acetone/H₂O, 24h, 100 °C (f) AIBN, dioxane, 7h, 90 °C (g) HBr, AcOH, 5h, 120 °C.^[7]

The advantages of these types of matrices are combined by modifying and polymerizing SOMs, resulting in polymers with the corresponding SOMs as sidechains (P(SOMs)). Here, vinyl groups were introduced to two standard SOMs 2,5-dihydroxybenzoic acid (DHB) and 7-methoxy-1-methyl-9H-pyrido[3,4-*b*]indole (harmine). The vinyl groups allowed radical polymerization to obtain the polyethylene-based polymers poly(vinyl dihydroxybenzoic acid) P(VDHB) and poly(vinyl harmine) P(VHar) (Figure 27 and Figure 28). All synthesized compounds were characterized via ¹H-/¹³C-NMR and FTIR analysis, and the P(SOMs) were further analyzed with MALDI MS and GPC to calculate their polymeric properties (*M_n*, *M_w*, *Đ*, and *DP*).

Figure 28: Synthesis of poly(vinyl harmine) (**P(VHar)**). (a) NBS, DCM, 24h, rt (b) $\text{Pd}(\text{PPh}_3)_4$, Na_2CO_3 , dioxane/ H_2O , 24h, 100 °C (c) AIBN, dioxane, 7h, 90 °C.^[7]

As expected from their high molar mass, the P(SOMs) were found to be fully vacuum stable (Figure 29, black and red plots). UV/Vis measurements of SOMs and P(SOMs) as thin layers indicated similar relative absorptions at the laser wavelength (355 nm).

Figure 29: Vacuum stability under HV ($\sim 10^{-7}$ mbar) over 24 h (average of 3 measurements, straight plots) and 72 h (single measurements, dashed plots) of DHB (yellow), **P(VDHB)** (red), harmine (blue) and **P(VHar)** (black). The balance error of 0.01 mg is visualized relative to the sample weight as error bars.^[7]

The matrix performance of both P(SOMs) regarding analyte scope, ionization efficiency and dual polarity mode suitability were tested and found to be comparable to their corresponding SOMs. In single analyte measurements, five commercially available lipids, five (oligo-)saccharides and two essential amino acids were analyzed with SOMs and P(SOMs) in both polarity modes, using equal measurement conditions and parameters (e.g., Figure 30, *myristic acid* (MA) and Figure 31, *fructose* (FT)).

Figure 30: Reflectron positive mode MALDI MS spectra of the sample plate (violet), the pure analyte MA (yellow), the pure DHB (green), DHB coated on MA (red), the pure *P*(VDHB) (blue), and *P*(VDHB) coated on MA (black). Peaks: $m/z = 23$ $[Na]^+$; 39 $[K]^+$; 57 $[MA1=C_4H_9]^+$; 88 $[MA2=C_4H_8O_2]^+$; 137 $[DHB-OH]^+$; 154 $[DHB]^+$; 177 $[DHB+Na]^+$; 193 $[DHB+K]^+$; 215 $[unknown\ DHB\ signal]^+$; 251 $[MA+Na]^+$; 267 $[MA+K]^+$. Laser intensity: 30%.^[7]

Figure 31: Reflectron negative mode MALDI MS spectra of the sample plate (violet), the pure analyte FT (yellow), the pure harmine (green), harmine coated on FT (red), the pure *P*(VHar) (blue), and *P*(VHar) coated on FT (black). Peaks: $m/z = 35$ $[Cl]^-$; 71 $[FT1=C_3H_3O_2]^-$; 101 $[FT2=C_4H_5O_3]^-$; 119 $[FT3=C_4H_7O_4]^-$; 130 $[FT4=C_6H_{10}O_3]^-$; 179 $[FT-H]^-$; 211 $[Har-H]^-$. Laser intensity: 30%.^[7]

Lipids were chosen as this analyte class is often investigated using DHB, while saccharides are typical targets for analyses using harmine, and amino acids are bio-relevant standard analytes of MALDI MS.

In addition to single analyte measurements, MALDI MSI measurements of breast cancer xenografts were carried out in both polarity modes (DHB_{positive}: Figure 32; harmine_{negative}: Figure 33) to evaluate the suitability of both P(SOMs) as matrices for clinical investigations. The resulting heat maps indicate a good extraction of small bio-analytes from the tissue by both P(SOMs).

Figure 32: Positive mode MALDI MSI measurements of a cryosection from a patient-derived breast cancer xenograft tumor using **P(VDHB)** as matrix. Top (from left to right): optical picture, ionic intensity heat maps of ions at m/z values 114.0, 147.0, 666.5, 762.4. Bottom (from left to right): ionic intensity heat maps of the ions at m/z values 794.4, 798.5, 874.4, 920.5.^[7]

Figure 33: Negative mode MALDI MSI measurements of a cryosection from a patient-derived breast cancer xenograft tumor using **P(VHar)** as matrix. Top (from left to right) optical pictures, ionic intensity maps of ions at m/z values 87.0, 89.1, 124.0, 145.1. Bottom (from left to right) ionic intensity heat maps of the ions at m/z values 283.2, 303.2, 835.4, 887.5.^[7]

The summarized results are:

- SOMs can be modified and polymerized to obtain P(SOMs) with inert backbones and functional sidechains.

- SOMs suffer from mass loss in HV ($\sim 10^{-7}$ mbar), while P(SOMs) fully retain their mass.
- Thin layers of SOMs and P(SOMs) have similar relative absorptions at the laser wavelength (355 nm).
- Matrix performances (i.e., analyte scope, ionization efficiency, dual polarity mode suitability) of P(SOMs) are comparable to the corresponding SOMs.
- P(SOMs) are fully *MALDI silent*: They show no matrix related signals in the low mass range.
- MSI heat maps of measurements using P(SOM)s indicate the successful extraction and ionization of small biomolecules on biological tissues in both polarity modes.

The results demonstrate that it is possible to combine of the advantages of SOMs and PMs by modifying and polymerizing high-performance SOMs. Also, this shows that not only polymers with conjugated backbones (CPMs) but also with conjugated sidechains (P(SOMs)) are suitable as MALDI matrices.

These results were published in the journal *ACS Applied Polymer Materials* in 2021 (chapter 7.3).^[7]

5. Conclusion & Outlook

The reliable and reproducible detection of LMWCs with MALDI MS and MALDI MSI is of interest for various medical and pathological research fields, such as metabolomics and lipidomics. However, standard matrix systems (small organic matrices, SOMs) suffer from several drawbacks, i.e., they produce high background interferences in the low mass range, often exhibit low vacuum stability, homogeneous thin-layer coating can be challenging, and they are rarely dual polarity mode suitable (chapter 2.1.3). These setbacks can reduce the spot-to-spot reproducibility of the MALDI measurements and can impede the comparison, e.g., of two MALDI MSI heat maps measured from different tissue samples. The development of matrices with enhanced performances might allow MALDI MS and MSI measurements without the mentioned drawbacks of SOMs. In this thesis, two different strategies to PMs were explored: (i) Investigating polymers with conjugated backbones and inert solubilizing sidechains, and (ii) modifying and polymerizing standard SOMs to obtain P(SOMs) with non-conjugated backbones and functional sidechains. Polymers are promising matrix candidates to enable reproducible LMWC analytics (chapter 2.2.2) as their increased molar mass ensures vacuum stability in HV and potentially allows *MALDI silent* measurements. Furthermore, their molecular structure can be modified, e.g., by introducing conjugated π -electron systems in the backbone or the sidechains to enhance UV absorption. Also, by adding acidic/basic functional groups, the ionization efficiency and dual polarity mode suitability can be possibly increased.

To explore the first strategy, different conjugated polymers were investigated as MALDI matrices for LMWC analytics (chapter 4.2). The conjugated backbone ensures a sufficient UV absorption, while the alkyl sidechains ensure solubility. Four of five investigated CPMs allowed LMWC detection in MALDI MS measurements of single and complex samples. They showed low background interferences (*MALDI silent*), were suitable for dual polarity mode measurements and their matrix performances were comparable to standard SOMs (DHB and 9AA). However, polymers can only adopt a semi-crystalline morphology, i.e., forming crystalline and amorphous domains. For SOMs, high crystallinity is described to be mandatory for a sufficient analyte incorporation and good matrix performance. Hence, for PMs, the analyte incorporation is expected to take place in the crystalline polymer parts, and the matrix performance is expected to be dependent on the degree of crystallinity. To investigate this, two fully amorphous

CPMs were synthesized, and their matrix performance compared to a structurally related semi-crystalline CPM (chapter 4.3). The results showed that also these amorphous CPMs are *MALDI silent*, dual polarity mode matrices, and that their matrix performance is comparable to the semi-crystalline CPM **PNDI(T2)**. This indicates that for CPMs, analyte incorporation takes place in the amorphous domains, and that – unlike for SOMs – for CPMs improving the crystallinity will not result in enhanced matrix performance.

The second strategy explored here was the polymerization of standard SOMs. SOMs have broad analyte scopes and high ionization efficiencies, while PMs are vacuum stable and expected to be *MALDI silent*. Two standard SOMs, DHB and harmine, were modified with vinyl groups and polymerized via radical polymerization. The resulting polyethylene-based polymers have chemically inert backbones and functional SOM sidechains. These P(SOMs) show similar matrix performances (i.e., analyte scopes, ionization efficiencies and dual polarity mode suitability) as the corresponding SOMs, while being *MALDI silent* and vacuum stable (chapter 4.4). Additionally, the P(SOMs) were successfully used for MALDI MSI measurements of breast cancer xenografts, showing excellent ion extraction from tissues in both polarity modes. Hence, the polymerization of SOMs is a promising strategy not only to improve SOMs but also to achieve high performing PMs for reproducible LMWC detection in MALDI MS and MALDI MSI measurements.

For comparing the matrix performances of the investigated PMs with standard matrices, the graphing software OriginPro was used, and a script was written to visualize, process, and evaluate the mass spectra (chapter 4.1).

However, measurements with CPMs as well as P(SOMs) showed some setbacks regarding important properties of MALDI matrices: the solubilizing alkyl sidechains of the CPMs can fragment at higher laser intensities, resulting in matrix related signals at $m/z \leq 85$ (chapter 4.2). The sidechains are important to ensure sufficient solubility in non-polar solvents, allowing wet sample preparation methods (chapter 2.1.1). To avoid fragmentation while ensuring high solubility, more stable sidechains might be used, e.g., polyethylene glycol. This might result in fully *MALDI silent* measurements, and furthermore, polyethylene glycol sidechains would also enable solubility in polar solvents. For matrix application during MALDI MSI tissue sample preparations, polar solvents can be preferable because of better ion extraction rates. Apart from the sidechains, the solubility of polymers is also influenced by the backbone length, which

can be adjusted by controlling the reaction conditions during the polymerization. Ideally, the PM's backbone is long enough to ensure *MALDI silence* and vacuum stability and short enough to guarantee high solubility for wet sample preparation. Especially *MALDI silent* properties of the PMs are important to avoid overlapping of matrix and analyte signals in the low mass range (chapter 2.1.3). Apart from solubility and *MALDI silence*, a further important property of PMs for MALDI MS and MSI is the scope of detectable bio-analytes. Although it was shown that various small bio-compounds ($M < 1000$ Da) are measurable using PMs, the detection of oligomeric and polymeric bio-compounds with larger molar masses ($M > 1000$; e.g., peptides and proteins), is still challenging. Hence, broadening the analyte scope towards larger compounds requires further investigations and optimizations of, e.g., the sample preparation methods or measurement parameters. Also, the utilization of shorter polymer chains might be advantageous.

Furthermore, the CPMs can be potentially improved by adding acidic/basic sidechains to increase the ionization efficiency. They can also be modified for enhanced detection of specific compounds, e.g., aldehyde-containing compounds like saccharides by introducing amino groups. Such functionalization could be performed, e.g., via click-chemistry, ideally allowing to choose a sidechain for a targeted analyte and click the sidechain to a standard polymer backbone, resulting in a specialized CPM.

Regarding the investigation on the MALDI process, a better understanding of the ionization mechanism of CPMs is desirable. So far, for small, conjugated oligomers, an analyte ionization via charge transfer was reported.^[4] Hence, this ionization mechanism is also likely for the presented CPMs (chapter 4.2 and 4.3), as they are only functionalized with solubilizing alkyl sidechains. Yet, the measurements carried out here indicate ionization via ion transfer, as the detected signals mainly originated from (de-)protonated analyte ions or cluster ions with sodium $[\text{Na}]^+$ or potassium $[\text{K}]^+$, as previously shown in the literature.^[3] A deeper investigation might contribute to a deeper understanding of the CPMs' ionization mechanisms during the MALDI process.

Finally, to allow more research on the scope and potential of PMs, the polymers need to be accessible for a broad research community in a sufficient quantity and consistent quality. This might be achieved by using simpler synthetic pathways and enabling straightforward upscaling.

6. List of References

- [1] B. Tieke, *Makromolekulare Chemie: Eine Einführung*, John Wiley & Sons **2005**, [9783527662289](#).
- [2] Y.-S. Lin and Y.-C. Chen, *Anal. Chem.* **2002**, *74*, 5793-5798, [10.1021/ac020418a](#).
- [3] L. J. Soltzberg and P. Patel, *Rapid Commun. Mass Spectrom.* **2004**, *18*, 1455-1458, [10.1002/rcm.1505](#).
- [4] A. Woldegiorgis, F. v. Kieseritzky, E. Dahlstedt, J. Hellberg, T. Brinck and J. Roeraade, *Rapid Commun. Mass Spectrom.* **2004**, *18*, 841–852, [10.1002/rcm.1412](#).
- [5] K. Horatz, M. Giampà, Y. Karpov, K. Sahre, H. Bednarz, A. Kiriya, B. Voit, K. Niehaus, N. Hadjichristidis, D. L. Michels and F. Lissel, *J. Am. Chem. Soc.* **2018**, *140*, 11416–11423, [10.1021/jacs.8b06637](#).
- [6] K. Horatz, K. Ditte, T. Prenveille, K.-N. Zhang, D. Jehnichen, A. Kiriya, B. Voit and F. Lissel, *ChemPlusChem* **2019**, *84*, 1338-1345, [10.1002/cplu.201900203](#).
- [7] K. Horatz, M. Giampà, Z. Qiao, S. A. Moestue and F. Lissel, *ACS Appl. Polym. Mater.* **2021**, *3*, 4234–4244, [10.1021/acsapm.1c00665](#).
- [8] M. Karas, D. Bachmann, U. Bahr and F. Hillenkamp, *Int. J. Mass Spectrom. Ion Processes* **1987**, *78*, 53-68, [10.1016/0168-1176\(87\)87041-6](#).
- [9] K. Tanaka, H. Waki, Y. Ido, S. Akita, Y. Yoshida and T. Yoshida, *Rapid Commun. Mass Spectrom.* **1988**, *2*, 151-153, [10.1002/rcm.1290020802](#).
- [10] R. C. Beavis and B. T. Chait, *Rapid Commun. Mass Spectrom.* **1989**, *3*, 436-439, [10.1002/rcm.1290031208](#).
- [11] R. M. Caprioli, T. B. Farmer and J. Gile, *Anal. Chem.* **1997**, *69*, 4751-4760, [10.1021/ac970888i](#).
- [12] J. L. Norris and R. M. Caprioli, *Chem. Rev.* **2013**, *113*, 2309-2342, [10.1021/cr3004295](#).
- [13] F. Hillenkamp, T. W. Jaskolla and M. Karas, *MALDI MS: A Practical Guide to Instrumentation, Methods, and Applications (The MALDI Process and Method)*, Wiley-VCH Verlag GmbH & Co. KGaA **2014**, [10.1002/9783527335961.ch1](#).
- [14] J. Leopold, Y. Popkova, K. M. Engel and J. Schiller, *Biomolecules* **2018**, *8*, [10.3390/biom8040173](#).
- [15] R. Knochenmuss, *MALDI Ionization Mechanisms: an Overview*, John Wiley & Sons, Inc. **2012**, 147–183, [9780470588901](#).

-
- [16] K. G. Owens and S. D. Hanton, *Conventional MALDI sample preparation*, 175, John Wiley & Sons, Inc. **2011**, [10.1002/9780470567234.ch6](https://doi.org/10.1002/9780470567234.ch6).
- [17] K. Kriegsmann, R. Longuespée, M. Hundemer, C. Zgorzelski, R. Casadonte, K. Schwamborn, W. Weichert, P. Schirmacher, A. Harms, D. Kazdal, J. Leichsenring, A. Stenzinger, A. Warth, M. Fresnais, J. Kriegsmann and M. Kriegsmann, *Proteom. Clin. Appl.* **2019**, *13*, [10.1002/prca.201800035](https://doi.org/10.1002/prca.201800035).
- [18] F. Tian, R. Liu, C. Fan, Y. Sun, X. Huang, Z. Nie, X. Zhao and X. Pu, *Metabolites* **2020**, *10*, *27*, [10.3390/metabo10010027](https://doi.org/10.3390/metabo10010027).
- [19] C. D. Calvano, A. Monopoli, T. R. I. Cataldi and F. Palmisano, *Anal. Bioanal. Chem.* **2018**, *410*, 4015–4038, [10.1007/s00216-018-1014-x](https://doi.org/10.1007/s00216-018-1014-x).
- [20] M. B. O'Rourke, S. P. Djordjevic and M. P. Padula, *Mass Spectrom. Rev.* **2018**, *37*, 217–228, [10.1002/mas.21515](https://doi.org/10.1002/mas.21515).
- [21] J. Axelsson, A.-M. Hoberg, C. Waterson, P. Myatt, G. L. Shield, J. Varney, D. M. Haddleton and P. J. Derrick, *Rapid Commun. Mass Spectrom.* **1997**, *11*, 209–213, [10.1002/\(SICI\)1097-0231\(19970131\)11:2<209::AID-RCM734>3.0.CO;2-E](https://doi.org/10.1002/(SICI)1097-0231(19970131)11:2<209::AID-RCM734>3.0.CO;2-E).
- [22] P. Huang, C.-Y. Huang, T.-C. Lin, L.-E. Lin, E. Yang, C. Lee, C.-C. Hsu and P.-T. Chou, *Anal. Chem.* **2020**, *92*, 7139–7145, [10.1021/acs.analchem.0c00570](https://doi.org/10.1021/acs.analchem.0c00570).
- [23] M. Nazari and D. C. Muddiman, *Analyst* **2016**, *121*, 595–605, [10.1039/c5an01513h](https://doi.org/10.1039/c5an01513h).
- [24] H.-Z. Tang, Y.-L. Ma, F. Liu, F. Liu, Z.-W. Liu, J.-W. Li, H.-Y. Zhou and Z.-X. Gao, *Int. J. Mass Spectrom.* **2017**, *417*, 34–39, [10.1016/j.ijms.2017.04.004](https://doi.org/10.1016/j.ijms.2017.04.004).
- [25] M. Mank, B. Stahl and G. Boehm, *Anal. Chem.* **2004**, *76*, 2938–2950, [10.1021/ac030354j](https://doi.org/10.1021/ac030354j).
- [26] H. He, L. Qin, Y. Zhang, M. Han, J. Li, Y. Liu, K. Qiu, X. Dai, Y. Li, M. Zeng, H. Guo, Y. Zhou and X. Wang, *Anal. Chem.* **2019**, *91*, 2634–2643, [10.1021/acs.analchem.8b03522](https://doi.org/10.1021/acs.analchem.8b03522).
- [27] F. O. Ayorinde, P. Hambricht, T. N. Porter and J. Q. L. Keith, *Rapid Commun. Mass Spectrom.* **1999**, *13*, 2474–2479, [10.1002/\(SICI\)1097-0231\(19991230\)13:24<2474::AID-RCM814>3.0.CO;2-0](https://doi.org/10.1002/(SICI)1097-0231(19991230)13:24<2474::AID-RCM814>3.0.CO;2-0).
- [28] H. N. Abdelhamida, J. Gopala and H.-F. Wu, *Anal. Chim. Acta* **2013**, *767*, 104–111, [10.1016/j.aca.2012.12.054](https://doi.org/10.1016/j.aca.2012.12.054).
- [29] M. C. Bernier, V. H. Wysocki and S. Dagan, *J. Mass Spectrom.* **2015**, *50*, 891–898, [10.1002/jms.3597](https://doi.org/10.1002/jms.3597).
-

-
- [30] X. Dong, J. Cheng, J. Li and Y. Wang, *Anal. Chem.* **2010**, 82, 6208-6214, [10.1021/ac101022m](#).
- [31] Y. Liu, J. Liu, P. Yin, M. Gao, C. Deng and X. J. Zhang, *Mass Spectrom.* **2011**, 46, 804-815, [10.1002/jms.1952](#).
- [32] R. Lv, R. Shi, E. Wu, C. Ma, R. Guo, J. Li, F. Ouyang, Q. Fang, L. Hu, C. Sun, Y. Liu, Z. Kang and J. Liu, *Talanta* **2020**, 220, 121371, [10.1016/j.talanta.2020.121371](#).
- [33] M. Karas, D. Bachmann and F. Hillenkamp, *Anal. Chem.* **1985**, 57, 2935-2939, [10.1021/ac00291a042](#).
- [34] A. Croxatto, G. Prod'hom and G. Greub, *FEMS Microbiol. Rev.* **2012**, 36, 380-407, [10.1111/j.1574-6976.2011.00298.x](#).
- [35] A. E. Clark, E. J. Kaleta, A. Arora and D. M. Wolk, *Clin. Microbiol. Rev.* **2013**, 26, 547-603, [10.1128/CMR.00072-12](#).
- [36] M. A. M. Rodrigo, O. Zitka, S. Krizkova, A. Moullick, V. Adam and R. Kizek, *J. Pharm. Biomed. Anal.* **2014**, 95, 245-255,
- [37] M. Karas and F. Hillenkamp, *Anal. Chem.* **1988**, 60, 2299-2301, [10.1021/ac00171a028](#).
- [38] M. Aichler and A. Walch, *Lab. Invest.* **2015**, 95, 422-431, [10.1038/labinvest.2014.156](#).
- [39] K. Strupat, M. Karas and F. Hillenkamp, *Int. J. Mass Spectrom. Ion Processes* **1991**, 111, 89-102, [10.1016/0168-1176\(91\)85050-V](#).
- [40] V. Horneffer, M. Glückmann, R. Krüger, M. Karas, K. Strupat and F. Hillenkamp, *Int. J. Mass Spectrom.* **2006**, 249-250, 426-432, [10.1016/j.ijms.2005.11.012](#).
- [41] I. K. Perera, J. Perkins and S. Kantartzoglou, *Rapid Commun. Mass Spectrom.* **1995**, 9, 180-187, [10.1002/rcm.1290090215](#).
- [42] R. C. Murphy, J. A. Hankin, R. M. Barkley and K. A. Z. Berry, *Biochim. Biophys. Acta* **2011**, 1811, 970-975, [10.1016/j.bbailip.2011.04.012](#).
- [43] O. Vorm, P. Roepstorff and M. Mann, *Anal. Chem.* **1994**, 66, 3281-3287, [10.1021/ac00091a044](#).
- [44] M. Heinrich, C. Schäffer, P. Messner and G. Allmaier, *Eur. J. Mass Spectrom.* **2008**, 14, 401-409, [10.1255/ejms.965](#).
- [45] Y. Dai, R. M. Whittall and L. Li, *Anal. Chem.* **1999**, 71, 1087-1091, [10.1021/ac980684h](#).
-

-
- [46] B. O. Keller and L. Li, *J. Am. Soc. Mass Spectrom.* **2006**, *17*, 780-785, [10.1016/j.jasms.2006.02.012](https://doi.org/10.1016/j.jasms.2006.02.012).
- [47] M. Kussmann, E. Nordhoff, H. Rahbek-Nielsen, S. Haebel, M. Rossel-Larsen, L. Jakobsen, J. Gobom, E. Mirgorodskaya, A. Kroll-Kristensen, L. Palmp and P. Roepstorff, *J. Mass Spectrom.* **1997**, *32*, 593-601, [10.1002/\(SICI\)1096-9888\(199706\)32:6<593::AID-JMS511>3.0.CO;2-D](https://doi.org/10.1002/(SICI)1096-9888(199706)32:6<593::AID-JMS511>3.0.CO;2-D).
- [48] M. A. R. Meier, R. Hoogenboom, M. W. M. Fijten, M. Schneider and U. S. Schubert, *J. Comb. Chem.* **2003**, *5*, 269-374, [10.1021/cc020101o](https://doi.org/10.1021/cc020101o).
- [49] M. Sadeghi, Z. Olumee, X. Tang, A. Vertes, Z.-X. Jiang, A. J. Henderson, H. S. Lee and C. R. Prasad, *Rapid Commun. Mass Spectrom.* **1997**, *11*, 393-397, [10.1002/\(SICI\)1097-0231\(19970228\)11:4<393::AID-RCM868>3.0.CO;2-S](https://doi.org/10.1002/(SICI)1097-0231(19970228)11:4<393::AID-RCM868>3.0.CO;2-S).
- [50] M. F. Mirabelli and R. Zenobi, *J. Am. Soc. Mass Spectrom.* **2017**, *28*, 1676-1686, [10.1007/s13361-017-1677-0](https://doi.org/10.1007/s13361-017-1677-0).
- [51] M. Karas and R. Krüger, *Chem. Rev.* **2003**, *103*, 427-439, [10.1021/cr010376a](https://doi.org/10.1021/cr010376a).
- [52] T. W. Jaskolla and M. Karas, *J. Am. Soc. Mass Spectrom.* **2011**, *22*, 976-988, [10.1007/s13361-011-0093-0](https://doi.org/10.1007/s13361-011-0093-0).
- [53] F. T. Vert, I. Z. Sanchez and A. O. Torrent, *J. Photochem.* **1983**, *23*, 355-368, [10.1016/0047-2670\(83\)87109-3](https://doi.org/10.1016/0047-2670(83)87109-3).
- [54] S.-P. Liang, I.-C. Lu, S.-T. Tsai, J.-L. Chen, Y. T. Lee and C.-K. Ni, *J. Am. Soc. Mass Spectrom.* **2017**, *28*, 2235-2245, [10.1007/s13361-017-1734-8](https://doi.org/10.1007/s13361-017-1734-8).
- [55] T. D. McCarley, R. L. McCarley and P. A. Limbach, *Anal. Chem.* **1998**, *70*, 4376-4379, [10.1021/ac980527i](https://doi.org/10.1021/ac980527i).
- [56] D. W. Koppenaal, C. J. Barinaga, M. B. Denton, R. P. Sperline, G. M. Hieftje, G. D. Schilling, F. J. Andrade and J. H. Barnes, *Anal. Chem.* **2005**, *77*, 418-427, [10.1021/ac053495p](https://doi.org/10.1021/ac053495p).
- [57] S. Hosseini and S. O. Martinez-Chapa, *Fundamentals of MALDI-ToF-MS Analysis - Applications in Biodiagnosis, Tissue Engineering and Drug Delivery*, Springer **2017**, [978-981-10-2356-9](https://doi.org/10.1007/978-981-10-2356-9).
- [58] P. Chaurand, M. Stoeckli and R. M. Caprioli, *Anal. Chem.* **1999**, *71*, 5263-5270, [10.1021/ac990781q](https://doi.org/10.1021/ac990781q).
- [59] C. Eriksson, N. Masaki, I. Yao, T. Hayasaka and M. Setou, *Mass Spectrom.* **2013**, *2*, S0022, [10.5702/massspectrometry.S0022](https://doi.org/10.5702/massspectrometry.S0022).
- [60] S. Shimma and Y. Sugiura, *Mass Spectrom.* **2014**, *3*, S0029, [10.5702/massspectrometry.s0029](https://doi.org/10.5702/massspectrometry.s0029).
-

-
- [61] H.-Y. J. Wang, S. N. Jackson, J. Post and A. S. Woods, *Int. J. of Mass Spec.* **2008**, 278, 143-149, [10.1016/j.ijms.2008.04.005](https://doi.org/10.1016/j.ijms.2008.04.005).
- [62] D. L. Baluya, T. J. Garrett and R. A. Yost, *Anal. Chem.* **2007**, 79, 6862-6867, [10.1021/ac070958d](https://doi.org/10.1021/ac070958d).
- [63] P. Juhasz, C. E. Costello and K. Biemann, *J. Am. Soc. Mass. Spectrom.* **1993**, 4, 399-409, [10.1016/1044-0305\(93\)85005-I](https://doi.org/10.1016/1044-0305(93)85005-I).
- [64] J. H. Gross, *Mass Spectrometry: a textbook* (2. ed), Springer **2011**, [9783642107092](https://doi.org/10.1007/9783642107092).
- [65] T. Fukai, J. Kuroda and T. Nomura, *J. Am. Soc. Mass Spectrom.* **2000**, 11, 458–463, [10.1016/S1044-0305\(00\)00109-4](https://doi.org/10.1016/S1044-0305(00)00109-4).
- [66] B. A. Boughton, O. R. B. Thomas, N. J. Demarais, D. Trede, S. E. Swearer and A. C. Grey, *J. Mass Spectrom.* **2020**, 55, e4460, [10.1002/jms.4460](https://doi.org/10.1002/jms.4460).
- [67] J. S. H. Wang, M. Freitas-Andrade, J. F. Bechberger, C. C. Naus, K. K. C. Yeung and S. N. Whitehead, *Rapid Commun. Mass Spectrom.* **2018**, 32, 951–958, [10.1002/rcm.8115](https://doi.org/10.1002/rcm.8115).
- [68] S. Schulz, M. Becker, M. R. Groseclose, S. Schadt and C. Hopf, *Curr. Opin. Biotechnol.* **2019**, 55, 51-59, [10.1016/j.copbio.2018.08.003](https://doi.org/10.1016/j.copbio.2018.08.003).
- [69] Y. Chen, W. Tang, A. Gordon and B. Li, *J. Am. Soc. Mass Spectrom.* **2020**, 31, 1066–1073, [10.1021/jasms.0c00003](https://doi.org/10.1021/jasms.0c00003).
- [70] S. Guo, W. Tang, Y. Hu, Y. Chen, A. Gordon, B. Li and P. Li, *Anal. Chem.* **2020**, 92, 1431–1438, [10.1021/acs.analchem.9b04618](https://doi.org/10.1021/acs.analchem.9b04618).
- [71] M. Dufresne, N. H. Patterson, J. L. Norris and R. M. Caprioli, *Anal. Chem.* **2019**, 91, 12928–12934, [10.1021/acs.analchem.9b02974](https://doi.org/10.1021/acs.analchem.9b02974).
- [72] F. Xiang and R. C. Beavis, *Org. Mass Spec.* **1993**, 28, 1424-1429, [10.1002/oms.1210281211](https://doi.org/10.1002/oms.1210281211).
- [73] R. L. Vermillion-Salsbury and D. M. Hercules, *Rapid Commun. Mass Spectrom.* **2002**, 16, 1575-1581, [10.1002/rcm.750](https://doi.org/10.1002/rcm.750).
- [74] R. C. Beavis, T. Chaudhary and B. T. Chait, *Organic Mass Spectrometry* **1992**, 27, 156-158, [10.1002/oms.1210270217](https://doi.org/10.1002/oms.1210270217).
- [75] H. Nonami, S. Fukui and R. Erra-Balsells, *J. Mass Spectrom.* **1997**, 32, 287-296, [10.1002/\(SICI\)1096-9888\(199703\)32:3<287::AID-JMS478>3.0.CO;2-%23](https://doi.org/10.1002/(SICI)1096-9888(199703)32:3<287::AID-JMS478>3.0.CO;2-%23).
- [76] Y. V. Vasil'ev, O. G. Khvostenko, A. V. Streletskii, O. V. Boltalina, S. G. Kotsiris and T. Drewello, *J. Phys. Chem. A.* **2006**, 110, 5967–5972, [10.1021/jp060568f](https://doi.org/10.1021/jp060568f)
-

-
- [77] W. J. Perry, N. H. Patterson, B. M. Prentice, E. K. Neumann, R. M. Caprioli and J. M. Spraggins, *J. Mass Spectrom.* **2020**, 55, 4491, [10.1002/jms.4491](#).
- [78] S. Bourcier, S. Bouchonnet and Y. Hoppilliard, *Int. J. Mass Spectrom. Ion Processes* **2001**, 210/211, 59-69, [10.1016/s1387-3806\(01\)00446-8](#).
- [79] S.-S. Choi and S.-H. Ha, *Bull. Korean Chem. Soc.* **2007**, 28, 2508-2510, [10.5012/bkcs.2007.28.12.2508](#).
- [80] A. Lee, H.-J. Yang, Y. Kim and J. Kim, *Bull. Korean Chem. Soc.* **2009**, 30, 1127-1130, [10.5012/bkcs.2009.30.5.1127](#).
- [81] J. Schiller, R. Süß, B. Fuchs, M. Müller, M. Petkovic, O. Zschörnig and H. Waschipky, *Eur. Biophys. J.* **2007**, 36, 517–527, [10.1007/s00249-006-0090-6](#).
- [82] Y. Wei, Y. Zhang, Y. Lin, L. Li, J. Liu, Z. Wang, S. Xiong and Z. Zhao, *Analyst* **2015**, 140, 1298-1305, [10.1039/c4an01964d](#).
- [83] B. O. Keller, J. Suib, A. B. Young and R. M. Whittall, *Anal. Chim. Acta* **2008**, 627, 71-81, [10.1016/j.aca.2008.04.043](#).
- [84] M. Giampa, M. B. Lissel, T. Patschkowski, J. Fuchser, V. H. Hans, O. Gembruch, H. Bednarza and K. Niehaus, *Chem. Commun.* **2016**, 52, 9801-9804, [10.1039/c6cc02387h](#).
- [85] C. Janfelt, N. Wellner, H. S. Hansen and S. H. Hansen, *J. Mass Spectrom.* **2013**, 48, 361–366 [10.1002/jms.3166](#).
- [86] Z. Qiao and F. Lissel, *Chem. Asian J.* **2021**, 16, 1-12, [10.1002/asia.202100044](#)
- [87] C. Jun, H. Tianxi and F. Xuemei, *Talanta* **2012**, 100, 419–424, [10.1016/j.talanta.2012.08.011](#).
- [88] M.-C. Tseng, R. Obena, Y.-W. Lu, P.-C. Lin, P.-Y. Lin, Y.-S. Yen, J.-T. Lin, L.-D. Huang, K.-L. Lu, L.-L. Lai, C.-C. Lin and Y.-J. Chen, *J. Am. Soc. Mass Spectrom.* **2010**, 21, 1930–1939, [10.1016/j.jasms.2010.08.001](#).
- [89] J. Son and S. Cha, *Bull. Korean Chem. Soc.* **2014**, 35, 1409-1412, [10.5012/bkcs.2014.35.5.1409](#).
- [90] A. Monopoli, C. D. Calvano, A. Nacciab and F. Palmisano, *Chem. Commun.* **2014**, 50, 4322, [10.1039/c4cc01185f](#).
- [91] M. L. Manier, J. M. Spraggins, M. L. Reyzer, J. L. Norris and R. M. Caprioli, *J. Mass Spectrom.* **2014**, 49, 665-673, [10.1002/jms.3411](#).
- [92] T. Fujita and T. Fujino, *Chem. Lett.* **2013**, 42, 350-351, [10.1246/cl.121232](#).
- [93] Q. Zhou, A. Fülöp and C. Hopf, *Anal. Bioanal. Chem.* **2020**, [10.1007/s00216-020-03023-7](#).
-

-
- [94] H. Zhao, Y. Li, J. Wang, M. Cheng, Z. Zhao, H. Zhang, C. Wang, J. Wang, Y. Qiao and J. Wang, *Appl. Mater. Interfaces* **2018**, *10*, 37732–37742, [10.1021/acsami.8b14643](https://doi.org/10.1021/acsami.8b14643).
- [95] N. Dittrich, J. Zuber, P. Rathack and M. Otto, *Int. J. Mass Spectrom. Ion Processes* **2018**, *433*, 31–39, [10.1016/j.ijms.2018.08.002](https://doi.org/10.1016/j.ijms.2018.08.002).
- [96] Y. Yang and Y. Xia, *Microchim. Acta* **2020**, *187*, 445, [10.1007/s00604-020-04397-2](https://doi.org/10.1007/s00604-020-04397-2).
- [97] H. Liu, Y. Zhou, J. Wang, C. Xiong, J. Xue, L. Zhan and Z. Nie, *Anal. Chem.* **2018**, *90*, 729–736, [10.1021/acs.analchem.7b02710](https://doi.org/10.1021/acs.analchem.7b02710).
- [98] L. Ling, C. Xiao, Y. Ma, L. Jiang, S. Wang, L. Guo, S. Jiang and X. Guo, *Anal. Chem.* **2019**, *91*, 8801–8807, [10.1021/acs.analchem.9b01434](https://doi.org/10.1021/acs.analchem.9b01434).
- [99] B. Li, R. Sun, A. Gordon, J. Ge, Y. Zhang, P. Li and H. Yang, *Anal. Chem.* **2019**, *91*, 8221–8228, [10.1021/acs.analchem.9b00803](https://doi.org/10.1021/acs.analchem.9b00803).
- [100] F. O. Ayorinde, K. Garvin and K. Saeed, *Rapid Commun. Mass Spectrom.* **2000**, *14*, 608–615, [10.1002/\(SICI\)1097-0231\(20000415\)14:7<608::AID-RCM918>3.0.CO;2-4](https://doi.org/10.1002/(SICI)1097-0231(20000415)14:7<608::AID-RCM918>3.0.CO;2-4).
- [101] F. O. Ayorinde, D. Z. Bezabeh and I. G. Delves, *Rapid Commun. Mass Spectrom.* **2003**, *17*, 1735–1742, [10.1002/rcm.1113](https://doi.org/10.1002/rcm.1113).
- [102] Y.-T. Chen and Y.-C. Ling, *J. Mass Spectrom.* **2002**, *37*, 716–730, [10.1002/jms.332](https://doi.org/10.1002/jms.332).
- [103] S. Zhang, J. Liu, Y. Chen, S. Xiong, G. Wang, J. Chen and G. Yang, *J. Am. Soc. Mass Spectrom.* **2010**, *21*, 154–160, [10.1016/j.jasms.2009.09.024](https://doi.org/10.1016/j.jasms.2009.09.024).
- [104] S. Seiffert, C. Kummerlöwe and N. Vennemann, *Lechner · Gehrke · Nordmeier - Makromolekulare Chemie*, *6*, Springer-Verlag GmbH **2020**, [978-3-662-61109-8](https://doi.org/10.1007/978-3-662-61109-8).
- [105] R. Giri, L. Nayak and M. Rahaman, *Mater. Res. Innov.* **2021**, *25*, 104–132, [10.1080/14328917.2020.1728073](https://doi.org/10.1080/14328917.2020.1728073).
- [106] X. Tang, Y. H. Lui, A. R. Merhi, B. Chen, S. Ding, B. Zhang and S. Hu, *Appl. Mater. Interfaces* **2017**, *9*, 44429–44440, [10.1021/acsami.7b1184](https://doi.org/10.1021/acsami.7b1184).
- [107] J. Huang and G. Yu, *Mater. Chem. Front.* **2021**, *5*, 76, [10.1039/d0qm00509f](https://doi.org/10.1039/d0qm00509f).
- [108] J. Mei and Z. Bao, *Chem. Mater.* **2014**, *26*, 604–615, [10.1021/cm4020805](https://doi.org/10.1021/cm4020805).
- [109] Y.-h. Shin, H. Komber, D. Caiola, M. Cassinelli, H. Sun, D. Stegerer, M. Schreiter, K. Horatz, F. Lissel, X. Jiao, C. R. McNeill, S. Cimò, C. Bertarelli, S.
-

- Fabiano, M. Caironi and M. Sommer, *Macromolecules* **2020**, 53, 5158-5168, [10.1021/acs.macromol.0c00657](https://doi.org/10.1021/acs.macromol.0c00657).
- [110] K. Norrman, A. Ghanbari-Siahkali and N. B. Larsen, *Annu. Rep. Prog. Chem.* **2005**, 101, 174-201, [10.1039/b408857n](https://doi.org/10.1039/b408857n).
- [111] K. Ditte, J. Perez, S. Chae, M. Hambsch, M. Al-Hussein, H. Komber, P. Formanek, S. C. B. Mannsfeld, A. Fery, A. Kiriya and F. Lissel, *Adv. Mater.* **2020**, 2005416, [10.1002/adma.202005416](https://doi.org/10.1002/adma.202005416).
- [112] A. Woldegiorgis, P. Löwenhielm, A. Björk and J. Roeraade, *Rapid Commun. Mass Spectrom.* **2004**, 18, 2904–2912, [10.1002/rcm.1707](https://doi.org/10.1002/rcm.1707).
- [113] *flexAnalysis 3.4 User Manual*, Bruker Daltonik GmbH, **2011**, <https://researchservices.pitt.edu/sites/default/files/flexAnalysis%20User%20Manual.pdf>, accessed at **25.01.2021**.
- [114] <http://www.sing-group.org/mass-up/index.php>, accessed at **25.01.2021**.
- [115] <https://lablicate.com/platform/openchrom>, accessed at **25.01.2021**.
- [116] P. Wenig and J. Odermatt, *Bioinformatics* **2010**, 11, 405, [10.1186/1471-2105-11-405](https://doi.org/10.1186/1471-2105-11-405).
- [117] H. López-Fernández, H. M. Santos, J. L. Capelo, F. Fdez-Riverola, D. Glez-Peña and M. Reboiro-Jato, *Bioinformatics* **2015**, 16, 318, [10.1186/s12859-015-0752-4](https://doi.org/10.1186/s12859-015-0752-4).
- [118] D. Blake, *Origin 6.0 Reviewer's Guide*, Microcal Software, Inc., **1998**, <https://www.originlab.com/pdfs/revguide.pdf>, accessed at **26.01.2021**.
- [119] <https://www.originlab.com/pdfs/RevGuide.pdf>, accessed at **02.12.2020**.
- [120] <https://www.originlab.com>, accessed at **02.12.2020**.

7. Publications

7.1. Conjugated Polymers as a New Class of Dual-Mode Matrices for MALDI Mass Spectrometry and Imaging (Publication 1)

September 12, 2018
Volume 140
Number 36
pubs.acs.org/JACS

J | A | C | S
JOURNAL OF THE AMERICAN CHEMICAL SOCIETY

Conjugated Polymers as a New Class of Dual-Mode Matrices for MALDI Mass Spectrometry and Imaging

Kilian Horatz,^{†,‡} Marco Giampà,[§] Yevhen Karpov,[†] Karin Sahre,[†] Hanna Bednarz,[§] Anton Kiri,^{†,§} Brigitte Voit,^{†,‡} Karsten Niehaus,[§] Nikos Hadjichristidis,[#] Dominik L. Michels,[⊥] and Franziska Lissel^{*,†,§}

[†]Institute of Macromolecular Chemistry, Leibniz-Institut für Polymerforschung Dresden e.V., Hohe Straße 6, 01069 Dresden, Germany

[‡]Organic Chemistry of Polymers, Technische Universität Dresden, Mommsenstrasse 4, 01062 Dresden, Germany

[§]Center for Biotechnology and Department for Proteome and Metabolome Research, Faculty of Biology, Bielefeld University, Universitätsstraße 25, 33615 Bielefeld, Germany

[#]Physical Sciences and Engineering Division, KAUST Catalysis Center, Polymer Synthesis Laboratory, King Abdullah University of Science and Technology (KAUST), Thuwal 23955, Kingdom of Saudi Arabia

[⊥]Computer, Electrical and Mathematical Sciences and Engineering Division, KAUST Visual Computing Center, Computational Sciences Group, King Abdullah University of Science and Technology (KAUST), Thuwal 23955, Kingdom of Saudi Arabia

Supporting Information

ABSTRACT: Matrix-assisted laser desorption/ionization mass spectrometry (MALDI MS) and MALDI MS imaging are ubiquitous analytical methods in medical, pharmaceutical, biological, and environmental research. Currently, there is a strong interest in the investigation of low molecular weight compounds (LMWCs), especially to trace and understand metabolic pathways, requiring the development of new matrix systems that have favorable optical properties and a high ionization efficiency and that are MALDI silent in the LMWC area. In this paper, five conjugated polymers, poly{[N,N'-bis(2-octyldodecyl)-naphthalene-1,4,5,8-bis(dicarboximide)-2,6-diyl]-alt-5,5'-(2,2'-bithiophene)} (PNDI(T2)), poly(3-dodecylthiophene-2,5-diyl) (P3DDT), poly{[2,3-bis(3-octyloxyphenyl)quinoxaline-5,8-diyl]-alt-(thiophene-2,5-diyl)} (PTQ1), poly{[N,N'-bis(2-octyldodecyl)-isoindigo-5,5'-diyl]-alt-5,5'-(2,2'-bithiophene)} (PII(T2)), and poly(9,9-di-*n*-octylfluorenyl-2,7-diyl) (P9OFI) are investigated as matrices. The polymers have a strong optical absorption, are solution processable, and can be coated into thin films, allowing a vast reduction in the amount of matrix used. All investigated polymers function as matrices in both positive and negative mode MALDI, classifying them as rare dual-mode matrices, and show a very good analyte ionization ability in both modes. PNDI(T2), P3DDT, PTQ1, and PII(T2) are MALDI silent in the full measurement range ($m/z > 150k$), except at high laser intensities. In MALDI MS experiments of single analytes and a complex biological sample, the performance of the polymers was found to be as good as two commonly used matrices (2,5-DHB for positive and 9AA for negative mode measurements). The detection limit of two standard analytes was determined as being below 164 pmol for reserpine and below 245 pmol for cholic acid. Additionally P3DDT was used successfully in first MALDI MS imaging experiments allowing the visualization of the tissue morphology of rat brain sections.

INTRODUCTION

MALDI MS and MALDI MS Imaging. Since the development of “matrix-assisted laser desorption/ionization mass spectrometry” (MALDI MS) in the 1980s,^{1–3} this type of mass spectrometry evolved into an important analysis tool for medical and biochemical investigations.^{4–8} Compared with other types of MS, e.g., electrospray ionization (ESI), secondary ion mass spectrometry (SIMS), or fast atom bombardment (FAB), MALDI MS exhibits several advantages: As the analyte is integrated into a protecting matrix, it is softly ionized, and high molecular weight compounds (e.g., proteins,

lipids, synthetic polymers) can be measured without destroying or fragmentizing them.^{9,10} The method does not require the labeling of the analytic targets and permits measurements down to attomolar concentrations.

MALDI mass spectrometry imaging (MS imaging) is an emerging clinicopathological imaging tool based on MALDI MS.^{11–13} It enables determining and visualizing the spatial distribution of molecules in a tissue sample, for example, in a

Received: June 23, 2018

Published: August 8, 2018

Scheme 1. Molecular Structures of (a) Poly{[N,N'-bis(2-octyldodecyl)-naphthalene-1,4,5,8-bis(dicarboximide)-2,6-diyl]-alt-5,5'-(2,2'-bithiophene)} (PNDI(T2)), (b) Poly(3-dodecylthiophene-2,5-diyl) (P3DDT), (c) Poly{[2,3-bis(3-octyloxyphenyl)quinoxaline-5,8-diyl]-alt-(thiophene-2,5-diyl)} (PTQ1), (d) Poly{[N,N'-bis(2-octyldodecyl)-isoindigo-5,5'-diyl]-alt-5,5'-(2,2'-bithiophene)} (PII(T2)), and (e) Poly(9,9-di-n-octylfluorenyl-2,7-diyl) (P9OFI)

cancer section, and is a versatile method with a high potential for medical,^{14,15} pharmaceutical,¹⁶ and environmental¹⁷ research. MALDI MS imaging and MALDI MS share the same working principles, but where MALDI MS establishes which molecular masses are present in the sample as a whole, MALDI MS imaging introduces spatial resolution by coating the matrix on the surface of a sample and then taking a series of measurements along a predefined grid. Each collected spectrum gives a detailed summary of the molecules present at the grid point, and combining the spectra allows the visualization of concentration gradients of molecules of interest and the creation of molecular maps of the measured tissue. It is also possible to construct 3D molecular maps by dividing an organ into several slices and merging all recorded spectra.¹⁸

Currently there is a drive to establish MALDI protocols for the analysis of low molecular weight compounds (LMWCs).¹⁷ This interest can be attributed to intensified efforts to study the metabolic profile of biological and pathological samples. The bulk of metabolites have a low molecular weight, and tracing their occurrence and composition is an important angle in the research on the formation and progression of diseases such as cancer,¹⁶ in drug development,¹⁵ or for the detection¹⁷ of pollutants and toxins.

Matrix Requirements. Choosing the right matrix for a given sample or analyte is a crucial step in the MALDI process. The ideal matrix has a strong optical absorption in the spectral region of the commonly used UV and IR lasers and a high ionization efficiency and is “MALDI silent”, i.e., does not lead to matrix-related peaks in the spectrum.^{12,19}

A MALDI MS imaging matrix ideally meets two additional conditions.^{12,20} (1) Homogenous matrix layer: The commonly used crystalline matrices need to exhibit a crystal density sufficient to allow measurements of the complete surface without gap regions, and the individual crystal has to be small relative to the grid size to avoid creating images that contain artifacts, as the lateral resolution is limited by the size of the matrix crystals. (2) Increased vacuum stability: Due to the high number of collected spectra, MALDI MS imaging measurements can take several hours to complete, especially in high-mass and/or high lateral resolution modes. To ensure comparability between different samples and also different positions on the same sample, the matrix cover not only needs to be highly uniform but also needs to be fully stable in ultrahigh vacuum (UHV) over an extended period of time.

Current state-of-the-art matrices were developed for specific analytes in regular MALDI MS using a “trial-and-error” approach.^{12,19} They are usually small molecules with conjugated systems and acidic or basic functional groups, e.g., 2,5-dihydroxybenzoic acid (2,5-DHB),²¹ 3,5-dimethoxy-4-hydroxycinnamic acid (sinapinic acid), α -cyano-4-hydroxycinnamic acid (α -CHCA), and so on. These small-molecule matrices desorb and ionize themselves, leading to matrix-related signals in the low molecular weight part of the MS spectrum. Due to the commonly used high matrix/analyte ratio, the matrix-related peaks can be intense and the resulting spectral interference as well as detector saturation often impedes the analysis of LMWCs.¹⁹ Several methods were explored to avoid the interferences from the matrix in the low-mass range. One is to change the ratio of analyte and matrix in order to have a complete suppression of matrix peaks;^{22,23} another way is to derivatize selected targets to shift their mass to a higher mass range where matrix-related peaks are not present.^{24–26}

Most well-established matrices are suitable for either positive or negative mode measurements, but not for both. For samples available only in low quantity, e.g., fine needle biopsies, or for studies investigating numerous samples in high-throughput analyses, it would be furthermore advantageous to have matrix systems available that are “dual mode” and allow high-quality measurements in both positive and negative mode.²⁷ As yet, only a few matrices are known that enable such dual-mode measurements.^{28–30} Furthermore, some of the small molecular matrices exhibit a comparably low vacuum stability, which influences the homogeneity of the layer especially during long-term MALDI MS imaging measurements and can falsify the resulting ion maps.

Consequently the development of new matrix systems is necessary to solve these problems.^{19,31,32} Rational design and synthesis of new crystalline small-molecule matrices as well as the deployment of novel compound classes, e.g., graphene,^{33,34} graphene oxide,^{34,35} nanoparticles,³⁶ metal oxides,³⁷ or ionic liquids,³⁸ are currently explored.

Conjugated Polymers as Matrices. MALDI MS is a standard tool for the analysis of polymeric materials.^{39–41} The utilization of polymers as MALDI matrices instead of small molecules could solve the described matrix problems: Conjugated polymers have a large π -system enabling them to interact with light of different wavelengths, which allows a versatile utilization of the matrices with different instruments

Table 1. UV/Vis Absorption of the Investigated Polymers

	PNDI(T2)	P3DDT	PTQ1	PII(T2)	P9OFI
absorption maxima [nm] ^{a,b}	392 (100%), 697 (81%)	523 (100%), 549 (95%), 601 (61%)	363 (95%), 622 (100%)	413 (67%), 626 (100%), 686 (89%)	386 (100%), 432 (13%)
relative absorption at 355 nm ^b	82%	23%	93%	48%	72%

^aUV/vis spectra of drop-casted films (see SI). ^bRelative to the maximum absorption of each compound.

and laser wavelengths. Due to their high molecular weight and chemical stability, they are nonvolatile and potentially allow measurement of spectra without matrix peaks (MALDI silent). Standard conjugated polymers used in organic electronics carry alkyl side chains, are soluble in organic solvents, and can be coated into semicrystalline thin films with a surface roughness in the low nanometer range using established methods,^{42,43} enabling homogeneous and defined matrix layers without gap regions. The molecular structure of the polymer backbone influences the physical properties (e.g., optical absorption, vacuum stability), and the polymer side chains can be used to modulate the chemical properties (e.g., solubility, analyte extraction ability). The large variability of the backbone and the possibility to introduce functional side chains, together with their overall ease of processing, make conjugated polymers very promising candidates to realize a novel class of functional high-performance matrices for MALDI MS and MALDI MS imaging. Interestingly, the research into polymers as functional matrix systems is so far limited to one report using poly(3-octylthiophene-2,5-diyl) for the measurements of acids.¹⁴ Investigations of other polymer systems have not been reported, and the potential of polymers as functional dual-mode matrices as well as the measurement of complex samples was never explored. Also, to the best of our knowledge, polymer matrices have never been used in MALDI MS imaging.

In this paper, five different conjugated polymers are investigated regarding their suitability as matrices for both negative and positive MALDI measurements. The polymers (see Scheme 1) poly{[N,N'-bis(2-octyldodecyl)-naphthalene-1,4,5,8-bis(dicarboximide)-2,5-diyl]-alt-5,5'-(2,2'-bithiophene)} (PNDI(T2)), poly(3-dodecylthiophene-2,5-diyl) (P3DDT), poly{[2,3-bis(3-octyloxyphenyl)quinoxaline-5,8-diyl]-alt-(thiophene-2,5-diyl)} (PTQ1), poly{[N,N'-bis(2-octyldodecyl)-isoindigo-5,5'-diyl]-alt-5,5'-(2,2'-bithiophene)} (PII(T2)), and poly(9,9-di-*n*-octylfluorenyl-2,7-diyl) (P9OFI) are well-explored semiconducting polymers. UV/vis spectra of drop-casted polymer films are recorded to determine the absorption at the MALDI laser wavelength, and the polymers are measured as analytes in MALDI experiments with and without secondary matrix to investigate their behavior. Then the performance of the polymers as matrices is tested in MALDI measurements of different pure LMWC analytes as well as of a complex biological sample and compared against two well-established matrices, 2,5-DHB for positive mode and 9AA for negative mode MALDI MS. The detection limit of standard analytes in both positive and negative mode is determined. Finally, P3DDT is tested in MALDI MS imaging experiments.

EXPERIMENTAL SECTION

MALDI MS spectra were recorded with a Bruker Autoflex speed MALDI-TOF/TOF equipped with a Smartbeam-II Nd:YAG laser (355 nm). The software tools flexControl 3.3 and flexAnalysis 3.3 were used. All MS measurements of analytes were carried out in reflectron mode, and the instrument was calibrated with red

phosphorus⁴⁵ at the beginning of each measurement cycle (see Supporting Information for further details). PNDI(T2)⁴² and PII(T2)⁴⁶ were synthesized in our laboratory; all other compounds are commercially available. All polymers except PTQ1 were purified with Soxhlet extraction before use. Layers of analytes and matrices were drop-coated from suitable solvents using the thin layer method, a variation of the dried droplet method (see Supporting Information).^{47,48} The polymers were coated from toluene solutions. For each matrix the best layer thickness combination was qualitatively determined by testing several combinations of analyte and matrix concentration (see Supporting Information for further details). The MALDI MS imaging experiments were carried out in negative reflectron mode on a Bruker ultrafleXtreme MALDI-TOF instrument equipped with a Smartbeam-II Nd:YAG laser (355 nm) and using flexImaging 4.1 software. P3DDT was spray-coated from an acetonitrile/chloroform solution using a Bruker ImagePrep instrument. All experimental details are described in the Supporting Information.

RESULTS AND DISCUSSION

Favorable optical properties are one of the basic prerequisites for MALDI matrices, and small-molecule matrices are often chosen based on an absorption maximum close to the laser wavelength (e.g., 2,5-DHB: local maximum at 324 nm with 18% of $\lambda_{\max}$ = 208 nm).^{21,49} All investigated polymers show good to very high relative absorptions $A_{\text{Nd:YAG}}$ at 355 nm, the wavelength of the commonly used Nd:YAG laser (Table 1 and Figure S4), thereby meeting the optical criteria. PNDI(T2) ($\lambda_{\max}$ = 392 nm, $A_{\text{Nd:YAG}}$ = 82%), PTQ1 ($\lambda_{\max}$ = 622 nm, $A_{\text{Nd:YAG}}$ = 93%), and P9OFI ($\lambda_{\max}$ = 386 nm, $A_{\text{Nd:YAG}}$ = 72%) have very good optical properties with regard to the wavelength of the Nd:YAG laser, while PII(T2) ($\lambda_{\max}$ = 626 nm, $A_{\text{Nd:YAG}}$ = 48%) and P3DDT ($\lambda_{\max}$ = 523 nm, $A_{\text{Nd:YAG}}$ = 23%) absorb less strongly, but still sufficiently well.

MALDI experiments can be carried out in linear or reflectron mode. The former allows measuring up to high molecular weights, but enables only a limited resolution and is commonly used for the analysis of polymers, while the reflectron mode permits a higher resolution but a limited m/z range and is preferred for metabolomics and peptidomics analyses and for MALDI MS imaging experiments.

To analyze the integrity of polymer structure, MALDI measurements of the polymers as analytes were carried out in positive linear mode using *trans*-2-[3-(4-*tert*-butylphenyl)-2-methyl-2-propenylidene]malononitrile (DCTB) as secondary matrix under the addition of 1 vol % of sodium trifluoroacetate (NaTFA), and the resulting spectra show typical patterns in the range up to m/z = 14k (Figures S78–S87) associated with the molecular mass distributions. Please note that in the corresponding figures the range below m/z = 1000 is deliberately not shown, as the intense DCTB-related peaks would dwarf the peaks of the polymeric analytes (Figure S77). The linear mode measurements were carried out with a laser intensity of 80%, which is significantly higher than the intensities used for the later polymer matrix experiments.

Positive and negative MALDI experiments of only the polymers without secondary matrix were carried out in linear as well as reflectron mode. Without a secondary matrix, PNDI(T2), P3DDT, PTQ1, and PII(T2) are MALDI silent in both positive and negative linear mode unless very high laser intensities are used. Similar results are obtained in positive and negative reflectron mode measurements: At low laser intensities the matrices are MALDI silent in both positive and negative mode (Figures S88 and S90), and only for high laser intensities are alkyl chain fragments visible at $m/z \leq 85$ (Figures S89 and S91).

In positive reflectron mode, alkali cations were visible at $m/z = 23$ (Na^+) and 39 (K^+) in some measurements, while in negative reflectron mode halide ions were sometimes detected at $m/z = 39$ (Cl^-) and $79/81$ (Br^-). These ions were observed already for measurements of the pure sample plate and are commonly found artifacts. While 2,5-DHB and 9AA tend to form clusters with these ions, thereby leading to more peaks in the LMWC range (Figures S5–S10), this was not observed for the polymer matrices.

Even without secondary matrix, P9OFI shows polymer-related patterns in both positive and negative mode measurements (Figures S88–S91). Test measurements with pure analytes using P9OFI as matrix in both positive and negative reflectron mode allowed the detection of the analytes, but also showed peaks relating to the polymer backbone (Figures S35 and S36). The polymer was not further investigated.

All measurements using the polymers as matrix were carried out in the high-resolution reflectron mode commonly used for metabolic analyses and MALDI MS imaging experiments. Analytes and polymer matrices were deposited in layers on the sample plate using the thin layer method^{47,48} with the matrix coated on the analyte. This sample preparation corresponds to the device structure commonly found in MALDI MS imaging experiments,¹² and coating the matrix as a thin film on the analyte requires significantly less matrix material compared to the classic sample preparation by cocrystallizing (mixing). Comparing layered and mixed sample preparation revealed that the sensitivity of the former outperforms the latter (Figures S92–S95).

Positive Reflectron Mode. Three compounds were measured in positive reflectron mode MALDI MS to assess the potential of the polymeric matrices, and the results of these experiments are compared against 2,5-DHB: reserpine (RP), an indole alkaloid used as MS standard, tetraethylammonium chloride (TEAC), which is used in histological assays to block potassium channels, and coumaphos (CP), a phosphorothioate-based insecticide used to control varroa mite infestation in honeybee colonies (see Table S2 for the analyte structures).

The results of these measurements are visualized in graphs containing four stacked spectra depicting, top down, (a) only the sample plate, (b) a layer of pure analyte without matrix, (c) a layer of pure matrix without analyte, and (d) an analyte layer coated with a matrix layer (see Figure 1). All polymers allow MALDI MS measurements of the analytes in positive and negative mode (see Figures 1 and 2 and Supporting Information). Comparing spectra (b) and (d) shows that the polymer is essential for successful measurements, as without it the analytes are not detected.

PNDI(T2), P3DDT, and PTQ1 were drop-coated from solutions with a very low concentration of 0.1 mg/mL, while for PII(T2) a concentration of 10 mg/mL was used (see

Figure 1. Reflectron positive mode MALDI MS spectra of (a) sample plate (green), (b) only the pure analyte RP (red), (c) only the pure matrix PNDI(T2) (blue), and (d) PNDI(T2) coated on RP (black). Peaks: $m/z = 607$ [$\text{RP} - \text{H}^+$]; 395 RP fragment [RP3^+]; 39 [K^+].

Figure 2. Reflectron negative mode MALDI MS spectra of (a) sample plate (green), (b) only the pure analyte CiA (red), (c) only the pure matrix PNDI(T2) (blue), and (d) PNDI(T2) coated on CiA (black). Peaks: $m/z = 191$ [$\text{CiA} - \text{H}^+$]; 87 CiA fragment [CiA1^-]; 42 side chain fragment [C_3H_6^-].

Supporting Information for further information). Reserpine was measured with low laser intensities between 10% and 40% (PNDI(T2) Figure S11; P3DDT Figure S17; PTQ1 Figure S23; PII(T2) Figure S29). The already reported phenomenon of hydride elimination was observed in all polymer measurements, leading to a molecular mass signal at $m/z = 607$.⁵⁰ In addition to this peak, RP fragments are observable in the PNDI(T2) and P3DDT measurements.

TEAC was detected as [TEA^+] at very low laser intensities of 10–20% (PNDI(T2) Figure S12; P3DDT Figure S18; PTQ1 Figure S24; PII(T2) Figure S30). Interestingly, in both PNDI(T2) and P3DDT measurements a fragment is observable despite a laser intensity of 10%, while both PTQ1 and PII(T2) experiments with 20% solely show the molecular mass peak of [TEA^+] at $m/z = 130$.

In comparison to RP and TEAC, the detection of CP required higher laser intensities of 20–50% for all four polymer matrices (PNDI(T2) Figure S13; P3DDT Figure S19; PTQ1

Figure S25; PII(T2) Figure S31). Fragments of CP were not detectable, but in the PTQ1 measurement, a peak at $m/z = 401$ could correspond to $[M + K]^+$.

Negative Reflectron Mode. For the assessment of the polymer matrices in negative reflectron mode MALDI MS, three metabolites were investigated (see Table S2 for molecular structures): cholic acid (ChA), an important primary bile acid with a steroid structure; citric acid (CiA), the central intermediate of the citric acid cycle; and L-phenylalanine (PA), an essential amino acid monitored in newborns to detect phenylketonuria. The results of the polymer matrix measurements were compared against measurements using 9AA.

The experiments establish that the polymers are essential for the detection of the analytes under the investigated conditions and substantiate that PNDI(T2), P3DDT, PTQ1, and PII(T2) function as negative mode matrices, classifying them as new dual-mode matrices. P9OFI similarly allows analyte detection in both modes, but was not further investigated due to the observed polymer-related peaks.

PII(T2), PNDI(T2), and PTQ1 were coated from solutions with a very low concentration of 0.1 mg/mL, and P3DDT was coated with a concentration of 1 mg/mL (see Supporting Information for further information).

ChA was measured with laser intensities of 30–40%, and the molecular mass peak was detected in every experiment. Also alkyl side chains at $m/z = 42$ and 57 are visible (PNDI(T2) Figure S14; P3DDT Figure S20; PTQ1 Figure S26; PII(T2) Figure S32).

CiA was detected with very low laser intensities of 10–30% (PNDI(T2) Figure 2; P3DDT Figure S21; PTQ1 Figure S27; PII(T2) Figure S33). A fragment of CiA is observable in the spectra of PNDI(T2), PTQ1, and PII(T2) at $m/z = 87$.

The measurements of PA were carried out with laser intensities of 30–40%. PTQ1 gives a very good spectrum (Figure S28) containing solely the molecular mass peak. In the spectra of PII(T2), PNDI(T2), and P3DDT additionally low-intensity peaks relating to alkyl side chains are visible at $m/z = 42$ and 57 (PNDI(T2) Figure S16; P3DDT Figure S22; PII(T2) Figure S34).

It has to be noted that optical properties and MALDI matrix suitability (required laser intensities, analyte ionization ability) are not directly correlated. While in both positive and negative mode there is a clear trend that higher laser intensities cause more distinct side chain signals, the intensity threshold seems to be influenced by several factors.

Detection Limit. Using the polymer matrices PNDI(T2), P3DDT, PTQ1, and PII(T2), concentration-dependent studies were carried out in both positive and negative reflectron mode MALDI MS to evaluate the sensitivity of the polymer matrices and the potential for quantitative analyses. Similarly to the previous measurements, first the analyte was drop-coated from a solution of a defined concentration; then the polymer was coated over the analyte. The concentration of the analyte was stepwise reduced until a detection was no longer possible, while all experimental parameters were kept constant. In positive mode, RP was measured with a laser intensity of 20% (PNDI(T2)), 40% (P3DDT) and 30% (PTQ1, PII(T2)). The limit of detection (LOD) for RP was found to be below 16 pmol for PNDI(T2) and below 164 pmol for P3DDT, PTQ1, and PII(T2) under the chosen conditions (PNDI(T2): Figures S53–S55; P3DDT: Figures S59–S61; PTQ1: Figures S65–S67; PII-

(T2): Figures S71–S73). Using a laser intensity of 50% (PNDI(T2), PII(T2)) and 30% (P3DDT, PTQ1), ChA was similarly measured in the negative mode, and the LOD was found to be below 245 pmol for all polymer matrices (PNDI(T2): Figures S56–S58; P3DDT: Figures S62–S64; PTQ1: Figures S68–S70; PII(T2): Figures S74–S76). Both LOD values were confirmed in three independent measurement series. Although the measurement conditions were not optimized, all polymer matrices show a very good sensitivity in both positive and negative mode LOD experiments, further substantiating the potential of conjugated polymers as excellent dual-mode matrices.

Complex Sample. To determine whether the polymer matrices are also suitable to measure complex biological samples containing a high number of metabolites in low individual concentration, a commercially available *Ginkgo biloba* (G.B.) extract was investigated in both positive and negative mode reflectron MALDI MS with PNDI(T2), P3DDT, PTQ1, and PII(T2). G.B. was chosen as complex sample as the extract contains a rich variety of flavonoids and terpene lactones, and studies^{51–53} using different MS methods identified several key components. All four investigated polymers allow recording complex spectra of G.B. in both positive and negative mode (PNDI(T2) Figures 3 and 4;

Figure 3. Reflectron positive mode MALDI MS spectra of (a) sample plate (green), (b) only the analyte G.B. (red), (c) only the pure matrix PNDI(T2) (blue), and (d) PNDI(T2) coated on G.B. (black). $m/z = 39$ corresponds to $[K]^+$.

P3DDT Figures S44–S46; PTQ1 Figures S47–S49; PII(T2) Figure S50–S52), confirming the suitability of the conjugated polymers for the dual-mode analysis of biological samples. The peak count and peak intensity were further analyzed.

In the positive mode, the number of detected peaks (with $I > 100$) and the average peak intensity using either of the polymers are fully consistent with the results obtained with 2,5-DHB (see Table 2, Figures S37 and S38). When the intensity threshold is raised ($I > 10k$), the number of detected peaks is in the same range for 2,5-DHB and PNDI(T2) (120 and 125, respectively), while P3DDT, PTQ1, and PII(T2) have significantly fewer high-intensity peaks (26, 20, and 43, respectively) and consequently achieve a better average peak intensity.

The prepared samples were subsequently used for negative mode measurements. With regard to the number of detected

Figure 4. Reflectron negative mode MALDI MS spectra of (a) sample plate (green), (b) only the analyte G.B. (red), (c) only the pure matrix PNDI(T2) (blue), and (d) PNDI(T2) coated on G.B. (black). Peaks: $m/z = 79/81$ $[\text{Br}]^-$; 42 side chain fragment $[\text{C}_3\text{H}_6]^-$.

peaks and the average peak intensity, PII(T2), PNDI(T2), and PTQ1 are comparable to 9AA (see Table 2, Figures S39 and S40), while P3DDT records about 25% less peaks, but significantly more high-intensity peaks (50% increase). Again PII(T2), PNDI(T2), and PTQ1 lead to fewer high-intensity peaks but an improved average peak intensity.

The experiments substantiate that the investigated polymers are efficient dual-mode matrices and suitable for the analysis of complex biological samples. It has to be noted that in the measurements with 2,5-DHB and 9AA the most intense peaks are caused by the small-molecule matrices.

MALDI MS Imaging. First MALDI MS imaging experiments of fresh-frozen rat brain sections⁵⁴ using P3DDT as matrix were performed. One coronal and one sagittal section were spray-coated with the polymer and measured in reflectron negative mode. Each section was sampled at approximately 85 834 and 51 445 individual locations or pixels for sagittal and coronal, respectively. Each pixel corresponds to a full mass spectrum with an m/z range of 0–2000.

Figure 5 depicts heat maps visualizing the spatial distribution of eight different ions in a single coronal brain section. The ions have masses ranging from 101.0 Da (± 0.3 Da) to 904.7 Da (± 0.3 Da), thereby covering the range of LMWCs central to metabolomics. In Figure 6, the spatial distributions of eight different ions in a single sagittal brain section are visualized. In order to represent the broad range of detected metabolites, the

depicted ions have molecular masses from 95.9 Da (± 0.1 Da) to 904.5 Da (± 0.1 Da).

In both data sets, different brain regions are clearly distinguishable, and the visualization results are well correlated to the tissue morphology, confirming the high potential of the used conjugated polymers as matrices in MALDI MS imaging. Indeed, the combination of the high molecular weight and the resulting stability together with the extensive π -conjugation and high UV absorption of the polymer leads to a complete absence of matrix-related peaks in the m/z range of 60–2000 in the mass spectra. This causes a significant improvement of the detection and evaluation of tissue-specific ions in the data set.

CONCLUSIONS AND OUTLOOK

MALDI MS is a ubiquitous analytical tool, and MALDI MS imaging has a high potential to become a standard imaging method in medical, pharmaceutical, and biological research. Both methods are increasingly used for the analysis of LMWCs, for example, metabolites, requiring new matrix systems that do not lead to matrix-related signals in the low-molecular-weight region and have an increased vacuum stability.

In this paper, five different conjugated polymers, PNDI(T2), P3DDT, PTQ1, PII(T2), and P9OFI, were tested as matrix systems in positive and negative mode MALDI MS. All investigated polymers have a high molecular weight and are fully vacuum stable. They are solution processable, can be coated into thin films, and show good to very good absorption behavior in the optical range of the commonly used Nd:YAG laser, thereby meeting all prerequisites for a MALDI matrix.

Using defined LMWCs as well as a complex *Ginkgo biloba* extract as analytes, we could show that all polymers function as matrices for positive as well as negative mode MALDI, classifying them as rare dual-mode matrices. Drop-coating thin films of the polymers on the analyte in a variation of the thin layer method allowed a vast reduction in the amount of matrix used without compromising the analytical result.

PNDI(T2), P3DDT, PTQ1, and PII(T2) are MALDI silent in the full measurement range in both positive (upper instrument limit $m/z = 200\text{k}$) and negative mode (upper instrument limit $m/z = 160\text{k}$), except at high laser intensities, when peaks at $m/z \leq 85$ are visible, which are attributable to fragments (in source decay) of the solubilizing alkyl chains. P9OFI allows the detection of analytes, but as the main chain of the polymer fragmentizes and backbone-related peaks are visible in both positive and negative mode spectra, this polymer was not further investigated. The performance of

Table 2. Number of Peaks and Average Peak Intensity in Reflectron Mode Measurements of G.B.

matrix	reflectron positive mode			reflectron negative mode		
	no. of peaks P_{100} with $I > 100$ (av peak intensity)	no. of peaks P_{10k} with $I > 10k$ (av peak intensity)	most intense peak m/z (peak intensity)	no. of peaks P_{100} with $I > 100$ (av peak intensity)	no. of peaks P_{10k} with $I > 10k$ (av peak intensity)	most intense peak m/z (peak intensity)
DHB	2980 (3.5×10^3)	120 (4.3×10^4)	215 (3.2×10^5) ^a			
9AA				2972 (2.7×10^3)	110 (3.4×10^4)	193 (7.3×10^5) ^a
PNDI(T2)	2980 (3.3×10^3)	125 (4.4×10^4)	104 (5.7×10^5)	2966 (7.6×10^2)	16 (2.0×10^4)	301 (3.2×10^4)
P3DDT	2986 (2.8×10^3)	26 (4.8×10^4)	104 (5.0×10^5)	2260 (3.3×10^3)	155 (3.6×10^4)	210 (3.3×10^5)
PII(T2)	2976 (2.6×10^3)	43 (5.4×10^4)	104 (6.6×10^5)	2960 (7.6×10^2)	19 (1.9×10^4)	210 (3.5×10^4)
PTQ1	2977 (1.8×10^3)	20 (5.7×10^4)	104 (5.5×10^5)	2513 (1.5×10^3)	66 (2.8×10^4)	210 (1.1×10^5)

^aMatrix-related peak.

Figure 5. Exemplary ion images of selected mass channels acquired by MALDI MS imaging of a coronal rat brain section in negative mode and by using P3DDT as matrix. The distributions of the ionic species are visualized as heat maps. Ionic species (m/z) from top left to bottom right: 101.0 Da (± 0.3 Da); 113.1 Da (± 0.3 Da); 124.0 Da (± 0.2 Da); 210.9 Da (± 0.3 Da); 225.2 Da (± 0.3 Da); 309.5 Da (± 0.3 Da); 384.7 Da (± 0.3 Da); 904.7 Da (± 0.3 Da).

Figure 6. Exemplary ion images of selected mass channels acquired by MALDI MS imaging of a sagittal rat brain section in negative mode and by using P3DDT as matrix. The distributions of the ionic species are visualized as heat maps. Ionic species (m/z) from top left to bottom right: 97.1 Da (± 0.3 Da); 113.1 Da (± 0.3 Da); 124.0 Da (± 0.3 Da); 153.1 Da (± 0.3 Da); 181.0 Da (± 0.3 Da); 225.3 Da (± 0.3 Da); 309.3 Da (± 0.1 Da); 904.5 Da (± 0.3 Da).

PNDI(T2), P3DDT, PTQ1, and PII(T2) as matrices was tested in reflectron mode, which permits a higher resolution and is preferred for metabolomics and MALDI MS imaging experiments. The efficacy when measuring single LMWC analytes is very good and in all cases comparable with or better than the performance of the well-established matrices 2,5-DHB (positive mode) and 9AA (negative mode). The measurement of a G.B. extract as an example of a complex biological sample similarly confirmed that the analyte ionization ability and sensitivity of the polymer matrices are comparable with or better than those of the commonly used matrices 2,5-DHB and 9AA. MALDI MS imaging experiments were performed on coronal and sagittal rat brain sections, and the visualization results are well correlated to the tissue morphology.

In summary, the presented study shows that conjugated polymers are rare dual-mode MALDI matrices combining vacuum stability and a very good analyte ionization ability with being MALDI silent. Due to the inherent structural versatility of both polymer backbone and side chains, conjugated polymers are a highly promising molecule class to design new functional dual-mode matrix systems for both MALDI MS and MALDI MS imaging.

Currently, we are carrying out more detailed studies on the suitability of selected polymer systems as MALDI MS imaging matrices, as well as exploring structural variations to further improve the analyte ionization ability.

■ ASSOCIATED CONTENT

⑤ Supporting Information

The Supporting Information is available free of charge on the ACS Publications website at DOI: 10.1021/jacs.8b06637.

Details about the used chemicals, the employed instrumentation, and experimental settings, as well as results of the UV/vis as well as all MALDI MS experiments (PDF)

■ AUTHOR INFORMATION

Corresponding Author

*lissel@ipfdd.de

ORCID

Anton Kiri: 0000-0002-4263-9377

Nikos Hadjichristidis: 0000-0003-1442-1714

Franziska Lissel: 0000-0003-0254-4565

Notes

The authors declare no competing financial interest.

■ ACKNOWLEDGMENTS

The work was supported by KAUST Center Partnership Fund OSR-2017-CPF-3322-01. F.L. thanks the VCI for a Liebig Fellowship. The authors gratefully acknowledge the support of Wolfgang Heidrich in establishing the collaboration infrastructure between KAUST's VCC and the IPF. Anastasia Khrenova is thanked for helping with the graphics.

REFERENCES

- (1) Karas, M.; Bachmann, D.; Hillenkamp, F. *Anal. Chem.* **1985**, *57*, 2935–2939.
- (2) Karas, M.; Bachmann, D.; Bahr, U.; Hillenkamp, F. *Int. J. Mass Spectrom. Ion Processes* **1987**, *78*, 53–68.
- (3) Tanaka, K.; Waki, H.; Ido, Y.; Akita, S.; Yoshida, Y.; Yoshida, T.; Matsuo, T. *Rapid Commun. Mass Spectrom.* **1988**, *2*, 151–153.
- (4) Clark, A. E.; Kaleta, E. J.; Arora, A.; Wolk, D. M. *Clin Microbiol Rev.* **2013**, *26*, 547–603.
- (5) Croxatto, A.; Prod'homme, G.; Greub, G. *FEMS Microbiol Rev.* **2012**, *36*, 380–407.
- (6) Rodrigo, M. A. M.; Zitka, O.; Krizkova, S.; Moulick, A.; Adam, V.; Kizek, R. *J. Pharm. Biomed. Anal.* **2014**, *95*, 245–255.
- (7) Patel, R. *Clin. Chem.* **2015**, *61*, 100–111.
- (8) Duncan, M. W.; Nedelkov, D.; Walsh, R.; Hattan, S. *J. Clin. Chem.* **2016**, *62*, 134–143.
- (9) Karas, M.; Hillenkamp, F. *Anal. Chem.* **1988**, *60*, 2299–2301.
- (10) Meier, M. A. R.; Hoogenboom, R.; Fijten, M. W. M.; Schneider, M.; Schubert, U. S. *J. Comb. Chem.* **2003**, *5*, 369–374.
- (11) Chaurand, P.; Stoeckli, M.; Caprioli, R. M. *Anal. Chem.* **1999**, *71*, 5263–5270.
- (12) Norris, J. L.; Caprioli, R. M. *Chem. Rev.* **2013**, *113*, 2309–2342.
- (13) Crecelius, A. C.; Schubert, U. S.; von Eggeling, F. *Analyst* **2015**, *140*, 5806–5820.
- (14) Aichler, M.; Walch, A. *Lab. Invest.* **2015**, *95*, 422–431.
- (15) Rauser, S.; Marquardt, C.; Balluff, B.; Deininger, S.-O.; Albers, C.; Belau, E.; Hartmer, R.; Suckau, D.; Specht, K.; Ebert, M. P.; Schmitt, M.; Aubele, M.; Höfler, H.; Walch, A. *J. Proteome Res.* **2010**, *9*, 1854–1863.
- (16) Reyzer, M. L.; Hsieh, Y.; Ng, K.; Korfmacher, W. A.; Caprioli, R. M. *J. Mass Spectrom.* **2003**, *38*, 1081–1092.
- (17) Lagarrigue, M.; Caprioli, R. M.; Pineau, C. *J. Proteomics* **2016**, *144*, 133–139.
- (18) Seeley, E. H.; Caprioli, R. M. *Anal. Chem.* **2012**, *84*, 2105–2110.
- (19) Calvano, C. D.; Monopoli, A.; Cataldi, T. R. I.; Palmisano, F. *Anal. Bioanal. Chem.* **2018**, *410*, 4015–4038.
- (20) Li, S.; Zhang, Y.; Liu, J. A.; Han, J.; Guan, M.; Yang, H.; Lin, Y.; Xiong, S.; Zhao, Z. *Sci. Rep.* **2016**, *6*, 37903.
- (21) Strupat, K.; Karas, M.; Hillenkamp, F. *Int. J. Mass Spectrom. Ion Processes* **1991**, *111*, 89–102.
- (22) Knochenmuss, R.; Dubois, F.; Dale, M. J.; Zenobi, R. *Rapid Commun. Mass Spectrom.* **1996**, *10*, 871–877.
- (23) McCombie, G.; Knochenmuss, R. *Anal. Chem.* **2004**, *76*, 4990–4997.
- (24) Monopoli, A.; Calvano, C. D.; Nacci, A.; Palmisano, F. *Chem. Commun. (Cambridge, U. K.)* **2014**, *50*, 4322–4324.
- (25) Shariatgorji, M.; Nilsson, A.; Källback, P.; Karlsson, O.; Zhang, X.; Svenningsson, P.; Andren, P. E. *J. Am. Soc. Mass Spectrom.* **2015**, *26*, 934–939.
- (26) Manier, M. L.; Spraggins, J. M.; Reyzer, M. L.; Norris, J. L.; Caprioli, R. M. *J. Mass Spectrom.* **2014**, *49*, 665–673.
- (27) Janfelt, C.; Wellner, N.; Hansen, H. S.; Hansen, S. H. *J. Mass Spectrom.* **2013**, *48*, 361–366.
- (28) Anderson, D. M. G.; Ablonczy, Z.; Koutalos, Y.; Spraggins, J.; Crouch, R. K.; Caprioli, R. M.; Schey, K. L. *J. Am. Soc. Mass Spectrom.* **2014**, *25*, 1394–1403.
- (29) Ketting, H.; Vens-Cappell, S.; Soltwisch, J.; Pirk, A.; Haier, J.; Muthing, J.; Dreisewerd, K. *Anal. Chem.* **2014**, *86*, 7798–7805.
- (30) Thomas, A.; Charbonneau, J. L.; Fournaise, E.; Chaurand, P. *Anal. Chem.* **2012**, *84*, 2048–2054.
- (31) Giampà, M.; Lissel, M. B.; Patschkowski, T.; Fuchser, J.; Hans, V. H.; Gembruch, O.; Bednarczyk, H.; Niehaus, K. *Chem. Commun. (Cambridge, U. K.)* **2016**, *52*, 9801–9804.
- (32) Muthu, M.; Chun, S.; Wu, H.-F.; Duncan, M. W.; Gopal, J. *J. Mass Spectrom.* **2018**, *53*, 525–540.
- (33) Dong, X.; Cheng, J.; Li, J.; Wang, Y. *Anal. Chem.* **2010**, *82*, 6208–6214.
- (34) Liu, Y.; Liu, J.; Yin, P.; Gao, M.; Deng, C.; Zhang, X. *J. Mass Spectrom.* **2011**, *46*, 804–815.
- (35) Wang, Z.; Cai, Y.; Wang, Y.; Zhou, X.; Zhang, Y.; Lu, H. *Sci. Rep.* **2017**, *7*, 44466.
- (36) Marsico, A. L. M.; Czeran, B.; Duncan, B.; Elci, S. G.; Jiang, Y.; Onasch, T. B.; Wormhoudt, J.; Rotello, V. M.; Vachet, R. W. *J. Am. Soc. Mass Spectrom.* **2015**, *26*, 1931–1937.
- (37) Bernier, M. C.; Wysocki, V. H.; Dagan, S. *J. Mass Spectrom.* **2015**, *50*, 891–898.
- (38) Towers, M. W.; Cramer, R. *Ionic Liquids and Other Liquid Matrices for Sensitive MALDI MS Analysis*. In *Advances in MALDI and Laser-Induced Soft Ionization Mass Spectrometry*; Cramer, R., Ed.; Springer International Publishing: Cham, 2016; pp 51–64.
- (39) Pasch, H.; Schrepp, W. *MALDI-TOF Mass Spectrometry of Synthetic Polymers*; Springer Science & Business Media, 2013.
- (40) Oh, J.; Kuk, J.; Lee, T.; Ye, J.; Paik, H.-J.; Lee, H. W.; Chang, T. *ACS Macro Lett.* **2017**, *6*, 758–761.
- (41) Liu, J.; Loewe, R. S.; McCullough, R. D. *Macromolecules* **1999**, *32*, 5777–5785.
- (42) Tkachov, R.; Karpov, Y.; Senkovskyy, V.; Raguzin, I.; Zessin, J.; Lederer, A.; Stamm, M.; Voit, B.; Beryozkina, T.; Bakulev, V.; Zhao, W.; Facchetti, A.; Kiri, A. *Macromolecules* **2014**, *47*, 3845–3851.
- (43) Karpov, Y.; Zhao, W.; Raguzin, I.; Beryozkina, T.; Bakulev, V.; Al-Hussein, M.; Häußler, L.; Stamm, M.; Voit, B.; Facchetti, A.; Tkachov, R.; Kiri, A. *ACS Appl. Mater. Interfaces* **2015**, *7*, 12478–12487.
- (44) Soltzberg, L. J.; Patel, P. *Rapid Commun. Mass Spectrom.* **2004**, *18*, 1455–1458.
- (45) Sládková, K.; Houska, J.; Havel, J. *Rapid Commun. Mass Spectrom.* **2009**, *23*, 3114–3118.
- (46) Karpov, Y.; Maiti, J.; Tkachov, R.; Beryozkina, T.; Bakulev, V.; Liu, W.; Komber, H.; Lappan, U.; Al-Hussein, M.; Stamm, M.; Voit, B.; Kiri, A. *Polym. Chem.* **2016**, *7*, 2691–2697.
- (47) Kussmann, M.; Nordhoff, E.; Rahbek-Nielsen, H.; Haebel, S.; Rossel-Larsen, M.; Jakobsen, L.; Gobom, J.; Mirgorodskaya, E.; Kroll-Kristensen, A.; Palm, L.; Roepstorff, P. *J. Mass Spectrom.* **1997**, *32*, 593–601.
- (48) Ole, V.; Roepstorff, P.; Mann, M. *Anal. Chem.* **1994**, *66*, 3281–3287.
- (49) Nielsen, K. *Journal of Chromatography A* **2003**, *1002*, 111–136.
- (50) Awad, H.; Stoudemayer, M. J.; Usher, L.; Amster, I. J.; Cohen, A.; Das, U.; Whittall, R. M.; Dimmock, J.; El-Aneel, A. *J. Mass Spectrom.* **2014**, *49*, 1139–1147.
- (51) Beck, S.; Stengel, J. *Phytochemistry* **2016**, *130*, 201–206.
- (52) Ding, S.; Dudley, E.; Plummer, S.; Tang, J.; Newton, R. P.; Brenton, A. G. *Phytochemistry* **2008**, *69*, 1555–1564.
- (53) Ding, S.; Dudley, E.; Song, Q.; Plummer, S.; Tang, J.; Newton, R. P.; Brenton, A. G. *Rapid Commun. Mass Spectrom.* **2008**, *22*, 766–772.
- (54) Animal tissue used in this study was kindly provided by the animal facility of the Bielefeld University. It represents tissues from animals that have been sacrificed in the context of other studies with no connection to the experiments presented in this article.

7.2. Amorphous Conjugated Polymers as Efficient Dual-Mode MALDI Matrices for Low-Molecular-Weight Analytes (Publication 2)

Amorphous Conjugated Polymers as Efficient Dual-Mode MALDI Matrices for Low-Molecular-Weight Analytes

Kilian Horatz,^[a, b] Kristina Ditte,^[a, b] Thomas Prenveille,^[a, b, c] Ke-Nan Zhang,^[a, b] Dieter Jehnichen,^[a] Anton Kiri,^[a] Brigitte Voit,^[a, b] and Franziska Lissel^{✉[a, b]}

Matrix-assisted laser desorption/ionization mass spectrometry (MALDI MS) usually employs highly crystalline small-molecule matrices, and the analyte is interpreted as being co-crystallized with the matrix. We recently showed that semi-crystalline polymers are efficient matrices for the detection of low-molecular-weight compounds (LMWCs) in MALDI MS and MALDI MS Imaging, and are dual-mode, i.e., enabling both positive and negative modes. The matrix performances of two fluorene/naphthalene diimide co-polymers P(TNDIT-FI(C₄C₇)) and P(TNDIT-FI(C₁₀C₈)) were investigated and compared. Both are

fully amorphous according to XRD measurements, show high relative absorption values at the wavelength of common MALDI lasers ($\lambda_{\text{Nd:YAG}} = 355 \text{ nm}$: C₄C₇ = 73 %; C₁₀C₈ = 67 %), and are solution processable. As matrices, they are dual-mode, and enable the detection of LMWCs while being mostly MALDI-silent. Compared with semicrystalline polymer matrices, the amorphous matrices give similar or better signal intensities, thus indicating that analyte inclusion takes place in the amorphous part of the polymer matrix.

Introduction

"Matrix assisted laser desorption/ionization mass spectrometry" (MALDI MS) was developed in the 1980s to record mass spectra of high molecular weight proteins,^[1] and since then became a standard tool to analyze synthetic and biological macromolecules in medical, pathological and environmental research.^[2] MALDI is a soft ionization method; the analyte is incorporated in the crystal lattice of an excess of matrix molecules, and when irradiated with a pulsed laser beam, the matrix absorbs the energy, ionizes the analyte and protects it from degradation during the desorption process.^[3] Although the ionization and desorption processes were intensively studied, they are as yet not fully understood, and several theories are described for MALDI ionization mechanisms, which can be mainly divided into primary ionization mechanisms (e.g. polar fluid model, cluster model) and secondary ionization mechanisms (e.g. proton transfer, electron transfer).^[4] In contrast to other types of mass spectrometry (e.g. electrospray ionization (ESI), fast atom bombardment (FAB), secondary ion mass spectrometry (SIMS), surface assisted laser ionization/desorption (SALDI),^[5] and

desorption/ionization on porous silicon (DIOS)^[6], MALDI rarely leads to analyte fragmentation, and produces predominantly single charged ion peaks,^[4c,7] simplifying peak assignments. Further advantages are a high detection sensitivity (often below femtomolar analyte concentration), a fast and simple sample preparation and high tolerance towards salts and buffers.^[3] While developed and predominantly used for the analysis of high molecular weight compounds (HMWCs) such as proteins and polymers,^[1,4] currently there is a pronounced interest to establish methods to reliably and reproducibly measure low-molecular-weight compounds (LMWCs) with molar masses below 1000 Da.^[8] This interest is driven by recent developments such as MALDI MS Imaging (MALDI MSI), a novel tool to investigate surfaces,^[2a,b,9] for example tissue slices, and address complex questions, e.g. regarding metabolomic changes in brain tissue affected by stroke.^[10]

MALDI Matrix Systems

MALDI matrix systems are designed to meet several physicochemical requirements: sufficient absorption at the wavelength of commonly used MALDI matrix laser systems (Nd:YAG) ($\lambda_{\text{Nd:YAG}} = 355 \text{ nm}$), establishing a crystal lattice to incorporate the analyte, vacuum stability for long term measurements, a good ionization efficiency and homogeneous distribution for reliable and reproducible measurements.^[3] Additionally, regarding the analyses of LMWCs the matrix should ideally not yield matrix related signals (e.g. molecule ion, clusters, fragments, etc.) in the low mass range ($m/z \leq 1000 \text{ Da}$). The correct choice of a suitable MALDI matrix system is essential for successful MALDI MS measurements and often challenging.^[3] Different classes of analytes necessitate different matrix systems.^[3,9a] Nowadays, the most widely used matrix systems are small organic matrices (SOMs), e.g. 2,5-dihydroxybenzoic acid (2,5-DHB),^[11] 4-hydroxy-

[a] K. Horatz,^{*} K. Ditte,^{*} Dr. T. Prenveille, K.-N. Zhang, Dr. D. Jehnichen, Dr. A. Kiri, Prof. Dr. B. Voit, Dr. F. Lissel
Leibniz-Institut für Polymerforschung Dresden e.V.
Hohe Straße 6, 01069 Dresden (Germany)
E-mail: lissel@ipfdd.de

[b] K. Horatz,^{*} K. Ditte,^{*} Dr. T. Prenveille, K.-N. Zhang, Prof. Dr. B. Voit, Dr. F. Lissel
Technische Universität Dresden
Mommensenstraße, 01062 Dresden (Germany)

[c] Dr. T. Prenveille
Arkema S.A. Colombes, France

[*] These authors contributed equally to this work.

Supporting information for this article is available on the WWW under <https://doi.org/10.1002/cplu.201900203>

This article is part of a Special Issue on " π -Conjugated (Macro)molecules and their Applications".

α -cyano cinnamic acid (HCCA), 3,5-dimethoxy-4-hydroxy-cinnamic acid (sinapinic acid) and 9-aminoacridine (9AA),^[12] which are suitable for analyses of biochemical compounds like proteins, peptides, lipids and drugs.^[9a] SOMs usually contain alternating double bonds and/or conjugated ring systems to ensure a sufficient and suitable UV absorption, and acidic or basic functional groups to (de-)protonate the analyte and provide a good ionization efficiency.

During the preparation of the MALDI sample, SOMs incorporate the analyte into their crystal lattice, and ideally, a homogenous and defect free crystal lattice is targeted to ensure a high ionization efficiency and successful measurements,^[11] and the correlation between the matrix crystal forms and the analyte signal distribution was intensively studied.^[13] Crystal formation and analyte incorporation are influenced by the sample preparation method and several method related parameters;^[3,14] e.g. it was reported that the commonly used air-dried-droplet method leads to poor homogeneity, while electrospray sample preparation gives evenly sized matrix crystals with well-distributed analyte incorporation.^[15] Furthermore, a low vapor pressure of the used solvent and a longer drying time improves the crystal formation and promotes homogeneous incorporation of the analyte.^[13b,16] However, the formation of crystal layers with a homogeneous analyte distribution is difficult to achieve when using SOMs, often resulting in analyte-rich “hotspots” and analyte-poor areas.^[13b,17]

Usually, SOMs are suitable for either positive or negative detection modes but not for both simultaneously,^[18] and only few examples of “dual-mode” matrices were published so far,^[19] requiring a double sample preparation to correlate both modes. Regarding LMWCs, SOMs can lead to a number of matrix related signals (e.g. molecule ion, clusters, fragments, etc.) in the low mass range ($m/z \leq 1000$ Da), often complicating or preventing LMWC analytics. A number of approaches were explored to avoid intense matrix signal generation by SOMs, e.g. matrix suppression effects,^[20] cyclodextrin-supported 2,5-DHB matrix,^[17b] improved reactive functional groups,^[21] and chemical derivatization.^[22] Also, new classes of matrix systems were developed and successfully established, e.g. nanoparticles,^[23] ionic liquids,^[24] graphene,^[25] doped graphene,^[26] graphene oxides,^[25b,27] graphite (GALDI-MS),^[28] metal oxides,^[29] mostly targeting polar and non-polar, biochemical LMWCs (e.g. amino acids, nucleosides, steroids, etc.).

Conjugated Polymers as MALDI Matrix

Recently, our group showed that conjugated polymers function as matrices in MALDI MS and MSI, revealing a sufficient absorption at 355 nm, few interferences in the LMWC area, high vacuum stability, homogeneous layering and dual-mode suitability.^[30] Especially poly([N,N'-bis(2-octyldodecyl)-naphthalene-1,4,5,8-bis(dicarbox-imide)-2,5-diyl]-*alt*-5,5'-(2,2'-bithiophene)) (PNDI(T2)) revealed an excellence performance as MALDI matrix. Interestingly, despite its excellent optical properties, poly(9,9-di-*n*-octylfluorenyl-2,7-diyl) (P9OFI) was the only

investigated polymer not “MALDI-silent”, instead exhibiting signals associated with a backbone degeneration.^[30]

Many conjugated polymers possess semi-crystalline structures due to a coexistence of crystalline regions induced by π - π stacking, and amorphous parts between crystallites forming due to coiling and folding of the polymer chains. For both MALDI active homopolymers PNDI(T2) and P9OFI, a semi-crystalline film morphology was reported,^[31] in itself already a deviation from the usual highly crystalline structure of SOMs. As mentioned above, the intermolecular interaction between analyte and matrix in form of an incorporation of the analyte into the matrix' crystal lattice was reported to be important for the ionization process.^[4a,b] This leads to the question whether a crystalline structure is – at least for polymeric matrices – indeed a prerequisite, or if analytes can also be embedded into the amorphous parts of the polymer film. The degree of polymer crystallinity can be tuned, e.g. by sidechain engineering, giving the unique possibility to explore the influence of crystallinity on matrix performance.

In this paper, we target to understand the necessity and influence of matrix crystallinity by investigating and comparing the performances of fully amorphous conjugated polymers with a structurally related semi-crystalline polymer. To obtain highly soluble polymers with an amorphous morphology, long and branched alkyl sidechains are introduced in both monomer units. The co-polymers are characterized by ¹H-NMR, GPC, CV, XRD, UV/VIS and MALDI MS. Then they are tested as MALDI matrices for the measurements of single LMWCs as well as a complex biological extract in both positive and negative mode prepared via “dried-droplet” sample preparation. Furthermore, measurements of spin-coated samples using the amorphous co-polymers and the established, semi-crystalline polymer PNDI (T2) as matrices were carried out to compare differences in matrix performance between amorphous and semi-crystalline polymer morphologies.

Results and Discussion

Design, Synthesis, and Characterization

A fluorene/NDI co-polymer was targeted due to the excellent optical properties of both parent homopolymers, and also to test whether incorporating fluorene in a co-polymer backbone can increase MALDI stability. To realize a fully amorphous morphology, long and branched alkyl sidechains were introduced on both monomers (see Scheme 1).

An NDI(T2) monomer with 2-octyldodecyl sidechains was polymerized with a 9,9-bis(2-ethylhexyl)fluorene. The structure of the resulting P(TNDIT-FI(C₄C₂)) was confirmed by ¹H-NMR (see Figure S4 in the Supporting Information), yet the solubility of the polymer was surprisingly low, leading to a low yield and also a comparably low number-average molecular weight of 25.1 kg/mol with a polydispersity (PDI) of 1.7 based on GPC (see Table S1). A second polymer was synthesized using the same NDI(T2) core, but a fluorene monomer with longer 2-octyldodecyl sidechains, matching the NDI(T2). The resulting P(TNDIT-FI

Scheme 1. Synthetic route for NDI(T2) and fluorine co-polymer **P(TNDIT-FI(C₄C₂))** and **P(TNDIT-FI(C₁₀C₈))**. The alkyl side-chains **R₁** and **R₂** are defined in this scheme.

(C₁₀C₈) is highly soluble, and the main fraction, isolated from *n*-hexane, has a number-average molecular weights of 42.1 kg/mol with a PDI of 2.2 (see Table S1). The ¹H-NMR is consistent with the expected structure (see Figure S5).

To further confirm the structural integrity of the co-polymers, MALDI MS measurements of the polymers as analytes were carried out in linear positive mode using trans-2-[3-(4-tert-butylphenyl)-2-methyl-2-propenyldiene]-malononitrile (DCTB) as secondary matrix and sodium trifluoroacetate (NaTFA) as additive (see Figures S30-34). The MALDI spectra reveal characteristic signal patterns with constant, repeating intervals of ~1380 m/z (± 7 m/z) for P(TNDIT-FI(C₄C₂)) and ~1717 m/z (± 7 m/z) for P(TNDIT-FI(C₁₀C₈)), consistent with the expected structures.

Several ionization pathways for MALDI have been described, involving either a proton- or an electron-transfer between the matrix and the analyte.^[4] The latter produces ionized radical molecules, without changing the analytes' molar mass. This electron transfer process is influenced by the electrochemical behaviors of the matrix, which were analyzed for the P(TNDIT-FI(C₄C₂)) and P(TNDIT-FI(C₁₀C₈)) co-polymers by thin-film cyclic voltammetry measurements. For both co-polymers, two reversible reductions are found, with redox potentials at -0.93 V and -1.25 V for P(TNDIT-FI(C₄C₂)) and -0.95 V and -1.29 V for P(TNDIT-FI(C₁₀C₈)) (see Figures S7 and S8). Corresponding HOMO/LUMO energies for P(TNDIT-FI(C₄C₂)) and P(TNDIT-FI(C₁₀C₈)) are given in the Tables S6 and S7. As expected, the variation of the alkyl side-chains has no significant influence on the redox potentials.

Crystallinity

X-ray diffraction (XRD) was utilized for morphological studies of the synthesized co-polymers. The samples were prepared by spin-coating from toluene following thin film preparation for grazing incidence X-ray diffraction methods, and additionally

measurements on drop-coated samples were carried out to allow a direct comparison with the films prepared for MALDI analyses. The GIWAXS curves of the drop- and spin-coated co-polymer samples similar patterns with a single, broad scattering maximum at around 20° diffraction angle, characteristic for disordered polymer chains (amorphous state). In comparison of the different side chain lengths an additional scattering maximum at around 4.7° (~ 1.9 nm) indicates an intramolecular phase separation caused by the long alkyl side chains against the more aromatic main chain (see Figure 1). The sharp peak at

Figure 1. GIWAXS curves of P(TNDIT-FI(C₄C₂)) (drop-coated: red; spin-coated: green) and P(TNDIT-FI(C₁₀C₈)) (drop-coated: blue; spin-coated: pink).

around 26.7° diffraction angle for P(TNDIT-FI(C₄C₂)) (spin-coated) (green) sample is suggested to be an inorganic impurity on the silicon wafer substrate. According to these results, both co-polymers P(TNDIT-FI(C₄C₂)) and P(TNDIT-FI(C₁₀C₈)) have completely amorphous morphologies. However, solution-processed thin films show a lesser degree of crystallinity compared to melt-processed samples of the same material.^[32] The long alkyl side-chains used to induce an amorphous morphology of the co-polymers lower the melting point, potentially allowing a solvent-free processing of the material. Differential scanning calorimetry (DSC) experiments of P(TNDIT-FI(C₁₀C₈)) showed a glass transition at 38 °C, and while a melting transition was not observed within the accessible range of -80 °C to +350 °C (see Figure S10), the co-polymer was observed to become mechanically soft and ductile over temperatures of 150 °C, and can be drawn into fibers (see Figure S2 (left)). As thermogravimetric analysis (TGA) showed that degradation started only above 320 °C under inert atmosphere, we deduced that a solvent-free processing would be possible (see Figure S9). The polymer was pressed at 225 °C to form a homogenous film, followed by a rapid or slow cooling process (see Figure S2 (right)). No difference in thickness of the formed layer was observed after cooling, which corresponds to a fully amorphous structure of the polymer.

Additional X-ray measurements utilizing a two-dimensional detector (2D-WAXS) were carried out. The diffractogram (see Figure 2) revealed no preferred orientation of the co-polymer

Figure 2. 2D-WAXS diffractogram of the co-polymer P(TNDIT-FI(C₁₀C₈)).

chains, but instead an amorphous halo is observed at $2\theta = 20^\circ$, which is induced by branched alkyl side-chains.

Optical Properties

Sufficiently strong optical absorption in the range of the MALDI laser is a key criterion for MALDI matrices. The UV/Vis spectra of drop-casted thin films of the co-polymers P(TNDIT-FI(C₄C₂)) and P(TNDIT-FI(C₁₀C₈)) depict a general pattern of two bands with a global absorption maximum at ~ 390 nm and a local absorption maximum at ~ 634 nm (see Figure S6). Compared to the UV/VIS spectra of the parent homopolymers PNDI(T2) and P9OFI, the absorption maximum at ~ 390 nm is the same for homo- and co-polymers, whereas the co-polymers' absorption maximum at ~ 634 nm is blue-shifted compared with PNDI(T2) ($\Delta\lambda \approx 63$ nm). P9OFI does not display an absorption in this region. The relative absorptions (relative to the global maximum) at the commonly used MALDI laser (Nd:YAG) wavelength ($\lambda_{\text{laser}} = 355$ nm) are 73 % for P(TNDIT-FI(C₄C₂)) and 67 % for P(TNDIT-FI(C₁₀C₈)) (see Table S5). These values are in a good agreement with PNDI(T2) and P9OFI, which both show good absorption at the laser wavelength, 82 % and 72 % respectively.^[30]

Amorphous Polymers as MALDI Matrices

To test the suitability of the amorphous polymers as MALDI matrices, three different single LMWCs were measured in positive and negative reflectron mode (see Table S2 for chemical structures) prepared via "dried-droplet" sample preparation (see the Supporting Information for details): Reserpine (RP), an indole alkaloid; tetraethylammonium chloride (TEAC), a potassium channel blocker; coumaphos (CP), a phosphorothioate insecticide, were measured in positive reflectron mode. Cholic acid (ChA), a primary bile acid; citric acid (CiA), an

intermediate of the citric-acid-cycle; L-phenylalanine (PA), an essential amino acid, were used to assess the amorphous polymers in negative reflectron mode. To enable a direct comparison between amorphous and semi-crystalline polymer matrices, the single analytes RP (pos.) and CiA (neg.) and the complex analyte G.B. (for pos. and neg.) were furthermore spin-coated on ITO-coated glass slides with P(TNDIT-FI(C₁₀C₈)), P(TNDIT-FI(C₄C₂)) and PNDI(T2) as matrices. The sample preparation was carried out in an identical fashion for all spin-coated samples (see the Supporting Information for more details). The MALDI MS spectra of all LMWCs shown in the Supporting Information are plotted in four-stacked graphs, measured under equal conditions: a) MALDI MS spectrum of the pure MALDI sample plate, b) MALDI MS spectrum of the pure analyte, c) MALDI MS spectrum of the pure matrix and d) MALDI MS spectrum of the analyte coated with the matrix.

With both co-polymers, all "dried-droplet" LMWC samples in positive mode were reliably and reproducibly measured at low laser intensities: 10 % for RP (Figures S11 and S17) and TEAC (Figures S12 and S18) and 20 % for CP (Figures S13 and S19). Comparing spectra b) and d) confirm that the co-polymer matrices are a crucial prerequisite to detect the analytes. For both RP measurements a hydride elimination is observed, leading to a peak at $m/z = 607$ Da $[\text{RP-H}]^+$.^[30,33] In the CP spectrum of P(TNDIT-FI(C₁₀C₈)), two signals at $m/z = 307$ Da and 338 Da are visible in addition to the molecular mass peak, while the spectrum using P(TNDIT-FI(C₄C₂)) displays only the additional peak at 307 Da. Both signals are not observable in any other analyte spectra using the same matrices, and are therefore assigned to the analyte CP. Despite of the low laser intensities, fragments of RP and TEAC are observed, listed in Table S3, and fragments of the co-polymers' alkyl side-chains are visible at $m/z = 43, 57, 71$ and 85 Da. Some spectra show the alkali potassium at $m/z = 39$ Da $[\text{K}^+]$, which is a common artifact in positive mode measurements due to trace amounts in analytical solvents.^[34]

The measurements demonstrate that both co-polymers function as very good matrices for positive reflectron mode MALDI despite not showing any crystallinity. Also, beside the alkyl side-chain fragments, the co-polymers are "MALDI-silent" in positive mode measurements, enabling the analysis of LMWCs. A comparison between the measurements of RP using the amorphous co-polymer matrices and the established, semi-crystalline polymer matrix PNDI(T2) (see Figure 3), prepared via spin-coating, reveals that both amorphous co-polymers show higher RP signal intensities at $m/z = 607$ Da than the semi-crystalline PNDI(T2). The co-polymer P(TNDIT-FI(C₄C₂)) exhibits the highest RP signal intensity but also shows two fragments of RP at $m/z = 397$ Da $[\text{RP1}]^+$ and 395 Da $[\text{RP3}]^+$ (see Table S3), which are not detected with PNDI(T2) and P(TNDIT-FI(C₁₀C₈)). Moreover, despite the high laser-intensities used for the spin-coated measurements (40 %) compared with the "dried-droplet" measurements (10 %), no side-chain fragments ($m/z = 43, 57, 71, 85$ Da) are observable in either of the spin-coated measurements. The differences are likely caused by the change in sample preparation, as spin-coating gives more uniform analyte distribution. In summary, for the positive mode measurements

Figure 3. Reflectron positive mode MALDI MS spectra of spin coated samples: (a) PNDI(T2) coated on RP (red), (b) P(TNDIT-FI(C₄C₂)) coated on RP (blue) and (c) P(TNDIT-FI(C₁₀C₈)) coated on RP (black). Peaks: $m/z = 607$ [RP-H]⁺; 397 RP fragment [RP1]⁺; 395 RP fragment [RP3]⁺.

Figure 4. Reflectron negative mode MALDI MS spectra of spin-coated samples: (a) PNDI(T2) coated on CiA (red), (b) P(TNDIT-FI(C₄C₂)) coated on CiA (blue) and (c) P(TNDIT-FI(C₁₀C₈)) coated on CiA (black). Peaks: $m/z = 191$ [CiA-H]⁻; 155 [unknown]; 129 CiA fragment [CiA2]⁻; 111 [unknown]; 87 CiA fragment [CiA1]⁻; 85 side-chain fragment [C₃H₁₂]⁻; 57 side-chain fragment [C₃H₈]⁻; 42 side-chain fragment [C₃H₆]⁻; 26 Acetonitrile fragment [CN]⁻.

of RP the semi-crystallinity of PNDI(T2) exhibits no advantages compared with the amorphous co-polymers regarding analyte signal intensity.

Similarly, both co-polymers allowed reliable and reproducible measurements in negative mode, and all three LMWCs were detected at low laser intensities: ChA (Figures S14 and S20) and CiA at 20% (Figures S15 and S21) and PA at 30% (Figures S16 and S22). Here, a fragment is only observable in the measurement of CiA at $m/z = 87$ Da [CiA1]⁻ (see Table S3). Also only one alkyl side-chain fragment is visible at $m/z = 42$ Da. Similar to the positive mode measurements, bromide is observed at $m/z = 79/81$ Da [Br]⁻, which is originating from the analytical solvents used for sample preparation.^[34] The co-polymer matrices do not lead to molecule ions, clusters or backbone fragments, verifying that they are not decomposing and "MALDI silent" during negative measurements.

Spin-coated CiA samples were prepared to compare the performances of the amorphous conjugated polymer matrices and the established, semi-crystalline PNDI(T2) matrix (see Figure 4). Similar to the positive mode measurements, both measurements with the amorphous co-polymer matrices show higher CiA signal intensities at $m/z = 191$ Da compared to the semi-crystalline PNDI(T2), with P(TNDIT-FI(C₄C₂)) revealing the highest intensity. As observed for the positive mode measurements of RP, the negative mode measurement of CiA using the semi-crystalline PNDI(T2) does not lead to side-chain related signals despite a higher laser intensity (60%) compared to the "dried-droplet" sample measurements (20%), and for the amorphous P(TNDIT-FI(C₁₀C₈)), only a weak signal at $m/z = 42$ Da is observed. On the other hand, several additional signals are observed in the P(TNDIT-FI(C₄C₂))-spectrum, which are assigned to be CiA fragments ($m/z = 87$ Da [CiA1]⁻ and 129 Da [CiA2]⁻) or side-chain fragments from the polymer matrix ($m/z = 42$ Da [C₃H₆]⁻, 57 Da [C₄H₈]⁻ and 85 Da [C₆H₁₂]⁻), as well as solvent-based impurities from the cleaning process of the sample plate. The spectrum also contains signals with unknown origins

($m/z = 111, 155$ Da). A comparison of matrix-analyte spectra with pure matrix spectra measured on the same sample shows that the unknown signals at $m/z = 111$ and 155 Da are not matrix fragments or matrix impurities (see Figure S29).

In summary, the negative mode measurements of CiA are consistent with the results of the positive mode measurements; the semi-crystalline morphology of PNDI(T2) does not constitute an advantage with regards to the intensity of the analyte signals.

To explore the potential of the amorphous polymer matrices to analyze complex biological samples with regard to MALDI MSI analytics, a commercially available Ginkgo Biloba extract (G.B.) was measured in positive and negative mode. G.B. contains a high number of different LMWCs, e.g. flavonoids and terpene lactones, and was already investigated using different MS methods.^[35] The measurements of the "dried-droplet" samples using both co-polymer matrices were performed at low laser intensities, at 30% for positive mode and 20% for negative mode, comparable with the single LMWC analytics. In positive mode, the average number of signals with $l \leq 100$ is 2982 with an average intensity $\bar{I} = 1.22 \times 10^3$ for P(TNDIT-FI(C₄C₂)), and 2339 for P(TNDIT-FI(C₁₀C₈)) with $\bar{I} = 5.02 \times 10^2$ (see Figures S23 and S25). In negative mode, the average number of signals with $l \leq 100$ is 1595 with $\bar{I} = 2.56 \times 10^2$ for P(TNDIT-FI(C₄C₂)), and 1426 for P(TNDIT-FI(C₁₀C₈)) with $\bar{I} = 3.24 \times 10^2$ (see Figures S24 and S26). The signal numbers in positive mode are in the same range as those reported for the parent homopolymers P9OFI and PNDI(T2),^[23] whereas the signal numbers in negative mode are lower, possibly because of the lower laser intensity used for the experiment. Furthermore, in negative mode the most intense signals are assigned to the G.B. analyte and not to the matrix, as usually found for SOMs. For positive mode, this is not the case for P(TNDIT-FI(C₁₀C₈)) due to a alkyl side-chain signal below $m/z = 100$.

Spin-coated samples of the Ginkgo Biloba extract were prepared to enable a direct comparison between the established semi-crystalline, polymeric matrix PNDI(T2) and the amorphous co-polymers (see Figure 5 for positive mode, Figure 6 for negative mode).

Figure 5. Reflectron positive mode MALDI MS spectra of spin-coated samples: a) PNDI(T2) coated on G.B. (red), b) P(TNDIT-FI(C₄C₂)) coated on G.B. (blue) and c) P(TNDIT-FI(C₁₀C₈)) coated on G.B. (black).

Figure 6. Reflectron negative mode MALDI MS spectra of spin-coated samples: a) PNDI(T2) coated on G.B. (red), b) P(TNDIT-FI(C₄C₂)) coated on G.B. (blue) and c) P(TNDIT-FI(C₁₀C₈)) coated on G.B. (black).

In positive mode measurements the analyte signal intensities ($m/z > 100$) detected with the amorphous co-polymer matrices are increased compared with the semi-crystalline PNDI(T2) matrix, and the P(TNDIT-FI(C₁₀C₈)) spectrum exhibits the highest intensities. The same is observed for the negative mode measurements: Here, the increase of analyte signal intensities is even larger than for the positive mode measurements, resulting in more detected signals in the area of $m/z = 400$ Da–600 Da and around 750 Da. For both polarity modes, higher laser intensities (80%) were used compared with the measurements of the “dried-droplet” samples (pos.: 20%; neg.: 30%), and

interestingly, unlike for the single analyte measurements, side-chain signals are detected. The results confirm the dual-mode ability of the co-polymers, and the comparison to the semi-crystalline PNDI(T2) matrix shows that their lack of crystallinity is not detrimental for the number of detected signals or the average signal intensity.

In summary, despite their amorphous morphologies both co-polymers reveal excellent and “MALDI-silent” performances for single and complex LMWC analytics in both positive and negative mode. In contrast to the parent homopolymer P9OFI, signals related to the co-polymers’ backbones are not observable, as co-polymerizing fluorene with the MALDI-stable NDI(T2) leads to a stable co-polymer.^[30] Moreover, all spin-coated single and complex LMWC measurements in positive as well as negative mode revealed that the semi-crystalline nature of PNDI(T2) exhibits no advantages in matrix performances compared to the amorphous co-polymer matrices: but at least for the polymers investigated so far, an amorphous morphology is advantageous. Furthermore, for all “dried-droplet” measurements, only a very low amount of matrix is required ($m_{\text{Matrix}}/m_{\text{LMWC}} = 0.1$; $m_{\text{Matrix}}/m_{\text{G.B.}} = 1$), leading to simple and cost-efficient sample preparations.

Conclusion

Currently there is a strong interest in establishing reliable and reproducible MALDI methods for the analysis of biological and biochemical LMWCs, necessitating a conscious matrix design. Targeted properties are high absorption at laser wavelength, low or no matrix-related signals in the LMWC area, and ideally the ability to support measurements in both positive and negative mode. For the most commonly used small organic matrices (SOMs), a full crystalline matrix structure is reported to be an obligate prerequisite for successful MALDI measurements. Here, we designed and investigated two amorphous fluorene-NDI co-polymers as MALDI matrices for LMWCs and compared them with the semi-crystalline polymer matrix PNDI(T2). Both P(TNDIT-FI(C₄C₂)) and P(TNDIT-FI(C₁₀C₈)) are fully amorphous based on XRD experiments, and have a high absorption at the wavelength of commonly used MALDI laser systems. The co-polymers are solution- and in the case of P(TNDIT-FI(C₁₀C₈)) melt-processable, allowing to form homogeneous films. Additionally, both co-polymers reveal similar electrochemical behaviors with two reduction potentials for good ionization efficiencies and are vacuum stable. Despite their amorphous morphologies, both co-polymers function as excellent MALDI matrices during analyses of “dried-droplet” single and complex LMWCs. Reliable and reproducible MALDI spectra were recorded at low laser intensities (10%–30%) using a very low amount of matrix. Furthermore, in contrast to the parent homopolymer P9OFI, the co-polymers’ backbones do not fragment and are not visible in the low mass range. Consequently, a co-polymerization of the stable NDI(T2) unit with the instable fluorene monomer leads to MALDI stable co-polymers. The comparison between the newly synthesized amorphous co-polymers and the semi-crystalline PNDI(T2) using spin-coated

samples showed that for the LMWC samples investigated here the detected analyte signal intensities are higher when using the amorphous co-polymer matrices compared of the semi-crystalline PNDI(T2) matrix.

This leads to the conclusion that a (semi-)crystalline morphology of the organic or polymeric matrix is not obligate to enable reliable and reproducible MALDI MS measurements. These results significantly broaden the design rules for new and efficient polymeric MALDI matrices, and also offer further opportunities to explore the relationship between crystallinity and matrix performance.

Experimental Section

All monomers (see Scheme 1) were synthesized according to literature,^[36] and polymerization was performed via standard Suzuki cross coupling polymerization.^[37] The polymers were characterized with ¹H-NMR (500 MHz, CDCl₃), gel permeation chromatography (GPC) (CHCl₃, 40 °C), thin-film cyclovoltammetry (CV) and linear positive mode MALDI MS (0–200k Da). Crystallinity was investigated with grazing-incidence wide-angle X-ray scattering (GIWAXS) X-ray diffraction (XRD) on spin and drop casted layers, and additionally P (TNDIT-Fl(C₁₀C₈)) was investigated with thermal gravimetry analyses (TGA), differential scanning calorimetry (DSC) and 2D wide angle X-ray scattering (2D-WAXS). The absorption behavior was tested with UV/VIS on drop-casted layers. More details on synthetic procedures and characterization data can be found in the Supporting Information.

All MALDI MS spectra were recorded with a Bruker Autoflex Speed MALDI-TOF/TOF instrument using a Smartbeam-II Nd:YAG laser ($\lambda_{\text{laser}} = 355 \text{ nm}$). For measurements and evaluations, the software tools flexControl 3.4 and flexAnalysis 3.4 from Bruker Daltonics GmbH were used. All MALDI measurements of the LMWCs were carried out in positive or negative reflectron mode, whereas for the analytical MALDI measurements of the co-polymers the linear positive mode was used. Before measuring, the instrument was calibrated with red phosphorus for reflectron mode or with a 2,5-DHB/PEG solution for linear mode (see Supporting Information for details). The LMWC MALDI samples were prepared according to the “dried-droplet” method^[38] and spin-coating method using self-established parameters.^[39] The co-polymers were dissolved in toluene ($c = 0.1 \text{ mg/ml}$) whereas the analytes were dissolved in suitable solvents ($c = 1 \text{ mg/ml}$), mostly in tetrahydrofuran. For the spin-coating method the same solutions with $c = 10 \text{ mg/ml}$ were used. Further detailed information is given in the Supporting Information.

Acknowledgements

F.L. thanks the FCI for a Liebig Fellowship. K.-N.Z. gratefully acknowledges a scholarship under the Chinese State Scholarship Fund for outstanding students abroad by the China Scholarship Council (CSC). The work was in part supported by KAUST Center Partnership Fund OSR-2017-CPF-3322-01.

Conflict of Interest

The authors declare no conflict of interest.

Keywords: amorphous morphology • conjugated polymers • MALDI mass spectrometry • low-molecular-weight analytes • matrices

- [1] a) M. Karas, D. Bachmann, F. Hillenkamp, *Anal. Chem.* **1985**, *57*, 2935–2939; b) M. Karas, D. Bachmann, U. Bahr, F. Hillenkamp, *Int. J. Mass Spectrom. Ion Processes*, **1987**, *78*, 53–68; c) K. Tanaka, H. Waki, Y. Ido, S. Akita, Y. Yoshida, T. Yoshida, *Rapid Commun. Mass Spectrom.* **1988**, *2*, 151–153.
- [2] a) A. E. Clark, E. J. Kaleta, A. Arora, D. M. Wolk, *Clin. Microbiol. Rev.* **2013**, *26*, 547–603; b) A. Croxatto, G. Prod'homme, G. Greub, *FEMS Microbiol. Rev.* **2012**, *36*, 380–407; c) M. W. Duncan, D. Nedelkov, R. Walsh, S. Hattian, *J. Clin. Chem. Commun.* **2016**, *62*, 134–143; d) R. Patel, *Clin. Chem.* **2015**, *61*, 100–111; e) M. A. M. Rodrigo, O. Zitka, S. Krizkova, A. Moullick, V. Adam, R. Kizek, *J. Pharm. Biomed. Anal.* **2014**, *95*, 245–255.
- [3] K. G. Owens, S. D. Hanton, *Conventional MALDI sample preparation*, Vol. 175, p. 129–158, John Wiley & Sons, Inc., **2011**.
- [4] a) R. Knochenmuss, *Electrospray MALDI Mass Spectrom.* **2012**, 147–183; b) M. Karas, R. Krüger, *Chem. Rev.* **2003**, *103*, 427–439; c) T. W. Jaskolla, M. Karas, *J. Am. Soc. Mass Spectrom.* **2011**, *22*, 976–988.
- [5] R. Arakawa, H. Kawasaki, *Anal. Sci.* **2010**, *26*, 1229–1240.
- [6] Q. Liu, V. L. Brown, L. He, *Porous Silicon for Biomedical Applications*, Vol. 68, p. 270–285, Elsevier, **2014**.
- [7] a) M. Karas, F. Hillenkamp, *Anal. Chem.* **1988**, *60*, 2299–2301; b) M. A. R. Meier, R. Hoogenboom, M. W. M. Fijten, M. Schneider, U. S. Schubert, *J. Comb. Chem.* **2003**, *5*, 269–374; c) V. Karbach, R. Knochenmuss, *Rapid Commun. Mass Spectrom.* **1998**, *12*, 968–974.
- [8] a) M. Lagarrigue, R. M. Caprioli, C. J. Pineau, *Proteomics* **2016**, *144*, 133–139; b) C. D. Calvano, A. Monopoli, T. R. I. Cataldi, F. Palmisano, *Anal. Bioanal. Chem.* **2018**, *410*, 4015–4038.
- [9] a) J. L. Norris, R. M. Caprioli, *Chem. Rev.* **2013**, *113*, 2309–2342; b) P. Chaurand, M. Stoekli, R. M. Caprioli, *Anal. Chem.* **1999**, *71*, 5263–5270.
- [10] a) H. Liu, R. Chen, J. Wang, S. Chen, C. Xiong, J. Wang, J. Hou, Q. He, N. Zhang, Z. Nie, L. Mao, *Anal. Chem.* **2014**, *86*, 10114–10121; b) J. S. H. Wang, M. Freitas-Andrade, J. F. Bechberger, C. C. Naus, K. K. C. Yeung, S. N. Whitehead, *Rapid Commun. Mass Spectrom.* **2018**, *32*, 951–958.
- [11] K. Strupat, M. Karas, F. Hillenkamp, *Int. J. Mass Spectrom. Ion Processes* **1991**, *111*, 89–102.
- [12] R. L. Vermillion-Salsbury, D. M. Hercules, *Rapid Commun. Mass Spectrom.* **2002**, *16*, 1575–1581.
- [13] a) V. Horneffer, A. Forsmann, K. Strupat, F. Hillenkamp, U. Kubitscheck, *Anal. Chem.* **2001**, *73*, 1016–1022; b) T. Nishikaze, H. Okumura, H. Jinmei, J. Amano, *Mass Spectrom.* **2012**, *1*, A0006.
- [14] M. Körsen, A. Pelster, K. Dreisewerd, H. F. Arlinghaus, *J. Am. Soc. Mass Spectrom.* **2016**, *27*, 277–284.
- [15] J. Axelsson, A.-M. Hoberg, C. Waterson, P. Myatt, G. L. Shield, J. Varney, D. M. Haddleton, P. J. Derrick, *Rap. Commun. in Mass Spec.* **1997**, *11*, 209–213.
- [16] S. L. Cohen, B. T. Chait, *Anal. Chem.* **1996**, *68*, 31–37.
- [17] a) Y. Dai, R. M. Whittall, L. Li, *Anal. Chem.* **1996**, *68*, 2494–2500; b) T. Fujita, T. Fujino, *Chem. Lett.* **2013**, *42*, 350–351.
- [18] M. Nazari, D. C. Muddiman, *Analyst* **2016**, *121*, 595–605.
- [19] a) D. M. G. Anderson, Z. Ablonczy, Y. Koutalos, J. Spraggins, R. K. Crouch, R. M. Caprioli, K. L. Schey, *J. Am. Soc. Mass Spectrom.* **2014**, *25*, 1394–1403; b) A. Thomas, J. L. Charbonneau, E. Fournaise, P. Chaurand, *Anal. Chem.* **2012**, *84*, 2048–2054; c) H. Kettling, S. Vens-Cappell, J. Soltwisch, A. Pirkel, J. Haier, J. Mühling, K. Dreisewerd, *Anal. Chem.* **2014**, *86*, 7798–7805.
- [20] a) R. Knochenmuss, F. Dubois, M. J. Dale, R. Zenobi, *Rapid Commun. Mass Spectrom.* **1996**, *10*, 871–877; b) G. McCombie, R. Knochenmuss, *Anal. Chem.* **2004**, *76*, 4990–4997.
- [21] A. Monopoli, C. D. Calvano, A. Nacci, F. Palmisano, *Chem. Commun.* **2014**, *50*, 4322.
- [22] M. L. Manier, J. M. Spraggins, M. L. Reyzer, J. L. Norris, R. M. Caprioli, *J. Mass Spectrom.* **2014**, *49*, 665–673.
- [23] A. L. M. Marsico, B. Creran, B. Duncan, S. G. Elci, Y. Jiang, T. B. Onasch, J. Wormhoudt, V. M. Rotello, R. W. Vachet, *J. Am. Soc. Mass Spectrom.* **2015**, *26*, 1931–1937.
- [24] M. W. Towers, R. Cramer, *Springer International Publishing: Cham* **2016**, pp 51–64.
- [25] a) X. Dong, J. Cheng, J. Li, Y. Wang, *Anal. Chem.* **2010**, *82*, 6208–6214; b) Y. Liu, J. Liu, P. Yin, M. Gao, C. Deng, X. J. Zhang, *Mass Spectrom.* **2011**, *46*, 804–815.

- [26] H. Zhao, Y. Li, J. Wang, M. Cheng, Z. Zhao, H. Zhang, C. Wang, J. Wang, Y. Qiao, J. Wang, *Appl. Mater. Interfaces* **2018**, *10*, 37732–37742.
- [27] Z. Wang, Y. Cai, Y. Wang, X. Zhou, Y. Zhang, H. Lu, *Sci. Rep.* **2017**, *7*, 44466.
- [28] N. Dittrich, J. Zuber, P. Rathack, M. Otto, *Int. J. Mass Spectrom. Ion Processes* **2018**, *433*, 31–39.
- [29] M. C. Bernier, V. H. Wysocki, S. Dagan, *J. Mass Spectrom.* **2015**, *50*, 891–898.
- [30] K. Horatz, M. Giampà, Y. Karpov, K. Sahre, H. Bednarz, A. Kiri, B. Voit, K. Niehaus, N. Hadjichristidis, D. L. Michels, F. Lissel, *J. Am. Chem. Soc.* **2018**, *140*, 11416–11423.
- [31] a) J. Wang, M. Ueda, T. Higashihara, *J. Polym. Sci. Part A* **2014**, *52*, 1139–1148; b) X. H. Li, X. Z. Shen, X. Gaob, Y. Y. Weng, *RSC Adv.* **2017**, *7*, 55885–55890.
- [32] Y. Zhao, X. Zhao, Y. Zang, C.-a. Di, Y. Diao, J. Mei, *Macromolecules* **2015**, *48*, 2048–2053.
- [33] H. Awad, M. J. Stoudemayer, L. Usher, I. J. Amster, A. Cohen, U. Das, R. M. Whittall, J. Dimmock, A. El-Aneed, *J. Mass Spectrom.* **2014**, *49*, 1139–1147.
- [34] B. O. Keller, J. Suib, A. B. Young, R. M. Whittall, *Anal. Chim. Acta* **2008**, *627*, 71–81.
- [35] a) S. Beck, J. Stengel, *Phytochemistry* **2016**, *130*, 201–206; b) S. Ding, E. Dudley, S. Plummer, J. Tang, R. P. Newton, A. G. Brenton, *Phytochemistry* **2008**, *69*, 1555–1564; c) S. Ding, E. Dudley, Q. Song, S. Plummer, J. Tang, R. P. Newton, A. G. Brenton, *Rapid Commun. Mass Spectrom.* **2008**, *22*, 766–772.
- [36] a) R. Tkachov, Y. Karpov, V. Senkovskyy, I. Raguzin, J. Zessin, A. Lederer, M. Stamm, B. Voit, T. Beryozkina, V. Bakulev, W. Zhao, A. Facchetti, A. Kiri, *Macromolecules* **2014**, *47*, 3845–3851; b) Y. Karpov, J. Maiti, R. Tkachov, T. Beryozkina, V. Bakulev, W. Liu, H. Komber, U. Lappan, M. Al-Hussein, M. Stamm, B. Voit, A. Kiri, *Polym. Chem.* **2016**, *7*, 2691–2697; c) R. Yang, R. Tian, Q. Hou, W. Yang, Y. Cao, *Macromolecules* **2003**, *36*, 7453–7460.
- [37] a) R. Tkachov, H. Komber, S. Rauch, A. Lederer, U. Oertel, L. Häußler, B. Voit, A. Kiri, *Macromolecules* **2014**, *47*, 4994–5001; b) T. Prenville, Master Thesis *Synthesis and characterization of melt-processable semi-conductive polymers*, **2015**, TU Dresden/Leibniz Institut für Polymerforschung Dresden e.V.
- [38] a) V. Ole, P. Roepstorff, M. Mann, *Anal. Chem.* **1994**, *66*, 3281–3287; b) M. Kussmann, E. Nordhoff, H. Rahbek-Nielsen, S. Haebel, M. Rossel-Larsen, L. Jakobsen, J. Gobom, E. Mirgorodskaya, A. Kroll-Kristensen, L. Palmi, P. Roepstorff, *J. Mass Spectrom.* **1997**, *32*, 593–601; c) Y. Dai, R. M. Whittall, L. Li, *Anal. Chem.* **1999**, *71*, 1087–1091.
- [39] I. K. Perera, J. Perkins, S. Kantartzoglou, *Rapid Commun. Mass Spectrom.* **1995**, *9*, 180–187.

Manuscript received: March 30, 2019
Revised manuscript received: June 14, 2019
Accepted manuscript online: June 18, 2019

FULL PAPERS

The answer is out there: The matrix is crucial for MALDI measurements and good organic matrices are usually highly crystalline. Two amorphous co-polymers P(TNDIT-FI (C_4C_2)) and P(TNDIT-FI($C_{10}C_8$)) were investigated regarding their suitability as MALDI matrices for the analysis of single and complex low-molecular-weight compounds. Both co-polymers show excellent performance as dual-mode MALDI matrices, suggesting that analyte incorporation takes place in the amorphous phase.

K. Horatz, K. Ditte, Dr. T. Prenveille, K.-N. Zhang, Dr. D. Jehnichen, Dr. A. Kiri, Prof. Dr. B. Voit, Dr. F. Lissel*

1 – 9

Amorphous Conjugated Polymers as Efficient Dual-Mode MALDI Matrices for Low-Molecular-Weight Analytes

7.3. Polymerization as a Strategy to Improve Small Organic Matrices for Low Molecular Weight Compound Analytics with MALDI MS and MALDI MS Imaging (Publication 3)

ACS APPLIED
POLYMER MATERIALS

pubs.acs.org/acsapm

Article

Polymerization as a Strategy to Improve Small Organic Matrices for Low-Molecular-Weight Compound Analytics with MALDI MS and MALDI MS Imaging

Kilian Horatz, Marco Giampà, Zhi Qiao, Siver A. Moestue, and Franziska Lissel*

Cite This: *ACS Appl. Polym. Mater.* 2021, 3, 4234–4244

Read Online

ACCESS |

Metrics & More

Article Recommendations

Supporting Information

ABSTRACT: Matrix-assisted laser desorption/ionization mass spectrometry (MALDI MS) and the corresponding visualizing technique MALDI MS imaging (MSI) are potent and widely used analytical methods in medical and pathological research. In recent years, the investigation of low-molecular-weight compounds (LMWCs) such as metabolites has moved increasingly into the focus. MALDI techniques require a matrix system, and small organic matrices (SOMs) are commonly used. While SOMs offer multiple advantages, such as broad analyte scopes and high ionization efficiencies, they also suffer from drawbacks, e.g., strong background interferences in the low-mass area ($m/z < 1000$) and low vacuum stability, which is particularly detrimental for LMWC analytics with high vacuum (HV) MALDI MS and MSI. Here, we apply polymerization as a strategy to alleviate these drawbacks while retaining the multiple advantages of SOMs. Vinyl groups were introduced to two SOMs, the state-of-the-art positive mode matrix 2,5-dihydroxybenzoic acid (DHB) as well as one of the few known dual polarity mode matrices, 7-methoxy-1-methyl-9H-pyrido[3,4-*b*]indole (harmine), and radical polymerization was performed to obtain polyethylene-based P(SOMs) carrying the corresponding SOMs as side chains. Compared to the corresponding SOMs, the synthesized P(SOMs) maintain optical properties in the solid state and have competitive performances regarding analyte scopes, ionization efficiencies, and dual polarity mode suitability. Additionally, both P(SOMs) are HV stable ($\sim 10^{-7}$ mbar) and reveal no background interferences in the low-mass area (MALDI-silent). To assess a potential application in a clinical workflow, the P(SOMs) were applied on breast cancer xenografts and MALDI MSI measurements were carried out, demonstrating their ability to produce and spatially resolve positive and negative tissue-related ions directly from the cancer tissue. Polymerization is shown to be a promising strategy to make state-of-the-art SOMs MALDI silent and vacuum stable and yield easily handled matrices for clinical workflows.

KEYWORDS: MALDI MS imaging, tumor xenograft, polymer matrix, metabolomics, free-radical polymerization

INTRODUCTION

“Matrix-assisted laser desorption/ionization mass spectrometry” (MALDI MS) and the corresponding visualization technique MALDI MS imaging (MSI) are important analytical tools, e.g., for biochemical and medical research.^{1–3} In contrast to similar MS techniques, for instance, secondary ion mass spectrometry (SIMS) or electrospray ionization (ESI), the analytes are incorporated in a protecting matrix, which allows us to ionize them softly, preventing fragmentation or degradation. Commonly used matrix systems are small organic matrices (SOMs), which fulfill the prerequisites of the MALDI process, namely, strong absorption of laser radiation, high crystallinity, and acidic/basic functional groups to assist analyte ionization. Additional advantages are their broad analyte scope and excellent ionization efficiencies, leading to a wide acceptance in the research community. SOMs can produce matrix-related signals, such as molecule ions, clusters, and

fragments, which hinder noise-free (MALDI silent) measurements below $m/z = 1000$. While such cluster signals can be used as internal standards (lock mass) to avoid mass shifts,⁴ MALDI-silent matrix systems are preferable: they allow us to freely choose sample-specific standards, and furthermore, the occurrence of multiple strong matrix-related signals can be detrimental for certain measurements, especially when analyzing low-molecular-weight compounds (LMWCs), e.g., in drug assays or disease research.^{5–7} Also, SOMs often exhibit

Received: May 31, 2021

Accepted: June 22, 2021

Published: July 7, 2021

low stability in high vacuum (HV) ($\sim 10^{-7}$ mbar) due to their low molar mass, which can result in an inhomogeneous matrix layer during long-term MALDI MSI measurements.⁸ A number of developments on the instrument side were reported in recent years, most notably innovations moving away from the HV centered instrumentation: For example, modern trapped ion mobility spectrometry (TIMS), orbitrap, and Fourier transform ion cyclotron resonance (FTICR) instruments allow a confident assignment of LMWCs under a subatmospheric (subAP) ion source pressure, preventing the sublimation of non-vacuum-stable matrices or analytes.^{9–11} Moreover, the detection and accurate identification of metabolites in complex tissue samples are performed using high-mass resolving power mass spectrometers, such as FTICR that require parametrical optimization and longer acquisition time, which is a setback especially in clinical and routine settings based on MSI procedures.^{12,13} However, recently, a subspace method was proposed to accelerate the data acquisition in FTICR MSI.¹⁴ MALDI orthogonal time-of-flight (TOF) instruments can also operate under atmospheric pressure (AP MALDI MS), hence not requiring vacuum-stable matrix systems.¹⁵ Yet, it was reported that significantly more mass signals are detectable with HV MALDI MS, decreasing the scope and applicability of AP MALDI MS.¹⁶ Consequently, developing strategies to improve SOMs by alleviating the disadvantages connected with their vacuum instability while retaining the multiple advantageous properties could be especially beneficial for standard operators in clinical settings.

Classic SOMs for MALDI MS are organic molecules with a molar mass of <500 Da, which contain aromatic (hetero)-cycles. The conjugated π -electron system allows sufficient absorption in the UV region, the heterocyclic structure or other acidic/basic functional groups increase the ionization efficiency by (de-)protonating the analytes, and a highly crystalline morphology ensures an efficient analyte incorporation. A good example is 2,5-dihydroxybenzoic acid (DHB), which is an excellent MALDI matrix and has been intensively studied regarding analyte scope, ionization efficiency, cluster formation, and MALDI sample preparation properties.^{17–21} As DHB contains an acidic functional group, it is predominantly used for positive mode measurements, but rare cases of experiments in negative mode were reported, albeit yielding lower signal intensities.^{22,23} Because of the wide range of accessible analytes and the superior matrix performance, DHB is an ubiquitously used standard matrix for modern tissue analytics with MALDI MSI and is frequently employed as a benchmark during the development or improvement of matrix systems.^{24,25} Nevertheless, DHB produces a number of matrix-related signals below $m/z = 1000$ (e.g., molecule ions, fragments, and clusters) and tends to sublime in HV.⁸ Several strategies were developed to counteract these disadvantages and enhance DHB's efficacy for LMWC analytics with MALDI MS and MSI. Adding co-matrix systems or covalently incorporating DHB in a sol–gel silicon-dioxide polymer decreased the background interferences.^{26–28} Regarding the HV stability of DHB, it was reported that the sublimation rate of DHB is highest in the first hour under HV and that the sublimation rate can be decreased by converting DHB into an ionic liquid.^{8,29}

7-Methoxy-1-methyl-9H-pyrido[3,4-*b*]indole (harmine) is an effective matrix used for MALDI analytics of (sulfated) oligosaccharides, proteins, and carbohydrates.^{30–33} While it is predominantly used for negative mode, harmine also facilitates

positive mode measurements, making it a rare dual polarity mode matrix. This is caused by the presence of a weakly acidic (indolic) and a basic (pyridinic) functional group, which results in the ionization of the analyte in the gas phase by protonation as well as deprotonation.³¹ The dual polarity mode ability is rare for SOMs, and especially when only low analyte amounts are available, it can be advantageous compared to predominantly mono-mode matrices such as DHB. The ionization efficiency of harmine as a negative mode MALDI matrix is increased when ammonium chloride is added, which is advantageous for, e.g., the efficient screening of carbohydrates.^{32,33} Yet, like most SOMs, harmine produces many matrix-related signals in the low-mass area, such as molecule ions and clusters, and tends to sublime in HV.

Next to the efforts to improve well-known SOMs, research is carried out to develop SOMs with enhanced MALDI matrix performances, e.g., to enable a specific analyte scope, to introduce suitability for dual polarity mode, increased ionization efficiencies, or low background interferences in the low-mass area.^{25,34,35} Among others, a specific analyte scope was achieved by employing reactive functional groups such as amine-groups, which react only with compounds carrying aldehyde groups.³⁶ Dual mode suitability was realized by developing structures containing acidic and basic functional groups, to ionize the analytes either by protonation or deprotonation.^{37,38} Utilizing high-molecular-weight matrices can enable LMWC measurements, as shown for, e.g., porphyrins,³⁹ oligothiophenes,^{40,41} and polythiophenes.⁴² We recently reported that conjugated homo- and heteropolymers are promising matrices for LMWC analytics with MALDI MS and MSI as they combine very low background interferences with HV stability and good ionization efficiencies in positive and negative modes.⁴³ Furthermore, different from classic SOMs, for semicrystalline conjugated polymers, the analyte incorporation seems to take place in the amorphous region.⁴⁴

Here, we apply polymerization as a new strategy to improve established SOMs. Adding vinyl groups to two well-known SOMs, DHB and harmine, followed by free-radical polymerization, resulted in two polymeric small organic matrices (P(SOMs)) carrying the respective SOMs as side chains: poly(vinyl dihydroxybenzoic acid) (P(VDHB)) and poly(vinyl harmine) (P(VHar)). Their optical properties, vacuum stabilities, and performances as MALDI matrices were tested and compared with the corresponding SOMs. Finally, the P(SOMs) were applied as matrices to acquire MALDI MSI heatmaps of patient-derived xenograft tumors, to evaluate their potential for application in preclinical studies of breast cancer. The results show that P(SOMs) retain the matrix performances of the corresponding SOMs, while exhibiting additional properties (i.e., making them vacuum stable and MALDI silent), and demonstrate that polymerization can be a promising strategy for improving SOMs.

■ EXPERIMENTAL SECTION

All precursors, reagents, and solvents were purchased from commercial suppliers and used without further purification. Detailed information about syntheses and experiments (NMR, FTIR, and UV–vis) is given in the Supporting Information. Number-average and mass-average molecular weights as well as dispersity and the degree of polymerization (M_n , M_w , $\bar{D}$, and DP) were first investigated with gel permeation chromatography (GPC) in dimethylacetamide (DMAc), resulting in absolute values as a molar-mass-sensitive light scattering detector was used.⁴⁵ Also, MALDI MS was used for the determination of the aforementioned polymeric properties, applying *trans*-2-[3-(4-

tert-butylphenyl)-2-methyl-2-propenyldiene]malono-nitrile (DCTB), a well-established secondary matrix for synthetic polymers such as polyvinyl polymers.⁴⁶ Additionally, the MALDI MS measurements were used to analyze the molar mass of the repeating unit and the end groups of the polymers.

All single analyte and limit of detection (LoD) MALDI MS measurements were performed on a Bruker Autoflex Speed MALDI TOF/TOF instrument equipped with a Smartbeam II/Nd:YAG laser ($\lambda = 355$ nm). The commercially available lipid, saccharide, and amino acid samples were prepared by spotting a drop of sample solution ($V_{\text{drop}} = 1 \mu\text{L}$) on the sample target, followed by a drop of matrix ($V_{\text{drop}} = 1 \mu\text{L}$) ("layered" dried-droplet method; see Section 2.6.4 in the Supporting Information for a detailed information). The concentrations of matrix and analyte solutions are depicted in the individual mass spectra. All analytes were measured in positive as well as negative reflectron modes to assess and compare the dual polarity mode suitability of SOMs and P(SOMs). LoD measurements of choline chloride (CC), phosphocholine (PC), and glycerophosphorylcholine (GPC) were carried out in both polarity modes to compare the ionization efficiencies of SOMs and P(SOMs) (three individual measurements per polarity mode and per sample). Additionally, the LoD measurements of the P(SOMs) in negative mode were compared to those of the standard SOM 9-aminoacridine (9AA). Detailed information about the molar mass, molecular structures, and proposed fragments of all analytes is given in Tables S1 and S2. The resulting spectra were plotted in six-stacked graphs for a direct comparison between SOMs and P(SOMs) with, from top to bottom, mass spectra of the pure sample plate (violet), the pure analyte (yellow), the pure SOM (green), the analyte coated with SOM (red), the pure P(SOM) (blue), and the analyte coated with P(SOM) (black). All spectra within a stacked plot were measured under the exactly same conditions to allow comparison.

The MAS98.06 patient-derived breast cancer xenograft models were established at the Institute of Cancer Research, Oslo University Hospital, as previously described.⁴⁷ Tumor tissue was orthotopically implanted into 6 week old female Hsd:ATHymic Nude-Foxn1nu mice. The animals were kept under pathogen-free conditions and received supplementary 17- β -estradiol (4 $\mu\text{g}/\text{mL}$) in the drinking water. When the tumors reached a size of 200–300 mm³, the mice received CB-839 (200 mg/kg) or placebo two times daily for 28 days, before tumors were harvested and snap-frozen in liquid nitrogen until further processing. All procedures and experiments involving animals were approved by the Norwegian Animal Research Authority (FOTS IDs: 7713 and 9126) and carried out according to the European Convention for the Protection of Vertebrates Used for Scientific Purposes (directive 2010/63/EU). The biopsy was divided in two equal parts: one part was formalin-fixed and paraffin-embedded. Sections of 4 μm were stained using a standard protocol for hematoxylin and eosin (H&E) staining to obtain high-quality morphological information, which was not achievable with the corresponding fresh-frozen section. The other part was stored at -80°C and used for the following MALDI MSI procedure. Cryosections of 10 μm , obtained using a cryostat (Leica CM 3050S), were mounted onto ITO slides (MALDI MSI diagnostic) and stored at -80°C . Prior to the matrix application, the sections were vacuum-dried for 10 min. A matrix solution (0.5 mg/mL P(VDHB) dissolved in MeOH/H₂O/acetone (50/30/20) and P(VHar) in MeOH/H₂O (75/25), respectively) was applied using an M5-HTX sprayer and the following parameters: temperature = 80°C ; number of passes = 6, flow rate = 0.075 mL/min, velocity = 1500 mm/s, track spacing = 2, pattern = criss cross, pressure = 10 psi, gas flow rate = 3, and dry time = 40 s. All MALDI MSI measurements were performed using a RapifleX MALDI TissueTyper mass spectrometer (Bruker Daltonics, Bremen, Germany) equipped with a Smartbeam 3D laser under "Single" mode and with a raster rate of 1.25. The analyzer was operated in negative and positive reflectron modes, and the laser was fired at a repetition rate of 10 kHz. External calibration was performed using red phosphorous dispersed in acetone, spotted beside the tissue section. Positive mode measurements were acquired in the calibrated range from $m/z = 61$ to 1400, whereas negative mode measurements were acquired in the range

from $m/z = 40$ to 1200. Further instrumental details are shown in Table S4. The samples were rastered at a lateral resolution of $70 \times 70 \mu\text{m}^2$ (x, y) with a laser scan range of 66 μm per pixel. The MALDI MSI heatmaps of histological sections were processed and evaluated using SCiLS Lab, version 2020a Pro, and were normalized to the total ion count (TIC). A weak denoising of the heatmaps was performed for a better visualization.

RESULTS AND DISCUSSION

Design, Synthesis, and Characterization. A polyethylene backbone was chosen because of its high stability and chemical inertness, as well as the facile accessibility via radical polymerization. DHB and 7-methoxy-1-methyl-9H-pyrido[3,4-*b*]indole (harmine) were chemically functionalized with vinyl groups and used as monomers to obtain the corresponding polymers.

For the synthesis of poly(vinyl dihydroxybenzoic acid) (P(VDHB); Scheme 1, top), 1,4-dibromo-2,5-dimethoxyben-

Scheme 1. Synthesis of P(VDHB) (Top) and P(VHar) (Bottom)^a

^a(a) *n*-BuLi, DMF, Et₂O, 1 h, 0°C to rt; (b) KMnO₄, acetone/H₂O, 24 h, rt; (c) H₂SO₄, methanol, 24 h, 80°C ; (d) Pd(PPh₃)₄, Na₂CO₃, dioxane/H₂O, 72 h, 100°C ; (e) KOH, acetone/H₂O, 24 h, 100°C ; (f) AIBN, dioxane, 7 h, 90°C ; (g) HBr, AcOH, 5 h, 120°C ; (h) NBS, DCM, 24 h, rt; (i) Pd(PPh₃)₄, Na₂CO₃, dioxane/H₂O, 24 h, 100°C ; and (j) AIBN, dioxane, 7 h, 90°C .

zene (1) was formylated with *n*-butyllithium and dimethylformamide in diethyl ether to yield 4-bromo-2,5-dimethoxybenzaldehyde (2). In the ¹H NMR spectrum, a new signal at 10.27 ppm is assigned to the aldehyde proton, and the carbonyl carbon gives rise to a new signal at 188.15 ppm in the ¹³C NMR (Figures S11 and S12). The aldehyde was oxidized with potassium permanganate in acetone and water to give 4-bromo-2,5-dimethoxybenzoic acid (3), leading to the disappearance of the aldehyde proton and the appearance of a new signal at 12.82 ppm ascribed to the acid proton (Figure S13). In the ¹³C NMR spectrum, the carbonyl carbon C-7 is shifted to 166.43 ppm due to the lower core shielding caused by the addition of oxygen and increased oxidation state (Figure S14). The acidic group was protected by esterification with

sulfuric acid and methanol, giving 4-bromo-2,5-dimethoxymethylbenzoate (**4**). The success of this reaction was verified by an additional singlet at 3.79 ppm, belonging to the methyl ester-group, and the vanishing of the acid proton signal in the ^1H NMR spectrum (Figure S15). In the ^{13}C NMR spectrum, an additional signal at 52.06 ppm is observed, consistent with the ester formation (Figure S16). Then, **4** was vinylated via Suzuki cross coupling reaction with vinylboronic acid pinacol ester in dioxane to give 4-vinyl-2,5-dimethoxymethylbenzoate (**5**). In the ^1H NMR spectrum, three additional doublet-of-doublet signals were observed at 5.41, 6.00, and 6.95 ppm, ascribed to the protons of the vinyl group (Figures S17 and S18). The ^{13}C NMR spectrum exhibits an additional large signal at 130.36 ppm, assigned to the two carbons of the vinyl group (Figure S19). Then, the acidic group of **5** was deprotected by hydrolysis with potassium hydroxide in acetone and water to give the monomer 4-vinyl-2,5-dimethoxybenzoic acid (**6**). Its structure was verified by the vanishing of the singlet at 3.79 ppm in the ^1H NMR spectrum as well as the signal at 51.85 ppm in the ^{13}C NMR spectrum (Figures S20–S22). Afterward, **6** was polymerized to poly(vinyl dimethoxybenzoic acid) (**P(VDMB)**) via free-radical polymerization using azo-bis(isobutyronitrile) (AIBN) as an initiator in dioxane. The ^1H NMR spectrum of **P(VDMB)** exhibited broad and partly overlapping signals (Figure S23). FTIR spectra of both the monomer **6** and the polymer **P(VDMB)** were recorded and compared (Figures S38, S39, and S43). The spectrum of **6** exhibits small signals in the range from 1800 to 2100 cm^{-1} , related to overtones of the phenyl ring. These signals vanished completely in the **P(VDMB)** spectrum due to signal broadening. Additionally, the signals of the C–C double bond stretching vibration at 1610 and 1628 cm^{-1} in the monomer spectrum vanished and instead only one signal at 1612 cm^{-1} was observed in the polymer spectrum. The signals belonging to the bending vibrations of the vinyl group (914, 968, and 991 cm^{-1}) were no longer present in the polymer spectrum, indicating a successful polymerization. MALDI MS measurements of **P(VDMB)** were performed with DCTB as a secondary matrix. The averaged molar masses determined by MALDI MS are $\bar{M}_n = 2826$ g/mol and $\bar{M}_w = 3045$ g/mol, resulting in an averaged $\bar{D}$ of 1.08 and an averaged DP of 15 (Figure S54 and Table S11). The $\bar{D}$ is close to 1 and thus represents a narrow size distribution of the polymer chains. This is unexpected for free-radical polymerization, which usually results in broader chain length distributions ($\bar{D} \geq 2$) due to side reactions, such as recombination and disproportion.⁴⁸ An explanation for the narrow $\bar{D}$ could be that not all **P(VDMB)** chains are detected with MALDI MS, maybe due to too large chain lengths. The repeating unit was confirmed ($\bar{M}_{\text{RU,exp.}} = 208.42 \pm 0.22$ g/mol and $\bar{M}_{\text{RU,theo.}} = 208.21$ g/mol). GPC measurements of **P(VDMB)** were not successful (Figures S46 and S47). A possible explanation is the adsorption of the polymer on the stationary phase because of the acidic group. Finally, to obtain a polymer with free hydroxy groups, the methoxy groups of **P(VDMB)** were cleaved quantitatively postpolymerization using acetic acid and hydrobromic acid. The resulting poly(vinyl dihydroxybenzoic acid) (**P(VDHB)**) again shows broad and partly overlapping signals in the ^1H NMR spectrum (Figure S24). The integrals of the hydroxy groups cannot be used to calculate the conversion ratio because of their exchanging character. Hence, to provide an accurate conversion ratio, additionally, an FTIR spectrum of **P(VDHB)** was recorded and compared with that of

P(VDMB) (Figures S39, S40, and S44). The successful ether cleavage was verified by the complete vanishing of the signal at 1030 cm^{-1} in the **P(VDHB)** spectrum, assigned to the bending vibrations of the ether. At the same time, a signal from around 1750–3700 cm^{-1} in the **P(VDHB)** spectrum, which is broader and much more intense than the signal from 2300–3700 cm^{-1} in the **P(VDMB)** spectrum, is ascribed to the hydroxy groups and indicates a high conversion ratio. Despite trying different matrices and mixing ratios, we were not able to detect the polymer **P(VDHB)** with MALDI MS. Nevertheless, the polymeric properties of **P(VDHB)** ($\bar{M}_n$, $\bar{M}_w$, $\bar{D}$, and $\bar{DP}$) are assumed to be equal to those of **P(VDMB)** as the postpolymerization modification does not influence them. More detailed information including the tested measurement parameters are given in Table S10. Like **P(VDMB)**, **P(VDHB)** was not detected in GPC experiments, probably due to the presence of the acidic group (Figures S48 and S49). UV–vis measurements revealed a blue-shift of the absorption maximum of **P(VDMB)** compared to **P(VDHB)** (Figures S32 and S33). This can be explained by the change in the auxochrome (methoxy to hydroxy), which influences the electron density within the phenyl ring.

P(VHar) was obtained in a facile three-step reaction starting from commercially available harmine (Scheme 1, bottom). First, harmine was brominated with *N*-bromosuccinimide in dichloromethane to obtain 8-bromo-harmine (**7**). The chemical shifts and intensities of the signals observed in the ^1H and ^{13}C NMR spectra (Figures S25–S27) are consistent with literature reports.⁴⁹ Next, **7** was reacted under Suzuki conditions with vinylboronic acid pinacol ester in dioxane to give 8-vinyl-harmine (**8**). In the ^1H NMR spectrum, three additional doublet-of-doublet signals were detected at 5.58, 6.20, and 7.27 ppm, originating from the three protons of the vinyl group (Figures S28 and S29). In the ^{13}C NMR spectrum, two additional signals from the vinyl group are observed at 115.93 and 128.60 ppm, and the signal of C-8, where the coupling was taking place, shifted to 108.89 ppm from 92.36 ppm for **7** (Figure S30).

Finally, the monomer was polymerized to the targeted polymer **P(VHar)** via free-radical polymerization in dioxane with AIBN as an initiator. Consistent with a polymeric structure, the ^1H NMR spectrum of **P(VHar)** reveals broad and overlapping signals (Figure S31). FTIR spectra of the monomer **8** and the polymer **P(VHar)** were collected and compared (Figures S41, S42, and S45). Similar to **P(VDMB)**, the small signals in the spectrum of **8** in the range from 1670 to 1890 cm^{-1} , related to overtones of the phenyl ring, vanished in the polymer spectrum due to signal broadening. Similarly, the signals at 827, 901, 928, and 989 cm^{-1} in the monomer spectrum, assigned to bending vibrations of the vinyl group, vanished completely after polymerization. The signals at 1610 and 1626 cm^{-1} (stretching vibrations of C–C double bonds) are no longer detected. Instead, one signal at 1618 cm^{-1} is observed in the polymer spectrum. MALDI MS measurements with the secondary matrix DCTB resulted in averaged molar masses of $\bar{M}_n = 3031$ g/mol and $\bar{M}_w = 3610$ g/mol, leading to an averaged $\bar{D}$ of 1.19 and a $\bar{DP}^*$ of 15. This $\bar{D}$ is larger than the $\bar{D}$ of **P(VDMB)** but still represents a narrower distribution than expected for free-radical polymerization.⁴⁸ Also, similar to **P(VDMB)**, the small $\bar{DP}$ is unusual for free-radical polymerization but not problematic for the targeted application as MALDI matrix for LMWC analytics as long as all low mass chains (which potentially could overlap signals of analytical

targets) are removed. The molar mass of the repeating unit was confirmed ($\bar{M}_{\text{RU,exp.}} = 238.26 \pm 0.16$ g/mol and $\bar{M}_{\text{RU,theo.}} = 238.29$ g/mol) (Figure S55 and Table S12). GPC measurements in DMAc gave higher values (Figures S50 and S52; Table S8): $\bar{M}_n = 12\,100$ g/mol, $\bar{M}_w = 13\,000$ g/mol, $\bar{D} = 1.07$, and $\bar{DP} = 55$, representing rather the formation of polymers than oligomers, as suggested from the MALDI MS results. A possible explanation for the difference is that not all polymer chains detected with GPC are also detected with MALDI MS. An extrapolation of the molar mass curve showed that the low-intensity peak at 8.59 min elution time likely corresponds to the chain masses observed in the MALDI mass spectrum (Figures S51 and S53; Table S9), indicating that only short polymer chains are detected with MALDI MS. Hence, the $\bar{DP}$'s calculated from the MALDI MS spectra are most probably not accurate, showing that the synthesized compounds are rather polymers than oligomers. To detect the longer polymer chains, a linear mode measurement was carried out (Figure S56) using identical parameters. Here, three signal series are detected in the same mass range as the reflectron mode spectrum but with less resolved signals. Again, longer chains of P(VHar) could not be detected.

Optical Properties. To be suitable as a MALDI matrix, a compound must absorb efficiently at the wavelength of the equipped laser (here, $\lambda_{\text{laser}} = 355$ nm). To compare the optical properties of the P(SOMs) with those of the related SOMs, the relative absorbance of all compounds was recorded as drop-casted thin films (Figure S32), as this is similar to the MALDI sample preparation conditions. Next, the mass attenuation coefficient (MAC)⁵⁰ was calculated from UV–vis spectra collected from methanol solution to directly compare the absorption efficiencies of SOMs and P(SOMs) at the laser wavelength (Figure S33). Detailed information about sample preparation, measurement conditions, calculations, and all results are given in Tables S5–S7. As thin films, DHB and P(VDHB) reveal an identical relative absorption of $A_{\text{rel},355} = 62\%$ at the laser wavelength, and the precursor-polymer P(VDMB) exhibits similar $A_{\text{rel},355} = 61\%$, indicating equal absorption efficiencies. Thin film samples of harmine show a higher relative absorbance ($A_{\text{rel},355} = 51\%$) than P(VHar) ($A_{\text{rel},355} = 42\%$), suggesting that harmine absorbs more efficiently than P(VHar) in a solid state.

In solution, DHB and P(VDHB) exhibit narrow averaged MAC_{355} (DHB: 1993.6 m²/kg and P(VDHB): 1760.2 m²/kg). The MAC of P(VDHB) is thus slightly lower (−11.7%), but still high enough to indicate good absorption efficiency. The precursor-polymer P(VDMB), however, exhibits a much lower MAC_{355} of 132.7 m²/kg (−93.3%), probably due to the different auxochrome as described above. The MAC_{355} is 265.6 m²/kg for harmine and 729.3 m²/kg (+274.6%) for P(VHar), indicating a higher absorption efficiency of P(VHar).

Vacuum Stability. MALDI MS measurements in standard instruments are generally performed under HV ($\sim 10^{-7}$ mbar). During longer MALDI MSI measurements, SOMs can sublime, which can change the amount of matrix on individual measurement spots. Polymers are generally vacuum stable due to their high molar mass. The vacuum stability of the SOMs and P(SOMs) was measured by keeping them as thin films under HV and determining the weight loss after defined intervals (Figure 1). For both DHB and harmine, the largest mass loss occurred within the first 30 min (DHB: 5.3% and harmine: 6.7%). Afterward, the mass loss rate slowed

Figure 1. Vacuum stability under HV ($\sim 10^{-7}$ mbar) over 24 h (average of three measurements, straight plots) and 72 h (single measurement, dashed plots) of DHB (yellow), P(VDHB) (red), harmine (blue), and P(VHar) (black). The balance error of 0.01 mg is visualized relative to the sample weight as error bars.

down, resulting in an exponential decrease during the first 6 h. For DHB, this is consistent with the literature.⁸ Measurements from 0 to 24 h showed that the exponential decrease continued for both SOMs, with DHB and harmine experiencing 16% and 14% total loss, respectively. A measurement after 72 h showed a total loss of 29% for DHB and 28% for harmine (Figures S34 and S36, respectively). As expected, these results confirm that both SOMs sublime, a potential disadvantage for long-term MALDI MSI measurements under HV. Also, the weight loss curves suggest that when using DHB or harmine, it might be advantageous to start MALDI MSI experiments after 2 h in HV to utilize the lower sublimation rate to achieve a good spot-to-spot reproducibility.⁵¹ In line with expectations, neither P(VDHB) nor P(VHar) showed significant mass changes over time under HV conditions: even after 24 h and 72 h, the masses were fully retained (Figures S35 and S37), indicating that the P(SOMs) are completely vacuum stable. At the same time, the new polymer matrices might also be used for experiments using MALDI ion sources with different pressures, such as AP MALDI.

P(SOMs) as MALDI Matrices. Despite potential drawbacks regarding vacuum stability and interferences in the low-mass area, SOMs are excellent matrices due to their broad analyte scope and high ionization efficiencies. Consequently, new matrix systems need to be benchmarked against this. Polymerization should ideally retain performances regarding scopes and efficiencies and alleviate the background interferences and vacuum instability. To verify this hypothesis, the matrix performances regarding analyte scopes, dual polarity mode suitability, background interferences, and ionization efficiencies of SOMs and P(SOMs) were compared using identical MALDI MS measurement conditions. Five commercially available lipids were used as DHB analytes (myristic acid (MA), linoleic acid, docosahexaenoic acid, dipalmitoyl-glycerol, and dipalmitoyl-glycero-phosphocholine) and five saccharides as harmine analytes (fructose (FT), glucuronic acid, sucrose, maltotriose, and α -cyclodextrin (α -CD)). Additionally, essential amino acids (methionine and phenylalanine) were used as analytes, as this molecule class is biologically relevant for MALDI MS investigations. All analytes

Figure 2. Reflectron positive mode (30% laser intensity) MALDI mass spectra of the sample plate (violet), the pure analyte MA (yellow), the pure SOM DHB (green), DHB coated on MA (red), the pure P(SOM) P(VDHB) (blue), and P(VDHB) coated on MA (black). Peaks: m/z = 23 $[\text{Na}]^+$; 39 $[\text{K}]^+$; 57 $[\text{MA1} = \text{C}_4\text{H}_9]^+$; 88 $[\text{MA2} = \text{C}_4\text{H}_8\text{O}_2]^+$; 137 $[\text{DHB} - \text{OH}]^+$; 154 $[\text{DHB}]^+$; 177 $[\text{DHB} + \text{Na}]^+$; 193 $[\text{DHB} + \text{K}]^+$; 215 $[\text{unknown DHB signal}]^+$; 251 $[\text{MA} + \text{Na}]^+$; and 267 $[\text{MA} + \text{K}]^+$.

Figure 3. Reflectron negative mode (30% laser intensity) MALDI mass spectra of the sample plate (violet), the pure analyte FT (yellow), the pure SOM harmine (green), harmine coated on FT (red), the pure P(SOM) P(VHar) (blue), and P(VHar) coated on FT (black). Peaks: m/z = 35 $[\text{Cl}]^-$; 71 $[\text{FT1} = \text{C}_3\text{H}_5\text{O}_2]^-$; 101 $[\text{FT2} = \text{C}_4\text{H}_5\text{O}_3]^-$; 119 $[\text{FT3} = \text{C}_4\text{H}_7\text{O}_4]^-$; 130 $[\text{FT4} = \text{C}_6\text{H}_{10}\text{O}_3]^-$; 179 $[\text{FT} - \text{H}]^-$; and 211 $[\text{Har} - \text{H}]^-$.

were used for single analyte measurements with DHB and P(VDHB) as well as harmine and P(VHar) in both positive and negative reflectron modes. Furthermore, three choline derivatives (CC, PC, and GIPC), which are biologically relevant metabolites for cancer research, were used for LoD measurements, again in both polarity modes.⁵² All measurements were repeated two times for statistical validation. The

molecular masses and structures as well as potential fragmentations of all analytes are given in Tables S1 and S2.

MALDI MS Measurements of Single Analytes—P(VDHB). All analytes detected with DHB were also found using P(VDHB), indicating a comparable analyte scope (Figures S57–S68). Despite the slightly lower MAC, P(VDHB) required an identical or lower laser intensity for the detection of all tested analytes than DHB, which indicates a

slightly increased ionization efficiency of P(VDHB) (Figure 2). α -CD in negative mode is not observed with either DHB or P(VDHB), despite testing different sample preparation and measurement conditions (detailed information is given in the Supporting Information). In all DHB spectra, matrix-related signals originating from molecule ions, fragments, and clusters were detected. The P(VDHB) spectra, however, reveal no matrix-related signals (Figure 2). P(VDHB) is thus a MALDI-silent dual polarity mode matrix under the used measurement conditions.

MALDI MS Measurements of Single Analytes—P(VHar).

Apart from α -CD, which was only detected with harmine at high laser intensity (90%) in negative mode but not detected with P(VHar) despite testing different conditions (detailed information are given in Figure S74), P(VHar) shows a similar analyte scope (Figures 3 and S69–S81). Again, the P(SOM) required an identical or lower laser intensity to detect the analytes, indicating slightly enhanced ionization efficiency. Like harmine, P(VHar) is suitable for measurements in both positive and negative modes. In both polarity modes, harmine produces matrix-related signals, i.e., molecule ions and clusters. P(VHar), however, produces no such signals, indicating that P(VHar) is a MALDI-silent dual polarity mode matrix.

LoD Measurements. To compare the ionization efficiencies of SOMs (including 9AA) and P(SOMs), the LoD measurements were carried out using different choline derivatives, which are biologically relevant metabolites in cancer research (Table 1). For the evaluation, the lowest

positive mode LoD measurements of PC using DHB and P(VDHB), the signal at $m/z = 122$ was still detected at a concentrations of 540 pmol, whereas in negative mode, the signal at $m/z = 161$ was observed down to a concentration of 5.4 nmol (Figures S88–S91). Similarly, when measuring GIPC in the positive mode, the signal at $m/z = 139$ was still observed at a concentration of 390 pmol and in the negative mode the signal at $m/z = 242$ was detected down to a concentration of 3.9 nmol, using either DHB or P(VDHB) (Figures S98–S101). Consistent with the CC measurements, these results indicate comparable ionization efficiencies for DHB and P(VDHB). A similar trend was observed for harmine and P(VHar). While the detection in positive mode is more sensitive, SOM and P(SOM) detect PC and GIPC down to the same concentrations (Figures S92–S95 and S102–S105). When using 9AA negative mode LoD measurements, PC was detected down to 540 fmol and GIPC down to 39 pmol (Figures S87 and S97). Hence, this indicates that the ionization efficiency of this standard matrix for negative mode measurements is enhanced compared to those of the SOMs and corresponding P(SOMs).

MALDI MSI. After verifying that the performance of the P(SOMs) is comparable to the corresponding SOMs regarding analyte scope, ionization efficiencies, and dual polarity mode suitability, their ability to detect multiple compounds directly on biologically relevant tissue sections was tested: P(VDHB) and P(VHar) were applied as matrices to analyze cryosections of tumor tissue from a breast cancer xenograft model by MALDI MSI. Figure 4 (top) shows the heatmaps of MALDI generated cations using P(VDHB) and the corresponding optical picture. The spatial localization of positive ions ($m/z = 114.0, 147.0, 666.5, 762.4, 794.4, 798.5, 874.5$, and 920.4) is tissue-specific, as the ions are localized mainly in the tissue.

Table 1. Calculated Minimal Molar Amounts of CC, PC, and GIPC Detected with DHB/P(VDHB), Harmine/P(VHar), and 9AA

	DHB/P(VDHB)		harmine/P(VHar)		9AA
	positive	negative	positive	negative	
CC	960 amol	n/a	9.6 pmol	n/a	n/a
PC	540 pmol	5.4 nmol	540 pmol	5.4 nmol	540 fmol
GIPC	390 pmol	3.9 nmol	390 pmol	3.9 nmol	39 pmol

detectable molar amounts of analytes were calculated by dividing the product of the solution concentration and drop volume, used for sample preparation, by the molar mass of the analyte (see eq S2). Because of the ionic character, CC was detected only in positive mode as the cationic molecule ion of choline ($m/z = 104$) and a fragment ($m/z = 86$).⁵³ These ions were also observed without matrix, but with DHB and P(VDHB), they were more intense and still found at concentrations of 960 amol (Figures S82 and S83). With harmine and P(VHar), CC was observed down to concentrations of 9.6 pmol (Figures S84 and S85). In these experiments, both P(SOMs) have the same detection limit for CC as those of their corresponding SOMs. MALDI MS measurements of pure PC and GIPC were first carried out without matrix in both polarity modes to determine the most promising analyte signals for the LoD measurements (Figures S86 and S96, respectively). High laser intensities (90%) were needed to obtain suitable signals, which were afterward filtered according to their intensities and their ability to uniquely represent the individual analyte. For PC, the signals at $m/z = 122$ (positive mode) and $m/z = 161$ (negative mode) were chosen and, for GIPC, the signals at $m/z = 139$ (positive mode) and $m/z = 242$ (negative mode) were chosen. All LoD measurements were performed using 40% laser intensity. In

Figure 4. Positive mode (top) and negative mode (bottom) MALDI MSI measurements of a cryosection from a patient-derived breast cancer xenograft tumor using P(VDHB) as a matrix. Positive mode: top (from left to right)—optical picture and ionic intensity heatmaps of ions at m/z values 114.0, 147.0, 666.5, and 762.4 and bottom (from left to right)—ionic intensity heatmaps of the ions at m/z values 794.4, 798.5, 874.4, and 920.4. Negative mode: top (from left to right)—optical picture and ionic intensity heatmaps of ions at m/z values 103.0, 255.2, 279.2, and 281.2 and bottom (from left to right)—ionic intensity heatmaps of the ions at m/z values 283.2, 702.4, 708.4, and 764.4.

Although the tissue section predominantly consists of viable neoplastic tissue, the ions display different spatial concentrations. A H&E-stained tissue section (Figure S108) indicates that the imaged specimen displays significant structural heterogeneity and contains both cancer cells and interstitial stromal cells. Furthermore, it has previously been demonstrated that the MAS98.06 patient-derived tumors display a marked physiological heterogeneity in tissue perfusion and oxygenation levels, which in turn regulate the metabolic features of the tumor tissue.^{54–56} Spatial variability in metabolite concentrations is, therefore, an expected finding and confirms that MALDI MSI is a suitable method for studies of this phenomenon.

Similar to single analyte measurements, sufficient heatmaps using P(VDHB) in negative mode are measured and are depicted in Figure 4 (bottom), showing the spatial distribution of negative ions localized on the tissue ($m/z = 103.0, 255.2, 279.2, 281.2, 283.2, 702.4, 708.4, \text{ and } 764.4$). Like in positive mode, the negative mode heatmaps display spatial variability in ion concentration. The experiments verify the dual polarity mode suitability for P(VDHB) also as MALDI MSI matrix. More detailed pictures of the heatmaps are depicted in Figures S115–S130. Similarly, P(VHar)-assisted laser ionization is also able to produce positive (Figure 5 (top), $m/z = 86.0, 104.1,$

(Tables S13–S16).^{57–64} The database search indicated that choline ($m/z = 104.1$), phosphocholine ($m/z = 184.1$), and glycerophosphocholine ($m/z = 258.1$) could be detected in positive mode and that known cancer biomarkers such as lactate ($m/z = 89.1$) and glutamine ($m/z = 145.1$) could be detected in negative mode.^{63,65}

These observations suggest that an ionization process assisted by P(VDHB) and P(VHar) occurs specifically on the tissue, showing that both P(SOMs) are well suitable as dual polarity mode matrices for MALDI MSI experiments. The sample preparation via spray coating resulted in a smooth and homogeneous matrix coverage of the tissue section (Figure S107). Contrary to SOMs, the spraying process of the P(SOMs) required only very low concentrated matrix solutions (0.5 mg/mL), resulting in lower material consumption and less contamination of the sprayer machine. Additionally, in contrast to previously reported conjugated polymer matrices,⁴³ the P(SOMs) are well soluble in polar/protic solvents such as methanol, supporting analyte extraction from the tissue and leading to better signal intensities. Similarly, for the corresponding SOMs, the solution composition influences the analyte extraction ability of the P(SOMs):⁶⁶ the production of ions in the range from $m/z = 700$ to $m/z = 960$ increases when the proportion of water is increased (Figure S106). Furthermore, the spatial localization of the exemplary ion at $m/z = 846.4$ was improved by adding acetone in the matrix solution. Although SOMs often require additives, e.g., ammonium sulfate, to improve the detection limits for MALDI MSI experiments, the used P(SOM) solutions, while being prepared without any additives, still led to sufficient signal intensities.^{67,68} A further comparison between P(VDHB) and DHB shows that the ions at $m/z = 72.1, 268.9, 402.7, 469.1, \text{ and } 486.6$ can only be visualized using P(VDHB) (Figures S109 and S111–S114). Three ions ($m/z = 268.9, 469.1, \text{ and } 486.6$) measured with DHB were completely delocalized on the tissue; their ion maps showed either high ion concentration outside the tissue or a too homogeneous ion distribution on the tissue. The other two ions ($m/z = 72.1$ and 402.7) were not detected using DHB as a matrix, giving black ions maps. Here, the MALDI-silent properties of the P(SOMs), indicated by the single analyte measurements, are also observed: a P(VDHB) heatmap of $m/z = 155.0$ (protonated molecule ion of DHB) is completely black and hence shows no signal distribution, while a heatmap derived from using DHB is colorful (Figure S110). The use of MALDI-silent P(SOMs) gives higher control over the experiment by allowing the free choice of sample-specific internal standards/calibrants and prevents the detector from damage caused by strong matrix molecule bombardment when using higher laser intensities, which is often required to enhance the ionic intensities during MALDI MSI experiments.⁶⁹ Although spectrometers working in subAP- to atmospheric pressure are now available, HV-MALDI TOF machines are still the most commonly used instruments for (pre-)clinical studies and require vacuum stable matrices like the presented P(SOMs).

CONCLUSIONS

Retaining the advantages of SOMs, such as broad analyte scope and high ionization efficiencies, while inducing MALDI silence and vacuum stability would allow us to reliably and reproducibly measure LMWCs with MALDI MS and MSI. Here, we apply polymerization as a new strategy to achieve

Figure 5. Positive mode (top) and negative mode (bottom) MALDI MSI measurements of a cryosection from a patient-derived breast cancer xenograft tumor using P(VHar) as a matrix. Positive mode: top (from left to right)—optical picture and ionic intensity maps of ions at m/z values 86.0, 104.1, 118.1, and 184.1 and bottom (from left to right)—ionic intensity heatmaps of the ions at m/z values 258.1, 504.3, 798.4, and 874.3. Negative mode: top (from left to right)—optical picture and ionic intensity maps of ions at m/z values 87.0, 89.1, 124.0, and 145.1 and bottom (from left to right)—ionic intensity heatmaps of the ions at m/z values 283.2, 303.2, 835.4, and 887.5.

118.1, 184.1, 258.1, 504.3, 798.4, and 874.3) as well as negative ions (Figure 5 (bottom), $m/z = 87.0, 89.1, 124.0, 145.1, 283.2, 303.2, 835.4, \text{ and } 887.5$) directly on the tissue section. P(VHar) thus enables dual polarity mode measurements also in MALDI MSI experiments. Enlarged pictures of the heatmaps are shown in Figures S131–S146. As MALDI TOF MSI measurements cannot be used for formal identification of the detected ions, the HMDB database was used to assign known biological entities to the m/z ratios

this. Two high-performance SOMs (DHB and harmine) were modified to contain a vinyl group and polymerized to the corresponding polyethylene-based P(SOMs), carrying SOMs as side chains. The P(SOMs) showed very good absorption behavior at the laser wavelength ($\lambda = 355$ nm) and high stability in HV ($\sim 10^{-7}$ mbar). Single analyte measurements of several commercially available lipids, saccharides, and amino acids indicated that both P(SOMs) have a similar analyte scope and slightly higher ionization efficiencies compared to their corresponding SOMs while being MALDI silent. Especially the ability for MALDI-silent LMWC detection not only increases the experiment control and reproducibility but also reduces instrumentation damage. Also, the P(SOMs) were successfully applied as matrices for MALDI MSI experiments of breast cancer xenograft tumors and allowed measurements of known cancer biomarkers in dual polarity mode without the utilization of additives. The obtained results are a promising foundation for future research efforts, e.g., for investigating the intrapatient and interpatient reproducibility to enable new clinical MSI routines. In summary, the polymerization of SOMs is a promising new strategy to obtain high-performance MALDI matrix systems, which maintain broad analyte scopes and high ionization efficiencies of the state-of-the-art SOMs while showing enhanced properties such as HV stability and MALDI silence.

■ ASSOCIATED CONTENT

Supporting Information

The Supporting Information is available free of charge at <https://pubs.acs.org/doi/10.1021/acsapm.1c00665>.

Details about the used chemicals, the employed instrumentation, and experimental settings and resulting data (PDF)

■ AUTHOR INFORMATION

Corresponding Author

Franziska Lissel – Institute of Macromolecular Chemistry, Leibniz-Institut für Polymerforschung Dresden e.V., Dresden, Saxony 01069, Germany; Faculty of Chemistry and Food Chemistry, Technische Universität Dresden, Dresden, Saxony 01062, Germany; Institute of Organic Chemistry and Macromolecular Chemistry, Friedrich Schiller University Jena, Jena, Thuringia 07743, Germany; orcid.org/0000-0003-0254-4565; Email: lissel@ipfdd.de

Authors

Kilian Horatz – Institute of Macromolecular Chemistry, Leibniz-Institut für Polymerforschung Dresden e.V., Dresden, Saxony 01069, Germany; Faculty of Chemistry and Food Chemistry, Technische Universität Dresden, Dresden, Saxony 01062, Germany

Marco Giampà – MR Cancer Group, Department of Clinical and Molecular Medicine, Norwegian University of Science and Technology, Trondheim 7030, Norway; orcid.org/0000-0002-8847-9604

Zhi Qiao – Institute of Macromolecular Chemistry, Leibniz-Institut für Polymerforschung Dresden e.V., Dresden, Saxony 01069, Germany; Faculty of Chemistry and Food Chemistry, Technische Universität Dresden, Dresden, Saxony 01062, Germany

Siver A. Moestue – MR Cancer Group, Department of Clinical and Molecular Medicine, Norwegian University of

Science and Technology, Trondheim 7030, Norway;
Department of Pharmacy, Nord University, Bodø 8026, Norway

Complete contact information is available at:
<https://pubs.acs.org/doi/10.1021/acsapm.1c00665>

Author Contributions

The manuscript was written through contributions of all authors. All authors have given approval to the final version of the manuscript.

Notes

The authors declare no competing financial interest.

■ ACKNOWLEDGMENTS

F.L. thanks the *Fonds der Chemischen Industrie* (FCI) for a Liebig Fellowship. Z.Q. thanks the China Scholarship Council for a scholarship (no. 201908080098). The authors thank C. Harnisch and P. Treppe for providing the GPC measurements and Dr. M. Malanin for FTIR spectra. MALDI MSI experiments were supported by the Norwegian Cancer Society (Grant no. 198069). The authors thank the Cellular and Molecular Imaging Core Facility (CMIC) for the tissue sectioning.

■ REFERENCES

- (1) Norris, J. L.; Caprioli, R. M. Analysis of Tissue Specimens by Matrix-Assisted Laser Desorption/Ionization Imaging Mass Spectrometry in Biological and Clinical Research. *Chem. Rev.* **2013**, *113*, 2309–2342.
- (2) Vaysse, P.-M.; Heeren, R. M. A.; Porta, T.; Balluff, B. Mass spectrometry imaging for clinical research—Latest developments, applications, and current limitations. *Analyst* **2017**, *142*, 2690–2712.
- (3) Llombart, V.; Trejo, S. A.; Bronsoms, S.; Morancho, A.; Ma, F.; Faura, J.; Garcia-Berrosos, T.; Simats, A.; Rosell, A.; Canals, F.; Hernandez-Guillamon, M.; Montaner, J. Profiling and identification of new proteins involved in brain ischemia using MALDI-imaging-mass-spectrometry. *J. Proteomics* **2017**, *152*, 243–253.
- (4) Römpf, A.; Spengler, B. Mass spectrometry imaging with high resolution in mass and space. *Histochem. Cell Biol.* **2013**, *139*, 759–783.
- (5) Wang, J. S. H.; Freitas-Andrade, M.; Bechberger, J. F.; Naus, C. C.; Yeung, K. K. C.; Whitehead, N. Matrix-assisted laser desorption/ionization imaging mass spectrometry of intraperitoneally injected danegaptide (ZP1609) for treatment of stroke-reperfusion injury in mice. *Rapid Commun. Mass Spectrom.* **2018**, *32*, 951–958.
- (6) Liu, H.; Chen, R.; Wang, J.; Chen, S.; Xiong, C.; Wang, J.; Hou, J.; He, Q.; Zhang, N.; Nie, Z.; Mao, L. 1,5-Diaminonaphthalene Hydrochloride Assisted Laser Desorption/Ionization Mass Spectrometry Imaging of Small Molecules in Tissues Following Focal Cerebral Ischemia. *Anal. Chem.* **2014**, *86*, 10114–10121.
- (7) Weigt, D.; Sammour, D. A.; Ulrich, T.; Munteanu, B.; Hopf, C. Automated analysis of lipid drug-response markers by combined fast and high-resolution whole cell MALDI mass spectrometry biotyping. *Sci. Rep.* **2018**, *8*, No. 11260.
- (8) Giampa, M.; Lissel, M. B.; Patschkowski, T.; Fuchser, J.; Hans, V. H.; Gembruch, O.; Bednarza, H.; Niehaus, K. Maleic anhydride proton sponge as a novel MALDI matrix for the visualization of small molecules (<250 m/z) in brain tumors by routine MALDI TOF imaging mass spectrometry. *Chem. Commun.* **2016**, *52*, 9801–9804.
- (9) Belov, M. E.; Ellis, S. R.; Dilillo, M.; Paine, M. R. L.; Danielson, W. F.; Anderson, G. A.; Graaf, E. L. D.; Eijkel, G. B.; Heeren, R. M. A.; McDonnell, L. A. Design and Performance of a Novel Interface for Combined Matrix-Assisted Laser Desorption Ionization at Elevated Pressure and Electrospray Ionization with Orbitrap Mass Spectrometry. *Anal. Chem.* **2017**, *89*, 7493–7501.

- (10) Soltwisch, J.; Heijs, B.; Koch, A.; Vens-Cappell, S.; Höndorf, J.; Dreisewerd, K. MALDI-2 on a Trapped Ion Mobility Quadrupole Time-of-Flight Instrument for Rapid Mass Spectrometry Imaging and Ion Mobility Separation of Complex Lipid Profiles. *Anal. Chem.* **2020**, *92*, 8697–8703.
- (11) Moskovets, E.; Misharin, A.; Laiko, V.; Doroshenko, V. A comparative study on the analytical utility of atmospheric and low-pressure MALDI sources for the mass spectrometric characterization of peptides. *Methods* **2016**, *104*, 21–32.
- (12) Ferey, J.; Marguet, F.; Laquerrière, A.; Marret, S.; Schmitz-Afonso, I.; Bekri, S.; Afonso, C.; Tebani, A. A new optimization strategy for MALDI FTICR MS tissue analysis for untargeted metabolomics using experimental design and data modeling. *Anal. Bioanal. Chem.* **2019**, *411*, 3891–3903.
- (13) Scigelova, M.; Hornshaw, M.; Giannakopoulos, A.; Makarov, A. Fourier Transform Mass Spectrometry. *Mol. Cell. Proteomics* **2011**, *10*, No. M111.009431.
- (14) Xie, Y. R.; Castro, D. C.; Lam, F.; Sweedler, J. V. Accelerating Fourier Transform-Ion Cyclotron Resonance Mass Spectrometry Imaging Using a Subspace Approach. *J. Am. Soc. Mass Spectrom.* **2020**, *31*, 2338–2347.
- (15) Hale, O. J.; Morris, M.; Jones, B.; Reynolds, C. K.; Cramer, R. Liquid Atmospheric Pressure Matrix-Assisted Laser Desorption/Ionization Mass Spectrometry Adds Enhanced Functionalities to MALDI MS Profiling for Disease Diagnostics. *ACS Omega* **2019**, *4*, 12759–12765.
- (16) Keller, C.; Maeda, J.; Jayaraman, D.; Chakraborty, S.; Sussman, M. R.; Harris, J. M.; Ané, J.-M.; Li, L. Comparison of Vacuum MALDI and AP-MALDI Platforms for the Mass Spectrometry Imaging of Metabolites Involved in Salt Stress in *Medicago truncatula*. *Front. Plant Sci.* **2018**, *9*, 1–15.
- (17) Strupat, K.; Karas, M.; Hillenkamp, F. 2,5-Dihydroxybenzoic acid: A new matrix for laser desorption-ionization mass spectrometry. *Int. J. Mass Spectrom. Ion Processes* **1991**, *111*, 89–102.
- (18) Bourcier, S.; Bouchonnet, S.; Hoppilliard, Y. Ionization of 2,5-dihydroxybenzoic acid (DHB) matrix-assisted laser desorption ionization experiments and theoretical study. *Int. J. Mass Spectrom. Ion Processes* **2001**, *210–211*, 59–69.
- (19) Choi, S.-S.; Ha, S.-H. Influence of Silver Salt Types on Formation of Silver Cluster Ions in MALDI with DHB as Matrix. *Bull. Korean Chem. Soc.* **2007**, *28*, 2508–2510.
- (20) Schiller, J.; Süß, R.; Fuchs, B.; Müller, M.; Petkovic, M.; Zschörnig, O.; Waschpky, H. The suitability of different DHB isomers as matrices for the MALDI-TOF MS analysis of phospholipids: Which isomer for what purpose? *Eur. Biophys. J.* **2007**, *36*, 517–527.
- (21) Wei, Y.; Zhang, Y.; Lin, Y.; Li, L.; Liu, J.; Wang, Z.; Xiong, S.; Zhao, Z. A uniform 2,5-dihydroxybenzoic acid layer as a matrix for MALDI-FTICR MS-based lipidomics. *Analyst* **2015**, *140*, 1298–1305.
- (22) Colsch, B.; Woods, A. S. Localization and imaging of sialylated glycosphingolipids in brain tissue sections by MALDI mass spectrometry. *Glycobiology* **2010**, *20*, 661–667.
- (23) Perry, W. J.; Patterson, N. H.; Prentice, B. M.; Neumann, E. K.; Caprioli, R. M.; Spraggins, J. M. Uncovering matrix effects on lipid analyses in MALDI imaging mass spectrometry experiments. *J. Mass Spectrom.* **2020**, *55*, No. e4491.
- (24) Dufresne, M.; Patterson, N. H.; Norris, J. L.; Caprioli, R. M. Combining Salt Doping and Matrix Sublimation for High Spatial Resolution MALDI Imaging Mass Spectrometry of Neutral Lipids. *Anal. Chem.* **2019**, *91*, 12928–12934.
- (25) He, H.; Qin, L.; Zhang, Y.; Han, M.; Li, J.; Liu, Y.; Qiu, K.; Dai, X.; Li, Y.; Zeng, M.; Guo, H.; Zhou, Y.; Wang, X. 3,4-Dimethoxycinnamic Acid as a Novel Matrix for Enhanced In Situ Detection and Imaging of Low-Molecular-Weight Compounds in Biological Tissues by MALDI-MSI. *Anal. Chem.* **2019**, *91*, 2634–2643.
- (26) Lin, Y.-S.; Chen, Y.-C. Laser Desorption/Ionization Time-of-Flight Mass Spectrometry on Sol-Gel-Derived 2,5-Dihydroxybenzoic Acid Film. *Anal. Chem.* **2002**, *74*, 5793–5798.
- (27) Tseng, M.-C.; Obena, R.; Lu, Y.-W.; Lin, P.-C.; Lin, P.-Y.; Yen, Y.-S.; Lin, J.-T.; Huang, L.-D.; Lu, K.-L.; Lai, L.-L.; Lin, C.-C.; Chen, Y.-J. Dihydrobenzoic Acid Modified Nanoparticle as a MALDI-TOF MS Matrix for Soft Ionization and Structure Determination of Small Molecules with Diverse Structures. *J. Am. Soc. Mass Spectrom.* **2010**, *21*, 1930–1939.
- (28) Son, J.; Cha, S. MALDI Mass Spectrometric Analysis of Nonderivatized Steroids Using Cyclodextrin-Supported 2,5-Dihydroxybenzoic Acid as Matrix. *Bull. Korean Chem. Soc.* **2014**, *35*, 1409–1412.
- (29) Abdelhamida, H. N.; Gopala, J.; Wu, H.-F. Synthesis and application of ionic liquid matrices (ILMs) for effective pathogenic bacteria analysis in matrix assisted laser desorption/ionization (MALDI-MS). *Anal. Chim. Acta* **2013**, *767*, 104–111.
- (30) Nonami, H.; Fukui, S.; Erra-Balsells, R. β -Carboline Alkaloids as Matrices for Matrix-Assisted Ultraviolet Laser Desorption Time-of-Flight Mass Spectrometry of Proteins and Sulfated Oligosaccharides: A Comparative Study Using Phenylcarbonyl Compounds, Carbazoles and Classical Matrices. *J. Mass Spectrom.* **1997**, *32*, 287–296.
- (31) Nonami, H.; Tanaka, K.; Fukuyama, Y.; Erra-Balsells, R. β -Carboline Alkaloids as Matrices for UV-Matrix-Assisted Laser Desorption/Ionization Time-of-Flight Mass Spectrometry in Positive and Negative Ion Modes. Analysis of Proteins of High Molecular Mass, and of Cyclic and Acyclic Oligosaccharides. *Rapid Commun. Mass Spectrom.* **1998**, *12*, 285–296.
- (32) Yamagaki, T.; Suzuki, H.; Tachibana, K. Solid-Phase Fluorescence and Ionization Efficiency in Negative-Ion Matrix-Assisted Laser Desorption/Ionization of Neutral Oligosaccharides: Interaction Between-Carboline Matrix and Ammonium Salt. *J. Am. Soc. Mass Spectrom.* **2007**, *18*, 714–723.
- (33) Ropartz, D.; Bodet, P. E.; Przybylski, C.; Gonnet, F.; Daniel, R.; Fer, M.; Helbert, W.; Bertrand, D.; Rogniaux, H. Performance evaluation on a wide set of matrix-assisted laser desorption ionization matrices for the detection of oligosaccharides in a high-throughput mass spectrometric screening of carbohydrate depolymerizing enzymes. *Rapid Commun. Mass Spectrom.* **2011**, *25*, 2059–2070.
- (34) Liu, H.; Zhou, Y.; Wang, J.; Xiong, C.; Xue, J.; Zhan, L.; Nie, Z. N-Phenyl-2-naphthylamine as a Novel MALDI Matrix for Analysis and in Situ Imaging of Small Molecules. *Anal. Chem.* **2018**, *90*, 729–736.
- (35) Qiao, Z.; Lissel, F. MALDI Matrices for the Analysis of Low Molecular Weight Compounds: Rational Design, Challenges and Perspectives. *Chem. Asian J.* **2021**, *16*, 868–878.
- (36) Ling, L.; Xiao, C.; Ma, Y.; Jiang, L.; Wang, S.; Guo, L.; Jiang, S.; Guo, X. 2-Phenyl-3-(p-aminophenyl) Acrylonitrile: A Reactive Matrix for Sensitive and Selective Analysis of Glycans by MALDI-MS. *Anal. Chem.* **2019**, *91*, 8801–8807.
- (37) Li, B.; Sun, R.; Gordon, A.; Ge, J.; Zhang, Y.; Li, P.; Yang, H. 3-Aminophthalhydrazide (Luminol) As a Matrix for Dual-Polarity MALDI MS Imaging. *Anal. Chem.* **2019**, *91*, 8221–8228.
- (38) Huang, P.; Huang, C.-Y.; Lin, T.-C.; Lin, L.-E.; Yang, E.; Lee, C.; Hsu, C.-C.; Chou, P.-T. Toward the Rational Design of Universal Dual Polarity Matrix for MALDI Mass Spectrometry. *Anal. Chem.* **2020**, *92*, 7139–7145.
- (39) Ayorinde, F. O.; Hambright, P.; Porter, T. N.; Keith, Q. L. Use of meso-tetrakis(pentafluorophenyl)porphyrin as a matrix for low molecular weight alkylphenol ethoxylates in laser desorption/ionization time-of-flight mass spectrometry. *Rapid Commun. Mass Spectrom.* **1999**, *13*, 2474–2479.
- (40) Woldegiorgis, A.; Kieseritzky, F. V.; Dahlstedt, E.; Hellberg, J.; Brinck, T.; Roeraade, J. Polymer-assisted laser desorption/ionization analysis of small molecular weight compounds. *Rapid Commun. Mass Spectrom.* **2004**, *18*, 841–852.
- (41) Woldegiorgis, A.; Löwenhielm, P.; Björk, A.; Roeraade, J. Matrix-assisted and polymer-assisted laser desorption/ionization time-of-flight mass spectrometric analysis of low molecular weight polystyrenes and polyethylene glycols. *Rapid Commun. Mass Spectrom.* **2004**, *18*, 2904–2912.

- (42) Soltzberg, L. J.; Patel, P. Small molecule matrix-assisted laser desorption/ionization time-of-flight mass spectrometry using a polymer matrix. *Rapid Commun. Mass Spectrom.* **2004**, *18*, 1455–1458.
- (43) Horatz, K.; Giampà, M.; Karpov, Y.; Sahre, K.; Bednarz, H.; Kiriy, A.; Voit, B.; Niehaus, K.; Hadjichristidis, N.; Michels, D. L.; Lissel, F. Conjugated Polymers as a New Class of Dual-Mode Matrices for MALDI Mass Spectrometry and Imaging. *J. Am. Chem. Soc.* **2018**, *140*, 11416–11423.
- (44) Horatz, K.; Ditte, K.; Prenville, T.; Zhang, K.-N.; Jehnichen, D.; Kiriy, A.; Voit, B.; Lissel, F. Amorphous Conjugated Polymers as Efficient Dual-Mode MALDI Matrices for Low-Molecular-Weight Analytes. *ChemPlusChem* **2019**, *84*, 1338–1345.
- (45) Heinzmann, G. *Relativ oder Absolut? Bewertung von Messergebnissen in der Gelpermationschromatographie*. www.analytik-news.de (accessed 2015-10-1).
- (46) Chendo, C.; Rollet, M.; Phan, T. N. T.; Viel, S.; Gigmes, D.; Charles, L. Successful MALDI mass spectrometry of poly(4-vinylpyridine) using a solvent-free sample preparation. *Int. J. Mass Spectrom.* **2015**, *376*, 90–96.
- (47) Bergamaschi, A.; Hjortland, G. O.; Triulzi, T.; Sørle, T.; Johnsen, H.; Ree, A. H.; Russnes, H. G. I.; Tronnes, S.; Mælandsmo, G. M.; Fodstad, O.; Borresen-Dale, A.-L.; Engebraaten, O. Molecular profiling and characterization of luminal-like and basal-like in vivo breast cancer xenograft models. *Mol. Oncol.* **2009**, *3*, 469–482.
- (48) Su, W.-F. Radical Chain Polymerization. In *Principles of Polymer Design and Synthesis*; Lecture Notes in Chemistry, Vol. 82; Springer: Berlin, Heidelberg, 2013; pp 137–183.
- (49) Ponce, M. A.; Erra-Balsells, R. Synthesis and Isolation of Bromo- β -carboline obtained by Bromination of β -Carboline Alkaloids. *J. Heterocyclic Chem.* **2001**, *38*, No. 1087.
- (50) McNaught, A. D.; Wilkinson, A. *IUPAC Compendium of Chemical Terminology*, 2nd ed.; Blackwell Scientific Publications: Oxford, UK, 1997.
- (51) Astigarraga, E.; Barreda-Gomez, G.; Lombardero, L.; Fresnedo, O.; Castano, F.; Giral, M. T.; Ochoa, B.; Rodriguez-Puertas, R.; Fernandez, J. A. Profiling and Imaging of Lipids on Brain and Liver Tissue by Matrix-Assisted Laser Desorption/Ionization Mass Spectrometry Using 2-Mercaptobenzothiazole as a Matrix. *Anal. Chem.* **2008**, *80*, 9105–9114.
- (52) Glunde, K.; Jie, C.; Bhujwalla, Z. M. Molecular Causes of the Aberrant Choline Phospholipid Metabolism in Breast Cancer. *Cancer Res.* **2004**, *64*, 4270–4276.
- (53) Shariatgorji, M.; Nilsson, A.; Goodwin, R. J. A.; Källback, P.; Schintu, N.; Zhang, X.; Crossman, A. R.; Bezard, E.; Svenningsson, P.; Andren, P. E. Direct Targeted Quantitative Molecular Imaging of Neurotransmitters in Brain Tissue Sections. *Neuron* **2014**, *84*, 697–707.
- (54) Huuse, E. M.; Moestue, S. A.; Lindholm, E. M.; Bathen, T. F.; Nalwoga, H.; Krüger, K.; Bofin, A.; Mælandsmo, G. M.; Akslen, L. A.; Engebraaten, O.; Gribbestad, I. S. In Vivo MRI and Histopathological Assessment of Tumor Microenvironment in Luminal-Like and Basal-Like Breast Cancer Xenografts. *J. Magn. Reson. Imaging* **2012**, *35*, 1098–1107.
- (55) Carmona-Fontaine, C.; Deforet, M.; Akkari, L.; Thompson, C. B.; Joyce, J. A.; Xavier, J. B. Metabolic origins of spatial organization in the tumor microenvironment. *Proc. Natl. Acad. Sci. U. S. A.* **2017**, *114*, 2934–2939.
- (56) Helmlinger, G.; Yuan, F.; Dellian, M.; Jain, R. K. Interstitial pH and pO₂ gradients in solid tumors in vivo: High-resolution measurements reveal a lack of correlation. *Nat. Med.* **1997**, *3*, 177–182.
- (57) Sun, C.; Wang, F.; Zhang, Y.; Yu, J.; Wang, X. Mass spectrometry imaging-based metabolomics to visualize the spatially resolved reprogramming of carnitine metabolism in breast cancer. *Theranostics* **2020**, *10*, 7070–7082.
- (58) Jiang, L.; Chughtai, K.; Purvine, S.; Bhujwalla, Z. M.; Raman, V.; Paša-Toli, L.; Heeren, R. M. A.; Glunde, K. MALDI-mass spectrometric imaging reveals hypoxia-driven lipids and proteins in a breast tumor model. *Anal. Chem.* **2015**, *87*, S947–S956.
- (59) Angerer, T. B.; Magnusson, Y.; Landberg, G.; Fletcher, J. S. Lipid Heterogeneity Resulting from Fatty Acid Processing in the Human Breast Cancer Microenvironment Identified by GCIB-ToF-SIMS Imaging. *Anal. Chem.* **2016**, *88*, 11946–11954.
- (60) Holzlechner, M.; Eugenin, E.; Prideaux, B. Mass spectrometry imaging to detect lipid biomarkers and disease signatures in cancer. *Cancer Rep.* **2019**, *2*, No. e1229.
- (61) Mattaini, K. R.; Sullivan, M. R.; Heiden, M. G. V. The importance of serine metabolism in cancer. *J. Cell. Biol.* **2016**, *214*, 249–257.
- (62) Dai, C.; Arceo, J.; Arnold, J.; Sreekumar, A.; Dovichi, N. J.; Li, J.; Littlepage, L. E. Metabolomics of oncogene-specific metabolic reprogramming during breast cancer. *Cancer Metab.* **2018**, *6*, 1–17.
- (63) Grinde, M. T.; Hilmarsdottir, B.; Tunset, H. M.; Henriksen, I. M.; Kim, J.; Haugen, M. H.; Rye, M. B.; Mælandsmo, G. M.; Moestue, S. A. Glutamine to proline conversion is associated with response to glutaminase inhibition in breast cancer. *Breast Cancer Res.* **2019**, *21*, 1–13.
- (64) O'Rourke, M. B.; Smith, C. C.; Monte, S. M. D. L.; Sutherland, G. T.; Padula, M. P. Higher Mass Accuracy MALDI-TOF/TOF Lipid Imaging of Human Brain Tissue in Alzheimer's Disease. *Curr. Protoc. Mol. Biol.* **2019**, *126*, No. e86.
- (65) Glunde, K.; Jiang, L.; Moestue, S. A.; Gribbestad, I. S. MRS and MRSI guidance in molecular medicine: Targeting and monitoring of choline and glucose metabolism in cancer. *NMR Biomed.* **2011**, *24*, 673–690.
- (66) Bashir, S.; Mutter, R.; Derrick, P. J. Parameterizing Matrix-Assisted Laser Desorption/Ionization: Effect of Solvents and Co-Additives on Analyte Peak Intensities. *Eur. J. Mass Spectrom.* **2004**, *10*, 487–493.
- (67) Sugiyama, E.; Masaki, N.; Matsushita, S.; Setou, M. Ammonium Sulfate Improves Detection of Hydrophilic Quaternary Ammonium Compounds through Decreased Ion Suppression in Matrix-Assisted Laser Desorption/Ionization Imaging Mass Spectrometry. *Anal. Chem.* **2015**, *87*, 11176–11181.
- (68) AlMasoud, N.; Correa, E.; Trivedi, D. K.; Goodacre, R. Fractional Factorial Design of MALDI-TOF-MS Sample Preparations for the Optimized Detection of Phospholipids and Acylglycerols. *Anal. Chem.* **2016**, *88*, 6301–6308.
- (69) Guenther, S.; Koestler, M.; Schulz, O.; Spengler, B. Laser spot size and laser power dependence of ion formation in high resolution MALDI imaging. *Int. J. Mass Spectrom. Ion Processes* **2010**, *294*, 7–15.

8. Appendix

8.1. LabTalk script (OriginPro 2020b)

1 //This script was written and edited by Kilian Horatz in context of the Ph.D.-thesis "Development of novel polymer matrices for MALDI MS and MALDI MS Imaging". Sharing or editing of this script without permission is strictly forbidden (E-Mail: kilianhoratz@online.de).

2

8.1.1. Section "plotting" (L. 3 - 2673)

```
3 [plotting]
4 newbook;
5 %V = %H;
6 kk=1;
7 getyesno (Do you want to plot normalized data?) yesno
  (Normalize);
8 dlgChkList inames="Small Analytes (0 - 3k Da)|Big An-
  alytes (1k - 100k Da)" osel:=SoA title:"Analyte size"
  desc:"Check the box of the analyte-option you want to
  choose. Check only ONE box" olist:=AB;
9 if (AB.GetLength() == 1) {
10   dlgChkList inames:"Multiple single spectra|Stacked
    spectra (2 - 6)" osel:=NoS title:"Plot type"
    desc:"Check the box of the plotting-option you want to
    choose. Check only ONE box" olist:=CB;
11   if (CB.GetLength() == 1) {
12     //Multiple single spektra
13     if (NoS == 1) {
14       //Start of the "Multiple single spectra" if command
15       type -b "All files you will import will be plotted singly in
        individual spectra.";
16       //Import of data
17       impasc menu:=multiple
18       options.Names.FNameToBkLnOnly:=1
19       options.FileStruct.NumericSeparator:=0
20       options.scripts.ScriptAfterEachImport:={
21         //Datasheet editing
22         %V=%H;
23         wks.col1.Iname$ = m/z;
24         wks.col2.Iname$ = Absolute Intensity;
25         wks.col2.comment$ = wks.name$;
26         MH11=max(col(2));
27         %Z = wks.col2.comment$;
28         %N = %V_D;
29         //Plotting imported data and sizing graph
30         plotxy iy:=(1,2) plot:=200 color:=color(0,0,0);
31         %W = %H;
32         layer.unit=1;
33         layer.width=85;
34         layer.height=85;
35         layer.left=12;
36         layer.top=5;
37         y1.fsize=18.5;
38         xb.fsize=18.5;
39         //Sizing the labels and change their position
40         layer1.y1.label.pt=14.5;
41         layer1.x1.label.pt=14.5;
42         layer1.y1.label.offsetH=20;
43         //Editing of x/y-axis
44         layer.x1.labels subtype=1;
45         layer.x1.showaxes=3;
46         layer.x1.ticks=1 + 4;
47         layer.y1.labels subtype=2;
48         layer.y1.showaxes=3;
49         layer.y1.ticks=1 + 4;
50         //Activation & design of the layer grids
51         layer1.x.showGrids=1;
52         layer1.x.grid.majorColor=color(192,192,192);
53         layer1.x.grid.majorType=1;
54         //Peakpicking and Peak Plotting
55         pkfind dir:=p method:=max npts:=10 filter:=h
          value:=MH11 ocenter_x:=<new> ocenter_y:=<new>
          ocenter:=<none>;
56         win -a %V;
57         wks.col3.Iname$=m/z;
58         wks.col4.Iname$=Absolut Intensity;
59         wks.col4.comment$=Maxima Spectrum 1;
60         plotxy iy:=(3,4) plot:=201 color:=color(255,0,0)
          ogl:=[%W]1! rescale:=0;
61         win -a %W;
62         set %N -k 2;
63         set %N -z 4;
64         set %N -q 1;
65         set %N -qm 1;
66         set %N -qb;
67         set %N -qc 2;
68         set %N -qp 4;
69         set %N -qr 90;
70         set %N -qy 20;
71         set %N -qs 14;
72         themeApply2g theme:= "MS";
73         //Editoring of Legend
74         label -r legend;
75         label -s -n leglay1 lb(%Z);
76         leglay1.fsize = 15;
77         leglay1.color=color(0,0,0);
78         leglay1.background=1;
79         leglay1.y = layer1.y.to + leglay1.dy / 1.3;
80         leglay1.x = layer1.x.from + leglay1.dx / 2;
81         //Add scrollbar and edit its apperance
82         axis_scrollbar axis:=right begin:=0 end:=145000;
83         yscrollbar.color=color(0,0,0);
84         yscrollbar.fillcolor=color(0,0,0);
85         yscrollthumb.color=color(0,0,0);
86         yscrollthumb.fillcolor=color(0,0,0);
87         yscrollbar.states=0;
88         yscrollbar.height=4150;
89         yscrollbar.top=240;
90         yscrollbar.width=150;
91         yscrollbar.left=6262;
92         yscrollbar.states=64;
93         yscrollbar.states=128;
94         yscrollbar.states=1;
95         yscrollbar.states=2;
96         yscrollbar.states=4;
97         yscrollbar.states=8;
98         yscrollbar.states=16;
99         yscrollbar.states=32;
100        yscrollbar.mouse=0;
101        //Add x-scroller and edit its apperance
102        axis_scroller xsize:=0.6 ysize:=0.8;
103        XMin.color=color(0,0,0);
104        XMin.fillcolor=color(0,0,0);
105        XMax.color=color(0,0,0);
106        XMax.fillcolor=color(0,0,0);
107        //Setting of axis range
108        layer.y1.from = -7000;
109        layer.y1.inc = 5000;
110        layer.y1.to = 270000;
111        if (SoA == 1) {
112          layer.x1.from = 0;
113          layer.x1.inc = 500;
114          layer.x1.to = 3000;
115        };
116        if (SoA == 2) {
```

Appendix

LabTalk script (OriginPro 2020b)

```
117 layer.x1.from = 1000;
118 layer.x1.inc = 10000;
119 layer.x1.to = 100000;
120 };
121 page.longname$ = %Z;
122 page.title = 1;
123 //Small final changes
124 win -a %V;
125 wks.col4.comment$ = "Peaks";
126 wks.name$=Sheet1;
127 page.title = 1;
128 win -ia;
129 win -a %W;
130 speedmode index:=project sm:=off;
131 };
132 };
133 //Stacked plots (2 - 6 plots)
134 if (NoS == 2) {
135 //Start of the "Stacked plots (2 - 6 spectra)" if command
136 type -b "The number of imported spectra is equal to the
number of stacked spectra. The first chosen file will be
plotted in the upper spectrum (and so on). Please do
not choose more then 6 files.";
137 //Import and preparation of data
138 newbook;
139 impasc
140 menu:=multiple
141 options.Names.FNameToBkLnOnly:=1
142 options.Mode:=3
143 options.FileStruct.NumericSeparator:=0
144 options.scripts.ScriptAfterEachImport:={
145 %L=%H;
146 if (kk == 6){
147 //6. Data file
148 %U=wks.name$;
149 wks.name$=Sheet1;
150 if (yesno == 1) {
151 normalize iy:=col(2) method:=range100 oy:=<input>;
152 };
153 colcopy irng:=[%L]Sheet1!col(1)
orng:=[%V]Sheet1!col(11);
154 colcopy irng:=[%L]Sheet1!col(2)
orng:=[%V]Sheet1!col(12);
155 win -c %L;
156 };
157 if (kk == 5){
158 //5. Data file
159 %T=wks.name$;
160 wks.name$=Sheet1;
161 if (yesno == 1) {
162 normalize iy:=col(2) method:=range100 oy:=<input>;
163 };
164 colcopy irng:=[%L]Sheet1!col(1)
orng:=[%V]Sheet1!col(9);
165 colcopy irng:=[%L]Sheet1!col(2)
orng:=[%V]Sheet1!col(10);
166 win -c %L;
167 kk++;
168 };
169 if (kk == 4){
170 //4. Data file
171 %S=wks.name$;
172 wks.name$=Sheet1;
173 if (yesno == 1) {
174 normalize iy:=col(2) method:=range100 oy:=<input>;
175 };
176 colcopy irng:=[%L]Sheet1!col(1)
orng:=[%V]Sheet1!col(7);
177 colcopy irng:=[%L]Sheet1!col(2)
orng:=[%V]Sheet1!col(8);
178 win -c %L;
179 kk++;
180 };
181 if (kk == 3){
182 //3. Data file
183 %R=wks.name$;
184 wks.name$=Sheet1;
185 if (yesno == 1) {
186 normalize iy:=col(2) method:=range100 oy:=<input>;
187 };
188 colcopy irng:=[%L]Sheet1!col(1)
orng:=[%V]Sheet1!col(5);
189 colcopy irng:=[%L]Sheet1!col(2)
orng:=[%V]Sheet1!col(6);
190 win -c %L;
191 kk++;
192 };
193 if (kk == 2){
194 //2. Data file
195 %Q=wks.name$;
196 wks.name$=Sheet1;
197 if (yesno == 1) {
198 normalize iy:=col(2) method:=range100 oy:=<input>;
199 };
200 colcopy irng:=[%L]Sheet1!col(1)
orng:=[%V]Sheet1!col(3);
201 colcopy irng:=[%L]Sheet1!col(2)
orng:=[%V]Sheet1!col(4);
202 win -c %L;
203 kk++;
204 };
205 if (kk == 1){
206 //1. Data file
207 %P=wks.name$;
208 wks.name$=Sheet1;
209 if (yesno == 1) {
210 normalize iy:=col(2) method:=range100 oy:=<input>;
211 };
212 colcopy irng:=[%L]Sheet1!col(1)
orng:=[%V]Sheet1!col(1);
213 colcopy irng:=[%L]Sheet1!col(2)
orng:=[%V]Sheet1!col(2);
214 win -c %L;
215 kk++;
216 };
217 };
218 //Stackplot of 6 spectra
219 if (exist(col(12)) == 1) {
220 newbook dummy;
221 //Enter information about plotted spectra, measurement
parameter and peakpicking parameter
222 col(a)[1]$="Spektrum 1";
223 col(a)[2]$="Spektrum 2";
224 col(a)[3]$="Spektrum 3";
225 col(a)[4]$="Spektrum 4";
226 col(a)[5]$="Spektrum 5";
227 col(a)[6]$="Spektrum 6";
228 col(b)[1]$=%P;
229 col(b)[2]$=%Q;
230 col(b)[3]$=%R;
231 col(b)[4]$=%S;
232 col(b)[5]$=%T;
233 col(b)[6]$=%U;
234 wcolwidth irng:=Col(2) width:=100;
235 win -z;
236 getn
237 (First Spectrum:) :@G
238 (Matrix) NoM61$
239 (Sample) NoS61$
240 (Additive) NoA61$
241 (Laser-Intensity) LI61$
242 (-) :@G
243 (Second Spectrum:) :@G
244 (Matrix) NoM62$
245 (Sample) NoS62$
246 (Additive) NoA62$
247 (Laser-Intensity) LI62$
248 (-) :@G
```

Appendix

LabTalk script (OriginPro 2020b)

```

249 (Third Spectrum:) :@G
250 (Matrix) NoM63$
251 (Sample) NoS63$
252 (Additive) NoA63$
253 (Laser-Intensity) LI63$
254 (-) :@G
255 (Fourth Spectrum:) :@G
256 (Matrix) NoM64$
257 (Sample) NoS64$
258 (Additive) NoA64$
259 (Laser-Intensity) LI64$
260 (-) :@G
261 (Fifth Spectrum:) :@G
262 (Matrix) NoM65$
263 (Sample) NoS65$
264 (Additive) NoA65$
265 (Laser-Intensity) LI65$
266 (-) :@G
267 (Sixth Spectrum:) :@G
268 (Matrix) NoM66$
269 (Sample) NoS66$
270 (Additive) NoA66$
271 (Laser-Intensity) LI66$
272 (-) :@G
273 (General Parameters:) :@G
274 (Detection mode) DM6$
275 (Sample preparation) SP6$
276 (-) :@G
277 (Please insert the measurement parameters for plot la-
    beling and the abbreviations used for the data file
    names of the last three spectra.);
278 win -c;
279 MH61=max(col(2));
280 MH62=max(col(4));
281 MH63=max(col(6));
282 MH64=max(col(8));
283 MH65=max(col(10));
284 MH66=max(col(12));
285 //Editing of data book
286 win -a %V;
287 wks.col1.lname$ = m/z;
288 wks.col2.lname$ = Absolute Intensity;
289 wks.col2.comment$ = "Matrix: %(NoM61$) ; Analyte:
    %(NoS61$) ; Additive: %(NoA61$)";
290 wks.col3.lname$ = m/z;
291 wks.col4.lname$ = Absolute Intensity;
292 wks.col4.comment$ = "Matrix: %(NoM62$) ; Analyte:
    %(NoS62$) ; Additive: %(NoA62$)";
293 wks.col5.lname$ = m/z;
294 wks.col6.lname$ = Absolute Intensity;
295 wks.col6.comment$ = "Matrix: %(NoM63$) ; Analyte:
    %(NoS63$) ; Additive: %(NoA63$)";
296 wks.col7.lname$ = m/z;
297 wks.col8.lname$ = Absolute Intensity;
298 wks.col8.comment$ = "Matrix: %(NoM64$) ; Analyte:
    %(NoS64$) ; Additive: %(NoA64$)";
299 wks.col9.lname$ = m/z;
300 wks.col10.lname$ = Absolute Intensity;
301 wks.col10.comment$ = "Matrix: %(NoM65$) ; Analyte:
    %(NoS65$) ; Additive: %(NoA65$)";
302 wks.col11.lname$ = m/z;
303 wks.col12.lname$ = Absolute Intensity;
304 wks.col12.comment$ = "Matrix: %(NoM66$) ; Analyte:
    %(NoS66$) ; Additive: %(NoA66$)";
305 //Stack plotting
306 plotxy iy:=(1,2) plot:=200 color:=color(128,0,128);
307 %W = %H;
308 layadd igp:=%W type:=normal userdef:=1 linkto:=1
    xaxis:=straight yaxis:=straight;
309 layadd igp:=%W type:=normal userdef:=1 linkto:=1
    xaxis:=straight yaxis:=straight;
310 layadd igp:=%W type:=normal userdef:=1 linkto:=1
    xaxis:=straight yaxis:=straight;
311 layadd igp:=%W type:=normal userdef:=1 linkto:=1
    xaxis:=straight yaxis:=straight;
312 layadd igp:=%W type:=normal userdef:=1 linkto:=1
    xaxis:=straight yaxis:=straight;
313 win -a %V;
314 plotxy iy:=(3,4) plot:=200 color:=color(128,128,0)
    ogl:=[%W]2!;
315 plotxy iy:=(5,6) plot:=200 color:=color(0,128,0)
    ogl:=[%W]3!;
316 plotxy iy:=(7,8) plot:=200 color:=color(255,0,0)
    ogl:=[%W]4!;
317 plotxy iy:=(9,10) plot:=200 color:=color(0,0,255)
    ogl:=[%W]5!;
318 plotxy iy:=(11,12) plot:=200 color:=color(0,0,0)
    ogl:=[%W]6!;
319 win -a %W;
320 //Sizing of the layers
321 layer1.unit=1;
322 layer1.width=88.2;
323 layer1.height=14.6;
324 layer1.left=9.3;
325 layer1.top=3.6;
326 layer2.unit=1;
327 layer2.width=88.2;
328 layer2.height=14.6;
329 layer2.left=9.3;
330 layer2.top=18.2;
331 layer3.unit=1;
332 layer3.width=88.2;
333 layer3.height=14.6;
334 layer3.left=9.3;
335 layer3.top=32.8;
336 layer4.unit=1;
337 layer4.width=88.2;
338 layer4.height=14.6;
339 layer4.left=9.3;
340 layer4.top=47.4;
341 layer5.unit=1;
342 layer5.width=88.2;
343 layer5.height=14.6;
344 layer5.left=9.3;
345 layer5.top=62;
346 layer6.unit=1;
347 layer6.width=88.2;
348 layer6.height=14.6;
349 layer6.left=9.3;
350 layer6.top=76.6;
351 //Sizing the labels and change their position
352 layer1.y1.label.pt=34;
353 layer1.x2.label.pt=34;
354 layer2.y1.label.pt=34;
355 layer3.y1.label.pt=34;
356 layer4.y1.label.pt=34;
357 layer4.y1.label.pt=34;
358 layer5.y1.label.pt=34;
359 layer6.y1.label.pt=34;
360 layer6.x1.label.pt=34;
361 layer1.x2.label.offsetV=-30;
362 layer1.y1.label.offsetH=20;
363 layer2.y1.label.offsetH=20;
364 layer3.y1.label.offsetH=20;
365 layer4.y1.label.offsetH=20;
366 layer5.y1.label.offsetH=20;
367 layer6.y1.label.offsetH=20;
368 //Scaling of the x/y-axis
369 if (SoA == 1) {
370 layer.x1.from = 0;
371 layer.x1.inc = 500;
372 layer.x1.to = 3000;
373 };
374 if (SoA == 2) {
375 layer.x1.from = 1000;
376 layer.x1.inc = 10000;
377 layer.x1.to = 100000;

```


Appendix

LabTalk script (OriginPro 2020b)

```
378 };
379 if (yesno == 1) {
380 layer1.y1.from = -15;
381 layer1.y1.inc = 30;
382 layer1.y1.to = 140;
383 };
384 if (yesno == 0) {
385 layer1.y1.from = -15000;
386 layer1.y1.inc = 100000;
387 layer1.y1.to = 250000;
388 };
389 //Design of the layers
390 layer1.x1.labelssubtype=1;
391 layer1.x1.showaxes=3;
392 if (yesno == 1) {
393 layer1.y1.labelssubtype=1;
394 };
395 if (yesno == 0) {
396 layer1.y1.labelssubtype=2;
397 };
398 layer1.y1.showaxes=3;
399 layer1.x1.ticks=1 + 2 + 4 + 8;
400 layer1.x2.ticks = 2 + 8;
401 layer1.y1.ticks=1 + 4;
402 layer1.y1.label.decPlaces=0;
403 layer1.x.showLabels=2;
404 if (yesno == 1) {
405 layer2.y1.labelssubtype=1;
406 };
407 if (yesno == 0) {
408 layer2.y1.labelssubtype=2;
409 };
410 layer2.y1.showaxes=3;
411 layer2.y1.ticks=1 + 4;
412 layer2.x1.ticks=1 + 2 + 4 + 8;
413 layer2.y1.label.decPlaces=0;
414 layer2.x.showLabels=0;
415 if (yesno == 1) {
416 layer3.y1.labelssubtype=1;
417 };
418 if (yesno == 0) {
419 layer3.y1.labelssubtype=2;
420 };
421 layer3.y1.showaxes=3;
422 layer3.y1.ticks=1 + 4;
423 layer3.x1.ticks=1 + 2 + 4 + 8;
424 layer3.y1.label.decPlaces=0;
425 layer3.x.showLabels=0;
426 if (yesno == 1) {
427 layer4.y1.labelssubtype=1;
428 };
429 if (yesno == 0) {
430 layer4.y1.labelssubtype=2;
431 };
432 layer4.y1.showaxes=3;
433 layer4.y1.ticks=1 + 4;
434 layer4.x1.ticks=1 + 2 + 4 + 8;
435 layer4.y1.label.decPlaces=0;
436 layer4.x.showLabels=0;
437 if (yesno == 1) {
438 layer5.y1.labelssubtype=1;
439 };
440 if (yesno == 0) {
441 layer5.y1.labelssubtype=2;
442 };
443 layer5.y1.showaxes=3;
444 layer5.y1.ticks=1 + 4;
445 layer5.x1.ticks=1 + 2 + 4 + 8;
446 layer5.y1.label.decPlaces=0;
447 layer5.x.showLabels=0;
448 if (yesno == 1) {
449 layer6.y1.labelssubtype=1;
450 };
451 if (yesno == 0) {
452 layer6.y1.labelssubtype=2;
453 };
454 layer6.y1.showaxes=3;
455 layer6.y1.ticks=1 + 4;
456 layer6.x1.ticks=1 + 4;
457 layer6.y1.label.decPlaces=0;
458 //Activation & design of the layer grids
459 layer1.x.showGrids=1;
460 layer1.x.grid.majorColor=color(192,192,192);
461 layer1.x.grid.majorType=1;
462 layer2.x.showGrids=1;
463 layer2.x.grid.majorColor=color(192,192,192);
464 layer2.x.grid.majorType=1;
465 layer3.x.showGrids=1;
466 layer3.x.grid.majorColor=color(192,192,192);
467 layer3.x.grid.majorType=1;
468 layer4.x.showGrids=1;
469 layer4.x.grid.majorColor=color(192,192,192);
470 layer4.x.grid.majorType=1;
471 layer5.x.showGrids=1;
472 layer5.x.grid.majorColor=color(192,192,192);
473 layer5.x.grid.majorType=1;
474 layer6.x.showGrids=1;
475 layer6.x.grid.majorColor=color(192,192,192);
476 layer6.x.grid.majorType=1;
477 //FIRST plot editing (layer 1)
478 %W!page.active = 1;
479 //Peakpicking and peak plotting (layer 1)
480 pkfind dir:=p method:=max npts:=10 filter:=h
value:=MH61 ocenter_x:=<new> ocenter_y:=<new>
ocenter:=<none>;
481 win -a %V;
482 wks.col13.lname$=m/z;
483 wks.col14.lname$=Absolut Intensity;
484 wks.col14.comment$=Maxima Spectrum 1;
485 plotxy iy:=(13,14) plot:=201 color:=color(128,0,128)
ogl:=[%W]! rescale:=0;
486 win -a %W;
487 %J = %V_N;
488 set %J -k 2;
489 set %J -z 4;
490 set %J -q 1;
491 set %J -qm 1;
492 set %J -qb;
493 set %J -qc 10;
494 set %J -qp 4;
495 set %J -qr 90;
496 set %J -qy 20;
497 set %J -qs 14;
498 themeApply2g theme:= "MS";
499 //Delete axis labels and make plot line thicker (layer 1)
500 label -r xb;
501 label -r yl;
502 %J = %V_B;
503 set %J -w 1000;
504 //Replacing and editing of the legend label (layer 1)
505 label -r legend;
506 label -s -n leglay1 lb(%(1Y, @LC));
507 leglay1.fsize = 20;
508 leglay1.color=color(128,0,128);
509 leglay1.background=1;
510 leglay1.y = layer1.y.to - leglay1.dy / 2;
511 leglay1.x = layer1.x.from + leglay1.dx / 2;
512 //Creating and editing of Laser-Intensity label (layer 1)
513 label -s -n lasintlay1 lb(\p80(Laser-Intensity: %(LI61$)));
514 lasintlay1.rotate= 90;
515 lasintlay1.y = (layer.y.to - layer.y.from) / 2 - 15000;
516 lasintlay1.x = layer.x.to + lasintlay1.dx / 2;
517 //Add scrollbar and edit its appearance (layer 1)
518 axis_scrollbar axis:=left begin:=%(layer1.y1.from$)
end:=%(layer1.y1.to$);
519 yscrollbar.color=color(128,0,128);
520 yscrollbar.fillcolor=color(128,0,128);
521 yscrollthumb.color=color(128,0,128);
```

Appendix

LabTalk script (OriginPro 2020b)

```
522 yscrollthumb.fillcolor=color(128,0,128);
523 yscrollbar.states=0;
524 yscrollbar.height=710;
525 yscrollbar.top=177;
526 yscrollbar.width=125;
527 yscrollbar.left=6290;
528 yscrollbar.states=64;
529 yscrollbar.states=128;
530 yscrollbar.states=1;
531 yscrollbar.states=2;
532 yscrollbar.states=4;
533 yscrollbar.states=8;
534 yscrollbar.states=16;
535 yscrollbar.states=32;
536 yscrollbar.mouse=0;
537 //Rescaling of y-axis (layer 1)
538 if (yesno == 1) {
539   layer1.y1.from = -15;
540   layer1.y1.inc = 30;
541   layer1.y1.to = 140;
542 };
543 if (yesno == 0) {
544   layer1.y1.from = -15000;
545   layer1.y1.inc = 100000;
546   layer1.y1.to = 250000;
547 };
548 //SECOND plot editing (layer 2)
549 %W!page.active = 2;
550 //Peakpicking and peak plotting (layer 2)
551 pkfind dir:=p method:=max npts:=10 filter:=h
value:=MH62 ocenter_x:=<new> ocenter_y:=<new>
ocenter:=<none>;
552 win -a %V;
553 wks.col15.Iname$=m/z;
554 wks.col16.Iname$=Absolut Intensity;
555 wks.col16.comment$=Maxima Spectrum 2;
556 plotxy iy:=(15,16) plot:=201 color:=color(128,128,0)
ogl:=[%W]2! rescale:=0;
557 win -a %W;
558 %K = %V_P;
559 set %K -k 2;
560 set %K -z 4;
561 set %K -q 1;
562 set %K -qm 1;
563 set %K -qb;
564 set %K -qc 8;
565 set %K -qp 4;
566 set %K -qr 90;
567 set %K -qy 20;
568 set %K -qs 14;
569 themeApply2g theme:= "MS";
570 //Delete axis labels and make plot line thicker (layer 2)
571 label -r xb;
572 label -r yl;
573 %K = %V_D;
574 set %K -w 1000;
575 //Replacing and editing of the legend label (layer 2)
576 label -r legend;
577 label -s -n leglay2 \b(%(1Y, @LC));
578 leglay2.fsize = 20;
579 leglay2.color=color(128,128,0);
580 leglay2.background=1;
581 leglay2.y = layer2.y.to - leglay2.dy / 2;
582 leglay2.x = layer2.x.from + leglay2.dx / 2;
583 //Creating and editing of Laser-Intensity label (layer 2)
584 label -s -n lasintlay2 \b\p80(Laser-Intensity: %(LI62$));
585 lasintlay2.rotate= 90;
586 lasintlay2.y = (layer.y.to - layer.y.from) / 2 - 15000;
587 lasintlay2.x = layer.x.to + lasintlay2.dx / 2;
588 //Add scrollbar and edit its appearance (layer 2)
589 axis_scrollbar axis:=left begin:=(layer2.y1.from$)
end:=(layer2.y1.to$);
590 yscrollbar.color=color(128,128,0);
591 yscrollbar.fillcolor=color(128,128,0);
592 yscrollthumb.color=color(128,128,0);
593 yscrollthumb.fillcolor=color(128,128,0);
594 yscrollbar.states=0;
595 yscrollbar.height=710;
596 yscrollbar.top=897;
597 yscrollbar.width=125;
598 yscrollbar.left=6290;
599 yscrollbar.states=64;
600 yscrollbar.states=128;
601 yscrollbar.states=1;
602 yscrollbar.states=2;
603 yscrollbar.states=4;
604 yscrollbar.states=8;
605 yscrollbar.states=16;
606 yscrollbar.states=32;
607 yscrollbar.mouse=0;
608 //Rescaling of y-axis (layer 2)
609 if (yesno == 1) {
610   layer1.y1.from = -15;
611   layer1.y1.inc = 30;
612   layer1.y1.to = 140;
613 };
614 if (yesno == 0) {
615   layer1.y1.from = -15000;
616   layer1.y1.inc = 100000;
617   layer1.y1.to = 250000;
618 };
619 //THIRD plot editing (layer 3)
620 %W!page.active = 3;
621 //Peakpicking and peak plotting (layer 3)
622 pkfind dir:=p method:=max npts:=10 filter:=h
value:=MH63 ocenter_x:=<new> ocenter_y:=<new>
ocenter:=<none>;
623 win -a %V;
624 wks.col17.Iname$=m/z;
625 wks.col18.Iname$=Absolut Intensity;
626 wks.col18.comment$=Maxima Spectrum 3;
627 plotxy iy:=(17,18) plot:=201 color:=color(0,128,0)
ogl:=[%W]3! rescale:=0;
628 win -a %W;
629 %L = %V_R;
630 set %L -k 2;
631 set %L -z 4;
632 set %L -q 1;
633 set %L -qm 1;
634 set %L -qb;
635 set %L -qc 12;
636 set %L -qp 4;
637 set %L -qr 90;
638 set %L -qy 20;
639 set %L -qs 14;
640 themeApply2g theme:= "MS";
641 //Delete axis labels and make plot line thicker (layer 3)
642 label -r xb;
643 label -r yl;
644 %L = %V_F;
645 set %L -w 1000;
646 //Replacing and editing of the legend label (layer 3)
647 label -r legend;
648 label -s -n leglay3 \b(%(1Y, @LC));
649 leglay3.fsize = 20;
650 leglay3.color=color(0,128,0);
651 leglay3.background=1;
652 leglay3.y = layer3.y.to - leglay3.dy / 2;
653 leglay3.x = layer3.x.from + leglay3.dx / 2;
654 //Creating and editing of Laser-Intensity label (layer 3)
655 label -s -n lasintlay3 \b\p80(Laser-Intensity: %(LI63$));
656 lasintlay3.rotate= 90;
657 lasintlay3.y = (layer.y.to - layer.y.from) / 2 - 15000;
658 lasintlay3.x = layer.x.to + lasintlay3.dx / 2;
659 //Add scrollbar and edit its appearance (layer 3)
660 axis_scrollbar axis:=left begin:=(layer3.y1.from$)
end:=(layer3.y1.to$);
661 yscrollbar.color=color(0,128,0);
```

Appendix

LabTalk script (OriginPro 2020b)

```
662 yscrollbar.fillcolor=color(0,128,0);
663 yscrollthumb.color=color(0,128,0);
664 yscrollthumb.fillcolor=color(0,128,0);
665 yscrollbar.states=0;
666 yscrollbar.height=710;
667 yscrollbar.top=1617;
668 yscrollbar.width=125;
669 yscrollbar.left=6290;
670 yscrollbar.states=64;
671 yscrollbar.states=128;
672 yscrollbar.states=1;
673 yscrollbar.states=2;
674 yscrollbar.states=4;
675 yscrollbar.states=8;
676 yscrollbar.states=16;
677 yscrollbar.states=32;
678 yscrollbar.mouse=0;
679 //Rescaling of y-axis (layer 3)
680 if (yesno == 1) {
681 layer1.y1.from = -15;
682 layer1.y1.inc = 30;
683 layer1.y1.to = 140;
684 };
685 if (yesno == 0) {
686 layer1.y1.from = -15000;
687 layer1.y1.inc = 100000;
688 layer1.y1.to = 250000;
689 };
690 //FOURTH plot editing (layer 4)
691 %W!page.active = 4;
692 //Peakpicking and peak plotting (layer 4)
693 pkfind dir:=p method:=max npts:=10 filter:=h
value:=MH64 ocenter_x:=<new> ocenter_y:=<new>
ocenter:=<none>;
694 win -a %V;
695 wks.col19.lname$=m/z;
696 wks.col20.lname$=Absolut Intensity;
697 wks.col20.comment$=Maxima Spectrum 4;
698 plotxy iy:=(19,20) plot:=201 color:=color(255,0,0)
ogl:=[%W]4! rescale:=0;
699 win -a %W;
700 %M = %V_T;
701 set %M -k 2;
702 set %M -z 4;
703 set %M -q 1;
704 set %M -qm 1;
705 set %M -qb;
706 set %M -qc 2;
707 set %M -qp 4;
708 set %M -qr 90;
709 set %M -qy 20;
710 set %M -qs 14;
711 themeApply2g theme:= "MS";
712 //Delete axis labels and make plot line thicker (layer 4)
713 label -r xb;
714 label -r yl;
715 %M = %V_H;
716 set %M -w 1000;
717 //Replacing and editing of the legend label (layer 4)
718 label -r legend;
719 label -s -n leglay4 \b(1Y, @LC));
720 leglay4.fsize = 20;
721 leglay4.color=color(255,0,0);
722 leglay4.background=1;
723 leglay4.y = layer4.y.to - leglay4.dy / 2;
724 leglay4.x = layer4.x.from + leglay4.dx / 2;
725 //Creating and editing of Laser-Intensity label (layer 4)
726 label -s -n lasintlay4 \b(p80(Laser-Intensity: %(LI64$)));
727 lasintlay4.rotate= 90;
728 lasintlay4.y = (layer.y.to - layer.y.from) / 2 - 15000;
729 lasintlay4.x = layer.x.to + lasintlay4.dx / 2;
730 //Add scrollbar and edit its apperance (layer 4)
731 axis_scrollbar axis:=left begin:=%(layer4.y1.from$)
end:=%(layer4.y1.to$);
732 yscrollbar.color=color(255,0,0);
733 yscrollbar.fillcolor=color(255,0,0);
734 yscrollthumb.color=color(255,0,0);
735 yscrollthumb.fillcolor=color(255,0,0);
736 yscrollbar.states=0;
737 yscrollbar.height=710;
738 yscrollbar.top=2337;
739 yscrollbar.width=125;
740 yscrollbar.left=6290;
741 yscrollbar.states=64;
742 yscrollbar.states=128;
743 yscrollbar.states=1;
744 yscrollbar.states=2;
745 yscrollbar.states=4;
746 yscrollbar.states=8;
747 yscrollbar.states=16;
748 yscrollbar.states=32;
749 yscrollbar.mouse=0;
750 //Rescaling of y-axis (layer 4)
751 if (yesno == 1) {
752 layer1.y1.from = -15;
753 layer1.y1.inc = 30;
754 layer1.y1.to = 140;
755 };
756 if (yesno == 0) {
757 layer1.y1.from = -15000;
758 layer1.y1.inc = 100000;
759 layer1.y1.to = 250000;
760 };
761 //FIFTH plot editing (layer 5)
762 %W!page.active = 5;
763 //Peakpicking and peak plotting (layer 5)
764 pkfind dir:=p method:=max npts:=10 filter:=h
value:=MH65 ocenter_x:=<new> ocenter_y:=<new>
ocenter:=<none>;
765 win -a %V;
766 wks.col21.lname$=m/z;
767 wks.col22.lname$=Absolut Intensity;
768 wks.col22.comment$=Maxima Spectrum 5;
769 plotxy iy:=(21,22) plot:=201 color:=color(0,0,255)
ogl:=[%W]5! rescale:=0;
770 win -a %W;
771 %N = %V_V;
772 set %N -k 2;
773 set %N -z 4;
774 set %N -q 1;
775 set %N -qm 1;
776 set %N -qb;
777 set %N -qc 4;
778 set %N -qp 4;
779 set %N -qr 90;
780 set %N -qy 20;
781 set %N -qs 14;
782 themeApply2g theme:= "MS";
783 //Delete axis labels and make plot line thicker (layer 5)
784 label -r xb;
785 label -r yl;
786 %N = %V_J;
787 set %N -w 1000;
788 //Replacing and editing of the legend label (layer 5)
789 label -r legend;
790 label -s -n leglay5 \b(1Y, @LC));
791 leglay5.fsize = 20;
792 leglay5.color=color(0,0,255);
793 leglay5.background=1;
794 leglay5.y = layer5.y.to - leglay5.dy / 2;
795 leglay5.x = layer5.x.from + leglay5.dx / 2;
796 //Creating and editing of Laser-Intensity label (layer 5)
797 label -s -n lasintlay5 \b(p80(Laser-Intensity: %(LI65$)));
798 lasintlay5.rotate= 90;
799 lasintlay5.y = (layer.y.to - layer.y.from) / 2 - 15000;
800 lasintlay5.x = layer.x.to + lasintlay5.dx / 2;
801 //Add scrollbar and edit its apperance (layer 5)
```

Appendix

LabTalk script (OriginPro 2020b)

```
802 axis_scrollbar axis:=left begin:=%(layer5.y1.from$)
end:=%(layer5.y1.to$);
803 yscrollbar.color=color(0,0,255);
804 yscrollbar.fillcolor=color(0,0,255);
805 yscrollthumb.color=color(0,0,255);
806 yscrollthumb.fillcolor=color(0,0,255);
807 yscrollbar.states=0;
808 yscrollbar.height=710;
809 yscrollbar.top=3057;
810 yscrollbar.width=125;
811 yscrollbar.left=6290;
812 yscrollbar.states=64;
813 yscrollbar.states=128;
814 yscrollbar.states=1;
815 yscrollbar.states=2;
816 yscrollbar.states=4;
817 yscrollbar.states=8;
818 yscrollbar.states=16;
819 yscrollbar.states=32;
820 yscrollbar.mouse=0;
821 //Rescaling of y-axis (layer 5)
822 if (yesno == 1) {
823 layer1.y1.from = -15;
824 layer1.y1.inc = 30;
825 layer1.y1.to = 140;
826 };
827 if (yesno == 0) {
828 layer1.y1.from = -15000;
829 layer1.y1.inc = 100000;
830 layer1.y1.to = 250000;
831 };
832 //SIXTH plot editing (layer 6)
833 %W!page.active = 6;
834 //Peakpicking and peak plotting (layer 6)
835 pkfind dir:=p method:=max npts:=10 filter:=h
value:=MH66 ocenter_x:=<new> ocenter_y:=<new>
ocenter:=<none>;
836 win -a %V;
837 wks.col23.lname$=m/z;
838 wks.col24.lname$=Absolut Intensity;
839 wks.col24.comment$=Maxima Spectrum 6;
840 plotxy iy:=(23,24) plot:=201 color:=color(0,0,0)
ogl:=[%W]6! rescale:=0;
841 win -a %W;
842 %O = %V_X;
843 set %O -k 2;
844 set %O -z 4;
845 set %O -q 1;
846 set %O -qm 1;
847 set %O -qb;
848 set %O -qc 0;
849 set %O -qp 4;
850 set %O -qr 90;
851 set %O -qy 20;
852 set %O -qs 14;
853 theme Apply2g theme:= "MS";
854 //Edit axis labels and make plot line thicker (layer 6)
855 yl.fsize=42;
856 xb.fsize=42;
857 yl.top=1980;
858 %O = %V_L;
859 set %O -w 1000;
860 //Replacing and editing of the legend label (layer 6)
861 label -r legend;
862 label -s -n leglay6 lb(%(1Y, @LC));
863 leglay6.fsize = 20;
864 leglay6.color=color(0,0,0);
865 leglay6.background=1;
866 leglay6.y = layer6.y.to - leglay6.dy / 2;
867 leglay6.x = layer6.x.from + leglay6.dx / 2;
868 //Creating and editing of Laser-Intensity label (layer 6)
869 label -s -n lasintlay6 lb(p80(Laser-Intensity: %(LI66$)));
870 lasintlay6.rotate= 90;
871 lasintlay6.y = (layer.y.to - layer.y.from) / 2 - 15000;
872 lasintlay6.x = layer.x.to + lasintlay6.dx / 2;
873 //Add scrollbar and edit its appearance (layer 6)
874 axis_scrollbar axis:=left begin:=%(layer6.y1.from$)
end:=%(layer6.y1.to$);
875 yscrollbar.color=color(0,0,0);
876 yscrollbar.fillcolor=color(0,0,0);
877 yscrollthumb.color=color(0,0,0);
878 yscrollthumb.fillcolor=color(0,0,0);
879 yscrollbar.states=0;
880 yscrollbar.height=710;
881 yscrollbar.top=3777;
882 yscrollbar.width=125;
883 yscrollbar.left=6290;
884 yscrollbar.states=64;
885 yscrollbar.states=128;
886 yscrollbar.states=1;
887 yscrollbar.states=2;
888 yscrollbar.states=4;
889 yscrollbar.states=8;
890 yscrollbar.states=16;
891 yscrollbar.states=32;
892 yscrollbar.mouse=0;
893 //Rescaling of y-axis (layer 6)
894 if (yesno == 1) {
895 layer1.y1.from = -15;
896 layer1.y1.inc = 30;
897 layer1.y1.to = 140;
898 };
899 if (yesno == 0) {
900 layer1.y1.from = -15000;
901 layer1.y1.inc = 100000;
902 layer1.y1.to = 250000;
903 };
904 //Add x-scroller and edit its appearance
905 axis_scroller xsize:=0.3 ysize:=-1.5;
906 XMin.color=color(0,0,0);
907 XMin.fillcolor=color(0,0,0);
908 XMax.color=color(0,0,0);
909 XMax.fillcolor=color(0,0,0);
910 //Adding measurement parameter labels
911 label -s -d 4130 4900 -n detectionmode \p80(Detection-
mode: %(DM6$));
912 label -s -d 10 4900 -n sampleprep \p80(Sample-prepa-
ration: %(SP6$));
913 win -a %V;
914 win -ia;
915 win -a %W;
916 win -z;
917 type -b "Frist spectrum: \n%P \nSecond spectrum:
\n%Q \nThird spectrum: \n%R \nFourth spectrum: \n%S
\nFifth spectrum: \n%T \nSixth spectrum: \n%U";
speedmode index:=project sm:=off;
918 break 1;
919 };
920 };
921 //Stackplot of 5 spectra
922 if (exist(col(10)) == 1) {
923 newbook dummy;
924 //Enter information about plotted spectra, measurement
parameter and peakpicking parameter
925 col(a)[1]$="Spektrum 1";
926 col(a)[2]$="Spektrum 2";
927 col(a)[3]$="Spektrum 3";
928 col(a)[4]$="Spektrum 4";
929 col(a)[5]$="Spektrum 5";
930 col(b)[1]$=%P;
931 col(b)[2]$=%Q;
932 col(b)[3]$=%R;
933 col(b)[4]$=%S;
934 col(b)[5]$=%T;
935 wcolwidth irng:=Col(2) width:=100;
936 win -z;
937 getn
938 (First Spectrum:) :@G
939 (Matrix) NoM51$
```


Appendix

LabTalk script (OriginPro 2020b)

```

940 (Sample) NoS51$
941 (Additive) NoA51$
942 (Laser-Intensity) LI51$
943 (-) :@G
944 (Second Spectrum:) :@G
945 (Matrix) NoM52$
946 (Sample) NoS52$
947 (Additive) NoA52$
948 (Laser-Intensity) LI52$
949 (-) :@G
950 (Third Spectrum:) :@G
951 (Matrix) NoM53$
952 (Sample) NoS53$
953 (Additive) NoA53$
954 (Laser-Intensity) LI53$
955 (-) :@G
956 (Fourth Spectrum:) :@G
957 (Matrix) NoM54$
958 (Sample) NoS54$
959 (Additive) NoA54$
960 (Laser-Intensity) LI54$
961 (-) :@G
962 (Fifth Spectrum:) :@G
963 (Matrix) NoM55$
964 (Sample) NoS55$
965 (Additive) NoA55$
966 (Laser-Intensity) LI55$
967 (-) :@G
968 (General Parameters:) :@G
969 (Detection mode) DM5$
970 (Sample preparation) SP5$
971 (-) :@G
972 (Please insert the measurement parameters for plot la-
    beling and the abbreviations used for the data file
    names of the last two spectra.);
973 win -c;
974 MH51=max(col(2));
975 MH52=max(col(4));
976 MH53=max(col(6));
977 MH54=max(col(8));
978 MH55=max(col(10));
979 //Editing of data book
980 win -a %V;
981 wks.col1.lname$ = m/z;
982 wks.col2.lname$ = Absolute Intensity;
983 wks.col2.comment$ = "Matrix: %(NoM51$) ; Analyte:
    %(NoS51$) ; Additive: %(NoA51$)";
984 wks.col3.lname$ = m/z;
985 wks.col4.lname$ = Absolute Intensity;
986 wks.col4.comment$ = "Matrix: %(NoM52$) ; Analyte:
    %(NoS52$) ; Additive: %(NoA52$)";
987 wks.col5.lname$ = m/z;
988 wks.col6.lname$ = Absolute Intensity;
989 wks.col6.comment$ = "Matrix: %(NoM53$) ; Analyte:
    %(NoS53$) ; Additive: %(NoA53$)";
990 wks.col7.lname$ = m/z;
991 wks.col8.lname$ = Absolute Intensity;
992 wks.col8.comment$ = "Matrix: %(NoM54$) ; Analyte:
    %(NoS54$) ; Additive: %(NoA54$)";
993 wks.col9.lname$ = m/z;
994 wks.col10.lname$ = Absolute Intensity;
995 wks.col10.comment$ = "Matrix: %(NoM55$) ; Analyte:
    %(NoS55$) ; Additive: %(NoA55$)";
996 //Stack plotting
997 plotxy iy:=(1,2) plot:=200 color:=color(128,128,0);
998 %W = %H;
999 layadd igp:=%W type:=normal userdef:=1 linkto:=1
    xaxis:=straight yaxis:=straight;
1000 layadd igp:=%W type:=normal userdef:=1 linkto:=1
    xaxis:=straight yaxis:=straight;
1001 layadd igp:=%W type:=normal userdef:=1 linkto:=1
    xaxis:=straight yaxis:=straight;
1002 layadd igp:=%W type:=normal userdef:=1 linkto:=1
    xaxis:=straight yaxis:=straight;
1003 win -a %V;
1004 plotxy iy:=(3,4) plot:=200 color:=color(0,128,0)
    ogl:=[%W]2;
1005 plotxy iy:=(5,6) plot:=200 color:=color(255,0,0)
    ogl:=[%W]3;
1006 plotxy iy:=(7,8) plot:=200 color:=color(0,0,255)
    ogl:=[%W]4;
1007 plotxy iy:=(9,10) plot:=200 color:=color(0,0,0)
    ogl:=[%W]5;
1008 win -a %W;
1009 //Sizing of the layers
1010 layer1.unit=1;
1011 layer1.width=88.2;
1012 layer1.height=17.5;
1013 layer1.left=9.3;
1014 layer1.top=3.6;
1015 layer2.unit=1;
1016 layer2.width=88.2;
1017 layer2.height=17.5;
1018 layer2.left=9.3;
1019 layer2.top=21.1;
1020 layer3.unit=1;
1021 layer3.width=88.2;
1022 layer3.height=17.5;
1023 layer3.left=9.3;
1024 layer3.top=38.6;
1025 layer4.unit=1;
1026 layer4.width=88.2;
1027 layer4.height=17.5;
1028 layer4.left=9.3;
1029 layer4.top=56.1;
1030 layer5.unit=1;
1031 layer5.width=88.2;
1032 layer5.height=17.5;
1033 layer5.left=9.3;
1034 layer5.top=73.6;
1035 //Sizing the labels and change their position
1036 layer1.y1.label.pt=31;
1037 layer1.x2.label.pt=31;
1038 layer2.y1.label.pt=31;
1039 layer3.y1.label.pt=31;
1040 layer4.y1.label.pt=31;
1041 layer5.y1.label.pt=31;
1042 layer5.x1.label.pt=31;
1043 layer1.x2.label.offsetV=-30;
1044 layer1.y1.label.offsetH=20;
1045 layer2.y1.label.offsetH=20;
1046 layer3.y1.label.offsetH=20;
1047 layer4.y1.label.offsetH=20;
1048 layer5.y1.label.offsetH=20;
1049 //Scaling of the x/y-layers
1050 if (SoA == 1) {
1051 layer.x1.from = 0;
1052 layer.x1.inc = 500;
1053 layer.x1.to = 3000;
1054 };
1055 if (SoA == 2) {
1056 layer.x1.from = 1000;
1057 layer.x1.inc = 10000;
1058 layer.x1.to = 100000;
1059 };
1060 if (yesno == 1) {
1061 layer1.y1.from = -10;
1062 layer1.y1.inc = 30;
1063 layer1.y1.to = 140;
1064 };
1065 if (yesno == 0) {
1066 layer1.y1.from = -15000;
1067 layer1.y1.inc = 100000;
1068 layer1.y1.to = 250000;
1069 };
1070 //Design of the layers
1071 layer1.x1.labels subtype=1;
1072 layer1.x1.showaxes=3;

```

Appendix

LabTalk script (OriginPro 2020b)

```
1073 if (yesno == 1) {
1074 layer1.y1.labelssubtype=1;
1075 };
1076 if (yesno == 0) {
1077 layer1.y1.labelssubtype=2;
1078 };
1079 layer1.y1.showaxes=3;
1080 layer1.x1.ticks=1 + 2 + 4 + 8;
1081 layer1.x2.ticks = 2 + 8;
1082 layer1.y1.ticks=1 + 4;
1083 layer1.y1.label.decPlaces=0;
1084 layer1.x.showLabels=2;
1085 if (yesno == 1) {
1086 layer2.y1.labelssubtype=1;
1087 };
1088 if (yesno == 0) {
1089 layer2.y1.labelssubtype=2;
1090 };
1091 layer2.y1.showaxes=3;
1092 layer2.y1.ticks=1 + 4;
1093 layer2.x1.ticks=1 + 2 + 4 + 8;
1094 layer2.y1.label.decPlaces=0;
1095 layer2.x.showLabels=0;
1096 if (yesno == 1) {
1097 layer3.y1.labelssubtype=1;
1098 };
1099 if (yesno == 0) {
1100 layer3.y1.labelssubtype=2;
1101 };
1102 layer3.y1.showaxes=3;
1103 layer3.y1.ticks=1 + 4;
1104 layer3.x1.ticks=1 + 2 + 4 + 8;
1105 layer3.y1.label.decPlaces=0;
1106 layer3.x.showLabels=0;
1107 if (yesno == 1) {
1108 layer4.y1.labelssubtype=1;
1109 };
1110 if (yesno == 0) {
1111 layer4.y1.labelssubtype=2;
1112 };
1113 layer4.y1.showaxes=3;
1114 layer4.y1.ticks=1 + 4;
1115 layer4.x1.ticks=1 + 2 + 4 + 8;
1116 layer4.y1.label.decPlaces=0;
1117 layer4.x.showLabels=0;
1118 if (yesno == 1) {
1119 layer5.y1.labelssubtype=1;
1120 };
1121 if (yesno == 0) {
1122 layer5.y1.labelssubtype=2;
1123 };
1124 layer5.y1.showaxes=3;
1125 layer5.y1.ticks=1 + 4;
1126 layer5.x1.ticks=1 + 4;
1127 layer5.y1.label.decPlaces=0;
1128 //Activation & design of the layer grids
1129 layer1.x.showGrids=1;
1130 layer1.x.grid.majorColor=color(192,192,192);
1131 layer1.x.grid.majorType=1;
1132 layer2.x.showGrids=1;
1133 layer2.x.grid.majorColor=color(192,192,192);
1134 layer2.x.grid.majorType=1;
1135 layer3.x.showGrids=1;
1136 layer3.x.grid.majorColor=color(192,192,192);
1137 layer3.x.grid.majorType=1;
1138 layer4.x.showGrids=1;
1139 layer4.x.grid.majorColor=color(192,192,192);
1140 layer4.x.grid.majorType=1;
1141 layer5.x.showGrids=1;
1142 layer5.x.grid.majorColor=color(192,192,192);
1143 layer5.x.grid.majorType=1;
1144 //FIRST plot editing (layer 1)
1145 %W!page.active = 1;
1146 //Peakpicking and peak plotting (layer 1)

1147 pkfind dir:=p method:=max npts:=10 filter:=h
value:=MH51 ocenter_x:=<new> ocenter_y:=<new>
ocenter:=<none>;
1148 win -a %V;
1149 wks.col11.name$=m/z;
1150 wks.col12.name$=Absolut Intensity;
1151 wks.col12.comment$=Maxima Spectrum 1;
1152 plotxy iy:=(11,12) plot:=201 color:=color(128,128,0)
ogl:=[%W]1! rescale:=0;
1153 win -a %W;
1154 %J = %V_L;
1155 set %J -k 2;
1156 set %J -z 4;
1157 set %J -q 1;
1158 set %J -qm 1;
1159 set %J -qb;
1160 set %J -qc 8;
1161 set %J -qp 4;
1162 set %J -qr 90;
1163 set %J -qy 20;
1164 set %J -qs 14;
1165 themeApply2g theme= "MS";
1166 //Delete axis labels and make plot line thicker (layer 1)
1167 label -r xb;
1168 label -r yl;
1169 %J = %V_B;
1170 set %J -w 1000;
1171 //Replacing and editing of the legend label (layer 1)
1172 label -r legend;
1173 label -s -n leglay1 lb(%(1Y, @LC));
1174 leglay1.fsize = 20;
1175 leglay1.color=color(128,128,0);
1176 leglay1.background=1;
1177 leglay1.y = layer1.y.to - leglay1.dy / 2;
1178 leglay1.x = layer1.x.from + leglay1.dx / 2;
1179 //Creating and editing of Laser-Intensity label (layer 1)
1180 label -s -n lasintlay1 lb(\p80(Laser-Intensity: %(LI51$)));
1181 lasintlay1.rotate= 90;
1182 lasintlay1.y = (layer.y.to - layer.y.from) / 2 - 10000;
1183 lasintlay1.x = layer.x.to + lasintlay1.dx / 2;
1184 //Add scrollbar and edit its appearance (layer 1)
1185 axis_scrollbar axis:=left begin:=%(layer1.y1.from$)
end:=%(layer1.y1.to$);
1186 yscrollbar.color=color(128,128,0);
1187 yscrollbar.fillcolor=color(128,128,0);
1188 yscrollthumb.color=color(128,128,0);
1189 yscrollthumb.fillcolor=color(128,128,0);
1190 yscrollbar.states=0;
1191 yscrollbar.height=850;
1192 yscrollbar.top=182;
1193 yscrollbar.width=125;
1194 yscrollbar.left=6290;
1195 yscrollbar.states=64;
1196 yscrollbar.states=128;
1197 yscrollbar.states=1;
1198 yscrollbar.states=2;
1199 yscrollbar.states=4;
1200 yscrollbar.states=8;
1201 yscrollbar.states=16;
1202 yscrollbar.states=32;
1203 yscrollbar.mouse=0;
1204 //Rescaling of y-axis (layer 1)
1205 if (yesno == 1) {
1206 layer1.y1.from = -10;
1207 layer1.y1.inc = 30;
1208 layer1.y1.to = 140;
1209 };
1210 if (yesno == 0) {
1211 layer1.y1.from = -15000;
1212 layer1.y1.inc = 100000;
1213 layer1.y1.to = 250000;
1214 };
1215 //SECOND plot editing (layer 2)
1216 %W!page.active = 2;
```

Appendix

LabTalk script (OriginPro 2020b)

```
1217 //Peakpicking and peak plotting (layer 2)
1218 pkfind dir:=p method:=max npts:=10 filter:=h
value:=MH52 ocenter_x:=<new> ocenter_y:=<new>
ocenter:=<none>;
1219 win -a %V;
1220 wks.col13.Iname$=m/z;
1221 wks.col14.Iname$=Absolut Intensity;
1222 wks.col14.comment$=Maxima Spectrum 2;
1223 plotxy iy:=(13,14) plot:=201 color:=color(0,128,0)
ogl:=[%W]2! rescale:=0;
1224 win -a %W;
1225 %K = %V_N;
1226 set %K -k 2;
1227 set %K -z 4;
1228 set %K -q 1;
1229 set %K -qm 1;
1230 set %K -qb;
1231 set %K -qc 12;
1232 set %K -qp 4;
1233 set %K -qr 90;
1234 set %K -qy 20;
1235 set %K -qs 14;
1236 themeApply2g theme:= "MS";
1237 //Delete axis labels and make plot line thicker (layer 2)
1238 label -r xb;
1239 label -r yl;
1240 %K = %V_D;
1241 set %K -w 1000;
1242 //Replacing and editing of the legend label (layer 2)
1243 label -r legend;
1244 label -s -n leglay2 \b(%(1Y, @LC));
1245 leglay2.fsize = 20;
1246 leglay2.color=color(0,128,0);
1247 leglay2.background=1;
1248 leglay2.y = layer2.y.to - leglay2.dy / 2;
1249 leglay2.x = layer2.x.from + leglay2.dx / 2;
1250 //Creating and editing of Laser-Intensity label (layer 2)
1251 label -s -n lasintlay2 \b(\p80(Laser-Intensity: %(LI52$)));
1252 lasintlay2.rotate= 90;
1253 lasintlay2.y = (layer.y.to - layer.y.from) / 2 - 10000;
1254 lasintlay2.x = layer.x.to + lasintlay2.dx / 2;
1255 //Add scrollbar and edit its apperance (layer 2)
1256 axis_scrollbar axis:=left begin:=%(layer2.y1.from$)
end:=%(layer2.y1.to$);
1257 yscrollbar.color=color(0,128,0);
1258 yscrollbar.fillcolor=color(0,128,0);
1259 yscrollthumb.color=color(0,128,0);
1260 yscrollthumb.fillcolor=color(0,128,0);
1261 yscrollbar.states=0;
1262 yscrollbar.height=850;
1263 yscrollbar.top=1045;
1264 yscrollbar.width=125;
1265 yscrollbar.left=6290;
1266 yscrollbar.states=64;
1267 yscrollbar.states=128;
1268 yscrollbar.states=1;
1269 yscrollbar.states=2;
1270 yscrollbar.states=4;
1271 yscrollbar.states=8;
1272 yscrollbar.states=16;
1273 yscrollbar.states=32;
1274 yscrollbar.mouse=0;
1275 //Rescaling of y-axis (layer 2)
1276 if (yesno == 1) {
1277 layer1.y1.from = -10;
1278 layer1.y1.inc = 30;
1279 layer1.y1.to = 140;
1280 };
1281 if (yesno == 0) {
1282 layer1.y1.from = -15000;
1283 layer1.y1.inc = 100000;
1284 layer1.y1.to = 250000;
1285 };
1286 //THIRD plot editing (layer 3)
1287 %W!page.active = 3;
1288 //Peakpicking and peak plotting (layer 3)
1289 pkfind dir:=p method:=max npts:=10 filter:=h
value:=MH53 ocenter_x:=<new> ocenter_y:=<new>
ocenter:=<none>;
1290 win -a %V;
1291 wks.col15.Iname$=m/z;
1292 wks.col16.Iname$=Absolut Intensity;
1293 wks.col16.comment$=Maxima Spectrum 3;
1294 plotxy iy:=(15,16) plot:=201 color:=color(255,0,0)
ogl:=[%W]3! rescale:=0;
1295 win -a %W;
1296 %L = %V_P;
1297 set %L -k 2;
1298 set %L -z 4;
1299 set %L -q 1;
1300 set %L -qm 1;
1301 set %L -qb;
1302 set %L -qc 2;
1303 set %L -qp 4;
1304 set %L -qr 90;
1305 set %L -qy 20;
1306 set %L -qs 14;
1307 themeApply2g theme:= "MS";
1308 //Delete axis labels and make plot line thicker (layer 3)
1309 label -r xb;
1310 label -r yl;
1311 %L = %V_F;
1312 set %L -w 1000;
1313 //Replacing and editing of the legend label (layer 3)
1314 label -r legend;
1315 label -s -n leglay3 \b(%(1Y, @LC));
1316 leglay3.fsize = 20;
1317 leglay3.color=color(255,0,0);
1318 leglay3.background=1;
1319 leglay3.y = layer3.y.to - leglay3.dy / 2;
1320 leglay3.x = layer3.x.from + leglay3.dx / 2;
1321 //Creating and editing of Laser-Intensity label (layer 3)
1322 label -s -n lasintlay3 \b(\p80(Laser-Intensity: %(LI53$)));
1323 lasintlay3.rotate= 90;
1324 lasintlay3.y = (layer.y.to - layer.y.from) / 2 - 10000;
1325 lasintlay3.x = layer.x.to + lasintlay3.dx / 2;
1326 //Add scrollbar and edit its apperance (layer 3)
1327 axis_scrollbar axis:=left begin:=%(layer3.y1.from$)
end:=%(layer3.y1.to$);
1328 yscrollbar.color=color(255,0,0);
1329 yscrollbar.fillcolor=color(255,0,0);
1330 yscrollthumb.color=color(255,0,0);
1331 yscrollthumb.fillcolor=color(255,0,0);
1332 yscrollbar.states=0;
1333 yscrollbar.height=850;
1334 yscrollbar.top=1908;
1335 yscrollbar.width=125;
1336 yscrollbar.left=6290;
1337 yscrollbar.states=64;
1338 yscrollbar.states=128;
1339 yscrollbar.states=1;
1340 yscrollbar.states=2;
1341 yscrollbar.states=4;
1342 yscrollbar.states=8;
1343 yscrollbar.states=16;
1344 yscrollbar.states=32;
1345 yscrollbar.mouse=0;
1346 //Rescaling of y-axis (layer 3)
1347 if (yesno == 1) {
1348 layer1.y1.from = -10;
1349 layer1.y1.inc = 30;
1350 layer1.y1.to = 140;
1351 };
1352 if (yesno == 0) {
1353 layer1.y1.from = -15000;
1354 layer1.y1.inc = 100000;
1355 layer1.y1.to = 250000;
1356 };
```

Appendix

LabTalk script (OriginPro 2020b)

```
1357 //FOURTH plot editing (layer 4)
1358 %W|page.active = 4;
1359 //Peakpicking and peak plotting (layer 4)
1360 pkfind dir:=p method:=max npts:=10 filter:=h
    value:=MH54 ocenter_x:=<new> ocenter_y:=<new>
    ocenter:=<none>;
1361 win -a %V;
1362 wks.col17.lname$=m/z;
1363 wks.col18.lname$=Absolut Intensity;
1364 wks.col18.comment$=Maxima Spectrum 4;
1365 plotxy iy:=(17,18) plot:=201 color:=color(0,0,255)
    ogl:=[%W]4! rescale:=0;
1366 win -a %W;
1367 %M = %V_R;
1368 set %M -k 2;
1369 set %M -z 4;
1370 set %M -q 1;
1371 set %M -qm 1;
1372 set %M -qb;
1373 set %M -qc 4;
1374 set %M -qp 4;
1375 set %M -qr 90;
1376 set %M -qy 20;
1377 set %M -qs 14;
1378 themeApply2g theme:= "MS";
1379 //Delete axis labels and make plot line thicker (layer 4)
1380 label -r xb;
1381 label -r yl;
1382 %M = %V_H;
1383 set %M -w 1000;
1384 //Replacing and editing of the legend label (layer 4)
1385 label -r legend;
1386 label -s -n leglay4 lb(%(1Y, @LC));
1387 leglay4.fsize = 20;
1388 leglay4.color=color(0,0,255);
1389 leglay4.background=1;
1390 leglay4.y = layer4.y.to - leglay4.dy / 2;
1391 leglay4.x = layer4.x.from + leglay4.dx / 2;
1392 //Creating and editing of Laser-Intensity label (layer 4)
1393 label -s -n lasintlay4 lb(lp80(Laser-Intensity: %(LI54$)));
1394 lasintlay4.rotate= 90;
1395 lasintlay4.y = (layer.y.to - layer.y.from) / 2 - 10000;
1396 lasintlay4.x = layer.x.to + lasintlay4.dx / 2;
1397 //Add scrollbar and edit its apperance (layer 4)
1398 axis_scrollbar axis:=left begin:=%(layer4.y1.from$)
    end:=%(layer4.y1.to$);
1399 scrollbar.color=color(0,0,255);
1400 scrollbar.fillcolor=color(0,0,255);
1401 scrollbarthumb.color=color(0,0,255);
1402 scrollbarthumb.fillcolor=color(0,0,255);
1403 scrollbar.states=0;
1404 scrollbar.height=850;
1405 scrollbar.top=2771;
1406 scrollbar.width=125;
1407 scrollbar.left=6290;
1408 scrollbar.states=64;
1409 scrollbar.states=128;
1410 scrollbar.states=1;
1411 scrollbar.states=2;
1412 scrollbar.states=4;
1413 scrollbar.states=8;
1414 scrollbar.states=16;
1415 scrollbar.states=32;
1416 scrollbar.mouse=0;
1417 //Rescaling of y-axis (layer 4)
1418 if (yesno == 1) {
1419 layer1.y1.from = -10;
1420 layer1.y1.inc = 30;
1421 layer1.y1.to = 140;
1422 };
1423 if (yesno == 0) {
1424 layer1.y1.from = -15000;
1425 layer1.y1.inc = 100000;
1426 layer1.y1.to = 250000;
1427 };
1428 //FIFTH plot editing (layer 5)
1429 %W|page.active = 5;
1430 //Peakpicking and peak plotting (layer 5)
1431 pkfind dir:=p method:=max npts:=10 filter:=h
    value:=MH55 ocenter_x:=<new> ocenter_y:=<new>
    ocenter:=<none>;
1432 win -a %V;
1433 wks.col19.lname$=m/z;
1434 wks.col20.lname$=Absolut Intensity;
1435 wks.col20.comment$=Maxima Spectrum 5;
1436 plotxy iy:=(19,20) plot:=201 color:=color(0,0,0)
    ogl:=[%W]5! rescale:=0;
1437 win -a %W;
1438 %N = %V_T;
1439 set %N -k 2;
1440 set %N -z 4;
1441 set %N -q 1;
1442 set %N -qm 1;
1443 set %N -qb;
1444 set %N -qc 1;
1445 set %N -qp 4;
1446 set %N -qr 90;
1447 set %N -qy 20;
1448 set %N -qs 14;
1449 themeApply2g theme:= "MS";
1450 //Edit axis labels and make plot line thicker (layer 5)
1451 yl.fsize=39;
1452 xb.fsize=39;
1453 yl.top=1980;
1454 %N = %V_J;
1455 set %N -w 1000;
1456 //Replacing and editing of the legend label (layer 5)
1457 label -r legend;
1458 label -s -n leglay5 lb(%(1Y, @LC));
1459 leglay5.fsize = 20;
1460 leglay5.color=color(0,0,0);
1461 leglay5.background=1;
1462 leglay5.y = layer5.y.to - leglay5.dy / 2;
1463 leglay5.x = layer5.x.from + leglay5.dx / 2;
1464 //Creating and editing of Laser-Intensity label (layer 5)
1465 label -s -n lasintlay5 lb(lp80(Laser-Intensity: %(LI55$)));
1466 lasintlay5.rotate= 90;
1467 lasintlay5.y = (layer.y.to - layer.y.from) / 2 - 10000;
1468 lasintlay5.x = layer.x.to + lasintlay5.dx / 2;
1469 //Add scrollbar and edit its apperance (layer 5)
1470 axis_scrollbar axis:=left begin:=%(layer5.y1.from$)
    end:=%(layer5.y1.to$);
1471 scrollbar.color=color(0,0,0);
1472 scrollbar.fillcolor=color(0,0,0);
1473 scrollbarthumb.color=color(0,0,0);
1474 scrollbarthumb.fillcolor=color(0,0,0);
1475 scrollbar.states=0;
1476 scrollbar.height=850;
1477 scrollbar.top=3634;
1478 scrollbar.width=125;
1479 scrollbar.left=6290;
1480 scrollbar.states=64;
1481 scrollbar.states=128;
1482 scrollbar.states=1;
1483 scrollbar.states=2;
1484 scrollbar.states=4;
1485 scrollbar.states=8;
1486 scrollbar.states=16;
1487 scrollbar.states=32;
1488 scrollbar.mouse=0;
1489 //Rescaling of y-axis (layer 5)
1490 if (yesno == 1) {
1491 layer1.y1.from = -10;
1492 layer1.y1.inc = 30;
1493 layer1.y1.to = 140;
1494 };
1495 if (yesno == 0) {
1496 layer1.y1.from = -15000;
```


Appendix

LabTalk script (OriginPro 2020b)

```
1497 layer1.y1.inc = 100000;
1498 layer1.y1.to = 250000;
1499 };
1500 //Add x-scroller and edit its appearance
1501 axis_scroller xsize:=0.3 ysize:=-1.5;
1502 XMin.color=color(0,0,0);
1503 XMin.fillcolor=color(0,0,0);
1504 XMax.color=color(0,0,0);
1505 XMax.fillcolor=color(0,0,0);
1506 //Adding measurement parameter labels
1507 label -s -d 4130 4900 -n detectionmode \p80(Detection-
mode: %(DM5$));
1508 label -s -d 10 4900 -n sampleprep \p80(Sample-prepa-
ration: %(SP5$));
1509 win -a %V;
1510 win -ia;
1511 win -a %W;
1512 win -z;
1513 type -b "First spectrum: \n%P \nSecond spectrum:
\n%Q \nThird spectrum: \n%R \nFourth spectrum: \n%S
\nFifth spectrum: \n%T";
1514 speedmode index:=project sm:=off;
1515 break 1;
1516 };
1517 //Stackplot of 4 spectra
1518 if (exist(col(8))) == 1 {
1519 newbook dummy;
1520 //Enter information about plotted spectra, measurement
parameter and peakpicking parameter
1521 col(a)[1]$="Spektrum 1";
1522 col(a)[2]$="Spektrum 2";
1523 col(a)[3]$="Spektrum 3";
1524 col(a)[4]$="Spektrum 4";
1525 col(b)[1]$=%P;
1526 col(b)[2]$=%Q;
1527 col(b)[3]$=%R;
1528 col(b)[4]$=%S;
1529 wcolwidth irng:=Col(2) width:=100;
1530 win -z;
1531 getn
1532 (First Spectrum:) :@G
1533 (Matrix) NoM41$
1534 (Sample) NoS41$
1535 (Additive) NoA41$
1536 (Laser-Intensity) LI41$
1537 (-) :@G
1538 (Second Spectrum:) :@G
1539 (Matrix) NoM42$
1540 (Sample) NoS42$
1541 (Additive) NoA42$
1542 (Laser-Intensity) LI42$
1543 (-) :@G
1544 (Third Spectrum:) :@G
1545 (Matrix) NoM43$
1546 (Sample) NoS43$
1547 (Additive) NoA43$
1548 (Laser-Intensity) LI43$
1549 (-) :@G
1550 (Fourth Spectrum:) :@G
1551 (Matrix) NoM44$
1552 (Sample) NoS44$
1553 (Additive) NoA44$
1554 (Laser-Intensity) LI44$
1555 (-) :@G
1556 (General Parameters:) :@G
1557 (Detection mode) DM4$
1558 (Sample preparation) SP4$
1559 (-) :@G
1560 (Please insert the measurement parameters for plot la-
beling and the abbreviations used for the data file
names of the last two spectra.);
1561 win -c;
1562 MH41=max(col(2));
1563 MH42=max(col(4));
1564 MH43=max(col(6));
1565 MH44=max(col(8));
1566 //Editoring of data book
1567 win -a %V;
1568 wks.col1.lname$ = m/z;
1569 wks.col2.lname$ = Absolute Intensity;
1570 wks.col2.comment$ = "Matrix: %(NoM41$) ; Analyte:
%(NoS41$) ; Additive: %(NoA41$)";
1571 wks.col3.lname$ = m/z;
1572 wks.col4.lname$ = Absolute Intensity;
1573 wks.col4.comment$ = "Matrix: %(NoM42$) ; Analyte:
%(NoS42$) ; Additive: %(NoA42$)";
1574 wks.col5.lname$ = m/z;
1575 wks.col6.lname$ = Absolute Intensity;
1576 wks.col6.comment$ = "Matrix: %(NoM43$) ; Analyte:
%(NoS43$) ; Additive: %(NoA43$)";
1577 wks.col7.lname$ = m/z;
1578 wks.col8.lname$ = Absolute Intensity;
1579 wks.col8.comment$ = "Matrix: %(NoM44$) ; Analyte:
%(NoS44$) ; Additive: %(NoA44$)";
1580 //Stack plotting
1581 plotxy iy:=(1,2) plot:=200 color:=color(0,128,0);
1582 %W = %H;
1583 layadd igp:=%W type:=normal userdef:=1 linkto:=1
xaxis:=straight yaxis:=straight;
1584 layadd igp:=%W type:=normal userdef:=1 linkto:=1
xaxis:=straight yaxis:=straight;
1585 layadd igp:=%W type:=normal userdef:=1 linkto:=1
xaxis:=straight yaxis:=straight;
1586 win -a %V;
1587 plotxy iy:=(3,4) plot:=200 color:=color(255,0,0)
ogl:=[%W]2!;
1588 plotxy iy:=(5,6) plot:=200 color:=color(0,0,255)
ogl:=[%W]3!;
1589 plotxy iy:=(7,8) plot:=200 color:=color(0,0,0)
ogl:=[%W]4!;
1590 win -a %W;
1591 //Sizing of the layers
1592 layer1.unit=1;
1593 layer1.width=88.2;
1594 layer1.height=21.9;
1595 layer1.left=9.3;
1596 layer1.top=3.6;
1597 layer2.unit=1;
1598 layer2.width=88.2;
1599 layer2.height=21.9;
1600 layer2.left=9.3;
1601 layer2.top=25.5;
1602 layer3.unit=1;
1603 layer3.width=88.2;
1604 layer3.height=21.9;
1605 layer3.left=9.3;
1606 layer3.top=47.4;
1607 layer4.unit=1;
1608 layer4.width=88.2;
1609 layer4.height=21.9;
1610 layer4.left=9.3;
1611 layer4.top=69.3;
1612 //Sizing the labels and change their position
1613 layer1.y1.label.pt=28;
1614 layer1.x2.label.pt=28;
1615 layer2.y1.label.pt=28;
1616 layer3.y1.label.pt=28;
1617 layer4.y1.label.pt=28;
1618 layer4.x1.label.pt=28;
1619 layer1.x2.label.offsetV=-30;
1620 layer1.y1.label.offsetH=20;
1621 layer2.y1.label.offsetH=20;
1622 layer3.y1.label.offsetH=20;
1623 layer4.y1.label.offsetH=20;
1624 //Scaling of the x/y-layers
1625 if (SoA == 1) {
1626 layer.x1.from = 0;
1627 layer.x1.inc = 500;
```

Appendix

LabTalk script (OriginPro 2020b)

```
1628 layer.x1.to = 3000;
1629 };
1630 if (SoA == 2) {
1631 layer.x1.from = 1000;
1632 layer.x1.inc = 10000;
1633 layer.x1.to = 100000;
1634 };
1635 if (yesno == 1) {
1636 layer1.y1.from = -8;
1637 layer1.y1.inc = 30;
1638 layer1.y1.to = 140;
1639 };
1640 if (yesno == 0) {
1641 layer1.y1.from = -15000;
1642 layer1.y1.inc = 100000;
1643 layer1.y1.to = 250000;
1644 };
1645 //Design of the layer
1646 layer1.x1.labelssubtype=1;
1647 layer1.x1.showaxes=3;
1648 if (yesno == 1) {
1649 layer1.y1.labelssubtype=1;
1650 };
1651 if (yesno == 0) {
1652 layer1.y1.labelssubtype=2;
1653 };
1654 layer1.y1.showaxes=3;
1655 layer1.x1.ticks=1 + 2 + 4 + 8;
1656 layer1.x2.ticks = 2 + 8;
1657 layer1.y1.ticks=1 + 4;
1658 layer1.y1.label.decPlaces=0;
1659 layer1.x.showLabels=2;
1660 if (yesno == 1) {
1661 layer2.y1.labelssubtype=1;
1662 };
1663 if (yesno == 0) {
1664 layer2.y1.labelssubtype=2;
1665 };
1666 layer2.y1.showaxes=3;
1667 layer2.y1.ticks=1 + 4;
1668 layer2.x1.ticks=1 + 2 + 4 + 8;
1669 layer2.y1.label.decPlaces=0;
1670 layer2.x.showLabels=0;
1671 if (yesno == 1) {
1672 layer3.y1.labelssubtype=1;
1673 };
1674 if (yesno == 0) {
1675 layer3.y1.labelssubtype=2;
1676 };
1677 layer3.y1.showaxes=3;
1678 layer3.y1.ticks=1 + 4;
1679 layer3.x1.ticks=1 + 2 + 4 + 8;
1680 layer3.y1.label.decPlaces=0;
1681 layer3.x.showLabels=0;
1682 if (yesno == 1) {
1683 layer4.y1.labelssubtype=1;
1684 };
1685 if (yesno == 0) {
1686 layer4.y1.labelssubtype=2;
1687 };
1688 layer4.y1.showaxes=3;
1689 layer4.y1.ticks=1 + 4;
1690 layer4.x1.ticks=1 + 4;
1691 layer4.y1.label.decPlaces=0;
1692 //Activation & design of the layer grids
1693 layer1.x.showGrids=1;
1694 layer1.x.grid.majorColor=color(192,192,192);
1695 layer1.x.grid.majorType=1;
1696 layer2.x.showGrids=1;
1697 layer2.x.grid.majorColor=color(192,192,192);
1698 layer2.x.grid.majorType=1;
1699 layer3.x.showGrids=1;
1700 layer3.x.grid.majorColor=color(192,192,192);
1701 layer3.x.grid.majorType=1;
1702 layer4.x.showGrids=1;
1703 layer4.x.grid.majorColor=color(192,192,192);
1704 layer4.x.grid.majorType=1;
1705 //FIRST plot editing (layer 1)
1706 %W!page.active = 1;
1707 //Peakpicking and peak plotting (layer 1)
1708 pkfind dir:=p method:=max npts:=10 filter:=h
value:=MH41 ocenter_x:=<new> ocenter_y:=<new>
ocenter:=<none>;
1709 win -a %V;
1710 wks.col9.Iname$=m/z;
1711 wks.col10.Iname$=Absolut Intensity;
1712 wks.col10.comment$=Maxima Spectrum 1;
1713 plotxy iy:=(9,10) plot:=201 color:=color(0,128,0)
ogl:=[%W]1! rescale:=0;
1714 win -a %W;
1715 %J = %V_J;
1716 set %J -k 2;
1717 set %J -z 4;
1718 set %J -q 1;
1719 set %J -qm 1;
1720 set %J -qb;
1721 set %J -qc 12;
1722 set %J -qp 4;
1723 set %J -qr 90;
1724 set %J -qy 20;
1725 set %J -qs 14;
1726 themeApply2g theme:= "MS";
1727 //Delete axis labels and make plot line thicker (layer 1)
1728 label -r xb;
1729 label -r yl;
1730 %J = %V_B;
1731 set %J -w 1000;
1732 //Replacing and editing of the legend label (layer 1)
1733 label -r legend;
1734 label -s -n leglay1 lb(%(1Y, @LC));
1735 leglay1.fsize = 20;
1736 leglay1.color=color(0,128,0);
1737 leglay1.background=1;
1738 leglay1.y = layer1.y.to - leglay1.dy / 2;
1739 leglay1.x = layer1.x.from + leglay1.dx / 2;
1740 //Creating and editing of Laser-Intensity label (layer 1)
1741 label -s -n lasintlay1 lb(\p80(Laser-Intensity: %(LI41$)));
1742 lasintlay1.rotate= 90;
1743 lasintlay1.y = (layer.y.to - layer.y.from) / 2;
1744 lasintlay1.x = layer.x.to + lasintlay1.dx / 2;
1745 //Add scrollbar and edit its apperance (layer 1)
1746 axis_scrollbar axis:=left begin:=%(layer1.y1.from$)
end:=%(layer1.y1.to$);
1747 yscrollbar.color=color(0,128,0);
1748 yscrollbar.fillcolor=color(0,128,0);
1749 yscrollthumb.color=color(0,128,0);
1750 yscrollthumb.fillcolor=color(0,128,0);
1751 yscrollbar.states=0;
1752 yscrollbar.height=1070;
1753 yscrollbar.top=180;
1754 yscrollbar.width=125;
1755 yscrollbar.left=6290;
1756 yscrollbar.states=64;
1757 yscrollbar.states=128;
1758 yscrollbar.states=1;
1759 yscrollbar.states=2;
1760 yscrollbar.states=4;
1761 yscrollbar.states=8;
1762 yscrollbar.states=16;
1763 yscrollbar.states=32;
1764 yscrollbar.mouse=0;
1765 //Rescaling of y-axis (layer 1)
1766 if (yesno == 1) {
1767 layer1.y1.from = -8;
1768 layer1.y1.inc = 30;
1769 layer1.y1.to = 140;
1770 };
1771 if (yesno == 0) {
```


Appendix

LabTalk script (OriginPro 2020b)

```
1772 layer1.y1.from = -15000;
1773 layer1.y1.inc = 100000;
1774 layer1.y1.to = 250000;
1775 };
1776 //SECOND plot editing (layer 2)
1777 %W!page.active = 2;
1778 //Peakpicking and peak plotting (layer 2)
1779 pkfind dir:=p method:=max npts:=10 filter:=h
value:=MH42 ocenter_x:=<new> ocenter_y:=<new>
ocenter:=<none>;
1780 win -a %V;
1781 wks.col11.lname$=m/z;
1782 wks.col12.lname$=Absolut Intensity;
1783 wks.col12.comment$=Maxima Spectrum 2;
1784 plotxy iy:=(11,12) plot:=201 color:=color(255,0,0)
ogl:=[%W]2! rescale:=0;
1785 win -a %W;
1786 %K = %V_L;
1787 set %K -k 2;
1788 set %K -z 4;
1789 set %K -q 1;
1790 set %K -qm 1;
1791 set %K -qb;
1792 set %K -qc 2;
1793 set %K -qp 4;
1794 set %K -qr 90;
1795 set %K -qy 20;
1796 set %K -qs 14;
1797 themeApply2g theme:= "MS";
1798 //Delete axis labels and make plot line thicker (layer 2)
1799 label -r xb;
1800 label -r yl;
1801 %K = %V_D;
1802 set %K -w 1000;
1803 //Replacing and editing of the legend label (layer 2)
1804 label -r legend;
1805 label -s -n leglay2 \b(%(1Y, @LC));
1806 leglay2.fsize = 20;
1807 leglay2.color=color(255,0,0);
1808 leglay2.background=1;
1809 leglay2.y = layer2.y.to - leglay2.dy / 2;
1810 leglay2.x = layer2.x.from + leglay2.dx / 2;
1811 //Creating and editing of Laser-Intensity label (layer 2)
1812 label -s -n lasintlay2 \b\p80(Laser-Intensity: %(LI42$));
1813 lasintlay2.rotate = 90;
1814 lasintlay2.y = (layer.y.to - layer.y.from) / 2;
1815 lasintlay2.x = layer.x.to + lasintlay2.dx / 2;
1816 //Add scrollbar and edit its apperance (layer 2)
1817 axis_scrollbar axis:=left begin:=(layer2.y1.from$)
end:=(layer2.y1.to$);
1818 yscrollbar.color=color(255,0,0);
1819 yscrollbar.fillcolor=color(255,0,0);
1820 yscrollthumb.color=color(255,0,0);
1821 yscrollthumb.fillcolor=color(255,0,0);
1822 yscrollbar.states=0;
1823 yscrollbar.height=1070;
1824 yscrollbar.top=1260;
1825 yscrollbar.width=125;
1826 yscrollbar.left=6290;
1827 yscrollbar.states=64;
1828 yscrollbar.states=128;
1829 yscrollbar.states=1;
1830 yscrollbar.states=2;
1831 yscrollbar.states=4;
1832 yscrollbar.states=8;
1833 yscrollbar.states=16;
1834 yscrollbar.states=32;
1835 yscrollbar.mouse=0;
1836 //Rescaling of y-axis (layer 2)
1837 if (yesno == 1) {
1838 layer1.y1.from = -8;
1839 layer1.y1.inc = 30;
1840 layer1.y1.to = 140;
1841 };
1842 if (yesno == 0) {
1843 layer1.y1.from = -15000;
1844 layer1.y1.inc = 100000;
1845 layer1.y1.to = 250000;
1846 };
1847 //THIRD plot editing (layer 3)
1848 %W!page.active = 3;
1849 //Peakpicking and peak plotting (layer 3)
1850 pkfind dir:=p method:=max npts:=10 filter:=h
value:=(MH43/2) ocenter_x:=<new> ocenter_y:=<new>
ocenter:=<none>;
1851 win -a %V;
1852 wks.col13.lname$=m/z;
1853 wks.col14.lname$=Absolut Intensity;
1854 wks.col14.comment$=Maxima Spectrum 3;
1855 plotxy iy:=(13,14) plot:=201 color:=color(0,0,255)
ogl:=[%W]3! rescale:=0;
1856 win -a %W;
1857 %L = %V_N;
1858 set %L -k 2;
1859 set %L -z 4;
1860 set %L -q 1;
1861 set %L -qm 1;
1862 set %L -qb;
1863 set %L -qc 4;
1864 set %L -qp 4;
1865 set %L -qr 90;
1866 set %L -qy 20;
1867 set %L -qs 14;
1868 themeApply2g theme:= "MS";
1869 //Delete axis labels and make plot line thicker (layer 3)
1870 label -r xb;
1871 label -r yl;
1872 %L = %V_F;
1873 set %L -w 1000;
1874 //Replacing and editing of the legend label (layer 3)
1875 label -r legend;
1876 label -s -n leglay3 \b(%(1Y, @LC));
1877 leglay3.fsize = 20;
1878 leglay3.color=color(0,0,255);
1879 leglay3.background=1;
1880 leglay3.y = layer3.y.to - leglay3.dy / 2;
1881 leglay3.x = layer3.x.from + leglay3.dx / 2;
1882 //Creating and editing of Laser-Intensity label (layer 3)
1883 label -s -n lasintlay3 \b\p80(Laser-Intensity: %(LI43$));
1884 lasintlay3.rotate = 90;
1885 lasintlay3.y = (layer.y.to - layer.y.from) / 2;
1886 lasintlay3.x = layer.x.to + lasintlay3.dx / 2;
1887 //Add scrollbar and edit its apperance (layer 3)
1888 axis_scrollbar axis:=left begin:=(layer3.y1.from$)
end:=(layer3.y1.to$);
1889 yscrollbar.color=color(0,0,255);
1890 yscrollbar.fillcolor=color(0,0,255);
1891 yscrollthumb.color=color(0,0,255);
1892 yscrollthumb.fillcolor=color(0,0,255);
1893 yscrollbar.states=0;
1894 yscrollbar.height=1070;
1895 yscrollbar.top=2340;
1896 yscrollbar.width=125;
1897 yscrollbar.left=6290;
1898 yscrollbar.states=64;
1899 yscrollbar.states=128;
1900 yscrollbar.states=1;
1901 yscrollbar.states=2;
1902 yscrollbar.states=4;
1903 yscrollbar.states=8;
1904 yscrollbar.states=16;
1905 yscrollbar.states=32;
1906 yscrollbar.mouse=0;
1907 //Rescaling of y-axis (layer 3)
1908 if (yesno == 1) {
1909 layer1.y1.from = -8;
1910 layer1.y1.inc = 30;
1911 layer1.y1.to = 140;
```

Appendix

LabTalk script (OriginPro 2020b)

```
1912 };
1913 if (yesno == 0) {
1914   layer1.y1.from = -15000;
1915   layer1.y1.inc = 100000;
1916   layer1.y1.to = 250000;
1917 };
1918 //FOURTH plot editing (layer 4)
1919 %W!page.active = 4;
1920 pkfind dir:=p method:=max npts:=10 filter:=h
    value:=MH44 ocenter_x:=<new> ocenter_y:=<new>
    ocenter:=<none>;
1921 win -a %V;
1922 wks.col15.lname$=m/z;
1923 wks.col16.lname$=Absolut Intensity;
1924 wks.col16.comment$=Maxima Spectrum 4;
1925 plotxy iy:=(15,16) plot:=201 color:=color(0,0,0)
    ogl:=[%W]4! rescale:=0;
1926 win -a %W;
1927 %M = %V_P;
1928 set %M -k 2;
1929 set %M -z 4;
1930 set %M -q 1;
1931 set %M -qm 1;
1932 set %M -qb;
1933 set %M -qc 1;
1934 set %M -qp 4;
1935 set %M -qr 90;
1936 set %M -qy 20;
1937 set %M -qs 14;
1938 themeApply2g theme:= "MS";
1939 //Edit axis labels and make plot line thicker (layer 4)
1940 yl.fsize=39;
1941 xb.fsize=39;
1942 yl.top=1980;
1943 %M = %V_H;
1944 set %M -w 1000;
1945 //Replacing and editing of the legend label (layer 4)
1946 label -r legend;
1947 label -s -n leglay4 \b(%(1Y, @LC));
1948 leglay4.fsize = 20;
1949 leglay4.color=color(0,0,0);
1950 leglay4.background=1;
1951 leglay4.y = layer4.y.to - leglay4.dy / 2;
1952 leglay4.x = layer4.x.from + leglay4.dx / 2;
1953 //Creating and editing of Laser-Intensity label (layer 4)
1954 label -s -n lasintlay4 \b\p80(Laser-Intensity: %(LI44$));
1955 lasintlay4.rotate= 90;
1956 lasintlay4.y = (layer.y.to - layer.y.from) / 2;
1957 lasintlay4.x = layer.x.to + lasintlay4.dx / 2;
1958 //Add scrollbar and edit its appearance (layer 4)
1959 axis_scrollback axis:=left begin:=%(layer4.y1.from$)
    end:=%(layer4.y1.to$);
1960 yscrollbar.color=color(0,0,0);
1961 yscrollbar.fillcolor=color(0,0,0);
1962 yscrollthumb.color=color(0,0,0);
1963 yscrollthumb.fillcolor=color(0,0,0);
1964 yscrollbar.states=0;
1965 yscrollbar.height=1070;
1966 yscrollbar.top=3420;
1967 yscrollbar.width=125;
1968 yscrollbar.left=6290;
1969 yscrollbar.states=64;
1970 yscrollbar.states=128;
1971 yscrollbar.states=1;
1972 yscrollbar.states=2;
1973 yscrollbar.states=4;
1974 yscrollbar.states=8;
1975 yscrollbar.states=16;
1976 yscrollbar.states=32;
1977 yscrollbar.mouse=0;
1978 //Rescaling of y-axis (layer 4)
1979 if (yesno == 1) {
1980   layer1.y1.from = -8;
1981   layer1.y1.inc = 30;
1982   layer1.y1.to = 140;
1983 };
1984 if (yesno == 0) {
1985   layer1.y1.from = -15000;
1986   layer1.y1.inc = 100000;
1987   layer1.y1.to = 250000;
1988 };
1989 //Add x-scroller and edit its appearance
1990 axis_scroller xsize:=0.3 ysize:=-1.5;
1991 XMin.color=color(0,0,0);
1992 XMin.fillcolor=color(0,0,0);
1993 XMax.color=color(0,0,0);
1994 XMax.fillcolor=color(0,0,0);
1995 //Adding measurement parameter labels
1996 label -s -d 4130 4900 -n detectionmode \p80(Detection-
    mode: %(DM4$));
1997 label -s -d 10 4900 -n sampleprep \p80(Sample-prepa-
    ration: %(SP4$));
1998 win -a %V;
1999 win -ia;
2000 win -a %W;
2001 win -z;
2002 type -b "Frist spectrum: \n%P \nSecond spectrum:
    \n%Q \nThird spectrum: \n%R \nFourth spectrum:
    \n%S";
2003 speedmode index:=project sm:=off;
2004 break 1;
2005 };
2006 //Stackplot of 3 spectra
2007 if (exist(col(6)) == 1) {
2008   newbook dummy;
2009   //Enter information about plotted spectra, measurement
    parameter and peakpicking parameter
2010   col(a)[1]$="Spektrum 1";
2011   col(a)[2]$="Spektrum 2";
2012   col(a)[3]$="Spektrum 3";
2013   col(b)[1]$=%P;
2014   col(b)[2]$=%Q;
2015   col(b)[3]$=%R;
2016   wcolwidth irng:=Col(2) width:=100;
2017   win -z;
2018   getn
2019   (TOP Spectrum:) :@G
2020   (Matrix) NoM31$
2021   (Sample) NoS31$
2022   (Additive) NoA31$
2023   (Laser-Intensity) LI31$
2024   (-) :@G
2025   (MIDDLE Spectrum:) :@G
2026   (Matrix) NoM32$
2027   (Sample) NoS32$
2028   (Additive) NoA32$
2029   (Laser-Intensity) LI32$
2030   (-) :@G
2031   (BOTTOM Spectrum:) :@G
2032   (Matrix) NoM33$
2033   (Sample) NoS33$
2034   (Additive) NoA33$
2035   (Laser-Intensity) LI33$
2036   (-) :@G
2037   (General Parameters:) :@G
2038   (Detection mode) DM3$
2039   (Sample preparation) SP3$
2040   (-) :@G
2041   (Please insert the measurement parameters for plot la-
    beling and the abbreviations used for the data file
    names of the individual spectra.);
2042   win -c;
2043   MH31=max(col(2));
2044   MH32=max(col(4));
2045   MH33=max(col(6));
2046   //Editing of data book
2047   win -a %V;
2048   wks.col1.lname$ = m/z;
```

Appendix

LabTalk script (OriginPro 2020b)

```
2049 wks.col2.lname$ = Absolute Intensity;
2050 wks.col2.comment$ = "Matrix: %(NoM31$) ; Analyte:
      %(NoS31$) ; Additive: %(NoA31$)";
2051 wks.col3.lname$ = m/z;
2052 wks.col4.lname$ = Absolute Intensity;
2053 wks.col4.comment$ = "Matrix: %(NoM32$) ; Analyte:
      %(NoS32$) ; Additive: %(NoA32$)";
2054 wks.col5.lname$ = m/z;
2055 wks.col6.lname$ = Absolute Intensity;
2056 wks.col6.comment$ = "Matrix: %(NoM33$) ; Analyte:
      %(NoS33$) ; Additive: %(NoA33$)";
2057 //Stack plotting
2058 plotxy iy:=(1,2) plot:=200 color:=color(255,0,0);
2059 %W = %H;
2060 layadd igp:=%W type:=normal userdef:=1 linkto:=1
      xaxis:=straight yaxis:=straight;
2061 layadd igp:=%W type:=normal userdef:=1 linkto:=1
      xaxis:=straight yaxis:=straight;
2062 win -a %V;
2063 plotxy iy:=(3,4) plot:=200 color:=color(0,0,255)
      ogl:=[%W]2!;
2064 plotxy iy:=(5,6) plot:=200 color:=color(0,0,0)
      ogl:=[%W]3!;
2065 win -a %W;
2066 //Sizing of the layers
2067 layer1.unit=1;
2068 layer1.width=88;
2069 layer1.height=29.1;
2070 layer1.left=9.5;
2071 layer1.top=3.6;
2072 layer2.unit=1;
2073 layer2.width=88;
2074 layer2.height=29.1;
2075 layer2.left=9.5;
2076 layer2.top=32.7;
2077 layer3.unit=1;
2078 layer3.width=88;
2079 layer3.height=29.1;
2080 layer3.left=9.5;
2081 layer3.top=61.8;
2082 //Sizing the labels and change their position
2083 layer1.y1.label.pt=24;
2084 layer1.x2.label.pt=24;
2085 layer2.y1.label.pt=24;
2086 layer3.y1.label.pt=24;
2087 layer3.x1.label.pt=24;
2088 layer1.x2.label.offsetV=-30;
2089 layer1.y1.label.offsetH=20;
2090 layer2.y1.label.offsetH=20;
2091 layer3.y1.label.offsetH=20;
2092 //Scaling of the x/y-axis
2093 if (SoA == 1) {
2094   layer.x1.from = 0;
2095   layer.x1.inc = 500;
2096   layer.x1.to = 3000;
2097 };
2098 if (SoA == 2) {
2099   layer.x1.from = 1000;
2100   layer.x1.inc = 10000;
2101   layer.x1.to = 100000;
2102 };
2103 if (yesno == 1) {
2104   layer1.y1.from = -7;
2105   layer1.y1.inc = 30;
2106   layer1.y1.to = 140;
2107 };
2108 if (yesno == 0) {
2109   layer1.y1.from = -15000;
2110   layer1.y1.inc = 100000;
2111   layer1.y1.to = 270000;
2112 };
2113 //Design of the layer
2114 layer1.x1.labels subtype=1;
2115 layer1.x1.showaxes=3;
2116 if (yesno == 1) {
2117   layer1.y1.labels subtype=1;
2118 };
2119 if (yesno == 0) {
2120   layer1.y1.labels subtype=2;
2121 };
2122 layer1.y1.showaxes=3;
2123 layer1.x1.ticks=1 + 2 + 4 + 8;
2124 layer1.x2.ticks=2 + 8;
2125 layer1.y1.ticks=1 + 4;
2126 layer1.x.showLabels=2;
2127 layer1.y1.label.decPlaces=0;
2128 if (yesno == 1) {
2129   layer2.y1.labels subtype=1;
2130 };
2131 if (yesno == 0) {
2132   layer2.y1.labels subtype=2;
2133 };
2134 layer2.y1.showaxes=3;
2135 layer2.y1.ticks=1 + 4;
2136 layer2.x1.ticks=1 + 2 + 4 + 8;
2137 layer2.x.showLabels=0;
2138 layer2.y1.label.decPlaces=0;
2139 if (yesno == 1) {
2140   layer3.y1.labels subtype=1;
2141 };
2142 if (yesno == 0) {
2143   layer3.y1.labels subtype=2;
2144 };
2145 layer3.y1.showaxes=3;
2146 layer3.y1.ticks=1 + 4;
2147 layer3.x1.ticks=1 + 4;
2148 layer3.y1.label.decPlaces=0;
2149 //Activation & design of the layer grids
2150 layer1.x.showGrids=1;
2151 layer1.x.grid.majorColor=color(192,192,192);
2152 layer1.x.grid.majorType=1;
2153 layer2.x.showGrids=1;
2154 layer2.x.grid.majorColor=color(192,192,192);
2155 layer2.x.grid.majorType=1;
2156 layer3.x.showGrids=1;
2157 layer3.x.grid.majorColor=color(192,192,192);
2158 layer3.x.grid.majorType=1;
2159 //FIRST plot editing (layer 1)
2160 %W!page.active = 1;
2161 //Peakpicking and peak plotting (layer 1)
2162 pkfind dir:=p method:=max npts:=10 filter:=h
      value:=MH31 ocenter_x:=<new> ocenter_y:=<new>
      ocenter:=<none>;
2163 win -a %V;
2164 wks.col7.lname$=m/z;
2165 wks.col8.lname$=Absolut Intensity;
2166 wks.col8.comment$=Maxima Spectrum 1;
2167 plotxy iy:=(7,8) plot:=201 color:=color(255,0,0)
      ogl:=[%W]1! rescale:=0;
2168 win -a %W;
2169 %J = %V_H;
2170 %J -k 2;
2171 %J z 4;
2172 set %J -q 1;
2173 set %J -qm 1;
2174 set %J -qb;
2175 set %J -qc 2;
2176 set %J -qp 4;
2177 set %J -qr 90;
2178 set %J -qy 20;
2179 set %J -qs 14;
2180 themeApply2g theme="MS";
2181 //Delete axis labels and make plot line thicker (layer 1)
2182 label -r xb;
2183 label -r yl;
2184 %J = %V_B;
2185 set %J -w 1000;
2186 //Replacing and editing of the legend label (layer 1)
```

Appendix

LabTalk script (OriginPro 2020b)

```
2187 label -r legend;
2188 label -s -n leglay1 \b(%(1Y, @LC));
2189 leglay1.fsize = 20;
2190 leglay1.color=color(255,0,0);
2191 leglay1.background=1;
2192 leglay1.y = layer1.y.to - leglay1.dy / 2;
2193 leglay1.x = layer1.x.from + leglay1.dx / 2;
2194 //Creating and editing of Laser-Intensity label (layer 1)
2195 label -s -n lasintlay1 \p100(Laser-Intensity: %(LI31$));
2196 lasintlay1.rotate= 90;
2197 lasintlay1.y = (layer.y.to - layer.y.from) / 2;
2198 lasintlay1.x = layer.x.to + lasintlay1.dx / 2;
2199 //Add scrollbar and edit its apperance (layer 1)
2200 axis_scrollbar axis:=left begin:=%(layer1.y1.from$)
end:=%(layer1.y1.to$);
2201 yscrollbar.color=color(255,0,0);
2202 yscrollbar.fillcolor=color(255,0,0);
2203 yscrollthumb.color=color(255,0,0);
2204 yscrollthumb.fillcolor=color(255,0,0);
2205 yscrollbar.states=0;
2206 yscrollbar.height=1430;
2207 yscrollbar.top=175;
2208 yscrollbar.width=120;
2209 yscrollbar.left=6290;
2210 yscrollbar.states=64;
2211 yscrollbar.states=128;
2212 yscrollbar.states=1;
2213 yscrollbar.states=2;
2214 yscrollbar.states=4;
2215 yscrollbar.states=8;
2216 yscrollbar.states=16;
2217 yscrollbar.states=32;
2218 yscrollbar.mouse=0;
2219 //Rescaling of y-axis (layer 1)
2220 if (yesno == 1) {
2221 layer1.y1.from = -7;
2222 layer1.y1.inc = 30;
2223 layer1.y1.to = 140;
2224 };
2225 if (yesno == 0) {
2226 layer1.y1.from = -15000;
2227 layer1.y1.inc = 100000;
2228 layer1.y1.to = 270000;
2229 };
2230 //SECOND plot editing (layer 2)
2231 %W|page.active = 2;
2232 //Peakpicking and peak plotting (layer 2)
2233 pkfind dir:=p method:=max npts:=10 filter:=h
value:=MH32 ocenter_x:=<new> ocenter_y:=<new>
ocenter:=<none>;
2234 win -a %V;
2235 wks.col9.Iname$=m/z;
2236 wks.col10.Iname$=Absolut Intensity;
2237 wks.col10.comment$=Maxima Spectrum 2;
2238 plotxy iy:=(9,10) plot:=201 color:=color(0,0,255)
ogl:=[%W|2] rescale:=0;
2239 win -a %W;
2240 %K = %V_J;
2241 set %K -k 2;
2242 set %K -z 4;
2243 set %K -q 1;
2244 set %K -qm 1;
2245 %K -qb;
2246 %K qc 4;
2247 set %K -qp 4;
2248 set %K -qr 90;
2249 set %K -qy 20;
2250 set %K -qs 14;
2251 themeApply2g theme:= "MS";
2252 //Delete axis labels and make plot line thicker (layer 2)
2253 label -r xb;
2254 label -r yl;
2255 %K = %V_D;
2256 set %K -w 1000;

2257 //Replacing and editing of the legend label (layer 2)
2258 label -r legend;
2259 label -s -n leglay2 \b(%(1Y, @LC));
2260 leglay2.fsize = 20;
2261 leglay2.color=color(0,0,255);
2262 leglay2.background=1;
2263 leglay2.y = layer2.y.to - leglay2.dy / 2;
2264 leglay2.x = layer2.x.from + leglay2.dx / 2;
2265 //Creating and editing of Laser-Intensity label (layer 2)
2266 label -s -n lasintlay2 \p100(Laser-Intensity: %(LI32$));
2267 lasintlay2.rotate= 90;
2268 lasintlay2.y = (layer.y.to - layer.y.from) / 2;
2269 lasintlay2.x = layer.x.to + lasintlay2.dx / 2;
2270 //Add scrollbar and edit its apperance (layer 2)
2271 axis_scrollbar axis:=left begin:=%(layer2.y1.from$)
end:=%(layer2.y1.to$) layer:=2;
2272 yscrollbar.color=color(0,0,255);
2273 yscrollbar.fillcolor=color(0,0,255);
2274 yscrollthumb.color=color(0,0,255);
2275 yscrollthumb.fillcolor=color(0,0,255);
2276 yscrollbar.states=0;
2277 yscrollbar.height=1425;
2278 yscrollbar.top=1615;
2279 yscrollbar.width=120;
2280 yscrollbar.left=6290;
2281 yscrollbar.states=64;
2282 yscrollbar.states=128;
2283 yscrollbar.states=1;
2284 yscrollbar.states=2;
2285 yscrollbar.states=4;
2286 yscrollbar.states=8;
2287 yscrollbar.states=16;
2288 yscrollbar.states=32;
2289 yscrollbar.mouse=0;
2290 //Rescaling of y-axis (layer 2)
2291 if (yesno == 1) {
2292 layer1.y1.from = -7;
2293 layer1.y1.inc = 30;
2294 layer1.y1.to = 140;
2295 };
2296 if (yesno == 0) {
2297 layer1.y1.from = -15000;
2298 layer1.y1.inc = 100000;
2299 layer1.y1.to = 270000;
2300 };
2301 //THIRD plot editing (layer 3)
2302 %W|page.active = 3;
2303 //Peakpicking and peak plotting (layer 3)
2304 pkfind dir:=p method:=max npts:=10 filter:=h
value:=MH33 ocenter_x:=<new> ocenter_y:=<new>
ocenter:=<none>;
2305 win -a %V;
2306 wks.col11.Iname$=m/z;
2307 wks.col12.Iname$=Absolut Intensity;
2308 wks.col12.comment$=Maxima Spectrum 3;
2309 plotxy iy:=(11,12) plot:=201 color:=color(0,0,0)
ogl:=[%W|3] rescale:=0;
2310 win -a %W;
2311 %L = %V_L;
2312 set %L -k 2;
2313 set %L -z 4;
2314 set %L -q 1;
2315 set %L -qm 1;
2316 set %L -qb;
2317 set %L -qc 1;
2318 set %L -qp 4;
2319 set %L -qr 90;
2320 %L -qy 20;
2321 %L qs 14;
2322 themeApply2g theme:= "MS";
2323 //Edit axis labels and make plot line thicker (layer 3)
2324 yl.fsize=31;
2325 xb.fsize=31;
2326 yl.top=2100;
```


Appendix

LabTalk script (OriginPro 2020b)

```

2327 %L = %V_F;
2328 set %L -w 1000;
2329 //Replacing and editing of the legend label (layer 3)
2330 label -r legend;
2331 label -s -n leglay3 \b(%(1Y, @LC));
2332 leglay3.fsiz = 20;
2333 leglay3.color=color(0,0,0);
2334 leglay3.background=1;
2335 leglay3.y = layer3.y.to - leglay3.dy / 2;
2336 leglay3.x = layer3.x.from + leglay3.dx / 2;
2337 //Creating and editing of Laser-Intensity label (layer 3)
2338 label -s -n lasintlay3 \p100(Laser-Intensity: %(LI33$));
2339 lasintlay3.rotate = 90;
2340 lasintlay3.y = (layer.y.to - layer.y.from) / 2;
2341 lasintlay3.x = layer.x.to + lasintlay3.dx / 2;
2342 //Add scrollbar and edit its appearance (layer 3)
2343 axis_scrollbar axis:=left begin:=(layer3.y1.from$)
end:=(layer3.y1.to$) layer:=3;
2344 yscrollbar.color=color(0,0,0);
2345 yscrollbar.fillcolor=color(0,0,0);
2346 yscrollthumb.color=color(0,0,0);
2347 yscrollthumb.fillcolor=color(0,0,0);
2348 yscrollbar.states=0;
2349 yscrollbar.height=1425;
2350 yscrollbar.top=3050;
2351 yscrollbar.width=120;
2352 yscrollbar.left=6290;
2353 yscrollbar.states=64;
2354 yscrollbar.states=128;
2355 yscrollbar.states=1;
2356 yscrollbar.states=2;
2357 yscrollbar.states=4;
2358 yscrollbar.states=8;
2359 yscrollbar.states=16;
2360 yscrollbar.states=32;
2361 yscrollbar.mouse=0;
2362 //Rescaling of y-axis (layer 3)
2363 if (yesno == 1) {
2364 layer1.y1.from = -7;
2365 layer1.y1.inc = 30;
2366 layer1.y1.to = 140;
2367 };
2368 if (yesno == 0) {
2369 layer1.y1.from = -15000;
2370 layer1.y1.inc = 100000;
2371 layer1.y1.to = 270000;
2372 };
2373 //Add x-scroller and edit its appearance
2374 axis_scroller xsize:=0.4 ysize:=-1.1;
2375 XMin.color=color(0,0,0);
2376 XMin.fillcolor=color(0,0,0);
2377 XMax.color=color(0,0,0);
2378 XMax.fillcolor=color(0,0,0);
2379 //Adding measurement parameter labels
2380 label -s -d 4130 4900 -n detectionmode \p80(Detection-
mode: %(DM3$));
2381 label -s -d 10 4900 -n sampleprep \p80(Sample-prepa-
ration: %(SP3$));
2382 win -a %V;
2383 win -ia;
2384 win -a %W;
2385 win -z;
2386 type -b "Frist spectrum: \n%P \nSecond spectrum:
\n%Q \nThird spectrum: \n%R";
2387 speedmode index:=project sm:=off;
2388 break 1;
2389 };
2390 //Stackplot of 2 spectra
2391 if (exist(col(4)) == 1) {
2392 newbook dummy;
2393 //Enter information about plotted spectra, measurement
parameter and peakpicking parameter
2394 col(a)[1]$="Spektrum 1";
2395 col(a)[2]$="Spektrum 2";
2396 col(b)[1]$=%P;
2397 col(b)[2]$=%Q;
2398 wcolwidth img:=Col(2) width:=100;
2399 win -z;
2400 getn
2401 (TOP Spectrum:) :@G
2402 (Matrix) NoM21$
2403 (Sample) NoS21$
2404 (Additive) NoA21$
2405 (Laser-Intensity) LI21$
2406 (-) :@G
2407 (BOTTOM Spectrum:) :@G
2408 (Matrix) NoM22$
2409 (Sample) NoS22$
2410 (Additive) NoA22$
2411 (Laser-Intensity) LI22$
2412 (-) :@G
2413 (General Parameters:) :@G
2414 (Detection mode) DM2$
2415 (Sample preparation) SP2$
2416 (-) :@G
2417 (Please insert the measurement parameters for plot la-
beling and the abbreviations used for the data file
names of the individual spectra.);
2418 win -c;
2419 MH21=max(col(2));
2420 MH22=max(col(4));
2421 //Editing of data book
2422 win -a %V;
2423 wks.col1.lname$ = m/z;
2424 wks.col2.lname$ = Absolute Intensity;
2425 wks.col2.comment$ = "Matrix: %(NoM21$) ; Analyte:
%(NoS21$) ; Additive: %(NoA21$)";
2426 wks.col3.lname$ = m/z;
2427 wks.col4.lname$ = Absolute Intensity;
2428 wks.col4.comment$ = "Matrix: %(NoM22$) ; Analyte:
%(NoS22$) ; Additive: %(NoA22$)";
2429 //Stack plotting
2430 plotxy iy:=(1,2) plot:=200 color:=color(0,0,255);
2431 %W = %H;
2432 layadd igp:=%W type:=normal userdef:=1 linkto:=1
axis:=straight yaxis:=straight;
2433 win -a %V;
2434 plotxy iy:=(3,4) plot:=200 color:=color(0,0,0)
ogl:=[%W]2!;
2435 win -a %W;
2436 //Sizing of the layers
2437 layer1.unit = 1;
2438 layer1.width = 87.5;
2439 layer1.height = 43.5;
2440 layer1.left = 10;
2441 layer1.top = 4;
2442 layer2.unit = 1;
2443 layer2.width = 87.5;
2444 layer2.height = 43.5;
2445 layer2.left = 10;
2446 layer2.top = 47.5;
2447 //Sizing the labels and change their position
2448 layer1.y1.label.pt=20;
2449 layer1.x2.label.pt=20;
2450 layer2.y1.label.pt=20;
2451 layer2.x1.label.pt=20;
2452 layer1.x2.label.offsetV=-30;
2453 layer1.y1.label.offsetH=20;
2454 layer2.y1.label.offsetH=20;
2455 //Scaling of the x/y-axis
2456 if (SoA == 1) {
2457 layer.x1.from = 0;
2458 layer.x1.inc = 500;
2459 layer.x1.to = 3000;
2460 };
2461 if (SoA == 2) {
2462 layer.x1.from = 1000;
2463 layer.x1.inc = 10000;

```

Appendix

LabTalk script (OriginPro 2020b)

```
2464 layer.x1.to = 100000;
2465 };
2466 if (yesno == 1) {
2467 layer1.y1.from = -5;
2468 layer1.y1.inc = 30;
2469 layer1.y1.to = 140;
2470 };
2471 if (yesno == 0) {
2472 layer1.y1.from = -15000;
2473 layer1.y1.inc = 100000;
2474 layer1.y1.to = 250000;
2475 };
2476 //Design of the layer
2477 layer1.x1.labelssubtype=1;
2478 layer1.x1.showaxes=3;
2479 if (yesno == 1) {
2480 layer1.y1.labelssubtype=1;
2481 };
2482 if (yesno == 0) {
2483 layer1.y1.labelssubtype=2;
2484 };
2485 layer1.y1.showaxes=3;
2486 layer1.x1.ticks=1 + 2 + 4 + 8;
2487 layer1.x2.ticks = 2 + 8;
2488 layer1.y1.ticks=1 + 4;
2489 layer1.y1.label.decPlaces=0;
2490 layer1.x.showLabels=2;
2491 if (yesno == 1) {
2492 layer2.y1.labelssubtype=1;
2493 };
2494 if (yesno == 0) {
2495 layer2.y1.labelssubtype=2;
2496 };
2497 layer2.y1.showaxes=3;
2498 layer2.y1.ticks=1 + 4;
2499 layer2.x1.ticks=1 + 4;
2500 layer2.y1.label.decPlaces=0;
2501 //Activation & design of the layer grids
2502 layer1.x.showGrids=1;
2503 layer1.x.grid.majorColor=color(192,192,192);
2504 layer1.x.grid.majorType=1;
2505 layer2.x.showGrids=1;
2506 layer2.x.grid.majorColor=color(192,192,192);
2507 layer2.x.grid.majorType=1;
2508 //FIRST plot editing (layer 1)
2509 %W!page.active = 1;
2510 //Peakpicking and peak plotting (layer 1)
2511 pkfind dir:=p method:=max npts:=10 filter:=h
value:=MH21 ocenter_x:=<new> ocenter_y:=<new>
ocenter:=<none>;
2512 win -a %V;
2513 wks.col5.lname$=m/z;
2514 wks.col6.lname$=Absolut Intensity;
2515 wks.col6.comment$=Maxima Spectrum 1;
2516 plotxyiy:=(5,6) plot:=201 color:=color(0,0,255)
ogl:=[%W]1! rescale:=0;
2517 win -a %W;
2518 %J = %V_F;
2519 set %J -k 2;
2520 set %J -z4;
2521 set %J -q 1;
2522 set %J -qm 1;
2523 set %J -qb;
2524 set %J -qc 4;
2525 set %J -qp 4;
2526 set %J -qr 90;
2527 set %J -qy 20;
2528 set %J -qs 14;
2529 themeApply2g theme:="MS";
2530 //Delete axis labels and make plot line thicker (layer 1)
2531 label -r xb;
2532 label -r yl;
2533 %J = %V_B;
2534 set %J -w 1000;
2535 //Replacing and editing of the legend label (layer 1)
2536 label -r legend;
2537 label -s -n leglay1 lb(%(1Y, @LC));
2538 leglay1.fsize = 20;
2539 leglay1.color=color(0,0,255);
2540 leglay1.background=1;
2541 leglay1.y = layer1.y.to - leglay1.dy /2;
2542 leglay1.x = layer1.x.from + leglay1.dx /2;
2543 //Creating and editing of Laser-Intensity label (layer 1)
2544 label -s -n lasintlay1 lp80(Laser-Intensity: %(LI21$));
2545 lasintlay1.rotate= 90;
2546 lasintlay1.y = (layer.y.to - layer.y.from) / 2;
2547 lasintlay1.x = layer.x.to + lasintlay1.dx / 2 - 10;
2548 //Add scrollbar and edit its apperance (layer 1)
2549 axis_scrollbar axis:=left begin:=%(layer1.y1.from$)
end:=%(layer1.y1.to$);
2550 yscrollbar.color=color(0,0,255);
2551 yscrollbar.fillcolor=color(0,0,255);
2552 yscrollthumb.color=color(0,0,255);
2553 yscrollthumb.fillcolor=color(0,0,255);
2554 yscrollbar.states=0;
2555 yscrollbar.height=2140;
2556 yscrollbar.top=190;
2557 yscrollbar.width=125;
2558 yscrollbar.left=6290;
2559 yscrollbar.states=64;
2560 yscrollbar.states=128;
2561 yscrollbar.states=1;
2562 yscrollbar.states=2;
2563 yscrollbar.states=4;
2564 yscrollbar.states=8;
2565 yscrollbar.states=16;
2566 yscrollbar.states=32;
2567 yscrollbar.mouse=0;
2568 //Rescaling of y-axis (layer 1)
2569 if (yesno == 1) {
2570 layer1.y1.from = -5;
2571 layer1.y1.inc = 30;
2572 layer1.y1.to = 140;
2573 };
2574 if (yesno == 0) {
2575 layer1.y1.from = -15000;
2576 layer1.y1.inc = 100000;
2577 layer1.y1.to = 250000;
2578 };
2579 //SECOND plot editing (layer 2)
2580 %W!page.active = 2;
2581 //Peakpicking and peak plotting (layer 2)
2582 pkfind dir:=p method:=max npts:=10 filter:=h
value:=MH22 ocenter_x:=<new> ocenter_y:=<new>
ocenter:=<none>;
2583 win -a %V;
2584 wks.col7.lname$=m/z;
2585 wks.col8.lname$=Absolut Intensity;
2586 wks.col8.comment$=Maxima Spectrum 2;
2587 plotxy iy:=(7,8) plot:=201 color:=color(0,0,0)
ogl:=[%W]2! rescale:=0;
2588 win -a %W;
2589 %K = %V_H;
2590 set %K -k 2;
2591 set %K -z 4;
2592 set %K -q 1;
2593 set %K -qm 1;
2594 set %K -qb;
2595 set %K -qc 1;
2596 set %K -qp 4;
2597 set %K -qr 90;
2598 set %K -qy 20;
2599 set %K -qs 14;
2600 themeApply2g theme:="MS";
2601 //Edit axis labels and make plot line thicker (layer 2)
2602 yl.fsize=25;
2603 xb.fsize=25;
2604 yl.top=2400;
```



```
2605 %K = %V_D;
2606 set %K -w 1000;
2607 //Replacing and editing of the legend label (layer 2)
2608 label -r legend;
2609 label -s -n leglay2 \b(1Y, @LC));
2610 leglay2.fsize = 20;
2611 leglay2.color=color(0,0,0);
2612 leglay2.background=1;
2613 leglay2.y = layer2.y.to - leglay2.dy / 2;
2614 leglay2.x = layer2.x.from + leglay2.dx / 2;
2615 //Creating and editing of Laser-Intensity label (layer 2)
2616 label -s -n lasintlay2 \p80(Laser-Intensity: %(LI22$));
2617 lasintlay2.rotate= 90;
2618 lasintlay2.y = (layer.y.to - layer.y.from) / 2;
2619 lasintlay2.x = layer.x.to + lasintlay2.dx / 2 - 10;
2620 //Add scrollbar and edit its appearance (layer 1)
2621 axis_scrollbar axis:=left begin:=(layer2.y1.from$)
end:=(layer2.y1.to$);
2622 yscrollbar.color=color(0,0,0);
2623 yscrollbar.fillcolor=color(0,0,0);
2624 yscrollthumb.color=color(0,0,0);
2625 yscrollthumb.fillcolor=color(0,0,0);
2626 yscrollbar.states=0;
2627 yscrollbar.height=2130;
2628 yscrollbar.top=2340;
2629 yscrollbar.width=125;
2630 yscrollbar.left=6290;
2631 yscrollbar.states=64;
2632 yscrollbar.states=128;
2633 yscrollbar.states=1;
2634 yscrollbar.states=2;
2635 yscrollbar.states=4;
2636 yscrollbar.states=8;
2637 yscrollbar.states=16;
2638 yscrollbar.states=32;
2639 yscrollbar.mouse=0;
2640 //Rescaling of y-axis (layer 1)

2641 if (yesno == 1) {
2642 layer1.y1.from = -5;
2643 layer1.y1.inc = 30;
2644 layer1.y1.to = 140;
2645 };
2646 if (yesno == 0) {
2647 layer1.y1.from = -15000;
2648 layer1.y1.inc = 100000;
2649 layer1.y1.to = 250000;
2650 };
2651 //Add x-scroller and edit its appearance
2652 axis_scroller xsize:=0.5 ysize:=-1;
2653 XMin.color=color(0,0,0);
2654 XMin.fillcolor=color(0,0,0);
2655 XMax.color=color(0,0,0);
2656 XMax.fillcolor=color(0,0,0);
2657 //Adding measurement parameter labels
2658 label -s -d 4130 4930 -n detectionmode \p80(Detection-
mode: %(DM2$));
2659 label -s -d 10 4930 -n sampleprep \p80(Sample-prepa-
ration: %(SP2$));
2660 win -a %V;
2661 win -ia;
2662 win -a %W;
2663 win -z;
2664 type -b "Frist spectrum: \n%P \nSecond spectrum:
\n%Q";
2665 speedmode index:=project sm:=off;
2666 break 1;
2667 };
2668 };
2669 }
2670 else type -b "Please start again and choose only one
plot type!";
2671 }
2672 else type -b "Please start again and choose only one
analyte type!";
2673 };
```

8.1.2. Section "smooth" (L. 2675 - 3078)

```
2675 [smooth]
2676 %W=%H;
2677 range -w aa=%C;
2678 range bb = uid2range(aa.GetPage());$;
2679
2680 dlgChkList inames="Smoothing of all spectra|Smooth-
ing of marked spectrum" osel:=PPP title:"Smoothing"
desc:="For which spectra do you want to perform
smoothing? Check only ONE box" olist:=CB;
2681
2682 if (CB.GetLength() == 1) {
2683 //Smoothing of all spektra
2684 if (PPP == 1) {
2685 //Smoothing of all spektra in 1-stack plot
2686 if (page.nlayers == 1) {
2687 win -a %(bb.name$);
2688 page.active=1;
2689 %L = wks.col2.comment$;
2690 win -a %W;
2691 smooth iy:=[%(bb.name$)]Sheet1!col(2) oy:=<input>
method:=3;
2692 win -a %(bb.name$);
2693 page.active=1;
2694 wks.col2.comment$ = %L;
2695 win -a %W;
2696 };
2697 //Smoothing of all spektra in 2-stack plot
2698 if (page.nlayers == 2) {
2699 win -a %(bb.name$);
2700 page.active=1;
2701 %L = wks.col2.comment$;
2702 %M = wks.col4.comment$;
2703 win -a %W;

2704 smooth iy:=[%(bb.name$)]Sheet1!col(2) oy:=<input>
method:=3;
2705 smooth iy:=[%(bb.name$)]Sheet1!col(4) oy:=<input>
method:=3;
2706 win -a %(bb.name$);
2707 page.active=1;
2708 wks.col2.comment$ = %L;
2709 wks.col4.comment$ = %M;
2710 win -a %W;
2711 };
2712 //Smoothing of all spektra in 3-stack plot
2713 if (page.nlayers == 3) {
2714 win -a %(bb.name$);
2715 page.active=1;
2716 %L = wks.col2.comment$;
2717 %M = wks.col4.comment$;
2718 %N = wks.col6.comment$;
2719 win -a %W;
2720 smooth iy:=[%(bb.name$)]Sheet1!col(2) oy:=<input>
method:=3;
2721 smooth iy:=[%(bb.name$)]Sheet1!col(4) oy:=<input>
method:=3;
2722 smooth iy:=[%(bb.name$)]Sheet1!col(6) oy:=<input>
method:=3;
2723 win -a %(bb.name$);
2724 page.active=1;
2725 wks.col2.comment$ = %L;
2726 wks.col4.comment$ = %M;
2727 wks.col6.comment$ = %N;
2728 win -a %W;
2729 };
2730 //Smoothing of all spektra in 4-stack plot
2731 if (page.nlayers == 4) {
```

Appendix

LabTalk script (OriginPro 2020b)

```
2732 win -a %(bb.name$);
2733 page.active=1;
2734 %L = wks.col2.comment$;
2735 %M = wks.col4.comment$;
2736 %N = wks.col6.comment$;
2737 %O = wks.col8.comment$;
2738 win -a %W;
2739 smooth iy:=[%(bb.name$)]Sheet1!col(2) oy:=<input>
method:=3;
2740 smooth iy:=[%(bb.name$)]Sheet1!col(4) oy:=<input>
method:=3;
2741 smooth iy:=[%(bb.name$)]Sheet1!col(6) oy:=<input>
method:=3;
2742 smooth iy:=[%(bb.name$)]Sheet1!col(8) oy:=<input>
method:=3;
2743 win -a %(bb.name$);
2744 page.active=1;
2745 wks.col2.comment$ = %L;
2746 wks.col4.comment$ = %M;
2747 wks.col6.comment$ = %N;
2748 wks.col8.comment$ = %O;
2749 win -a %W;
2750 };
2751 //Smoothing of all spektra in 5-stack plot
2752 if (page.nlayers == 5) {
2753 win -a %(bb.name$);
2754 page.active=1;
2755 %L = wks.col2.comment$;
2756 %M = wks.col4.comment$;
2757 %N = wks.col6.comment$;
2758 %O = wks.col8.comment$;
2759 %P = wks.col10.comment$;
2760 win -a %W;
2761 smooth iy:=[%(bb.name$)]Sheet1!col(2) oy:=<input>
method:=3;
2762 smooth iy:=[%(bb.name$)]Sheet1!col(4) oy:=<input>
method:=3;
2763 smooth iy:=[%(bb.name$)]Sheet1!col(6) oy:=<input>
method:=3;
2764 smooth iy:=[%(bb.name$)]Sheet1!col(8) oy:=<input>
method:=3;
2765 smooth iy:=[%(bb.name$)]Sheet1!col(10) oy:=<input>
method:=3;
2766 win -a %(bb.name$);
2767 page.active=1;
2768 wks.col2.comment$ = %L;
2769 wks.col4.comment$ = %M;
2770 wks.col6.comment$ = %N;
2771 wks.col8.comment$ = %O;
2772 wks.col10.comment$ = %P;
2773 win -a %W;
2774 };
2775 //Smoothing of all spektra in 6-
stack plot
2776 if (page.nlayers == 6) {
2777 win -a %(bb.name$);
2778 page.active=1;
2779 %L = wks.col2.comment$;
2780 %M = wks.col4.comment$;
2781 %N = wks.col6.comment$;
2782 %O = wks.col8.comment$;
2783 %P = wks.col10.comment$;
2784 %Q = wks.col12.comment$;
2785 win -a %W;
2786 smooth iy:=[%(bb.name$)]Sheet1!col(2) oy:=<input>
method:=3;
2787 smooth iy:=[%(bb.name$)]Sheet1!col(4) oy:=<input>
method:=3;
2788 smooth iy:=[%(bb.name$)]Sheet1!col(6) oy:=<input>
method:=3;
2789 smooth iy:=[%(bb.name$)]Sheet1!col(8) oy:=<input>
method:=3;
2790 smooth iy:=[%(bb.name$)]Sheet1!col(10) oy:=<input>
method:=3;
2791 smooth iy:=[%(bb.name$)]Sheet1!col(12) oy:=<input>
method:=3;
2792 win -a %(bb.name$);
2793 page.active=1;
2794 wks.col2.comment$ = %L;
2795 wks.col4.comment$ = %M;
2796 wks.col6.comment$ = %N;
2797 wks.col8.comment$ = %O;
2798 wks.col10.comment$ = %P;
2799 wks.col12.comment$ = %Q;
2800 win -a %W;
2801 };
2802 };
2803 //Smoothing of marked spektrum
2804 if (PPP == 2){
2805 //Smoothing of marked spektrum in 1-stack plot
2806 if (page.nlayers == 1) {
2807 //Smoothing of layer 1
2808 if (page.active == 1) {
2809 win -a %(bb.name$);
2810 page.active=1;
2811 %L = wks.col2.comment$;
2812 win -a %W;
2813 smooth iy:=[%(bb.name$)]Sheet1!col(2) oy:=<input>
method:=3;
2814 win -a %(bb.name$);
2815 page.active=1;
2816 wks.col2.comment$ = %L;
2817 win -a %W;
2818 };
2819 };
2820 //Smoothing of marked spektrum in 2-stack plot
2821 if (page.nlayers == 2) {
2822 //Smoothing of layer 1
2823 if (page.active == 1) {
2824 win -a %(bb.name$);
2825 page.active=1;
2826 %L = wks.col2.comment$;
2827 win -a %W;
2828 smooth iy:=[%(bb.name$)]Sheet1!col(2) oy:=<input>
method:=3;
2829 win -a %(bb.name$);
2830 page.active=1;
2831 wks.col2.comment$ = %L;
2832 win -a %W;
2833 };
2834 //Smoothing of layer 2
2835 if (page.active == 2) {
2836 win -a %(bb.name$);
2837 page.active=1;
2838 %M = wks.col4.comment$;
2839 win -a %W;
2840 smooth iy:=[%(bb.name$)]Sheet1!col(4) oy:=<input>
method:=3;
2841 win -a %(bb.name$);
2842 page.active=1;
2843 wks.col4.comment$ = %M;
2844 %W;
2845 };
2846 };
2847 //Smoothing of marked spektrum in 3-stack plot
2848 if (page.nlayers == 3) {
2849 //Smoothing of layer 1
2850 if (page.active == 1) {
2851 win -a %(bb.name$);
2852 page.active=1;
2853 %L = wks.col2.comment$;
2854 win -a %W;
2855 smooth iy:=[%(bb.name$)]Sheet1!col(2) oy:=<input>
method:=3;
2856 win -a %(bb.name$);
2857 page.active=1;
2858 wks.col2.comment$ = %L;
2859 win -a %W;
```

Appendix

LabTalk script (OriginPro 2020b)

```
2860 };
2861 //Smoothing of layer 2
2862 if (page.active == 2) {
2863 win -a %(bb.name$);
2864 page.active=1;
2865 %M = wks.col4.comment$;
2866 win -a %W;
2867 smooth iy:=[%(bb.name$)]Sheet1!col(4) oy:=<input>
method:=3;
2868 win -a %(bb.name$);
2869 page.active=1;
2870 wks.col4.comment$ = %M;
2871 win -a %W;
2872 };
2873 //Smoothing of layer 3
2874 if (page.active == 3) {
2875 win -a %(bb.name$);
2876 page.active=1;
2877 %N = wks.col6.comment$;
2878 win -a %W;
2879 smooth iy:=[%(bb.name$)]Sheet1!col(6) oy:=<input>
method:=3;
2880 win -a %(bb.name$);
2881 page.active=1;
2882 wks.col6.comment$ = %N;
2883 win -a %W;
2884 };
2885 };
2886 //Smoothing of marked spektrum in 4-stack plot
2887 if (page.nlayers == 4) {
2888 //Smoothing of layer 1
2889 if (page.active == 1) {
2890 win -a %(bb.name$);
2891 page.active=1;
2892 %L = wks.col2.comment$;
2893 win -a %W;
2894 smooth iy:=[%(bb.name$)]Sheet1!col(2) oy:=<input>
method:=3;
2895 win -a %(bb.name$);
2896 page.active=1;
2897 wks.col2.comment$ = %L;
2898 win -a %W;
2899 };
2900 //Smoothing of layer 2
2901 if (page.active == 2) {
2902 win -a %(bb.name$);
2903 page.active=1;
2904 %M = wks.col4.comment$;
2905 win -a %W;
2906 smooth iy:=[%(bb.name$)]Sheet1!col(4) oy:=<input>
method:=3;
2907 win -a %(bb.name$);
2908 page.active=1;
2909 wks.col4.comment$ = %M;
2910 win -a %W;
2911 };
2912 //Smoothing of layer 3
2913 if (page.active == 3) {
2914 win -a %(bb.name$);
2915 page.active=1;
2916 %N = wks.col6.comment$;
2917 win -a %W;
2918 smooth iy:=[%(bb.name$)]Sheet1!col(6) oy:=<input>
method:=3;
2919 win -a %(bb.name$);
2920 page.active=1;
2921 wks.col6.comment$ = %N;
2922 win -a %W;
2923 };
2924 //Smoothing of layer 4
2925 if (page.active == 4) {
2926 win -a %(bb.name$);
2927 page.active=1;
2928 %O = wks.col8.comment$;
2929 win -a %W;
2930 smooth iy:=[%(bb.name$)]Sheet1!col(8) oy:=<input>
method:=3;
2931 win -a %(bb.name$);
2932 page.active=1;
2933 wks.col8.comment$ = %O;
2934 win -a %W;
2935 };
2936 };
2937 //Smoothing of marked spektrum in 5-stack plot
2938 if (page.nlayers == 5) {
2939 //Smoothing of layer 1
2940 if (page.active == 1) {
2941 win -a %(bb.name$);
2942 page.active=1;
2943 %L = wks.col2.comment$;
2944 win -a %W;
2945 smooth iy:=[%(bb.name$)]Sheet1!col(2) oy:=<input>
method:=3;
2946 win -a %(bb.name$);
2947 page.active=1;
2948 wks.col2.comment$ = %L;
2949 win -a %W;
2950 };
2951 //Smoothing of layer 2
2952 if (page.active == 2) {
2953 win -a %(bb.name$);
2954 page.active=1;
2955 %M = wks.col4.comment$;
2956 win -a %W;
2957 smooth iy:=[%(bb.name$)]Sheet1!col(4) oy:=<input>
method:=3;
2958 win -a %(bb.name$);
2959 page.active=1;
2960 wks.col4.comment$ = %M;
2961 win -a %W;
2962 };
2963 //Smoothing of layer 3
2964 if (page.active == 3) {
2965 win -a %(bb.name$);
2966 page.active=1;
2967 %N = wks.col6.comment$;
2968 win -a %W;
2969 smooth iy:=[%(bb.name$)]Sheet1!col(6) oy:=<input>
method:=3;
2970 win -a %(bb.name$);
2971 page.active=1;
2972 wks.col6.comment$ = %N;
2973 win -a %W;
2974 };
2975 //Smoothing of layer 4
2976 if (page.active == 4) {
2977 win -a %(bb.name$);
2978 page.active=1;
2979 %O = wks.col8.comment$;
2980 win -a %W;
2981 smooth iy:=[%(bb.name$)]Sheet1!col(8) oy:=<input>
method:=3;
2982 win -a %(bb.name$);
2983 page.active=1;
2984 wks.col8.comment$ = %O;
2985 win -a %W;
2986 };
2987 //Smoothing of layer 5
2988 if (page.active == 5) {
2989 win -a %(bb.name$);
2990 page.active=1;
2991 %P = wks.col10.comment$;
2992 win -a %W;
2993 smooth iy:=[%(bb.name$)]Sheet1!col(10) oy:=<input>
method:=3;
2994 win -a %(bb.name$);
2995 page.active=1;
2996 wks.col10.comment$ = %P;
```

```
2997 win -a %W;
2998 };
2999 };
3000 //Smoothing of marked spektrum in 6-stack plot
3001 if (page.nlayers == 6) {
3002 //Smoothing of layer 1
3003 if (page.active == 1) {
3004 win -a %(bb.name$);
3005 page.active=1;
3006 %L = wks.col2.comment$;
3007 win -a %W;
3008 smooth iy:=[%(bb.name$)]Sheet1!col(2) oy:=<input>
method:=3;
3009 win -a %(bb.name$);
3010 page.active=1;
3011 wks.col2.comment$ = %L;
3012 win -a %W;
3013 };
3014 //Smoothing of layer 2
3015 if (page.active == 2) {
3016 win -a %(bb.name$);
3017 page.active=1;
3018 %M = wks.col4.comment$;
3019 win -a %W;
3020 smooth iy:=[%(bb.name$)]Sheet1!col(4) oy:=<input>
method:=3;
3021 win -a %(bb.name$);
3022 page.active=1;
3023 wks.col4.comment$ = %M;
3024 win -a %W;
3025 };
3026 //Smoothing of layer 3
3027 if (page.active == 3) {
3028 win -a %(bb.name$);
3029 page.active=1;
3030 %N = wks.col6.comment$;
3031 win -a %W;
3032 smooth iy:=[%(bb.name$)]Sheet1!col(6) oy:=<input>
method:=3;
3033 win -a %(bb.name$);
3034 page.active=1;
3035 wks.col6.comment$ = %N;
3036 win -a %W;
3037 };
3038 //Smoothing of layer 4
3039 if (page.active == 4) {
3040 win -a %(bb.name$);
3041 page.active=1;
3042 %O = wks.col8.comment$;
3043 win -a %W;
3044 smooth iy:=[%(bb.name$)]Sheet1!col(8) oy:=<input>
method:=3;
3045 win -a %(bb.name$);
3046 page.active=1;
3047 wks.col8.comment$ = %O;
3048 win -a %W;
3049 };
3050 //Smoothing of layer 5
3051 if (page.active == 5) {
3052 win -a %(bb.name$);
3053 page.active=1;
3054 %P = wks.col10.comment$;
3055 win -a %W;
3056 smooth iy:=[%(bb.name$)]Sheet1!col(10) oy:=<input>
method:=3;
3057 win -a %(bb.name$);
3058 page.active=1;
3059 wks.col10.comment$ = %P;
3060 win -a %W;
3061 };
3062 //Smoothing of layer 6
3063 if (page.active == 6) {
3064 win -a %(bb.name$);
3065 page.active=1;
3066 %Q = wks.col12.comment$;
3067 win -a %W;
3068 smooth iy:=[%(bb.name$)]Sheet1!col(12) oy:=<input>
method:=3;
3069 win -a %(bb.name$);
3070 page.active=1;
3071 wks.col12.comment$ = %Q;
3072 win -a %W;
3073 };
3074 };
3075 };
3076 }
3077 else type -b "Please start again and choose only one
smoothing operation!";
3078 };
```

8.1.3. Section "baseline" (L. 3080 - 4686)

```
3080 [baseline]
3081 dlgChkList inames:="Manual Baseline Correction|Auto-
matic Baseline Correction" osel:=PPP title:="Baseline
Correction" desc:="Which type of baseline correction do
you want to perform? Check only ONE box" olist:=CB;
3082
3083 if (CB.GetLength() == 1) {
3084 //Manual baseline correction
3085 if (PPP == 1) {
3086 type -b "Use the data reader to define the data points of
the baseline. Press Enter after each point selection.
Press Esc when you are done and the baseline correc-
tion begins. It does not have to be the whole spect-
rum.";
3087 //Set variables, plot baseline in marked layer and start
data reader
3088 %V=%H;
3089 newbook dummy;
3090 %B=%H;
3091 win -a %V;
3092 //Baseline plotting in layer 1
3093 if (page.active == 1) {
3094 win -a %B;
3095 plotxy iy:=(1,2) plot:=202 color:=color(60,255,60)
ogl:=[%V]1! rescale:=0 legend:=0 size:=5;
3096 win -a %V;
3097 };
3098 //Baseline plotting in layer 2
3099 if (page.active == 2) {
3100 win -a %B;
3101 plotxy iy:=(1,2) plot:=202 color:=color(60,255,60)
ogl:=[%V]2! rescale:=0 legend:=0 size:=5;
3102 win -a %V;
3103 };
3104 //Baseline plotting in layer 3
3105 if (page.active == 3) {
3106 win -a %B;
3107 plotxy iy:=(1,2) plot:=202 color:=color(60,255,60)
ogl:=[%V]3! rescale:=0 legend:=0 size:=5;
3108 win -a %V;
3109 };
3110 //Baseline plotting in layer 4
3111 if (page.active == 4) {
3112 win -a %B;
3113 plotxy iy:=(1,2) plot:=202 color:=color(60,255,60)
ogl:=[%V]4! rescale:=0 legend:=0 size:=5;
3114 win -a %V;
3115 };
3116 //Baseline plotting in layer 5
3117 if (page.active == 5) {
```


Appendix

LabTalk script (OriginPro 2020b)

```
3118 win -a %B;
3119 plotxy iy:=(1,2) plot:=202 color:=color(60,255,60)
    ogl:=[%V]5! rescale:=0 legend:=0 size:=5;
3120 win -a %V;
3121 };
3122 //Baseline plotting in layer 6
3123 if (page.active == 6) {
3124 win -a %B;
3125 plotxy iy:=(1,2) plot:=202 color:=color(60,255,60)
    ogl:=[%V]6! rescale:=0 legend:=0 size:=5;
3126 win -a %V;
3127 };
3128 pp=0;
3129 dotool 3;
3130 dotool -z;
3131 def pointproc {
3132 pp++;
3133 %([%B]Sheet1, 1, pp)=X;
3134 %([%B]Sheet1, 2, pp)=Y;
3135 };
3136 def quittoolbox {
3137 layer -e %B_B;
3138 range -w aa=%C;
3139 range bb = uid2range(aa.GetPage());$;
3140 //Baseline correction of layer 1
3141 if (page.active == 1) {
3142 //Data acquisition
3143 win -a %B;
3144 wks.col1.format = 1;
3145 wks.col2.format = 1;
3146 win -a %([bb.name$])Sheet1, 1, 1);
3147 int datastart=list(([%B]Sheet1, 1, 1)).col(1);
3148 int dataend=list(([%B]Sheet1, 1, pp)).col(1);
3149 copydata irng:=[%([bb.name$])Sheet1!col(1)[datastart:dataend]
    orgng:=[%B]Sheet1!col(3);
3150 copydata irng:=[%([bb.name$])Sheet1!col(2)[datastart:dataend]
    orgng:=[%B]Sheet1!col(4);
3151 win -a %B;
3152 int aa=count(col(1),1);
3153 %K=col(1)[aa];
3154 pp=aa-1;
3155 //Calculation of baseline corrected spectrum
3156 for(int ii=1; ii<=pp; ii++){
3157 int datastart1=list(([%B]Sheet1, 1, ii)).col(3);
3158 int dataend1=list(([%B]Sheet1, 1, ii+1)).col(3);
3159 subtract_line iy:=(3,4)[datastart1:dataend1]
    x1:=[%([B]Sheet1, 1, ii)) y1:=[%([B]Sheet1, 2, ii))
    x2:=[%([B]Sheet1, 1, ii+1)) y2:=[%([B]Sheet1, 2,
    ii+1)) oy:=[new>;
3160 };
3161 for(int ii=1; ii<=pp; ii++){
3162 int NoP=count(col(e),1);
3163 if (exist(col(g)) == 1) {
3164 copydata irng:=(7,8) orgng:=(5,6)[NoP];
3165 delete col(g);
3166 delete col(g);
3167 };
3168 };
3169 //Check baseline and calculated spectrum
3170 plotxy iy:=(3,4) plot:=200 color:=color(0,0,0);
3171 %U=%H;
3172 win -a %B;
3173 wks.col1.Iname$ = m/z;
3174 wks.col2.Iname$ = Absolute Intensity;
3175 wks.col2.comment$ = "Baseline";
3176 wks.col3.Iname$ = m/z;
3177 wks.col4.Iname$ = Absolute Intensity;
3178 wks.col4.comment$ = "Original Spectrum";
3179 wks.col5.Iname$ = m/z;
3180 wks.col6.Iname$ = Absolute Intensity;
3181 wks.col6.comment$ = "Baselinecorrected Spectrum";
3182 plotxy iy:=(1,2) plot:=200 color:=color(255,0,0)
    ogl:=[%U]1!;
3183 plotxy iy:=(5,6) plot:=200 color:=color(0,0,255)
    ogl:=[%U]1!;
```

```
3184 win -a %U;
3185 layer.x.from=0;
3186 layer.x.to=%K;
3187 getyesno (Was the baseline correction successful?)
    yesno (Baselinecorrection);
3188 if (yesno == 0){
3189 type -b "Sorry, that it didn't work. Shit happens. Try
    again.";
3190 win -c;
3191 win -a %B;
3192 win -c;
3193 win -a %V;
3194 };
3195 //Replace old spectrum with new one and finish
3196 if (yesno == 1){
3197 win -c;
3198 win -a %B;
3199 int NoP=count(col(e),1);
3200 copydata irng:=[col(5)[1:NoP]
    orgng:=[%([bb.name$])Sheet1!col(1)[datastart:dataend];
3201 copydata irng:=[col(6)[1:NoP]
    orgng:=[%([bb.name$])Sheet1!col(2)[datastart:dataend];
3202 win -c;
3203 win -a %([bb.name$]);
3204 win -ia;
3205 win -a %V;
3206 win -z;
3207 page.active = 1;
3208 plot -l;
3209 %L=layer.x.from$;
3210 %M=layer.x.to$;
3211 %N=layer.y.from$;
3212 %O=layer.y.to$;
3213 rescale;
3214 layer.x.from=%L;
3215 layer.x.to=%M;
3216 layer.y.from=%N;
3217 layer.y.to=%O;
3218 page.active$=6;
3219 layer.x.posttype=1;
3220 layer.x.position=0;
3221 };
3222 };
3223 //Baseline correction of layer 2
3224 if (page.active == 2) {
3225 //Data acquisition
3226 win -a %B;
3227 wks.col1.format = 1;
3228 wks.col2.format = 1;
3229 win -a %([bb.name$]);
3230 int datastart=list(([%B]Sheet1, 1, 1)).col(3);
3231 int dataend=list(([%B]Sheet1, 1, pp)).col(3);
3232 copydata irng:=[%([bb.name$])Sheet1!col(3)[datastart:dataend]
    orgng:=[%B]Sheet1!col(3);
3233 copydata irng:=[%([bb.name$])Sheet1!col(4)[datastart:dataend]
    orgng:=[%B]Sheet1!col(4);
3234 win -a %B;
3235 range rx = 1;
3236 range ry = 2;
3237 int aa=count(col(a),1);
3238 %K=col(1)[aa];
3239 pp=aa-1;
3240 //Calculation of baseline corrected spectrum
3241 for(int ii=1; ii<=pp; ii++){
3242 int datastart1=list(([%B]Sheet1, 1, ii)).col(3);
3243 int dataend1=list(([%B]Sheet1, 1, ii+1)).col(3);
3244 subtract_line iy:=(3,4)[datastart1:dataend1]
    x1:=[%([B]Sheet1, 1, ii)) y1:=[%([B]Sheet1, 2, ii))
    x2:=[%([B]Sheet1, 1, ii+1)) y2:=[%([B]Sheet1, 2,
    ii+1)) oy:=[new>;
3245 };
3246 for(int ii=1; ii<=pp; ii++){
3247 int NoP=count(col(e),1);
3248 if (exist(col(g)) == 1) {
3249 copydata irng:=(7,8) orgng:=(5,6)[NoP];
```


Appendix

LabTalk script (OriginPro 2020b)

```
3250 delete col(g);
3251 delete col(g);
3252 };
3253 };
3254 //Check baseline and calculated spectrum
3255 plotxyiy:=(3,4) plot:=200color:=color(0,0,0);
3256 %U=%H;
3257 win -a %B;
3258 wks.col1.lname$ = m/z;
3259 wks.col2.lname$ = Absolute Intensity;
3260 wks.col2.comment$ = "Baseline";
3261 wks.col3.lname$ = m/z;
3262 wks.col4.lname$ = Absolute Intensity;
3263 wks.col4.comment$ = "Original Spectrum";
3264 wks.col5.lname$ = m/z;
3265 wks.col6.lname$ = Absolute Intensity;
3266 wks.col6.comment$ = "Baselinecorrected Spectrum";
3267 plotxyiy:=(1,2) plot:=200color:=color(255,0,0)
ogl:=[%U]1!;
3268 plotxyiy:=(5,6) plot:=200color:=color(0,0,255)
ogl:=[%U]1!;
3269 win -a %U;
3270 layer.x.from=0;
3271 layer.x.to=%K;
3272 getyesno(Was the baseline correction success-
ful?)yesno(Baselinecorrection);
3273 if (yesno == 0){
3274 type-b"Sorry, that it didn't work. Shit happens. Try
again.";
3275 win -c;
3276 win -a %B;
3277 win -c;
3278 win -a %V;
3279 };
3280 //Replace old spectrum with new one and finish
3281 if (yesno == 1){
3282 win -c;
3283 win -a %B;
3284 int NoP=count(col(e),1);
3285 copydata irng:=col(5)[1:NoP]
orgng:=[%(bb.name$)]Sheet1!col(3)[datastart:dataend];
3286 copydata irng:=col(6)[1:NoP]
orgng:=[%(bb.name$)]Sheet1!col(4)[datastart:dataend];
3287 win -c;
3288 win -a %(bb.name$);
3289 win -ia;
3290 win -a %V;
3291 win -z;
3292 page.active = 2;
3293 plot -l;
3294 %L=layer.x.from$;
3295 %M=layer.x.to$;
3296 %N=layer.y.from$;
3297 %O=layer.y.to$;
3298 rescale;
3299 layer.x.from=%L;
3300 layer.x.to=%M;
3301 layer.y.from=%N;
3302 layer.y.to=%O;
3303 page.active$=6;
3304 layer.x.postype=1;
3305 layer.x.position=0;
3306 };
3307 };
3308 //Baseline correction of layer 3
3309 if (page.active == 3) {
3310 //Data acquisition
3311 win -a %B;
3312 wks.col1.format = 1;
3313 wks.col2.format = 1;
3314 win -a %(bb.name$);
3315 int datastart=list(([%B]Sheet1, 1, 1)),col(5));
3316 int dataend=list(([%B]Sheet1, 1, pp)),col(5));
3317 copydata irng:=[%(bb.name$)]Sheet1!col(5)[datas-
tart:dataend] orgng:=[%B]Sheet1!col(3);
3318 copydata irng:=[%(bb.name$)]Sheet1!col(6)[datas-
tart:dataend] orgng:=[%B]Sheet1!col(4);
3319 win -a %B;
3320 int aa=count(col(a),1);
3321 %K=col(1)[aa];
3322 pp=aa-1;
3323 //Calculation of baseline corrected spectrum
3324 for(int ii=1; ii<=pp; ii++){
3325 int datastart1=list(([%B]Sheet1, 1, ii)),col(3));
3326 int dataend1=list(([%B]Sheet1, 1, ii+1)),col(3));
3327 subtract_line iy:=(3,4)[datastart1:dataend1]
x1:=[%B]Sheet1, 1, ii) y1:=[%B]Sheet1, 2, ii)
x2:=[%B]Sheet1, 1, ii+1) y2:=[%B]Sheet1, 2,
ii+1) oy:=[new];
3328 };
3329 for(int ii=1; ii<=pp; ii++){
3330 int NoP=count(col(5),1);
3331 if (exist(col(7)) == 1) {
3332 copydata irng:=(7,8) orgng:=(5,6)[NoP];
3333 delete col(7);
3334 delete col(7);
3335 };
3336 };
3337 //Check baseline and calculated spectrum
3338 plotxy iy:=(3,4) plot:=200 color:=color(0,0,0);
3339 %U=%H;
3340 win -a %B;
3341 wks.col1.lname$ = m/z;
3342 wks.col2.lname$ = Absolute Intensity;
3343 wks.col2.comment$ = "Baseline";
3344 wks.col3.lname$ = m/z;
3345 wks.col4.lname$ = Absolute Intensity;
3346 wks.col4.comment$ = "Original Spectrum";
3347 wks.col5.lname$ = m/z;
3348 wks.col6.lname$ = Absolute Intensity;
3349 wks.col6.comment$ = "Baselinecorrected Spectrum";
3350 plotxy iy:=(1,2) plot:=200 color:=color(255,0,0)
ogl:=[%U]1!;
3351 plotxy iy:=(5,6) plot:=200 color:=color(0,0,255)
ogl:=[%U]1!;
3352 win -a %U;
3353 layer.x.from=0;
3354 layer.x.to=%K;
3355 getyesno (Was the baseline correction successful?)
yesno (Baselinecorrection);
3356 if (yesno == 0){
3357 type -b "Sorry, that it didn't work. Shit happens. Try
again.";
3358 win -c;
3359 win -a %B;
3360 win -c;
3361 win -a %V;
3362 };
3363 //Replace old spectrum with new one and finish
3364 if (yesno == 1){
3365 win -c;
3366 win -a %B;
3367 int NoP=count(col(e),1);
3368 copydata irng:=col(5)[1:NoP]
orgng:=[%(bb.name$)]Sheet1!col(5)[datastart:dataend];
3369 copydata irng:=col(6)[1:NoP]
orgng:=[%(bb.name$)]Sheet1!col(6)[datastart:dataend];
3370 win -c;
3371 win -a %(bb.name$);
3372 win -ia;
3373 win -a %V;
3374 win -z;
3375 page.active = 3;
3376 plot -l;
3377 %L=layer.x.from$;
3378 %M=layer.x.to$;
3379 %N=layer.y.from$;
3380 %O=layer.y.to$;
```

Appendix

LabTalk script (OriginPro 2020b)

```
3381 rescale;
3382 layer.x.from=%L;
3383 layer.x.to=%M;
3384 layer.y.from=%N;
3385 layer.y.to=%O;
3386 page.active$=6;
3387 layer.x.posttype=1;
3388 layer.x.position=0;
3389 };
3390 };
3391 //Baseline correction of layer 4
3392 if (page.active == 4) {
3393 //Data acquisition
3394 win -a %B;
3395 wks.col1.format = 1;
3396 wks.col2.format = 1;
3397 win -a %(bb.name$);
3398 int datastart=list((%([%B]Sheet1, 1, 1)),col(7));
3399 int dataend=list((%([%B]Sheet1, 1, pp)),col(7));
3400 copydata irng=[%(bb.name$)]Sheet1!col(7)[datastart:dataend]
      orng=[%B]Sheet1!col(3);
3401 copydata irng=[%(bb.name$)]Sheet1!col(8)[datastart:dataend]
      orng=[%B]Sheet1!col(4);
3402 win -a %B;
3403 int aa=count(col(a),1);
3404 %K=col(1)[aa];
3405 pp=aa-1;
3406 //Calculation of baseline corrected spectrum
3407 for(int ii=1; ii<=pp; ii++){
3408 int datastart1=list((%([%B]Sheet1, 1, ii)),col(3));
3409 int dataend1=list((%([%B]Sheet1, 1, ii+1)),col(3));
3410 subtract_line iy=(3,4)[datastart1:dataend1]
      x1=(%([%B]Sheet1, 1, ii)) y1=(%([%B]Sheet1, 2, ii))
      x2=(%([%B]Sheet1, 1, ii+1)) y2=(%([%B]Sheet1, 2, ii+1))
      oy=<new>;
3411 };
3412 for(int ii=1; ii<=pp; ii++){
3413 int NoP=count(col(e),1);
3414 if (exist(col(g)) == 1) {
3415 copydata irng=(7,8) orng=(5,6)[NoP];
3416 delete col(g);
3417 delete col(g);
3418 };
3419 };
3420 //Check baseline and calculated spectrum
3421 plotxy iy=(3,4) plot:=200 color:=color(0,0,0);
3422 %U=%H;
3423 win -a %B;
3424 wks.col1.lname$ = m/z;
3425 wks.col2.lname$ = Absolute Intensity;
3426 wks.col2.comment$ = "Baseline";
3427 wks.col3.lname$ = m/z;
3428 wks.col4.lname$ = Absolute Intensity;
3429 wks.col4.comment$ = "Original Spectrum";
3430 wks.col5.lname$ = m/z;
3431 wks.col6.lname$ = Absolute Intensity;
3432 wks.col6.comment$ = "Baselinecorrected Spectrum";
3433 plotxy iy=(1,2) plot:=200 color:=color(255,0,0)
      ogl=[%U]1!;
3434 plotxy iy=(5,6) plot:=200 color:=color(0,0,255)
      ogl=[%U]1!;
3435 win -a %U;
3436 layer.x.from=0;
3437 layer.x.to=%K;
3438 getyesno (Was the baseline correction successful?)
      yesno (Baselinecorrection);
3439 if (yesno == 0){
3440 type -b "Sorry, that it didn't work. Shit happens. Try again.";
3441 win -c;
3442 win -a %B;
3443 win -c;
3444 win -a %V;
3445 };
3446 //Replace old spectrum with new one and finish
3447 if (yesno == 1){
3448 win -c;
3449 win -a %B;
3450 int NoP=count(col(e),1);
3451 copydata irng=[%(bb.name$)]Sheet1!col(7)[datastart:dataend];
      orng=[%(bb.name$)]Sheet1!col(7)[datastart:dataend];
3452 copydata irng=[%(bb.name$)]Sheet1!col(8)[datastart:dataend];
      orng=[%(bb.name$)]Sheet1!col(8)[datastart:dataend];
3453 win -c;
3454 win -a %V;
3455 page.active = 4;
3456 plot -I;
3457 %L=layer.x6.from$;
3458 %M=layer.x6.to$;
3459 rescale;
3460 layer.x6.from=%L;
3461 layer.x6.to=%M;
3462 page.active$=6;
3463 layer.x.posttype=1;
3464 layer.x.position=0;
3465 };
3466 };
3467 //Baseline correction of layer 5
3468 if (page.active == 5) {
3469 //Data acquisition
3470 win -a %B;
3471 wks.col1.format = 1;
3472 wks.col2.format = 1;
3473 win -a %(bb.name$);
3474 int datastart=list((%([%B]Sheet1, 1, 1)),col(9));
3475 int dataend=list((%([%B]Sheet1, 1, pp)),col(9));
3476 copydata irng=[%(bb.name$)]Sheet1!col(9)[datastart:dataend]
      orng=[%B]Sheet1!col(3);
3477 copydata irng=[%(bb.name$)]Sheet1!col(10)[datastart:dataend]
      orng=[%B]Sheet1!col(4);
3478 win -a %B;
3479 int aa=count(col(a),1);
3480 %K=col(1)[aa];
3481 pp=aa-1;
3482 //Calculation of baseline corrected spectrum
3483 for(int ii=1; ii<=pp; ii++){
3484 int datastart1=list((%([%B]Sheet1, 1, ii)),col(3));
3485 int dataend1=list((%([%B]Sheet1, 1, ii+1)),col(3));
3486 subtract_line iy=(3,4)[datastart1:dataend1]
      x1=(%([%B]Sheet1, 1, ii)) y1=(%([%B]Sheet1, 2, ii))
      x2=(%([%B]Sheet1, 1, ii+1)) y2=(%([%B]Sheet1, 2, ii+1))
      oy=<new>;
3487 };
3488 for(int ii=1; ii<=pp; ii++){
3489 int NoP=count(col(e),1);
3490 if (exist(col(g)) == 1) {
3491 copydata irng=(7,8) orng=(5,6)[NoP];
3492 delete col(g);
3493 delete col(g);
3494 };
3495 };
3496 //Check baseline and calculated spectrum
3497 plotxy iy=(3,4) plot:=200 color:=color(0,0,0);
3498 %U=%H;
3499 win -a %B;
3500 wks.col1.lname$ = m/z;
3501 wks.col2.lname$ = Absolute Intensity;
3502 wks.col2.comment$ = "Baseline";
3503 wks.col3.lname$ = m/z;
3504 wks.col4.lname$ = Absolute Intensity;
3505 wks.col4.comment$ = "Original Spectrum";
3506 wks.col5.lname$ = m/z;
3507 wks.col6.lname$ = Absolute Intensity;
3508 wks.col6.comment$ = "Baselinecorrected Spectrum";
3509 plotxy iy=(1,2) plot:=200 color:=color(255,0,0)
      ogl=[%U]1!;
3510 plotxy iy=(5,6) plot:=200 color:=color(0,0,255)
      ogl=[%U]1!;
3511 win -a %U;
```

Appendix

LabTalk script (OriginPro 2020b)

```
3512 layer.x.from=0;
3513 layer.x.to=%K;
3514 getyesno (Was the baseline correction successful?)
yesno (Baselinecorrection);
3515 if (yesno == 0){
3516 type -b "Sorry, that it didn't work. Shit happens. Try
again.";
3517 win -c;
3518 win -a %B;
3519 win -c;
3520 win -a %V;
3521 };
3522 //Replace old spectrum with new one and finish
3523 if (yesno == 1){
3524 win -c;
3525 win -a %B;
3526 int NoP=count(col(e),1);
3527 copydata irng:=col(5)[1:NoP]
orng:=[% (bb.name$)]Sheet1!col(9)[datastart:dataend];
3528 copydata irng:=col(6)[1:NoP]
orng:=[% (bb.name$)]Sheet1!col(10)[datastart:dataend];
3529 win -c;
3530 win -a % (bb.name$);
3531 win -ia;
3532 win -a %V;
3533 win -z;
3534 page.active = 5;
3535 plot -l;
3536 %L=layer.x.from$;
3537 %M=layer.x.to$;
3538 %N=layer.y.from$;
3539 %O=layer.y.to$;
3540 rescale;
3541 layer.x.from=%L;
3542 layer.x.to=%M;
3543 layer.y.from=%N;
3544 layer.y.to=%O;
3545 page.active$=6;
3546 layer.x.postype=1;
3547 layer.x.position=0;
3548 };
3549 };
3550 //Baseline correction of layer 6
3551 if (page.active == 6) {
3552 //Data acquisition
3553 win -a %B;
3554 wks.col1.format = 1;
3555 wks.col2.format = 1;
3556 win -a % (bb.name$);
3557 int datastart=list((% ([%B]Sheet1, 1, 1)),col(11));
3558 int dataend=list((% ([%B]Sheet1, 1, pp)),col(11));
3559 copydata irng:=[% (bb.name$)]Sheet1!col(11)[datastart:dataend] orng:=[%B]Sheet1!col(3);
3560 copydata irng:=[% (bb.name$)]Sheet1!col(12)[datastart:dataend] orng:=[%B]Sheet1!col(4);
3561 win -a %B;
3562 int aa=count(col(a),1);
3563 %K=col(1)[aa];
3564 pp=aa-1;
3565 //Calculation of baseline corrected spectrum
3566 for(int ii=1; ii<=pp; ii++){
3567 int datastart1=list((% ([%B]Sheet1, 1, ii)),col(3));
3568 int dataend1=list((% ([%B]Sheet1, 1, ii+1)),col(3));
3569 subtract_line iy:=(3,4)[datastart1:dataend1]
x1:=(% ([%B]Sheet1, 1, ii)) y1:=(% ([%B]Sheet1, 2, ii))
x2:=(% ([%B]Sheet1, 1, ii+1)) y2:=(% ([%B]Sheet1, 2, ii+1)) oy:=<new>;
3570 };
3571 for(int ii=1; ii<=pp; ii++){
3572 int NoP=count(col(e),1);
3573 if (exist(col(g)) == 1) {
3574 copydata irng:=(7,8) orng:=(5,6)[NoP];
3575 delete col(g);
3576 delete col(g);
3577 };
3578 };
3579 //Check baseline and calculated spectrum
3580 plotxy iy:=(3,4) plot:=200 color:=color(0,0,0);
3581 %U=%H;
3582 win -a %B;
3583 wks.col1.lname$ = m/z;
3584 wks.col2.lname$ = Absolute Intensity;
3585 wks.col2.comment$ = "Baseline";
3586 wks.col3.lname$ = m/z;
3587 wks.col4.lname$ = Absolute Intensity;
3588 wks.col4.comment$ = "Original Spectrum";
3589 wks.col5.lname$ = m/z;
3590 wks.col6.lname$ = Absolute Intensity;
3591 wks.col6.comment$ = "Baselinecorrected Spectrum";
3592 plotxy iy:=(1,2) plot:=200 color:=color(255,0,0)
ogl:=[%U]1!;
3593 plotxy iy:=(5,6) plot:=200 color:=color(0,0,255)
ogl:=[%U]1!;
3594 win -a %U;
3595 layer.x.from=0;
3596 layer.x.to=%K;
3597 getyesno (Was the baseline correction successful?)
yesno (Baselinecorrection);
3598 if (yesno == 0){
3599 type -b "Sorry, that it didn't work. Shit happens. Try
again.";
3600 win -c;
3601 win -a %B;
3602 win -c;
3603 win -a %V;
3604 };
3605 //Replace old spectrum with new one and finish
3606 if (yesno == 1){
3607 win -c;
3608 win -a %B;
3609 int NoP=count(col(e),1);
3610 copydata irng:=col(5)[1:NoP]
orng:=[% (bb.name$)]Sheet1!col(11)[datastart:dataend];
3611 copydata irng:=col(6)[1:NoP]
orng:=[% (bb.name$)]Sheet1!col(12)[datastart:dataend];
3612 win -c;
3613 win -a % (bb.name$);
3614 win -ia;
3615 win -a %V;
3616 win -z;
3617 page.active = 6;
3618 plot -l;
3619 %L=layer.x.from$;
3620 %M=layer.x.to$;
3621 %N=layer.y.from$;
3622 %O=layer.y.to$;
3623 rescale;
3624 layer.x.from=%L;
3625 layer.x.to=%M;
3626 layer.y.from=%N;
3627 layer.y.to=%O;
3628 layer.x.postype=1;
3629 layer.x.position=0;
3630 };
3631 };
3632 };
3633 };
3634 //Automatic baseline correction
3635 if (PPP == 2) {
3636 //Set variables and values for peakpicking
3637 type -b "This script will calculate and apply a baseline
based on the start- and endpoint of a peak picking area.
By varying the "Maximum Half Width" this area can be
edited to improve the baseline." ;
3638 getn
3639 (Peak picking parameters:) :@G
3640 (Local Points:) LP$
3641 (Minimum Height [absolut value]:) MH$
3642 (Maximum Half Width:) HW$
```

Appendix

LabTalk script (OriginPro 2020b)

```
3643 (-):@G
3644 (Please insert the parameters for the peak picking process.);
3645 %V=%H;
3646 newbook dummy;
3647 %B=%H;
3648 win -a %V;
3649 range -w aa=%C;
3650 range bb = uid2range(aa.GetPage());$;
3651 //Automatic baseline correction of layer 1
3652 if (page.active == 1) {
3653 //Calculation of baseline with peakpicking
3654 pkfind dir:=p method:=max npts:=%(LP$)
    hwidth:=%(HW$) filter:=h value:=%(MH$)
    oleft_x:=[%B]Sheet1!col(3) oleft_y:=[%B]Sheet1!col(4)
    oright_x:=[%B]Sheet1!col(5)
    oright_y:=[%B]Sheet1!col(6) ocenter:=<none>;
3655 win -a %B;
3656 copydata irng:=[%B]Sheet1!col(1)
    orng:=[%B]Sheet1!col(1);
3657 copydata irng:=[%B]Sheet1!col(2)
    orng:=[%B]Sheet1!col(2);
3658 copydata irng:=[%B]Sheet1!col(1)[1]
    orng:=[%B]Sheet1!col(7)[1];
3659 copydata irng:=[%B]Sheet1!col(2)[1]
    orng:=[%B]Sheet1!col(8)[1];
3660 int NoP=count(col(3),1);
3661 kk=2;
3662 for(int ii=2; ii<=NoP; ii++){
3663 copydata irng:=[%B]Sheet1!col(3)[ii]
    orng:=[%B]Sheet1!col(7)[kk];
3664 copydata irng:=[%B]Sheet1!col(4)[ii]
    orng:=[%B]Sheet1!col(8)[kk];
3665 copydata irng:=[%B]Sheet1!col(5)[ii]
    orng:=[%B]Sheet1!col(7)[kk+1];
3666 copydata irng:=[%B]Sheet1!col(6)[ii]
    orng:=[%B]Sheet1!col(8)[kk+1];
3667 kk+2;
3668 };
3669 int ik=count(col(1),1);
3670 int ij=count(col(7),1);
3671 copydata irng:=[%B]Sheet1!col(1)[ik]
    orng:=[%B]Sheet1!col(7)[ij+1];
3672 copydata irng:=[%B]Sheet1!col(2)[ik]
    orng:=[%B]Sheet1!col(8)[ij+1];
3673 wks.col7.format = 1;
3674 wks.col8.format = 1;
3675 int aa=count(col(7),1);
3676 %K=col(7)[aa];
3677 MaVY=max(col(8));
3678 //Check baseline
3679 plotxy iy:=(1,2) plot:=200 color:=color(0,0,0);
3680 %U=%H;
3681 win -a %B;
3682 wks.col1.lname$ = m/z;
3683 wks.col2.lname$ = Absolute Intensity;
3684 wks.col2.comment$ = "Original Spectrum";
3685 wks.col7.lname$ = m/z;
3686 wks.col8.lname$ = Absolute Intensity;
3687 wks.col8.comment$ = "Baseline";
3688 plotxy iy:=(7,8) plot:=200 color:=color(255,0,0)
    ogl:=[%U]1!;
3689 win -a %U;
3690 win -z;
3691 layer.x.from=0;
3692 layer.x.to=%K;
3693 layer.y.from=0;
3694 layer.y.to=MaVY;
3695 speedmode index:=project sm:=off;
3696 dlgChkList inames:="Not good / reset and exit|Not
    good / change parameters|Good, please go on"
    oset:=POP title:="Baseline Calculation" desc:="Was the
    baseline calculation successful? What do you want to
    do now? Check only ONE box" olist:=KB;
3697 if (KB.GetLength() == 1) {
3698 if (POP == 1) {
3699 win -c;
3700 win -a %B;
3701 win -c;
3702 win -a %V;
3703 };
3704 for (int ii=2; ii==POP; ii+0) {
3705 col(3)="";
3706 col(4)="";
3707 col(5)="";
3708 col(6)="";
3709 col(7)="";
3710 col(8)="";
3711 getn
3712 (Peak picking parameters):@G
3713 (Local Points:) LP$
3714 (Minimum Height [absolut value]:) MH$
3715 (Maximum Half Width:) HW$
3716 (-):@G
3717 (Please insert the parameters for the peak picking process.);
3718 pkfind dir:=p method:=max npts:=%(LP$)
    hwidth:=%(HW$) filter:=h value:=%(MH$)
    oleft_x:=[%B]Sheet1!col(3) oleft_y:=[%B]Sheet1!col(4)
    oright_x:=[%B]Sheet1!col(5)
    oright_y:=[%B]Sheet1!col(6) ocenter:=<none>;
3719 win -a %B;
3720 copydata irng:=[%B]Sheet1!col(1)[1]
    orng:=[%B]Sheet1!col(7)[1];
3721 copydata irng:=[%B]Sheet1!col(2)[1]
    orng:=[%B]Sheet1!col(8)[1];
3722 int NoP=count(col(3),1);
3723 kk=2;
3724 for(int ii=2; ii<=NoP; ii++){
3725 copydata irng:=[%B]Sheet1!col(3)[ii]
    orng:=[%B]Sheet1!col(7)[kk];
3726 copydata irng:=[%B]Sheet1!col(4)[ii]
    orng:=[%B]Sheet1!col(8)[kk];
3727 copydata irng:=[%B]Sheet1!col(5)[ii]
    orng:=[%B]Sheet1!col(7)[kk+1];
3728 copydata irng:=[%B]Sheet1!col(6)[ii]
    orng:=[%B]Sheet1!col(8)[kk+1];
3729 kk+2;
3730 };
3731 int ik=count(col(1),1);
3732 int ij=count(col(7),1);
3733 copydata irng:=[%B]Sheet1!col(1)[ik]
    orng:=[%B]Sheet1!col(7)[ij+1];
3734 copydata irng:=[%B]Sheet1!col(2)[ik]
    orng:=[%B]Sheet1!col(8)[ij+1];
3735 wks.col7.format = 1;
3736 wks.col8.format = 1;
3737 int aa=count(col(7),1);
3738 %K=col(7)[aa];
3739 MVY=max(col(8));
3740 win -a %U;
3741 win -z;
3742 layer.x.from=0;
3743 layer.x.to=%K;
3744 layer.y.from=0;
3745 layer.y.to=MVY;
3746 speedmode index:=project sm:=off;
3747 getyesno (Was the baseline calculation successful?)
    yesno (Baselinecorrection);
3748 if (yesno == 0) {
3749 type -b "Alright, try again.";
3750 };
3751 if (yesno == 1) {
3752 POP=3;
3753 };
3754 };
3755 if (POP == 3) {
3756 win -a %B;
3757 range rx = 7;
3758 range ry = 8;
```


Appendix

LabTalk script (OriginPro 2020b)

```
3759 int aa=count(col(7),1);
3760 pp=aa-1;
3761 //Calculation of the baseline-corrected spectrum
3762 for(int ii=1; ii<=pp; ii++){
3763   int datastart1=list(rx[ii],col(1));
3764   int dataend1=list(rx[ii+1],col(1));
3765   subtract_line iy:=(1,2)[datastart1:dataend1] x1:=rx[ii]
   y1:=ry[ii] x2:=rx[ii+1] y2:=ry[ii+1] oy:=<new>;
3766 };
3767 for(int ii=1; ii<=pp; ii++){
3768   int NoP=count(col(9),1);
3769   if (exist(col(K)) == 1) {
3770     copydata irng:=(11,12) orng:=(9,10)[NoP];
3771     delete col(K);
3772     delete col(K);
3773   };
3774 };
3775 //Check baseline and calculated spectrum
3776 plotxy iy:=(9,10) plot:=200 color:=color(0,0,255)
   ogl:=[%U]1! legend:=0 rescale:=0;
3777 win -a %B;
3778 wks.col9.lname$ = m/z;
3779 wks.col10.lname$ = Absolute Intensity;
3780 wks.col10.comment$ = "Baselinecorrected Spectrum";
3781 MiVY=min(col(10));
3782 win -a %U;
3783 win -z;
3784 layer.y.from=MiVY;
3785 getyesno (Was the baseline subtraction successful?)
   yesno (Baselinecorrection);
3786 if (yesno == 0){
3787   type -b "Sorry, that it didn't work. Shit happens. Try
   again.";
3788   win -c;
3789   win -a %B;
3790   win -c;
3791   win -a %V;
3792 };
3793 //Replace old spectrum with new one and finish
3794 if (yesno == 1){
3795   win -c;
3796   win -a %B;
3797   copydata irng:=col(9)
   orng:=[%(bb.name$)]Sheet1!col(1);
3798   copydata irng:=col(10)
   orng:=[%(bb.name$)]Sheet1!col(2);
3799   win -c;
3800   win -a %(bb.name$);
3801   win -ia;
3802   win -a %V;
3803   win -z;
3804   page.active = 1;
3805   plot -l;
3806   %L=layer.x.from$;
3807   %M=layer.x.to$;
3808   %N=layer.y.from$;
3809   %O=layer.y.to$;
3810   rescale;
3811   layer.x.from=%L;
3812   layer.x.to=%M;
3813   layer.y.from=%N;
3814   layer.y.to=%O;
3815   page.active$=6;
3816   layer.x.postype=1;
3817   layer.x.position=0;
3818 };
3819 };
3820 }
3821 else type -b "Please start again and choose only one
   operation you want to perform!";
3822 };
3823 //Automatic baseline correction of layer 2
3824 if (page.active == 2) {
3825   //Calculation of baseline with peakpicking
3826   pkfind dir:=p method:=max npts:=%(LP$)
   hwidth:=%(HW$) filter:=h value:=%(MH$)
   oleft_x:=[%B]Sheet1!col(3) oleft_y:=[%B]Sheet1!col(4)
   oright_x:=[%B]Sheet1!col(5)
   oright_y:=[%B]Sheet1!col(6) ocenter:=<none>;
3827   win -a %B;
3828   copydata irng:=[%(bb.name$)]Sheet1!col(3)
   orng:=[%B]Sheet1!col(1);
3829   copydata irng:=[%(bb.name$)]Sheet1!col(4)
   orng:=[%B]Sheet1!col(2);
3830   copydata irng:=[%B]Sheet1!col(1)[1]
   orng:=[%B]Sheet1!col(7)[1];
3831   copydata irng:=[%B]Sheet1!col(2)[1]
   orng:=[%B]Sheet1!col(8)[1];
3832   int NoP=count(col(3),1);
3833   kk=2;
3834   for(int ii=2; ii<=NoP; ii++){
3835     copydata irng:=[%B]Sheet1!col(3)[ii]
     orng:=[%B]Sheet1!col(7)[kk];
3836     copydata irng:=[%B]Sheet1!col(4)[ii]
     orng:=[%B]Sheet1!col(8)[kk];
3837     copydata irng:=[%B]Sheet1!col(5)[ii]
     orng:=[%B]Sheet1!col(7)[kk+1];
3838     copydata irng:=[%B]Sheet1!col(6)[ii]
     orng:=[%B]Sheet1!col(8)[kk+1];
3839     kk+2;
3840   };
3841   int ik=count(col(1),1);
3842   int ij=count(col(7),1);
3843   copydata irng:=[%B]Sheet1!col(1)[ik]
   orng:=[%B]Sheet1!col(7)[ij+1];
3844   copydata irng:=[%B]Sheet1!col(2)[ik]
   orng:=[%B]Sheet1!col(8)[ij+1];
3845   wks.col7.format = 1;
3846   wks.col8.format = 1;
3847   int aa=count(col(7),1);
3848   %K=col(7)[aa];
3849   MaVY=max(col(8));
3850   //Check baseline
3851   plotxy iy:=(1,2) plot:=200 color:=color(0,0,0);
3852   %U=%H;
3853   win -a %B;
3854   wks.col1.lname$ = m/z;
3855   wks.col2.lname$ = Absolute Intensity;
3856   wks.col2.comment$ = "Original Spectrum";
3857   wks.col7.lname$ = m/z;
3858   wks.col8.lname$ = Absolute Intensity;
3859   wks.col8.comment$ = "Baseline";
3860   plotxy iy:=(7,8) plot:=200 color:=color(255,0,0)
   ogl:=[%U]1!;
3861   win -a %U;
3862   win -z;
3863   layer.x.from=0;
3864   layer.x.to=%K;
3865   layer.y.from=0;
3866   layer.y.to=MaVY;
3867   speedmode index:=project sm:=off;
3868   dlgChkList inames:="Not good / reset and exit|Not
   good / change parameters|Good, please go on"
   oset:=POP title:="Baseline Calculation" desc:="Was the
   baseline calculation successful? What do you want to
   do now? Check only ONE box" olist:=KB;
3869   if (KB.GetLength() == 1) {
3870     if (POP == 1) {
3871       win -c;
3872       win -a %B;
3873       win -c;
3874       win -a %V;
3875     };
3876     for (int ii=2; ii==POP; ii+0) {
3877       col(3)="",
3878       col(4)="",
3879       col(5)="",
3880       col(6)="",
3881       col(7)="";
```


Appendix

LabTalk script (OriginPro 2020b)

```
3882 col(8)="";
3883 getn
3884 (Peak picking parameters:) :@G
3885 (Local Points:) LP$
3886 (Minimum Height [absolut value:] ) MH$
3887 (Maximum Half Width:) HW$
3888 (-) :@G
3889 (Please insert the parameters for the peak picking process.);
3890 pkfind dir:=p method:=max npts:=%(LP$)
hwidth:=%(HW$) filter:=h value:=%(MH$)
oleft_x:=[%B]Sheet1!col(3) oleft_y:=[%B]Sheet1!col(4)
oright_x:=[%B]Sheet1!col(5)
oright_y:=[%B]Sheet1!col(6) ocenter:=<none>;
3891 win -a %B;
3892 copydata irng:=[%B]Sheet1!col(1)[1]
orng:=[%B]Sheet1!col(7)[1];
3893 copydata irng:=[%B]Sheet1!col(2)[1]
orng:=[%B]Sheet1!col(8)[1];
3894 int NoP=count(col(3),1);
3895 kk=2;
3896 for(int ii=2; ii<=NoP; ii++){
3897 copydata irng:=[%B]Sheet1!col(3)[ii]
orng:=[%B]Sheet1!col(7)[kk];
3898 copydata irng:=[%B]Sheet1!col(4)[ii]
orng:=[%B]Sheet1!col(8)[kk];
3899 copydata irng:=[%B]Sheet1!col(5)[ii]
orng:=[%B]Sheet1!col(7)[kk+1];
3900 copydata irng:=[%B]Sheet1!col(6)[ii]
orng:=[%B]Sheet1!col(8)[kk+1];
3901 kk+2;
3902 };
3903 int ik=count(col(1),1);
3904 int ij=count(col(7),1);
3905 copydata irng:=[%B]Sheet1!col(1)[ik]
orng:=[%B]Sheet1!col(7)[ij+1];
3906 copydata irng:=[%B]Sheet1!col(2)[ik]
orng:=[%B]Sheet1!col(8)[ij+1];
3907 wks.col7.format = 1;
3908 wks.col8.format = 1;
3909 int aa=count(col(7),1);
3910 %K=col(7)[aa];
3911 MVY=max(col(8));
3912 win -a %U;
3913 win -z;
3914 layer.x.from=0;
3915 layer.x.to=%K;
3916 layer.y.from=0;
3917 layer.y.to=MVY;
3918 speedmode index:=project sm:=off;
3919 getyesno (Was the baseline calculation successful?)
yesno (Baselinecorrection);
3920 if (yesno == 0) {
3921 type -b "Alright, try again.";
3922 };
3923 if(yesno == 1) {
3924 POP=3;
3925 };
3926 };
3927 if (POP == 3) {
3928 win -a %B;
3929 range rx = 7;
3930 range ry = 8;
3931 int aa=count(col(7),1);
3932 pp=aa-1;
3933 //Calculation of the baseline-corrected spectrum
3934 for(int ii=1; ii<=pp; ii++){
3935 int datastart1=list(rx[ii],col(1));
3936 int dataend1=list(rx[ii+1],col(1));
3937 subtract_line iy:=(1,2)[datastart1:dataend1] x1:=(rx[ii]
y1:=ry[ii] x2:=(rx[ii+1] y2:=ry[ii+1] oy:=<new>;
3938 };
3939 for(int ii=1; ii<=pp; ii++){
3940 int NoP=count(col(9),1);
3941 if (exist(col(K)) == 1) {
3942 copydata irng:=(11,12) orng:=(9,10)[NoP];
3943 delete col(K);
3944 delete col(K);
3945 };
3946 };
3947 //Check baseline and calculated spectrum
3948 plotxy iy:=(9,10) plot:=200 color:=color(0,0,255)
ogl:=[%U]1! legend:=0 rescale:=0;
3949 win -a %B;
3950 wks.col9.lname$ = m/z;
3951 wks.col10.lname$ = Absolute Intensity;
3952 wks.col10.comment$ = "Baselinecorrected Spectrum";
3953 MiVY=min(col(10));
3954 win -a %U;
3955 win -z;
3956 layer.y.from=MiVY;
3957 getyesno (Was the baseline subtraction successful?)
yesno (Baselinecorrection);
3958 if (yesno == 0) {
3959 type -b "Sorry, that it didn't work. Shit happens. Try
again.";
3960 win -c;
3961 win -a %B;
3962 win -c;
3963 win -a %V;
3964 };
3965 //Replace old spectrum with new one and finish
3966 if (yesno == 1){
3967 win -c;
3968 win -a %B;
3969 copydata irng:=col(9)
orng:=[% (bb.name$)]Sheet1!col(3);
3970 copydata irng:=col(10)
orng:=[% (bb.name$)]Sheet1!col(4);
3971 win -c;
3972 win -a % (bb.name$);
3973 win -ia;
3974 win -a %V;
3975 win -z;
3976 page.active = 2;
3977 plot -I;
3978 %L=layer.x.from$;
3979 %M=layer.x.to$;
3980 %N=layer.y.from$;
3981 %O=layer.y.to$;
3982 rescale;
3983 layer.x.from=%L;
3984 layer.x.to=%M;
3985 layer.y.from=%N;
3986 layer.y.to=%O;
3987 page.active$=6;
3988 layer.x.posttype=1;
3989 layer.x.position=0;
3990 };
3991 };
3992 }
3993 else type -b "Please start again and choose only one
operation you want to perform!";
3994 };
3995 //Automatic baseline correction of layer 3
3996 if (page.active == 3) {
3997 //Calculation of baseline with peakpicking
3998 pkfind dir:=p method:=max npts:=%(LP$)
hwidth:=%(HW$) filter:=h value:=%(MH$)
oleft_x:=[%B]Sheet1!col(3) oleft_y:=[%B]Sheet1!col(4)
oright_x:=[%B]Sheet1!col(5)
oright_y:=[%B]Sheet1!col(6) ocenter:=<none>;
3999 win -a %B;
4000 copydata irng:=[% (bb.name$)]Sheet1!col(5)
orng:=[%B]Sheet1!col(1);
4001 copydata irng:=[% (bb.name$)]Sheet1!col(6)
orng:=[%B]Sheet1!col(2);
4002 copydata irng:=[%B]Sheet1!col(1)[1]
orng:=[%B]Sheet1!col(7)[1];
```

Appendix

LabTalk script (OriginPro 2020b)

```
4003 copydata irng:=[%B]Sheet1!col(2)[1]
    orng:=[%B]Sheet1!col(8)[1];
4004 int NoP=count(col(3),1);
4005 kk=2;
4006 for(int ii=2; ii<=NoP; ii++){
4007 copydata irng:=[%B]Sheet1!col(3)[ii]
    orng:=[%B]Sheet1!col(7)[kk];
4008 copydata irng:=[%B]Sheet1!col(4)[ii]
    orng:=[%B]Sheet1!col(8)[kk];
4009 copydata irng:=[%B]Sheet1!col(5)[ii]
    orng:=[%B]Sheet1!col(7)[kk+1];
4010 copydata irng:=[%B]Sheet1!col(6)[ii]
    orng:=[%B]Sheet1!col(8)[kk+1];
4011 kk+2;
4012 };
4013 int ik=count(col(1),1);
4014 int ij=count(col(7),1);
4015 copydata irng:=[%B]Sheet1!col(1)[ik]
    orng:=[%B]Sheet1!col(7)[ij+1];
4016 copydata irng:=[%B]Sheet1!col(2)[ik]
    orng:=[%B]Sheet1!col(8)[ij+1];
4017 wks.col7.format = 1;
4018 wks.col8.format = 1;
4019 int aa=count(col(7),1);
4020 %K=col(7)[aa];
4021 MaVY=max(col(8));
4022 //Check baseline
4023 plotxy iy:=(1,2) plot:=200 color:=color(0,0,0);
4024 %U=%H;
4025 win -a %B;
4026 wks.col1.Iname$ = m/z;
4027 wks.col2.Iname$ = Absolute Intensity;
4028 wks.col2.comment$ = "Original Spectrum";
4029 wks.col7.Iname$ = m/z;
4030 wks.col8.Iname$ = Absolute Intensity;
4031 wks.col8.comment$ = "Baseline";
4032 plotxy iy:=(7,8) plot:=200 color:=color(255,0,0)
    ogl:=[%U]1!;
4033 win -a %U;
4034 win -z;
4035 layer.x.from=0;
4036 layer.x.to=%K;
4037 layer.y.from=0;
4038 layer.y.to=MaVY;
4039 speedmode index:=project sm:=off;
4040 dlgChkList inames:="Not good / reset and exit|Not
    good / change parameters|Good, please go on"
    osel:=POP title:="Baseline Calculation" desc:="Was the
    baseline calculation successful? What do you want to
    do now? Check only ONE box" olist:=KB;
4041 if (KB.GetLength() == 1) {
4042 if (POP == 1) {
4043 win -c;
4044 win -a %B;
4045 win -c;
4046 win -a %V;
4047 };
4048 for (int ii=2; ii==POP; ii+0) {
4049 col(3)="";
4050 col(4)="";
4051 col(5)="";
4052 col(6)="";
4053 col(7)="";
4054 col(8)="";
4055 getn
4056 (Peak picking parameters:) :@G
4057 (Local Points:) LP$
4058 (Minimum Height [absolut value:] ) MH$
4059 (Maximum Half Width:) HW$
4060 (-) :@G
4061 (Please insert the parameters for the peak picking pro-
    cess.);
4062 pkfind dir:=p method:=max npts:=%(LP$)
    hwidth:=%(HW$) filter:=h value:=%(MH$)
    oleft_x:=[%B]Sheet1!col(3) oleft_y:=[%B]Sheet1!col(4)
    oright_x:=[%B]Sheet1!col(5)
    oright_y:=[%B]Sheet1!col(6) ocenter:=<none>;
4063 win -a %B;
4064 copydata irng:=[%B]Sheet1!col(1)[1]
    orng:=[%B]Sheet1!col(7)[1];
4065 copydata irng:=[%B]Sheet1!col(2)[1]
    orng:=[%B]Sheet1!col(8)[1];
4066 int NoP=count(col(3),1);
4067 kk=2;
4068 for(int ii=2; ii<=NoP; ii++){
4069 copydata irng:=[%B]Sheet1!col(3)[ii]
    orng:=[%B]Sheet1!col(7)[kk];
4070 copydata irng:=[%B]Sheet1!col(4)[ii]
    orng:=[%B]Sheet1!col(8)[kk];
4071 copydata irng:=[%B]Sheet1!col(5)[ii]
    orng:=[%B]Sheet1!col(7)[kk+1];
4072 copydata irng:=[%B]Sheet1!col(6)[ii]
    orng:=[%B]Sheet1!col(8)[kk+1];
4073 kk+2;
4074 };
4075 int ik=count(col(1),1);
4076 int ij=count(col(7),1);
4077 copydata irng:=[%B]Sheet1!col(1)[ik]
    orng:=[%B]Sheet1!col(7)[ij+1];
4078 copydata irng:=[%B]Sheet1!col(2)[ik]
    orng:=[%B]Sheet1!col(8)[ij+1];
4079 wks.col7.format = 1;
4080 wks.col8.format = 1;
4081 int aa=count(col(7),1);
4082 %K=col(7)[aa];
4083 MVY=max(col(8));
4084 win -a %U;
4085 win -z;
4086 layer.x.from=0;
4087 layer.x.to=%K;
4088 layer.y.from=0;
4089 layer.y.to=MVY;
4090 speedmode index:=project sm:=off;
4091 getyesno (Was the baseline calculation successful?)
    yesno (Baseline correction);
4092 if (yesno == 0) {
4093 type -b "Alright, try again.";
4094 };
4095 if (yesno == 1) {
4096 POP=3;
4097 };
4098 };
4099 if (POP == 3) {
4100 win -a %B;
4101 range rx = 7;
4102 range ry = 8;
4103 int aa=count(col(7),1);
4104 pp=aa-1;
4105 //Calculation of the baseline-corrected spectrum
4106 for(int ii=1; ii<=pp; ii++){
4107 int datastart1=list(rx[ii],col(1));
4108 int dataend1=list(rx[ii+1],col(1));
4109 subtract_line iy:=(1,2)[datastart1:dataend1] x1:=rx[ii]
    y1:=ry[ii] x2:=rx[ii+1] y2:=ry[ii+1] oy:=<new>;
4110 };
4111 for(int ii=1; ii<=pp; ii++){
4112 int NoP=count(col(9),1);
4113 if (exist(col(K)) == 1) {
4114 copydata irng:=(11,12) orng:=(9,10)[NoP];
4115 delete col(K);
4116 delete col(K);
4117 };
4118 };
4119 //Check baseline and calculated spectrum
4120 plotxy iy:=(9,10) plot:=200 color:=color(0,0,255)
    ogl:=[%U]1! legend:=0 rescale:=0;
4121 win -a %B;
4122 wks.col9.Iname$ = m/z;
4123 wks.col10.Iname$ = Absolute Intensity;
4124 wks.col10.comment$ = "Baselinecorrected Spectrum";
```

Appendix

LabTalk script (OriginPro 2020b)

```
4125 MiVY=min(col(10));
4126 win -a %U;
4127 win -z;
4128 layer.y.from=MiVY;
4129 getyesno (Was the baseline subtraction successful?)
yesno (Baselinecorrection);
4130 if (yesno == 0){
4131 type -b "Sorry, that it didn't work. Shit happens. Try
again.";
4132 win -c;
4133 win -a %B;
4134 win -c;
4135 win -a %V;
4136 };
4137 //Replace old spectrum with new one and finish
4138 if (yesno == 1){
4139 win -c;
4140 win -a %B;
4141 copydata irng:=col(9)
orng:=[%B]Sheet1!col(5);
4142 copydata irng:=col(10)
orng:=[%B]Sheet1!col(6);
4143 win -c;
4144 win -a %B;
4145 win -ia;
4146 win -a %V;
4147 win -z;
4148 page.active = 3;
4149 plot -l;
4150 %L=layer.x.from$;
4151 %M=layer.x.to$;
4152 %N=layer.y.from$;
4153 %O=layer.y.to$;
4154 rescale;
4155 layer.x.from=%L;
4156 layer.x.to=%M;
4157 layer.y.from=%N;
4158 layer.y.to=%O;
4159 page.active$=6;
4160 layer.x.posttype=1;
4161 layer.x.position=0;
4162 };
4163 };
4164 }
4165 else type -b "Please start again and choose only one
operation you want to perform!";
4166 };
4167 //Automatic baseline correction of layer 4
4168 if (page.active == 4) {
4169 //Calculation of baseline with peakpicking
4170 pkfind dir:=p method:=max npts:=%(LP$)
hwidth:=%(HW$) filter:=h value:=%(MH$)
oleft_x:=[%B]Sheet1!col(3) oleft_y:=[%B]Sheet1!col(4)
oright_x:=[%B]Sheet1!col(5)
oright_y:=[%B]Sheet1!col(6) ocenter:=<none>;
4171 win -a %B;
4172 copydata irng:=[%B]Sheet1!col(7)
orng:=[%B]Sheet1!col(1);
4173 copydata irng:=[%B]Sheet1!col(8)
orng:=[%B]Sheet1!col(2);
4174 copydata irng:=[%B]Sheet1!col(1)[1]
orng:=[%B]Sheet1!col(7)[1];
4175 copydata irng:=[%B]Sheet1!col(2)[1]
orng:=[%B]Sheet1!col(8)[1];
4176 int NoP=count(col(3),1);
4177 kk=2;
4178 for(int ii=2; ii<=NoP; ii++){
4179 copydata irng:=[%B]Sheet1!col(3)[ii]
orng:=[%B]Sheet1!col(7)[kk];
4180 copydata irng:=[%B]Sheet1!col(4)[ii]
orng:=[%B]Sheet1!col(8)[kk];
4181 copydata irng:=[%B]Sheet1!col(5)[ii]
orng:=[%B]Sheet1!col(7)[kk+1];
4182 copydata irng:=[%B]Sheet1!col(6)[ii]
orng:=[%B]Sheet1!col(8)[kk+1];
4183 kk+2;
4184 };
4185 int ik=count(col(1),1);
4186 int ij=count(col(7),1);
4187 copydata irng:=[%B]Sheet1!col(1)[ik]
orng:=[%B]Sheet1!col(7)[ij+1];
4188 copydata irng:=[%B]Sheet1!col(2)[ik]
orng:=[%B]Sheet1!col(8)[ij+1];
4189 wks.col7.format = 1;
4190 wks.col8.format = 1;
4191 int aa=count(col(7),1);
4192 %K=col(7)[aa];
4193 MaVY=max(col(8));
4194 //Check baseline
4195 plotxy iy:=(1,2) plot:=200 color:=color(0,0,0);
4196 %U=%H;
4197 win -a %B;
4198 wks.col1.lname$ = m/z;
4199 wks.col2.lname$ = Absolute Intensity;
4200 wks.col2.comment$ = "Original Spectrum";
4201 wks.col7.lname$ = m/z;
4202 wks.col8.lname$ = Absolute Intensity;
4203 wks.col8.comment$ = "Baseline";
4204 plotxy iy:=(7,8) plot:=200 color:=color(255,0,0)
ogl:=[%U]1!;
4205 win -a %U;
4206 win -z;
4207 layer.x.from=0;
4208 layer.x.to=%K;
4209 layer.y.from=0;
4210 layer.y.to=MaVY;
4211 speedmode index:=project sm:=off;
4212 dlgChkList inames:="Not good / reset and exit|Not
good / change parameters|Good, please go on"
osel:=POP title:="Baseline Calculation" desc:="Was the
baseline calculation successful? What do you want to
do now? Check only ONE box" olist=KB;
4213 if (KB.GetLength() == 1) {
4214 if (POP == 1) {
4215 win -c;
4216 win -a %B;
4217 win -c;
4218 win -a %V;
4219 };
4220 for (int ii=2; ii==POP; ii+0) {
4221 col(3)="",
4222 col(4)="",
4223 col(5)="",
4224 col(6)="",
4225 col(7)="",
4226 col(8)="",
4227 getn
4228 (Peak picking parameters:) :@G
4229 (Local Points:) LP$
4230 (Minimum Height [absolut value]:) MH$
4231 (Maximum Half Width:) HW$
4232 (-) :@G
4233 (Please insert the parameters for the peak picking pro-
cess.);
4234 pkfind dir:=p method:=max npts:=%(LP$)
hwidth:=%(HW$) filter:=h value:=%(MH$)
oleft_x:=[%B]Sheet1!col(3) oleft_y:=[%B]Sheet1!col(4)
oright_x:=[%B]Sheet1!col(5)
oright_y:=[%B]Sheet1!col(6) ocenter:=<none>;
4235 win -a %B;
4236 copydata irng:=[%B]Sheet1!col(1)[1]
orng:=[%B]Sheet1!col(7)[1];
4237 copydata irng:=[%B]Sheet1!col(2)[1]
orng:=[%B]Sheet1!col(8)[1];
4238 int NoP=count(col(3),1);
4239 kk=2;
4240 for(int ii=2; ii<=NoP; ii++){
4241 copydata irng:=[%B]Sheet1!col(3)[ii]
orng:=[%B]Sheet1!col(7)[kk];
```

Appendix

LabTalk script (OriginPro 2020b)

```
4242 copydata irng:=[%B]Sheet1!col(4)[ii]
    org:=[%B]Sheet1!col(8)[kk];
4243 copydata irng:=[%B]Sheet1!col(5)[ii]
    org:=[%B]Sheet1!col(7)[kk+1];
4244 copydata irng:=[%B]Sheet1!col(6)[ii]
    org:=[%B]Sheet1!col(8)[kk+1];
4245 kk+2;
4246 };
4247 int ik=count(col(1),1);
4248 int ij=count(col(7),1);
4249 copydata irng:=[%B]Sheet1!col(1)[ik]
    org:=[%B]Sheet1!col(7)[ij+1];
4250 copydata irng:=[%B]Sheet1!col(2)[ik]
    org:=[%B]Sheet1!col(8)[ij+1];
4251 wks.col7.format = 1;
4252 wks.col8.format = 1;
4253 int aa=count(col(7),1);
4254 %K=col(7)[aa];
4255 MVY=max(col(8));
4256 win -a %U;
4257 win -z;
4258 layer.x.from=0;
4259 layer.x.to=%K;
4260 layer.y.from=0;
4261 layer.y.to=MVY;
4262 speedmode index:=project sm:=off;
4263 getyesno (Was the baseline calculation successful?)
    yesno (Baselinecorrection);
4264 if (yesno == 0) {
4265     type -b "Alright, try again.";
4266 };
4267 if(yesno == 1) {
4268     POP=3;
4269 };
4270 };
4271 if (POP == 3) {
4272     win -a %B;
4273     range rx = 7;
4274     range ry = 8;
4275     int aa=count(col(7),1);
4276     pp=aa-1;
4277     //Calculation of the baseline-corrected spectrum
4278     for(int ii=1; ii<=pp; ii++){
4279         int datastart1=list(rx[ii],col(1));
4280         int dataend1=list(rx[ii+1],col(1));
4281         subtract_line iy:=(1,2)[datastart1:dataend1] x1:rx[ii]
            y1:ry[ii] x2:rx[ii+1] y2:ry[ii+1] oy:=<new>;
4282     };
4283     for(int ii=1; ii<=pp; ii++){
4284         int NoP=count(col(9),1);
4285         if (exist(col(K)) == 1) {
4286             copydata irng:=(11,12) org:=(9,10)[NoP];
4287             delete col(K);
4288             delete col(K);
4289         };
4290     };
4291     //Check baseline and calculated spectrum
4292     plotxy iy:=(9,10) plot:=200 color:=color(0,0,255)
        ogl:=[%U]1! legend:=0 rescale:=0;
4293     win -a %B;
4294     wks.col9.lname$ = m/z;
4295     wks.col10.lname$ = Absolute Intensity;
4296     wks.col10.comment$ = "Baselinecorrected Spectrum";
4297     MiVY=min(col(10));
4298     win -a %U;
4299     win -z;
4300     layer.y.from=MiVY;
4301     getyesno (Was the baseline subtraction successful?)
        yesno (Baselinecorrection);
4302     if (yesno == 0){
4303         type -b "Sorry, that it didn't work. Shit happens. Try
            again.";
4304         win -c;
4305         win -a %B;
4306         win -c;
4307         win -a %V;
4308     };
4309     //Replace old spectrum with new one and finish
4310     if (yesno == 1){
4311         win -c;
4312         win -a %B;
4313         copydata irng:=col(9)
            org:=[%B]Sheet1!col(7);
4314         copydata irng:=col(10)
            org:=[%B]Sheet1!col(8);
4315         win -c;
4316         win -a %B;
4317         win -ia;
4318         win -a %V;
4319         win -z;
4320         page.active = 4;
4321         plot -l;
4322         %L=layer.x.from$;
4323         %M=layer.x.to$;
4324         %N=layer.y.from$;
4325         %O=layer.y.to$;
4326         rescale;
4327         layer.x.from=%L;
4328         layer.x.to=%M;
4329         layer.y.from=%N;
4330         layer.y.to=%O;
4331         page.active$=6;
4332         layer.x.posttype=1;
4333         layer.x.position=0;
4334     };
4335 };
4336 }
4337 else type -b "Please start again and choose only one
    operation you want to perform!";
4338 };
4339 //Automatic baseline correction of layer 5
4340 if (page.active == 5) {
4341     //Calculation of baseline with peakpicking
4342     pkfind dir:=p method:=max npts:=%(LP$)
        hwidth:=%(HW$) filter:=h value:=%(MH$)
        oleft_x:=[%B]Sheet1!col(3) oleft_y:=[%B]Sheet1!col(4)
        oright_x:=[%B]Sheet1!col(5)
        oright_y:=[%B]Sheet1!col(6) ocenter:=<none>;
4343     win -a %B;
4344     copydata irng:=[%B]Sheet1!col(9)
        org:=[%B]Sheet1!col(1);
4345     copydata irng:=[%B]Sheet1!col(10)
        org:=[%B]Sheet1!col(2);
4346     copydata irng:=[%B]Sheet1!col(1)[1]
        org:=[%B]Sheet1!col(7)[1];
4347     copydata irng:=[%B]Sheet1!col(2)[1]
        org:=[%B]Sheet1!col(8)[1];
4348     int NoP=count(col(3),1);
4349     kk=2;
4350     for(int ii=2; ii<=NoP; ii++){
4351         copydata irng:=[%B]Sheet1!col(3)[ii]
            org:=[%B]Sheet1!col(7)[kk];
4352         copydata irng:=[%B]Sheet1!col(4)[ii]
            org:=[%B]Sheet1!col(8)[kk];
4353         copydata irng:=[%B]Sheet1!col(5)[ii]
            org:=[%B]Sheet1!col(7)[kk+1];
4354         copydata irng:=[%B]Sheet1!col(6)[ii]
            org:=[%B]Sheet1!col(8)[kk+1];
4355         kk+2;
4356     };
4357     int ik=count(col(1),1);
4358     int ij=count(col(7),1);
4359     copydata irng:=[%B]Sheet1!col(1)[ik]
        org:=[%B]Sheet1!col(7)[ij+1];
4360     copydata irng:=[%B]Sheet1!col(2)[ik]
        org:=[%B]Sheet1!col(8)[ij+1];
4361     wks.col7.format = 1;
4362     wks.col8.format = 1;
4363     int aa=count(col(7),1);
4364     %K=col(7)[aa];
```


Appendix

LabTalk script (OriginPro 2020b)

```
4365 MaVY=max(col(8));
4366 //Check baseline
4367 plotxy iy:=(1,2) plot:=200 color:=color(0,0,0);
4368 %U=%H;
4369 win -a %B;
4370 wks.col1.Iname$ = m/z;
4371 wks.col2.Iname$ = Absolute Intensity;
4372 wks.col2.comment$ = "Original Spectrum";
4373 wks.col7.Iname$ = m/z;
4374 wks.col8.Iname$ = Absolute Intensity;
4375 wks.col8.comment$ = "Baseline";
4376 plotxy iy:=(7,8) plot:=200 color:=color(255,0,0)
ogl:=[%U]1!;
4377 win -a %U;
4378 win -z;
4379 layer.x.from=0;
4380 layer.x.to=%K;
4381 layer.y.from=0;
4382 layer.y.to=MaVY;
4383 speedmode index:=project sm:=off;
4384 dlgChkList inames:="Not good / reset and exit|Not
good / change parameters|Good, please go on"
osel:=POP title:="Baseline Calculation" desc:="Was the
baseline calculation successful? What do you want to
do now? Check only ONE box" olist:=KB;
4385 if (KB.GetLength() == 1) {
4386 if (POP == 1) {
4387 win -c;
4388 win -a %B;
4389 win -c;
4390 win -a %V;
4391 };
4392 for (int ii=2; ii==POP; ii++) {
4393 col(3)="";
4394 col(4)="";
4395 col(5)="";
4396 col(6)="";
4397 col(7)="";
4398 col(8)="";
4399 getn
4400 (Peak picking parameters:) :@G
4401 (Local Points:) LP$
4402 (Minimum Height [absolut value:] ) MH$
4403 (Maximum Half Width:) HW$
4404 (-) :@G
4405 (Please insert the parameters for the peak picking pro-
cess.);
4406 pkfind dir:=p method:=max npts:=%(LP$)
hwidth:=%(HW$) filter:=h value:=%(MH$)
oleft_x:=[%B]Sheet1!col(3) oleft_y:=[%B]Sheet1!col(4)
oright_x:=[%B]Sheet1!col(5)
oright_y:=[%B]Sheet1!col(6) ocenter:=<none>;
4407 win -a %B;
4408 copydata irng:=[%B]Sheet1!col(1)[1]
ormg:=[%B]Sheet1!col(7)[1];
4409 copydata irng:=[%B]Sheet1!col(2)[1]
ormg:=[%B]Sheet1!col(8)[1];
4410 int NoP=count(col(3),1);
4411 kk=2;
4412 for(int ii=2; ii<=NoP; ii++){
4413 copydata irng:=[%B]Sheet1!col(3)[ii]
ormg:=[%B]Sheet1!col(7)[kk];
4414 copydata irng:=[%B]Sheet1!col(4)[ii]
ormg:=[%B]Sheet1!col(8)[kk];
4415 copydata irng:=[%B]Sheet1!col(5)[ii]
ormg:=[%B]Sheet1!col(7)[kk+1];
4416 copydata irng:=[%B]Sheet1!col(6)[ii]
ormg:=[%B]Sheet1!col(8)[kk+1];
4417 kk+2;
4418 };
4419 int ik=count(col(1),1);
4420 int ij=count(col(7),1);
4421 copydata irng:=[%B]Sheet1!col(1)[ik]
ormg:=[%B]Sheet1!col(7)[ij+1];
4422 copydata irng:=[%B]Sheet1!col(2)[ik]
ormg:=[%B]Sheet1!col(8)[ij+1];
4423 wks.col7.format = 1;
4424 wks.col8.format = 1;
4425 int aa=count(col(7),1);
4426 %K=col(7)[aa];
4427 MVY=max(col(8));
4428 win -a %U;
4429 win -z;
4430 layer.x.from=0;
4431 layer.x.to=%K;
4432 layer.y.from=0;
4433 layer.y.to=MVY;
4434 speedmode index:=project sm:=off;
4435 getyesno (Was the baseline calculation successful?)
yesno (Baselinecorrection);
4436 if (yesno == 0) {
4437 type -b "Alright, try again.";
4438 };
4439 if(yesno == 1) {
4440 POP=3;
4441 };
4442 };
4443 if (POP == 3) {
4444 win -a %B;
4445 range rx = 7;
4446 range ry = 8;
4447 int aa=count(col(7),1);
4448 pp=aa-1;
4449 //Calculation of the baseline-corrected spectrum
4450 for(int ii=1; ii<=pp; ii++){
4451 int datastart1=list(rx[ii],col(1));
4452 int dataend1=list(rx[ii+1],col(1));
4453 subtract_line iy:=(1,2)[datastart1:dataend1] x1:=rx[ii]
y1:=ry[ii] x2:=rx[ii+1] y2:=ry[ii+1] oy:=<new>;
4454 };
4455 for(int ii=1; ii<=pp; ii++){
4456 int NoP=count(col(9),1);
4457 if (exist(col(K)) == 1) {
4458 copydata irng:=(11,12) ormng:=(9,10)[NoP];
4459 delete col(K);
4460 delete col(K);
4461 };
4462 };
4463 //Check baseline and calculated spectrum
4464 plotxy iy:=(9,10) plot:=200 color:=color(0,0,255)
ogl:=[%U]1! legend:=0 rescale:=0;
4465 win -a %B;
4466 wks.col9.Iname$ = m/z;
4467 wks.col10.Iname$ = Absolute Intensity;
4468 wks.col10.comment$ = "Baselinecorrected Spectrum";
4469 MiVY=min(col(10));
4470 win -a %U;
4471 win -z;
4472 layer.y.from=MiVY;
4473 getyesno (Was the baseline subtraction successful?)
yesno (Baselinecorrection);
4474 if (yesno == 0){
4475 type -b "Sorry, that it didn't work. Shit happens. Try
again.";
4476 win -c;
4477 win -a %B;
4478 win -c;
4479 win -a %V;
4480 };
4481 //Replace old spectrum with new one and finish
4482 if (yesno == 1){
4483 win -c;
4484 win -a %B;
4485 copydata irng:=col(9)
ormng:=[%B]Sheet1!col(9);
4486 copydata irng:=col(10)
ormng:=[%B]Sheet1!col(10);
4487 win -c;
4488 win -a %B;
4489 win -a %B;
4490 win -a %B;
4491 win -a %B;
4492 win -a %B;
4493 win -a %B;
4494 win -a %B;
4495 win -a %B;
4496 win -a %B;
4497 win -a %B;
4498 win -a %B;
4499 win -a %B;
4500 win -a %B;
4501 win -a %B;
4502 win -a %B;
4503 win -a %B;
4504 win -a %B;
4505 win -a %B;
4506 win -a %B;
4507 win -a %B;
4508 win -a %B;
4509 win -a %B;
4510 win -a %B;
4511 win -a %B;
4512 win -a %B;
4513 win -a %B;
4514 win -a %B;
4515 win -a %B;
4516 win -a %B;
4517 win -a %B;
4518 win -a %B;
4519 win -a %B;
4520 win -a %B;
4521 win -a %B;
4522 win -a %B;
4523 win -a %B;
4524 win -a %B;
4525 win -a %B;
4526 win -a %B;
4527 win -a %B;
4528 win -a %B;
4529 win -a %B;
4530 win -a %B;
4531 win -a %B;
4532 win -a %B;
4533 win -a %B;
4534 win -a %B;
4535 win -a %B;
4536 win -a %B;
4537 win -a %B;
4538 win -a %B;
4539 win -a %B;
4540 win -a %B;
4541 win -a %B;
4542 win -a %B;
4543 win -a %B;
4544 win -a %B;
4545 win -a %B;
4546 win -a %B;
4547 win -a %B;
4548 win -a %B;
4549 win -a %B;
4550 win -a %B;
4551 win -a %B;
4552 win -a %B;
4553 win -a %B;
4554 win -a %B;
4555 win -a %B;
4556 win -a %B;
4557 win -a %B;
4558 win -a %B;
4559 win -a %B;
4560 win -a %B;
4561 win -a %B;
4562 win -a %B;
4563 win -a %B;
4564 win -a %B;
4565 win -a %B;
4566 win -a %B;
4567 win -a %B;
4568 win -a %B;
4569 win -a %B;
4570 win -a %B;
4571 win -a %B;
4572 win -a %B;
4573 win -a %B;
4574 win -a %B;
4575 win -a %B;
4576 win -a %B;
4577 win -a %B;
4578 win -a %B;
4579 win -a %B;
4580 win -a %B;
4581 win -a %B;
4582 win -a %B;
4583 win -a %B;
4584 win -a %B;
4585 win -a %B;
4586 win -a %B;
4587 win -a %B;
4588 win -a %B;
4589 win -a %B;
4590 win -a %B;
4591 win -a %B;
4592 win -a %B;
4593 win -a %B;
4594 win -a %B;
4595 win -a %B;
4596 win -a %B;
4597 win -a %B;
4598 win -a %B;
4599 win -a %B;
4600 win -a %B;
4601 win -a %B;
4602 win -a %B;
4603 win -a %B;
4604 win -a %B;
4605 win -a %B;
4606 win -a %B;
4607 win -a %B;
4608 win -a %B;
4609 win -a %B;
4610 win -a %B;
4611 win -a %B;
4612 win -a %B;
4613 win -a %B;
4614 win -a %B;
4615 win -a %B;
4616 win -a %B;
4617 win -a %B;
4618 win -a %B;
4619 win -a %B;
4620 win -a %B;
4621 win -a %B;
4622 win -a %B;
4623 win -a %B;
4624 win -a %B;
4625 win -a %B;
4626 win -a %B;
4627 win -a %B;
4628 win -a %B;
4629 win -a %B;
4630 win -a %B;
4631 win -a %B;
4632 win -a %B;
4633 win -a %B;
4634 win -a %B;
4635 win -a %B;
4636 win -a %B;
4637 win -a %B;
4638 win -a %B;
4639 win -a %B;
4640 win -a %B;
4641 win -a %B;
4642 win -a %B;
4643 win -a %B;
4644 win -a %B;
4645 win -a %B;
4646 win -a %B;
4647 win -a %B;
4648 win -a %B;
4649 win -a %B;
4650 win -a %B;
4651 win -a %B;
4652 win -a %B;
4653 win -a %B;
4654 win -a %B;
4655 win -a %B;
4656 win -a %B;
4657 win -a %B;
4658 win -a %B;
4659 win -a %B;
4660 win -a %B;
4661 win -a %B;
4662 win -a %B;
4663 win -a %B;
4664 win -a %B;
4665 win -a %B;
4666 win -a %B;
4667 win -a %B;
4668 win -a %B;
4669 win -a %B;
4670 win -a %B;
4671 win -a %B;
4672 win -a %B;
4673 win -a %B;
4674 win -a %B;
4675 win -a %B;
4676 win -a %B;
4677 win -a %B;
4678 win -a %B;
4679 win -a %B;
4680 win -a %B;
4681 win -a %B;
4682 win -a %B;
4683 win -a %B;
4684 win -a %B;
4685 win -a %B;
4686 win -a %B;
4687 win -a %B;
4688 win -a %B;
4689 win -a %B;
4690 win -a %B;
4691 win -a %B;
4692 win -a %B;
4693 win -a %B;
4694 win -a %B;
4695 win -a %B;
4696 win -a %B;
4697 win -a %B;
4698 win -a %B;
4699 win -a %B;
4700 win -a %B;
4701 win -a %B;
4702 win -a %B;
4703 win -a %B;
4704 win -a %B;
4705 win -a %B;
4706 win -a %B;
4707 win -a %B;
4708 win -a %B;
4709 win -a %B;
4710 win -a %B;
4711 win -a %B;
4712 win -a %B;
4713 win -a %B;
4714 win -a %B;
4715 win -a %B;
4716 win -a %B;
4717 win -a %B;
4718 win -a %B;
4719 win -a %B;
4720 win -a %B;
4721 win -a %B;
4722 win -a %B;
4723 win -a %B;
4724 win -a %B;
4725 win -a %B;
4726 win -a %B;
4727 win -a %B;
4728 win -a %B;
4729 win -a %B;
4730 win -a %B;
4731 win -a %B;
4732 win -a %B;
4733 win -a %B;
4734 win -a %B;
4735 win -a %B;
4736 win -a %B;
4737 win -a %B;
4738 win -a %B;
4739 win -a %B;
4740 win -a %B;
4741 win -a %B;
4742 win -a %B;
4743 win -a %B;
4744 win -a %B;
4745 win -a %B;
4746 win -a %B;
4747 win -a %B;
4748 win -a %B;
4749 win -a %B;
4750 win -a %B;
4751 win -a %B;
4752 win -a %B;
4753 win -a %B;
4754 win -a %B;
4755 win -a %B;
4756 win -a %B;
4757 win -a %B;
4758 win -a %B;
4759 win -a %B;
4760 win -a %B;
4761 win -a %B;
4762 win -a %B;
4763 win -a %B;
4764 win -a %B;
4765 win -a %B;
4766 win -a %B;
4767 win -a %B;
4768 win -a %B;
4769 win -a %B;
4770 win -a %B;
4771 win -a %B;
4772 win -a %B;
4773 win -a %B;
4774 win -a %B;
4775 win -a %B;
4776 win -a %B;
4777 win -a %B;
4778 win -a %B;
4779 win -a %B;
4780 win -a %B;
4781 win -a %B;
4782 win -a %B;
4783 win -a %B;
4784 win -a %B;
4785 win -a %B;
4786 win -a %B;
4787 win -a %B;
4788 win -a %B;
4789 win -a %B;
4790 win -a %B;
4791 win -a %B;
4792 win -a %B;
4793 win -a %B;
4794 win -a %B;
4795 win -a %B;
4796 win -a %B;
4797 win -a %B;
4798 win -a %B;
4799 win -a %B;
4800 win -a %B;
4801 win -a %B;
4802 win -a %B;
4803 win -a %B;
4804 win -a %B;
4805 win -a %B;
4806 win -a %B;
4807 win -a %B;
4808 win -a %B;
4809 win -a %B;
4810 win -a %B;
4811 win -a %B;
4812 win -a %B;
4813 win -a %B;
4814 win -a %B;
4815 win -a %B;
4816 win -a %B;
4817 win -a %B;
4818 win -a %B;
4819 win -a %B;
4820 win -a %B;
4821 win -a %B;
4822 win -a %B;
4823 win -a %B;
4824 win -a %B;
4825 win -a %B;
4826 win -a %B;
4827 win -a %B;
4828 win -a %B;
4829 win -a %B;
4830 win -a %B;
4831 win -a %B;
4832 win -a %B;
4833 win -a %B;
4834 win -a %B;
4835 win -a %B;
4836 win -a %B;
4837 win -a %B;
4838 win -a %B;
4839 win -a %B;
4840 win -a %B;
4841 win -a %B;
4842 win -a %B;
4843 win -a %B;
4844 win -a %B;
4845 win -a %B;
4846 win -a %B;
4847 win -a %B;
4848 win -a %B;
4849 win -a %B;
4850 win -a %B;
4851 win -a %B;
4852 win -a %B;
4853 win -a %B;
4854 win -a %B;
4855 win -a %B;
4856 win -a %B;
4857 win -a %B;
4858 win -a %B;
4859 win -a %B;
4860 win -a %B;
4861 win -a %B;
4862 win -a %B;
4863 win -a %B;
4864 win -a %B;
4865 win -a %B;
4866 win -a %B;
4867 win -a %B;
4868 win -a %B;
4869 win -a %B;
4870 win -a %B;
4871 win -a %B;
4872 win -a %B;
4873 win -a %B;
4874 win -a %B;
4875 win -a %B;
4876 win -a %B;
4877 win -a %B;
4878 win -a %B;
4879 win -a %B;
4880 win -a %B;
4881 win -a %B;
4882 win -a %B;
4883 win -a %B;
4884 win -a %B;
4885 win -a %B;
4886 win -a %B;
4887 win -a %B;
4888 win -a %B;
4889 win -a %B;
4890 win -a %B;
4891 win -a %B;
4892 win -a %B;
4893 win -a %B;
4894 win -a %B;
4895 win -a %B;
4896 win -a %B;
4897 win -a %B;
4898 win -a %B;
4899 win -a %B;
4900 win -a %B;
4901 win -a %B;
4902 win -a %B;
4903 win -a %B;
4904 win -a %B;
4905 win -a %B;
4906 win -a %B;
4907 win -a %B;
4908 win -a %B;
4909 win -a %B;
4910 win -a %B;
4911 win -a %B;
4912 win -a %B;
4913 win -a %B;
4914 win -a %B;
4915 win -a %B;
4916 win -a %B;
4917 win -a %B;
4918 win -a %B;
4919 win -a %B;
4920 win -a %B;
4921 win -a %B;
4922 win -a %B;
4923 win -a %B;
4924 win -a %B;
4925 win -a %B;
4926 win -a %B;
4927 win -a %B;
4928 win -a %B;
4929 win -a %B;
4930 win -a %B;
4931 win -a %B;
4932 win -a %B;
4933 win -a %B;
4934 win -a %B;
4935 win -a %B;
4936 win -a %B;
4937 win -a %B;
4938 win -a %B;
4939 win -a %B;
4940 win -a %B;
4941 win -a %B;
4942 win -a %B;
4943 win -a %B;
4944 win -a %B;
4945 win -a %B;
4946 win -a %B;
4947 win -a %B;
4948 win -a %B;
4949 win -a %B;
4950 win -a %B;
4951 win -a %B;
4952 win -a %B;
4953 win -a %B;
4954 win -a %B;
4955 win -a %B;
4956 win -a %B;
4957 win -a %B;
4958 win -a %B;
4959 win -a %B;
4960 win -a %B;
4961 win -a %B;
4962 win -a %B;
4963 win -a %B;
4964 win -a %B;
4965 win -a %B;
4966 win -a %B;
4967 win -a %B;
4968 win -a %B;
4969 win -a %B;
4970 win -a %B;
4971 win -a %B;
4972 win -a %B;
4973 win -a %B;
4974 win -a %B;
4975 win -a %B;
4976 win -a %B;
4977 win -a %B;
4978 win -a %B;
4979 win -a %B;
4980 win -a %B;
4981 win -a %B;
4982 win -a %B;
4983 win -a %B;
4984 win -a %B;
4985 win -a %B;
4986 win -a %B;
4987 win -a %B;
4988 win -a %B;
4989 win -a %B;
4990 win -a %B;
4991 win -a %B;
4992 win -a %B;
4993 win -a %B;
4994 win -a %B;
4995 win -a %B;
4996 win -a %B;
4997 win -a %B;
4998 win -a %B;
4999 win -a %B;
5000 win -a %B;
```


Appendix

LabTalk script (OriginPro 2020b)

```
4489 win -ia;
4490 win -a %V;
4491 win -z;
4492 page.active = 5;
4493 plot -l;
4494 %L=layer.x.from$;
4495 %M=layer.x.to$;
4496 %N=layer.y.from$;
4497 %O=layer.y.to$;
4498 rescale;
4499 layer.x.from=%L;
4500 layer.x.to=%M;
4501 layer.y.from=%N;
4502 layer.y.to=%O;
4503 page.active$=6;
4504 layer.x.posttype=1;
4505 layer.x.position=0;
4506 };
4507 };
4508 }
4509 else type -b "Please start again and choose only one
operation you want to perform!";
4510 };
4511 //Automatic baseline correction of layer 6
4512 if (page.active == 6) {
4513 //Calculation of baseline with peakpicking
4514 pkfind dir:=p method:=max npts:=%(LP$)
hwidth:=%(HW$) filter:=h value:=%(MH$)
oleft_x:=[%B]Sheet1!col(3) oleft_y:=[%B]Sheet1!col(4)
oright_x:=[%B]Sheet1!col(5)
oright_y:=[%B]Sheet1!col(6) ocenter:=<none>;
4515 win -a %B;
4516 copydata irng:=[%B]Sheet1!col(11)
omrg:=[%B]Sheet1!col(1);
4517 copydata irng:=[%B]Sheet1!col(12)
omrg:=[%B]Sheet1!col(2);
4518 copydata irng:=[%B]Sheet1!col(1)[1]
omrg:=[%B]Sheet1!col(7)[1];
4519 copydata irng:=[%B]Sheet1!col(2)[1]
omrg:=[%B]Sheet1!col(8)[1];
4520 int NoP=count(col(3),1);
4521 kk=2;
4522 for(int ii=2; ii<=NoP; ii++){
4523 copydata irng:=[%B]Sheet1!col(3)[ii]
omrg:=[%B]Sheet1!col(7)[kk];
4524 copydata irng:=[%B]Sheet1!col(4)[ii]
omrg:=[%B]Sheet1!col(8)[kk];
4525 copydata irng:=[%B]Sheet1!col(5)[ii]
omrg:=[%B]Sheet1!col(7)[kk+1];
4526 copydata irng:=[%B]Sheet1!col(6)[ii]
omrg:=[%B]Sheet1!col(8)[kk+1];
4527 kk+2;
4528 };
4529 int ik=count(col(1),1);
4530 int ij=count(col(7),1);
4531 copydata irng:=[%B]Sheet1!col(1)[ik]
omrg:=[%B]Sheet1!col(7)[ij+1];
4532 copydata irng:=[%B]Sheet1!col(2)[ik]
omrg:=[%B]Sheet1!col(8)[ij+1];
4533 wks.col7.format = 1;
4534 wks.col8.format = 1;
4535 int aa=count(col(7),1);
4536 %K=col(7)[aa];
4537 MaVY=max(col(8));
4538 //Check baseline
4539 plotxy iy:=(1,2) plot:=200 color:=color(0,0,0);
4540 %U=%H;
4541 win -a %B;
4542 wks.col1.Iname$ = m/z;
4543 wks.col2.Iname$ = Absolute Intensity;
4544 wks.col2.comment$ = "Original Spectrum";
4545 wks.col7.Iname$ = m/z;
4546 wks.col8.Iname$ = Absolute Intensity;
4547 wks.col8.comment$ = "Baseline";
4548 plotxy iy:=(7,8) plot:=200 color:=color(255,0,0)
ogl:=[%U]1!;
4549 win -a %U;
4550 win -z;
4551 layer.x.from=0;
4552 layer.x.to=%K;
4553 layer.y.from=0;
4554 layer.y.to=MaVY;
4555 speedmode index:=project sm:=off;
4556 dlgChkList inames:="Not good / reset and exit|Not
good / change parameters|Good, please go on"
osel:=POP title:="Baseline Calculation" desc:="Was the
baseline calculation successful? What do you want to
do now? Check only ONE box" olist=KB;
4557 if (KB.GetLength() == 1) {
4558 if (POP == 1) {
4559 win -c;
4560 win -a %B;
4561 win -c;
4562 win -a %V;
4563 };
4564 for (int ii=2; ii==POP; ii+0) {
4565 col(3)="" ;
4566 col(4)="" ;
4567 col(5)="" ;
4568 col(6)="" ;
4569 col(7)="" ;
4570 col(8)="" ;
4571 getn
4572 (Peak picking parameters:) :@G
4573 (Local Points:) LP$
4574 (Minimum Height [absolut value]:) MH$
4575 (Maximum Half Width:) HW$
4576 (-) :@G
4577 (Please insert the parameters for the peak picking pro-
cess.);
4578 pkfind dir:=p method:=max npts:=%(LP$)
hwidth:=%(HW$) filter:=h value:=%(MH$)
oleft_x:=[%B]Sheet1!col(3) oleft_y:=[%B]Sheet1!col(4)
oright_x:=[%B]Sheet1!col(5)
oright_y:=[%B]Sheet1!col(6) ocenter:=<none>;
4579 win -a %B;
4580 copydata irng:=[%B]Sheet1!col(1)[1]
omrg:=[%B]Sheet1!col(7)[1];
4581 copydata irng:=[%B]Sheet1!col(2)[1]
omrg:=[%B]Sheet1!col(8)[1];
4582 int NoP=count(col(3),1);
4583 kk=2;
4584 for(int ii=2; ii<=NoP; ii++){
4585 copydata irng:=[%B]Sheet1!col(3)[ii]
omrg:=[%B]Sheet1!col(7)[kk];
4586 copydata irng:=[%B]Sheet1!col(4)[ii]
omrg:=[%B]Sheet1!col(8)[kk];
4587 copydata irng:=[%B]Sheet1!col(5)[ii]
omrg:=[%B]Sheet1!col(7)[kk+1];
4588 copydata irng:=[%B]Sheet1!col(6)[ii]
omrg:=[%B]Sheet1!col(8)[kk+1];
4589 kk+2;
4590 };
4591 int ik=count(col(1),1);
4592 int ij=count(col(7),1);
4593 copydata irng:=[%B]Sheet1!col(1)[ik]
omrg:=[%B]Sheet1!col(7)[ij+1];
4594 copydata irng:=[%B]Sheet1!col(2)[ik]
omrg:=[%B]Sheet1!col(8)[ij+1];
4595 wks.col7.format = 1;
4596 wks.col8.format = 1;
4597 int aa=count(col(7),1);
4598 %K=col(7)[aa];
4599 MVY=max(col(8));
4600 win -a %U;
4601 win -z;
4602 layer.x.from=0;
4603 layer.x.to=%K;
4604 layer.y.from=0;
```

```

4605 layer.y.to=MVY;
4606 speedmode index:=project sm:=off;
4607 getyesno (Was the baseline calculation successful?)
4608 yesno (Baselinecorrection);
4608 if (yesno == 0) {
4609 type -b "Alright, try again.";
4610 };
4611 if(yesno == 1) {
4612 POP=3;
4613 };
4614 };
4615 if (POP == 3) {
4616 win -a %B;
4617 range rx = 7;
4618 range ry = 8;
4619 int aa=count(col(7),1);
4620 pp=aa-1;
4621 //Calculation of the baseline-corrected spectrum
4622 for(int ii=1; ii<=pp; ii++){
4623 int datastart1=list(rx[ii],col(1));
4624 int dataend1=list(rx[ii+1],col(1));
4625 subtract_line iy:=(1,2)[datastart1:dataend1] x1:=rx[ii]
y1:=ry[ii] x2:=rx[ii+1] y2:=ry[ii+1] oy:=<new>;
4626 };
4627 for(int ii=1; ii<=pp; ii++){
4628 int NoP=count(col(9),1);
4629 if (exist(col(K)) == 1) {
4630 copydata irng:=(11,12) orng:=(9,10)[NoP];
4631 delete col(K);
4632 delete col(K);
4633 };
4634 };
4635 //Check baseline and calculated spectrum
4636 plotxy iy:=(9,10) plot:=200 color:=color(0,0,255)
ogl:=[%U]1! legend:=0 rescale:=0;
4637 win -a %B;
4638 wks.col9.Iname$ = m/z;
4639 wks.col10.Iname$ = Absolute Intensity;
4640 wks.col10.comment$ = "Baselinecorrected Spectrum";
4641 MiVY=min(col(10));
4642 win -a %U;
4643 win -z;
4644 layer.y.from=MiVY;
4645 getyesno (Was the baseline subtraction successful?)
4646 yesno (Baselinecorrection);
4646 if (yesno == 0){
4647 type -b "Sorry, that it didn't work. Shit happens. Try
again.";
4648 win -c;
4649 win -a %B;
4650 win -c;
4651 win -a %V;
4652 };
4653 //Replace old spectrum with new one and finish
4654 if (yesno == 1){
4655 win -c;
4656 win -a %B;
4657 copydata irng:=col(9)
orng:=[%(bb.name$)]Sheet1!col(11);
4658 copydata irng:=col(10)
orng:=[%(bb.name$)]Sheet1!col(12);
4659 win -c;
4660 win -a %(bb.name$);
4661 win -ia;
4662 win -a %V;
4663 win -z;
4664 page.active = 6;
4665 plot -l;
4666 %L=layer.x.from$;
4667 %M=layer.x.to$;
4668 %N=layer.y.from$;
4669 %O=layer.y.to$;
4670 rescale;
4671 layer.x.from=%L;
4672 layer.x.to=%M;
4673 layer.y.from=%N;
4674 layer.y.to=%O;
4675 page.active$=6;
4676 layer.x.posttype=1;
4677 layer.x.position=0;
4678 };
4679 };
4680 }
4681 else type -b "Please start again and choose only one
operation you want to perform!";
4682 };
4683 };
4684 }
4685 else type -b "Please start again and choose only one
baseline correction operation!";
4686 };

```

8.1.4. Section "manualpeakpicking" (L. 4688 - 5147)

```

4688 [manualpeakpicking]
4689 type -b "Select a signal in a spectrum and press Enter
to save it. Repeat this for all other signals you want to
pick. When you are done, press ESC and the peak
picking begins.";
4690 //Set variables and start data reader to choose values
4691 @global=1;
4692 dataset dsx,dsy;
4693 pp=0;
4694 dotool 3;
4695 def pointproc {
4696 pp++;
4697 dsx[pp] = x;
4698 dsy[pp] = y;
4699 };
4700 //Define what happens to the chosen values
4701 def quittoolbox {
4702 range -w aa=%C;
4703 range bb = uid2range(aa.GetPage())$;
4704 %W=%H;
4705 //Manual peakpicking for 1-stack plot
4706 if (page.nlayers == 1) {
4707 //Manual peakpicking for layer 1
4708 if (page.active == 1) {
4709 //Extract values and insert them to the right place
4710 win -a %(bb.name$);
4711 page.active$=Sheet1;
4712 int NoP1=count(col(c),1);
4713 int NoP2=count(col(d),1);
4714 NoP1++;
4715 NoP2++;
4716 for(int ii=1; ii<=pp; ii++){
4717 cell(NoP1,3)=dsx[ii];
4718 cell(NoP2,4)=dsy[ii];
4719 NoP1++;
4720 NoP2++;
4721 };
4722 //Sort new values
4723 sort -c 4 4 %(bb.name$)_c;
4724 %(bb.name$)_c=sort(%(bb.name$)_c);
4725 };
4726 win -i;
4727 win -a %W;
4728 win -z;
4729 break 1;
4730 };
4731 //Manual peakpicking for 2-stack plot
4732 if (page.nlayers == 2) {
4733 //Manual peakpicking for layer 1
4734 if (page.active == 1) {

```

Appendix

LabTalk script (OriginPro 2020b)

```
4735 //Extract values and insert them to the right place
4736 win -a %(bb.name$);
4737 page.active$=Sheet1;
4738 int NoP1=count(col(e),1);
4739 int NoP2=count(col(f),1);
4740 NoP1++;
4741 NoP2++;
4742 for(int ii=1; ii<=pp; ii++){
4743 cell(NoP1,5)=dsx[iii];
4744 cell(NoP2,6)=dsy[iii];
4745 NoP1++;
4746 NoP2++;
4747 };
4748 //Sort new values
4749 sort -c 6 6 %(bb.name$)_e;
4750 %(bb.name$)_e=sort(%(bb.name$)_e);
4751 };
4752 //Manual peakpicking for layer 2
4753 if (page.active == 2) {
4754 //Extract values and insert them to the right place
4755 win -a %(bb.name$);
4756 page.active$=Sheet1;
4757 int NoP1=count(col(g),1);
4758 int NoP2=count(col(h),1);
4759 NoP1++;
4760 NoP2++;
4761 for(int ii=1; ii<=pp; ii++){
4762 cell(NoP1,7)=dsx[iii];
4763 cell(NoP2,8)=dsy[iii];
4764 NoP1++;
4765 NoP2++;
4766 };
4767 //Sort new values
4768 sort -c 8 8 %(bb.name$)_g;
4769 %(bb.name$)_g=sort(%(bb.name$)_g);
4770 };
4771 win -i;
4772 win -a %W;
4773 win -z;
4774 break 1;
4775 };
4776 //Manual peakpicking for 3-stack plot
4777 if (page.nlayers == 3) {
4778 //Manual peakpicking for layer 1
4779 if (page.active == 1) {
4780 //Extract values and insert them to the right place
4781 win -a %(bb.name$);
4782 page.active$=Sheet1;
4783 int NoP1=count(col(g),1);
4784 int NoP2=count(col(h),1);
4785 NoP1++;
4786 NoP2++;
4787 for(int ii=1; ii<=pp; ii++){
4788 cell(NoP1,7)=dsx[iii];
4789 cell(NoP2,8)=dsy[iii];
4790 NoP1++;
4791 NoP2++;
4792 };
4793 //Sort new values
4794 sort -c 8 8 %(bb.name$)_g;
4795 %(bb.name$)_g=sort(%(bb.name$)_g);
4796 };
4797 //Manual peakpicking for layer 2
4798 if (page.active == 2) {
4799 //Extract values and insert them to the right place
4800 win -a %(bb.name$);
4801 page.active$=Sheet1;
4802 int NoP1=count(col(i),1);
4803 int NoP2=count(col(j),1);
4804 NoP1++;
4805 NoP2++;
4806 for(int ii=1; ii<=pp; ii++){
4807 cell(NoP1,9)=dsx[iii];
4808 cell(NoP2,10)=dsy[iii];
4809 NoP1++;
4810 NoP2++;
4811 };
4812 //Sort new values
4813 sort -c 10 10 %(bb.name$)_i;
4814 %(bb.name$)_i=sort(%(bb.name$)_i);
4815 };
4816 //Manual peakpicking for layer 3
4817 if (page.active == 3) {
4818 //Extract values and insert them to the right place
4819 win -a %(bb.name$);
4820 page.active$=Sheet1;
4821 int NoP1=count(col(k),1);
4822 int NoP2=count(col(l),1);
4823 NoP1++;
4824 NoP2++;
4825 for(int ii=1; ii<=pp; ii++){
4826 cell(NoP1,11)=dsx[iii];
4827 cell(NoP2,12)=dsy[iii];
4828 NoP1++;
4829 NoP2++;
4830 };
4831 //Sort new values
4832 sort -c 12 12 %(bb.name$)_k;
4833 %(bb.name$)_k=sort(%(bb.name$)_k);
4834 };
4835 win -i;
4836 win -a %W;
4837 win -z;
4838 break 1;
4839 };
4840 //Manual peakpicking for 4-stack plot
4841 if (page.nlayers == 4) {
4842 //Manual peakpicking for layer 1
4843 if (page.active == 1) {
4844 //Extract values and insert them to the right place
4845 win -a %(bb.name$);
4846 page.active$=Sheet1;
4847 int NoP1=count(col(i),1);
4848 int NoP2=count(col(j),1);
4849 NoP1++;
4850 NoP2++;
4851 for(int ii=1; ii<=pp; ii++){
4852 cell(NoP1,9)=dsx[iii];
4853 cell(NoP2,10)=dsy[iii];
4854 NoP1++;
4855 NoP2++;
4856 };
4857 //Sort new values
4858 sort -c 10 10 %(bb.name$)_i;
4859 %(bb.name$)_i=sort(%(bb.name$)_i);
4860 };
4861 //Manual peakpicking for layer 2
4862 if (page.active == 2) {
4863 //Extract values and insert them to the right place
4864 win -a %(bb.name$);
4865 page.active$=Sheet1;
4866 int NoP1=count(col(k),1);
4867 int NoP2=count(col(l),1);
4868 NoP1++;
4869 NoP2++;
4870 for(int ii=1; ii<=pp; ii++){
4871 cell(NoP1,11)=dsx[iii];
4872 cell(NoP2,12)=dsy[iii];
4873 NoP1++;
4874 NoP2++;
4875 };
4876 //Sort new values
4877 sort -c 12 12 %(bb.name$)_k;
4878 %(bb.name$)_k=sort(%(bb.name$)_k);
4879 };
4880 //Manual peakpicking for layer 3
4881 if (page.active == 3) {
4882 //Extract values and insert them to the right place
```

Appendix

LabTalk script (OriginPro 2020b)

```
4883 win -a %(bb.name$);
4884 page.active$=Sheet1;
4885 int NoP1=count(col(m),1);
4886 int NoP2=count(col(n),1);
4887 NoP1++;
4888 NoP2++;
4889 for(int ii=1; ii<=pp; ii++){
4890 cell(NoP1,13)=dsx[ii];
4891 cell(NoP2,14)=dsy[ii];
4892 NoP1++;
4893 NoP2++;
4894 };
4895 //Sort new values
4896 sort -c 14 14 %(bb.name$)_m;
4897 %(bb.name$)_m=sort(%(bb.name$)_m);
4898 };
4899 //Manual peakpicking for layer 4
4900 if (page.active == 4) {
4901 //Extract values and insert them to the right place
4902 win -a %(bb.name$);
4903 page.active$=Sheet1;
4904 int NoP1=count(col(o),1);
4905 int NoP2=count(col(p),1);
4906 NoP1++;
4907 NoP2++;
4908 for(int ii=1; ii<=pp; ii++){
4909 cell(NoP1,15)=dsx[ii];
4910 cell(NoP2,16)=dsy[ii];
4911 NoP1++;
4912 NoP2++;
4913 };
4914 //Sort new values
4915 sort -c 16 16 %(bb.name$)_o;
4916 %(bb.name$)_o=sort(%(bb.name$)_o);
4917 };
4918 win -i;
4919 win -a %W;
4920 win -z;
4921 break 1;
4922 };
4923 //Manual peakpicking for 5-stack plot
4924 if (page.nlayers == 5) {
4925 //Manual peakpicking for layer 1
4926 if (page.active == 1) {
4927 //Extract values and insert them to the right place
4928 win -a %(bb.name$);
4929 page.active$=Sheet1;
4930 int NoP1=count(col(k),1);
4931 int NoP2=count(col(l),1);
4932 NoP1++;
4933 NoP2++;
4934 for(int ii=1; ii<=pp; ii++){
4935 cell(NoP1,11)=dsx[ii];
4936 cell(NoP2,12)=dsy[ii];
4937 NoP1++;
4938 NoP2++;
4939 };
4940 //Sort new values
4941 sort -c 12 12 %(bb.name$)_k;
4942 %(bb.name$)_k=sort(%(bb.name$)_k);
4943 };
4944 //Manual peakpicking for layer 2
4945 if (page.active == 2) {
4946 //Extract values and insert them to the right place
4947 win -a %(bb.name$);
4948 page.active$=Sheet1;
4949 int NoP1=count(col(m),1);
4950 int NoP2=count(col(n),1);
4951 NoP1++;
4952 NoP2++;
4953 for(int ii=1; ii<=pp; ii++){
4954 cell(NoP1,13)=dsx[ii];
4955 cell(NoP2,14)=dsy[ii];
4956 NoP1++;
4957 NoP2++;
4958 };
4959 //Sort new values
4960 sort -c 14 14 %(bb.name$)_m;
4961 %(bb.name$)_m=sort(%(bb.name$)_m);
4962 };
4963 //Manual peakpicking for layer 3
4964 if (page.active == 3) {
4965 //Extract values and insert them to the right place
4966 win -a %(bb.name$);
4967 page.active$=Sheet1;
4968 int NoP1=count(col(o),1);
4969 int NoP2=count(col(p),1);
4970 NoP1++;
4971 NoP2++;
4972 for(int ii=1; ii<=pp; ii++){
4973 cell(NoP1,15)=dsx[ii];
4974 cell(NoP2,16)=dsy[ii];
4975 NoP1++;
4976 NoP2++;
4977 };
4978 //Sort new values
4979 sort -c 16 16 %(bb.name$)_o;
4980 %(bb.name$)_o=sort(%(bb.name$)_o);
4981 };
4982 //Manual peakpicking for layer 4
4983 if (page.active == 4) {
4984 //Extract values and insert them to the right place
4985 win -a %(bb.name$);
4986 page.active$=Sheet1;
4987 int NoP1=count(col(q),1);
4988 int NoP2=count(col(r),1);
4989 NoP1++;
4990 NoP2++;
4991 for(int ii=1; ii<=pp; ii++){
4992 cell(NoP1,17)=dsx[ii];
4993 cell(NoP2,18)=dsy[ii];
4994 NoP1++;
4995 NoP2++;
4996 };
4997 //Sort new values
4998 sort -c 18 18 %(bb.name$)_q;
4999 %(bb.name$)_q=sort(%(bb.name$)_q);
5000 };
5001 //Manual peakpicking for layer 5
5002 if (page.active == 5) {
5003 //Extract values and insert them to the right place
5004 win -a %(bb.name$);
5005 page.active$=Sheet1;
5006 int NoP1=count(col(s),1);
5007 int NoP2=count(col(t),1);
5008 NoP1++;
5009 NoP2++;
5010 for(int ii=1; ii<=pp; ii++){
5011 cell(NoP1,19)=dsx[ii];
5012 cell(NoP2,20)=dsy[ii];
5013 NoP1++;
5014 NoP2++;
5015 };
5016 //Sort new values
5017 sort -c 20 20 %(bb.name$)_s;
5018 %(bb.name$)_s=sort(%(bb.name$)_s);
5019 };
5020 win -i;
5021 win -a %W;
5022 win -z;
5023 break 1;
5024 };
5025 //Manual peakpicking for 6-stack plot
5026 if (page.nlayers == 6) {
5027 //Manual peakpicking for layer 1
5028 if (page.active == 1) {
5029 //Extract values and insert them to the right place
5030 win -a %(bb.name$);
```



```
5031 page.active$=Sheet1;
5032 int NoP1=count(col(m),1);
5033 int NoP2=count(col(n),1);
5034 NoP1++;
5035 NoP2++;
5036 for(int ii=1; ii<=pp; ii++){
5037 cell(NoP1,13)=dsx[ii];
5038 cell(NoP2,14)=dsy[ii];
5039 NoP1++;
5040 NoP2++;
5041 };
5042 //Sort new values
5043 sort -c 14 14 %(bb.name$)_m;
5044 %(bb.name$)_m=sort(%(bb.name$)_m);
5045 };
5046 //Manual peakpicking for layer 2
5047 if (page.active == 2) {
5048 //Extract values and insert them to the right place
5049 win -a %(bb.name$);
5050 page.active$=Sheet1;
5051 int NoP1=count(col(o),1);
5052 int NoP2=count(col(p),1);
5053 NoP1++;
5054 NoP2++;
5055 for(int ii=1; ii<=pp; ii++){
5056 cell(NoP1,15)=dsx[ii];
5057 cell(NoP2,16)=dsy[ii];
5058 NoP1++;
5059 NoP2++;
5060 };
5061 //Sort new values
5062 sort -c 16 16 %(bb.name$)_o;
5063 %(bb.name$)_o=sort(%(bb.name$)_o);
5064 };
5065 //Manual peakpicking for layer 3
5066 if (page.active == 3) {
5067 //Extract values and insert them to the right place
5068 win -a %(bb.name$);
5069 page.active$=Sheet1;
5070 int NoP1=count(col(q),1);
5071 int NoP2=count(col(r),1);
5072 NoP1++;
5073 NoP2++;
5074 for(int ii=1; ii<=pp; ii++){
5075 cell(NoP1,17)=dsx[ii];
5076 cell(NoP2,18)=dsy[ii];
5077 NoP1++;
5078 NoP2++;
5079 };
5080 //Sort new values
5081 sort -c 18 18 %(bb.name$)_q;
5082 %(bb.name$)_q=sort(%(bb.name$)_q);
5083 };
5084 //Manual peakpicking for layer 4
5085 if (page.active == 4) {
5086 //Extract values and insert them to the right place
5087 win -a %(bb.name$);
5088 page.active$=Sheet1;
5089 int NoP1=count(col(s),1);
5090 int NoP2=count(col(t),1);
5091 NoP1++;
5092 NoP2++;
5093 for(int ii=1; ii<=pp; ii++){
5094 cell(NoP1,19)=dsx[ii];
5095 cell(NoP2,20)=dsy[ii];
5096 NoP1++;
5097 NoP2++;
5098 };
5099 //Sort new values
5100 sort -c 20 20 %(bb.name$)_s;
5101 %(bb.name$)_s=sort(%(bb.name$)_s);
5102 };
5103 //Manual peakpicking for layer 5
5104 if (page.active == 5) {
5105 //Extract values and insert them to the right place
5106 win -a %(bb.name$);
5107 page.active$=Sheet1;
5108 int NoP1=count(col(u),1);
5109 int NoP2=count(col(v),1);
5110 NoP1++;
5111 NoP2++;
5112 for(int ii=1; ii<=pp; ii++){
5113 cell(NoP1,21)=dsx[ii];
5114 cell(NoP2,22)=dsy[ii];
5115 NoP1++;
5116 NoP2++;
5117 };
5118 //Sort new values
5119 sort -c 22 22 %(bb.name$)_u;
5120 %(bb.name$)_u=sort(%(bb.name$)_u);
5121 };
5122 //Manual peakpicking for layer 6
5123 if (page.active == 6) {
5124 //Extract values and insert them to the right place
5125 win -a %(bb.name$);
5126 page.active$=Sheet1;
5127 int NoP1=count(col(w),1);
5128 int NoP2=count(col(x),1);
5129 NoP1++;
5130 NoP2++;
5131 for(int ii=1; ii<=pp; ii++){
5132 cell(NoP1,23)=dsx[ii];
5133 cell(NoP2,24)=dsy[ii];
5134 NoP1++;
5135 NoP2++;
5136 };
5137 //Sort new values
5138 sort -c 24 24 %(bb.name$)_w;
5139 %(bb.name$)_w=sort(%(bb.name$)_w);
5140 };
5141 win -i;
5142 win -a %W;
5143 win -z;
5144 break 1;
5145 };
5146 };
5147 };
```

8.1.5. Section “autopeakpicking” (L. 5149 - 5673)

```
5149 [autopeakpicking]
5150 %W=H;%
5151 range -w aa=C%;
5152 range bb =uid2range(aa.GetPage());$;
5153
5154 dlgChkList inames="Parameters for all spectra|Param-
eters for marked spectrum|Different parameters for dif-
ferent spectra" osel:=PPP title:"Auto peak picking"
desc:="For which spectra do you want to perform auto
peak picking? Check only ONE box" olist:=CB;
5155
5156 if(CB.GetLength()==1){
5157 //Autopeakpicking for all spectra
5158 if(PPP ==1){
5159 HW$=0.5;
5160 getn
5161 (Peak picking parameters:) :@G
5162 (Local Points:) LP$
5163 (Minimum Height [absolut value]:) MH$
5164 (Maximum Have Width:) HW$
5165 (-) :@G
5166 (Please insert the parameters for the peak picking pro-
cess.);}
```


Appendix

LabTalk script (OriginPro 2020b)

```
5167 //Peakpicking for layer 1
5168 if(page.nlayers == 1){
5169 %W!page.active = 1;
5170 pkfind dir:=p method:=maxnpts:=%(LP$)filter:=h
value:=%(MH$)ocen-
ter_x:=[%(bb.name$)]Sheet1!col(3)ocen-
ter_y:=[%(bb.name$)]Sheet1!col(4) ocenter:=<none>;
5171 };
5172 //Peakpicking for layer 2
5173 if(page.nlayers == 2){
5174 %W!page.active = 1;
5175 pkfind dir:=p method:=maxnpts:=%(LP$)filter:=h
value:=%(MH$)ocen-
ter_x:=[%(bb.name$)]Sheet1!col(5)ocen-
ter_y:=[%(bb.name$)]Sheet1!col(6) ocenter:=<none>;
5176 %W!page.active = 2;
5177 pkfind dir:=p method:=maxnpts:=%(LP$)filter:=h
value:=%(MH$)ocen-
ter_x:=[%(bb.name$)]Sheet1!col(7)ocen-
ter_y:=[%(bb.name$)]Sheet1!col(8) ocenter:=<none>;
5178 };
5179 //Peakpicking for layer 3
5180 if(page.nlayers == 3){
5181 %W!page.active = 1;
5182 pkfind dir:=p method:=maxnpts:=%(LP$)filter:=h
value:=%(MH$)ocen-
ter_x:=[%(bb.name$)]Sheet1!col(7)ocen-
ter_y:=[%(bb.name$)]Sheet1!col(8) ocenter:=<none>;
5183 %W!page.active = 2;
5184 pkfind dir:=p method:=maxnpts:=%(LP$)filter:=h
value:=%(MH$)ocen-
ter_x:=[%(bb.name$)]Sheet1!col(9)ocen-
ter_y:=[%(bb.name$)]Sheet1!col(10) ocenter:=<none>;
5185 %W!page.active = 3;
5186 pkfind dir:=p method:=maxnpts:=%(LP$)filter:=h
value:=%(MH$)ocen-
ter_x:=[%(bb.name$)]Sheet1!col(11)ocen-
ter_y:=[%(bb.name$)]Sheet1!col(12) ocenter:=<none>;
5187 };
5188 //Peakpicking for layer 4
5189 if(page.nlayers == 4){
5190 %W!page.active = 1;
5191 pkfind dir:=p method:=maxnpts:=%(LP$)filter:=h
value:=%(MH$)ocen-
ter_x:=[%(bb.name$)]Sheet1!col(9)ocen-
ter_y:=[%(bb.name$)]Sheet1!col(10) ocenter:=<none>;
5192 %W!page.active = 2;
5193 pkfind dir:=p method:=maxnpts:=%(LP$)filter:=h
value:=%(MH$)ocen-
ter_x:=[%(bb.name$)]Sheet1!col(11)ocen-
ter_y:=[%(bb.name$)]Sheet1!col(12) ocenter:=<none>;
5194 %W!page.active = 3;
5195 pkfind dir:=p method:=maxnpts:=%(LP$)filter:=h
value:=%(MH$)ocen-
ter_x:=[%(bb.name$)]Sheet1!col(13)ocen-
ter_y:=[%(bb.name$)]Sheet1!col(14) ocenter:=<none>;
5196 %W!page.active = 4;
5197 pkfind dir:=p method:=maxnpts:=%(LP$)filter:=h
value:=%(MH$)ocen-
ter_x:=[%(bb.name$)]Sheet1!col(15)ocen-
ter_y:=[%(bb.name$)]Sheet1!col(16) ocenter:=<none>;
5198 };
5199 //Peakpicking for layer 5
5200 if(page.nlayers == 5){
5201 %W!page.active = 1;
5202 pkfind dir:=p method:=maxnpts:=%(LP$)filter:=h
value:=%(MH$)ocen-
ter_x:=[%(bb.name$)]Sheet1!col(11)ocen-
ter_y:=[%(bb.name$)]Sheet1!col(12) ocenter:=<none>;
5203 %W!page.active = 2;
5204 pkfind dir:=p method:=maxnpts:=%(LP$)filter:=h
value:=%(MH$)ocen-
ter_x:=[%(bb.name$)]Sheet1!col(13)ocen-
ter_y:=[%(bb.name$)]Sheet1!col(14) ocen-
ter_y:=[%(bb.name$)]Sheet1!col(14) ocenter:=<none>;
5205 %W!page.active = 3;
5206 pkfind dir:=p method:=max npts:=%(LP$) filter:=h
value:=%(MH$) ocen-
ter_x:=[%(bb.name$)]Sheet1!col(15) ocen-
ter_y:=[%(bb.name$)]Sheet1!col(16) ocenter:=<none>;
5207 %W!page.active = 4;
5208 pkfind dir:=p method:=max npts:=%(LP$) filter:=h
value:=%(MH$) ocen-
ter_x:=[%(bb.name$)]Sheet1!col(17) ocen-
ter_y:=[%(bb.name$)]Sheet1!col(18) ocenter:=<none>;
5209 %W!page.active = 5;
5210 pkfind dir:=p method:=max npts:=%(LP$) filter:=h
value:=%(MH$) ocen-
ter_x:=[%(bb.name$)]Sheet1!col(19) ocen-
ter_y:=[%(bb.name$)]Sheet1!col(20) ocenter:=<none>;
5211 };
5212 //Peakpicking for layer 6
5213 if (page.nlayers == 6) {
5214 %W!page.active = 1;
5215 pkfind dir:=p method:=max npts:=%(LP$)
hwidth:=%(HW$) filter:=h value:=%(MH$) ocen-
ter_x:=[%(bb.name$)]Sheet1!col(13) ocen-
ter_y:=[%(bb.name$)]Sheet1!col(14)
oleft_x:=[%(bb.name$)]Sheet1!col(25)
oleft_y:=[%(bb.name$)]Sheet1!col(26)
oright_x:=[%(bb.name$)]Sheet1!col(27)
oright_y:=[%(bb.name$)]Sheet1!col(28) ocen-
ter:=<none>;
5216 %W!page.active = 2;
5217 pkfind dir:=p method:=max npts:=%(LP$)
hwidth:=%(HW$) filter:=h value:=%(MH$) ocen-
ter_x:=[%(bb.name$)]Sheet1!col(15) ocen-
ter_y:=[%(bb.name$)]Sheet1!col(16)
oleft_x:=[%(bb.name$)]Sheet1!col(29)
oleft_y:=[%(bb.name$)]Sheet1!col(30)
oright_x:=[%(bb.name$)]Sheet1!col(31)
oright_y:=[%(bb.name$)]Sheet1!col(32) ocen-
ter:=<none>;
5218 %W!page.active = 3;
5219 pkfind dir:=p method:=max npts:=%(LP$)
hwidth:=%(HW$) filter:=h value:=%(MH$) ocen-
ter_x:=[%(bb.name$)]Sheet1!col(17) ocen-
ter_y:=[%(bb.name$)]Sheet1!col(18)
oleft_x:=[%(bb.name$)]Sheet1!col(33)
oleft_y:=[%(bb.name$)]Sheet1!col(34)
oright_x:=[%(bb.name$)]Sheet1!col(35)
oright_y:=[%(bb.name$)]Sheet1!col(36) ocen-
ter:=<none>;
5220 %W!page.active = 4;
5221 pkfind dir:=p method:=max npts:=%(LP$)
hwidth:=%(HW$) filter:=h value:=%(MH$) ocen-
ter_x:=[%(bb.name$)]Sheet1!col(19) ocen-
ter_y:=[%(bb.name$)]Sheet1!col(20)
oleft_x:=[%(bb.name$)]Sheet1!col(37)
oleft_y:=[%(bb.name$)]Sheet1!col(38)
oright_x:=[%(bb.name$)]Sheet1!col(39)
oright_y:=[%(bb.name$)]Sheet1!col(40) ocen-
ter:=<none>;
5222 %W!page.active = 5;
5223 pkfind dir:=p method:=max npts:=%(LP$)
hwidth:=%(HW$) filter:=h value:=%(MH$) ocen-
ter_x:=[%(bb.name$)]Sheet1!col(21) ocen-
ter_y:=[%(bb.name$)]Sheet1!col(22)
oleft_x:=[%(bb.name$)]Sheet1!col(41)
oleft_y:=[%(bb.name$)]Sheet1!col(42)
oright_x:=[%(bb.name$)]Sheet1!col(43)
oright_y:=[%(bb.name$)]Sheet1!col(44) ocen-
ter:=<none>;
5224 %W!page.active = 6;
5225 pkfind dir:=p method:=max npts:=%(LP$)
hwidth:=%(HW$) filter:=h value:=%(MH$) ocen-
ter_x:=[%(bb.name$)]Sheet1!col(23) ocen-
ter_y:=[%(bb.name$)]Sheet1!col(24)
oleft_x:=[%(bb.name$)]Sheet1!col(45)
oleft_y:=[%(bb.name$)]Sheet1!col(46)
oright_x:=[%(bb.name$)]Sheet1!col(47)
```

Appendix

LabTalk script (OriginPro 2020b)

```
orignt_y:=[%(bb.name$)]Sheet1!col(48) ocenter:=<none>;
5226 };
5227 };
5228 //Autopeakpicking only for marked spektrum
5229 if (PPP == 2){
5230 //Autopeakpicking of marked spektrum in 1-stack plot
5231 if (page.nlayers == 1) {
5232 //1. Spektrum marked
5233 if (page.active == 1) {
5234 getn
5235 (Peak picking parameters:) :@G
5236 (Local Points:) LP11$
5237 (Minimum Height [absolut value]:) MH11$
5238 (-) :@G
5239 (Please insert the parameters for the peak picking process.);
5240 pkfind dir:=p method:=max npts:=%(LP11$) filter:=h
value:=%(MH11$) ocenter_x:=[%(bb.name$)]Sheet1!col(3) ocenter_y:=[%(bb.name$)]Sheet1!col(4) ocenter:=<none>;
5241 };
5242 };
5243 //Autopeakpicking of marked spektrum in 2-stack plot
5244 if (page.nlayers == 2) {
5245 //1. Spektrum marked
5246 if (page.active == 1) {
5247 getn
5248 (Peak picking parameters:) :@G
5249 (Local Points:) LP21$
5250 (Minimum Height [absolut value]:) MH21$
5251 (-) :@G
5252 (Please insert the parameters for the peak picking process.);
5253 pkfind dir:=p method:=max npts:=%(LP21$) filter:=h
value:=%(MH21$) ocenter_x:=[%(bb.name$)]Sheet1!col(5) ocenter_y:=[%(bb.name$)]Sheet1!col(6) ocenter:=<none>;
5254 };
5255 //2. Spektrum marked
5256 if (page.active == 2) {
5257 getn
5258 (Peak picking parameters:) :@G
5259 (Local Points:) LP22$
5260 (Minimum Height [absolut value]:) MH22$
5261 (-) :@G
5262 (Please insert the parameters for the peak picking process.);
5263 pkfind dir:=p method:=max npts:=%(LP22$) filter:=h
value:=%(MH22$) ocenter_x:=[%(bb.name$)]Sheet1!col(7) ocenter_y:=[%(bb.name$)]Sheet1!col(8) ocenter:=<none>;
5264 };
5265 };
5266 //Autopeakpicking of marked spektrum in 3-stack plot
5267 if (page.nlayers == 3) {
5268 //1. Spektrum marked
5269 if (page.active == 1) {
5270 getn
5271 (Peak picking parameters:) :@G
5272 (Local Points:) LP31$
5273 (Minimum Height [absolut value]:) MH31$
5274 (-) :@G
5275 (Please insert the parameters for the peak picking process.);
5276 pkfind dir:=p method:=max npts:=%(LP31$) filter:=h
value:=%(MH31$) ocenter_x:=[%(bb.name$)]Sheet1!col(7) ocenter_y:=[%(bb.name$)]Sheet1!col(8) ocenter:=<none>;
5277 };
5278 //2. Spektrum marked
5279 if (page.active == 2) {
5280 getn
5281 (Peak picking parameters:) :@G
5282 (Local Points:) LP32$
5283 (Minimum Height [absolut value]:) MH32$
5284 (-) :@G
5285 (Please insert the parameters for the peak picking process.);
5286 pkfind dir:=p method:=max npts:=%(LP32$) filter:=h
value:=%(MH32$) ocenter_x:=[%(bb.name$)]Sheet1!col(9) ocenter_y:=[%(bb.name$)]Sheet1!col(10) ocenter:=<none>;
5287 };
5288 //3. Spektrum marked
5289 if (page.active == 3) {
5290 getn
5291 (Peak picking parameters:) :@G
5292 (Local Points:) LP33$
5293 (Minimum Height [absolut value]:) MH33$
5294 (-) :@G
5295 (Please insert the parameters for the peak picking process.);
5296 pkfind dir:=p method:=max npts:=%(LP33$) filter:=h
value:=%(MH33$) ocenter_x:=[%(bb.name$)]Sheet1!col(11) ocenter_y:=[%(bb.name$)]Sheet1!col(12) ocenter:=<none>;
5297 };
5298 };
5299 //Autopeakpicking of marked spektrum in 4-stack plot
5300 if (page.nlayers == 4) {
5301 //1. Spektrum marked
5302 if (page.active == 1) {
5303 getn
5304 (Peak picking parameters:) :@G
5305 (Local Points:) LP41$
5306 (Minimum Height [absolut value]:) MH41$
5307 (-) :@G
5308 (Please insert the parameters for the peak picking process.);
5309 pkfind dir:=p method:=max npts:=%(LP41$) filter:=h
value:=%(MH41$) ocenter_x:=[%(bb.name$)]Sheet1!col(9) ocenter_y:=[%(bb.name$)]Sheet1!col(10) ocenter:=<none>;
5310 };
5311 //2. Spektrum marked
5312 if (page.active == 2) {
5313 getn
5314 (Peak picking parameters:) :@G
5315 (Local Points:) LP42$
5316 (Minimum Height [absolut value]:) MH42$
5317 (-) :@G
5318 (Please insert the parameters for the peak picking process.);
5319 pkfind dir:=p method:=max npts:=%(LP42$) filter:=h
value:=%(MH42$) ocenter_x:=[%(bb.name$)]Sheet1!col(11) ocenter_y:=[%(bb.name$)]Sheet1!col(12) ocenter:=<none>;
5320 };
5321 //3. Spektrum marked
5322 if (page.active == 3) {
5323 getn
5324 (Peak picking parameters:) :@G
5325 (Local Points:) LP43$
5326 (Minimum Height [absolut value]:) MH43$
5327 (-) :@G
5328 (Please insert the parameters for the peak picking process.);
5329 pkfind dir:=p method:=max npts:=%(LP43$) filter:=h
value:=%(MH43$) ocenter_x:=[%(bb.name$)]Sheet1!col(13) ocenter_y:=[%(bb.name$)]Sheet1!col(14) ocenter:=<none>;
5330 };
5331 //4. Spektrum marked
5332 if (page.active == 4) {
5333 getn
5334 (Peak picking parameters:) :@G
5335 (Local Points:) LP44$
```

Appendix

LabTalk script (OriginPro 2020b)

```
5336 (Minimum Height [absolut value]:) MH44$
5337 (-) :@G
5338 (Please insert the parameters for the peak picking process.);
5339 pkfind dir:=p method:=max npts:=%(LP44$) filter:=h
value:=%(MH44$) ocen-
ter_x:=[%(bb.name$)]Sheet1!col(15) ocen-
ter_y:=[%(bb.name$)]Sheet1!col(16) ocenter:=<none>;
5340 };
5341 };
5342 //Autopeakpicking of marked spektrum in 5-stack plot
5343 if (page.nlayers == 5) {
5344 //1. Spektrum marked
5345 if (page.active == 1) {
5346 getn
5347 (Peak picking parameters:) :@G
5348 (Local Points:) LP51$
5349 (Minimum Height [absolut value]:) MH51$
5350 (-) :@G
5351 (Please insert the parameters for the peak picking process.);
5352 pkfind dir:=p method:=max npts:=%(LP51$) filter:=h
value:=%(MH51$) ocen-
ter_x:=[%(bb.name$)]Sheet1!col(11) ocen-
ter_y:=[%(bb.name$)]Sheet1!col(12) ocenter:=<none>;
5353 };
5354 //2. Spektrum marked
5355 if (page.active == 2) {
5356 getn
5357 (Peak picking parameters:) :@G
5358 (Local Points:) LP52$
5359 (Minimum Height [absolut value]:) MH52$
5360 (-) :@G
5361 (Please insert the parameters for the peak picking process.);
5362 pkfind dir:=p method:=max npts:=%(LP52$) filter:=h
value:=%(MH52$) ocen-
ter_x:=[%(bb.name$)]Sheet1!col(13) ocen-
ter_y:=[%(bb.name$)]Sheet1!col(14) ocenter:=<none>;
5363 };
5364 //3. Spektrum marked
5365 if (page.active == 3) {
5366 getn
5367 (Peak picking parameters:) :@G
5368 (Local Points:) LP53$
5369 (Minimum Height [absolut value]:) MH53$
5370 (-) :@G
5371 (Please insert the parameters for the peak picking process.);
5372 pkfind dir:=p method:=max npts:=%(LP53$) filter:=h
value:=%(MH53$) ocen-
ter_x:=[%(bb.name$)]Sheet1!col(15) ocen-
ter_y:=[%(bb.name$)]Sheet1!col(16) ocenter:=<none>;
5373 };
5374 //4. Spektrum marked
5375 if (page.active == 4) {
5376 getn
5377 (Peak picking parameters:) :@G
5378 (Local Points:) LP54$
5379 (Minimum Height [absolut value]:) MH54$
5380 (-) :@G
5381 (Please insert the parameters for the peak picking process.);
5382 pkfind dir:=p method:=max npts:=%(LP54$) filter:=h
value:=%(MH54$) ocen-
ter_x:=[%(bb.name$)]Sheet1!col(17) ocen-
ter_y:=[%(bb.name$)]Sheet1!col(18) ocenter:=<none>;
5383 };
5384 //5. Spektrum marked
5385 if (page.active == 5) {
5386 getn
5387 (Peak picking parameters:) :@G
5388 (Local Points:) LP55$
5389 (Minimum Height [absolut value]:) MH55$
5390 (-) :@G
5391 (Please insert the parameters for the peak picking process.);
5392 pkfind dir:=p method:=max npts:=%(LP55$) filter:=h
value:=%(MH55$) ocen-
ter_x:=[%(bb.name$)]Sheet1!col(19) ocen-
ter_y:=[%(bb.name$)]Sheet1!col(20) ocenter:=<none>;
5393 };
5394 };
5395 //Autopeakpicking of marked spektrum in 6-stack plot
5396 if (page.nlayers == 6) {
5397 //1. Spektrum marked
5398 if (page.active == 1) {
5399 HW61$=0.5;
5400 getn
5401 (Peak picking parameters:) :@G
5402 (Local Points:) LP61$
5403 (Minimum Height [absolut value]:) MH61$
5404 (Maximum Half Width:) HW61$
5405 (-) :@G
5406 (Please insert the parameters for the peak picking process.);
5407 pkfind dir:=p method:=max npts:=%(LP61$)
hwidth:=%(HW61$) filter:=h value:=%(MH61$) ocen-
ter_x:=[%(bb.name$)]Sheet1!col(13) ocen-
ter_y:=[%(bb.name$)]Sheet1!col(14)
oleft_x:=[%(bb.name$)]Sheet1!col(25)
oleft_y:=[%(bb.name$)]Sheet1!col(26)
oright_x:=[%(bb.name$)]Sheet1!col(27)
oright_y:=[%(bb.name$)]Sheet1!col(28) ocen-
ter:=<none>;
5408 };
5409 //2. Spektrum marked
5410 if (page.active == 2) {
5411 HW62$=0.5;
5412 getn
5413 (Peak picking parameters:) :@G
5414 (Local Points:) LP62$
5415 (Minimum Height [absolut value]:) MH62$
5416 (Maximum Half Width:) HW62$
5417 (-) :@G
5418 (Please insert the parameters for the peak picking process.);
5419 pkfind dir:=p method:=max npts:=%(LP62$)
hwidth:=%(HW62$) filter:=h value:=%(MH62$) ocen-
ter_x:=[%(bb.name$)]Sheet1!col(15) ocen-
ter_y:=[%(bb.name$)]Sheet1!col(16)
oleft_x:=[%(bb.name$)]Sheet1!col(29)
oleft_y:=[%(bb.name$)]Sheet1!col(30)
oright_x:=[%(bb.name$)]Sheet1!col(31)
oright_y:=[%(bb.name$)]Sheet1!col(32) ocen-
ter:=<none>;
5420 };
5421 //3. Spektrum marked
5422 if (page.active == 3) {
5423 HW63$=0.5;
5424 getn
5425 (Peak picking parameters:) :@G
5426 (Local Points:) LP63$
5427 (Minimum Height [absolut value]:) MH63$
5428 (Maximum Half Width:) HW63$
5429 (-) :@G
5430 (Please insert the parameters for the peak picking process.);
5431 pkfind dir:=p method:=max npts:=%(LP63$)
hwidth:=%(HW63$) filter:=h value:=%(MH63$) ocen-
ter_x:=[%(bb.name$)]Sheet1!col(17) ocen-
ter_y:=[%(bb.name$)]Sheet1!col(18)
oleft_x:=[%(bb.name$)]Sheet1!col(33)
oleft_y:=[%(bb.name$)]Sheet1!col(34)
oright_x:=[%(bb.name$)]Sheet1!col(35)
oright_y:=[%(bb.name$)]Sheet1!col(36) ocen-
ter:=<none>;
5432 };
5433 //4. Spektrum marked
```


Appendix

LabTalk script (OriginPro 2020b)

```
5434 if (page.active == 4) {
5435   HW64$=0.5;
5436   getn
5437   (Peak picking parameters:) :@G
5438   (Local Points:) LP64$
5439   (Minimum Height [absolut value]:) MH64$
5440   (Maximum Half Width:) HW64$
5441   (-) :@G
5442   (Please insert the parameters for the peak picking process.);
5443   pkfind dir:=p method:=max npts:=%(LP64$)
   hwidth:=%(HW64$) filter:=h value:=%(MH64$) ocenter_x:=[%(bb.name$)]Sheet1!col(19) ocenter_y:=[%(bb.name$)]Sheet1!col(20)
   oleft_x:=[%(bb.name$)]Sheet1!col(37)
   oleft_y:=[%(bb.name$)]Sheet1!col(38)
   oright_x:=[%(bb.name$)]Sheet1!col(39)
   oright_y:=[%(bb.name$)]Sheet1!col(40) ocenter:=<none>;
5444 };
5445 //5. Spektrum marked
5446 if (page.active == 5) {
5447   HW65$=0.5;
5448   getn
5449   (Peak picking parameters:) :@G
5450   (Local Points:) LP65$
5451   (Minimum Height [absolut value]:) MH65$
5452   (Maximum Half Width:) HW65$
5453   (-) :@G
5454   (Please insert the parameters for the peak picking process.);
5455   pkfind dir:=p method:=max npts:=%(LP65$)
   hwidth:=%(HW65$) filter:=h value:=%(MH65$) ocenter_x:=[%(bb.name$)]Sheet1!col(21) ocenter_y:=[%(bb.name$)]Sheet1!col(22)
   oleft_x:=[%(bb.name$)]Sheet1!col(41)
   oleft_y:=[%(bb.name$)]Sheet1!col(42)
   oright_x:=[%(bb.name$)]Sheet1!col(43)
   oright_y:=[%(bb.name$)]Sheet1!col(44) ocenter:=<none>;
5456 };
5457 //6. Spektrum marked
5458 if (page.active == 6) {
5459   HW66$=0.5;
5460   getn
5461   (Peak picking parameters:) :@G
5462   (Local Points:) LP66$
5463   (Minimum Height [absolut value]:) MH66$
5464   (Maximum Half Width:) HW66$
5465   (-) :@G
5466   (Please insert the parameters for the peak picking process.);
5467   pkfind dir:=p method:=max npts:=%(LP66$)
   hwidth:=%(HW66$) filter:=h value:=%(MH66$) ocenter_x:=[%(bb.name$)]Sheet1!col(23) ocenter_y:=[%(bb.name$)]Sheet1!col(24)
   oleft_x:=[%(bb.name$)]Sheet1!col(45)
   oleft_y:=[%(bb.name$)]Sheet1!col(46)
   oright_x:=[%(bb.name$)]Sheet1!col(47)
   oright_y:=[%(bb.name$)]Sheet1!col(48) ocenter:=<none>;
5468 };
5469 };
5470 };
5471 //Autopeakpicking with different parameters for each spektrum
5472 if (PPP == 3) {
5473   //Autopeakpicking for 2-stack plot with individual parameters
5474   if (page.nlayers == 2) {
5475     getn
5476     (Spectrum 1:
5477     Peak picking parameters:) :@G
5479     (Local Points:) LP21$
5480     (Minimum Height [absolut value]:) MH21$
5481     (-) :@G
5482     (Spectrum 2:
5483     Peak picking parameters:) :@G
5485     (Local Points:) LP22$
5486     (Minimum Height [absolut value]:) MH22$
5487     (-) :@G
5488     (Please insert the parameters for the peak picking process.);
5489     %W!page.active = 1;
5490     pkfind dir:=p method:=max npts:=%(LP21$) filter:=h value:=%(MH21$) ocenter_x:=[%(bb.name$)]Sheet1!col(5) ocenter_y:=[%(bb.name$)]Sheet1!col(6) ocenter:=<none>;
5491     %W!page.active = 2;
5492     pkfind dir:=p method:=max npts:=%(LP22$) filter:=h value:=%(MH22$) ocenter_x:=[%(bb.name$)]Sheet1!col(7) ocenter_y:=[%(bb.name$)]Sheet1!col(8) ocenter:=<none>;
5493   };
5494   //Autopeakpicking for 3-stack plot with individual parameters
5495   if (page.nlayers == 3) {
5496     getn
5497     (Spectrum 1:
5498     Peak picking parameters:) :@G
5500     (Local Points:) LP31$
5501     (Minimum Height [absolut value]:) MH31$
5502     (-) :@G
5503     (Spectrum 2:
5504     Peak picking parameters:) :@G
5506     (Local Points:) LP32$
5507     (Minimum Height [absolut value]:) MH32$
5508     (-) :@G
5509     (Spectrum 3:
5510     Peak picking parameters:) :@G
5512     (Local Points:) LP33$
5513     (Minimum Height [absolut value]:) MH33$
5514     (-) :@G
5515     (Please insert the parameters for the peak picking process.);
5516     %W!page.active = 1;
5517     pkfind dir:=p method:=max npts:=%(LP31$) filter:=h value:=%(MH31$) ocenter_x:=[%(bb.name$)]Sheet1!col(7) ocenter_y:=[%(bb.name$)]Sheet1!col(8) ocenter:=<none>;
5518     %W!page.active = 2;
5519     pkfind dir:=p method:=max npts:=%(LP32$) filter:=h value:=%(MH32$) ocenter_x:=[%(bb.name$)]Sheet1!col(9) ocenter_y:=[%(bb.name$)]Sheet1!col(10) ocenter:=<none>;
5520     %W!page.active = 3;
5521     pkfind dir:=p method:=max npts:=%(LP33$) filter:=h value:=%(MH33$) ocenter_x:=[%(bb.name$)]Sheet1!col(11) ocenter_y:=[%(bb.name$)]Sheet1!col(12) ocenter:=<none>;
5522   };
5523   //Autopeakpicking for 4-stack plot with individual parameters
5524   if (page.nlayers == 4) {
5525     getn
5526     (Spectrum 1:
5527     Peak picking parameters:) :@G
5529     (Local Points:) LP41$
5530     (Minimum Height [absolut value]:) MH41$
5531     (-) :@G
5532     (Spectrum 2:
5533     Peak picking parameters:) :@G
5534     (Spectrum 3:
5535     Peak picking parameters:) :@G
5536     (Spectrum 4:
5537     Peak picking parameters:) :@G
5538     (Local Points:) LP42$
5539     (Minimum Height [absolut value]:) MH42$
5540     (Maximum Half Width:) HW42$
5541     (-) :@G
5542     (Please insert the parameters for the peak picking process.);
5543     %W!page.active = 1;
5544     pkfind dir:=p method:=max npts:=%(LP41$) filter:=h value:=%(MH41$) ocenter_x:=[%(bb.name$)]Sheet1!col(13) ocenter_y:=[%(bb.name$)]Sheet1!col(14) ocenter:=<none>;
5545     %W!page.active = 2;
5546     pkfind dir:=p method:=max npts:=%(LP42$) filter:=h value:=%(MH42$) ocenter_x:=[%(bb.name$)]Sheet1!col(15) ocenter_y:=[%(bb.name$)]Sheet1!col(16) ocenter:=<none>;
5547     %W!page.active = 3;
5548     pkfind dir:=p method:=max npts:=%(LP43$) filter:=h value:=%(MH43$) ocenter_x:=[%(bb.name$)]Sheet1!col(17) ocenter_y:=[%(bb.name$)]Sheet1!col(18) ocenter:=<none>;
5549     %W!page.active = 4;
5550     pkfind dir:=p method:=max npts:=%(LP44$) filter:=h value:=%(MH44$) ocenter_x:=[%(bb.name$)]Sheet1!col(19) ocenter_y:=[%(bb.name$)]Sheet1!col(20) ocenter:=<none>;
5551   };
5552 }
5553 }
```

Appendix

LabTalk script (OriginPro 2020b)

```
5535 (Local Points:) LP42$
5536 (Minimum Height [absolut value]:) MH42$
5537 (-) :@G
5538 (Spectrum 3:
5539
5540 Peak picking parameters:) :@G
5541 (Local Points:) LP43$
5542 (Minimum Height [absolut value]:) MH43$
5543 (-) :@G
5544 (Spectrum 4:
5545
5546 Peak picking parameters:) :@G
5547 (Local Points:) LP44$
5548 (Minimum Height [absolut value]:) MH44$
5549 (-) :@G
5550 (Please insert the parameters for the peak picking process.);
5551 %W!page.active = 1;
5552 pkfind dir:=p method:=max npts:=%(LP41$) filter:=h
value:=%(MH41$) ocenter_x:=[%(bb.name$)]Sheet1!col(9) ocenter_y:=[%(bb.name$)]Sheet1!col(10) ocenter:=<none>;
5553 %W!page.active = 2;
5554 pkfind dir:=p method:=max npts:=%(LP42$) filter:=h
value:=%(MH42$) ocenter_x:=[%(bb.name$)]Sheet1!col(11) ocenter_y:=[%(bb.name$)]Sheet1!col(12) ocenter:=<none>;
5555 %W!page.active = 3;
5556 pkfind dir:=p method:=max npts:=%(LP43$) filter:=h
value:=%(MH43$) ocenter_x:=[%(bb.name$)]Sheet1!col(13) ocenter_y:=[%(bb.name$)]Sheet1!col(14) ocenter:=<none>;
5557 %W!page.active = 4;
5558 pkfind dir:=p method:=max npts:=%(LP44$) filter:=h
value:=%(MH44$) ocenter_x:=[%(bb.name$)]Sheet1!col(15) ocenter_y:=[%(bb.name$)]Sheet1!col(16) ocenter:=<none>;
5559 };
5560 //Autopeakpicking for 5-stack plot with individual parameters
5561 if (page.nlayers == 5) {
5562 getn
5563 (Spectrum 1:
5564
5565 Peak picking parameters:) :@G
5566 (Local Points:) LP51$
5567 (Minimum Height [absolut value]:) MH51$
5568 (-) :@G
5569 (Spectrum 2:
5570
5571 Peak picking parameters:) :@G
5572 (Local Points:) LP52$
5573 (Minimum Height [absolut value]:) MH52$
5574 (-) :@G
5575 (Spectrum 3:
5576
5577 Peak picking parameters:) :@G
5578 (Local Points:) LP53$
5579 (Minimum Height [absolut value]:) MH53$
5580 (-) :@G
5581 (Spectrum 4:
5582
5583 Peak picking parameters:) :@G
5584 (Local Points:) LP54$
5585 (Minimum Height [absolut value]:) MH54$
5586 (-) :@G
5587 (Spectrum 5:
5588
5589 Peak picking parameters:) :@G
5590 (Local Points:) LP55$
5591 (Minimum Height [absolut value]:) MH55$
5592 (-) :@G
5593 (Please insert the parameters for the peak picking process.);
5594 %W!page.active = 1;
5595 pkfind dir:=p method:=max npts:=%(LP51$) filter:=h
value:=%(MH51$) ocenter_x:=[%(bb.name$)]Sheet1!col(11) ocenter_y:=[%(bb.name$)]Sheet1!col(12) ocenter:=<none>;
5596 %W!page.active = 2;
5597 pkfind dir:=p method:=max npts:=%(LP52$) filter:=h
value:=%(MH52$) ocenter_x:=[%(bb.name$)]Sheet1!col(13) ocenter_y:=[%(bb.name$)]Sheet1!col(14) ocenter:=<none>;
5598 %W!page.active = 3;
5599 pkfind dir:=p method:=max npts:=%(LP53$) filter:=h
value:=%(MH53$) ocenter_x:=[%(bb.name$)]Sheet1!col(15) ocenter_y:=[%(bb.name$)]Sheet1!col(16) ocenter:=<none>;
5600 %W!page.active = 4;
5601 pkfind dir:=p method:=max npts:=%(LP54$) filter:=h
value:=%(MH54$) ocenter_x:=[%(bb.name$)]Sheet1!col(17) ocenter_y:=[%(bb.name$)]Sheet1!col(18) ocenter:=<none>;
5602 %W!page.active = 5;
5603 pkfind dir:=p method:=max npts:=%(LP55$) filter:=h
value:=%(MH55$) ocenter_x:=[%(bb.name$)]Sheet1!col(19) ocenter_y:=[%(bb.name$)]Sheet1!col(20) ocenter:=<none>;
5604 };
5605 //Autopeakpicking for 6-stack plot with individual parameters
5606 if (page.nlayers == 6) {
5607 HW61$=0.5;
5608 HW62$=0.5;
5609 HW63$=0.5;
5610 HW64$=0.5;
5611 HW65$=0.5;
5612 HW66$=0.5;
5613 getn
5614 (Spectrum 1:
5615
5616 Peak picking parameters:) :@G
5617 (Local Points:) LP61$
5618 (Minimum Height [absolut value]:) MH61$
5619 (Maximum Half Width:) HW61$
5620 (-) :@G
5621 (Spectrum 2:
5622
5623 Peak picking parameters:) :@G
5624 (Local Points:) LP62$
5625 (Minimum Height [absolut value]:) MH62$
5626 (Maximum Half Width:) HW62$
5627 (-) :@G
5628 (Spectrum 3:
5629
5630 Peak picking parameters:) :@G
5631 (Local Points:) LP63$
5632 (Minimum Height [absolut value]:) MH63$
5633 (Maximum Half Width:) HW63$
5634 (-) :@G
5635 (Spectrum 4:
5636
5637 Peak picking parameters:) :@G
5638 (Local Points:) LP64$
5639 (Minimum Height [absolut value]:) MH64$
5640 (Maximum Half Width:) HW64$
5641 (-) :@G
5642 (Spectrum 5:
5643
5644 Peak picking parameters:) :@G
5645 (Local Points:) LP65$
5646 (Minimum Height [absolut value]:) MH65$
5647 (Maximum Half Width:) HW65$
5648 (-) :@G
5649 (Spectrum 6:
5650
5651 Peak picking parameters:) :@G
```



```

5652 (Local Points:) LP66$
5653 (Minimum Height [absolut value:] MH66$
5654 (Maximum Half Width:) HW66$
5655 (-) :@G
5656 (Please insert the parameters for the peak picking process.);
5657 %W!page.active = 1;
5658 pkfind dir:=p method:=max npts:=%(LP61$)
hwidth:=%(HW61$) filter:=h value:=%(MH61$) ocenter_x:=[%(bb.name$)]Sheet1!col(13) ocenter_y:=[%(bb.name$)]Sheet1!col(14) oleft_x:=[%(bb.name$)]Sheet1!col(25) oleft_y:=[%(bb.name$)]Sheet1!col(26) oright_x:=[%(bb.name$)]Sheet1!col(27) oright_y:=[%(bb.name$)]Sheet1!col(28) ocenter:=<none>;
5659 %W!page.active = 2;
5660 pkfind dir:=p method:=max npts:=%(LP62$)
hwidth:=%(HW62$) filter:=h value:=%(MH62$) ocenter_x:=[%(bb.name$)]Sheet1!col(15) ocenter_y:=[%(bb.name$)]Sheet1!col(16) oleft_x:=[%(bb.name$)]Sheet1!col(29) oleft_y:=[%(bb.name$)]Sheet1!col(30) oright_x:=[%(bb.name$)]Sheet1!col(31) oright_y:=[%(bb.name$)]Sheet1!col(32) ocenter:=<none>;
5661 %W!page.active = 3;
5662 pkfind dir:=p method:=max npts:=%(LP63$)
hwidth:=%(HW63$) filter:=h value:=%(MH63$) ocenter_x:=[%(bb.name$)]Sheet1!col(17) ocenter_y:=[%(bb.name$)]Sheet1!col(18) oleft_x:=[%(bb.name$)]Sheet1!col(33) oleft_y:=[%(bb.name$)]Sheet1!col(34) oright_x:=[%(bb.name$)]Sheet1!col(35) oright_y:=[%(bb.name$)]Sheet1!col(36) ocenter:=<none>;
5663 %W!page.active = 4;
5664 pkfind dir:=p method:=max npts:=%(LP64$)
hwidth:=%(HW64$) filter:=h value:=%(MH64$) ocenter_x:=[%(bb.name$)]Sheet1!col(19) ocenter_y:=[%(bb.name$)]Sheet1!col(20) oleft_x:=[%(bb.name$)]Sheet1!col(37) oleft_y:=[%(bb.name$)]Sheet1!col(38) oright_x:=[%(bb.name$)]Sheet1!col(39) oright_y:=[%(bb.name$)]Sheet1!col(40) ocenter:=<none>;
5665 %W!page.active = 5;
5666 pkfind dir:=p method:=max npts:=%(LP65$)
hwidth:=%(HW65$) filter:=h value:=%(MH65$) ocenter_x:=[%(bb.name$)]Sheet1!col(21) ocenter_y:=[%(bb.name$)]Sheet1!col(22) oleft_x:=[%(bb.name$)]Sheet1!col(41) oleft_y:=[%(bb.name$)]Sheet1!col(42) oright_x:=[%(bb.name$)]Sheet1!col(43) oright_y:=[%(bb.name$)]Sheet1!col(44) ocenter:=<none>;
5667 %W!page.active = 6;
5668 pkfind dir:=p method:=max npts:=%(LP66$)
hwidth:=%(HW66$) filter:=h value:=%(MH66$) ocenter_x:=[%(bb.name$)]Sheet1!col(23) ocenter_y:=[%(bb.name$)]Sheet1!col(24) oleft_x:=[%(bb.name$)]Sheet1!col(45) oleft_y:=[%(bb.name$)]Sheet1!col(46) oright_x:=[%(bb.name$)]Sheet1!col(47) oright_y:=[%(bb.name$)]Sheet1!col(48) ocenter:=<none>;
5669 };
5670 };
5671 }
5672 else type -b "Please start again and choose only one type of autpeakpicking!";
5673 };

```

8.1.6. Section "deletesignals" (L. 5675 - 5938)

```

5675 [deletesignals]
5676 type -b "Please use the data reader to choose the signals you want to delete. Press Enter after you marked the signals. When you are done press Esc and the deletion begins.";
5677 //Set variables and start data reader to choose values
5678 @global=1;
5679 dataset dsx;
5680 pp=0;
5681 dotool 3;
5682 def pointproc {
5683 pp++;
5684 dsx[pp] = X;
5685 };
5686 //Define what happens to the chosen values
5687 def quittoolbox {
5688 %W=%H;
5689 range -w aa=%C;
5690 range bb = uid2range(aa.GetPage());$;
5691 //Deletion of signals in 1-stack plot
5692 if (page.nlayers == 1) {
5693 win -a %(bb.name$);
5694 for(int ii=1; ii<=pp; ii++){
5695 int dels=index(dsx[ii].col(3));
5696 wks.deleteRows(dels, 1, 3, 4);
5697 };
5698 win -ia;
5699 win -a %W;
5700 win -z;
5701 };
5702 //Deletion of signals in 2-stack plot
5703 if (page.nlayers == 2) {
5704 //Deletion of signals in layer 1
5705 if (page.active == 1) {
5706 win -a %(bb.name$);
5707 for(int ii=1; ii<=pp; ii++){
5708 int dels=index(dsx[ii].col(5));
5709 wks.deleteRows(dels, 1, 5, 6);
5710 };
5711 win -ia;
5712 win -a %W;
5713 win -z;
5714 };
5715 //Deletion of signals in layer 2
5716 if (page.active == 2) {
5717 win -a %(bb.name$);
5718 for(int ii=1; ii<=pp; ii++){
5719 int dels=index(dsx[ii].col(7));
5720 wks.deleteRows(dels, 1, 7, 8);
5721 };
5722 win -ia;
5723 win -a %W;
5724 win -z;
5725 };
5726 };
5727 //Deletion of signals in 3-stack plot
5728 if (page.nlayers == 3) {
5729 //Deletion of signals in layer 1
5730 if (page.active == 1) {
5731 win -a %(bb.name$);
5732 for(int ii=1; ii<=pp; ii++){
5733 int dels=index(dsx[ii].col(7));
5734 wks.deleteRows(dels, 1, 7, 8);
5735 };
5736 win -ia;
5737 win -a %W;
5738 win -z;
5739 };

```

```
5740 //Deletion of signals in layer 2
5741 if (page.active == 2) {
5742 win -a %(bb.name$);
5743 for(int ii=1; ii<=pp; ii++){
5744 int dels=index(dsx[ii],col(9));
5745 wks.deleteRows(dels, 1, 9, 10);
5746 };
5747 win -ia;
5748 win -a %W;
5749 win -z;
5750 };
5751 //Deletion of signals in layer 3
5752 if (page.active == 3) {
5753 win -a %(bb.name$);
5754 for(int ii=1; ii<=pp; ii++){
5755 int dels=index(dsx[ii],col(11));
5756 wks.deleteRows(dels, 1, 11, 12);
5757 };
5758 win -ia;
5759 win -a %W;
5760 win -z;
5761 };
5762 };
5763 //Deletion of signals in 4-stack plot
5764 if (page.nlayers == 4) {
5765 //Deletion of signals in layer 1
5766 if (page.active == 1) {
5767 win -a %(bb.name$);
5768 for(int ii=1; ii<=pp; ii++){
5769 int dels=index(dsx[ii],col(9));
5770 wks.deleteRows(dels, 1, 9, 10);
5771 };
5772 win -ia;
5773 win -a %W;
5774 win -z;
5775 };
5776 //Deletion of signals in layer 2
5777 if (page.active == 2) {
5778 win -a %(bb.name$);
5779 for(int ii=1; ii<=pp; ii++){
5780 int dels=index(dsx[ii],col(11));
5781 wks.deleteRows(dels, 1, 11, 12);
5782 };
5783 win -ia;
5784 win -a %W;
5785 win -z;
5786 };
5787 //Deletion of signals in layer 3
5788 if (page.active == 3) {
5789 win -a %(bb.name$);
5790 for(int ii=1; ii<=pp; ii++){
5791 int dels=index(dsx[ii],col(13));
5792 wks.deleteRows(dels, 1, 13, 14);
5793 };
5794 win -ia;
5795 win -a %W;
5796 win -z;
5797 };
5798 //Deletion of signals in layer 4
5799 if (page.active == 4) {
5800 win -a %(bb.name$);
5801 for(int ii=1; ii<=pp; ii++){
5802 int dels=index(dsx[ii],col(15));
5803 wks.deleteRows(dels, 1, 15, 16);
5804 };
5805 win -ia;
5806 win -a %W;
5807 win -z;
5808 };
5809 };
5810 //Deletion of signals in 5-stack plot
5811 if (page.nlayers == 5) {
5812 //Deletion of signals in layer 1
5813 if (page.active == 1) {
5814 win -a %(bb.name$);
5815 for(int ii=1; ii<=pp; ii++){
5816 int dels=index(dsx[ii],col(11));
5817 wks.deleteRows(dels, 1, 11, 12);
5818 };
5819 win -ia;
5820 win -a %W;
5821 win -z;
5822 };
5823 //Deletion of signals in layer 2
5824 if (page.active == 2) {
5825 win -a %(bb.name$);
5826 for(int ii=1; ii<=pp; ii++){
5827 int dels=index(dsx[ii],col(13));
5828 wks.deleteRows(dels, 1, 13, 14);
5829 };
5830 win -ia;
5831 win -a %W;
5832 win -z;
5833 };
5834 //Deletion of signals in layer 3
5835 if (page.active == 3) {
5836 win -a %(bb.name$);
5837 for(int ii=1; ii<=pp; ii++){
5838 int dels=index(dsx[ii],col(15));
5839 wks.deleteRows(dels, 1, 15, 16);
5840 };
5841 win -ia;
5842 win -a %W;
5843 win -z;
5844 };
5845 //Deletion of signals in layer 4
5846 if (page.active == 4) {
5847 win -a %(bb.name$);
5848 for(int ii=1; ii<=pp; ii++){
5849 int dels=index(dsx[ii],col(17));
5850 wks.deleteRows(dels, 1, 17, 18);
5851 };
5852 win -ia;
5853 win -a %W;
5854 win -z;
5855 };
5856 //Deletion of signals in layer 5
5857 if (page.active == 5) {
5858 win -a %(bb.name$);
5859 for(int ii=1; ii<=pp; ii++){
5860 int dels=index(dsx[ii],col(19));
5861 wks.deleteRows(dels, 1, 19, 20);
5862 };
5863 win -ia;
5864 win -a %W;
5865 win -z;
5866 };
5867 };
5868 //Deletion of signals in 6-stack plot
5869 if (page.nlayers == 6) {
5870 //Deletion of signals in layer 1
5871 if (page.active == 1) {
5872 win -a %(bb.name$);
5873 for(int ii=1; ii<=pp; ii++){
5874 int dels=index(dsx[ii],col(13));
5875 wks.deleteRows(dels, 1, 13, 14);
5876 };
5877 win -ia;
5878 win -a %W;
5879 win -z;
5880 };
5881 //Deletion of signals in layer 2
5882 if (page.active == 2) {
5883 win -a %(bb.name$);
5884 for(int ii=1; ii<=pp; ii++){
5885 int dels=index(dsx[ii],col(15));
5886 wks.deleteRows(dels, 1, 15, 16);
5887 };
5888 win -ia;
```

```

5889 win -a %W;
5890 win -z;
5891 };
5892 //Deletion of signals in layer 3
5893 if (page.active == 3) {
5894 win -a %(bb.name$);
5895 for(int ii=1; ii<=pp; ii++){
5896 int dels=index(dsx[ii],col(17));
5897 wks.deleteRows(dels, 1, 17, 18);
5898 };
5899 win -ia;
5900 win -a %W;
5901 win -z;
5902 };
5903 //Deletion of signals in layer 4
5904 if (page.active == 4) {
5905 win -a %(bb.name$);
5906 for(int ii=1; ii<=pp; ii++){
5907 int dels=index(dsx[ii],col(19));
5908 wks.deleteRows(dels, 1, 19, 20);
5909 };
5910 win -ia;
5911 win -a %W;
5912 win -z;
5913 };
5914 //Deletion of signals in layer 5

```

```

5915 if (page.active == 5) {
5916 win -a %(bb.name$);
5917 for(int ii=1; ii<=pp; ii++){
5918 int dels=index(dsx[ii],col(21));
5919 wks.deleteRows(dels, 1, 21, 22);
5920 };
5921 win -ia;
5922 win -a %W;
5923 win -z;
5924 };
5925 //Deletion of signals in layer 6
5926 if (page.active == 6) {
5927 win -a %(bb.name$);
5928 for(int ii=1; ii<=pp; ii++){
5929 int dels=index(dsx[ii],col(23));
5930 wks.deleteRows(dels, 1, 23, 24);
5931 };
5932 win -ia;
5933 win -a %W;
5934 win -z;
5935 };
5936 };
5937 };
5938 };

```

8.1.7. Section "PDI" (L. 5940 - 7718)

```

5940 [PDI]
5941 type -b "This skript will calculate Mn, Mw, PDI, the de-
5942 gree of polymerization (DP) and the avarage molar
5943 weight of the repeating unit (incl. standard error). Before
5944 you start this script, make sure that you picked ONLY
5945 signals related to your polymer. This is the case when
5946 the differences between the picked signals are equal to
5947 the molar mass of the repeating unit.";
5948 %D=%H;
5949 //Calculation script for 1 layer plot
5950 if (page.nlayers == 1) {
5951 //Calculations for the 1 layer
5952 if (page.active == 1){
5953 range -w aa=%C;
5954 range bb = uid2range(aa.GetPage());$;
5955 win -a %(bb.name$);
5956 %V=%H;
5957 //Creating of a dummy workbook and importing of nec-
5958 essary data
5959 newbook dummy;
5960 %B=%H;
5961 win -a %V;
5962 colcopy irng:=[%V]Sheet1!Col(3)
5963 orng:=[%B]Sheet1!Col(1);
5964 colcopy irng:=[%V]Sheet1!Col(4)
5965 orng:=[%B]Sheet1!Col(2);
5966 win -a %B;
5967 //Calculations of Mn, Mw and PDI
5968 col(c)=col(a)*col(b);
5969 col(d)=col(a)^2*col(b);
5970 colstats 2 sum:=[%B]Sheet2!col(a)[1];
5971 copydata irng:=[%B]Sheet2!col(a)[1]
5972 orng:=[%B]Sheet1!col(5);
5973 page.active$=Sheet2;
5974 layer -d;
5975 colstats 3 sum:=[%B]Sheet2!col(a)[1];
5976 copydata irng:=[%B]Sheet2!col(a)[1]
5977 orng:=[%B]Sheet1!col(6);
5978 page.active$=Sheet2;
5979 layer -d;
5980 colstats 4 sum:=[%B]Sheet2!col(a)[1];
5981 copydata irng:=[%B]Sheet2!col(a)[1]
5982 orng:=[%B]Sheet1!col(7);
5983 page.active$=Sheet2;
5984 layer -d;
5985 col(h)=col(f)/col(e);
5986 col(i)=col(g)/col(f);
5987 col(j)=col(i)/col(h);
5988 col(a)=round(col(a),0);
5989 int NoS=count(col(a),1);
5990 for (ii = 1; ii<=NoS; ii++) {
5991 col(k)[ii]=col(a)[ii+1]-col(a)[ii];
5992 };
5993 colstats 11 mean:=col(l)[1] sem:=<new>;
5994 win -a %V;
5995 //Collecting and presenting of calculated data
5996 %P="Mn/Mw/PDI";
5997 if (exist([%V]!P) != 2) {
5998 newsheet name:=%P cols:=2;
5999 wks.col1.lname$ = "Name";
6000 wks.col2.lname$ = "1. Spectrum";
6001 }
6002 else {
6003 page.active$=%P;
6004 };
6005 copydata irng:=[%B]sheet1!col(h)[1]
6006 orng:=[%V]!P!col(B)[1] clear:=1;
6007 copydata irng:=[%B]sheet1!col(i)[1]
6008 orng:=[%V]!P!col(B)[2] clear:=1;
6009 copydata irng:=[%B]sheet1!col(j)[1]
6010 orng:=[%V]!P!col(B)[3] clear:=1;
6011 copydata irng:=[%B]sheet1!col(l)[1]
6012 orng:=[%V]!P!col(B)[4] clear:=1;
6013 copydata irng:=[%B]sheet1!col(n)[1]
6014 orng:=[%V]!P!col(B)[5] clear:=1;
6015 col(B)[6]=col(B)[2]/col(B)[4];
6016 Mn$=[%B]sheet1!col(h)[1]$;
6017 Mw$=[%B]sheet1!col(i)[1]$;
6018 PDI$=[%B]sheet1!col(j)[1]$;
6019 MEAN$=[%B]sheet1!col(l)[1]$;
6020 SEM$=[%B]sheet1!col(n)[1]$;
6021 DP$=[%V]!P!col(B)[6]$;
6022 col(A)[1]$=Mn;
6023 col(A)[2]$=Mw;
6024 col(A)[3]$=PDI;
6025 col(A)[4]$=Molar weight of repeating unit (mean);
6026 col(A)[5]$=Standard error of repeating unit;
6027 col(A)[6]$=Degree of Polymerization;
6028 worksheet -s 1;
6029 menu -e 35154;

```

Appendix

LabTalk script (OriginPro 2020b)

```

6013 wcolwidth irng:=[%V]%P!Col(1) width:=20;
6014 wcolwidth irng:=[%V]%P!Col(2) width:=10;
6015 page.active$=Sheet1;
6016 win -ia;
6017 win -c %B;
6018 win -a %D;
6019 win -z;
6020 type -b "Here are the values for the 1. Spectrum: \n Mn
\t\t\t\t%(Mn$) \n Mw \t\t\t\t%(Mw$) \n PDI \t\t\t\t
%(PDI$) \n Molar weight of repeating unit (mean) \t
%(MEAN$) \n Standard error of repeating unit \t
%(SEM$) \n Degree of Polymerization \t\t %(DP$)";
6021 };
6022 };
6023 //Calculation script for 2 layer plot
6024 if (page.nlayers == 2) {
6025 //Calculations for the 1 layer
6026 if (page.active == 1){
6027 range -w aa=%C;
6028 range bb = uid2range(aa.GetPage());$;
6029 win -a %(bb.name$);
6030 %V=%H;
6031 //Creating of a dummy workbook and importing of nec-
        essary data
6032 newbook dummy;
6033 %B=%H;
6034 win -a %V;
6035 colcopy irng:=[%V]Sheet1!Col(5)
        orng:=[%B]Sheet1!Col(1);
6036 colcopy irng:=[%V]Sheet1!Col(6)
        orng:=[%B]Sheet1!Col(2);
6037 win -a %B;
6038 //Calculations of Mn, Mw and PDI
6039 col(c)=col(a)*col(b);
6040 col(d)=col(a)^2*col(b);
6041 colstats 2 sum:=[%B]Sheet2!col(a)[1];
6042 copydata irng:=[%B]Sheet2!col(a)[1]
        orng:=[%B]Sheet1!col(5);
6043 page.active$=Sheet2;
6044 layer -d;
6045 colstats 3sum:=[%B]Sheet2!col(a)[1];
6046 copydata irng:=[%B]Sheet2!col(a)[1]
        orng:=[%B]Sheet1!col(6);
6047 page.active$=Sheet2;
6048 layer -d;
6049 colstats 4sum:=[%B]Sheet2!col(a)[1];
6050 copydata irng:=[%B]Sheet2!col(a)[1]
        orng:=[%B]Sheet1!col(7);
6051 page.active$=Sheet2;
6052 layer -d;
6053 col(h)=col(f)/col(e);
6054 col(i)=col(g)/col(f);
6055 col(j)=col(i)/col(h);
6056 col(a)=round(col(a),0);
6057 int NoS=count(col(a),1);
6058 for (ii = 1; ii <= NoS; ii++) {
6059 col(k)[ii]=col(a)[ii+1]-col(a)[ii];
6060 };
6061 colstats 11 mean:=col(l)[1] sem:=<new>;
6062 win -a %V;
6063 //Collecting and presenting of calculated data
6064 %P="Mn/Mw/PDI";
6065 if(exist([%V]%P!) != 2) {
6066 newsheet name:=%P cols:=3;
6067 wks.col1.name$="Name";
6068 wks.col2.name$="1. Spectrum";
6069 wks.col3.name$="2. Spectrum";
6070 }
6071 else {
6072 page.active$=%P;
6073 };
6074 copydata irng:=[%B]sheet1!col(h)[1]
        orng:=[%V]%P!col(B)[1] clear:=1;
6075 copydata irng:=[%B]sheet1!col(i)[1]
        orng:=[%V]%P!col(B)[2] clear:=1;
6076 copydata irng:=[%B]sheet1!col(j)[1]
        orng:=[%V]%P!col(B)[3] clear:=1;
6077 copydata irng:=[%B]sheet1!col(l)[1]
        orng:=[%V]%P!col(B)[4] clear:=1;
6078 copydata irng:=[%B]sheet1!col(n)[1]
        orng:=[%V]%P!col(B)[5] clear:=1;
6079 col(b)[6]=col(b)[2]/col(b)[4];
6080 Mn$=[%B]sheet1!col(h)[1]$;
6081 Mw$=[%B]sheet1!col(i)[1]$;
6082 PDI$=[%B]sheet1!col(j)[1]$;
6083 MEAN$=[%B]sheet1!col(l)[1]$;
6084 SEM$=[%B]sheet1!col(n)[1]$;
6085 DP$=[%V]%P!col(B)[6]$;
6086 col(A)[1]=$Mn;
6087 col(A)[2]=$Mw;
6088 col(A)[3]=$PDI;
6089 col(A)[4]=$Molar weight of repeating unit (mean);
6090 col(A)[5]=$Standard error of repeating unit;
6091 col(A)[6]=$Degree of Polymerization;
6092 worksheet -s 1;
6093 menu -e 35154;
6094 wcolwidth irng:=[%V]%P!Col(1) width:=20;
6095 wcolwidth irng:=[%V]%P!Col(2) width:=10;
6096 wcolwidth irng:=[%V]%P!Col(3) width:=10;
6097 page.active$=Sheet1;
6098 win -ia;
6099 win -c %B;
6100 win -a %D;
6101 win -z;
6102 type -b "Here are the values for the 1. Spectrum: \n Mn
\t\t\t\t%(Mn$) \n Mw \t\t\t\t%(Mw$) \n PDI \t\t\t\t
%(PDI$) \n Molar weight of repeating unit (mean) \t
%(MEAN$) \n Standard error of repeating unit \t
%(SEM$) \n Degree of Polymerization \t\t %(DP$)";
6103 };
6104 //Calculations for the 2 layer
6105 if (page.active == 2){
6106 range -w aa=%C;
6107 range bb = uid2range(aa.GetPage());$;
6108 win -a %(bb.name$);
6109 %V=%H;
6110 //Creating of a dummy workbook and importing of nec-
        essary data
6111 newbook dummy;
6112 %B=%H;
6113 win -a %V;
6114 colcopy irng:=[%V]Sheet1!Col(7)
        orng:=[%B]Sheet1!Col(1);
6115 colcopy irng:=[%V]Sheet1!Col(8)
        orng:=[%B]Sheet1!Col(2);
6116 win -a %B;
6117 //Calculations of Mn, Mw and PDI
6118 col(c)=col(a)*col(b);
6119 col(d)=col(a)^2*col(b);
6120 colstats 2 sum:=[%B]Sheet2!col(a)[1];
6121 copydata irng:=[%B]Sheet2!col(a)[1]
        orng:=[%B]Sheet1!col(5);
6122 page.active$=Sheet2;
6123 layer -d;
6124 colstats 3 sum:=[%B]Sheet2!col(a)[1];
6125 copydata irng:=[%B]Sheet2!col(a)[1]
        orng:=[%B]Sheet1!col(6);
6126 page.active$=Sheet2;
6127 layer -d;
6128 colstats 4 sum:=[%B]Sheet2!col(a)[1];
6129 copydata irng:=[%B]Sheet2!col(a)[1]
        orng:=[%B]Sheet1!col(7);
6130 page.active$=Sheet2;
6131 layer -d;
6132 col(h)=col(f)/col(e);
6133 col(i)=col(g)/col(f);
6134 col(j)=col(i)/col(h);
6135 col(a)=round(col(a),0);
6136 int NoS=count(col(a),1);
6137 for (ii = 1; ii <= NoS; ii++) {

```


Appendix

LabTalk script (OriginPro 2020b)

```
6138 col(k)[ii]=col(a)[ii+1]-col(a)[ii];
6139 };
6140 colstats 11 mean:=col(l)[1] sem:=<new>;
6141 win -a %V;
6142 //Collecting and presenting of calculated data
6143 %P="Mn/Mw/PDI";
6144 if (exist([%V]%P!) != 2) {
6145 newsheet name:=%P cols:=3;
6146 wks.col1.lname$ = "Name";
6147 wks.col2.lname$ = "1. Spectrum";
6148 wks.col3.lname$ = "2. Spectrum";
6149 }
6150 else {
6151 page.active$=%P;
6152 };
6153 copydata irng:=[%B]sheet1!col(h)[1]
  orng:=[%V]%P!col(c)[1] clear:=1;
6154 copydata irng:=[%B]sheet1!col(i)[1]
  orng:=[%V]%P!col(c)[2] clear:=1;
6155 copydata irng:=[%B]sheet1!col(j)[1]
  orng:=[%V]%P!col(c)[3] clear:=1;
6156 copydata irng:=[%B]sheet1!col(l)[1]
  orng:=[%V]%P!col(c)[4] clear:=1;
6157 copydata irng:=[%B]sheet1!col(n)[1]
  orng:=[%V]%P!col(c)[5] clear:=1;
6158 col(c)[6]=col(c)[2]/col(c)[4];
6159 Mn$=[%B]sheet1!col(h)[1];
6160 Mw$=[%B]sheet1!col(i)[1];
6161 PDI$=[%B]sheet1!col(j)[1];
6162 MEAN$=[%B]sheet1!col(l)[1];
6163 SEM$=[%B]sheet1!col(n)[1];
6164 DP$=[%V]%P!col(c)[6];
6165 col(A)[1]$=Mn;
6166 col(A)[2]$=Mw;
6167 col(A)[3]$=PDI;
6168 col(A)[4]$=Molar weight of repeating unit (mean);
6169 col(A)[5]$=Standard error of repeating unit;
6170 col(A)[6]$=Degree of Polymerization;
6171 worksheet -s 1;
6172 menu -e 35154;
6173 wcolwidth irng:=[%V]%P!Col(1) width:=20;
6174 wcolwidth irng:=[%V]%P!Col(2) width:=10;
6175 wcolwidth irng:=[%V]%P!Col(3) width:=10;
6176 page.active$=Sheet1;
6177 win -ia;
6178 win -c %B;
6179 win -a %D;
6180 win -z;
6181 type -b "Here are the values for the 2. Spectrum: \n Mn
\t\t\t\t%(Mn$) \n Mw \t\t\t\t%(Mw$) \n PDI \t\t\t\t
%(PDI$) \n Molar weight of repeating unit (mean) \t
%(MEAN$) \n Standard error of repeating unit \t
%(SEM$) \n Degree of Polymerization \t\t %(DP$)";
6182 };
6183 };
6184 //Calculation script for 3 layer plot
6185 if (page.nlayers == 3) {
6186 //Calculations for the 1 layer
6187 if (page.active == 1){
6188 range -w aa=%C;
6189 range bb = uid2range(aa.GetPage());$;
6190 win -a % (bb.name$);
6191 %V=%H;
6192 //Creating of a dummy workbook and importing of nec-
  essary data
6193 newbook dummy;
6194 %B=%H;
6195 win -a %V;
6196 colcopy irng:=[%V]Sheet1!Col(7)
  orng:=[%B]Sheet1!Col(1);
6197 colcopy irng:=[%V]Sheet1!Col(8)
  orng:=[%B]Sheet1!Col(2);
6198 win -a %B;
6199 //Calculations of Mn, Mw and PDI
6200 col(c)=col(a)*col(b);
6201 col(d)=col(a)^2*col(b);
6202 colstats 2 sum:=[%B]Sheet2!col(a)[1];
6203 copydata irng:=[%B]Sheet2!col(a)[1]
  orng:=[%B]Sheet1!col(5);
6204 page.active$=Sheet2;
6205 layer -d;
6206 colstats 3 sum:=[%B]Sheet2!col(a)[1];
6207 copydata irng:=[%B]Sheet2!col(a)[1]
  orng:=[%B]Sheet1!col(6);
6208 page.active$=Sheet2;
6209 layer -d;
6210 colstats 4 sum:=[%B]Sheet2!col(a)[1];
6211 copydata irng:=[%B]Sheet2!col(a)[1]
  orng:=[%B]Sheet1!col(7);
6212 page.active$=Sheet2;
6213 layer -d;
6214 col(h)=col(f)/col(e);
6215 col(i)=col(g)/col(f);
6216 col(j)=col(i)/col(h);
6217 col(a)=round(col(a),0);
6218 int NoS=count(col(a),1);
6219 for (ii = 1; ii<=NoS; ii++) {
6220 col(k)[ii]=col(a)[ii+1]-col(a)[ii];
6221 };
6222 colstats 11 mean:=col(l)[1] sem:=<new>;
6223 win -a %V;
6224 //Collecting and presenting of calculated data
6225 %P="Mn/Mw/PDI";
6226 if (exist([%V]%P!) != 2) {
6227 newsheet name:=%P cols:=4;
6228 wks.col1.lname$ = "Name";
6229 wks.col2.lname$ = "1. Spectrum";
6230 wks.col3.lname$ = "2. Spectrum";
6231 wks.col4.lname$ = "3. Spectrum";
6232 }
6233 else {
6234 page.active$=%P;
6235 };
6236 copydata irng:=[%B]sheet1!col(h)[1]
  orng:=[%V]%P!col(B)[1] clear:=1;
6237 copydata irng:=[%B]sheet1!col(i)[1]
  orng:=[%V]%P!col(B)[2] clear:=1;
6238 copydata irng:=[%B]sheet1!col(j)[1]
  orng:=[%V]%P!col(B)[3] clear:=1;
6239 copydata irng:=[%B]sheet1!col(l)[1]
  orng:=[%V]%P!col(B)[4] clear:=1;
6240 copydata irng:=[%B]sheet1!col(n)[1]
  orng:=[%V]%P!col(B)[5] clear:=1;
6241 col(B)[6]=col(B)[2]/col(B)[4];
6242 Mn$=[%B]sheet1!col(h)[1];
6243 Mw$=[%B]sheet1!col(i)[1];
6244 PDI$=[%B]sheet1!col(j)[1];
6245 MEAN$=[%B]sheet1!col(l)[1];
6246 SEM$=[%B]sheet1!col(n)[1];
6247 DP$=[%V]%P!col(B)[6];
6248 col(A)[1]$=Mn;
6249 col(A)[2]$=Mw;
6250 col(A)[3]$=PDI;
6251 col(A)[4]$=Molar weight of repeating unit (mean);
6252 col(A)[5]$=Standard error of repeating unit;
6253 col(A)[6]$=Degree of Polymerization;
6254 worksheet -s 1;
6255 menu -e 35154;
6256 wcolwidth irng:=[%V]%P!Col(1) width:=20;
6257 wcolwidth irng:=[%V]%P!Col(2) width:=10;
6258 wcolwidth irng:=[%V]%P!Col(3) width:=10;
6259 wcolwidth irng:=[%V]%P!Col(4) width:=10;
6260 page.active$=Sheet1;
6261 win -ia;
6262 win -c %B;
6263 win -a %D;
6264 win -z;
6265 type -b "Here are the values for the 1. Spectrum: \n Mn
\t\t\t\t%(Mn$) \n Mw \t\t\t\t%(Mw$) \n PDI \t\t\t\t
%(PDI$) \n Molar weight of repeating unit (mean) \t
```


Appendix

LabTalk script (OriginPro 2020b)

```
% (MEANS$) \n Standard error of repeating unit \t
%(SEM$) \n Degree of Polymerization \t\t %(DP$)";
6266 };
6267 //Calculations for the 2 layer
6268 if (page.active == 2){
6269 range -w aa=%C;
6270 range bb = uid2range(aa.GetPage());
6271 win -a %(bb.name$);
6272 %V=%H;
6273 //Creating of a dummy workbook and importing of necessary data
6274 newbook dummy;
6275 %B=%H;
6276 win -a %V;
6277 colcopy irng:=[%V]Sheet1!Col(9)
org:=[%B]Sheet1!Col(1);
6278 colcopy irng:=[%V]Sheet1!Col(10)
org:=[%B]Sheet1!Col(2);
6279 win -a %B;
6280 //Calculations of Mn, Mw and PDI
6281 col(c)=col(a)*col(b);
6282 col(d)=col(a)^2*col(b);
6283 colstats 2 sum:=[%B]Sheet2!col(a)[1];
6284 copydata irng:=[%B]Sheet2!col(a)[1]
org:=[%B]Sheet1!col(5);
6285 page.active$=Sheet2;
6286 layer -d;
6287 colstats 3 sum:=[%B]Sheet2!col(a)[1];
6288 copydata irng:=[%B]Sheet2!col(a)[1]
org:=[%B]Sheet1!col(6);
6289 page.active$=Sheet2;
6290 layer -d;
6291 colstats 4 sum:=[%B]Sheet2!col(a)[1];
6292 copydata irng:=[%B]Sheet2!col(a)[1]
org:=[%B]Sheet1!col(7);
6293 page.active$=Sheet2;
6294 layer -d;
6295 col(h)=col(f)/col(e);
6296 col(i)=col(g)/col(f);
6297 col(j)=col(i)/col(h);
6298 col(a)=round(col(a),0);
6299 int NoS=count(col(a),1);
6300 for (ii = 1; ii<=NoS; ii++) {
6301 col(k)[ii]=col(a)[ii+1]-col(a)[ii];
6302 };
6303 colstats 11 mean:=col(l)[1] sem:=<new>;
6304 win -a %V;
6305 //Collecting and presenting of calculated data
6306 %P="Mn/Mw/PDI";
6307 if (exist([%V]!P) != 2) {
6308 newsheet name:=%P cols:=4;
6309 wks.col1.lname$ = "Name";
6310 wks.col2.lname$ = "1. Spectrum";
6311 wks.col3.lname$ = "2. Spectrum";
6312 wks.col4.lname$ = "3. Spectrum";
6313 }
6314 else {
6315 page.active$=%P;
6316 };
6317 copydata irng:=[%B]sheet1!col(h)[1]
org:=[%V]!P!col(C)[1] clear:=1;
6318 copydata irng:=[%B]sheet1!col(i)[1]
org:=[%V]!P!col(C)[2] clear:=1;
6319 copydata irng:=[%B]sheet1!col(j)[1]
org:=[%V]!P!col(C)[3] clear:=1;
6320 copydata irng:=[%B]sheet1!col(l)[1]
org:=[%V]!P!col(C)[4] clear:=1;
6321 copydata irng:=[%B]sheet1!col(n)[1]
org:=[%V]!P!col(C)[5] clear:=1;
6322 col(c)[6]=col(c)[2]/col(c)[4];
6323 Mn$=[%B]sheet1!col(h)[1]$;
6324 Mw$=[%B]sheet1!col(i)[1]$;
6325 PDI$=[%B]sheet1!col(j)[1]$;
6326 MEANS$=[%B]sheet1!col(l)[1]$;
6327 SEM$=[%B]sheet1!col(n)[1]$;
6328 DP$=[%V]!P!col(c)[6]$;
6329 col(A)[1]$=Mn;
6330 col(A)[2]$=Mw;
6331 col(A)[3]$=PDI;
6332 col(A)[4]$=Molar weight of repeating unit (mean);
6333 col(A)[5]$=Standard error of repeating unit;
6334 col(A)[6]$=Degree of Polymerization;
6335 worksheet -s 1;
6336 menu -e 35154;
6337 wcolwidth irng:=[%V]!P!Col(1) width:=20;
6338 wcolwidth irng:=[%V]!P!Col(2) width:=10;
6339 wcolwidth irng:=[%V]!P!Col(3) width:=10;
6340 wcolwidth irng:=[%V]!P!Col(4) width:=10;
6341 page.active$=Sheet1;
6342 win -ia;
6343 win -c %B;
6344 win -a %D;
6345 win -z;
6346 type -b "Here are the values for the 2. Spectrum: \n Mn
\t\t\t\t %(Mn$) \n Mw \t\t\t\t %(Mw$) \n PDI \t\t\t\t
%(PDI$) \n Molar weight of repeating unit (mean) \t
%(MEANS$) \n Standard error of repeating unit \t
%(SEM$) \n Degree of Polymerization \t\t %(DP$)";
6347 };
6348 //Calculations for the 3 layer
6349 if (page.active == 3){
6350 range -w aa=%C;
6351 range bb = uid2range(aa.GetPage());
6352 win -a %(bb.name$);
6353 %V=%H;
6354 //Creating of a dummy workbook and importing of necessary data
6355 newbook dummy;
6356 %B=%H;
6357 win -a %V;
6358 colcopy irng:=[%V]Sheet1!Col(11)
org:=[%B]Sheet1!Col(1);
6359 colcopy irng:=[%V]Sheet1!Col(12)
org:=[%B]Sheet1!Col(2);
6360 win -a %B;
6361 //Calculations of Mn, Mw and PDI
6362 col(c)=col(a)*col(b);
6363 col(d)=col(a)^2*col(b);
6364 colstats 2 sum:=[%B]Sheet2!col(a)[1];
6365 copydata irng:=[%B]Sheet2!col(a)[1]
org:=[%B]Sheet1!col(5);
6366 page.active$=Sheet2;
6367 layer -d;
6368 colstats 3 sum:=[%B]Sheet2!col(a)[1];
6369 copydata irng:=[%B]Sheet2!col(a)[1]
org:=[%B]Sheet1!col(6);
6370 page.active$=Sheet2;
6371 layer -d;
6372 colstats 4 sum:=[%B]Sheet2!col(a)[1];
6373 copydata irng:=[%B]Sheet2!col(a)[1]
org:=[%B]Sheet1!col(7);
6374 page.active$=Sheet2;
6375 layer -d;
6376 col(h)=col(f)/col(e);
6377 col(i)=col(g)/col(f);
6378 col(j)=col(i)/col(h);
6379 col(a)=round(col(a),0);
6380 int NoS=count(col(a),1);
6381 for (ii = 1; ii<=NoS; ii++) {
6382 col(k)[ii]=col(a)[ii+1]-col(a)[ii];
6383 };
6384 colstats 11 mean:=col(l)[1] sem:=<new>;
6385 win -a %V;
6386 //Collecting and presenting of calculated data
6387 %P="Mn/Mw/PDI";
6388 if (exist([%V]!P) != 2) {
6389 newsheet name:=%P cols:=4;
6390 wks.col1.lname$ = "Name";
6391 wks.col2.lname$ = "1. Spectrum";
6392 wks.col3.lname$ = "2. Spectrum";
```

Appendix

LabTalk script (OriginPro 2020b)

```
6393 wks.col4.lname$ = "3. Spectrum";
6394 }
6395 else {
6396 page.active$=%P;
6397 };
6398 copydata irng:=[%B]sheet1!col(h)[1]
  org:=[%V]%P!col(D)[1] clear:=1;
6399 copydata irng:=[%B]sheet1!col(i)[1]
  org:=[%V]%P!col(D)[2] clear:=1;
6400 copydata irng:=[%B]sheet1!col(j)[1]
  org:=[%V]%P!col(D)[3] clear:=1;
6401 copydata irng:=[%B]sheet1!col(l)[1]
  org:=[%V]%P!col(D)[4] clear:=1;
6402 copydata irng:=[%B]sheet1!col(n)[1]
  org:=[%V]%P!col(D)[5] clear:=1;
6403 col(d)[6]=col(d)[2]/col(d)[4];
6404 Mn$=[%B]sheet1!col(h)[1]$;
6405 Mw$=[%B]sheet1!col(i)[1]$;
6406 PDI$=[%B]sheet1!col(j)[1]$;
6407 MEAN$=[%B]sheet1!col(l)[1]$;
6408 SEM$=[%B]sheet1!col(n)[1]$;
6409 DP$=[%V]%P!col(d)[6]$;
6410 col(A)[1]$=Mn;
6411 col(A)[2]$=Mw;
6412 col(A)[3]$=PDI;
6413 col(A)[4]$=Molar weight of repeating unit (mean);
6414 col(A)[5]$=Standard error of repeating unit;
6415 col(A)[6]$=Degree of Polymerization;
6416 worksheet -s 1;
6417 menu -e 35154;
6418 wcolwidth irng:=[%V]%P!Col(1) width:=20;
6419 wcolwidth irng:=[%V]%P!Col(2) width:=10;
6420 wcolwidth irng:=[%V]%P!Col(3) width:=10;
6421 wcolwidth irng:=[%V]%P!Col(4) width:=10;
6422 page.active$=Sheet1;
6423 win -ia;
6424 win -c %B;
6425 win -a %D;
6426 win -z;
6427 type -b "Here are the values for the 3. Spectrum: \n Mn
\t\t\t\t%(Mn$) \n Mw \t\t\t\t%(Mw$) \n PDI \t\t\t\t
%(PDI$) \n Molar weight of repeating unit (mean) \t
%(MEAN$) \n Standard error of repeating unit \t
%(SEM$) \n Degree of Polymerization \t\t %(DP$)";
6428 };
6429 };
6430 //Calculation script for 4 layer plot
6431 if (page.nlayers == 4) {
6432 //Calculations for the 1 layer
6433 if (page.active == 1){
6434 range -w aa=%C;
6435 range bb = uid2range(aa.GetPage());$;
6436 win -a %(bb.name$);
6437 %V=%H;
6438 //Creating of a dummy workbook and importing of nec-
  essary data
6439 newbook dummy;
6440 %B=%H;
6441 win -a %V;
6442 colcopy irng:=[%V]Sheet1!Col(9)
  org:=[%B]Sheet1!Col(1);
6443 colcopy irng:=[%V]Sheet1!Col(10)
  org:=[%B]Sheet1!Col(2);
6444 win -a %B;
6445 //Calculations of Mn, Mw and PDI
6446 col(c)=col(a)*col(b);
6447 col(d)=col(a)^2*col(b);
6448 colstats 2 sum:=[%B]Sheet2!col(a)[1];
6449 copydata irng:=[%B]Sheet2!col(a)[1]
  org:=[%B]Sheet1!col(5);
6450 page.active$=Sheet2;
6451 layer -d;
6452 colstats 3 sum:=[%B]Sheet2!col(a)[1];
6453 copydata irng:=[%B]Sheet2!col(a)[1]
  org:=[%B]Sheet1!col(6);
6454 page.active$=Sheet2;
6455 layer -d;
6456 colstats 4 sum:=[%B]Sheet2!col(a)[1];
6457 copydata irng:=[%B]Sheet2!col(a)[1]
  org:=[%B]Sheet1!col(7);
6458 page.active$=Sheet2;
6459 layer -d;
6460 col(h)=col(f)/col(e);
6461 col(i)=col(g)/col(f);
6462 col(j)=col(i)/col(h);
6463 col(a)=round(col(a),0);
6464 int NoS=count(col(a),1);
6465 for (ii = 1; ii<=NoS; ii++) {
6466 col(k)[ii]=col(a)[ii+1]-col(a)[ii];
6467 };
6468 colstats 11 mean:=col(l)[1] sem:=<new>;
6469 win -a %V;
6470 //Collecting and presenting of calculated data
6471 %P="Mn/Mw/PDI";
6472 if (exist([%V]%P!) != 2) {
6473 newsheet name:=%P cols:=5;
6474 wks.col1.lname$ = "Name";
6475 wks.col2.lname$ = "1. Spectrum";
6476 wks.col3.lname$ = "2. Spectrum";
6477 wks.col4.lname$ = "3. Spectrum";
6478 wks.col5.lname$ = "4. Spectrum";
6479 }
6480 else {
6481 page.active$=%P;
6482 };
6483 copydata irng:=[%B]sheet1!col(h)[1]
  org:=[%V]%P!col(B)[1] clear:=1;
6484 copydata irng:=[%B]sheet1!col(i)[1]
  org:=[%V]%P!col(B)[2] clear:=1;
6485 copydata irng:=[%B]sheet1!col(j)[1]
  org:=[%V]%P!col(B)[3] clear:=1;
6486 copydata irng:=[%B]sheet1!col(l)[1]
  org:=[%V]%P!col(B)[4] clear:=1;
6487 copydata irng:=[%B]sheet1!col(n)[1]
  org:=[%V]%P!col(B)[5] clear:=1;
6488 col(B)[6]=col(B)[2]/col(B)[4];
6489 Mn$=[%B]sheet1!col(h)[1]$;
6490 Mw$=[%B]sheet1!col(i)[1]$;
6491 PDI$=[%B]sheet1!col(j)[1]$;
6492 MEAN$=[%B]sheet1!col(l)[1]$;
6493 SEM$=[%B]sheet1!col(n)[1]$;
6494 DP$=[%V]%P!col(B)[6]$;
6495 col(A)[1]$=Mn;
6496 col(A)[2]$=Mw;
6497 col(A)[3]$=PDI;
6498 col(A)[4]$=Molar weight of repeating unit (mean);
6499 col(A)[5]$=Standard error of repeating unit;
6500 col(A)[6]$=Degree of Polymerization;
6501 worksheet -s 1;
6502 menu -e 35154;
6503 wcolwidth irng:=[%V]%P!Col(1) width:=20;
6504 wcolwidth irng:=[%V]%P!Col(2) width:=10;
6505 wcolwidth irng:=[%V]%P!Col(3) width:=10;
6506 wcolwidth irng:=[%V]%P!Col(4) width:=10;
6507 wcolwidth irng:=[%V]%P!Col(5) width:=10;
6508 page.active$=Sheet1;
6509 win -ia;
6510 win -c %B;
6511 win -a %D;
6512 win -z;
6513 type -b "Here are the values for the 1. Spectrum: \n Mn
\t\t\t\t%(Mn$) \n Mw \t\t\t\t%(Mw$) \n PDI \t\t\t\t
%(PDI$) \n Molar weight of repeating unit (mean) \t
%(MEAN$) \n Standard error of repeating unit \t
%(SEM$) \n Degree of Polymerization \t\t %(DP$)";
6514 };
6515 //Calculations for the 2 layer
6516 if (page.active == 2){
6517 range -w aa=%C;
6518 range bb = uid2range(aa.GetPage());$;
```

Appendix

LabTalk script (OriginPro 2020b)

```

6519 win -a %(bb.name$);
6520 %V=%H;
6521 //Creating of a dummy workbook and importing of necessary data
6522 newbook dummy;
6523 %B=%H;
6524 win -a %V;
6525 colcopy irng:=[%V]Sheet1!Col(11)
    orng:=[%B]Sheet1!Col(1);
6526 colcopy irng:=[%V]Sheet1!Col(12)
    orng:=[%B]Sheet1!Col(2);
6527 win -a %B;
6528 //Calculations of Mn, Mw and PDI
6529 col(c)=col(a)*col(b);
6530 col(d)=col(a)^2*col(b);
6531 colstats 2 sum:=[%B]Sheet2!col(a)[1];
6532 copydata irng:=[%B]Sheet2!col(a)[1]
    orng:=[%B]Sheet1!col(5);
6533 page.active$=Sheet2;
6534 layer -d;
6535 colstats 3 sum:=[%B]Sheet2!col(a)[1];
6536 copydata irng:=[%B]Sheet2!col(a)[1]
    orng:=[%B]Sheet1!col(6);
6537 page.active$=Sheet2;
6538 layer -d;
6539 colstats 4 sum:=[%B]Sheet2!col(a)[1];
6540 copydata irng:=[%B]Sheet2!col(a)[1]
    orng:=[%B]Sheet1!col(7);
6541 page.active$=Sheet2;
6542 layer -d;
6543 col(h)=col(f)/col(e);
6544 col(i)=col(g)/col(f);
6545 col(j)=col(i)/col(h);
6546 col(a)=round(col(a),0);
6547 int NoS=count(col(a),1);
6548 for (ii = 1; ii<=NoS; ii++) {
6549     col(k)[ii]=col(a)[ii+1]-col(a)[ii];
6550 };
6551 colstats 11 mean:=col(l)[1] sem:=<new>;
6552 win -a %V;
6553 //Collecting and presenting of calculated data
6554 %P="Mn/Mw/PDI";
6555 if (exist([%V]!P) != 2) {
6556     newsheet name:=%P cols:=5;
6557     wks.col1.Iname$ = "Name";
6558     wks.col2.Iname$ = "1. Spectrum";
6559     wks.col3.Iname$ = "2. Spectrum";
6560     wks.col4.Iname$ = "3. Spectrum";
6561     wks.col5.Iname$ = "4. Spectrum";
6562 }
6563 else {
6564     page.active$=%P;
6565 };
6566 copydata irng:=[%B]sheet1!col(h)[1]
    orng:=[%V]!P!col(C)[1] clear:=1;
6567 copydata irng:=[%B]sheet1!col(i)[1]
    orng:=[%V]!P!col(C)[2] clear:=1;
6568 copydata irng:=[%B]sheet1!col(j)[1]
    orng:=[%V]!P!col(C)[3] clear:=1;
6569 copydata irng:=[%B]sheet1!col(l)[1]
    orng:=[%V]!P!col(C)[4] clear:=1;
6570 copydata irng:=[%B]sheet1!col(n)[1]
    orng:=[%V]!P!col(C)[5] clear:=1;
6571 col(c)[6]=col(c)[2]/col(c)[4];
6572 Mn$=[%B]sheet1!col(h)[1];
6573 Mw$=[%B]sheet1!col(i)[1];
6574 PDI$=[%B]sheet1!col(j)[1];
6575 MEAN$=[%B]sheet1!col(l)[1];
6576 SEM$=[%B]sheet1!col(n)[1];
6577 DP$=[%V]!P!col(c)[6];
6578 col(A)[1]$=Mn;
6579 col(A)[2]$=Mw;
6580 col(A)[3]$=PDI;
6581 col(A)[4]$=Molar weight of repeating unit (mean);
6582 col(A)[5]$=Standard error of repeating unit;

6583 col(A)[6]$=Degree of Polymerization;
6584 worksheet -s 1;
6585 menu -e 35154;
6586 wcolwidth irng:=[%V]!P!Col(1) width:=20;
6587 wcolwidth irng:=[%V]!P!Col(2) width:=10;
6588 wcolwidth irng:=[%V]!P!Col(3) width:=10;
6589 wcolwidth irng:=[%V]!P!Col(4) width:=10;
6590 wcolwidth irng:=[%V]!P!Col(5) width:=10;
6591 page.active$=Sheet1;
6592 win -ia;
6593 win -c %B;
6594 win -a %D;
6595 win -z;
6596 type -b type -b "Here are the values for the 2. Spectrum:
    \n Mn \t\t\t\t %(Mn$) \n Mw \t\t\t\t %(Mw$) \n PDI
    \t\t\t\t %(PDI$) \n Molar weight of repeating unit (mean)
    \t %(MEAN$) \n Standard error of repeating unit \t
    %(SEM$) \n Degree of Polymerization \t\t %(DP$)";
6597 };
6598 //Calculations for the 3 layer
6599 if (page.active == 3){
6600     range -w aa=%C;
6601     range bb = uid2range(aa.GetPage())$;
6602     win -a %(bb.name$);
6603     %V=%H;
6604     //Creating of a dummy workbook and importing of necessary data
6605     newbook dummy;
6606     %B=%H;
6607     win -a %V;
6608     colcopy irng:=[%V]Sheet1!Col(13)
        orng:=[%B]Sheet1!Col(1);
6609     colcopy irng:=[%V]Sheet1!Col(14)
        orng:=[%B]Sheet1!Col(2);
6610     win -a %B;
6611     //Calculations of Mn, Mw and PDI
6612     col(c)=col(a)*col(b);
6613     col(d)=col(a)^2*col(b);
6614     colstats 2 sum:=[%B]Sheet2!col(a)[1];
6615     copydata irng:=[%B]Sheet2!col(a)[1]
        orng:=[%B]Sheet1!col(5);
6616     page.active$=Sheet2;
6617     layer -d;
6618     colstats 3 sum:=[%B]Sheet2!col(a)[1];
6619     copydata irng:=[%B]Sheet2!col(a)[1]
        orng:=[%B]Sheet1!col(6);
6620     page.active$=Sheet2;
6621     layer -d;
6622     colstats 4 sum:=[%B]Sheet2!col(a)[1];
6623     copydata irng:=[%B]Sheet2!col(a)[1]
        orng:=[%B]Sheet1!col(7);
6624     page.active$=Sheet2;
6625     layer -d;
6626     col(h)=col(f)/col(e);
6627     col(i)=col(g)/col(f);
6628     col(j)=col(i)/col(h);
6629     col(a)=round(col(a),0);
6630     int NoS=count(col(a),1);
6631     for (ii = 1; ii<=NoS; ii++) {
6632         col(k)[ii]=col(a)[ii+1]-col(a)[ii];
6633     };
6634     colstats 11 mean:=col(l)[1] sem:=<new>;
6635     win -a %V;
6636     //Collecting and presenting of calculated data
6637     %P="Mn/Mw/PDI";
6638     if (exist([%V]!P) != 2){
6639         newsheet name:=%P cols:=5;
6640         wks.col1.Iname$ = "Name";
6641         wks.col2.Iname$ = "1. Spectrum";
6642         wks.col3.Iname$ = "2. Spectrum";
6643         wks.col4.Iname$ = "3. Spectrum";
6644         wks.col5.Iname$ = "4. Spectrum";
6645     }
6646     else {
6647         page.active$=P%;

```

Appendix

LabTalk script (OriginPro 2020b)

```
6648 };
6649 copydata irng:=[%B]sheet1!col(h)[1]
    orng:=[%V]%P!col(D)[1] clear:=1;
6650 copydata irng:=[%B]sheet1!col(i)[1]
    orng:=[%V]%P!col(D)[2] clear:=1;
6651 copydata irng:=[%B]sheet1!col(j)[1]
    orng:=[%V]%P!col(D)[3] clear:=1;
6652 copydata irng:=[%B]sheet1!col(l)[1]
    orng:=[%V]%P!col(D)[4] clear:=1;
6653 copydata irng:=[%B]sheet1!col(n)[1]
    orng:=[%V]%P!col(D)[5] clear:=1;
6654 col(d)[6]=col(d)[2]/col(d)[4];
6655 Mn$=[%B]sheet1!col(h)[1]$;
6656 Mw$=[%B]sheet1!col(i)[1]$;
6657 PDI$=[%B]sheet1!col(j)[1]$;
6658 MEAN$=[%B]sheet1!col(l)[1]$;
6659 SEM$=[%B]sheet1!col(n)[1]$;
6660 DP$=[%V]%P!col(d)[6]$;
6661 col(A)[1]$=Mn;
6662 col(A)[2]$=Mw;
6663 col(A)[3]$=PDI;
6664 col(A)[4]$=Molar weight of repeating unit (mean);
6665 col(A)[5]$=Standard error of repeating unit;
6666 col(A)[6]$=Degree of Polymerization;
6667 worksheet -s 1;
6668 menu -e 35154;
6669 wcolwidth irng:=[%V]%P!Col(1) width:=20;
6670 wcolwidth irng:=[%V]%P!Col(2) width:=10;
6671 wcolwidth irng:=[%V]%P!Col(3) width:=10;
6672 wcolwidth irng:=[%V]%P!Col(4) width:=10;
6673 wcolwidth irng:=[%V]%P!Col(5) width:=10;
6674 page.active$=Sheet1;
6675 win -ia;
6676 win -c %B;
6677 win -a %D;
6678 win -z;
6679 type -b "Here are the values for the 3. Spectrum: \n Mn
\t\t\t\t%(Mn$) \n Mw \t\t\t\t%(Mw$) \n PDI \t\t\t\t
%(PDI$) \n Molar weight of repeating unit (mean) \t
%(MEAN$) \n Standard error of repeating unit \t
%(SEM$) \n Degree of Polymerization \t\t %(DP$)";
6680 };
6681 //Calculations for the 4 layer
6682 if(page.active == 4){
6683 range -w aa=C%;
6684 range bb = uid2range(aa.GetPage());$;
6685 win -a %(bb.name$);
6686 %V=H;%
6687 //Creating of a dummy workbook and importing of nec-
    essary data
6688 newbook dummy;
6689 %B=H;%
6690 win -a %V;
6691 colcopy
    irng:=[%V]Sheet1!Col(15)orng:=[%B]Sheet1!Col(1);
6692 colcopy
    irng:=[%V]Sheet1!Col(16)orng:=[%B]Sheet1!Col(2);
6693 win -a %B;
6694 //Calculations of Mn, Mw and PDI
6695 col(c)=col(a)*col(b);
6696 col(d)=col(a)^2*col(b);
6697 colstats 2sum:=[%B]Sheet2!col(a)[1];
6698 copydata irng:=[%B]Sheet2!col(a)[1]
    orng:=[%B]Sheet1!col(5);
6699 page.active$=Sheet2;
6700 layer -d;
6701 colstats 3 sum:=[%B]Sheet2!col(a)[1];
6702 copydata irng:=[%B]Sheet2!col(a)[1]
    orng:=[%B]Sheet1!col(6);
6703 page.active$=Sheet2;
6704 layer -d;
6705 colstats 4 sum:=[%B]Sheet2!col(a)[1];
6706 copydata irng:=[%B]Sheet2!col(a)[1]
    orng:=[%B]Sheet1!col(7);
6707 page.active$=Sheet2;
6708 layer -d;
6709 col(h)=col(f)/col(e);
6710 col(i)=col(g)/col(f);
6711 col(j)=col(i)/col(h);
6712 col(a)=round(col(a),0);
6713 int NoS=count(col(a),1);
6714 for (ii = 1; ii<=NoS; ii++) {
6715 col(k)[ii]=col(a)[ii+1]-col(a)[ii];
6716 };
6717 colstats 11 mean:=col(l)[1] sem:=<new>;
6718 win -a %V;
6719 //Collecting and presenting of calculated data
6720 %P="Mn/Mw/PDI";
6721 if (exist([%V]%P!) != 2) {
6722 newsheet name:=%P cols:=5;
6723 wks.col1.lname$ = "Name";
6724 wks.col2.lname$ = "1. Spectrum";
6725 wks.col3.lname$ = "2. Spectrum";
6726 wks.col4.lname$ = "3. Spectrum";
6727 wks.col5.lname$ = "4. Spectrum";
6728 }
6729 else {
6730 page.active$=%P;
6731 };
6732 copydata irng:=[%B]sheet1!col(h)[1]
    orng:=[%V]%P!col(E)[1] clear:=1;
6733 copydata irng:=[%B]sheet1!col(i)[1]
    orng:=[%V]%P!col(E)[2] clear:=1;
6734 copydata irng:=[%B]sheet1!col(j)[1]
    orng:=[%V]%P!col(E)[3] clear:=1;
6735 copydata irng:=[%B]sheet1!col(l)[1]
    orng:=[%V]%P!col(E)[4] clear:=1;
6736 copydata irng:=[%B]sheet1!col(n)[1]
    orng:=[%V]%P!col(E)[5] clear:=1;
6737 col(e)[6]=col(e)[2]/col(e)[4];
6738 Mn$=[%B]sheet1!col(h)[1]$;
6739 Mw$=[%B]sheet1!col(i)[1]$;
6740 PDI$=[%B]sheet1!col(j)[1]$;
6741 MEAN$=[%B]sheet1!col(l)[1]$;
6742 SEM$=[%B]sheet1!col(n)[1]$;
6743 DP$=[%V]%P!col(e)[6]$;
6744 col(A)[1]$=Mn;
6745 col(A)[2]$=Mw;
6746 col(A)[3]$=PDI;
6747 col(A)[4]$=Molar weight of repeating unit (mean);
6748 col(A)[5]$=Standard error of repeating unit;
6749 col(A)[6]$=Degree of Polymerization;
6750 worksheet -s 1;
6751 menu -e 35154;
6752 wcolwidth irng:=[%V]%P!Col(1) width:=20;
6753 wcolwidth irng:=[%V]%P!Col(2) width:=10;
6754 wcolwidth irng:=[%V]%P!Col(3) width:=10;
6755 wcolwidth irng:=[%V]%P!Col(4) width:=10;
6756 wcolwidth irng:=[%V]%P!Col(5) width:=10;
6757 page.active$=Sheet1;
6758 win -ia;
6759 win -c %B;
6760 win -a %D;
6761 win -z;
6762 type -b "Here are the values for the 4. Spectrum: \n Mn
\t\t\t\t%(Mn$) \n Mw \t\t\t\t%(Mw$) \n PDI \t\t\t\t
%(PDI$) \n Molar weight of repeating unit (mean) \t
%(MEAN$) \n Standard error of repeating unit \t
%(SEM$) \n Degree of Polymerization \t\t %(DP$)";
6763 };
6764 };
6765 //Calculation script for 5 layer plot
6766 if (page.nlayers == 5) {
6767 //Calculations for the 1 layer
6768 if (page.active == 1){
6769 range -w aa=%C;
6770 range bb = uid2range(aa.GetPage());$;
6771 win -a %(bb.name$);
6772 %V=H;
```


Appendix

LabTalk script (OriginPro 2020b)

```

6773 //Creating of a dummy workbook and importing of nec-
        essary data
6774 newbook dummy;
6775 %B=%H;
6776 win -a %V;
6777 colcopy irng:=[%V]Sheet1!Col(11)
        orng:=[%B]Sheet1!Col(1);
6778 colcopy irng:=[%V]Sheet1!Col(12)
        orng:=[%B]Sheet1!Col(2);
6779 win -a %B;
6780 //Calculations of Mn, Mw and PDI
6781 col(c)=col(a)*col(b);
6782 col(d)=col(a)^2*col(b);
6783 colstats 2 sum:=[%B]Sheet2!col(a)[1];
6784 copydata irng:=[%B]Sheet2!col(a)[1]
        orng:=[%B]Sheet1!col(5);
6785 page.active$=Sheet2;
6786 layer -d;
6787 colstats 3 sum:=[%B]Sheet2!col(a)[1];
6788 copydata irng:=[%B]Sheet2!col(a)[1]
        orng:=[%B]Sheet1!col(6);
6789 page.active$=Sheet2;
6790 layer -d;
6791 colstats 4 sum:=[%B]Sheet2!col(a)[1];
6792 copydata irng:=[%B]Sheet2!col(a)[1]
        orng:=[%B]Sheet1!col(7);
6793 page.active$=Sheet2;
6794 layer -d;
6795 col(h)=col(f)/col(e);
6796 col(i)=col(g)/col(f);
6797 col(j)=col(i)/col(h);
6798 col(a)=round(col(a),0);
6799 int NoS=count(col(a),1);
6800 for (ii = 1; ii<=NoS; ii++) {
6801     col(k)[ii]=col(a)[ii+1]-col(a)[ii];
6802 };
6803 colstats 11 mean:=col(l)[1] sem:=<new>;
6804 win -a %V;
6805 //Collecting and presenting of calculated data
6806 %P="Mn/Mw/PDI";
6807 if (exist([%V]%P!) != 2) {
6808     newsheet name:=%P cols:=6;
6809     wks.col1.Iname$ = "Name";
6810     wks.col2.Iname$ = "1. Spectrum";
6811     wks.col3.Iname$ = "2. Spectrum";
6812     wks.col4.Iname$ = "3. Spectrum";
6813     wks.col5.Iname$ = "4. Spectrum";
6814     wks.col6.Iname$ = "5. Spectrum";
6815 }
6816 else {
6817     page.active$=%P;
6818 };
6819 copydata irng:=[%B]sheet1!col(h)[1]
        orng:=[%V]%P!col(B)[1] clear:=1;
6820 copydata irng:=[%B]sheet1!col(i)[1]
        orng:=[%V]%P!col(B)[2] clear:=1;
6821 copydata irng:=[%B]sheet1!col(j)[1]
        orng:=[%V]%P!col(B)[3] clear:=1;
6822 copydata irng:=[%B]sheet1!col(l)[1]
        orng:=[%V]%P!col(B)[4] clear:=1;
6823 copydata irng:=[%B]sheet1!col(n)[1]
        orng:=[%V]%P!col(B)[5] clear:=1;
6824 col(B)[6]=col(B)[2]/col(B)[4];
6825 Mn$=[%B]sheet1!col(h)[1];
6826 Mw$=[%B]sheet1!col(i)[1];
6827 PDI$=[%B]sheet1!col(j)[1];
6828 MEAN$=[%B]sheet1!col(l)[1];
6829 SEM$=[%B]sheet1!col(n)[1];
6830 DP$=[%V]%P!col(B)[6];
6831 col(A)[1]$=Mn;
6832 col(A)[2]$=Mw;
6833 col(A)[3]$=PDI;
6834 col(A)[4]$=Molar weight of repeating unit (mean);
6835 col(A)[5]$=Standard error of repeating unit;
6836 col(A)[6]$=Degree of Polymerization;

6837 worksheet -s 1;
6838 menu -e 35154;
6839 wcolwidth irng:=[%V]%P!Col(1) width:=20;
6840 wcolwidth irng:=[%V]%P!Col(2) width:=10;
6841 wcolwidth irng:=[%V]%P!Col(3) width:=10;
6842 wcolwidth irng:=[%V]%P!Col(4) width:=10;
6843 wcolwidth irng:=[%V]%P!Col(5) width:=10;
6844 wcolwidth irng:=[%V]%P!Col(6) width:=10;
6845 page.active$=Sheet1;
6846 win -ia;
6847 win -c %B;
6848 win -a %D;
6849 win -z;
6850 type -b "Here are the values for the 1. Spectrum: \n Mn
        \n %Mn$ \n Mw \n %Mw$ \n PDI \n %PDI$ \n Molar weight of repeating unit (mean) \n
        %MEAN$ \n Standard error of repeating unit \n
        %SEM$ \n Degree of Polymerization \n %DP$";
6851 };
6852 //Calculations for the 2 layer
6853 if (page.active == 2){
6854     range -w aa=%C;
6855     range bb = uid2range(aa.GetPage());
6856     win -a %(bb.name$);
6857     %V=%H;
6858     //Creating of a dummy workbook and importing of nec-
        essary data
6859     newbook dummy;
6860     %B=%H;
6861     win -a %V;
6862     colcopy irng:=[%V]Sheet1!Col(13)
        orng:=[%B]Sheet1!Col(1);
6863     colcopy irng:=[%V]Sheet1!Col(14)
        orng:=[%B]Sheet1!Col(2);
6864     win -a %B;
6865     //Calculations of Mn, Mw and PDI
6866     col(c)=col(a)*col(b);
6867     col(d)=col(a)^2*col(b);
6868     colstats 2 sum:=[%B]Sheet2!col(a)[1];
6869     copydata irng:=[%B]Sheet2!col(a)[1]
        orng:=[%B]Sheet1!col(5);
6870     page.active$=Sheet2;
6871     layer -d;
6872     colstats 3 sum:=[%B]Sheet2!col(a)[1];
6873     copydata irng:=[%B]Sheet2!col(a)[1]
        orng:=[%B]Sheet1!col(6);
6874     page.active$=Sheet2;
6875     layer -d;
6876     colstats 4 sum:=[%B]Sheet2!col(a)[1];
6877     copydata irng:=[%B]Sheet2!col(a)[1]
        orng:=[%B]Sheet1!col(7);
6878     page.active$=Sheet2;
6879     layer -d;
6880     col(h)=col(f)/col(e);
6881     col(i)=col(g)/col(f);
6882     col(j)=col(i)/col(h);
6883     col(a)=round(col(a),0);
6884     int NoS=count(col(a),1);
6885     for (ii = 1; ii<=NoS; ii++) {
6886         col(k)[ii]=col(a)[ii+1]-col(a)[ii];
6887     };
6888     colstats 11 mean:=col(l)[1] sem:=<new>;
6889     win -a %V;
6890     //Collecting and presenting of calculated data
6891     %P="Mn/Mw/PDI";
6892     if (exist([%V]%P!) != 2) {
6893         newsheet name:=%P cols:=6;
6894         wks.col1.Iname$ = "Name";
6895         wks.col2.Iname$ = "1. Spectrum";
6896         wks.col3.Iname$ = "2. Spectrum";
6897         wks.col4.Iname$ = "3. Spectrum";
6898         wks.col5.Iname$ = "4. Spectrum";
6899         wks.col6.Iname$ = "5. Spectrum";
6900     }
6901     else {

```


Appendix

LabTalk script (OriginPro 2020b)

```
6902 page.active$=%P;
6903 };
6904 copydata irng:=[%B]sheet1!col(h)[1]
    orng:=[%V]%P!col(C)[1] clear:=1;
6905 copydata irng:=[%B]sheet1!col(i)[1]
    orng:=[%V]%P!col(C)[2] clear:=1;
6906 copydata irng:=[%B]sheet1!col(j)[1]
    orng:=[%V]%P!col(C)[3] clear:=1;
6907 copydata irng:=[%B]sheet1!col(l)[1]
    orng:=[%V]%P!col(C)[4] clear:=1;
6908 copydata irng:=[%B]sheet1!col(n)[1]
    orng:=[%V]%P!col(C)[5] clear:=1;
6909 col(c)[6]=col(c)[2]/col(c)[4];
6910 Mn$=[%B]sheet1!col(h)[1]$;
6911 Mw$=[%B]sheet1!col(i)[1]$;
6912 PDI$=[%B]sheet1!col(j)[1]$;
6913 MEAN$=[%B]sheet1!col(l)[1]$;
6914 SEM$=[%B]sheet1!col(n)[1]$;
6915 DP$=[%V]%P!col(c)[6]$;
6916 col(A)[1]$=Mn;
6917 col(A)[2]$=Mw;
6918 col(A)[3]$=PDI;
6919 col(A)[4]$=Molar weight of repeating unit (mean);
6920 col(A)[5]$=Standard error of repeating unit;
6921 col(A)[6]$=Degree of Polymerization;
6922 worksheet -s 1;
6923 menu -e 35154;
6924 wcolwidth irng:=[%V]%P!Col(1) width:=20;
6925 wcolwidth irng:=[%V]%P!Col(2) width:=10;
6926 wcolwidth irng:=[%V]%P!Col(3) width:=10;
6927 wcolwidth irng:=[%V]%P!Col(4) width:=10;
6928 wcolwidth irng:=[%V]%P!Col(5) width:=10;
6929 wcolwidth irng:=[%V]%P!Col(6) width:=10;
6930 page.active$=Sheet1;
6931 win -ia;
6932 win -c %B;
6933 win -a %D;
6934 win -z;
6935 type -b "Here are the values for the 2. Spectrum: \n Mn
\t\t\t\t\t%(Mn$) \n Mw \t\t\t\t\t%(Mw$) \n PDI \t\t\t\t\t
%(PDI$) \n Molar weight of repeating unit (mean) \t
%(MEAN$) \n Standard error of repeating unit \t
%(SEM$) \n Degree of Polymerization \t\t\t\t\t%(DP$)";
6936 };
6937 //Calculations for the 3 layer
6938 if (page.active == 3){
6939 range -w aa=%C;
6940 range bb = uid2range(aa.GetPage());$;
6941 win -a %(bb.name$);
6942 %V=%H;
6943 //Creating a dummy workbook and importing of nec-
    essary data
6944 newbook dummy;
6945 %B=%H;
6946 win -a %V;
6947 colcopy irng:=[%V]Sheet1!Col(15)
    orng:=[%B]Sheet1!Col(1);
6948 colcopy irng:=[%V]Sheet1!Col(16)
    orng:=[%B]Sheet1!Col(2);
6949 win -a %B;
6950 //Calculations of Mn, Mw and PDI
6951 col(c)=col(a)*col(b);
6952 col(d)=col(a)^2*col(b);
6953 colstats 2 sum:=[%B]Sheet2!col(a)[1];
6954 copydata irng:=[%B]Sheet2!col(a)[1]
    orng:=[%B]Sheet1!col(5);
6955 page.active$=Sheet2;
6956 layer -d;
6957 colstats 3 sum:=[%B]Sheet2!col(a)[1];
6958 copydata irng:=[%B]Sheet2!col(a)[1]
    orng:=[%B]Sheet1!col(6);
6959 page.active$=Sheet2;
6960 layer -d;
6961 colstats 4 sum:=[%B]Sheet2!col(a)[1];
6962 copydata irng:=[%B]Sheet2!col(a)[1]
    orng:=[%B]Sheet1!col(7);
6963 page.active$=Sheet2;
6964 layer -d;
6965 col(h)=col(f)/col(e);
6966 col(i)=col(g)/col(f);
6967 col(j)=col(i)/col(h);
6968 col(a)=round(col(a),0);
6969 int NoS=count(col(a),1);
6970 for (ii = 1; ii<=NoS; ii++) {
6971 col(k)[ii]=col(a)[ii+1]-col(a)[ii];
6972 };
6973 colstats 11 mean:=col(l)[1] sem:=<new>;
6974 win -a %V;
6975 //Collecting and presenting of calculated data
6976 %P="Mn/Mw/PDI";
6977 if (exist([%V]%P!) != 2) {
6978 newsheet name:=%P cols:=6;
6979 wks.col1.lname$ = "Name";
6980 wks.col2.lname$ = "1. Spectrum";
6981 wks.col3.lname$ = "2. Spectrum";
6982 wks.col4.lname$ = "3. Spectrum";
6983 wks.col5.lname$ = "4. Spectrum";
6984 wks.col6.lname$ = "5. Spectrum";
6985 }
6986 else {
6987 page.active$=%P;
6988 };
6989 copydata irng:=[%B]sheet1!col(h)[1]
    orng:=[%V]%P!col(D)[1] clear:=1;
6990 copydata irng:=[%B]sheet1!col(i)[1]
    orng:=[%V]%P!col(D)[2] clear:=1;
6991 copydata irng:=[%B]sheet1!col(j)[1]
    orng:=[%V]%P!col(D)[3] clear:=1;
6992 copydata irng:=[%B]sheet1!col(l)[1]
    orng:=[%V]%P!col(D)[4] clear:=1;
6993 copydata irng:=[%B]sheet1!col(n)[1]
    orng:=[%V]%P!col(D)[5] clear:=1;
6994 col(d)[6]=col(d)[2]/col(d)[4];
6995 Mn$=[%B]sheet1!col(h)[1]$;
6996 Mw$=[%B]sheet1!col(i)[1]$;
6997 PDI$=[%B]sheet1!col(j)[1]$;
6998 MEAN$=[%B]sheet1!col(l)[1]$;
6999 SEM$=[%B]sheet1!col(n)[1]$;
7000 DP$=[%V]%P!col(d)[6]$;
7001 col(A)[1]$=Mn;
7002 col(A)[2]$=Mw;
7003 col(A)[3]$=PDI;
7004 col(A)[4]$=Molar weight of repeating unit (mean);
7005 col(A)[5]$=Standard error of repeating unit;
7006 col(A)[6]$=Degree of Polymerization;
7007 worksheet -s 1;
7008 menu -e 35154;
7009 wcolwidth irng:=[%V]%P!Col(1) width:=20;
7010 wcolwidth irng:=[%V]%P!Col(2) width:=10;
7011 wcolwidth irng:=[%V]%P!Col(3) width:=10;
7012 wcolwidth irng:=[%V]%P!Col(4) width:=10;
7013 wcolwidth irng:=[%V]%P!Col(5) width:=10;
7014 wcolwidth irng:=[%V]%P!Col(6) width:=10;
7015 page.active$=Sheet1;
7016 win -ia;
7017 win -c %B;
7018 win -a %D;
7019 win -z;
7020 type -b "Here are the values for the 3. Spectrum: \n Mn
\t\t\t\t\t%(Mn$) \n Mw \t\t\t\t\t%(Mw$) \n PDI \t\t\t\t\t
%(PDI$) \n Molar weight of repeating unit (mean) \t
%(MEAN$) \n Standard error of repeating unit \t
%(SEM$) \n Degree of Polymerization \t\t\t\t\t%(DP$)";
7021 };
7022 //Calculations for the 4 layer
7023 if (page.active == 4){
7024 range -w aa=%C;
7025 range bb = uid2range(aa.GetPage());$;
7026 win -a %(bb.name$);
```

Appendix

LabTalk script (OriginPro 2020b)

```

2027 %V=%H;
2028 //Creating of a dummy workbook and importing of nec-
2029 essary data
2029 newbook dummy;
2030 %B=%H;
2031 win -a %V;
2032 colcopy irng:=[%V]Sheet1!Col(17)
2033 orng:=[%B]Sheet1!Col(1);
2033 colcopy irng:=[%V]Sheet1!Col(18)
2034 orng:=[%B]Sheet1!Col(2);
2034 win -a %B;
2035 //Calculations of Mn, Mw and PDI
2036 col(c)=col(a)*col(b);
2037 col(d)=col(a)^2*col(b);
2038 colstats 2 sum:=[%B]Sheet2!col(a)[1];
2039 copydata irng:=[%B]Sheet2!col(a)[1]
2040 orng:=[%B]Sheet1!col(5);
2040 page.active$=Sheet2;
2041 layer -d;
2042 colstats 3 sum:=[%B]Sheet2!col(a)[1];
2043 copydata irng:=[%B]Sheet2!col(a)[1]
2044 orng:=[%B]Sheet1!col(6);
2044 page.active$=Sheet2;
2045 layer -d;
2046 colstats 4 sum:=[%B]Sheet2!col(a)[1];
2047 copydata irng:=[%B]Sheet2!col(a)[1]
2048 orng:=[%B]Sheet1!col(7);
2048 page.active$=Sheet2;
2049 layer -d;
2050 col(h)=col(f)/col(e);
2051 col(i)=col(g)/col(f);
2052 col(j)=col(i)/col(h);
2053 col(a)=round(col(a),0);
2054 int NoS=count(col(a),1);
2055 for (ii = 1; ii<=NoS; ii++) {
2056 col(k)[ii]=col(a)[ii+1]-col(a)[ii];
2057 };
2058 colstats 11 mean:=col(l)[1] sem:=<new>;
2059 win -a %V;
2060 //Collecting and presenting of calculated data
2061 %P="Mn/Mw/PDI";
2062 if (exist([%V]%P!) != 2) {
2063 newsheet name:=%P cols:=6;
2064 wks.col1.Iname$ = "Name";
2065 wks.col2.Iname$ = "1. Spectrum";
2066 wks.col3.Iname$ = "2. Spectrum";
2067 wks.col4.Iname$ = "3. Spectrum";
2068 wks.col5.Iname$ = "4. Spectrum";
2069 wks.col6.Iname$ = "5. Spectrum";
2070 }
2071 else {
2072 page.active$=%P;
2073 };
2074 copydata irng:=[%B]sheet1!col(h)[1]
2075 orng:=[%V]%P!col(E)[1] clear:=1;
2075 copydata irng:=[%B]sheet1!col(i)[1]
2076 orng:=[%V]%P!col(E)[2] clear:=1;
2076 copydata irng:=[%B]sheet1!col(j)[1]
2077 orng:=[%V]%P!col(E)[3] clear:=1;
2077 copydata irng:=[%B]sheet1!col(l)[1]
2078 orng:=[%V]%P!col(E)[4] clear:=1;
2078 copydata irng:=[%B]sheet1!col(n)[1]
2079 orng:=[%V]%P!col(E)[5] clear:=1;
2079 col(e)[6]=col(e)[2]/col(e)[4];
2080 Mn$=[%B]sheet1!col(h)[1]$;
2081 Mw$=[%B]sheet1!col(i)[1]$;
2082 PDI$=[%B]sheet1!col(j)[1]$;
2083 MEAN$=[%B]sheet1!col(l)[1]$;
2084 SEM$=[%B]sheet1!col(n)[1]$;
2085 DP$=[%V]%P!col(e)[6]$;
2086 col(A)[1]$=Mn;
2087 col(A)[2]$=Mw;
2088 col(A)[3]$=PDI;
2089 col(A)[4]$=Molar weight of repeating unit (mean);
2090 col(A)[5]$=Standard error of repeating unit;
2091 col(A)[6]$=Degree of Polymerization;
2092 worksheet -s 1;
2093 menu -e 35154;
2094 wcolwidth irng:=[%V]%P!Col(1) width:=20;
2095 wcolwidth irng:=[%V]%P!Col(2) width:=10;
2096 wcolwidth irng:=[%V]%P!Col(3) width:=10;
2097 wcolwidth irng:=[%V]%P!Col(4) width:=10;
2098 wcolwidth irng:=[%V]%P!Col(5) width:=10;
2099 wcolwidth irng:=[%V]%P!Col(6) width:=10;
2100 page.active$=Sheet1;
2101 win -ia;
2102 win -c %B;
2103 win -a %D;
2104 win -z;
2105 type -b "Here are the values for the 4. Spectrum: \n Mn
2106 \t\t\t % (Mn$) \n Mw \t\t\t % (Mw$) \n PDI \t\t\t
2107 % (PDI$) \n Molar weight of repeating unit (mean) \t
2108 % (MEAN$) \n Standard error of repeating unit \t
2109 % (SEM$) \n Degree of Polymerization \t\t % (DP$)";
2106 };
2107 //Calculations for the 5 layer
2108 if (page.active == 5){
2109 range -w aa=%C;
2110 range bb = uid2range(aa.GetPage());$;
2111 win -a % (bb.name$);
2112 %V=%H;
2113 //Creating of a dummy workbook and importing of nec-
2114 essary data
2114 newbook dummy;
2115 %B=%H;
2116 win -a %V;
2117 colcopy irng:=[%V]Sheet1!Col(19)
2118 orng:=[%B]Sheet1!Col(1);
2118 colcopy irng:=[%V]Sheet1!Col(20)
2119 orng:=[%B]Sheet1!Col(2);
2119 win -a %B;
2120 //Calculations of Mn, Mw and PDI
2121 col(c)=col(a)*col(b);
2122 col(d)=col(a)^2*col(b);
2123 colstats 2 sum:=[%B]Sheet2!col(a)[1];
2124 copydata irng:=[%B]Sheet2!col(a)[1]
2125 orng:=[%B]Sheet1!col(5);
2125 page.active$=Sheet2;
2126 layer -d;
2127 colstats 3 sum:=[%B]Sheet2!col(a)[1];
2128 copydata irng:=[%B]Sheet2!col(a)[1]
2129 orng:=[%B]Sheet1!col(6);
2129 page.active$=Sheet2;
2130 layer -d;
2131 colstats 4 sum:=[%B]Sheet2!col(a)[1];
2132 copydata irng:=[%B]Sheet2!col(a)[1]
2133 orng:=[%B]Sheet1!col(7);
2133 page.active$=Sheet2;
2134 layer -d;
2135 col(h)=col(f)/col(e);
2136 col(i)=col(g)/col(f);
2137 col(j)=col(i)/col(h);
2138 col(a)=round(col(a),0);
2139 int NoS=count(col(a),1);
2140 for (ii = 1; ii<=NoS; ii++) {
2141 col(k)[ii]=col(a)[ii+1]-col(a)[ii];
2142 };
2143 colstats 11 mean:=col(l)[1] sem:=<new>;
2144 win -a %V;
2145 //Collecting and presenting of calculated data
2146 %P="Mn/Mw/PDI";
2147 if (exist([%V]%P!) != 2) {
2148 newsheet name:=%P cols:=6;
2149 wks.col1.Iname$ = "Name";
2150 wks.col2.Iname$ = "1. Spectrum";
2151 wks.col3.Iname$ = "2. Spectrum";
2152 wks.col4.Iname$ = "3. Spectrum";
2153 wks.col5.Iname$ = "4. Spectrum";
2154 wks.col6.Iname$ = "5. Spectrum";
2155 }

```

Appendix

LabTalk script (OriginPro 2020b)

```
7156 else {
7157 page.active$=%P;
7158 };
7159 copydata irng:=[%B]sheet1!col(h)[1]
    orng:=[%V]%P!col(F)[1] clear:=1;
7160 copydata irng:=[%B]sheet1!col(i)[1]
    orng:=[%V]%P!col(F)[2] clear:=1;
7161 copydata irng:=[%B]sheet1!col(j)[1]
    orng:=[%V]%P!col(F)[3] clear:=1;
7162 copydata irng:=[%B]sheet1!col(l)[1]
    orng:=[%V]%P!col(F)[4] clear:=1;
7163 copydata irng:=[%B]sheet1!col(n)[1]
    orng:=[%V]%P!col(F)[5] clear:=1;
7164 col(f)[6]=col(f)[2]/col(f)[4];
7165 Mn$=[%B]sheet1!col(h)[1]$;
7166 Mw$=[%B]sheet1!col(i)[1]$;
7167 PDI$=[%B]sheet1!col(j)[1]$;
7168 MEAN$=[%B]sheet1!col(l)[1]$;
7169 SEM$=[%B]sheet1!col(n)[1]$;
7170 DP$=[%V]%P!col(f)[6]$;
7171 col(A)[1]$=Mn;
7172 col(A)[2]$=Mw;
7173 col(A)[3]$=PDI;
7174 col(A)[4]$=Molar weight of repeating unit (mean);
7175 col(A)[5]$=Standard error of repeating unit;
7176 col(A)[6]$=Degree of Polymerization;
7177 worksheet -s 1;
7178 menu -e 35154;
7179 wcolwidth irng:=[%V]%P!Col(1) width:=20;
7180 wcolwidth irng:=[%V]%P!Col(2) width:=10;
7181 wcolwidth irng:=[%V]%P!Col(3) width:=10;
7182 wcolwidth irng:=[%V]%P!Col(4) width:=10;
7183 wcolwidth irng:=[%V]%P!Col(5) width:=10;
7184 wcolwidth irng:=[%V]%P!Col(6) width:=10;
7185 page.active$=Sheet1;
7186 win -ia;
7187 win -c %B;
7188 win -a %D;
7189 win -z;
7190 type -b "Here are the values for the 5. Spectrum: \n Mn
\t\t\t\t%(Mn$) \n Mw \t\t\t\t%(Mw$) \n PDI \t\t\t\t
%(PDI$) \n Molar weight of repeating unit (mean) \t
%(MEAN$) \n Standard error of repeating unit \t
%(SEM$) \n Degree of Polymerization \t\t %(DP$)";
7191 };
7192 };
7193 //Calculation script for 6 layer plot
7194 if (page.nlayers == 6) {
7195 //Calculations for the 1 layer
7196 if (page.active == 1){
7197 range -w aa=%C;
7198 range bb = uid2range(aa.GetPage());$;
7199 win -a %(bb.name$);
7200 %V=%H;
7201 //Creating of a dummy workbook and importing of nec-
    essary data
7202 newbook dummy;
7203 %B=%H;
7204 win -a %V;
7205 colcopy irng:=[%V]Sheet1!Col(13)
    orng:=[%B]Sheet1!Col(1);
7206 colcopy irng:=[%V]Sheet1!Col(14)
    orng:=[%B]Sheet1!Col(2);
7207 win -a %B;
7208 //Calculations of Mn, Mw and PDI
7209 col(c)=col(a)*col(b);
7210 col(d)=col(a)^2*col(b);
7211 colstats 2 sum:=[%B]Sheet2!col(a)[1];
7212 copydata irng:=[%B]Sheet2!col(a)[1]
    orng:=[%B]Sheet1!col(5);
7213 page.active$=Sheet2;
7214 layer -d;
7215 colstats 3 sum:=[%B]Sheet2!col(a)[1];
7216 copydata irng:=[%B]Sheet2!col(a)[1]
    orng:=[%B]Sheet1!col(6);
7217 page.active$=Sheet2;
7218 layer -d;
7219 colstats 4 sum:=[%B]Sheet2!col(a)[1];
7220 copydata irng:=[%B]Sheet2!col(a)[1]
    orng:=[%B]Sheet1!col(7);
7221 page.active$=Sheet2;
7222 layer -d;
7223 col(h)=col(f)/col(e);
7224 col(i)=col(g)/col(f);
7225 col(j)=col(i)/col(h);
7226 col(a)=round(col(a),0);
7227 int NoS=count(col(a),1);
7228 for (ii = 1; ii<=NoS; ii++) {
7229 col(k)[ii]=col(a)[ii+1]-col(a)[ii];
7230 };
7231 colstats 11 mean:=col(l)[1] sem:=<new>;
7232 win -a %V;
7233 //Collecting and presenting of calculated data
7234 %P="Mn/Mw/PDI";
7235 if (exist([%V]%P!) != 2) {
7236 newsheet name:=%P cols:=7;
7237 wks.col1.lname$ = "Name";
7238 wks.col2.lname$ = "1. Spectrum";
7239 wks.col3.lname$ = "2. Spectrum";
7240 wks.col4.lname$ = "3. Spectrum";
7241 wks.col5.lname$ = "4. Spectrum";
7242 wks.col6.lname$ = "5. Spectrum";
7243 wks.col7.lname$ = "6. Spectrum";
7244 }
7245 else {
7246 page.active$=%P;
7247 };
7248 copydata irng:=[%B]sheet1!col(h)[1]
    orng:=[%V]%P!col(B)[1] clear:=1;
7249 copydata irng:=[%B]sheet1!col(i)[1]
    orng:=[%V]%P!col(B)[2] clear:=1;
7250 copydata irng:=[%B]sheet1!col(j)[1]
    orng:=[%V]%P!col(B)[3] clear:=1;
7251 copydata irng:=[%B]sheet1!col(l)[1]
    orng:=[%V]%P!col(B)[4] clear:=1;
7252 copydata irng:=[%B]sheet1!col(n)[1]
    orng:=[%V]%P!col(B)[5] clear:=1;
7253 col(B)[6]=col(B)[2]/col(B)[4];
7254 Mn$=[%B]sheet1!col(h)[1]$;
7255 Mw$=[%B]sheet1!col(i)[1]$;
7256 PDI$=[%B]sheet1!col(j)[1]$;
7257 MEAN$=[%B]sheet1!col(l)[1]$;
7258 SEM$=[%B]sheet1!col(n)[1]$;
7259 DP$=[%V]%P!col(B)[6]$;
7260 col(A)[1]$=Mn;
7261 col(A)[2]$=Mw;
7262 col(A)[3]$=PDI;
7263 col(A)[4]$=Molar weight of repeating unit (mean);
7264 col(A)[5]$=Standard error of repeating unit;
7265 col(A)[6]$=Degree of Polymerization;
7266 worksheet -s 1;
7267 menu -e 35154;
7268 wcolwidth irng:=[%V]%P!Col(1) width:=20;
7269 wcolwidth irng:=[%V]%P!Col(2) width:=10;
7270 wcolwidth irng:=[%V]%P!Col(3) width:=10;
7271 wcolwidth irng:=[%V]%P!Col(4) width:=10;
7272 wcolwidth irng:=[%V]%P!Col(5) width:=10;
7273 wcolwidth irng:=[%V]%P!Col(6) width:=10;
7274 wcolwidth irng:=[%V]%P!Col(7) width:=10;
7275 page.active$=Sheet1;
7276 win -ia;
7277 win -c %B;
7278 win -a %D;
7279 win -z;
7280 type -b "Here are the values for the 1. Spectrum: \n Mn
\t\t\t\t%(Mn$) \n Mw \t\t\t\t%(Mw$) \n PDI \t\t\t\t
%(PDI$) \n Molar weight of repeating unit (mean) \t
%(MEAN$) \n Standard error of repeating unit \t
%(SEM$) \n Degree of Polymerization \t\t %(DP$)";
7281 };
```

Appendix

LabTalk script (OriginPro 2020b)

```
7282 //Calculations for the 2 layer
7283 if (page.active == 2){
7284   range -w aa=%C;
7285   range bb = uid2range(aa.GetPage());$;
7286   win -a %(bb.name$);
7287   %V=%H;
7288   //Creating a dummy workbook and importing of necessary data
7289   newbook dummy;
7290   %B=%H;
7291   win -a %V;
7292   colcopy irng:=[%V]Sheet1!Col(15)
   org:=[%B]Sheet1!Col(1);
7293   colcopy irng:=[%V]Sheet1!Col(16)
   org:=[%B]Sheet1!Col(2);
7294   win -a %B;
7295   //Calculations of Mn, Mw and PDI
7296   col(c)=col(a)*col(b);
7297   col(d)=col(a)^2*col(b);
7298   colstats 2 sum:=[%B]Sheet2!col(a)[1];
7299   copydata irng:=[%B]Sheet2!col(a)[1]
   org:=[%B]Sheet1!col(5);
7300   page.active$=Sheet2;
7301   layer -d;
7302   colstats 3 sum:=[%B]Sheet2!col(a)[1];
7303   copydata irng:=[%B]Sheet2!col(a)[1]
   org:=[%B]Sheet1!col(6);
7304   page.active$=Sheet2;
7305   layer -d;
7306   colstats 4 sum:=[%B]Sheet2!col(a)[1];
7307   copydata irng:=[%B]Sheet2!col(a)[1]
   org:=[%B]Sheet1!col(7);
7308   page.active$=Sheet2;
7309   layer -d;
7310   col(h)=col(f)/col(e);
7311   col(i)=col(g)/col(f);
7312   col(j)=col(i)/col(h);
7313   col(a)=round(col(a),0);
7314   int NoS=count(col(a),1);
7315   for (ii = 1; ii<=NoS; ii++) {
7316     col(k)[ii]=col(a)[ii+1]-col(a)[ii];
7317   };
7318   colstats 11 mean:=col(l)[1] sem:=<new>;
7319   win -a %V;
7320   //Collecting and presenting of calculated data
7321   %P="Mn/Mw/PDI";
7322   if (exist([%V]!P!) != 2) {
7323     newsheet name:=%P cols:=7;
7324     wks.col1.lname$ = "Name";
7325     wks.col2.lname$ = "1. Spectrum";
7326     wks.col3.lname$ = "2. Spectrum";
7327     wks.col4.lname$ = "3. Spectrum";
7328     wks.col5.lname$ = "4. Spectrum";
7329     wks.col6.lname$ = "5. Spectrum";
7330     wks.col7.lname$ = "6. Spectrum";
7331   }
7332   else {
7333     page.active$=%P;
7334   };
7335   copydata irng:=[%B]sheet1!col(h)[1]
   org:=[%V]!P!col(C)[1] clear:=1;
7336   copydata irng:=[%B]sheet1!col(i)[1]
   org:=[%V]!P!col(C)[2] clear:=1;
7337   copydata irng:=[%B]sheet1!col(j)[1]
   org:=[%V]!P!col(C)[3] clear:=1;
7338   copydata irng:=[%B]sheet1!col(l)[1]
   org:=[%V]!P!col(C)[4] clear:=1;
7339   copydata irng:=[%B]sheet1!col(n)[1]
   org:=[%V]!P!col(C)[5] clear:=1;
7340   col(c)[6]=col(c)[2]/col(c)[4];
7341   Mn$=[%B]sheet1!col(h)[1]$;
7342   Mw$=[%B]sheet1!col(i)[1]$;
7343   PDI$=[%B]sheet1!col(j)[1]$;
7344   MEAN$=[%B]sheet1!col(l)[1]$;
7345   SEM$=[%B]sheet1!col(n)[1]$;
7346   DP$=[%V]!P!col(c)[6]$;
7347   col(A)[1]$=Mn;
7348   col(A)[2]$=Mw;
7349   col(A)[3]$=PDI;
7350   col(A)[4]$=Molar weight of repeating unit (mean);
7351   col(A)[5]$=Standard error of repeating unit;
7352   col(A)[6]$=Degree of Polymerization;
7353   worksheet -s 1;
7354   menu -e 35154;
7355   wcolwidth irng:=[%V]!P!Col(1) width:=20;
7356   wcolwidth irng:=[%V]!P!Col(2) width:=10;
7357   wcolwidth irng:=[%V]!P!Col(3) width:=10;
7358   wcolwidth irng:=[%V]!P!Col(4) width:=10;
7359   wcolwidth irng:=[%V]!P!Col(5) width:=10;
7360   wcolwidth irng:=[%V]!P!Col(6) width:=10;
7361   wcolwidth irng:=[%V]!P!Col(7) width:=10;
7362   page.active$=Sheet1;
7363   win -ia;
7364   win -c %B;
7365   win -a %D;
7366   win -z;
7367   type -b "Here are the values for the 2. Spectrum: \n Mn
   \t\t\t\t %(Mn$) \n Mw \t\t\t\t %(Mw$) \n PDI \t\t\t\t
   %(PDI$) \n Molar weight of repeating unit (mean) \t
   %(MEAN$) \n Standard error of repeating unit \t
   %(SEM$) \n Degree of Polymerization \t\t %(DP$)";
7368 };
7369 //Calculations for the 3 layer
7370 if (page.active == 3){
7371   range -w aa=%C;
7372   range bb = uid2range(aa.GetPage());$;
7373   win -a %(bb.name$);
7374   %V=%H;
7375   //Creating a dummy workbook and importing of necessary data
7376   newbook dummy;
7377   %B=%H;
7378   win -a %V;
7379   colcopy irng:=[%V]Sheet1!Col(17)
   org:=[%B]Sheet1!Col(1);
7380   colcopy irng:=[%V]Sheet1!Col(18)
   org:=[%B]Sheet1!Col(2);
7381   win -a %B;
7382   //Calculations of Mn, Mw and PDI
7383   col(c)=col(a)*col(b);
7384   col(d)=col(a)^2*col(b);
7385   colstats 2 sum:=[%B]Sheet2!col(a)[1];
7386   copydata irng:=[%B]Sheet2!col(a)[1]
   org:=[%B]Sheet1!col(5);
7387   page.active$=Sheet2;
7388   layer -d;
7389   colstats 3 sum:=[%B]Sheet2!col(a)[1];
7390   copydata irng:=[%B]Sheet2!col(a)[1]
   org:=[%B]Sheet1!col(6);
7391   page.active$=Sheet2;
7392   layer -d;
7393   colstats 4 sum:=[%B]Sheet2!col(a)[1];
7394   copydata irng:=[%B]Sheet2!col(a)[1]
   org:=[%B]Sheet1!col(7);
7395   page.active$=Sheet2;
7396   layer -d;
7397   col(h)=col(f)/col(e);
7398   col(i)=col(g)/col(f);
7399   col(j)=col(i)/col(h);
7400   col(a)=round(col(a),0);
7401   int NoS=count(col(a),1);
7402   for (ii = 1; ii<=NoS; ii++) {
7403     col(k)[ii]=col(a)[ii+1]-col(a)[ii];
7404   };
7405   colstats 11 mean:=col(l)[1] sem:=<new>;
7406   win -a %V;
7407   //Collecting and presenting of calculated data
7408   %P="Mn/Mw/PDI";
7409   if (exist([%V]!P!) != 2) {
7410     newsheet name:=%P cols:=7;
```


Appendix

LabTalk script (OriginPro 2020b)

```
7411 wks.col1.lname$ = "Name";
7412 wks.col2.lname$ = "1. Spectrum";
7413 wks.col3.lname$ = "2. Spectrum";
7414 wks.col4.lname$ = "3. Spectrum";
7415 wks.col5.lname$ = "4. Spectrum";
7416 wks.col6.lname$ = "5. Spectrum";
7417 wks.col7.lname$ = "6. Spectrum";
7418 }
7419 else {
7420 page.active$=%P;
7421 };
7422 copydata irng:=[%B]sheet1!col(h)[1]
7423   orng:=[%V]%P!col(D)[1] clear:=1;
7423 copydata irng:=[%B]sheet1!col(i)[1]
7424   orng:=[%V]%P!col(D)[2] clear:=1;
7424 copydata irng:=[%B]sheet1!col(j)[1]
7425   orng:=[%V]%P!col(D)[3] clear:=1;
7425 copydata irng:=[%B]sheet1!col(l)[1]
7426   orng:=[%V]%P!col(D)[4] clear:=1;
7426 copydata irng:=[%B]sheet1!col(n)[1]
7427   orng:=[%V]%P!col(D)[5] clear:=1;
7427 col(d)[6]=col(d)[2]/col(d)[4];
7428 Mn$=[%B]sheet1!col(h)[1]$;
7429 Mw$=[%B]sheet1!col(i)[1]$;
7430 PDI$=[%B]sheet1!col(j)[1]$;
7431 MEAN$=[%B]sheet1!col(l)[1]$;
7432 SEM$=[%B]sheet1!col(n)[1]$;
7433 DP$=[%V]%P!col(d)[6]$;
7434 col(A)[1]$=Mn;
7435 col(A)[2]$=Mw;
7436 col(A)[3]$=PDI;
7437 col(A)[4]$=Molar weight of repeating unit (mean);
7438 col(A)[5]$=Standard error of repeating unit;
7439 col(A)[6]$=Degree of Polymerization;
7440 worksheet -s 1;
7441 menu -e 35154;
7442 wcolwidth irng:=[%V]%P!Col(1) width:=20;
7443 wcolwidth irng:=[%V]%P!Col(2) width:=10;
7444 wcolwidth irng:=[%V]%P!Col(3) width:=10;
7445 wcolwidth irng:=[%V]%P!Col(4) width:=10;
7446 wcolwidth irng:=[%V]%P!Col(5) width:=10;
7447 wcolwidth irng:=[%V]%P!Col(6) width:=10;
7448 wcolwidth irng:=[%V]%P!Col(7) width:=10;
7449 page.active$=Sheet1;
7450 win -ia;
7451 win -c %B;
7452 win -a %D;
7453 win -z;
7454 type -b "Here are the values for the 3. Spectrum: \n Mn
\t\t\t\t%(Mn$) \n Mw \t\t\t\t%(Mw$) \n PDI \t\t\t\t
%(PDI$) \n Molar weight of repeating unit (mean) \t
%(MEAN$) \n Standard error of repeating unit \t
%(SEM$) \n Degree of Polymerization \t\t %(DP$)";
7455 };
7456 //Calculations for the 4 layer
7457 if (page.active == 4){
7458 range -w aa=%C;
7459 range bb = uid2range(aa.GetPage());$;
7460 win -a %(bb.name$);
7461 %V=%H;
7462 //Creating of a dummy workbook and importing of nec-
7463   essary data
7463 newbook dummy;
7464 %B=%H;
7465 win -a %V;
7466 colcopy irng:=[%V]Sheet1!Col(19)
7467   orng:=[%B]Sheet1!Col(1);
7467 colcopy irng:=[%V]Sheet1!Col(20)
7468   orng:=[%B]Sheet1!Col(2);
7468 win -a %B;
7469 //Calculations of Mn, Mw and PDI
7470 col(c)=col(a)*col(b);
7471 col(d)=col(a)^2*col(b);
7472 colstats 2 sum:=[%B]Sheet2!col(a)[1];
7473 copydata irng:=[%B]Sheet2!col(a)[1]
7474   orng:=[%B]Sheet1!col(5);
7474 page.active$=Sheet2;
7475 layer -d;
7476 colstats 3 sum:=[%B]Sheet2!col(a)[1];
7477 copydata irng:=[%B]Sheet2!col(a)[1]
7478   orng:=[%B]Sheet1!col(6);
7478 page.active$=Sheet2;
7479 layer -d;
7480 colstats 4 sum:=[%B]Sheet2!col(a)[1];
7481 copydata irng:=[%B]Sheet2!col(a)[1]
7482   orng:=[%B]Sheet1!col(7);
7482 page.active$=Sheet2;
7483 layer -d;
7484 col(h)=col(f)/col(e);
7485 col(i)=col(g)/col(f);
7486 col(j)=col(i)/col(h);
7487 col(a)=round(col(a),0);
7488 int NoS=count(col(a),1);
7489 for (ii = 1; ii<=NoS; ii++) {
7490   col(k)[ii]=col(a)[ii+1]-col(a)[ii];
7491 };
7492 colstats 11 mean:=col(l)[1] sem:=<new>;
7493 win -a %V;
7494 //Collecting and presenting of calculated data
7495 %P="Mn/Mw/PDI";
7496 if (exist([%V]%P!) != 2) {
7497   newsheet name:=%P cols:=7;
7498   wks.col1.lname$ = "Name";
7499   wks.col2.lname$ = "1. Spectrum";
7500   wks.col3.lname$ = "2. Spectrum";
7501   wks.col4.lname$ = "3. Spectrum";
7502   wks.col5.lname$ = "4. Spectrum";
7503   wks.col6.lname$ = "5. Spectrum";
7504   wks.col7.lname$ = "6. Spectrum";
7505 }
7506 else {
7507   page.active$=%P;
7508 };
7509 copydata irng:=[%B]sheet1!col(h)[1]
7510   orng:=[%V]%P!col(E)[1] clear:=1;
7510 copydata irng:=[%B]sheet1!col(i)[1]
7511   orng:=[%V]%P!col(E)[2] clear:=1;
7511 copydata irng:=[%B]sheet1!col(j)[1]
7512   orng:=[%V]%P!col(E)[3] clear:=1;
7512 copydata irng:=[%B]sheet1!col(l)[1]
7513   orng:=[%V]%P!col(E)[4] clear:=1;
7513 copydata irng:=[%B]sheet1!col(n)[1]
7514   orng:=[%V]%P!col(E)[5] clear:=1;
7514 col(e)[6]=col(e)[2]/col(e)[4];
7515 Mn$=[%B]sheet1!col(h)[1]$;
7516 Mw$=[%B]sheet1!col(i)[1]$;
7517 PDI$=[%B]sheet1!col(j)[1]$;
7518 MEAN$=[%B]sheet1!col(l)[1]$;
7519 SEM$=[%B]sheet1!col(n)[1]$;
7520 DP$=[%V]%P!col(e)[6]$;
7521 col(A)[1]$=Mn;
7522 col(A)[2]$=Mw;
7523 col(A)[3]$=PDI;
7524 col(A)[4]$=Molar weight of repeating unit (mean);
7525 col(A)[5]$=Standard error of repeating unit;
7526 col(A)[6]$=Degree of Polymerization;
7527 worksheet -s 1;
7528 menu -e 35154;
7529 wcolwidth irng:=[%V]%P!Col(1) width:=20;
7530 wcolwidth irng:=[%V]%P!Col(2) width:=10;
7531 wcolwidth irng:=[%V]%P!Col(3) width:=10;
7532 wcolwidth irng:=[%V]%P!Col(4) width:=10;
7533 wcolwidth irng:=[%V]%P!Col(5) width:=10;
7534 wcolwidth irng:=[%V]%P!Col(6) width:=10;
7535 wcolwidth irng:=[%V]%P!Col(7) width:=10;
7536 page.active$=Sheet1;
7537 win -ia;
7538 win -c %B;
7539 win -a %D;
```



```

7664 col(k)[ii]=col(a)[ii+1]-col(a)[ii];
7665 };
7666 colstats 11 mean:=col(l)[1] sem:=<new>;
7667 win -a %V;
7668 //Collecting and presenting of calculated data
7669 %P="Mn/Mw/PDI";
7670 if (exist([%V]%P!) != 2) {
7671 newsheet name:=%P cols:=7;
7672 wks.col1.lname$ = "Name";
7673 wks.col2.lname$ = "1. Spectrum";
7674 wks.col3.lname$ = "2. Spectrum";
7675 wks.col4.lname$ = "3. Spectrum";
7676 wks.col5.lname$ = "4. Spectrum";
7677 wks.col6.lname$ = "5. Spectrum";
7678 wks.col7.lname$ = "6. Spectrum";
7679 }
7680 else {
7681 page.active$=%P;
7682 };
7683 copydata irng:=[%B]sheet1!col(h)[1]
omrg:=[%V]%P!col(G)[1] clear:=1;
7684 copydata irng:=[%B]sheet1!col(i)[1]
omrg:=[%V]%P!col(G)[2] clear:=1;
7685 copydata irng:=[%B]sheet1!col(j)[1]
omrg:=[%V]%P!col(G)[3] clear:=1;
7686 copydata irng:=[%B]sheet1!col(l)[1]
omrg:=[%V]%P!col(G)[4] clear:=1;
7687 copydata irng:=[%B]sheet1!col(n)[1]
omrg:=[%V]%P!col(G)[5] clear:=1;
7688 col(g)[6]=col(g)[2]/col(g)[4];
7689 Mn$=[%B]sheet1!col(h)[1]$;
7690 Mw$=[%B]sheet1!col(i)[1]$;
7691 PDI$=[%B]sheet1!col(j)[1]$;
7692 MEAN$=[%B]sheet1!col(l)[1]$;
7693 SEM$=[%B]sheet1!col(n)[1]$;
7694 DP$=[%V]%P!col(g)[6]$;
7695 col(A)[1]$=Mn;
7696 col(A)[2]$=Mw;
7697 col(A)[3]$=PDI;
7698 col(A)[4]$=Molar weight of repeating unit (mean);
7699 col(A)[5]$=Standard error of repeating unit;
7700 col(A)[6]$=Degree of Polymerization;
7701 worksheet -s 1;
7702 menu -e 35154;
7703 wcolwidth irng:=[%V]%P!Col(1) width:=20;
7704 wcolwidth irng:=[%V]%P!Col(2) width:=10;
7705 wcolwidth irng:=[%V]%P!Col(3) width:=10;
7706 wcolwidth irng:=[%V]%P!Col(4) width:=10;
7707 wcolwidth irng:=[%V]%P!Col(5) width:=10;
7708 wcolwidth irng:=[%V]%P!Col(6) width:=10;
7709 wcolwidth irng:=[%V]%P!Col(7) width:=10;
7710 page.active$=Sheet1;
7711 win -ia;
7712 win -c %B;
7713 win -a %D;
7714 win -z;
7715 type -b "Here are the values for the 6. Spectrum: \n Mn
\t\t\t\t % (Mn$) \n Mw \t\t\t\t % (Mw$) \n PDI \t\t\t\t
%(PDI$) \n Molar weight of repeating unit (mean) \t
%(MEAN$) \n Standard error of repeating unit \t
%(SEM$) \n Degree of Polymerization \t\t % (DP$)";
7716 };
7717 };
7718 };

```

8.1.8. Section “nachbearbeitung” (L. 7720 - 7751)

```

7720 [nachbearbeitung]
7721 type -b "This skript will reset all legends and the set-
tings of all peak plots.";
7722 %W = %H;
7723 %W!page.active = 1;
7724 leglay1.text$ = \b(%(1Y, @LC));
7725 themeApply2g theme:= "MS";
7726 if (exist([%W]2!,10) == 3) {
7727 %W!page.active = 2;
7728 leglay2.text$ = \b(%(1Y, @LC));
7729 themeApply2g theme:= "MS";
7730 };
7731 if (exist([%W]3!,10) == 3) {
7732 %W!page.active = 3;
7733 leglay3.text$ = \b(%(1Y, @LC));
7734 themeApply2g theme:= "MS";
7735 };
7736 if (exist([%W]4!,10) == 3) {
7737 %W!page.active = 4;
7738 leglay4.text$ = \b(%(1Y, @LC));
7739 themeApply2g theme:= "MS";
7740 };
7741 if (exist([%W]5!,10) == 3) {
7742 %W!page.active = 5;
7743 leglay5.text$ = \b(%(1Y, @LC));
7744 themeApply2g theme:= "MS";
7745 };
7746 if (exist([%W]6!,10) == 3) {
7747 %W!page.active = 6;
7748 leglay6.text$ = \b(%(1Y, @LC));
7749 themeApply2g theme:= "MS";
7750 };
7751 };

```

8.1.9. Section “hide” (L. 7753 - 7819)

```

7753 [hide]
7754 %W = %H;
7755 %W!page.active = 1;
7756 if (exist(yscrollbar,16) == 16){
7757 yscrollbar.show=0;
7758 yscrollthumb.show=0;
7759 };
7760 if (exist(XMin,16) == 16){
7761 XMin.show=0;
7762 XMax.show=0;
7763 };
7764 if (exist([%W]2!,10) == 3) {
7765 %W!page.active = 2;
7766 if (exist(yscrollbar,16) == 16){
7767 yscrollbar.show=0;
7768 yscrollthumb.show=0;
7769 };
7770 if (exist(XMin,16) == 16){
7771 XMin.show=0;
7772 XMax.show=0;
7773 };
7774 };
7775 if (exist([%W]3!,10) == 3) {
7776 %W!page.active = 3;
7777 if (exist(yscrollbar,16) == 16){
7778 yscrollbar.show=0;
7779 yscrollthumb.show=0;
7780 };
7781 if (exist(XMin,16) == 16){
7782 XMin.show=0;
7783 XMax.show=0;
7784 };
7785 };
7786 if (exist([%W]4!,10) == 3) {
7787 %W!page.active = 4;
7788 if (exist(yscrollbar,16) == 16){

```

```
7789 yscrollbar.show=0;
7790 yscrollthumb.show=0;
7791 };
7792 if (exist(XMin,16) == 16){
7793 XMin.show=0;
7794 XMax.show=0;
7795 };
7796 };
7797 if (exist([%W]5!,10) == 3) {
7798 %W!page.active = 5;
7799 if (exist(yscrollbar,16) == 16){
7800 yscrollbar.show=0;
7801 yscrollthumb.show=0;
7802 };
7803 if (exist(XMin,16) == 16){
7804 XMin.show=0;
7805 XMax.show=0;
7806 };
7807 };
7808 if (exist([%W]6!,10) == 3) {
7809 %W!page.active = 6;
7810 if (exist(yscrollbar,16) == 16){
7811 yscrollbar.show=0;
7812 yscrollthumb.show=0;
7813 };
7814 if (exist(XMin,16) == 16){
7815 XMin.show=0;
7816 XMax.show=0;
7817 };
7818 };
7819 };
```

8.1.10. Section “show” (L. 7821 - 7887)

```
7821 [show]
7822 %W = %H;
7823 %W!page.active = 1;
7824 if (exist(yscrollbar,16) == 16){
7825 yscrollbar.show=1;
7826 yscrollthumb.show=1;
7827 };
7828 if (exist(XMin,16) == 16){
7829 XMin.show=1;
7830 XMax.show=1;
7831 };
7832 if (exist([%W]2!,10) == 3) {
7833 %W!page.active = 2;
7834 if (exist(yscrollbar,16) == 16){
7835 yscrollbar.show=1;
7836 yscrollthumb.show=1;
7837 };
7838 if (exist(XMin,16) == 16){
7839 XMin.show=1;
7840 XMax.show=1;
7841 };
7842 };
7843 if (exist([%W]3!,10) == 3) {
7844 %W!page.active = 3;
7845 if (exist(yscrollbar,16) == 16){
7846 yscrollbar.show=1;
7847 yscrollthumb.show=1;
7848 };
7849 if (exist(XMin,16) == 16){
7850 XMin.show=1;
7851 XMax.show=1;
7852 };
7853 };
7854 if (exist([%W]4!,10) == 3) {
7855 %W!page.active = 4;
7856 if (exist(yscrollbar,16) == 16){
7857 yscrollbar.show=1;
7858 yscrollthumb.show=1;
7859 };
7860 if (exist(XMin,16) == 16){
7861 XMin.show=1;
7862 XMax.show=1;
7863 };
7864 };
7865 if (exist([%W]5!,10) == 3) {
7866 %W!page.active = 5;
7867 if (exist(yscrollbar,16) == 16){
7868 yscrollbar.show=1;
7869 yscrollthumb.show=1;
7870 };
7871 if (exist(XMin,16) == 16){
7872 XMin.show=1;
7873 XMax.show=1;
7874 };
7875 };
7876 if (exist([%W]6!,10) == 3) {
7877 %W!page.active = 6;
7878 if (exist(yscrollbar,16) == 16){
7879 yscrollbar.show=1;
7880 yscrollthumb.show=1;
7881 };
7882 if (exist(XMin,16) == 16){
7883 XMin.show=1;
7884 XMax.show=1;
7885 };
7886 };
7887 };
```

8.1.11. Section “splitspectrum” (L. 7889 - 8127)

```
7889 [splitspectrum]
7890 type -b "With this script you can put specified breaks into
the x-axis of your plot and so highlight special areas.";
7891 %W=%H;
7892 BX$=layer1.x1.from$;
7893 IX$=layer1.x1.inc$;
7894 EX$=layer1.x1.to$;
7895
7896 dlgChkList inames:="1|2|3|4" osel:=NoP title:="Area for
peak search" desc:="How many areas of the plot do you
want to highlight? Check only ONE box" olist:=CB;
7897
7898 if (CB.GetLength() == 1) {
7899 //Skript for 1 area
7900 if (NoP == 1) {
7901 getn
7902 (Signal 1) :@G
7903 (m/z) S11
7904 (+/-) S11PM
7905 (Increment) S11I
7906 (-) :@G
7907 (Please insert the m/z of the middle of the areas (= m/z +/-
50) you want to highlight.);
7908 layer1.x1.from=S11-S11PM;
7909 layer1.x1.inc = S11I;
7910 layer1.x1.to=S11+S11PM;
7911 };
7912 //Skript for 2 areas
7913 if (NoP == 2) {
7914 getn
7915 (Signal 1) :@G
7916 (m/z) S21
7917 (+/-) S21PM
7918 (Increment) S21I
```

Appendix

LabTalk script (OriginPro 2020b)

```
7919 (-):@G
7920 (Signal 2):@G
7921 (m/z) S22
7922 (+/-) S22PM
7923 (Increment) S22I
7924 (-):@G
7925 (Please insert the m/z of the signal you want to highlight
and give the size of the area around it (the areas must not
overlap) including the increment of the x-axis (recom-
mended --> 20% of area size value).);
7926 if (sqrt((S21-S22)^2) > (S21PM+S22PM)) {
7927 //General x axis manipulation
7928 layer1.x1.from=S21-S21PM;
7929 layer1.x1.inc = S21I;
7930 layer1.x1.to=S22+S22PM;
7931 layer1.x1.breaks.enable=1;
7932 layer1.x1.breaks.count=1;
7933 //Break 1
7934 layer1.x1.break1.from = $(S21+S21PM);
7935 layer1.x1.break1.to = $(S22-S22PM);
7936 layer1.x1.break1.pos = 50;
7937 layer1.x1.break1.inc = S22I;
7938 if (exist([%W]2!,10) == 3) {
7939 layer2.x1.break1.inc = S22I;
7940 };
7941 if (exist([%W]3!,10) == 3) {
7942 layer3.x1.break1.inc = S22I;
7943 };
7944 if (exist([%W]4!,10) == 3) {
7945 layer4.x1.break1.inc = S22I;
7946 };
7947 if (exist([%W]5!,10) == 3) {
7948 layer5.x1.break1.inc = S22I;
7949 };
7950 if (exist([%W]6!,10) == 3) {
7951 layer6.x1.break1.inc = S22I;
7952 };
7953 }
7954 else {
7955 type -b "Please start again and read the instructions care-
fully!";
7956 };
7957 };
7958 //Skript for 3 areas
7959 if (NoP == 3) {
7960 getn
7961 (Signal 1):@G
7962 (m/z) S31
7963 (+/-) S31PM
7964 (Increment) S31I
7965 (-):@G
7966 (Signal 2):@G
7967 (m/z) S32
7968 (+/-) S32PM
7969 (Increment) S32I
7970 (-):@G
7971 (Signal 3):@G
7972 (m/z) S33
7973 (+/-) S33PM
7974 (Increment) S33I
7975 (-):@G
7976 (Please insert the m/z of the signal you want to highlight
and give the size of the area around it (the areas must not
overlap) including the increment of the x-axis (recom-
mended --> 40% of area size value).);
7977 if (sqrt((S31-S32)^2) > (S31PM+S32PM) && sqrt((S31-
S33)^2) > (S31PM+S33PM) && sqrt((S32-S33)^2) >
(S32PM+S33PM)) {
7978 //General x axis manipulation
7979 layer1.x1.from=$(S31-S31PM);
7980 layer1.x1.inc = S31I;
7981 layer1.x1.to=$(S33+S33PM);
7982 layer1.x1.breaks.enable=1;
7983 layer1.x1.breaks.count=2;
7984 //Break 1
7985 layer1.x1.break1.from = $(S31+S31PM);
7986 layer1.x1.break1.to = $(S32-S32PM);
7987 layer1.x1.break1.pos = 33;
7988 layer1.x1.break1.inc = S32I;
7989 if (exist([%W]2!,10) == 3) {
7990 layer2.x1.break1.inc = S32I;
7991 };
7992 if (exist([%W]3!,10) == 3) {
7993 layer3.x1.break1.inc = S32I;
7994 };
7995 if (exist([%W]4!,10) == 3) {
7996 layer4.x1.break1.inc = S32I;
7997 };
7998 if (exist([%W]5!,10) == 3) {
7999 layer5.x1.break1.inc = S32I;
8000 };
8001 if (exist([%W]6!,10) == 3) {
8002 layer6.x1.break1.inc = S32I;
8003 };
8004 //Break 2
8005 layer1.x1.break2.from = $(S32+S32PM);
8006 layer1.x1.break2.to = $(S33-S33PM);
8007 layer1.x1.break2.pos = 66;
8008 layer1.x1.break2.inc = S33I;
8009 if (exist([%W]2!,10) == 3) {
8010 layer2.x1.break2.inc = S33I;
8011 };
8012 if (exist([%W]3!,10) == 3) {
8013 layer3.x1.break2.inc = S33I;
8014 };
8015 if (exist([%W]4!,10) == 3) {
8016 layer4.x1.break2.inc = S33I;
8017 };
8018 if (exist([%W]5!,10) == 3) {
8019 layer5.x1.break2.inc = S33I;
8020 };
8021 if (exist([%W]6!,10) == 3) {
8022 layer6.x1.break2.inc = S33I;
8023 };
8024 }
8025 else {
8026 type -b "Please start again and read the instructions care-
fully!";
8027 };
8028 };
8029 //Skript for 4 areas
8030 if (NoP == 4) {
8031 getn
8032 (Signal 1):@G
8033 (m/z) S41
8034 (+/-) S41PM
8035 (Increment) S41I
8036 (-):@G
8037 (Signal 2):@G
8038 (m/z) S42
8039 (+/-) S42PM
8040 (Increment) S42I
8041 (-):@G
8042 (Signal 3):@G
8043 (m/z) S43
8044 (+/-) S43PM
8045 (Increment) S43I
8046 (-):@G
8047 (Signal 4):@G
8048 (m/z) S44
8049 (+/-) S44PM
8050 (Increment) S44I
8051 (-):@G
8052 (Please insert the m/z of the signal you want to highlight
and give the size of the area around it (the areas must not
overlap) including the increment of the x-axis (recom-
mended --> 50% of area size value).);
8053 if (sqrt((S41-S42)^2) > (S41PM+S42PM) && sqrt((S41-
S43)^2) > (S41PM+S43PM) && sqrt((S41-S44)^2) >
(S41PM+S44PM) && sqrt((S42-S43)^2) >
```

```
(S42PM+S43PM) && sqrt((S42-S44)^2) >
(S42PM+S44PM) && sqrt((S43-S44)^2) >
(S43PM+S44PM)) {
8054 //General x axis manipulation
8055 layer1.x1.from=$(S41-S41PM);
8056 layer.x1.inc = S41I;
8057 layer1.x1.to=$(S44+S44PM);
8058 layer1.x1.breaks.enable=1;
8059 layer1.x1.breaks.count=3;
8060 //Break 1
8061 layer1.x1.break1.from = $(S41+S41PM);
8062 layer1.x1.break1.to = $(S42-S42PM);
8063 layer1.x1.break1.pos = 25;
8064 layer1.x1.break1.inc = S42I;
8065 if (exist([%W]2!,10) == 3) {
8066 layer2.x1.break1.inc = S42I;
8067 };
8068 if (exist([%W]3!,10) == 3) {
8069 layer3.x1.break1.inc = S42I;
8070 };
8071 if (exist([%W]4!,10) == 3) {
8072 layer4.x1.break1.inc = S42I;
8073 };
8074 if (exist([%W]5!,10) == 3) {
8075 layer5.x1.break1.inc = S42I;
8076 };
8077 if (exist([%W]6!,10) == 3) {
8078 layer6.x1.break1.inc = S42I;
8079 };
8080 //Break 2
8081 layer1.x1.break2.from = $(S42+S42PM);
8082 layer1.x1.break2.to = $(S43-S43PM);
8083 layer1.x1.break2.pos = 50;
8084 layer1.x1.break2.inc = S43I;
8085 if (exist([%W]2!,10) == 3) {
8086 layer2.x1.break2.inc = S43I;
8087 };
8088 if (exist([%W]3!,10) == 3) {
8089 layer3.x1.break2.inc = S43I;
8090 };
8091 if (exist([%W]4!,10) == 3) {
8092 layer4.x1.break2.inc = S43I;
8093 };
8094 if (exist([%W]5!,10) == 3) {
8095 layer5.x1.break2.inc = S43I;
8096 };
8097 if (exist([%W]6!,10) == 3) {
8098 layer6.x1.break2.inc = S43I;
8099 };
8100 //Break 3
8101 layer1.x1.break3.from = $(S43+S43PM);
8102 layer1.x1.break3.to = $(S44-S44PM);
8103 layer1.x1.break3.pos = 75;
8104 layer1.x1.break3.inc = S44I;
8105 if (exist([%W]2!,10) == 3) {
8106 layer2.x1.break3.inc = S44I;
8107 };
8108 if (exist([%W]3!,10) == 3) {
8109 layer3.x1.break3.inc = S44I;
8110 };
8111 if (exist([%W]4!,10) == 3) {
8112 layer4.x1.break3.inc = S44I;
8113 };
8114 if (exist([%W]5!,10) == 3) {
8115 layer5.x1.break3.inc = S44I;
8116 };
8117 if (exist([%W]6!,10) == 3) {
8118 layer6.x1.break3.inc = S44I;
8119 };
8120 }
8121 else {
8122 type -b "Please start again and read the instructions carefully!";
8123 };
8124 };
8125 }
8126 else type -b "Please start again and choose only one amount of highlighted areas!";
8127 };
```

8.1.12. Section “splitspectrumstop” (L. 8129 - 8135)

```
8129 [splitspectrumstop]
8130 //With this script you can reset your plots to the scales before introducing breaks (--> "Splitspektrum"-script).
8131 layer1.x1.breaks.enable=0;
8132 layer1.x1.from$=BX$;
8133 layer1.x1.inc$=IX$;
8134 layer1.x1.to$=EX$;
8135 };
```


8.2. Supporting Information

8.2.1. Conjugated Polymers as a New Class of Dual-Mode Matrices for MALDI Mass Spectrometry and Imaging (Publication 1)

Conjugated Polymers as a New Class of Dual-Mode Matrices for MALDI Mass Spectrometry and Imaging

Supporting Information

Kilian Horatz^{†‡}, Marco Giampà[§], Yevhen Karpov[†], Karin Sahre[†], Hanna Bednarz[§], Anton Kiriya[†], Brigitte Voit^{†‡}, Karsten Niehaus[§], Nikos Hadjichristidis[#], Dominik L. Michels[‡], and Franziska Lissel^{†*}

[†]Institute of Macromolecular Chemistry, Leibniz-Institut für Polymerforschung Dresden e.V., Hohe Straße 6, 01069 Dresden, Germany

[‡]Organic Chemistry of Polymers, Technische Universität Dresden, Mommsenstrasse 4, 01062 Dresden, Germany

[§]Center for Biotechnology and Department for Proteome and Metabolome Research, Faculty of Biology, Bielefeld University, Universitätsstraße 25, 33615 Bielefeld, Germany

[#]Physical Sciences and Engineering Division, KAUST Catalysis Center, Polymer Synthesis Laboratory, King Abdullah University of Science and Technology (KAUST), Thuwal 23955, Kingdom of Saudi Arabia

[‡]Computer, Electrical and Mathematical Sciences and Engineering Division, KAUST visual Computing Center, Computational Sciences Group, King Abdullah University of Science and Technology (KAUST), Thuwal 23955, Kingdom of Saudi Arabia

*lissel@ipfdd.de

1. Materials

The small molecular matrices 9-aminoacridine (9AA, $\geq 99.5\%$), 2,5-dihydroxybenzoic acid (2,5-DHB, $>99\%$), and trans-2-[3-(4-*tert*-butylphenyl)-2-methyl-2-propenyldene]malonitrile (DCTB, $>99\%$) were purchased commercially and used without further purification (see Figure S1 for the molecular structures).

The polymers (see Figure S2) poly(3-dodecylthiophene-2,5-diyl) (P3DDT), poly([2,3-bis-(3-octyloxyphenyl)quinoxaline-5,8-diyl]-*alt*-(thiophene-2,5-diyl)) (PTQ1), and poly(9,9-di-*n*-octylfluorenyl-2,7-diyl) (P9OFI) were purchased commercially, poly([N,N'-bis(2-octyldodecyl)-isoidigo-5,5'-diyl]-*alt*-5,5'-(2,2'-bithiophene)) (PII(T2)) and poly([N,N'-bis(2-octyldodecyl)-naphthalene-1,4,5,8-bis(dicarboximide)-2,6-diyl]-*alt*-5,5'-(2,2'-bithiophene)) (PNDI(T2)) were synthesized and purified according to literature^[1,2]. PTQ1 was used without further purification, all other polymers were further purified via Soxhlet extraction in acetone (P3DDT, PNDI(T2), PII(T2), P9OFI) and hexane (PNDI(T2), PII(T2)). M_n and PD values for polymers (see Table S1) were obtained by GPC (40 °C, chloroform) on Agilent 1260 Infinity System with ResiPore columns calibrated with polystyrene (PS) standards.

All solvents were purchased commercially in analytical grade ($>99\%$) and used without further purification. All small molecular analytes listed in Table S1 were purchased commercially ($>98\%$) and used without further purification. Table S2 lists observed peaks with their parent analyte and proposed molecular formula. As complex sample a *Ginkgo biloba* extract (secondary reference standard) was purchased commercially from Sigma Aldrich (CAS-No.: 90045-36-6, assay bilobalide: 155.19 mg/g; assay ginkgolide: 67.99 mg/g).

Table S1: Molecular weight and polydispersity of polymer matrices used in this work.

Polymer matrix	M_n * [kg/mol]	M_w * [kg/mol]	PD
PNDI(T2)	28.1	138.9	4.9
P3DDT	38.0	92.6	2.4
PTQ1	12.7	50.1	3.9
PII(T2)	20.3	81.3	4.0
P9OFI	29.2	215.6	7.4

*GPC, 40 °C, chloroform

Table of Content

1. Materials.....	S3
2. Analytical measurements.....	S7
2.1. UV/Vis.....	S7
2.2. MALDI MS	S9
2.2.1. MALDI instrument.....	S9
2.2.2. Calibration	S10
2.2.3. Sample plate cleaning procedure.....	S11
2.2.4. Sample preparation.....	S11
2.2.5. Single Matrix/Analyte Spectra	S12
2.2.6. Complex sample (Ginkgo Biloba).....	S50
2.2.7. Detection limit	S60
2.2.8. Polymers as analytes.....	S85
2.2.9. Comparison spectra	S91
2.3. MALDI MS Imaging.....	S99
3. References.....	S100

Figure S1: Molecular structures of 2,5-dihydroxybenzoic acid (2,5-DHB, left), trans-2-[3-(4-tert-butylphenyl)-2-methyl-2-propenyldene]malononitrile (DCTB, middle) and 9-aminoacridine (9AA, right).

Figure S2: Molecular structures of a) PNDI(T2), b) P3DDT, c) P9OFI, d) PTQI and e) PI(T2).

Table S2: Names, molecular formulas, weights and structures of the investigated analytes.

Name	Abbrev.	Molecular formula	Molecular weight [g/mol]	Molecular structure
Reserpine	RP	$C_{23}H_{40}N_2O_9$	608	
Tetraethyl-ammonium chloride	TEAC	$C_8H_{20}ClN$	165	
Coumaphos	CP	$C_{14}H_{16}ClO_4PS$	362	
Cholic acid	ChA	$C_{26}H_{46}O_5$	408	
Citric acid	CiA	$C_6H_8O_7$	192	
L-Phenyl-alanine	PA	$C_9H_9NO_2$	165	

2. Analytical measurements

2.1. UV/Vis

UV/Vis measurements were carried out on a Perkin Elmer "Lamda800" instrument. The parameters of the measurements are listed in Table S4. For the preparation of thin film samples multiple droplets ($V = 1 \mu\text{l}$; $c = 1 \text{ mg/ml}$ in chloroform) of the polymer solutions were dripped on a glass substrate as depicted in Figure S3 and dried in air until the solvent had evaporated (approximately 1 minute). The measurements were carried out in dual beam mode with an empty glass substrate as reference. The recorded UV/vis spectra are depicted in Figure S4, Table S5 lists the obtained absorption maxima.

Table S4: Parameters for the UV/vis measurements of the polymer matrices

Parameter	Setting
Wavelength area	250 - 800 nm
Scan Data Interval	1 nm
Integration Time	0.56 sec.
Scan Ordinate Mode	Absorption
Scan Speed	100 nm/min
Beam Mode	Dual beam
Slit	2
Slit Mode	Fix

Figure S3: Scheme depicting the sample preparation for the UV/vis measurements.

Table S3: Nomenclature of observed fragments of the investigated analytes.

Fragment	Parent analyte	Proposed molecular formula	Observed m/z [g/mol]
[RP1] ⁺	Reserpine RP	C ₂₃ H ₂₉ N ₅ O ₄	397
[RP2] ⁺ [3]	Reserpine RP	C ₁₀ H ₁₁ O ₄	195
[RP3] ⁺	Reserpine RP	(-)	395
[TEA] ⁺	Tetraethylammonium chloride TEAC	C ₈ H ₂₀ N	130
[TEAC] ⁺	Tetraethylammonium chloride TEAC	C ₁₅ H ₁₇ N	86
[CIA] ⁺	Citric acid CIA	C ₆ H ₈ O ₇	87

Figure S4: UV/Vis spectra of the polymer matrices a) PNDI(T2) (yellow), b) P3DDT (blue), c) PTQ1 (green), d) PI(T2) (red) and e) P9OFI (black).

Table S5: UV/Vis absorption maxima of the polymer matrices.

Polymer	PNDI(T2)	P3DDT	PTQ1	PI(T2)	P9OFI
Absorption maxima [nm] ^a	392 (100%), 697 (81%)	523 (100%), 549 (95%), 601 (61%)	363 (95%), 622 (100%)	413 (67%), 626 (100%), 686 (89%)	386 (100%), 432 (13%)

^a percentage in brackets is relative to the maximum absorption of each compound.

2.2. MALDI MS

2.2.1. MALDI instrument

Unless noted otherwise, all MALDI spectra were recorded with a "Bruker autoflex speed" MALDI-TOF/TOF in reflectron mode. The laser of this instrument is a smartbeam-II with a wavelength of 355 nm.^[6] The software for measuring and evaluating the spectra is FlexControl 3.3 and FlexAnalysis 3.3. All measurements were carried out using a standard set of parameters (see Table S6).

Table S6: Parameters for positive and negative reflectron & linear mode MALDI measurements.

Parameter	Reflectron		Linear	
	Positive	Negative	Positive	Negative
m/z range	0 - 3000	0 - 3000	0 - 200k	0 - 160k
Sample rate	4.0	4.0	0.5	0.5
Shots per measurement	500	500	1000	1000
Matrix suppression	Off	Off	Off	Off
Measurements per spectra	10	10	10	10
Random walk	Off	Off	Off	Off
Electronic gain	Enhanced	Enhanced	Enhanced	Enhanced
Realtime smooth	Off	Off	High	High
Laser-frequency	1000 Hz	1000 Hz	1000 Hz	1000 Hz
Laser-size	Large	Large	Large	Large

2.2.2. Calibration

2.2.2.1. Low molecular weight area (0 - 3000 m/z)

At the beginning of each set of low molecular weight MALDI measurements in reflectron mode and after changing the measurement polarity, the instrument was calibrated with red phosphorus ($c = 5 \text{ mg/ml}$ in acetonitrile) as described in literature.^[5] The phosphorus clusters used for the calibrations are listed in Table S7.

Table S7: Phosphorus clusters and corresponding m/z values used for the positive and negative mode calibration.

Phosphorus cluster	Positive [m/z]	Negative [m/z]
$P_{33}^{+/-}$	712.39597	712.39706
$P_{28}^{+/-}$	774.34349	774.34459
$P_{27}^{+/-}$	836.29101	836.29211
$P_{41}^{+/-}$	1269.92367	1269.92477
$P_{49}^{+/-}$	1517.71377	1517.71486

2.2.2.2. High molecular weight area (1k - 14k m/z)

At the beginning of each set of high molecular weight MALDI measurements in linear positive mode, the instrument was calibrated with a 1:1 mixture of 2,5-dihydroxybenzoic acid (2,5-DHB) and polyethylene glycol (PEG) ($M_w = 6650 \text{ Da}$) in acetone as described in literature.^[6]

2.2.3. Sample plate cleaning procedure

The sample plate was cleaned using the following procedure:

1. Wipe sample plate with wet paper tissue (each once: Chloroform, acetonitrile, acetone);
2. Ultra-sonication for 10 min. at rt in 150 ml Milli-Q H_2O ;
3. Rinse with 50 ml acetone;
4. Ultra-sonication for 10 min. at rt in acetonitrile / acetone (100 ml / 50 ml);
5. Rinse with 50 mL acetone;
6. Ultra-sonication for 10 min. at rt in chloroform / ethanol (100 ml / 50 ml);
7. Rinse with 50 ml ethanol;
8. Rinse with 50 ml chloroform.

2.2.4. Sample preparation

The polymer matrices PNDI(T2), P3DDT, PH(T2), PTQ1 and P9OFI were dissolved in toluene and the small molecular matrices 2,5-DHB and 9AA in acetone. The analytes cholic acid (ChA), citric acid (CiA), reserpine (RP) and coumaphos (CP) were dissolved in acetone. L-Phenylalanine (PA) was dissolved in methanol, tetraethylammonium chloride (TEAC) in acetonitrile. The samples were prepared with on a standard sample plate (Bruker "MTP 384 target plate ground steel BC" Part-No.: 8280784).^[7-8] Preparation method was the thin layer method, a variation of the dried droplet method: First the individual analyte solutions were dripped on the plate ($V = 1 \mu\text{l}$) and dried in air for several seconds to several minutes until the solvent had completely evaporated. Then the individual matrices were coated by dripping the corresponding solutions ($V = 1 \mu\text{l}$) on top of the analytes and dried in air for around 1 - 2 minutes.^[7]

2.2.5. Single Matrix/Analyte Spectra

For each matrix the best layer thickness combination was qualitatively determined by testing several combinations of analyte and matrix concentration. Concentrations were varied from 0.1 mg/ml, 1 mg/ml, and 10 mg/ml for both the matrix solution and the analyte solution. Reserpine (RP) was used for the positive mode, and cholic acid (ChA) for the negative mode, and the determined best layer thickness was used for all subsequent measurements of the matrix.

Tables listing the tested concentrations are given at the beginning of each section discussing a matrix system. The combination marked in yellow was qualitatively determined to be the best and subsequently used.

2.2.5.1. 2,5-Dihydroxybenzoic acid (2,5-DHB)

Table S8: Qualitative comparison of different layer thickness combinations for positive mode measurements with 2,5-DHB.

c (Matrix) [mg/ml]	c (Analyte) [mg/ml]	Comments about the resulting spectra
10	10	good
	1	good
	0.1	ok
1	10	bad
	1	ok
	0.1	bad
0.1	10	very bad
	1	ok
	0.1	bad

2.2.5.1.1. Reserpine (RP)

Table S9: Observed peaks in the 2,5-DHB - RP spectrum and proposed species.

Peak m/z	Origin ^a
137	[DHB-OH] ⁺
154	[DHB] ⁺
177	[DHB+Na] ⁺
215	(-) ^b
397	[RP] ⁺
607	[RP-H] ⁺
609	[RP+H] ⁺

^a see Table S2 and Table S3 for abbreviations; ^b clear assignment not possible;

Table S10: Observed peaks in the 2,5-DHB - TEAC spectrum and proposed species.

Signal m/z	Origin ^a
86	[TEAC] ⁺
130	[TEA] ⁺
137	[DHB-OH] ⁺
154	[DHB] ⁺
177	[DHB+Na] ⁺
193	[DHB+K] ⁺
215	(-) ^b
347	[2xDHB+K] ⁺

^a see Table S2 and Table S3 for abbreviations; ^b clear assignment not possible;

Table S11: Observed peaks in the 2,5-DHB - CP spectrum and proposed species.

Signal m/z	Origin ^a
137	[DHB-OH] ⁺
154	[DHB] ⁺
177	[DHB+Na] ⁺
215	(-) ^b
347	[2xDHB+K] ⁺
363	[CP+H] ⁺

^a see Table S2 and Table S3 for abbreviations; ^b clear assignment not possible;

2.2.5.2. 9-Aminoacridine (9AA)

Table S12: Qualitative comparison of different layer thickness combinations for negative mode measurements with 9AA.

c (Matrix) [mg/ml]	c (Analyte) [mg/ml]	Comments about the resulting spectra
10	10	ok
	1	ok
	0.1	very bad
1	10	good
	1	good
	0.1	very bad
0.1	10	bad
	1	ok
	0.1	ok

2.2.5.2.1. Cholic acid (ChA)

Figure S8: Reflection negative mode MALDI MS spectra of a) sample plate (green), b) only the analyte ChA (red), c) only the pure matrix 9AA (blue) and d) 9AA coated on ChA (black).

Table S13: Observed peaks in the 9AA - ChA spectrum and proposed species.

Signal m/z	Origin*
35	Cl ⁻
89 ^[9]	Matrix related signal
97 ^[10]	Matrix related signal
193	[9AA-H] ⁻
229 ^[9]	[9AA+Cl] ⁻
407	[ChA-H] ⁻

* see Table S2 and Table S3 for abbreviations.

Table S14: Observed peaks in the 9AA - CiA spectrum and proposed species.

Signal m/z	Origin ^a
89 ^[9]	Matrix related signal
97 ^[9]	Matrix related signal
191	[CiA-H] ⁺
193	[9AA-H] ⁺
229 ^[9]	[9AA+Cl] ⁺

^a see Table S2 and Table S3 for abbreviations;

Table S15: Observed peaks in the 9AA - PA spectrum and proposed species.

Signal m/z	Origin ^a
35	Cl ⁻
89 ^[9]	Matrix related signal
97 ^[9]	Matrix related signal
164	[PA-H] ⁺
193	[9AA-H] ⁺
229 ^[9]	[9AA+Cl] ⁺

^a see Table S2 and Table S3 for abbreviations;

2.2.5.3. Poly[*(N,N'*-bis(2-octyldodecyl)-naphthalene-1,4,5,8-bis(dicarboximide)-2,6-diyl)-*alt*-5,5'-(2,2'-bithiophene)] (PNDI(T2))

Table S16: Qualitative comparison of different layer thickness combinations for positive mode measurements with PNDI(T2).

c (Matrix) [mg/ml]	c (Analyte) [mg/ml]	Comments about the resulting spectra
10	10	good
	1	good
	0.1	bad
1	10	very good
	1	very good
	0.1	good
0.1	10	very good
	1	very good
	0.1	ok

Table S17: Qualitative comparison of different layer thickness combinations for negative mode measurements with PNDI(T2).

c (Matrix) [mg/ml]	c (Analyte) [mg/ml]	Comments about the resulting spectra
10	10	bad
	1	bad
	0.1	analyte not visible
1	10	bad
	1	bad
	0.1	bad
0.1	10	bad
	1	ok
	0.1	bad

2.2.5.3.1. Reserpine (RP)

Table S18: Observed peaks in the PNDI(T2) - RP spectrum and proposed species.

Signal m/z	Origin ^a
39	K ⁺
395	[RP3] ⁺
607	[RP-H] ⁺

^a see Table S2 and Table S3 for abbreviations;

Table S19: Observed peaks in the PNDI(T2) - TEAC spectrum and proposed species.

Signal m/z	Origin ^a
86	[TEAC] ⁺
130	[TEA] ⁺

^a see Table S2 and Table S3 for abbreviations;

- S22 -

Table S20: Observed peaks in the PNDI(T2) - CP spectrum and proposed species.

Signal m/z	Origin ^a
43	[C ₂ H ₃] ⁺
57	[C ₂ H ₅] ⁺
71	[C ₂ H ₇] ⁺
85	[C ₂ H ₉] ⁺
363	[Coumaphos+H] ⁺

^a see Table S2 and Table S3 for abbreviations;

- S23 -

2.2.5.3.4. Cholic acid (Cha)

Table S21: Observed peaks in the PND(T2) - Cha spectrum and proposed species.

Signal m/z	Origin ^a
35	Cl ⁻
42	[C ₃ H ₆] ⁺
79/81	Br ⁻
407	[Cha-H] ⁺

^a see Table S2 and Table S3 for abbreviations;

2.2.5.3.5. Citric acid (Cia)

Table S22: Observed peaks in the PND(T2) - Cia spectrum and proposed species.

Signal m/z	Origin ^a
42	[C ₃ H ₆] ⁺
87	[C ₃ H ₆] ⁺
191	[Cia-H] ⁺

^a see Table S2 and Table S3 for abbreviations;

2.2.5.4. Poly(3-dodecylthiophene-2,5-diyl) (P3DDT)

Table S24: Qualitative comparison of different layer thickness combinations for positive mode measurements with P3DDT.

c (Matrix) [mg/ml]	c (Analyte) [mg/ml]	Comments about the resulting spectra
10	10	bad
	1	bad
	0.1	bad
1	10	bad
	1	bad
	0.1	very bad
0.1	10	ok
	1	bad
	0.1	bad

Table S25: Qualitative comparison of different layer thickness combinations for negative mode measurements with P3DDT.

c (Matrix) [mg/ml]	c (Analyte) [mg/ml]	Comments about the resulting spectra
10	10	bad
	1	very bad
	0.1	very bad
1	10	ok
	1	bad
	0.1	very bad
0.1	10	bad
	1	bad
	0.1	very bad

2.2.5.3.6. L-Phenylalanine (PA)

Table S23: Observed peaks in the PND(T2) - PA spectrum and proposed species.

Signal m/z	Origin ^a
35	Cl ⁻
42	[C ₃ H ₆] ⁻
57	[C ₄ H ₈ O] ⁻
79/81	Br ⁻
164	[PA-H] ⁺

^a see Table S2 and Table S3 for abbreviations;

2.2.5.4.1. Reserpine (RP)

Table S26: Observed peaks in the P3DDT - RP spectrum and proposed species.

Signal m/z	Origin ^a
39	K ⁺
195	[RP2] ⁺
397	[RP1] ⁺
607	[RP-H] ⁺

^a see Table S2 and Table S3 for abbreviations.

2.2.5.4.2. Tetraethylammonium chloride (TEAC)

Table S27: Observed peaks in the P3DDT - TEAC spectrum and proposed species.

Signal m/z	Origin ^a
86	[TEAC1] ⁺
130	[TEA] ⁺

^a see Table S2 and Table S3 for abbreviations.

Table S28: Observed peaks in the P3DDT - CP spectrum and proposed species.

Signal m/z	Origin ^a
39	K ⁺
43	[C ₃ H ₃] ⁺
57	[C ₄ H ₆] ⁺
363	[CP+H] ⁺

^a see Table S2 and Table S3 for abbreviations;

Table S29: Observed peaks in the P3DDT - ChA spectrum and proposed species.

Signal m/z	Origin ^a
33	HS ⁻
57	[C ₄ H ₆] ⁻
79/81	Br ⁻
407	[ChA-H] ⁻

^a see Table S2 and Table S3 for abbreviations;

2.2.5.4.5. Citric acid (CIA)

Table S30: Observed peaks in the P3DDT - CIA spectrum and proposed species.

Signal m/z	Origin ^a
81	Br ⁻
191	[CIA-H] ⁺

^a see Table S2 and Table S3 for abbreviations;

- S32 -

2.2.5.4.6. L-Phenylalanine (PA)

Table S31: Observed peaks in the P3DDT - PA spectrum and proposed species.

Signal m/z	Origin ^a
33	HS ⁻
57	[C ₄ H ₆] ⁻
79/81	Br ⁻
164	[PA-H] ⁺

^a see Table S2 and Table S3 for abbreviations;

- S33 -

2.2.5.5. Poly{[2,3-bis(3-octyloxyphenyl)quinoxaline-5,8-diyl]-alt-(thiophene-2,5-diyl)} (PTQ1)

Table S32: Qualitative comparison of different layer thickness combinations for positive mode measurements with PTQ1.

c (Matrix) [mg/ml]	c (Analyte) [mg/ml]	Comments about the resulting spectra
10	10	ok
	1	ok
	0.1	very bad
1	10	ok
	1	bad
	0.1	bad
0.1	10	good
	1	ok
	0.1	bad

Table S33: Qualitative comparison of different layer thickness combinations for negative mode measurements with PTQ1.

c (Matrix) [mg/ml]	c (Analyte) [mg/ml]	Comments about the resulting spectra
10	10	ok
	1	bad
	0.1	Analyte not visible
1	10	bad
	1	very bad
	0.1	very bad
0.1	10	ok
	1	bad
	0.1	very bad

Table S34: Observed peaks in the PTQ1 - RP spectrum and proposed species.

Signal m/z	Origin ^a
39	K ⁺
195	[RP2] ⁺
395	[RP3] ⁺
397	[RP1] ⁺
607	[RP-H] ⁺

^a see Table S2 and Table S3 for abbreviations;

2.2.5.2. Tetraethylammonium chloride (TEAC)

Table S35: Observed peaks in the PTQ1 - TEAC spectrum and proposed species.

Signal m/z	Origin ^a
39	K ⁺
130	[TEA] ⁺

^a see Table S2 and Table S3 for abbreviations;

2.2.5.3. Coumaphos (CP)

Table S36: Observed peaks in the PTQ1 - CP spectrum and proposed species.

Signal m/z	Origin ^a
39	K ⁺
43	[C ₆ H] ⁺
363	[CP+H] ⁺
401	[CP+K] ⁺

^a see Table S2 and Table S3 for abbreviations;

Table S37: Observed peaks in the PTQI - ChA spectrum and proposed species.

Signal m/z	Origin ^a
57	[C ₂ H ₅ O] ⁺
407	[ChA-H] ⁺

^a see Table S2 and Table S3 for abbreviations;

Table S38: Observed peaks in the PTQI - CiA spectrum and proposed species.

Signal m/z	Origin ^a
87	[CiA] ⁺
191	[CiA-H] ⁺

^a see Table S2 and Table S3 for abbreviations;

2.2.5.5.6. L-Phenylalanine (PA)

Table S39: Observed peaks in the PTQ1 - PA spectrum and proposed species.

Signal m/z	Origin ^a
164	[PA-H] ⁺

^a see Table S2 and Table S3 for abbreviations.

2.2.5.6. Poly([N,N'-bis(2-octyldodecyl)-isoindigo-5,5'-diyl]-alt-5,5'-(2,2'-bithiophene)) (PII(T2))

Table S40: Qualitative comparison of different layer thickness combinations for positive mode measurements with PII(T2).

c (Matrix) [mg/ml]	c (Analyte) [mg/ml]	Comments about the resulting spectra
10	10	ok
	1	bad
	0.1	very bad
1	10	bad
	1	bad
	0.1	bad
0.1	10	very bad
	1	bad
	0.1	bad

Table S41: Qualitative comparison of different layer thickness combinations for negative mode measurements with PII(T2).

c (Matrix) [mg/ml]	c (Analyte) [mg/ml]	Comments about the resulting spectra
10	10	very bad
	1	Analyte not visible
	0.1	Analyte not visible
1	10	Analyte not visible
	1	very bad
	0.1	Analyte not visible
0.1	10	bad
	1	ok
	0.1	very bad

Table S42: Observed peaks in the PIH(T2) - RP spectrum and proposed species.

Signal m/z	Origin ^a
39	K ⁺
607	[RP-H] ⁺

^a see Table S2 and Table S3 for abbreviations;

Table S43: Observed peaks in the PIH(T2) - TEAC spectrum and proposed species.

Signal m/z	Origin ^a
130	[TEA] ⁺

^a see Table S2 and Table S3 for abbreviations;

2.2.5.6.3. Coumaphos (CP)

Table S44: Observed peaks in the PII(T2) - CP spectrum and proposed species.

Signal m/z	Origin ^a
23	Na ⁺
29	[C ₂ H ₅] ⁺
39	K ⁺
43	[C ₂ H ₇] ⁺
57	[C ₂ H ₉] ⁺
71	[C ₃ H ₁₁] ⁺
85	[C ₄ H ₁₃] ⁺
363	[CP+H] ⁺

^a see Table S2 and Table S3 for abbreviations;

- S44 -

2.2.5.6.4. Cholic acid (ChA)

Table S45: Observed peaks in the PII(T2) - ChA spectrum and proposed species.

Signal m/z	Origin ^a
26	(-) ^b
42	[C ₂ H ₆] ⁻
79/81	Br ⁻
268	(-) ^b
407	[ChA-H] ⁻

^a see Table S2 and Table S3 for abbreviations; ^b clear assignment not possible;

- S45 -

Table S46: Observed peaks in the PIH(T2) - CiA spectrum and proposed species.

Signal m/z	Origin ^a
79/81	Br ⁻
87	[CiA] ⁻
191	[CiA-H] ⁻

^a see Table S2 and Table S3 for abbreviations;

Table S47: Observed peaks in the PIH(T2) - PA spectrum and proposed species.

Signal m/z	Origin ^a
26	(⁻) ^b
35	Cl ⁻
42	[C ₃ H ₆] ⁻
57	[C ₄ H ₆] ⁻
63	(⁻) ^b
79/81	Br ⁻
164	[PA-H] ⁻

^a see Table S2 and Table S3 for abbreviations; ^b clear assignment not possible;

2.2.5.7. Poly(9,9-di-n-octylfluorenyl-2,7-diyl) (P9OFI)

2.2.5.7.1. Reserpine (RP)

Figure S35: Reflectron positive mode MALDI MS spectra of a) sample plate (green), b) only the analyte RP (red), c) only the pure matrix P9OFI (blue) and d) P9OFI coated on RP (black).

Table S48: Observed peaks in the P9OFI - RP spectrum and proposed species. Polymer related peaks are not given.

Signal m/z	Origin ^a
57	[C ₄ H ₉] ⁺
226	(-) ^b
395	[RP] ⁺
607	[RP] ⁺

^a see Table S2 and Table S3 for abbreviations; ^b clear assignment not possible;

2.2.5.7.2. Cholic acid (Cha)

Figure S36: Reflectron negative mode MALDI MS spectra of a) sample plate (green), b) only the analyte Cha (red), c) only the pure matrix P9OFI (blue) and d) P9OFI coated on Cha (black).

Table S49: Observed peaks in the P9OFI - Cha spectrum and proposed species. Polymer related peaks are not given.

Signal m/z	Origin ^a
35	Cl ⁻
63	(-) ^b
79/81	Br ⁻
93	(-) ^b
407	[Cha-H] ⁺

^a see Table S2 and Table S3 for abbreviations; ^b clear assignment not possible;

2.2.6.2. 9-Aminoacridine (9AA)

Figure S39: Reflectron negative mode MALDI MS spectra of a) sample plate (green), b) only the analyte G.B. (red), c) only the pure matrix 9AA (blue) and d) 9AA coated on G.B. (black).

Figure S40: Reflectron negative mode MALDI MS spectra of 9AA coated on G.B.

- S51 -

2.2.6. Complex sample (Gingko Biloba)

2.2.6.1. 2,5-Dihydroxybenzoic acid (2,5-DHB)

Figure S37: Reflectron positive mode MALDI MS spectra of a) sample plate (green), b) only the analyte G.B. (red), c) only the pure matrix 2,5-DHB (blue) and d) 2,5-DHB coated on G.B. (black).

Figure S38: Reflectron positive mode MALDI MS spectra of 2,5-DHB coated on G.B.

- S50 -

2.2.6.3. Poly([N,N'-bis(2-octyldodecyl)-naphthalene-1,4,5,8-bis(dicarboximide)-2,6-diyl]-di-5,5'-[2,2'-bithiophene]) (PNDI(T2))

Figure S41: Reflectron positive mode MALDI MS spectra of a) sample plate (green), b) only the analyte G.B. (red), c) only the pure matrix PNDI(T2) (blue) and d) PNDI(T2) coated on G.B. (black).

Figure S42: Reflectron positive mode MALDI MS spectra of PNDI(T2) coated on G.B.

Figure S43: Reflectron negative mode MALDI MS spectra of a) sample plate (green), b) only the analyte G.B. (red), c) only the pure matrix PNDI(T2) (blue) and d) PNDI(T2) coated on G.B. (black).

2.2.6.4. Poly(3-dodecylthiophene-2,5-diyl) (P3DDT)

Figure S46: Reflectron negative mode MALDI MS spectra of a) sample plate (green), b) only the analyte G.B. (red), c) only the pure matrix P3DDT (blue) and d) P3DDT coated on G.B. (black).

Figure S44: Reflectron positive mode MALDI MS spectra of a) sample plate (green), b) only the analyte G.B. (red), c) only the pure matrix P3DDT (blue) and d) P3DDT coated on G.B. (black).

Figure S45: Reflectron positive mode MALDI MS spectra of P3DDT coated on G.B.

2.2.6.5. Poly([2,2-bis(3-octyloxyphenyl)quinoxaline-5,8-diyl]-*alt*-(thiophene-2,5-diyl)) (PTQ1)

Figure S47: Reflectron positive mode MALDI MS spectra of a) sample plate (green), b) only the analyte G.B. (red), c) only the pure matrix PTQ1 (blue) and d) PTQ1 coated on G.B. (black).

Figure S48: Reflectron positive mode MALDI MS spectra of PTQ1 coated on G.B.

- S56 -

Figure S49: Reflectron negative mode MALDI MS spectra of a) sample plate (green), b) only the analyte G.B. (red), c) only the pure matrix PTQ1 (blue) and d) PTQ1 coated on G.B. (black).

- S57 -

- S58 -

- S59 -

2.2.7. Detection limit

For the determination of the efficiency of the polymer matrices PNDI(T2), P3DDT, PTQ1 and PI(T2), limit of detection measurements with Reserpine (RP) for positive mode and Chloric acid (ChA) for negative mode were carried out. After identifying the lowest measurable analyte solution concentration using the same laser intensity, the measurements were carried out 3 times in total with individually prepared solutions.

2.2.7.1. Poly[*N,N'*-bis(2-octyldodecyl)-naphthalene-1,4,5,8-bis(dicarboximide)-2,6-diyl]-*alt*-5,5'-[4,2'-*bithiophene*] (PNDI(T2))

2.2.7.1.1. Reserpine (RP)

1. Reserpine sample

Figure S53: Reflection positive mode MALDI MS spectra of a) PNDI(T2) coated on RP (0.1 mg/ml) (red) PNDI(T2) coated on RP (0.01 mg/ml) (blue) c) PNDI(T2) coated on RP (0.001 mg/ml) (black). Please note the differences in intensity scaling.

Table S50: Reserpine (RP) limit of detection (LOD) measurements with a PNDI(T2) matrix: Observed peaks the spectra of the first dilution series and proposed assignments.

Spectrum a (0.1 mg/ml RP)		Spectrum b (0.01 mg/ml RP)		Spectrum c (0.001 mg/ml RP)	
Signal m/z	Origin*	Signal m/z	Origin*	Signal m/z	Origin*
29	[C ₅ H ₅] ⁺	39	K ⁺	43	[C ₄ H ₃] ⁺
43	[C ₅ H ₃] ⁺	43	[C ₄ H ₃] ⁺	57	[C ₄ H ₃] ⁺
57	[C ₄ H ₃] ⁺	57	[C ₄ H ₃] ⁺	71	[C ₄ H ₃] ⁺
71	[C ₄ H ₃] ⁺	71	[C ₄ H ₃] ⁺		
85	[C ₄ H ₃] ⁺	395	[RP ₂] ⁺		
195	[RP ₂] ⁺	607	[RP-H] ⁺		
395	[RP ₂] ⁺				
397	[RP] ⁺				
607	[RP-H] ⁺				

* see Table S2 and Table S3 for abbreviations;

3. Reserpine sample

Figure S55: Reflectron positive mode MALDI MS spectra of a) PNDI(T2) coated on RP (0.1 mg/ml) (red) b) PNDI(T2) coated on RP (0.01 mg/ml) (blue) c) PNDI(T2) coated on RP (0.001 mg/ml) (black). Please note the differences in intensity scaling.

Table S52: Reserpine (RP) limit of detection (LOD) measurements with a PNDI(T2) matrix: Observed peaks in the spectra of the third dilution series and proposed assignments.

Spectrum a (0.1 mg/ml RP)		Spectrum b (0.01 mg/ml RP)		Spectrum c (0.001 mg/ml RP)	
Signal m/z	Origin ^a	Signal m/z	Origin ^a	Signal m/z	Origin ^a
43	[C ₃ H ₇] ⁺	29	[C ₂ H ₅] ⁺	39	K ⁺
57	[C ₄ H ₉] ⁺	43	[C ₃ H ₇] ⁺	43	[C ₃ H ₇] ⁺
71	[C ₄ H ₉] ⁺	57	[C ₄ H ₉] ⁺	57	[C ₄ H ₉] ⁺
85	[C ₄ H ₉] ⁺	71	[C ₄ H ₉] ⁺	71	[C ₄ H ₉] ⁺
195	[RP2] ⁺	395	[RP3] ⁺		
395	[RP3] ⁺	397	[RP1] ⁺		
397	[RP1] ⁺	607	[RP-H] ⁺		
607	[RP-H] ⁺				

^a see Table S2 and Table S3 for abbreviations;

2. Reserpine sample

Figure S54: Reflectron positive mode MALDI MS spectra of a) PNDI(T2) coated on RP (0.1 mg/ml) (red) b) PNDI(T2) coated on RP (0.01 mg/ml) (blue) c) PNDI(T2) coated on RP (0.001 mg/ml) (black). Please note the differences in intensity scaling.

Table S51: Reserpine (RP) limit of detection (LOD) measurements with a PNDI(T2) matrix: Observed peaks in the spectra of the second dilution series and proposed assignments.

Spectrum a (0.1 mg/ml RP)		Spectrum b (0.01 mg/ml RP)		Spectrum c (0.001 mg/ml RP)	
Signal m/z	Origin ^a	Signal m/z	Origin ^a	Signal m/z	Origin ^a
43	[C ₃ H ₇] ⁺	29	[C ₂ H ₅] ⁺	39	K ⁺
57	[C ₄ H ₉] ⁺	43	K ⁺	43	[C ₃ H ₇] ⁺
71	[C ₄ H ₉] ⁺	43	[C ₃ H ₇] ⁺	57	[C ₄ H ₉] ⁺
195	[RP2] ⁺	57	[C ₄ H ₉] ⁺	71	[C ₄ H ₉] ⁺
395	[RP3] ⁺	71	[C ₄ H ₉] ⁺		
397	[RP1] ⁺	85	[C ₄ H ₉] ⁺		
607	[RP-H] ⁺	607	[RP-H] ⁺		

^a see Table S2 and Table S3 for abbreviations;

2.2.7.1.2. Cholic Acid (ChA)

1. Cholic acid sample

Figure S56: Reflection negative mode MALDI MS spectra of a) PND(T2) coated on 1. ChA (1 mg/ml) (red) b) PND(T2) coated on 1. ChA (0.1 mg/ml) (blue) c) PND(T2) coated on 1. ChA (0.01 mg/ml) (black). Please note the differences in intensity scaling.

Table S53: Cholic acid (ChA) limit of detection (LOD) measurements with a PND(T2) matrix: Observed peaks in the spectra of the first dilution series and proposed assignments.

Spectrum a (1 mg/ml ChA)		Spectrum b (0.1 mg/ml ChA)		Spectrum c (0.01 mg/ml ChA)	
Signal m/z	Origin ^a	Signal m/z	Origin ^a	Signal m/z	Origin ^a
26	(-) ^b	26	(-) ^b	26	(-) ^b
35	Cl ⁻	35	Cl ⁻	35	Cl ⁻
42	[C ₂₄ H ₄₆ O] ⁻	42	[C ₂₄ H ₄₆ O] ⁻	42	[C ₂₄ H ₄₆ O] ⁻
57	[C ₂₄ H ₄₆ O] ⁻	57	[C ₂₄ H ₄₆ O] ⁻		
79/81	Br ⁻	79/81	Br ⁻		
407	[ChA-H] ⁺	407	[ChA-H] ⁺		

^a see Table S2 and Table S3 for abbreviations; ^b clear assignment not possible.

2. Cholic acid sample

Figure S57: Reflection negative mode MALDI MS spectra of a) PND(T2) coated on 2. ChA (1 mg/ml) (red) b) PND(T2) coated on 2. ChA (0.1 mg/ml) (blue) c) PND(T2) coated on 2. ChA (0.01 mg/ml) (black). Please note the differences in intensity scaling.

Table S54: Cholic acid (ChA) limit of detection (LOD) measurements with a PND(T2) matrix: Observed peaks in the spectra of the second dilution series and proposed assignments.

Spectrum a (1 mg/ml ChA)		Spectrum b (0.1 mg/ml ChA)		Spectrum c (0.01 mg/ml ChA)	
Signal m/z	Origin ^a	Signal m/z	Origin ^a	Signal m/z	Origin ^a
26	(-) ^b	26	(-) ^b	26	(-) ^b
35	Cl ⁻	35	Cl ⁻	42	[C ₂₄ H ₄₆ O] ⁻
42	[C ₂₄ H ₄₆ O] ⁻	42	[C ₂₄ H ₄₆ O] ⁻	48	(-) ^b
57	[C ₂₄ H ₄₆ O] ⁻	57	[C ₂₄ H ₄₆ O] ⁻	60	(-) ^b
79/81	Br ⁻	79/81	Br ⁻	72	(-) ^b
407	[ChA-H] ⁺	407	[ChA-H] ⁺	84	(-) ^b
				96	(-) ^b

^a see Table S2 and Table S3 for abbreviations; ^b clear assignment not possible.

2.2.7.2.2. Poly(3-dodecylthiophene-2,5-diyl) (P3DDT)

2.2.7.2.1. Reserpine (RP)

1. Reserpine sample

Figure S59: Reflectron positive mode MALDI MS spectra of a) P3DDT coated on RP (1 mg/ml) (red) b) P3DDT coated on RP (0.1 mg/ml) (blue) c) P3DDT coated on RP (0.01 mg/ml) (black). Please note the differences in intensity scaling.

Table S56: Reserpine (RP) limit of detection (LOD) measurements with a P3DDT matrix: Observed peaks in the spectra of the first dilution series and proposed assignments.

Spectrum a (1 mg/ml RP)		Spectrum b (0.1 mg/ml RP)		Spectrum c (0.01 mg/ml RP)	
Signal m/z	Origin ^a	Signal m/z	Origin ^a	Signal m/z	Origin ^a
39	K ⁺	39	K ⁺	39	K ⁺
43	[C ₄ H ₇] ⁺	43	[C ₄ H ₇] ⁺	74	K ⁺
57	[C ₄ H ₆] ⁺	57	[C ₄ H ₆] ⁺		
74	(^c) ^b	74	(^c) ^b		
195	[RP2] ⁺	395	[RP3] ⁺		
395	[RP3] ⁺	397	[RP1] ⁺		
397	[RP1] ⁺	607	[RP-H] ⁺		
607	[RP-H] ⁺				

^a see Table S2 and Table S3 for abbreviations; ^b clear assignment not possible;

- S67 -

3. Cholic acid sample

Figure S58: Reflectron negative mode MALDI MS spectra of a) PNDI(T2) coated on 3. ChA (1 mg/ml) (red) b) PNDI(T2) coated on 3. ChA (0.1 mg/ml) (blue) c) PNDI(T2) coated on 3. ChA (0.01 mg/ml) (black). Please note the differences in intensity scaling.

Table S55: Cholic acid (ChA) limit of detection (LOD) measurements with a PNDI(T2) matrix: Observed peaks in the spectra of the third dilution series and proposed assignments.

Spectrum a (1 mg/ml ChA)		Spectrum b (0.1 mg/ml ChA)		Spectrum c (0.01 mg/ml ChA)	
Signal m/z	Origin ^a	Signal m/z	Origin ^a	Signal m/z	Origin ^a
26	(^c) ^b	26	(^c) ^b	42	[C ₄ H ₆] ⁺
35	Cl ⁻	35	Cl ⁻	63	(^c) ^b
42	[C ₄ H ₇] ⁻	42	[C ₄ H ₇] ⁻	79	Br ⁻
57	[C ₄ H ₆] ⁻	57	[C ₄ H ₆] ⁻		
79/81	Br ⁻	79/81	Br ⁻		
407	[ChA-H] ⁺	407	[ChA-H] ⁺		

^a see Table S2 and Table S3 for abbreviations; ^b clear assignment not possible;

- S66 -

2. Reserpine sample

Figure S60: Reflectron positive mode MALDI MS spectra of a) P3DDT coated on RP (1 mg/ml) (red) b) P3DDT coated on RP (0.1 mg/ml) (blue) c) P3DDT coated on RP (0.01 mg/ml) (black). Please note the differences in intensity scaling.

Table S57: Reserpine (RP) limit of detection (LOD) measurements with a P3DDT matrix: Observed peaks in the spectra of the second dilution series and proposed assignments.

Spectrum a (1 mg/ml RP)		Spectrum b (0.1 mg/ml RP)		Spectrum c (0.01 mg/ml RP)	
Signal m/z	Origin ^a	Signal m/z	Origin ^a	Signal m/z	Origin ^a
39	K ⁺	39	K ⁺	39	K ⁺
43	[C ₃ H ₇] ⁺	43	[C ₃ H ₇] ⁺		
57	[C ₃ H ₆] ⁺	57	[C ₃ H ₆] ⁺		
74	(-) ^b	74	(-) ^b		
195	[RP2] ⁺	195	[RP2] ⁺		
395	[RP3] ⁺	397	[RP1] ⁺		
397	[RP1] ⁺	607	[RP-H] ⁺		
607	[RP-H] ⁺				

^a see Table S2 and Table S3 for abbreviations; ^b clear assignment not possible.

3. Reserpine sample

Figure S61: Reflectron positive mode MALDI MS spectra of a) P3DDT coated on RP (1 mg/ml) (red) b) P3DDT coated on RP (0.1 mg/ml) (blue) c) P3DDT coated on RP (0.01 mg/ml) (black). Please note the differences in intensity scaling.

Table S58: Reserpine (RP) limit of detection (LOD) measurements with a P3DDT matrix: Observed peaks in the spectra of the third dilution series and proposed assignments.

Spectrum a (1 mg/ml RP)		Spectrum b (0.1 mg/ml RP)		Spectrum c (0.01 mg/ml RP)	
Signal m/z	Origin ^a	Signal m/z	Origin ^a	Signal m/z	Origin ^a
39	K ⁺	39	K ⁺	39	K ⁺
43	[C ₃ H ₇] ⁺	43	[C ₃ H ₇] ⁺		
74	(-) ^b	57	[C ₃ H ₆] ⁺		
195	[RP2] ⁺	195	[RP2] ⁺		
395	[RP3] ⁺	395	[RP3] ⁺		
397	[RP1] ⁺	397	[RP1] ⁺		
607	[RP-H] ⁺	607	[RP-H] ⁺		

^a see Table S2 and Table S3 for abbreviations; ^b clear assignment not possible.

2.2.7.2.2. Cholic Acid (ChA)

1. Cholic acid sample

Figure S62: Reflectron negative mode MALDI MS spectra of a) P3DDT coated on 1. ChA (1 mg/ml) (red) b) P3DDT coated on 1. ChA (0.1 mg/ml) (blue) c) P3DDT coated on 1. ChA (0.01 mg/ml) (black). Please note the differences in intensity scaling.

Table S59: Cholic acid (ChA) limit of detection (LOD) measurements with a P3DDT matrix: Observed peaks in the spectra of the first dilution series and proposed assignments.

Spectrum a (1 mg/ml ChA)		Spectrum b (0.1 mg/ml ChA)		Spectrum c (0.01 mg/ml ChA)	
Signal m/z	Origin ^a	Signal m/z	Origin ^a	Signal m/z	Origin ^a
33	HS ⁻	33	HS ⁻	33	HS ⁻
57	[C ₂₄ H ₄₉] ⁻	57	[C ₂₄ H ₄₉] ⁻	57	[C ₂₄ H ₄₉] ⁻
407	[ChA-H] ⁻	407	[ChA-H] ⁻	407	[ChA-H] ⁻

^a see Table S2 and Table S3 for abbreviations.

2. Cholic acid sample

Figure S63: Reflectron negative mode MALDI MS spectra of a) P3DDT coated on 1. ChA (1 mg/ml) (red) b) P3DDT coated on 1. ChA (0.1 mg/ml) (blue) c) P3DDT coated on 1. ChA (0.01 mg/ml) (black). Please note the differences in intensity scaling.

Table S60: Cholic acid (ChA) limit of detection (LOD) measurements with a P3DDT matrix: Observed peaks in the spectra of the second dilution series and proposed assignments.

Spectrum a (1 mg/ml ChA)		Spectrum b (0.1 mg/ml ChA)		Spectrum c (0.01 mg/ml ChA)	
Signal m/z	Origin ^a	Signal m/z	Origin ^a	Signal m/z	Origin ^a
33	HS ⁻	33	HS ⁻	33	HS ⁻
57	[C ₂₄ H ₄₉] ⁻	57	[C ₂₄ H ₄₉] ⁻	57	[C ₂₄ H ₄₉] ⁻
79/81	Br ⁻	79/81	Br ⁻	79	Br ⁻
407	[ChA-H] ⁻	407	[ChA-H] ⁻		

^a see Table S2 and Table S3 for abbreviations.

3. Cholic acid sample

Figure S64: Reflectron negative mode MALDI MS spectra of a) P3DDT coated on 1. CHA (1 mg/ml) (red) b) P3DDT coated on 1. CHA (0.1 mg/ml) (blue) c) P3DDT coated on 1. CHA (0.01 mg/ml) (black). Please note the differences in intensity scaling.

Table S61: Cholic acid (Cha) limit of detection (LOD) measurements with a P3DDT matrix: Observed peaks in the spectra of the third dilution series and proposed assignments.

Spectrum a (1 mg/ml Cha)		Spectrum b (0.1 mg/ml Cha)		Spectrum c (0.01 mg/ml Cha)	
Signal m/z	Origin ^a	Signal m/z	Origin ^a	Signal m/z	Origin ^a
33	HS	33	HS	33	HS
57	[C ₁₈ H ₃₅ O] ⁺	57	[C ₁₈ H ₃₅ O] ⁺	57	[C ₁₈ H ₃₅ O] ⁺
79/81	Br ⁻	79/81	Br ⁻	79/81	Br ⁻
407	[Cha-H] ⁺	407	[Cha-H] ⁺		

^a see Table S2 and Table S3 for abbreviations;

- S72 -

2.2.7.3. Poly[2,3-bis(3-octyloxyphenyl)quinoxaline-5,8-diyl]-di-(thiophene-2,5-diyl) (PTQ1)

2.2.7.3.1. Reserpine (RP)

1. Reserpine sample

Figure S65: Reflectron positive mode MALDI MS spectra of a) PTQ1 coated on RP (1 mg/ml) (red) b) PTQ1 coated on RP (0.1 mg/ml) (blue) c) PTQ1 coated on RP (0.01 mg/ml) (black). Please note the differences in intensity scaling.

Table S62: Reserpine (RP) limit of detection (LOD) measurements with a PTQ1 matrix: Observed peaks in the spectra of the first dilution series and proposed assignments.

Spectrum a (1 mg/ml RP)		Spectrum b (0.1 mg/ml RP)		Spectrum c (0.01 mg/ml RP)	
Signal m/z	Origin ^a	Signal m/z	Origin ^a	Signal m/z	Origin ^a
195	[RP2] ⁺	74	(-) ^b	39	K ⁺
395	[RP3] ⁺	395	[RP3] ⁺		
397	[RP1] ⁺	397	[RP1] ⁺	74	(-) ^b
607	[RP-H] ⁺	607	[RP-H] ⁺		

^a see Table S2 and Table S3 for abbreviations; ^b clear assignment not possible;

- S73 -

3. Reserpine sample

Figure S67: Reflectron positive mode MALDI MS spectra of a) PTQ1 coated on RP (1 mg/ml) (red) b) PTQ1 coated on RP (0.1 mg/ml) (blue) c) PTQ1 coated on RP (0.01 mg/ml) (black). Please note the differences in intensity scaling.

Table S64: Reserpine (RP) limit of detection (LOD) measurements with a PTQ1 matrix: Observed peaks in the spectra of the third dilution series and proposed assignments.

Spectrum a (1 mg/ml RP)		Spectrum b (0.1 mg/ml RP)		Spectrum c (0.01 mg/ml RP)	
Signal m/z	Origin ^a	Signal m/z	Origin ^a	Signal m/z	Origin ^a
39	K ⁺	39	K ⁺	43	[C ₃ H ₇] ⁺
74	(-) ^b	43	[C ₃ H ₇] ⁺		
195	[RP2] ⁺	195	[RP2] ⁺		
395	[RP3] ⁺	395	[RP3] ⁺		
397	[RP1] ⁺	397	[RP1] ⁺		
607	[RP-H] ⁺	607	[RP-H] ⁺		
		630	[RP+Na] ⁺		

^a see Table S2 and Table S3 for abbreviations; ^b clear assignment not possible;

2. Reserpine sample

Figure S66: Reflectron positive mode MALDI MS spectra of a) PTQ1 coated on RP (1 mg/ml) (red) b) PTQ1 coated on RP (0.1 mg/ml) (blue) c) PTQ1 coated on RP (0.01 mg/ml) (black). Please note the differences in intensity scaling.

Table S63: Reserpine (RP) limit of detection (LOD) measurements with a PTQ1 matrix: Observed peaks in the spectra of the second dilution series and proposed assignments.

Spectrum a (1 mg/ml RP)		Spectrum b (0.1 mg/ml RP)		Spectrum c (0.01 mg/ml RP)	
Signal m/z	Origin ^a	Signal m/z	Origin ^a	Signal m/z	Origin ^a
39	K ⁺	395	[RP3] ⁺	39	K ⁺
43	[C ₃ H ₇] ⁺	397	[RP1] ⁺	74	(-) ^b
74	(-) ^b	607	[RP-H] ⁺		
195	[RP2] ⁺	630	[RP+Na] ⁺		
395	[RP3] ⁺				
397	[RP1] ⁺				
607	[RP-H] ⁺				

^a see Table S2 and Table S3 for abbreviations; ^b clear assignment not possible;

2.2.7.3.2. Cholic Acid (ChA)

1. Cholic acid sample

Figure S68: Reflectron negative mode MALDI MS spectra of a) PTQ1 coated on L. ChA (1 mg/ml) (red) b) PTQ1 coated on L. ChA (0.1 mg/ml) (blue) c) PTQ1 coated on L. ChA (0.01 mg/ml) (black). Please note the differences in intensity scaling.

Table S65: Cholic acid (ChA) limit of detection (LOD) measurements with a PTQ1 matrix: Observed peaks in the spectra of the first dilution series and proposed assignments.

Spectrum a (1 mg/ml ChA)		Spectrum b (0.1 mg/ml ChA)		Spectrum c (0.01 mg/ml ChA)	
Signal m/z	Origin ^a	Signal m/z	Origin ^a	Signal m/z	Origin ^a
46	(-) ^b	46	(-) ^b	33	HS
407	[ChA-H] ⁺	407	[ChA-H] ⁺		

^a see Table S2 and Table S3 for abbreviations; ^b clear assignment not possible.

2. Cholic acid sample

Figure S69: Reflectron negative mode MALDI MS spectra of a) PTQ1 coated on L. ChA (1 mg/ml) (red) b) PTQ1 coated on L. ChA (0.1 mg/ml) (blue) c) PTQ1 coated on L. ChA (0.01 mg/ml) (black). Please note the differences in intensity scaling.

Table S66: Cholic acid (ChA) limit of detection (LOD) measurements with a PTQ1 matrix: Observed peaks in the spectra of the second dilution series and proposed assignments.

Spectrum a (1 mg/ml ChA)		Spectrum b (0.1 mg/ml ChA)		Spectrum c (0.01 mg/ml ChA)	
Signal m/z	Origin ^a	Signal m/z	Origin ^a	Signal m/z	Origin ^a
46	(-) ^b	46	(-) ^b		
407	[ChA-H] ⁺	57	[C ₁₈ H ₃₃ O] ⁺		
		79	B ⁺		
		407	[ChA-H] ⁺		

^a see Table S2 and Table S3 for abbreviations; ^b clear assignment not possible.

2.2.7.4. Poly[*N,N'*-bis(2-octyldodecyl)-isoidigo-5,5'-diyl]-*adt*-5,5'-(2,2'-bithiophene)] (PH(T2))

2.2.7.4.1. Reserpine (RP)

1. Reserpine sample

Figure S71: Reflectron positive mode MALDI MS spectra of a) PH(T2) coated on RP (1 mg/ml) (red) b) PH(T2) coated on RP (0.1 mg/ml) (blue) c) PH(T2) coated on RP (0.01 mg/ml) (black). Please note the differences in intensity scaling.

Table S68: Reserpine (RP) limit of detection (LOD) measurements with a PH(T2) matrix: Observed peaks in the spectra of the first dilution series and proposed assignments.

Spectrum a (1 mg/ml RP)	Spectrum b (0.1 mg/ml RP)		Spectrum c (0.01 mg/ml RP)	
Signal m/z	Origin ^a	Signal m/z	Origin ^a	Signal m/z
57	[C ₈ H ₉] ⁺	43	[C ₃ H ₇] ⁺	23
607	[RP-H] ⁺	57	[C ₄ H ₉] ⁺	39
		71	[C ₃ H ₇] ⁺	43
		607	[RP-H] ⁺	57
			(-) ^b	69

^a see Table S2 and Table S3 for abbreviations; ^b clear assignment not possible;

3. Cholic acid sample

Figure S70: Reflectron negative mode MALDI MS spectra of a) PTQI coated on L. Cha (1 mg/ml) (red) b) PTQI coated on L. Cha (0.1 mg/ml) (blue) c) PTQI coated on L. Cha (0.01 mg/ml) (black). Please note the differences in intensity scaling.

Table S67: Cholic acid (Cha) limit of detection (LOD) measurements with a PTQI matrix: Observed peaks in the spectra of the third dilution series and proposed assignments.

Spectrum a (1 mg/ml Cha)	Spectrum b (0.1 mg/ml Cha)		Spectrum c (0.01 mg/ml Cha)	
Signal m/z	Origin ^a	Signal m/z	Origin ^a	Signal m/z
46	(-) ^b	46	(-) ^b	
407	[Cha-H] ⁺	407	[Cha-H] ⁺	

^a see Table S2 and Table S3 for abbreviations; ^b clear assignment not possible;

2. Reserpine sample

Figure S72: Reflectron positive mode MALDI MS spectra of a) PII(T2) coated on RP (1 mg/ml) (red) b) PII(T2) coated on RP (0.1 mg/ml) (blue) c) PII(T2) coated on RP (0.01 mg/ml) (black). Please note the differences in intensity scaling.

Table S69: Reserpine (RP) limit of detection (LOD) measurements with a PII(T2) matrix: Observed peaks in the spectra of the second dilution series and proposed assignments.

Spectrum a (1 mg/ml RP)		Spectrum b (0.1 mg/ml RP)		Spectrum c (0.01 mg/ml RP)	
Signal m/z	Origin ^a	Signal m/z	Origin ^a	Signal m/z	Origin ^a
607	[RP-H] ⁺	39	K ⁺	23	Na ⁺
		43	[C ₃ H ₇] ⁺	39	K ⁺
		57	[C ₃ H ₆] ⁺	43	[C ₃ H ₇] ⁺
		71	[C ₃ H ₁₁] ⁺	57	[C ₃ H ₆] ⁺
		85	[C ₃ H ₁₃] ⁺	71	[C ₃ H ₁₁] ⁺
		195	[RP2] ⁺		
		607	[RP-H] ⁺		

^a see Table S2 and Table S3 for abbreviations;

3. Reserpine sample

Figure S73: Reflectron positive mode MALDI MS spectra of a) PII(T2) coated on RP (1 mg/ml) (red) b) PII(T2) coated on RP (0.1 mg/ml) (blue) c) PII(T2) coated on RP (0.01 mg/ml) (black). Please note the differences in intensity scaling.

Table S70: Reserpine (RP) limit of detection (LOD) measurements with a PII(T2) matrix: Observed peaks in the spectra of the third dilution series and proposed assignments.

Spectrum a (1 mg/ml RP)		Spectrum b (0.1 mg/ml RP)		Spectrum c (0.01 mg/ml RP)	
Signal m/z	Origin ^a	Signal m/z	Origin ^a	Signal m/z	Origin ^a
607	[RP-H] ⁺	43	[C ₃ H ₇] ⁺	23	Na ⁺
		57	[C ₃ H ₆] ⁺	29	[C ₃ H ₅] ⁺
		607	[RP-H] ⁺	39	K ⁺
				43	[C ₃ H ₇] ⁺
				57	[C ₃ H ₆] ⁺
				71	[C ₃ H ₁₁] ⁺

^a see Table S2 and Table S3 for abbreviations;

2. Cholic acid sample

Figure S75: Reflectron negative mode MALDI MS spectra of a) PIH(T2) coated on 1. ChA (1 mg/ml) (red) b) PIH(T2) coated on 1. ChA (0.1 mg/ml) (blue) c) PIH(T2) coated on 1. ChA (0.01 mg/ml) (black). Please note the differences in intensity scaling.

Table S72: Cholic acid (ChA) limit of detection (LOD) measurements with a PIH(T2) matrix. Observed peaks in the spectra of the second dilution series and proposed assignments.

Spectrum a (1 mg/ml ChA)		Spectrum b (0.1 mg/ml ChA)		Spectrum c (0.01 mg/ml ChA)	
Signal m/z	Origin ^a	Signal m/z	Origin ^a	Signal m/z	Origin ^a
42	[C ₃ H ₆] ⁺	42	[C ₃ H ₆] ⁺	48	(-) ^b
79/81	Br ⁻	79/81	Br ⁻	60	(-) ^b
407	[ChA-H] ⁺	407	[ChA-H] ⁺	72	(-) ^b
				96	(-) ^b

^a see Table S2 and Table S3 for abbreviations; ^b clear assignment not possible;

2.2.7.4.2. Cholic Acid (ChA)

1. Cholic acid sample

Figure S74: Reflectron negative mode MALDI MS spectra of a) PIH(T2) coated on 1. ChA (1 mg/ml) (red) b) PIH(T2) coated on 1. ChA (0.1 mg/ml) (blue) c) PIH(T2) coated on 1. ChA (0.01 mg/ml) (black). Please note the differences in intensity scaling.

Table S71: Cholic acid (ChA) limit of detection (LOD) measurements with a PIH(T2) matrix. Observed peaks in the spectra of the first dilution series and proposed assignments.

Spectrum a (1 mg/ml ChA)		Spectrum b (0.1 mg/ml ChA)		Spectrum c (0.01 mg/ml ChA)	
Signal m/z	Origin ^a	Signal m/z	Origin ^a	Signal m/z	Origin ^a
42	[C ₃ H ₆] ⁺	26	(-) ^b	26	(-) ^b
79/81	Br ⁻	42	[C ₃ H ₆] ⁺	42	[C ₃ H ₆] ⁺
407	[ChA-H] ⁺	79/81	Br ⁻	79/81	Br ⁻
				407	[ChA-H] ⁺

^a see Table S2 and Table S3 for abbreviations; ^b clear assignment not possible;

3. Cholic acid sample

Figure S76: Reflectron negative mode MALDI MS spectra of a) PII(T2) coated on L. Cha (1 mg/ml) (red) b) PII(T2) coated on L. Cha (0.1 mg/ml) (blue) c) PII(T2) coated on L. Cha (0.01 mg/ml) (black). Please note the differences in intensity scaling.

Table S73: Cholic acid (Cha) limit of detection (LOD) measurements with a PII(T2) matrix: Observed peaks in the spectra of the third dilution series and proposed assignments.

Spectrum a (1 mg/ml Cha)		Spectrum b (0.1 mg/ml Cha)		Spectrum c (0.01 mg/ml Cha)	
Signal m/z	Origin ^a	Signal m/z	Origin ^a	Signal m/z	Origin ^a
42	[C ₁₄ H ₃₀ O] ⁺	79/81	Br ⁻	42	[C ₁₄ H ₃₀ O] ⁺
79/81	[C ₁₄ H ₃₀ O] ⁺	407	[Cha-H] ⁺	48	(-) ^b
407	[Cha-H] ⁺			60	(-) ^b
				72	(-) ^b
				79/81	Br ⁻
				96	(-) ^b

2.2.8. Polymers as analytes

In the following part the spectra of the polymers as analytes with trans-2-[3-(4-tert-butylphenyl)-2-methyl-2-propenylidene]malononitrile (DCTB) as secondary matrix plus 1 vol-% of sodium trifluoroacetate (NaTFA) as additive are shown. The measurements were carried out with a laser intensity of 80%.

Figure S77: Linear positive mode MALDI MS spectra of a) pure DCTB matrix plus 1 vol-% of NaTFA (blue) and b) DCTB matrix plus 1 vol-% of NaTFA coated on PNDI(T2) (black) from m/z = 0 - 1000.

2.2.8.1. PND(T2)

Figure S78: Linear positive mode MALDI MS spectra of a) pure DCTB matrix plus 1 vol-% of NaTFA (blue) and b) DCTB matrix plus 1 vol-% of NaTFA coated on PND(T2) (black) from $m/z = 1k - 150k$.

2.2.8.2. P3DDT

Figure S80: Linear positive mode MALDI MS spectra of a) pure DCTB matrix plus 1 vol-% of NaTFA (blue) and b) DCTB matrix plus 1 vol-% of NaTFA coated on P3DDT (black) from $m/z = 1k - 150k$.

Figure S79: Linear positive mode MALDI MS spectra of a) pure DCTB matrix plus 1 vol-% of NaTFA (blue) and b) DCTB matrix plus 1 vol-% of NaTFA coated on PND(T2) (black) from $m/z = 1k - 14k$.

Figure S81: Linear positive mode MALDI MS spectra of a) pure DCTB matrix plus 1 vol-% of NaTFA (blue) and b) DCTB matrix plus 1 vol-% of NaTFA coated on P3DDT (black) from $m/z = 1k - 14k$.

2.2.8.3. PTQ1

Figure S82: Linear positive mode MALDI MS spectra of a) pure DCTB matrix plus 1 vol-% of NaTFA (blue) and b) DCTB matrix plus 1 vol-% of NaTFA coated on PTQ1 (black) from $m/z = 1k - 150k$.

Figure S83: Linear positive mode MALDI MS spectra of a) pure DCTB matrix plus 1 vol-% of NaTFA (blue) and b) DCTB matrix plus 1 vol-% of NaTFA coated on PTQ1 (black) from $m/z = 1k - 14k$.

- S88 -

2.2.8.4. PIU(72)

Figure S84: Linear positive mode MALDI MS spectra of a) pure DCTB matrix plus 1 vol-% of NaTFA (blue) and b) DCTB matrix plus 1 vol-% of NaTFA coated on PIU(72) (black) from $m/z = 1k - 150k$.

Figure S85: Linear positive mode MALDI MS spectra of a) pure DCTB matrix plus 1 vol-% of NaTFA (blue) and b) DCTB matrix plus 1 vol-% of NaTFA coated on PIU(72) (black) from $m/z = 1k - 14k$.

- S89 -

2.2.9. Comparison spectra

2.2.9.1. Positive matrices

Laser Intensity: 10%
Mode: Reflectron positive
Figure S88: Reflectron positive mode MALDI MS spectra of different pure matrices at 10% laser intensity: a) 2,5-DHB (purple), b) PNDI(T2) (orange), c) P3DDT (green), d) PTQ1 (red), e) PI(T2) (blue) and f) P9OF1 (black). Please note the differences in intensity scaling in spectrum f). The concentration of the polymer solutions was 10 mg/ml.

2.2.8.5. P9OF1

Figure S86: Linear positive mode MALDI MS spectra of a) pure DCTB matrix plus 1 vol-% of NaTFA (blue) and b) DCTB matrix plus 1 vol-% of NaTFA coated on P9OF1 (black) from $m/z = 1k - 150k$.

2.2.9.2. Negative matrices

Table S74: Observed peaks in the positive mixed vs. layered PNDI(T2) comparison spectra and proposed species.

Spectrum a)		Spectrum b)	
Signal m/z	Origin ^a	Signal m/z	Origin ^a
43	[C ₃ H] ⁺	43	[C ₃ H] ⁺
57	[C ₄ H ₆] ⁺	57	[C ₄ H ₆] ⁺
71	[C ₅ H ₁₁] ⁺	71	[C ₅ H ₁₁] ⁺
195	[RP2] ⁺	195	[RP2] ⁺
395	[RP3] ⁺	395	[RP3] ⁺
607	[RP-H] ⁺	607	[RP1] ⁺

^a see Table S2 and Table S3 for abbreviations;

2.2.9.3.2. P3DDT

Table S75: Observed peaks in the positive mixed vs. layered P3DDT comparison spectra and proposed species.

Spectrum a)			Spectrum b)		
Signal m/z	Origin ^a		Signal m/z	Origin ^a	
39	K ⁺		39	K ⁺	
43	[C ₄ H] ⁺		43	[C ₄ H] ⁺	
57	[RP-H] ⁺		57	[C ₄ H ₆] ⁺	
195	[RP ₂] ⁺		195	[RP ₂] ⁺	
395	[RP ₃] ⁺		395	[RP ₃] ⁺	
607	[RP] ⁺		607	[RP] ⁺	

^a see Table S2 and Table S3 for abbreviations.

2.2.9.3.3. PTQ1

Table S76: Observed peaks in the positive mixed vs. layered PTQ1 comparison spectra and proposed species.

Spectrum a)			Spectrum b)		
Signal m/z	Origin ^a		Signal m/z	Origin ^a	
39	K ⁺		39	K ⁺	
43	[C ₄ H] ⁺		43	[C ₄ H] ⁺	
57	[C ₄ H ₆] ⁺		57	[C ₄ H ₆] ⁺	
195	[RP ₂] ⁺		195	[RP ₂] ⁺	
395	[RP ₃] ⁺		395	[RP ₃] ⁺	
607	[RP-H] ⁺		607	[RP-H] ⁺	

^a see Table S2 and Table S3 for abbreviations.

2.2.9.3.4. PIH(T2)

Table S77: Observed peaks in the positive mixed vs. layered PIH(T2) comparison spectra and proposed species.

Spectrum a)		Spectrum b)	
Signal m/z	Origin ^a	Signal m/z	Origin ^a
39	K ⁺	43	[C ₃ H ₇] ⁺
43	[C ₃ H ₇] ⁺	57	[C ₄ H ₉] ⁺
57	[C ₄ H ₉] ⁺	195	[RP2] ⁺
71	[C ₅ H ₁₁] ⁺	607	[RP-H] ⁺
195	[RP2] ⁺		
395	[RP3] ⁺		
607	[RP-H] ⁺		

^a see Table S2 and Table S3 for abbreviations;

2.3. MALDI MS Imaging

Animal tissue used in this study was kindly provided by the animal facility of the Bielefeld University. It represents tissues from animals sacrificed in the context of other studies with no connection to the experiments presented in this manuscript.

Cryo-sections of rat brain (Wistar rat, 12µm of thickness) were sliced using a cryo-microtome (Kryostat CM1850, Leica Biosystem), mounted onto an indium tin oxide coated conductive glass slide (ITO; Bruker Daltonics, Bremen, Germany) and immediately dried in a vacuum desiccator for 24 hours.

The samples were scanned (Plussec optiLab 850, Ahrensburg, Germany) with an image resolution of 7200 dpi. The sections were coated with a P3DDT matrix (5 mg/mL in 50% (v/v) acetonitrile/chloroform) using an ImagePrep instrument (Bruker Daltonics, Bremen, Germany) with the following parameters: 40 cycles, spray power 16%, modulation 20%, spray time 3.35 seconds, incubation time 5 seconds and dry time 14 seconds.

Mass spectrometric analyses were performed in the reflectron negative mode accelerating potential on a time-of-flight mass spectrometer (Bruker ultraflexXtreme; Bruker Daltonics, Bremen, Germany) with the following parameters: ion source 1 - voltage 20.01 kV, ion source 2 - voltage 17.70 kV, reflectron 1 - voltage 21.20 kV, reflectron 2 - voltage 10.80 kV and lens voltage 6.50 kV. The mass spectrometer was equipped with a smartbeam-II laser (Nd:YAG 355 nm) capable of operating at a repetition rate of 1000 Hz with optimized delayed extraction time and laser beam size was set to large. The laser energy was optimized for signal-to-noise in each preparation. Mass spectral data sets were acquired over a whole section using flexImaging™ software (Bruker Daltonics, Bremen, Germany) in the mass range of m/z 0 to 2000, 80 laser shots per spectrum and with a sample rate of 4 GHz. For both the coronal and sagittal sections, a raster step size of 30 µm was used.

For device calibration in the negative mode, 1 µL of red phosphorus (5mg/mL) in acetone was spotted both onto an Anchor Chip 384 MALDI target (Bruker Daltonics, Bremen, Germany) and onto the ITO slides outside the brain sections.

After data acquisition, molecular images were reconstructed using the flexImaging software. Each data set consists of approximately 85834 (sagittal) and 51445 (coronal) individual sampling locations, or pixels. Data sets were normalized using the flexImaging software. For display purposes, image sets were interpolated and pixel intensities were rescaled in flexImaging in order to utilize the entire dynamic range.

3. References

- [1] R. Tkachov, Y. Karpov, V. Senkovskyy, I. Raguzin, J. Zessin, A. Lederer, M. Stamm, B. Voit, T. Beryozkina, V. Bakulev, W. Zhao, A. Facchetti and A. Kiriy, *Macromolecules*, **2014**, *47*, 3845-3851, [10.1021/ma5007667](#).
- [2] Y. Karpov, J. Maiß, R. Tkachov, T. Beryozkina, V. Bakulev, W. Liu, H. Komber, U. Lappan, M. Al-Hussein, M. Stamm, B. Voit and A. Kiriy, *Polym. Chem.*, **2016**, *7*, 2691-2697, [10.1039/c6py00055f](#).
- [3] B. Hong, W. Cheng, J. Wu and C. Zhao, *Chromatographia*, **2010**, *72*, 841-847, [10.1365/s10337-010-1760-1](#).
- [4] A. Holle and R. Ketterlinus, *Innovative smartbeam Laser Technology Enhances MALDI-TOF Based Proteomics Applications*, Technical Note #TN-12, Bruker Daltonics GmbH, **2008**, 1-4.
- [5] K. Sladkova, J. Houska and J. Havel, *Rapid Commun. Mass Spectrom.*, **2009**, *23*, 3114-3118, [10.1002/rcm.4230](#).
- [6] B. Desmazières, W. Buchmann, P. Terrier and J. Tortajada, *Anal. Chem.*, **2008**, *80*, 783-792, [10.1021/ac0715367](#).
- [7] M. Karas and F. Hillenkamp, *Anal. Chem.*, **1988**, *60*, 2299-2301, [10.1021/ac00171a028](#).
- [8] M. A. R. Meier, R. Hoogenboom, M. W. M. Fijten, M. Schneider and U. S. Schubert, *J. Comb. Chem.*, **2003**, *5*, 269-374, [10.1021/cc020101o](#).
- [9] R. L. Vermillion-Salsbury and D. M. Hercules, *Rapid Commun. Mass Spectrom.*, **2002**, *16*, 1575-1581, [10.1002/rcm.750](#).

8.2.2. Amorphous Conjugated Polymers as Efficient Dual-Mode MAL-DI Matrices for Low-Molecular-Weight Analytes (Publication 2)

CHEMPLUSCHEM

Supporting Information

© Copyright Wiley-VCH Verlag GmbH & Co. KGaA, 69451 Weinheim, 2019

Amorphous Conjugated Polymers as Efficient Dual-Mode MALDI Matrices for Low-Molecular-Weight Analytes

Kilian Horatz⁺, Kristina Ditté⁺, Thomas Prenveille, Ke-Nan Zhang, Dieter Jehnichen, Anton Kiriya, Brigitte Voit, and Franziska Lissel* This article is part of a Special Issue on “ π -Conjugated (Macro)molecules and their Applications”.

Table of Content

1. Materials.....	2
1.1. Syntheses.....	4
1.1.1. Synthesis of P(TNDIT-Fl(C ₆ C ₂)).....	4
1.1.2. Synthesis of P(TNDIT-Fl(C ₁₀ C ₈)).....	4
2. Instrumental & Sample Preparation Information.....	5
2.1. NMR.....	5
2.2. UV/VIS.....	5
2.3. XRD.....	5
2.4. Cyclovoltammetry.....	5
2.5. Thermogravimetric analysis (TGA).....	5
2.6. Differential scanning calorimetry (DSC).....	5
2.7. MALDI.....	6
2.7.1. MALDI instrument and measurement conditions.....	6
2.7.2. Calibration.....	6
2.7.3. Sample plate cleaning procedure.....	6
2.7.4. Sample preparation.....	7
2.7.5. Sample measurements.....	Fehler! Textmarke nicht definiert.
3. Analytical measurements.....	8
3.1. NMR.....	8
3.1.1. P(TNDIT-Fl(C ₆ C ₂)).....	8
3.1.2. P(TNDIT-Fl(C ₁₀ C ₈)).....	9
3.2. UV/VIS.....	10
3.3. Cyclovoltammetry (CV).....	11
3.3.1. P(TNDIT-Fl(C ₆ C ₂)).....	11
3.3.2. P(TNDIT-Fl(C ₁₀ C ₈)).....	12
3.4. Thermogravimetric analysis (TGA).....	13
3.5. Differential scanning calorimetry (DSC).....	14
3.6. MALDI MS.....	15
3.6.1. Single Matrix/Analyte Spectra.....	15
3.6.2. Complex sample (Ginkgo Biloba).....	27
3.6.3. Close-up of spin-coated sample measurements.....	29
3.6.4. Polymers as analytes.....	31
4. References	34

1. Materials

The polymer matrices (see Figure S1) poly(4-(5-(9,9-bis(2-ethylhexyl)-9H-fluoren-2-yl)thiophen-2-yl)-2,7-bis(2-octyldodecyl)-9-(thiophen-2-yl)benzofuran-1,3,6,8(2H,7H)-tetraone) (PTNDIT-FI(C₆C₂)) and poly(4-(5-(9,9-bis(2-octyldodecyl)-9H-fluoren-2-yl)thiophen-2-yl)-2,7-bis(2-octyldodecyl)-9-(thiophen-2-yl)benzofuran-1,3,6,8(2H,7H)-tetraone) (PTNDIT-FI(C₆C₃)) were synthesized and purified according to chapter 1.1 and to literature.^{11,21} M_w , M_n and PD values for polymers (see Table S1) were obtained by GPC (40 °C, chloroform) on Agilent 1260 Infinity System with ResiPore columns calibrated with polystyrene (PS) standards.

All solvents were purchased commercially in analytical grade (>99%) and used without further purification. All small molecular analyses listed in Table S2 were purchased commercially (>98%) and used without further purification. As complex sample a *Ginkgo biloba* extract (secondary reference standard) was purchased commercially from Sigma Aldrich (CAS-No.: 90045-36-6, assay bilobalide: 155.19 mg/g; assay ginkgolide: 67.99 mg/g). Table S3 lists observed peaks with their parent analyte and proposed molecular formula.

Table S1: Molecular weight and polydispersity of polymer matrices used in this work.

Polymer matrices	M_n [kg/mol]	M_w [kg/mol]	PD
PTNDIT-FI(C ₆ C ₂)	25.1	42.4	1.7
PTNDIT-FI(C ₆ C ₃)	42.1	93.7	2.2
PNDIT(T2)	28.1 ^[8]	138.9 ^[8]	4.9 ^[3]
P9OFI	29.2 ^[8]	215.6 ^[8]	7.4 ^[3]

*GPC, 40 °C, chloroform

Table S2: Small molecular analyses.

Name	Abbrev.	Molecular formula	Molecular weight [g/mol]	Molecular structure
Reserpine	RP	C ₃₃ H ₄₀ N ₂ O ₉	608	
Tetraethylammonium chloride	TEAC	C ₈ H ₂₀ ClN	185	
Coumarophos	CP	C ₁₄ H ₁₆ ClO ₃ PS	362	
Cholic acid	ChA	C ₂₆ H ₄₆ O ₅	408	
Citric acid	CiA	C ₆ H ₈ O ₇	192	
L-Phenyl-alanine	PA	C ₉ H ₉ NO ₂	165	
Red phosphorus	(-)	P _x	31*x	(-)

Table S3: Proposed fragments of small molecular analyses.^[3]

Fragment	Parent analyte	Proposed molecular formula	Observed m/z
[RP1] ⁺	Reserpine	C ₂₅ H ₃₂ N ₂ O ₄	397
[RP2] ⁺	Reserpine	C ₁₀ H ₁₁ O ₄	195 ^[4]
[RP3] ⁺	Reserpine	(-)	395
[TEA] ⁺	Tetraethylammonium chloride	C ₈ H ₂₀ N	130
[TEAC] ⁺	Tetraethylammonium chloride	C ₈ H ₁₇ N	86
[CiA] ⁺	Citric acid	C ₃ H ₅ O ₃	87
[CiA2] ⁺	Citric acid	C ₄ H ₆ O ₄	129

2. Instrumental & Sample Preparation Information

2.1. NMR

A Bruker Avance III 500 spectrometer was used to record proton ¹H-NMR spectra at 500.13 MHz. The spectra were referenced to the non-deuterated solvent signal of chloroform-d (CDCl₃: δ(H)=7.26 ppm). Abbreviations were used for ¹H-NMR spectra data as follows: s - singlet, d - doublet, dd - doublet of doublet, t - triplet, qua - quartet, qui - quintet, sep - septet, m - multiplet and br - broad signal.

2.2. UV/VIS

UV/VIS measurements were carried out on drop-casted films with a Perkin Elmer "Lambda800" instrument. Parameters and sample preparation method were chosen according to literature.^[3] The recorded UV/VIS spectra are depicted in Figure S6. Table S5 lists the obtained absorption maxima.

2.3. XRD

For the XRD sample preparation the copolymers P(TNDIT-FI(C₁₀C₃)) and P(TNDIT-FI(C₁₀C₆)) were dissolved in toluene (c = 10 mg/ml) and spin-drop-coated on SiO₂-coated silicon wafers. For drop-coating multiple drops (V = 1 µl) of the individual copolymer solutions were dripped on the silicon wafers till they were completely covered and dried for several minutes in air. For spin-coating the wafers were covered with the individual copolymer solutions (V = 100 µl) and spin-coated with the following parameters:

- Speed: 800 rpm
- Acceleration: 50 rpm/s
- Time: 180 sec.

The X-ray scattering curves were recorded using X-ray 2-cycle diffractometer XRD 3003 T/T (GE Sensing & Inspection Technologies GmbH, Ahrensburg/Germany; former: Seifert GmbH) with reflectometry equipment and multilayer monochromator (Cu-K_α radiation, λ = 1.5418 Å). Asymmetric reflection measurements have been made with a small angle of incidence (GIWAXS). The grazing incidence geometry (fix incidence angles α ≤ 0.2°) was combined with a long secondary Soler collimator (angular divergence 0.7°) for higher intensity (of small and thin samples [substrate: Si wafer]). The 2θ measurements were carried out from 3 to 60° with a constant Δ2θ = 0.05°.

2.4. Cyclic voltammetry

CV measurements were performed under nitrogen using a deoxygenated 0.1M solution of tetra-n-butylammonium hexafluorophosphate (NBu₄PF₆) in acetonitrile. The three-electrode setup comprised of silver/silver ion reference electrode, a platinum wire as counter electrode and a platinum disk as a working electrode was utilized. P(TNDIT-FI(C₁₀C₂)) and P(TNDIT-FI(C₁₀C₃)) were dissolved in toluene (c = 0.1 mg/ml) and drop-coated on the working electrode prior to each measurement. All voltammograms were referenced to the ferrocene/ferrocenium (Fc) redox couple as an external standard. HOMO and LUMO energies were calculated with the following equations (1) and (2):

$$E_{HOMO} = -(5.09 + E_{onset,ox.} - \frac{E_{1_{Fc}}}{2})[eV] \quad (1)$$

$$E_{LUMO} = -(5.09 + E_{onset,red.} - \frac{E_{1_{Fc}}}{2})[eV] \quad (2)$$

2.5. Thermogravimetric analysis (TGA)

The TGA and DTG measurements were performed on a TA instrument TGA Q5000 between 40 °C and 800 °C at 10 K/min. under inert (N₂) atmosphere.

2.6. Differential scanning calorimetry (DSC)

The DSC graphs were measured with two different instruments. The first one is a NETZSCH DSC 204 F1 Phoenix 240-1 at 10 K/min between -160 °C and 300 °C. Between the heating and cooling cycles, an isotherm of 6 minutes at -160 °C and an isotherm of 2 minutes were carried out. The other DSC device is a TA instrument DSC Q2000 between -80 °C and 350 °C at 10 K/min. Here, isotherms were carried out for 2 minutes at -80 °C and 350 °C.

1.1. Syntheses

1.1.1. Synthesis of P(TNDIT-FI(C₁₀C₂))

In a round bottom 2-neck flask, 2,2'-(9,9-bis(2-ethylhexyl)-9H-fluorene-2,7-diyl)bis(4,4,5,5-tetramethyl-1,3,2-dioxaborolane) (38.7 mg, 0.0337 mmol) and dibromo-naphthalenedimide bis-thiophene (23.58 mg, 0.0337 mmol) were placed under inert conditions. THF (3 mL) and 2M aqueous solution of Na₂CO₃ (1.35 mL) were added. The solution was degassed for 15 min with a nitrogen flow, immediately after palladium catalyst (1.24 mg, 0.0031 mmol) prepared by mixing of Pd(CH₃CN)₂Cl₂ and P(Bu₃)₃ (1 equivalent of each) was added to the reaction mixture. It was then stirred at room temperature for 4 hours. Afterwards, the organic phase was separated and poured into a mixture of 2M HCl (0.5 mL) and methanol (4 mL) with stirring. The precipitate was collected by filtration and subjected to fractionation in a Soxhlet extractor using methanol, acetone, hexane and chloroform.

1.1.2. Synthesis of P(TNDIT-FI(C₁₀C₃))

In a round-bottom 2-neck flask, 2,2'-(9,9-bis(2-octyldodecyl)-9H-fluorene-2,7-diyl)bis(4,4,5,5-tetramethyl-1,3,2-dioxaborolane) (1.015 g, 1.04 mmol) and dibromo-naphthalenedimide bis-thiophene (1.155 g, 1.01 mmol) were placed under inert conditions. THF (120 mL) and 2M aqueous solution of Na₂CO₃ (42 mL) were added. The solution was degassed for 15 min with a nitrogen flow, immediately after palladium catalyst (45 mg, 0.12 mmol) prepared by mixing of Pd(CH₃CN)₂Cl₂ and P(Bu₃)₃ (1 equivalent of each) was added to the reaction mixture. It was then stirred at room temperature for 48 hours. Afterwards, the organic phase was separated and poured into a mixture of 2M HCl (25 mL) and methanol (200 mL) with stirring. The precipitate was collected by filtration, washed with water and methanol, and dried under reduced pressure overnight to give P(TNDIT-FI(C₁₀C₃)) as a dark-blue solid (1.94 g). The polymer was washed with methanol and subjected to fractionation in a Soxhlet extractor using acetone, hexane and chloroform. Final fractions were extracted with hexane (1.083 g) and chloroform (468 mg). As additional purification step, the polymer was precipitated in ethyl acetate. The precipitate was collected by centrifuging, dissolved in chloroform and poured into methanol. The solid was then filtered to give a final product as a dark-blue solid.

Figure S2: Fiber (left) and thin film (right) sample of the copolymer P(TNDIT-FI(C₁₀C₃))

Drawing of fiber like samples:

The polymer loaded in a round bottom flask was heated at 230 °C in a metal bath. A tweezer was plunged into the melt, as soon as the polymer was completely molten. Afterwards, the tweezer was taken out of the melt and slowly pulled to draw a fiber sample.

Pressing polymer films:

A SPECAC hydraulic press associated with electrically heated platens, preheated to 225 °C, was used to press the polymer films. Between polyimide foils the polymer was placed, to avoid contact with the platens. After closing of the mold, the film was pressed at 225 °C with a pressure of 4.25 tons for 3 minutes. Then the mold was either cooled in a cooling device or directly opened, depending on the desired crystallinity.

- S5 -

- S4 -

2.7. MALDI

2.7.1. MALDI instrument and measurement conditions

All MALDI spectra were recorded with a "Bruker autoflex speed" MALDI-TOF/TOF in reflection mode. The laser of this instrument is a smartbeam-II with a wavelength of 355 nm.^[9] The software for measuring and evaluating the spectra is flexControl 3.4 and flexAnalysis 3.4. All measurements were carried out using a standard set of parameters, which are known to literature.^[9] These parameters are shown in Table S4. All spectra depicted in the paper and the SI are a combination of 10 spectra à 500 shots, resulting in a total shot number of 5000 per spectrum. All 10 measurements per spectrum were measured on different spots of the dried droplet spot and were chosen randomly by the operator. When a spectrum exhibited a good quality (e.g. high S/N-ratio, less impurities, etc.) it was added to the final combination. The best measurements could be obtained from the edges of the dried droplet spot.

Table S4: Parameters for positive and negative reflection & positive linear mode MALDI measurements.

Parameter	Reflection		Linear
	Positive	Negative	Positive
m/z range	0 - 3000	0 - 3000	0 - 200k
Sample rate	2.0 (2.5°)	2.0 (2.5°)	0.5
Shots per measurement	500	500	1000
Matrix suppression	Off	Off	Off
Measurements per spectra	10	10	10
Random walk	Off	Off	Off
Detector gain	100x (51x°)	100x (51x°)	100x
Electronic gain	Enhanced	Enhanced	Enhanced
Realtime smooth	Off	Off	High
Laser-frequency	1000 Hz	1000 Hz	1000 Hz
Laser-size	Large	Large	Large

^afor measurements of spin-coated samples G1A and G1B.

2.7.2. Calibration

2.7.2.1. Low molecular weight area (0-3000 m/z)

At the beginning of each set of low molecular weight MALDI measurements in reflection mode and after changing the measurement polarity, the instrument was calibrated as described in literature.^[3, 6]

2.7.2.2. High molecular weight area (1k - 14k m/z)

At the beginning of each set of high molecular weight MALDI measurements in linear positive mode, the instrument was calibrated as described in literature.^[3, 7]

2.7.3. Sample plate cleaning procedure

The sample plate was cleaned before every new series of measurements using the following procedure, which was slightly changed compared to literature.^[9]

1. Wipe sample plate with wet paper tissue (each once: Chloroform, acetonitrile, acetone);
2. Clean with a sponge/steel-polishing solution "Mir Muscle, Stahl-Fix classic" © SC Johnson;
3. Ultra-sonication for 10 min, at rt in Milli-Q H₂O / acetonitrile (100 ml / 50 ml);
4. Rinse with 50 ml acetonitrile;
5. Ultra-sonication for 10 min, at rt in acetone / tetrahydrofuran (100 ml / 50 ml);
6. Rinse with 50ml tetrahydrofuran;
7. Rinse with 50ml acetone;
8. Ultra-sonication for 10 min, at rt in chloroform / ethanol (100 ml / 50 ml);
9. Rinse with 50 ml ethanol;
10. Rinse with 50 ml chloroform.

2.7.4. Sample preparation

2.7.4.1. Dried-droplet method

The samples were prepared with the dried-droplet method on a standard sample plate (Bruker "MTP 384 target plate ground steel BC" Part-No.: 8280784).^[8,9] The polymer matrices P(TNDIT-Fl(C₆C₂)) and P(TNDIT-Fl(C₆C₂)) were dissolved in toluene. The analytes cholic acid (ChA), citric acid (CA), reserpine (RP), coumarins (CP) and Ginkgo Biloba (G.B.) were dissolved in tetrahydrofuran. L-Phenylalanine (PA) was dissolved in ethanol, tetraethylammonium chloride (TEAC) in acetonitrile. First the individual analyte solutions were dripped on the plate (V = 0.5 µl) and dried in air for several seconds to several minutes until the solvent had completely evaporated. Then the individual matrices were coated by dripping the corresponding solutions (V = 0.5 µl) on top of the analytes and dry them in air for around 1 - 2 minutes.^[9]

2.7.4.2. Spin-coating method

The samples were prepared on ITO-coated glass slides from Bruker (Part-No.: 8237001), which were clamped on a "MTP TLC Adapter" (Part-No.: 8255595) from Bruker for the measurements. All polymer matrices were dissolved in toluene, whereas reserpine was dissolved in acetone and citric acid and Ginkgo Biloba were dissolved in tetrahydrofuran. All solutions were filtered with a 1 µm PTFE filter. For the spin-coating process of the analytes and the matrices the following parameters were used:

- Speed:
 - Analyte: 500 rpm
 - Matrix: 500 rpm
- Acceleration:
 - Analyte: 50 rpm/s
 - Matrix: 50 rpm/s
- Time:
 - Analyte: 60 s
 - Matrix: 60 s
- Solution:
 - Analyte: V = 100 µl with c = 10 mg/ml
 - Matrix: V = 150 µl with c = 10 mg/ml

Figure S3: ITO-coated glass slides with the polymer matrices coated on reserpine via spin-coating: PNDI(12) (left), P(TNDIT-Fl(C₆C₂)) (middle), P(TNDIT-Fl(C₆C₂)) (right).

3. Analytical measurements

3.1. NMR

3.1.1. $P(TNDIT-F(C_4C_2))$

- S8 -

3.1.2. $P(TNDIT-F(C_{10}C_8))$

- S9 -

Figure S6: UV/VIS spectra of the polymer matrices a) PTNDIT-FI(C₄C₂) (red), b) PTNDIT-FI(C₆C₈) (green), c) PNDI(T2) (black, dashed)^[9] and d) P9OFI (blue, dashed)^[9].

Table S5: UV/VIS absorption maxima of the polymer matrices.

Polymer	PTNDIT-FI(C ₄ C ₂)	PTNDIT-FI(C ₆ C ₈)	PNDI(T2) ^[9]	P9OFI ^[9]
Absorption maxima [nm] ^a	389 (100%) 633 (51%)	389 (100%) 635 (46%)	392 (100%) 697 (81%)	386 (100%) 432 (13%)
Relative absorption at 355 nm [%] ^a	73%	67%	82%	72%

^a relative to the global absorption maximum of each compound.

Figure S7: Cyclic voltammogram of PTNDIT-FI(C₄C₂). Left: reduction curve (blue); Right: oxidation curve (red); fitting lines for E_{onset} calculations (green).

Table S6: Parameters of the copolymer PTNDIT-FI(C₄C₂).

Parameter	Value
$E_{onset,ox}$	1.12 V
$E_{onset,red}$	-0.91 V
HOMO	-6.09 eV
LUMO	-4.06 eV

Figure S9: TGA curve for the copolymer P(TNDIT-FI(C₁₀C₈)).

Figure S8: Cyclic voltammogram of P(TNDIT-FI(C₁₀C₈)). Left: reduction curve (blue); Right: oxidation curve (red); fitting lines for E_{onset} calculations (green).

Table S7: Parameters of the copolymer P(TNDIT-FI(C₁₀C₈)).

Parameter	Value
E _{onset,ox}	1.16 V
E _{onset,red}	-0.93 V
HOMO	-6.13 eV
LUMO	-4.03 eV

3.5. Differential scanning calorimetry (DSC)

Figure S10: DSC curve of the copolymer P(TNDIT-Fl(C₆₀C₆)) from -80 °C - 350 °C.

Table S8: Values for several parameters of the copolymer P(TNDIT-Fl(C₆₀C₆)) determined with DSC.

Parameter	Value
T _g	38 °C
T _m	-

- S15 -

Signal m/z	Origin
69	[C ₆ H ₇] ⁺
696	[RP-H] ⁺
697	[RP-1] ⁺
691	[RP2] ⁺
75	[C ₆ H ₉] ⁺
87	[C ₆ H ₇] ⁺

Table S9: Observed peaks in the P(TNDIT-Fl(C₆C₂)) - RP spectrum and proposed species.

3.6. MALDI MS
3.6.1. Single Matrix/Analyte Spectra
3.6.1.1. P(TNDIT-Fl(C₆C₂))
3.6.1.1.1. Reserpine (RP)

Table S11: Observed peaks in the $P(TNDIT-FI(C_4C_2))$ - CP spectrum and proposed species.

Signal m/z	Origin
43	$[C_3H_7]^+$
57	$[C_4H_9]^+$
71	$[C_5H_{11}]^+$
307	(-)
363	$[CP+H]^+$

- S17 -

Table S10: Observed peaks in the $P(TNDIT-FI(C_4C_2))$ - TEAC spectrum and proposed species.

Signal m/z	Origin
130	$[TEA]^+$

- S16 -

Table S12: Observed peaks in the PTNDIT-FI(C₆C₂) - ChA spectrum and proposed species.

Signal m/z	Origin
42	[C ₆ H ₆] ⁺
81	Br
407	[ChA-H] ⁺

- S18 -

Table S13: Observed peaks in the PTNDIT-FI(C₆C₂) - CItA spectrum and proposed species.

Signal m/z	Origin
42	[C ₆ H ₆] ⁺
87	[CItA] ⁺
191	[CItA-H] ⁺

- S19 -

Table S15: Observed peaks in the PII(T2) - RP spectrum and proposed species.

Signal m/z	Origin
43	[C ₃ H ₇] ⁺
57	[C ₄ H ₉] ⁺
195	[RP2] ⁺
397	[RP1] ⁺
607	[RP-H] ⁺

- S21 -

Table S14: Observed peaks in the P(TNDIT-FI(C₄C₂)) - PA spectrum and proposed species.

Signal m/z	Origin
42	[C ₃ H ₆] ⁻
79/81	Br ⁻
164	[PA-H] ⁻

- S20 -

Table S16: Observed peaks in the $P(TNDIT-FI(C_{10}C_8))$ - TEAC spectrum and proposed species.

Signal m/z	Origin
86	$[TEAC]^+$
130	$[TEA]^+$

Table S17: Observed peaks in the $P(TNDIT-FI(C_{10}C_8))$ - CP spectrum and proposed species.

Signal m/z	Origin
39	K^+
43	$[C_3H_7]^+$
57	$[C_4H_9]^+$
71	$[C_5H_{11}]^+$
85	$[C_6H_{13}]^+$
307	$(-)$
338	$(-)$
363	$[CP+H]^+$

Table S19: Observed peaks in the $P(TNDIT-FI(C_{10}C_8))$ - CiA spectrum and proposed species.

Signal m/z	Origin
42	$[C_3H_6]^+$
87	$[CiA]^+$
191	$[CiA-H]^+$

- S25 -

Table S18: Observed peaks in the $P(TNDIT-FI(C_{10}C_8))$ - ChA spectrum and proposed species.

Signal m/z	Origin
42	$[C_3H_6]^+$
79.81	Br
407	$[ChA-H]^+$

- S24 -

Figure S22: Reflection negative mode MALDI MS spectra of a) sample plate (green), b) only the analyte PA (red), c) only the pure matrix P(TNDIT-FI(C₁₀C₃)) (blue) and d) P(TNDIT-FI(C₁₀C₃)) coated on PA (black).

Table S20: Observed peaks in the P(TNDIT-FI(C₁₀C₃)) - PA spectrum and proposed species.

Signal m/z	Origin
42	[C ₃ H ₆] ⁺
79.81	Br ⁻
164	[PA+H] ⁺

- S26 -

Figure S23: Reflection positive mode MALDI MS spectra of a) sample plate (green), b) only the analyte G.B. (red), c) only the pure matrix P(TNDIT-FI(C₄C₂)) (blue) and d) P(TNDIT-FI(C₄C₂)) coated on G.B. (black).

Figure S24: Reflection negative mode MALDI MS spectra of a) sample plate (green), b) only the analyte G.B. (red), c) only the pure matrix P(TNDIT-FI(C₄C₂)) (blue) and d) P(TNDIT-FI(C₄C₂)) coated on G.B. (black).

- S27 -

3.6.3. Close-up of spin-coated sample measurements

Figure S27: Close-up of reflection positive mode MALDI MS spectra of spin coated samples: (a) PND(T2) coated on RP (red), (b) P(TNDIT-FI(C₄C₂)) coated on RP (blue) and (c) P(TNDIT-FI(C₁₀C₈)) coated on RP (black). Peaks: m/z = 605 [unknown], 606 [unknown], 607 [RP-H⁺], 608 [RP⁺], 609 [RP+H⁺], 610 [unknown].

- S29 -

Figure S25: Reflection positive mode MALDI MS spectra of a) sample plate (green), b) only the analyte G.B. (red), c) only the pure matrix P(TNDIT-FI(C₁₀C₈)) (blue) and d) P(TNDIT-FI(C₁₀C₈)) coated on G.B. (black).

Figure S26: Reflection negative mode MALDI MS spectra of a) sample plate (green), b) only the analyte G.B. (red), c) only the pure matrix P(TNDIT-FI(C₁₀C₈)) (blue) and d) P(TNDIT-FI(C₁₀C₈)) coated on G.B. (black).

- S28 -

3.6.4. Polymers as analytes

In the following part the baseline corrected spectra of the polymers as analytes with trans-2-(3-(4-tert-butylphenyl)-2-methyl-2-propenylidene)malonitrile (DCTB) as secondary matrix plus 1 vol.-% of sodium trifluoroacetate (NaTFA) as additive are shown. The measurements were carried out with a laser intensity of 80%.

Figure S31: Linear positive mode MALDI MS spectra of a) pure DCTB (red), b) DCTB mixed with P(TNDIT-FI(C₄C₂)) (10:1) (blue) and c) pure P(TNDIT-FI(C₄C₂)) (black) from m/z = 1k - 200k.

Figure S32: Linear positive mode MALDI MS spectra of a) pure DCTB (red) and b) DCTB mixed with P(TNDIT-FI(C₄C₂)) (10:1) (blue) and c) pure P(TNDIT-FI(C₄C₂)) (black) from m/z = 1k - 20k.

Figure S33: Linear positive mode MALDI MS spectra of a) pure DCTB (red), b) DCTB mixed with P(TNDIT-FI(C₁₀C₈)) (10:1) (blue) and c) pure P(TNDIT-FI(C₁₀C₈)) (black) from m/z = 1k - 200k.

Figure S34: Linear positive mode MALDI MS spectra of a) pure DCTB (red) and b) DCTB mixed with P(TNDIT-FI(C₁₀C₈)) (10:1) (blue) and c) pure P(TNDIT-FI(C₁₀C₈)) (black) from m/z = 1k - 20k.

4. References

- [1] R. Tachov, H. Komber, S. Rauch, A. Lederer, U. Oertel, L. Häußler, B. Voit and A. Kiry, *Macromolecules*, **2014**, *47*, 4994-5001, [10.1021/ma5006588](#).
- [2] T. Prenville, *Synthesis and characterization of melt-processable semiconductive polymers*, 2015.
- [3] K. Horatz, M. Giampà, Y. Karpov, K. Sahre, H. Bednarz, A. Kiry, B. Voit, K. Niehaus, N. Hadjichristidis, D. L. Michels and F. Lissel, *J. Am. Chem. Soc.*, **2018**, *140*, 11416–11423, [10.1021/acs.8b06637](#).
- [4] B. Hong, W. Cheng, J. Wu and C. Zhao, *Chromatographia*, **2010**, *72*, 841-847, [10.1365/s10337-010-1760-1](#).
- [5] A. Holle and R. Ketterlinus, *Innovative smartbeam Laser Technology Enhances MALDI-TOF Based Proteomics Applications*, Technical Note #TN-12, Bruker Daltonics GmbH, **2008**, 1-4.
- [6] K. Sladkova, J. Houska and J. Havel, *Rapid Commun. Mass Spectrom.*, **2009**, *23*, 3114-3118, [10.1002/rcm.4230](#).
- [7] B. Desmazières, W. Buchmann, P. Ternier and J. Tortajada, *Anal. Chem.*, **2008**, *80*, 783-792, [10.1021/ac0715367](#).
- [8] M. Karas and F. Hillenkamp, *Anal. Chem.*, **1988**, *60*, 2299-2301, [10.1021/ac00171a028](#).
- [9] M. A. R. Meier, R. Hoogenboom, M. W. M. Fijlen, M. Schneider and U. S. Schubert, *J. Comb. Chem.*, **2003**, *5*, 269-374, [10.1021/cc020101o](#).

8.2.3. Polymerization as a Strategy to Improve Small Organic Matrices for Low Molecular Weight Compound Analytics with MALDI MS and MALDI MS Imaging (Publication 3)

Supporting Information

Polymerization as a Strategy to Improve Small Organic Matrices for Low Molecular Weight Compound Analytics with MALDI MS and MALDI MS Imaging

Kilian Horatz^{†‡}, Marco Giampà[§], Zhi Qiao^{†‡}, Siver A. Moestue[§] and Franziska Lissel^{†‡*}

[†]Institute of Macromolecular Chemistry, Leibniz-Institut für Polymerforschung Dresden e.V., Dresden, Saxony, 01069, Germany.

[‡]Faculty of Chemistry and Food Chemistry, Technische Universität Dresden, Dresden, Saxony, 01062, Germany.

[§]MR Cancer Group, Department of Clinical and Molecular Medicine, Norwegian University of Science and Technology, Trondheim, 7030, Norway.

[§]Department of Pharmacy, Nord University, Bodø, 8026, Norway.

[†]Institute of Organic Chemistry and Macromolecular Chemistry, Friedrich Schiller University Jena, Jena, Thuringia, 07743, Germany.

*lissel@ipfdd.de

Table of Contents

1. Materials	2
1.1. Syntheses	4
2. Instrumental & Sample Preparation Information	7
2.1. Nuclear magnetic resonance (NMR)	7
2.2. UV/VIS	7
2.3. Vacuum stability	7
2.4. Fourier-transform infrared spectroscopy (FTIR)	7
2.5. Gel permeation chromatography (GPC)	8
2.6. MALDI MS	8
2.7. MALDI Imaging	9
3. Analytical measurements	10
3.1. NMR	10
3.2. UV/VIS	17
3.3. Vacuum stability	18
3.4. FTIR	19
3.5. GPC	20
3.6. MALDI MS	22
4. References	58

1. Materials

The polymer matrices poly(vinyl dihydroxybenzoic acid) (PVDHB) and poly(vinyl harmine) (P(VHar)) were synthesized and purified according to subchapter 1.1 and the literature cited therein.

The small organic matrices (SOMs) 2,5-dihydroxybenzoic acid (DHB), 9-aminoacridine (9AA) and 7-methoxy-1-methyl-9H-pyrido[3,4-b]indole (harmine) were purchased from Sigma Aldrich. DHB was purchased in MALDI-suitable purity, whereas harmine was recrystallized from refluxing MeOH before being used as MALDI matrix.

Unless noted otherwise, solvents were purchased in analytical grade (>99%) and used without further purification. All analytes listed in Table S1 were purchased (with >98% purity) and used without further purification. Table S2 lists observed peaks with their parent analyte and proposed molecular formula.

Table S1: Molecular formula, weight, and structure of the investigated analytes.

Name	Abbrev.	Molecular formula	Molecular weight [g/mol]	Molecular structure
Saccharides				
Fructose	FT	C ₆ H ₁₂ O ₆	180
Glucuronic acid	GA	C ₆ H ₁₀ O ₇	194
Sucrose	SR	C ₁₂ H ₂₂ O ₁₁	342
Maltotriose	MT	C ₁₈ H ₃₂ O ₁₆	504
α-Cyclodextrin	α-CD	C ₃₆ H ₆₆ O ₃₀	972
Lipids				
Myristic acid	MA	C ₁₄ H ₂₈ O ₂	228
Linoleic acid	LA	C ₁₈ H ₃₂ O ₂	280
Docosahexanoic acid	DA	C ₂₂ H ₄₂ O ₂	328
Dipalmitoyl-Glycerol	DG	C ₅₅ H ₁₀₈ O ₅	568

Dipalmitoyl-Glycerophosphocholine	DGPC	C ₄₀ H ₈₀ NO ₈ P	734	
Choline-derivatives				
Choline chloride	CC	C ₅ H ₁₄ ClNO	139 (104 ⁺)	
Phosphocholine	PC	C ₅ H ₁₄ NO ₄ P	183	
Glycerophosphorylcholine	GIPC	C ₈ H ₂₀ NO ₆ P	257	
Amino acids				
Methionine	ME	C ₅ H ₁₁ NO ₂ S	149	
Phenylalanine	PA	C ₉ H ₉ NO ₂	165	

Table S2: Proposed fragments of the analyses.

Fragment	Parent analyte	Proposed molecular formula	Observed m/z [g/mol]
[FT1] ⁺	Fructose (FT)	C ₆ H ₁₂ O ₂	71
[FT2] ⁺	Fructose (FT)	C ₆ H ₁₂ O ₃	101
[FT3] ⁺	Fructose (FT)	C ₆ H ₁₂ O ₄	119
[FT4] ⁺	Fructose (FT)	C ₆ H ₁₂ O ₃	130
[GA1] ⁺	Glucuronic acid (GA)	C ₆ H ₁₀ O ₂	58
[GA2] ⁺	Glucuronic acid (GA)	C ₆ H ₁₀ O ₂	71
[GA3] ⁺	Glucuronic acid (GA)	C ₆ H ₁₀ O ₃	87
[GA4] ⁺	Glucuronic acid (GA)	C ₆ H ₁₀ O ₄	130
[SR1] ⁺	Sucrose (SR)	C ₁₂ H ₂₂ O	57
[SR2] ⁺	Sucrose (SR)	C ₁₂ H ₂₂ O ₂	71
[SR3] ⁺	Sucrose (SR)	C ₁₂ H ₂₂ O ₃	203
[MT1] ⁺	Maltotriose (MT)	C ₃₄ H ₆₂ O ₂	71
[MT2] ⁺	Maltotriose (MT)	C ₃₄ H ₆₂ O ₁₁	341
[MA1] ⁺	Myristic acid (MA)	C ₁₄ H ₂₆	57
[MA2] ⁺	Myristic acid (MA)	C ₁₄ H ₂₆ O ₂	88
[MA3] ⁺	Myristic acid (MA)	C ₁₄ H ₂₆ O ₃	722
[MA4] ⁺	Myristic acid (MA)	C ₁₄ H ₂₆ O ₆	737
[LA1] ⁺	Linoleic acid (LA)	C ₁₈ H ₃₂ O ₆	877/878/879
[LA2] ⁺	Linoleic acid (LA)	C ₁₈ H ₃₂ O ₈	893/894
[DG1] ⁺	Dipalmitoyl-Glycerol (DG)	C ₃₆ H ₇₀ O ₂	255
[DG2] ⁺	Dipalmitoyl-Glycerol (DG)	C ₃₆ H ₇₀ O ₃	295
[DGPC1] ⁺	Dipalmitoyl-Glycerol-Phosphocholine (DGPC)	C ₃₆ H ₇₀ O ₂	255
[CC1] ⁺	Choline chloride (CC)	C ₅ H ₁₂ N	86

- S3 -

1.1. Syntheses

1.1.1. Poly(2,5-dihydroxybenzoic acid) (P(VDHB))

1.1.1.1. 4-Bromo-2,5-dimethoxybenzaldehyde (2)

A 2-necked round bottom flask equipped with a reflux condenser was evacuated, heated, and flooded with N₂ three times. The flask was charged with 1,4-dibromo-2,5-dimethoxybenzene (**1**) (2 g, 6.8 mmol, 1 eq.) and again evacuated and flooded thrice with N₂. Then, dry diethylether (Et₂O, 68 ml) was added and the solution cooled to 0 °C. Afterwards, *n*-butyllithium (4.43 ml, 1.6 mol/l, 7.1 mmol, 1.05 eq.) was added dropwise over 10 min. The solution was stirred at 0 °C for 15 min and dry dimethylformamide (DMF, 0.75 ml, 10.1 mmol, 1.5 eq.) was added dropwise. The reaction was stirred at 0 °C for 15 min and refluxed for 15 min at 45 °C. The reaction was poured into 10% HCl (aq.), extracted with dichloromethane (DCM), dried over Na₂SO₄, filtered, and concentrated to a yellow solid. Purification by silica gel chromatography (hexane/EtOAc = 10:1) afforded the product as a pale-yellow solid (R_f ≈ 0.37). Recrystallization in *i*-PrOH resulted in pale-yellow crystals (Yield: 53%).¹

¹H-NMR (DMSO-*d*₆, 500 MHz) (see Figure S11):

δ [ppm] = 3.83 (s, 3H, H-8), 3.89 (s, 3H, H-9), 7.25 (s, 1H, H-3), 7.51 (s, 1H, H-6), 10.27 (s, 1H, H-7).

¹³C-NMR (DMSO-*d*₆, 125 MHz) (see Figure S12):

δ [ppm] = 56.52 (C-8), 56.75 (C-9), 109.34 (C-6), 118.26 (C-4), 119.54 (C-3), 123.67 (C-1), 149.73 (C-5), 155.8 (C-2), 188.14 (C-7).

1.1.1.2. 4-Bromo-2,5-dimethoxybenzoic acid (3)

To a solution of **2** (1 g, 4.1 mmol, 1 eq.) in acetone (58 ml), a solution of potassium permanganate (1.29 g, 8.2 mmol, 2 eq.) in water (41 ml) was added. The reaction was stirred overnight at room temperature. Afterwards, a solution of potassium hydroxide (0.46 g, 8.2 mmol, 2 eq.) in water (8 ml) was added and stirred for 30 min. After concentrating in vacuo, the crude, black residue was dispersed in DCM/MeOH (1:1) and filtered over celite (prewashed with DCM/MeOH (1:1)). The resulting white solid was dispersed in ethyl acetate and filtered. The resulting white solid was dispersed in 1M HCl, extracted with ethyl acetate, dried over Na₂SO₄, and concentrated in vacuo, resulting in the product as white solid (Yield: 79%). This synthesis was performed as variation from literature.²

¹H-NMR (DMSO-*d*₆, 500 MHz) (see Figure S13):

δ [ppm] = 3.79 (s, 3H, H-8), 3.82 (s, 3H, H-10), 7.32 (s, 1H, H-3), 7.37 (s, 1H, H-6), 12.82 (br, 1H, H-8).

¹³C-NMR (DMSO-*d*₆, 125 MHz) (see Figure S14):

δ [ppm] = 56.60 (C-9), 56.71 (C-10), 113.93 (C-6), 114.89 (C-4), 117.96 (C-3), 121.13 (C-1), 149.04 (C-5), 152.37 (C-2), 166.43 (C-7).

1.1.1.3. 4-Bromo-2,5-dimethoxybenzoate (4)

3 (1 g, 3.8 mmol, 1 eq.) was dissolved in MeOH in the presence of concentrated sulfuric acid (0.3 ml, 5.7 mmol, 1.5 eq.) and stirred at 80 °C overnight. After cooling to room temperature (rt) the MeOH was evaporated completely, and the crude residue was transferred to a separating funnel with ethyl acetate. The organic phase was extracted with 1M KOH (aq.) three times and then was dried over Na₂SO₄. Removing the solvent in vacuo gave the product as white solid (yield: 95%). This synthesis was performed as variation from literature.³

¹H-NMR (DMSO-*d*₆, 500 MHz) (see Figure S15):

δ [ppm] = 3.79 (s, 3H, H-8), 3.80 (s, 3H, H-9), 3.82 (s, 3H, H-10), 7.40 (s, 1H, H-6), 7.32 (s, 1H, H-3).

¹³C-NMR (DMSO-*d*₆, 125 MHz) (see Figure S16):

δ [ppm] = 52.06 (C-8), 56.65 (C-9), 56.73 (C-10), 113.85 (C-6), 115.59 (C-4), 118.04 (C-3), 119.75 (C-1), 149.08 (C-5), 152.42 (C-2), 165.41 (C-7).

Figure S1: Molecular structure of **2**.

Figure S2: Molecular structure of **3**.

Figure S3: Molecular structure of **4**.

- S4 -

1.1.1.1.7. Poly(2,5-dihydroxy-4-vinylbenzoic acid) (P(VDHB))

A 1-neck round bottom flask was charged with **P(VDMB)** (0.05 g) in acetic acid (4.44 ml). Then hydrobromic acid (48% in H₂O) (9.61 ml) was added and the reaction mixture was stirred for 5 h at 120 °C. After cooling to rt, the solvent was completely removed in vacuo. The crude residue was dissolved in MeOH and precipitated two times in chloroform / Et₂O (2:1 vol-ratio) resulting in the product as brownish solid (yield: 80 wt. %). This synthesis was performed as variation from literature.⁷

¹H-NMR (DMSO-d₆, 500 MHz) (see Figure S24):
 δ [ppm] = 1.10-1.69 (br, 2H, H-11), 1.76-3.47 (br, 1H, H-10), 5.08-6.67 (br, 1H, H-1), 6.67-7.43 (br, 1H, H-6), 7.58-9.07 (br, 1H, H-12), 9.43-11.35 (br, 1H, H-9), 12.12-13.61 (br, 1H, H-8).

Figure S7: Molecular structure of **P(VDHB)**.

1.1.1.2. Poly(7-methoxy-1-methyl-9H-pyrido[3,4-b]indole) (P(VHAr))

1.1.2.1. 8-Bromo-7-methoxy-1-methyl-9H-pyrido[3,4-b]indole (7)

To a solution of 7-methoxy-1-methyl-9H-pyrido[3,4-b]indole (harmine) (0.7 g, 3.3 mmol, 1 eq.) in DCM (280 ml), a solution of *N*-bromosuccinimide (0.56 g, 3.1 mmol, 0.95 eq.) in DCM (60 ml) was added dropwise over 1 h. The reaction was stirred overnight in the absence of light at rt under air. The crude mixture was washed with 0.1M NaOH (aq.) and then with water. The organic layer was dried over Na₂SO₄, filtered, and evaporated in vacuo to give a pale-yellow solid residue. The residue was separated by column chromatography with pure ethyl acetate as eluent to give the product as white solid (*R*_f ≈ 0.29; yield: 36%).⁸

Figure S8: Molecular structure of **7**.

1.1.2.2. 8-Vinyl-7-methoxy-1-methyl-9H-pyrido[3,4-b]indole (8)

A 2-neck round bottom flask was equipped with a septum and a reflux condenser and evacuated, heated, and flooded with N₂ three times. **7** (0.5 g, 1.7 mmol, 1 eq.) was dissolved in 1,4-dioxane (16 ml) and sodium carbonate (0.546 g, 5.2 mmol, 3 eq.) was dissolved in water (2.15 ml). Both mixtures were combined with vinylboronic acid pinacol ester (0.44 ml, 2.6 mmol, 1.5 eq.) and bubbled with N₂ for 30 min. Afterwards, tetrakis(triphenylphosphine)palladium(0) (0.0992 g, 0.0992 g, 0.1 mmol, 0.05 eq.) was added. After stirring at 100 °C for 24 h, the solution was cooled to rt and transferred to a separation funnel with water and ethyl acetate. The crude product was extracted with ethyl acetate, dried over Na₂SO₄, concentrated and purified via column chromatography on silica with ethyl acetate as eluent (*R*_f ≈ 0.17; yield: 50%). This synthesis was performed as variation from literature.⁴

1.1.1.1.6. Poly(2,5-dimethoxy-4-vinylbenzoic acid) (P(VDMB))

A 2-neck round bottom flask was equipped with a reflux-condenser and a septum was evacuated, heated, and flooded with N₂ three times. Afterwards, the flask was charged with **6** (0.2 g, 1 mmol, 1 eq.) and 2,2'-azobis(2-methylpropanitrile) (0.0032 g, 19.2 · 10⁻³ mmol, 0.02 eq.) and evacuated and flooded with N₂ three times. After adding dry dioxane (4.15 ml), the solution was stirred at 90 °C for 7 h under N₂ atmosphere. Afterwards, the crude reaction mixture was cooled to rt and concentrated in Et₂O. After filtration the polymer was purified by soxhlet using Et₂O and finally extracted using acetone, resulting in the pure product as white solid (yield: 72 wt. %). This synthesis was performed as variation from literature.⁶

Figure S9: Molecular structure of **8**.

1.1.1.1.5. 2,5-Dimethoxy-4-vinylbenzoic acid (6)

¹H-NMR (DMSO-d₆, 500 MHz) (see Figure S25 and Figure S21):
 δ [ppm] = 2.78 (s, 3H, H-10), 3.94 (s, 3H, H-11), 5.57-5.60 (dd, 1H, H-13"), 6.19-6.23 (dd, 1H, H-13'), 7.04-7.06 (d, 1H, H-6), 7.24-7.30 (dd, 1H, H-12), 7.82-7.83 (d, 1H, H-4), 8.06-8.07 (d, 1H, H-5), 8.16-8.17 (d, 1H, H-3), 10.92 (s, 1H, H-9).
¹³C-NMR (DMSO-d₆, 125 MHz) (see Figure S30):
 δ [ppm] = 20.64 (C-10), 56.09 (C-11), 105.42 (C-6), 108.89 (C-8), 111.84 (C-4), 115.93 (C-13), 118.39 (C-4a), 121.54 (C-5), 127.65 (C-4b), 128.60 (C-12), 134.96 (C-8a), 138.06 (C-9a), 140.27 (C-3), 141.76 (C-1), 157.95 (C-7).

1.1.1.1.4. 4-Vinyl-2,5-dimethoxymethylbenzoate (5)

A 2-neck round bottom flask was equipped with a septum and a reflux condenser and evacuated, heated, and flooded with N₂ three times. **4** (1 g, 3.6 mmol, 1 eq.) was dissolved in 1,4-dioxane (34 ml) and sodium carbonate was dissolved in water (4.5 ml). Both mixtures were combined with vinylboronic acid pinacol ester (1.54 ml, 9.1 mmol, 2.5 eq.) and stirred at room temperature while bubbling with N₂ for 30 min. Then, tetrakis(triphenylphosphine)palladium(0) (0.336 g, 0.336 mmol, 0.08 eq.) was added. After stirring at 100 °C for around 72 h the solution was cooled to rt, the crude mixture was transferred to a separation funnel using water and ethyl acetate, and the aqueous phase was extracted three times with ethyl acetate. The combined organic phases were dried over Na₂SO₄, concentrated, and purified via column chromatography (hexane / EtOAc = 4:1; *R*_f = 0.45). The resulted yellow oil was recrystallized by adding a few drops of *n*-PrOH and cooling to 20 °C, resulting in the pure product as white solid (yield: 60%). This synthesis was performed as variation from literature.⁴

Figure S4: Molecular structure of **5**.

1.1.1.1.5. 2,5-Dimethoxy-4-vinylbenzoic acid (6)

¹H-NMR (DMSO-d₆, 500 MHz) (see Figure S17 and Figure S18):
 δ [ppm] = 3.79-3.81 (m, 9H, H-8/9/12), 5.40-5.42 (dd, 1H, H-11"), 5.98-6.02 (dd, 1H, H-11'), 6.92-6.98 (dd, 1H, H-10), 7.25-7.26 (d, 2H, H-6/3)-
¹³C-NMR (DMSO-d₆, 125 MHz) (see Figure S19):
 δ [ppm] = 51.85 (C-8), 56.01 (C-9), 56.42 (C-6), 110.82 (C-6), 113.50 (C-4), 117.57 (C-3), 119.58 (C-1), 130.36 (C-10/11), 149.77 (C-5), 152.51 (C-2), 165.73 (C-7).

1.1.1.1.5. 2,5-Dimethoxy-4-vinylbenzoic acid (6)

5 (0.5 g, 2.3 mmol, 1 eq.) was dissolved in acetone (24.2 ml) and potassium hydroxide (25.3 g, 450 mmol, 200 eq.) was dissolved in water (22.5 ml). Both solutions were combined and stirred vigorously overnight at 100 °C. After cooling to rt, the crude reaction mixture was transferred to a separation funnel with 1M KOH (aq.) and ethyl acetate. After extraction of the aqueous phase with ethyl acetate, the aqueous phase was acidified with 12M HCl to pH 1-2. The acidified solution was extracted with ethyl acetate and the combined organic phases were dried over Na₂SO₄. Removing the solvent in vacuo gave a solid residue, which was recrystallized in boiling water to yield the product as white solid (yield: 67%). This synthesis was performed as variation from literature.⁵

Figure S5: Molecular structure of **6**.

1.1.1.1.5. 2,5-Dimethoxy-4-vinylbenzoic acid (6)

¹H-NMR (DMSO-d₆, 500 MHz) (see Figure S20 and Figure S21):
 δ [ppm] = 3.78-3.82 (d, 6H, H-9/12), 5.38-5.40 (dd, 1H, H-11"), 5.96-6.00 (dd, 1H, H-11'), 6.92-6.98 (dd, 1H, H-10), 7.23-7.26 (d, 2H, H-6/3), 12.57 (br, 1H, H-8).
¹³C-NMR (DMSO-d₆, 125 MHz) (see Figure S22):
 δ [ppm] = 56.00 (C-9), 56.48 (C-12), 110.83 (C-6), 113.65 (C-4), 117.32 (C-3), 120.82 (C-1), 129.97 (C-11), 130.46 (C-10), 149.83 (C-5), 152.50 (C-2), 166.76 (C-7).

1.1.1.1.6. Poly(2,5-dimethoxy-4-vinylbenzoic acid) (P(VDMB))

A 2-neck round-bottom flask equipped with a reflux-condenser and a septum was evacuated, heated, and flooded with N₂ three times. Afterwards, the flask was charged with **6** (0.2 g, 1 mmol, 1 eq.) and 2,2'-azobis(2-methylpropanitrile) (0.0032 g, 19.2 · 10⁻³ mmol, 0.02 eq.) and evacuated and flooded with N₂ three times. After adding dry dioxane (4.15 ml), the solution was stirred at 90 °C for 7 h under N₂ atmosphere. Afterwards, the crude reaction mixture was cooled to rt and concentrated in Et₂O. After filtration the polymer was purified by soxhlet using Et₂O and finally extracted using acetone, resulting in the pure product as white solid (yield: 72 wt. %). This synthesis was performed as variation from literature.⁶

Figure S6: Molecular structure of **P(VDMB)**.

1.1.1.1.6. Poly(2,5-dimethoxy-4-vinylbenzoic acid) (P(VDMB))

¹H-NMR (DMSO-d₆, 500 MHz) (see Figure S23):
 δ [ppm] = 0.83-1.97 (br, 2H, H-11), 2.22-2.77 (br, 1H, H-10), 3.08-3.67 (br, 6H, H-9/12), 4.78-6.71 (br, 1H, H-3), 6.71-7.62 (br, 1H, H-6), 9.93-13.67 (br, 1H, H-8).

- S5 -

1.1.2.3. Poly(8-vinyl-7-methoxy-1-methyl-9H-pyrido[3,4-b]indole) (P(VHn))^s

A 2-neck round-bottom flask was equipped with a reflux-condenser was evacuated, heated, and flooded with N₂ three times. **8** (0.2 g, 0.8 mmol, 1 eq.) and azobis(2-methylpropanitrile) were dissolved in dry 1,4-dioxane (3.63 ml). The solution was heated at 90 °C and stirred for 7h under nitrogen atmosphere. Afterwards, the solution was cooled to room temperature, transferred to a one neck flask with chloroform and concentrated to dryness. After precipitation in Et₂O, the polymer was purified via Soxhlet in Et₂O and extracted with chloroform. The polymer was obtained as a white/yellowish brittle solid (yield: 24 wt. %). This synthesis was performed as variation from literature.⁶

¹H-NMR (DMSO-*d*₆, 500 MHz) (see Figure S31):
δ [ppm] = 0.25-1.45 (br, 2H, H-13), 1.60-3.15 (br, 4H, H-10/12), 3.48-4.92 (br, 3H, H-11), 5.77-8.88 (br, 4H, H-3/4/5/6).

2. Instrumental & Sample Preparation Information

2.1. Nuclear magnetic resonance (NMR)

A Bruker Avance III 500 spectrometer was used to record ¹H-NMR spectra at 500.13 MHz and ¹³C-NMR at 125.03 MHz at room temperature. The spectra were referenced to the non-deuterated solvent signal of DMSO (DMSO-*d*₆: δ(¹H) = 2.5 ppm, δ(¹³C) = 39.52 ppm). Abbreviations for ¹H-NMR spectra: s - singlet, d - doublet, dd - doublet of doublet, t - triplet, qua - quartet, and br - broad signal. All measured spectra are depicted in subchapter 3.1.

2.2. UV/Vis

UV/Vis measurements were carried out on a Perkin Elmer "Lambda800" instrument using two different sample preparation methods: (1) As drop-casted films (concentration = 1 mg/ml) on glass slide; (2) In MeOH solutions to determine the mass attenuation coefficient (MAC) at the laser wavelength of the MALDI-laser (λ_{laser} = 355 nm). For each concentration, two spectra were recorded from individual solutions prepared from different stock solutions (10 stock solutions per sample). The final sample spectra are the calculated mean of the two measured spectra. The mass attenuation coefficients (MAC) were calculated with the following formula (S1):

$$\epsilon = \frac{A}{c \cdot d} \quad (S1)$$

with ϵ = mass attenuation coefficient [$\frac{m^2}{kg}$]; A = Absorbance [a.u.]; c = concentration [$\frac{kg}{m^3}$]; d = diameter of cuvette [m]. Used here: d = 0.01 m; c = 0.01, 0.025, 0.05, 0.075 and 0.1 kg/m³.

To determine the MAC, the mean values from the averaged spectra of all concentrations were calculated. All spectra are depicted in subchapter 3.2, Table S5 to Table S7 give additional data.

2.3. Vacuum stability

For the determination of the vacuum stability, the compounds were drop-casted (100 µl of a 10 mg/ml solution) on a glass slide and kept in a MALDI instrument at a pressure of around 3·10⁻⁷ mbar. The sample weight was determined using a "Sartorius" "CPA225D" balance (Δm = 0.01 mg). The mass of the deposited sample at t = 0h is set as 100% and all other values were calculated relative to this value. For the measurements up to 24 h, three individual measurements were carried out for each matrix. For each matrix, one additional measurement was performed after 72 h. Before reading the mass values from the balance, the operator waited up to 5 minutes to ensure correct values. The plots of the obtained averaged values are shown in subchapter 3.3.

2.4. Fourier-transform infrared spectroscopy (FTIR)

All FTIR-measurements were carried out on a "Vertex 80v" from Bruker Daltonics equipped with a Golden Gate Diamond ATR Unit (SPECAC) and an MCT-detector. The spectra were recorded in the range of 4000 - 600 cm⁻¹ with a resolution of 4 cm⁻¹ and 100 measurements per spectra. After measuring, all spectra were baseline-corrected, and the diamond bands were subtracted. The single sample spectra and corresponding comparison spectra are shown in subchapter 3.4.

2.5. Gel permeation chromatography (GPC)

GPC measurements were performed using an Agilent PL-GPC 220 equipped with a Polar-Gel-M (300 x 7.5 mm) column and a HPLC-pump 1200 from Agilent Technologies. Dimethylacetamide (DMAc) with 3 mg/ml LiCl was used as eluent at 1 ml/min flow rate. The samples were detected with a light scattering (LS) MiniDAWN - Detector TEOSS II from Wyatt Technologies and simultaneously a refractive index (RI) KNAUER - Detector K-2301. All samples were measured twice, and the resulting values were averaged. The resulting data is shown in subchapter 3.5.

2.6. MALDI MS

2.6.1. Instrumental parameters "Autoflex"

All MALDI spectra were recorded with a "Bruker Autoflex speed" MALDI-TOF/TOF with a smartbeam-II laser with a wavelength of 355 nm.⁹ The used software is FlexControl 3.4. For the evaluation and visualization of the spectra, the Originlab Pro 2020b was used. All measurements were carried out using a standard set of parameters¹⁰ listed in Table S3. All spectra depicted in the manuscript and the SI are a combination of 10 spectra a 500 shots, resulting in a total shot number of 5000 per spectrum. All 10 measurements per spectrum were measured on different spots of the dried droplet spot and were chosen randomly by the operator. When a spectrum exhibited a good quality (e.g. high S/N-ratio, less impurities, etc.), it was added to the final combination.

Table S3: Parameters for positive and negative reflection mode and positive linear mode MALDI measurements.

Parameter	Reflection			Linear		
	Positive	Negative	Positive	Negative	Positive	Linear
m/z range	0 - 3000/20000	0 - 3000/20000	0 - 20000	0 - 20000	0 - 20000	0 - 20000
Sample rate	2.0	2.0	2.0	2.0	2.0	2.0
Shots per measurement	500	500	500	500	500	500
Matrix suppression	Off	Off	Off	Off	Off	Off
Measurements per spectra	10	10	10	10	10	10
Random walk	Off	Off	Off	Off	Off	Off
Detector gain	100x	100x	100x	100x	100x	100x
Electronic gain	Enhanced	Enhanced	Enhanced	Enhanced	Enhanced	Enhanced
Realtime smooth	Off	Off	Off	Off	Off	Off
Laser-frequency	1000 Hz	1000 Hz	1000 Hz	1000 Hz	1000 Hz	1000 Hz
Laser-size	Large	Large	Large	Large	Large	Large

2.6.2. Calibration

2.6.2.1. Low molecular weight area (m/z = 0 - 3000)

At the beginning of each set of low molecular weight MALDI measurements in reflection mode and after changing the measurement polarity, the instrument was calibrated as described in literature.¹⁰⁻¹¹

2.6.2.2. High molecular weight area (m/z = 0 - 20k)

At the beginning of each set of high molecular weight MALDI measurements in reflection positive mode, the instrument was calibrated with PEG650 as described in literature.^{10, 12}

2.6.3. Sample plate cleaning procedure

The sample plate was cleaned before every new series of measurements using a modified literature protocol:¹⁰

1. Wipe the sample plate with MeOH-soaked paper tissue;
2. Ultra-sonication for 10 min at rt in 0.1M NaOH (aq.) / i-PrOH (100 ml / 50 ml);
3. Rinse with 50 ml Milli-Q H₂O;
4. Rinse with 50 ml i-PrOH;
5. Ultra-sonication for 10 min at rt in MeOH / acetone (100 ml / 50 ml);
6. Rinse with 50ml acetone;
7. Rinse with 50ml MeOH;
8. Ultra-sonication for 10 min at rt in hexane / chloroform (100 ml / 50 ml);
9. Rinse with 50 ml chloroform;
10. Rinse with 50 ml hexane.

3. Analytical measurements

3.1. NMR

3.1.1. 4-bromo-2,5-dimethoxy-benzaldehyde (2)

Figure S11: ¹H-NMR spectrum of **2** in DMSO at 500 MHz.

Figure S12: ¹³C-NMR spectrum of **2** in DMSO at 125 MHz.

3.1.2. 4-bromo-2,5-dimethoxybenzoic acid (3)

Figure S13: ¹H-NMR spectrum of **3** in DMSO at 500 MHz.

- S10 -

2.6.4. Sample preparation

Dried-droplet method

The samples were prepared with the dried-droplet method on a standard sample plate (Bruker "MTP 384 target plate ground steel BC" Part-No.: 8280784).^{13,14} For the "polymer as matrix" measurements (see subchapter 3.6.2) the "layered" dried-droplet method was used for sample preparation. The matrices harmine, **PVDHB** and **PVHar** were dissolved in MeOH and water in a volume ratio of 3:1, whereas DHB was dissolved in acetone with a concentration depicted in the individual spectra. The analytes myristic acid (MA), linoleic acid (LA), docosahexaenoic acid (DA), and dipalmitoyl-glycerol (DG) were dissolved in acetone; the analytes fructose (FT), glucuronic acid (GA), sucrose (SR), maltotriose (MT), α -cyclodextrin (α -CD), methionine (ME) and phenylalanine (PA) were dissolved in MeOH and water with a volume ratio of 3:1; the analyte dipalmitoyl-glycero-phosphocholine (DGPC) was dissolved in MeOH and water with a volume ratio of 4:1; the analyte choline chloride (CC) was dissolved in acetonitrile; the analytes phosphocholine (PC) and glycerophosphorylcholine (GPC) were dissolved in acetonitrile and water with a volume ratio of 2:1. The analyte concentrations are depicted in the individual spectra. First the analyte solution was dripped on the plate ($V = 1 \mu\text{l}$) and dried in air for several seconds to several min until the solvent had completely evaporated. Then the matrix was coated by dripping the corresponding solutions ($V = 1 \mu\text{l}$) on top of the analyte and dry it in air for several seconds.¹³ For the "polymer as analyte" measurements (see subchapter 3.6.1) the "mixed" dried-droplet method was used for sample preparation. The polymers **PVDMB** and **PVHar** were dissolved in MeOH as well as the secondary matrix DCTB ($c = 1 \text{ mg/ml}$). For both polymers the mixing ratio was 10:1. After dropping $1 \mu\text{l}$ of matrix/analyte-solution on the sample plate the solvent was evaporated in air for several seconds. The measurements of **PVDHB** were performed according to the parameters listed in Table S10.

2.7. MALDI Imaging

2.7.1. Matrix solutions preparation

The matrix solutions were prepared as follow:

PVDHB: 5 mg of **PVDHB** were dissolved in 5 mL MeOH (HPLC grade, Sigma Aldrich). Afterwards 2 mL of acetone (HPLC grade, Sigma Aldrich) and 3 mL H₂O (HPLC grade, Fischer Scientific) were added to obtain the final solution.

PVHar: 5 mg of **PVHar** were dissolved in 7.5 mL MeOH (HPLC grade, Sigma Aldrich). Afterwards 2.5 mL H₂O (HPLC grade, Fischer Scientific) were added to obtain the final solution.

2.7.2. Instrumental parameters "Rapiflex"

Table S4: Instrumental parameters for MALDI Imaging measurements in reflectron mode.

Parameter	PVDHB		PVHar	
m/z range	Positive	Negative	Positive	Negative
Shots per measurement	61 - 1401	40 - 1200	61 - 1401	40 - 1200
	300	1000	150	100
Matrix suppression	48.7951	31.9987	48.7951	31.9987
	2345 V	2276 V	2276 V	2276 V
DetectorGain	2345 V	2276 V	2276 V	2276 V
	10 kHz	10 kHz	10 kHz	10 kHz
Spot size	70 μm	70 μm	70 μm	70 μm
	20.000 kV	20.000 kV	20.000 kV	20.000 kV
IonSource1	11.600 kV	11.600 kV	2.610 kV	11.600 kV
	20.840 kV	20.840 kV	20.840 kV	20.840 kV
Reflector	1.085 kV	1.085 kV	1.085 kV	1.085 kV
	8.7 kV	8.7 kV	8.7 kV	8.7 kV
Reflector3	110 ns	110 ns	110 ns	110 ns
	0.326519	0.326519	0.326519	0.326519
PieDelay	64	60	55	61
	66 μm	66 μm	66 μm	66 μm
DeflectionConstant	64	60	55	61
	66 μm	66 μm	66 μm	66 μm
PowerUV	64	60	55	61
	66 μm	66 μm	66 μm	66 μm
Scan size X / Y	64	60	55	61
	66 μm	66 μm	66 μm	66 μm

- S9 -

Figure S14: ^{13}C -NMR spectrum of **3** in DMSO at 125 MHz.

Figure S15: ^1H -NMR spectrum of **4** in DMSO at 500 MHz.

Figure S16: ^{13}C -NMR spectrum of **4** in DMSO at 125 MHz.

- S11 -

3.1.4. 4-vinyl-2,5-dimethoxybenzoate (**5**)

Figure S17: ^1H -NMR spectrum of **5** in DMSO at 500 MHz.

Figure S18: Close-up of ^1H -NMR spectrum of **5** in DMSO at 500 MHz.

Figure S19: ^{13}C -NMR spectrum of **5** in DMSO at 125 MHz.

- S12 -

3.1.5. 4-vinyl-2,5-dimethoxybenzoic acid (**6**)

Figure S20: ^1H -NMR spectrum of **6** in DMSO at 500 MHz.

Figure S21: Close-up of ^1H -NMR spectrum of **6** in DMSO at 500 MHz.

Figure S22: ^{13}C -NMR spectrum of **6** in DMSO at 125 MHz.

- S13 -

3.1.6. P(VDMB)

Figure S23: ^1H -NMR spectrum of P(VDMB) in DMSO at 500 MHz.

3.1.7. P(VDHB)

3.2. UV/VIS

3.2.1. Drop-casted films

Figure S32: UV/VIS spectra of drop-casted films of DHB (yellow), P(VDMB) (green), P(VDHB) (red), harmine (blue) and P(VHar) (black). All spectra were normalized from 0 to 1.

Table S5: UV/VIS absorption maxima of drop-casted SOMs and PMs.

Polymer	Absorption maxima [nm] ^a	Relative absorption at 355 nm [%] ^b
DHB	328 (100%)	62
P(VDMB)	321 (100%)	61
P(VDHB)	332 (100%)	62
harmine	306 (88%) 330 (93%)	51
P(VHar)	344 (100%) 307 (100%) 335 (67%)	42

^a relative to the global absorption maximum of each compound.

3.2.2. Solution

Figure S33: MAC plot of DHB (yellow), P(VDMB) (green), P(VDHB) (red), harmine (blue) and P(VHar) (black). Here, all concentrations were averaged for final results.

Table S6: Mass attenuation coefficients (MACs) of the absorption maxima of SOMs and PMs as MeOH solutions.

Concentration [g/l]	Maximum mass attenuation coefficient [m²/kg]				
	DHB	P(VDMB)	P(VDHB)	harmine	P(VHar)
0.01	5067.6 (335)	2924.2 (319)	3816.9 (335)	9772.5 (301) 3615.0 (324) 3158.0 (337)	5098.0 (306) 2107.2 (332) 1552.7 (345)
	3368.2 (335)	2980.0 (319)	3554.9 (335)	10744.8 (301) 3883.4 (324) 3387.5 (337)	4729.7 (306) 1923.7 (332) 1403.4 (345)
0.05	3660.0 (335)	2621.9 (319)	2885.5 (335)	7497.9 (301) 3725.4 (324) 3266.1 (337)	5755.2 (306) 2329.7 (335) 1693.7 (345)
	3775.7 (335)	2392.4 (319)	2552.5 (335)	9677.7 (301) 3756.8 (324) 3270.1 (337)	4119.1 (306) 1667.9 (332) 1212.8 (345)
0.1	3083.7 (335)	2256.9 (319)	2526.5 (335)	10464.1 (301) 3869.6 (324) 3377.9 (337)	4503.5 (306) 1730.3 (332) 1258.9 (345)
	3791.0 (335)	2635.1 (319)	3067.3 (335)	9631.4 (301) 3770.0 (324) 3291.9 (337)	4841.1 (306) 1951.8 (332) 1424.3 (345)
average					

Table S7: MACs of SOMs and P(SOMs) as MeOH solutions at laser wavelength (355 nm).

Concentration [g/l]	Mass attenuation coefficient at 355 nm [m²/kg]				
	DHB	P(VDMB)	P(VDHB)	harmine	P(VHar)
0.01	2615.4	137.9	2147.7	343.1	820.1
	1766.9	144.7	2041.9	203.7	716.7
0.025	1941.4	130.1	1667.9	182.4	729.3
	2003.9	118.2	1483.8	412.6	613.7
0.1	1640.3	132.8	1459.7	186.4	636.8
	1993.6	132.7	1760.2	265.6	729.3

3.3. Vacuum stability

Figure S34: Plot of relative masses of DHB during high vacuum in MALDI instrument.

Figure S35: Plot of relative masses of P(VDHB) during high vacuum in MALDI instrument.

Figure S36: Plot of relative masses of hamine during high vacuum in MALDI instrument.

3.4. FTIR

3.4.1. Single spectra

Figure S38: FTIR spectrum of 6.

Figure S40: FTIR spectrum of P(VDHB).

Figure S37: Plot of relative masses of P(VHer) during high vacuum in MALDI instrument.

Figure S39: FTIR spectrum of P(VDMB).

Figure S41: FTIR spectrum of 8.

Figure S42: FTIR spectrum of P(VHer).

Figure S44: FTIR spectra of P(VDMB) (blue) and P(VDHB) (red) highlighting two areas which show reaction related differences.

3.5. GPC

3.5.1. P(VDMB)

Figure S46: GPC-chromatogram of P(VDMB) analyzed by light scattering (LS).

Figure S43: FTIR spectra of 6 (blue) and P(VDMB) (red) highlighting three areas which show reaction related differences.

Figure S45: FTIR spectra of 8 (blue) and P(VHer) (red) highlighting three areas which show reaction related differences.

Figure S47: GPC-chromatogram of P(VDMB) analyzed by refractive index (RI).

Table S8: GPC results for P(VHar).

Name	M_n [g/mol]	M_w [g/mol]	$\bar{P}$	DP	$dn/dc_{calc.}$
P(VHar)-1	12000	12900	1.08	-54	0.177
P(VHar)-2	12200	13100	1.07	-55	0.177
Averaged	12100	13000	1.07	-55	0.177

Explanation for the different values determined with GPC and MALDI MS:

Table S9: Molar masses of P(VHar) at 8.59 min elution time.

Name	$M_{theo.}$ [g/mol] @ 8.59 min.
P(VHar)-1	1718.30684 = 1718
P(VHar)-2	1705.79589 = 1706
Averaged	1712.05137 = 1712

When comparing the calculated value ($M_{theo.}=1712$ g/mol) with the MALDI MS spectrum of P(VHar) (Figure S55) one can see that the calculated value is close to the maximum signal of the MALDI spectrum (ca. $m/z = 1900$). The difference can have three reasons: 1) the elution volume is not completely linear, which could cause deviations; 2) the $dn/dc_{calc.}$ value can deviate and 3) the extrapolation of the molar mass curves is not an accurate method and only gives theoretical values.

3.6. MALDI MS

3.6.1. Polymers as analytes

3.6.1.1. P(VDHB)

Although the precursor polymer (P(VDHB)) is observable with MALDI MS (see Figure S54), we were not able to detect the polymer P(VDHB) with MALDI MS. In Table S10 sample preparation parameters are listed, which were tried out and failed. The SOM, the polymer analyte and the additive were always dissolved in the same solvent (MeOH) with the same concentration. The vol.-% unit of the additive refers to the volume of the matrix solution.

Table S10: Tested parameters to detect P(VDHB), none of which were successful.

Matrix	Additive	Sample preparation	Detection mode
DCTB (1 mg/ml)	1 vol.-% NaTFA	Mixed (10:1)	Reflection positive/negative
DCTB (1 mg/ml)	1 vol.-% NaTFA	Mixed (1:1)	Reflection positive/negative
DCTB (10 mg/ml)	1 vol.-% NaTFA	Mixed (10:1)	Reflection positive/negative
DCTB (10 mg/ml)	(-)	Mixed (10:1)	Reflection positive/negative
DHB (1 mg/ml)	1 vol.-% NaTFA	Mixed (10:1)	Reflection positive/negative
9AA (1 mg/ml)	1 vol.-% NaTFA	Mixed (10:1)	Reflection positive/negative
Dithranol (1 mg/ml)	1 vol.-% NaTFA	Mixed (10:1)	Reflection positive/negative
Dithranol (1 mg/ml)	1 vol.-% NaTFA	Mixed (1:1)	Reflection positive/negative

3.5.2. P(VDHB)

Figure S48: GPC-chromatogram of P(VDHB) analyzed by light scattering (LS).

3.5.3. P(VHar)

Figure S49: GPC-chromatogram of P(VHar) analyzed by light scattering (LS).

Figure S50: GPC-chromatogram of P(VHar) analyzed by light scattering (LS).

Figure S48: GPC-chromatogram of P(VDHB) analyzed by refractive index (RI).

Figure S49: GPC-chromatogram of P(VHar) analyzed by refractive index (RI).

Figure S51: Zoomed GPC-chromatogram of P(VHar) analyzed by light scattering (LS).

Figure S53: Molar mass curve of P(VHar) analyzed by light scattering (LS).

Figure S54: Reflection negative mode MALDI MS spectra of DCTB mixed with P(VDMB) (10:1) (blue) and pure DCTB matrix (black). All measurements were performed with 100% laser intensity. Both spectra were smoothed and baseline corrected.

Table S11: Polymeric information about P(VDMB) calculated from MALDI MS spectrum.

	Blue	Red	Green	Average
M_n [g/mol]	2863	2724	2891	2826
M_w [g/mol]	3111	2944	3079	3045
$\bar{P}$	1.09	1.08	1.07	1.08
DP	~15	~14	~15	15
$M_{n,theo}$ [g/mol]	208.37	208.38	208.50	208.42
$\Delta M_{n,theo}$	0.11	0.22	0.34	0.22
End group	IBN-[VDMB] $_n$ -IBN	IBN-[VDMB] $_n$ -H ⁺	IBN-[VDMB] $_n$ -H ⁺	IBN-[VDMB] $_n$ -IBN + CN ⁻
IBN = isobutyronitrile, $M_{n,theo}$ = 208.21 g/mol

The MALDI mass spectrum exhibits three signal series representing two different end group combinations, which are expected for free radical polymerization: (1) isobutyronitrile (IBN) on both sides and (2) IBN and disproportionated monomer (Figure S44 and Table S11). Both combinations are ionized by deprotonation. End group combination (2) was also detected as CN⁻-cluster.

- S23 -

Figure S55: Reflection positive mode MALDI MS spectra of DCTB mixed with P(VHar) (10:1) (blue) and pure DCTB matrix (black). All measurements were performed with 80% laser intensity. Both spectra were smoothed and baseline corrected.

Table S12: Polymeric information about P(VHar) calculated from MALDI MS spectrum.

	Blue	Red	Green	Average
M_n [g/mol]	2947	3485	2661	3031
M_w [g/mol]	3632	4093	3104	3610
$\bar{P}$	1.23	1.17	1.16	1.19
DP	~15	~17	~13	15
$M_{n,theo}$ [g/mol]	238.35	238.28	238.16	238.26
$\Delta M_{n,theo}$	0.2	0.15	0.14	0.16
End group	IBN-[VHar] $_n$ -H ⁺	IBN-[VHar] $_n$ -IBN + H ⁺	VHar-[VHar] $_n$ -VHar	VHar-[VHar] $_n$ -VHar + H ⁺
IBN = isobutyronitrile, $M_{n,theo}$ = 238.29 g/mol

In MALDI MS, three different end group combinations were detected: (1) IBN and disproportionated monomer, (2) IBN on both sides and (3) disproportionated monomer on both sides (Figure S45 and Table S12). All were ionized by protonation and represent end groups expected for radical polymerization.

- S24 -

3.6.2. Polymers as matrices

3.6.2.1. DHB / PVDHB - single analyte

Analyte: Fructose (FT)

Figure S59: Reflection positive (left) and negative (right) mode MALDI MS spectra of the sample plate (violet), only the analyte SR (yellow), only the pure matrix DHB (green), DHB coated on SR (red), only the pure matrix PVDHB (blue) and PVDHB coated on SR (black). Positive and negative mode measurements were performed with 30% and 40% laser intensity, respectively. Please mind the differences in Y-axis scale, $m/z = 23$ [Na]⁺, 39 [K]⁺, 57 [SR+H]⁺, 71 [SR+H]⁺, 153 [DHB+H]⁺, 203 [SR+H]⁺, 315 [matrix]⁺, 341 [SR+H]⁺, 365 [SR+Na]⁺.

Figure S60: Reflection positive (left) and negative (right) mode MALDI MS spectra of the sample plate (violet), only the analyte MT (yellow), only the pure matrix DHB (green), DHB coated on MT (red), only the pure matrix PVDHB (blue) and PVDHB coated on MT (black). All measurements were performed with 40% laser intensity. Spectrum in positive mode was baseline corrected. Please mind the differences in Y-axis scale, $m/z = 23$ [Na]⁺, 39 [K]⁺, 71 [MT+H]⁺, 153 [DHB+H]⁺, 177 [DHB+Na]⁺, 193 [DHB+K]⁺, 307 [DHB+DHB+H]⁺, 315 [matrix]⁺, 341 [MT+H]⁺, 527 [MT+Na]⁺.

^a see Table S1 and Table S2 for abbreviations; ^b clear assignment not possible

- S27 -

Figure S61: Reflection positive mode MALDI MS spectra of the sample plate (violet), only the analyte α-CD (yellow), only the pure matrix DHB (green), DHB coated on α-CD (red), only the pure matrix PVDHB (blue) and PVDHB coated on α-CD (black). All measurements were performed with 40% laser intensity, $m/z = 23$ [Na]⁺, 39 [K]⁺, 177 [DHB+Na]⁺, 995 [α-CD+Na]⁺.

It was not possible to detect α-CD (neither molecule ion at $m/z = 971$ nor any fragments) in negative mode using DHB or PVDHB as matrices. All samples were prepared via previously described "layered" dried-droplet method. The measurements were carried out using different matrix concentrations (0.1 and 0.5 mg/ml), analyte concentrations (0.1, 0.5, 1, 5 and 10 mg/ml) and laser intensities (20-90%).

Figure S62: Reflection positive (left) and negative (right) mode MALDI MS spectra of the sample plate (violet), only the analyte MA (yellow), only the pure matrix DHB (green), DHB coated on MA (red), only the pure matrix PVDHB (blue) and PVDHB coated on MA (black). All measurements were performed with 30% laser intensity. All positive mode spectra were baseline corrected. Please mind the differences in Y-axis scale, $m/z = 23$ [Na]⁺, 39 [K]⁺, 57 [MA+H]⁺, 89 [MA+H]⁺, 137 [DHB+H]⁺, 154 [DHB+H]⁺, 177 [DHB+Na]⁺, 193 [DHB+K]⁺, 215 [MA+H]⁺, 227 [MA+H]⁺, 251 [MA+Na]⁺, 267 [MA+K]⁺, 315 [matrix]⁺, 722 [MA+H]⁺.

^a see Table S1 and Table S2 for abbreviations; ^b clear assignment not possible

- S28 -

Figure S67: Reflection positive (left) and negative (right) mode MALDI MS spectra of the sample plate (violet), only the analyte ME (yellow), only the pure matrix DHB (green), DHB coated on ME (red), only the pure matrix P(VDHB) (blue) and P(VDHB) coated on ME (black). Positive and negative mode measurements were performed with 40% and 30% laser intensity, respectively. $m/z = 23$ [Na]⁺, 39 [K]⁺, 148 [ME-H]⁺, 153 [DHB-H]⁺, 172 [ME+Na]⁺, 177 [DHB+Na]⁺.

Figure S68: Reflection positive (left) and negative (right) mode MALDI MS spectra of the sample plate (violet), only the analyte PA (yellow), only the pure matrix DHB (green), DHB coated on PA (red), only the pure matrix P(VDHB) (blue) and P(VDHB) coated on PA (black). Positive and negative mode measurements were performed with 40% and 30% laser intensity, respectively. $m/z = 23$ [Na]⁺, 39 [K]⁺, 153 [DHB-H]⁺, 164 [PA-H]⁺, 177 [DHB+Na]⁺, 188 [PA+Na]⁺.

^a see Table S1 and Table S2 for abbreviations; ^b clear assignment not possible

- S31 -

3.6.2.2. Harmine / P(VHar) - single analyte Analyte: Fructose (FT)

Figure S69: Reflection positive (left) and negative (right) mode MALDI MS spectra of the sample plate (violet), only the analyte FT (yellow), only the pure matrix harmine (green), harmine coated on FT (red), only the pure matrix P(VHar) (blue) and P(VHar) coated on FT (black). All measurements were performed with 50% laser intensity. $m/z = 55$ [Cl]⁺, 71 [FT-H]⁺, 101 [FT+Na]⁺, 119 [FT-H]⁺, 130 [FT+Na]⁺, 203 [FT+Na]⁺, 211 [Har-H]⁺, 212 [Har]⁺, 423 [Har+Har]⁺.

Figure S70: Reflection positive (left) and negative (right) mode MALDI MS spectra of the sample plate (violet), only the analyte GA (yellow), only the pure matrix harmine (green), harmine coated on GA (red), only the pure matrix P(VHar) (blue) and P(VHar) coated on GA (black). Positive and negative mode measurements were performed with 50% and 10% laser intensity, respectively. All negative mode spectra were baseline corrected. $m/z = 58$ [GA-H]⁺, 71 [GA+Na]⁺, 87 [GA-H]⁺, 130 [GA+Na]⁺, 193 [GA-H]⁺, 211 [Har-H]⁺, 212 [Har]⁺, 213 [Har-H]⁺, 217 [GA+Na]⁺, 247 [Har+Cl]⁺, 423 [Har+Har]⁺.

^a see Table S1 and Table S2 for abbreviations; ^b clear assignment not possible

- S32 -

Figure S73: Reflection positive mode MALDI MS spectra of the sample plate (violet), only the analyte α -CD (yellow), only the pure matrix harmin (green), harmin coated on α -CD (red), only the pure matrix **PivHar** (blue) and **PivHar** coated on α -CD (black). All measurements were performed with 40% laser intensity. $m/z = 23$ [Na^+], 39 [K^+], 213 [Har-H^+], 422 [Har-Har^+], 995 [$\alpha\text{-CD}+\text{Na}^+$].

It was not possible to detect α -CD (neither molecule ion at $m/z = 971$ nor any fragments) in negative mode using **PVHA** as matrix. With **hmg** it was only possible with a matrix concentration of 0.5 mg/ml and an analyte concentration of 1 mg/ml at very high laser intensity (90%) (Figure S74). All samples were prepared via previously described "layered" dried-droplet method. The measurements were carried out using different matrix concentrations (0.1 and 0.5 mg/ml), analyte concentrations (0.1, 0.5, 1, 5 and 10 mg/ml) and laser intensities (20–90%).

Figure S74: Reflectron negative mode MALDI MS spectra of harmine coated on α -CD. This measurement was performed with 90% laser intensity.

^a see Table S1 and Table S2 for abbreviations; ^b clear assignment not possible

Figure S71: Reflection positive (left) and negative (right) mode MALDI-MS spectra of the sample plate (violet), only the analyte SR (yellow), only the pure matrix hamine (green), hamine coated on SR (red), only the pure matrix **PwHAr** (blue) and **PwHAr** coated on SR (black). All measurements were performed with 40% laser intensity, m/z = 23 [Na], 39 [K], 71 [S₂]⁺, 211 [Har-H⁺], 212 [Har⁺], 213 [Har-H⁺], 341 [SR-H⁺], 365 [SR-Na⁺], 407 [m+trif].

Analyte: Maltotriose (MT)

Figure S72: Reflection positive (left) and negative (right) mode MALDI MS spectra of the sample plate (violet), only the analyte MT (yellow), only the pure matrix harmine (green), harmine coated on MT (red), only the pure matrix PIVhar (blue) and PIVhar coated on MT (black). All measurements were performed with 40% laser intensity, m/z = 23 [Na]⁺, 39 [K]⁺, 71 [Har+H]⁺, 212 [Har]⁺, 227 [matrix]⁺, 341 [MT]⁺, 421 [Har+Har]⁺, 422 [Har+Har]⁺, 527 [MT+Na]⁺.

^a see Table S1 and Table S2 for abbreviations; ^b clear assignment not possible

Supporting Information

Figure S75: Relative positive mode MALDI-MS speed of the sample plate (violet), only the analyte MA (yellow), only the pure matrix trimine (green), trimine coated on MA (red), only the pure matrix *PV(PhH)* (blue) and *PV(PhH)* coated on MA (black). Positive and negative mode measurements were performed with 30% and 20% laser intensity, respectively. *m/z* = 23 [Na⁺], 35 [Cl⁻], 39 [K⁺], 211 [PhH-H⁺], 213 [PhH-H⁺], 227 [matrix⁺], 227 [MA-H⁺], 251 [MA-Na⁺], 422 [Hh-H-Ha⁺], 53 [MAH⁺].

Analyte: Linoleic acid (LA)

Figure S76. Reflection positive (left) and negative (right) mode MALDI-MS spectra of the sample plate (violet), on the analyte LA (yellow), the pure matrix *harmine* (green), *harmine* coated on LA (red), only the pure matrix *PyH4* (blue) and *PyH4* coated on LA (black). All measurements were performed with 20% laser intensity. *m/z* [234] [361] [367] [211] [364] $\cdot$ 1, [212] $\cdot$ 2, [213] $\cdot$ 3, [214] $\cdot$ 4, [215] $\cdot$ 5, [216] $\cdot$ 6, [217] $\cdot$ 7, [218] $\cdot$ 8, [219] $\cdot$ 9, [220] $\cdot$ 10, [221] $\cdot$ 11, [222] $\cdot$ 12, [223] $\cdot$ 13, [224] $\cdot$ 14, [225] $\cdot$ 15, [226] $\cdot$ 16, [227] $\cdot$ 17, [228] $\cdot$ 18, [229] $\cdot$ 19, [230] $\cdot$ 20, [231] $\cdot$ 21, [232] $\cdot$ 22, [233] $\cdot$ 23, [234] $\cdot$ 24, [235] $\cdot$ 25, [236] $\cdot$ 26, [237] $\cdot$ 27, [238] $\cdot$ 28, [239] $\cdot$ 29, [240] $\cdot$ 30, [241] $\cdot$ 31, [242] $\cdot$ 32, [243] $\cdot$ 33, [244] $\cdot$ 34, [245] $\cdot$ 35, [246] $\cdot$ 36, [247] $\cdot$ 37, [248] $\cdot$ 38, [249] $\cdot$ 39, [250] $\cdot$ 40, [251] $\cdot$ 41, [252] $\cdot$ 42, [253] $\cdot$ 43, [254] $\cdot$ 44, [255] $\cdot$ 45, [256] $\cdot$ 46, [257] $\cdot$ 47, [258] $\cdot$ 48, [259] $\cdot$ 49, [260] $\cdot$ 50, [261] $\cdot$ 51, [262] $\cdot$ 52, [263] $\cdot$ 53, [264] $\cdot$ 54, [265] $\cdot$ 55, [266] $\cdot$ 56, [267] $\cdot$ 57, [268] $\cdot$ 58, [269] $\cdot$ 59, [270] $\cdot$ 60, [271] $\cdot$ 61, [272] $\cdot$ 62, [273] $\cdot$ 63, [274] $\cdot$ 64, [275] $\cdot$ 65, [276] $\cdot$ 66, [277] $\cdot$ 67, [278] $\cdot$ 68, [279] $\cdot$ 69, [280] $\cdot$ 70, [281] $\cdot$ 71, [282] $\cdot$ 72, [283] $\cdot$ 73, [284] $\cdot$ 74, [285] $\cdot$ 75, [286] $\cdot$ 76, [287] $\cdot$ 77, [288] $\cdot$ 78, [289] $\cdot$ 79, [290] $\cdot$ 80, [291] $\cdot$ 81, [292] $\cdot$ 82, [293] $\cdot$ 83, [294] $\cdot$ 84, [295] $\cdot$ 85, [296] $\cdot$ 86, [297] $\cdot$ 87, [298] $\cdot$ 88, [299] $\cdot$ 89, [300] $\cdot$ 90, [301] $\cdot$ 91, [302] $\cdot$ 92, [303] $\cdot$ 93, [304] $\cdot$ 94, [305] $\cdot$ 95, [306] $\cdot$ 96, [307] $\cdot$ 97, [308] $\cdot$ 98, [309] $\cdot$ 99, [310] $\cdot$ 100, [311] $\cdot$ 101, [312] $\cdot$ 102, [313] $\cdot$ 103, [314] $\cdot$ 104, [315] $\cdot$ 105, [316] $\cdot$ 106, [317] $\cdot$ 107, [318] $\cdot$ 108, [319] $\cdot$ 109, [320] $\cdot$ 110, [321] $\cdot$ 111, [322] $\cdot$ 112, [323] $\cdot$ 113, [324] $\cdot$ 114, [325] $\cdot$ 115, [326] $\cdot$ 116, [327] $\cdot$ 117, [328] $\cdot$ 118, [329] $\cdot$ 119, [330] $\cdot$ 120, [331] $\cdot$ 121, [332] $\cdot$ 122, [333] $\cdot$ 123, [334] $\cdot$ 124, [335] $\cdot$ 125, [336] $\cdot$ 126, [337] $\cdot$ 127, [338] $\cdot$ 128, [339] $\cdot$ 129, [340] $\cdot$ 130, [341] $\cdot$ 131, [342] $\cdot$ 132, [343] $\cdot$ 133, [344] $\cdot$ 134, [345] $\cdot$ 135, [346] $\cdot$ 136, [347] $\cdot$ 137, [348] $\cdot$ 138, [349] $\cdot$ 139, [350] $\cdot$ 140, [351] $\cdot$ 141, [352] $\cdot$ 142, [353] $\cdot$ 143, [354] $\cdot$ 144, [355] $\cdot$ 145, [356] $\cdot$ 146, [357] $\cdot$ 147, [358] $\cdot$ 148, [359] $\cdot$ 149, [360] $\cdot$ 150, [361] $\cdot$ 151, [362] $\cdot$ 152, [363] $\cdot$ 153, [364] $\cdot$ 154, [365] $\cdot$ 155, [366] $\cdot$ 156, [367] $\cdot$ 157, [368] $\cdot$ 158, [369] $\cdot$ 159, [370] $\cdot$ 160, [371] $\cdot$ 161, [372] $\cdot$ 162, [373] $\cdot$ 163, [374] $\cdot$ 164, [375] $\cdot$ 165, [376] $\cdot$ 166, [377] $\cdot$ 167, [378] $\cdot$ 168, [379] $\cdot$ 169, [380] $\cdot$ 170, [381] $\cdot$ 171, [382] $\cdot$ 172, [383] $\cdot$ 173, [384] $\cdot$ 174, [385] $\cdot$ 175, [386] $\cdot$ 176, [387] $\cdot$ 177, [388] $\cdot$ 178, [389] $\cdot$ 179, [390] $\cdot$ 180, [391] $\cdot$ 181, [392] $\cdot$ 182, [393] $\cdot$ 183, [394] $\cdot$ 184, [395] $\cdot$ 185, [396] $\cdot$ 186, [397] $\cdot$ 187, [398] $\cdot$ 188, [399] $\cdot$ 189, [400] $\cdot$ 190, [401] $\cdot$ 191, [402] $\cdot$ 192, [403] $\cdot$ 193, [404] $\cdot$ 194, [405] $\cdot$ 195, [406] $\cdot$ 196, [407] $\cdot$ 197, [408] $\cdot$ 198, [409] $\cdot$ 199, [410] $\cdot$ 200, [411] $\cdot$ 201, [412] $\cdot$ 202, [413] $\cdot$ 203, [414] $\cdot$ 204, [415] $\cdot$ 205, [416] $\cdot$ 206, [417] $\cdot$ 207, [418] $\cdot$ 208, [419] $\cdot$ 209, [420] $\cdot$ 210, [421] $\cdot$ 211, [422] $\cdot$ 212, [423] $\cdot$ 213, [424] $\cdot$ 214, [425] $\cdot$ 215, [426] $\cdot$ 216, [427] $\cdot$ 217, [428] $\cdot$ 218, [429] $\cdot$ 219, [430] $\cdot$ 220, [431] $\cdot$ 221, [432] $\cdot$ 222, [433] $\cdot$ 223, [434] $\cdot$ 224, [435] $\cdot$ 225, [436] $\cdot$ 226, [437] $\cdot$ 227, [438] $\cdot$ 228, [439] $\cdot$ 229, [440] $\cdot$ 230, [441] $\cdot$ 231, [442] $\cdot$ 232, [443] $\cdot$ 233, [444] $\cdot$ 234, [445] $\cdot$ 235, [446] $\cdot$ 236, [447] $\cdot$ 237, [448] $\cdot$ 238, [449] $\cdot$ 2

^a see Table S1 and Table S2 for abbreviations; ^b clear assignment not possible

- 535 -

Figure S77: Reflection positive (left) and negative (right) mode MALDI/MS spectra of the sample plate (violet), only the analyte DA (yellow), only the pure matrix (harmine (green), harmine coated on DA (red), only the pure matrix *PivHaeI* (blue) and *PivHaeII* coated on DA (black). Positive and negative mode measurements were performed with 30% and 20% laser intensity, respectively. $m/z = 23$ [Na]⁺, 211 [H⁺H⁺]⁺, 212 [H⁺H⁺]⁺, 213 [H⁺H⁺H⁺]⁺, 327 [DA+H]⁺, 205 [H⁺Na]⁺, 351 [DA⁺Na]⁺, 421 [H⁺H⁺H⁺]⁺.

Analysat: Dipalmitoyl-Glycerol (DG)

Figure S78. Relative (left) positive (left) and negative (right) model MALDI-MS spectra of the sample plate (yellow), only the analyte DG (yellow), only the pure matrix *harmine* (green), *harmine* coated on DG (red), only the pure matrix *P1(har)* (blue) and *P1(har)* coated on DG (black). All measurements were performed with 50% laser intensity. Please note the differences in Y-axis scale, m/z [23 (Na⁺), 38 (K⁺), 211 (Har-H⁺), 213 (Har-H⁺), 227 (maltin⁺), 255 (DG-H⁺), 422 (Har-H⁺), 551 (DG-Na⁺), 607 (DG-H⁺), 633 (Har-H⁺), 655 (DG-H⁺), 657 (DG-H⁺), 659 (DG-H⁺), 661 (DG-H⁺), 663 (Har-H⁺), 665 (DG-H⁺), 667 (DG-H⁺), 669 (DG-H⁺), 671 (DG-H⁺), 673 (DG-H⁺), 675 (DG-H⁺), 677 (DG-H⁺), 679 (DG-H⁺), 681 (DG-H⁺), 683 (DG-H⁺), 685 (DG-H⁺), 687 (DG-H⁺), 689 (DG-H⁺), 691 (DG-H⁺), 693 (DG-H⁺), 695 (DG-H⁺), 697 (DG-H⁺), 699 (DG-H⁺), 701 (DG-H⁺), 703 (DG-H⁺), 705 (DG-H⁺), 707 (DG-H⁺), 709 (DG-H⁺), 711 (DG-H⁺), 713 (DG-H⁺), 715 (DG-H⁺), 717 (DG-H⁺), 719 (DG-H⁺), 721 (DG-H⁺), 723 (DG-H⁺), 725 (DG-H⁺), 727 (DG-H⁺), 729 (DG-H⁺), 731 (DG-H⁺), 733 (DG-H⁺), 735 (DG-H⁺), 737 (DG-H⁺), 739 (DG-H⁺), 741 (DG-H⁺), 743 (DG-H⁺), 745 (DG-H⁺), 747 (DG-H⁺), 749 (DG-H⁺), 751 (DG-H⁺), 753 (DG-H⁺), 755 (DG-H⁺), 757 (DG-H⁺), 759 (DG-H⁺), 761 (DG-H⁺), 763 (DG-H⁺), 765 (DG-H⁺), 767 (DG-H⁺), 769 (DG-H⁺), 771 (DG-H⁺), 773 (DG-H⁺), 775 (DG-H⁺), 777 (DG-H⁺), 779 (DG-H⁺), 781 (DG-H⁺), 783 (DG-H⁺), 785 (DG-H⁺), 787 (DG-H⁺), 789 (DG-H⁺), 791 (DG-H⁺), 793 (DG-H⁺), 795 (DG-H⁺), 797 (DG-H⁺), 799 (DG-H⁺), 801 (DG-H⁺), 803 (DG-H⁺), 805 (DG-H⁺), 807 (DG-H⁺), 809 (DG-H⁺), 811 (DG-H⁺), 813 (DG-H⁺), 815 (DG-H⁺), 817 (DG-H⁺), 819 (DG-H⁺), 821 (DG-H⁺), 823 (DG-H⁺), 825 (DG-H⁺), 827 (DG-H⁺), 829 (DG-H⁺), 831 (DG-H⁺), 833 (DG-H⁺), 835 (DG-H⁺), 837 (DG-H⁺), 839 (DG-H⁺), 841 (DG-H⁺), 843 (DG-H⁺), 845 (DG-H⁺), 847 (DG-H⁺), 849 (DG-H⁺), 851 (DG-H⁺), 853 (DG-H⁺), 855 (DG-H⁺), 857 (DG-H⁺), 859 (DG-H⁺), 861 (DG-H⁺), 863 (DG-H⁺), 865 (DG-H⁺), 867 (DG-H⁺), 869 (DG-H⁺), 871 (DG-H⁺), 873 (DG-H⁺), 875 (DG-H⁺), 877 (DG-H⁺), 879 (DG-H⁺), 881 (DG-H⁺), 883 (DG-H⁺), 885 (DG-H⁺), 887 (DG-H⁺), 889 (DG-H⁺), 891 (DG-H⁺), 893 (DG-H⁺), 895 (DG-H⁺), 897 (DG-H⁺), 899 (DG-H⁺), 901 (DG-H⁺), 903 (DG-H⁺), 905 (DG-H⁺), 907 (DG-H⁺), 909 (DG-H⁺), 911 (DG-H⁺), 913 (DG-H⁺), 915 (DG-H⁺), 917 (DG-H⁺), 919 (DG-H⁺), 921 (DG-H⁺), 923 (DG-H⁺), 925 (DG-H⁺), 927 (DG-H⁺), 929 (DG-H⁺), 931 (DG-H⁺), 933 (DG-H⁺), 935 (DG-H⁺), 937 (DG-H⁺), 939 (DG-H⁺), 941 (DG-H⁺), 943 (DG-H⁺), 945 (DG-H⁺), 947 (DG-H⁺), 949 (DG-H⁺), 951 (DG-H⁺), 953 (DG-H⁺), 955 (DG-H⁺), 957 (DG-H⁺), 959 (DG-H⁺), 961 (DG-H⁺), 963 (DG-H⁺), 965 (DG-H⁺), 967 (DG-H⁺), 969 (DG-H⁺), 971 (DG-H⁺), 973 (DG-H⁺), 975 (DG-H⁺), 977 (DG-H⁺), 979 (DG-H⁺), 981 (DG-H⁺), 983 (DG-H⁺), 985 (DG-H⁺), 987 (DG-H⁺), 989 (DG-H⁺), 991 (DG-H⁺), 993 (DG-H⁺), 995 (DG-H⁺), 997 (DG-H⁺), 999 (DG-H⁺), 1001 (DG-H⁺), 1003 (DG-H⁺), 1005 (DG-H⁺), 1007 (DG-H⁺), 1009 (DG-H⁺), 1011 (DG-H⁺), 1013 (DG-H⁺), 1015 (DG-H⁺), 1017 (DG-H⁺), 1019 (DG-H⁺), 1021 (DG-H⁺), 1023 (DG-H⁺), 1025 (DG-H⁺), 1027 (DG-H⁺), 1029 (DG-H⁺), 1031 (DG-H⁺), 1033 (DG-H⁺), 1035 (DG-H⁺), 1037 (DG-H⁺), 1039 (DG-H⁺), 1041 (DG-H⁺), 1043 (DG-H⁺), 1045 (DG-H⁺), 1047 (DG-H⁺), 1049 (DG-H⁺), 1051 (DG-H⁺), 1053 (DG-H⁺), 1055 (DG-H⁺), 1057 (DG-H⁺), 1059 (DG-H⁺), 1061 (DG-H⁺), 1063 (DG-H⁺), 1065 (DG-H⁺), 1067 (DG-H⁺), 1069 (DG-H⁺), 1071 (DG-H⁺), 1073 (DG-H⁺), 1075 (DG-H⁺), 1077 (DG-H⁺), 1079 (DG-H⁺), 1081 (DG-H⁺), 1083 (DG-H⁺), 1085 (DG-H⁺), 1087 (DG-H⁺), 1089 (DG-H⁺), 1091 (DG-H⁺), 1093 (DG-H⁺), 1095 (DG-H⁺), 1097 (DG-H⁺), 1099 (DG-H⁺), 1101 (DG-H⁺), 1103 (DG-H⁺), 1105 (DG-H⁺), 1107 (DG-H⁺), 1109 (DG-H⁺), 1111 (DG-H⁺), 1113 (DG-H⁺), 1115 (DG-H⁺), 1117 (DG-H⁺), 1119 (DG-H⁺), 1121 (DG-H⁺), 1123 (DG-H⁺), 1125 (DG-H⁺), 1127 (DG-H⁺), 1129 (DG-H⁺), 1131 (DG-H⁺), 1133 (DG-H⁺), 1135 (DG-H⁺), 1137 (DG-H⁺), 1139 (DG-H⁺), 1141 (DG-H⁺), 1143 (DG-H⁺), 1145 (DG-H⁺), 1147 (DG-H⁺), 1149 (DG-H⁺), 1151 (DG-H⁺), 1153 (DG-H⁺), 1155 (DG-H⁺), 1157 (DG-H⁺), 1159 (DG-H⁺), 1161 (DG-H⁺), 1163 (DG-H⁺), 1165 (DG-H⁺), 1167 (DG-H⁺), 1169 (DG-H⁺), 1171 (DG-H⁺), 1173 (DG-H⁺), 1175 (DG-H⁺), 1177 (DG-H⁺), 1179 (DG-H⁺), 1181 (DG-H⁺), 1183 (DG-H⁺), 1185 (DG-H⁺), 1187

^a see Table S1 and Table S2 for abbreviations; ^b clear assignment not possible

- 536 -

3.6.3.1. Choline chloride

3.6.3.1.1. DHB / PVDHB

Figure S82: Full reflection positive mode MALDI MS spectra of pure DHB (green), DHB coated on CC (0.001 mg/ml) (red), pure PVDHB (blue) and PVDHB coated on CC (0.001 mg/ml) (black). All spectra were measured with 30% laser intensity.

1. Choline chloride (CC) sample (positive mode)

Figure S83: First reflection positive mode MALDI MS spectra of CC (top: 10^3 mg/ml, middle: 10^{-4} mg/ml, bottom: 10^{-5} mg/ml) coated with DHB (left) and PVDHB (right). All spectra were measured with 30% laser intensity. Please mind the differences in Y-axis scale. $m/z = 23$ [Na]⁺, 39 [K]⁺, 86 [CC]⁺, 104 [CC], 154 [DHB]⁺, 177 [DHB+Na]⁺, 193 [DHB+K]⁺.

^a see Table S1 and Table S2 for abbreviations; ^b clear assignment not possible

- S39 -

3.6.3.1.2. Harmine / PVHAr

Full spectrum of choline chloride (CC):

Figure S84: Full reflection positive mode MALDI MS spectra of alpure harmine (green), harmine coated on CC (0.01 mg/ml) (red), pure PVHAr (blue) and PVHAr coated on CC (0.01 mg/ml) (black). All spectra were measured with 30% laser intensity. Please mind the differences in Y-axis scale.

1. Choline chloride (CC) sample (positive mode)

Figure S85: First reflection positive mode MALDI MS spectra of CC (top: 10^3 mg/ml, middle: 10^{-4} mg/ml, bottom: 10^{-5} mg/ml) coated with harmine (left) and PVHAr (right). All spectra were measured with 30% laser intensity. Please mind the differences in Y-axis scale. $m/z = 23$ [Na]⁺, 39 [K]⁺, 86 [CC]⁺, 104 [CC], 213 [HAr+H]⁺, 227 [HAr+P], 235 [HAr+Na]⁺, 422/423 [HAr+HAr]⁺, 437 [HAr+P].

^a see Table S1 and Table S2 for abbreviations; ^b clear assignment not possible

- S40 -

3.6.3.2.1. DHB / P(VDHB)

Full spectra of Phosphocholine (PC):

Figure S88: Full reflection positive mode MALDI MS spectra of pure DHB (green), DHB coated on PC (1 mg/ml) (red), pure P(VDHB) (blue) and P(VDHB) coated on PC (1 mg/ml) (black). All spectra were measured with 40% laser intensity. Please mind differences in Y-axis scale.

Figure S89: Full reflection negative mode MALDI MS spectra of pure DHB (green), DHB coated on PC (2 mg/ml) (red), pure P(VDHB) (blue) and P(VDHB) coated on PC (2 mg/ml) (black). All spectra were measured with 40% laser intensity.

3.6.3.2. Phosphocholine chloride (PC)

The molecule ion of PC at $m/z = 184$ was not detectable with any matrix. Hence, PC was measured without matrix (LDI MS) to see which signals can be expected and belong to PC although they might not be assignable.

Figure S86: LDI spectra of pure PC in positive mode (blue) and negative mode (black). All spectra were measured with 100% laser intensity. Please note the different Y-axis scales.

According to these spectra the favored signal for LoD-measurements in positive mode is at $m/z = 122$ and in negative mode at $m/z = 161$. Nevertheless, also the other observable signals are of interest.

Figure S87: First reflection negative mode MALDI MS spectra of PC (top: 10^{-4} mg/ml, middle: 10^{-5} mg/ml, bottom: 10^{-6} mg/ml) coated with 9AA. All spectra were measured with 40% laser intensity. $m/z = 161$ (anale)^a, 163 (anale)^a, 165 (anale)^a, 193 (9AA-H⁺).

^a see Table S1 and Table S2 for abbreviations; ^b clear assignment not possible; ^c see Figure S86

1. Phosphocholine (PC) sample (positive mode)

Figure S90. First reflection positive mode MALDI MS spectra of PC (top: 10^3 mg/ml, middle: 10^2 mg/ml, bottom: 10^1 mg/ml) coated with DHB (left) and PIVdHB (right). All spectra were measured with 40% laser intensity. Please mind the differences in Y-axis scale. $m/z = 23$ [Na]⁺, 39 [K]⁺, 58 [analyte]⁺, 72 [analyte]⁺, 86 [CC⁺]⁺, 104 [CC⁺]⁺, 122 [analyte]⁺, 137 [DHB-H]⁺, 154 [DHB]⁺, 177 [DHB+Na]⁺, 193 [DHB+K]⁺.

1. Phosphocholine (PC) sample (negative mode)

Figure S91. First reflection negative mode MALDI MS spectra of PC (top: 10^3 mg/ml, middle: 10^2 mg/ml, bottom: 10^1 mg/ml) coated with DHB (left) and PIVdHB (right). All spectra were measured with 40% laser intensity. Please mind the differences in Y-axis scale. $m/z = 153$ [DHB-H]⁻, 154 [DHB-H]⁻, 161 [analyte]⁻, 165 [analyte]⁻, 169 [analyte]⁻, 171 [analyte]⁻, 175 [analyte]⁻, 251 [analyte]⁻, 307 [DHB-DHB-H]⁻, 315 [m+H]⁻.

^a see Table S1 and Table S2 for abbreviations; ^b clear assignment not possible; ^c see Figure S86

- S43 -

3.6.3.2.2. Harmine / PIVHarm

Full spectra of Phosphocholine (PC):

Figure S92. Full reflection positive mode MALDI MS spectra of pure harmine (green), harmine coated on PC (red), pure PIVHarm (blue) and PIVHarm coated on PC (black). All spectra were measured with 40% laser intensity. Please mind differences in Y-axis scale.

Figure S93. Full reflection negative mode MALDI MS spectra of pure harmine (green), harmine coated on PC (red), pure PIVHarm (blue) and PIVHarm coated on PC (black). All spectra were measured with 40% laser intensity.

- S44 -

3.6.3.3. Glycerophosphorylcholine
The molecule ion of GIPC at $m/z = 257$ was not detectable with any matrix. Hence, GIPC was measured without matrix (LDI MS) to see which signals can be expected and belong to GIPC although they might not be assignable.

Figure S96: MALDI-spectra of pure GIPC in positive mode (blue) and negative mode (black). All spectra were measured with 100% laser intensity. Please note the different Y-axis scales.

According to these spectra the favored signal for LoD-measurements in positive mode is at $m/z = 139$ and in negative mode at $m/z = 242$. Nevertheless, also the other observable signals are of interest.

Figure S97: First reflectron negative mode MALDI MS spectra of GIPC (top: 10^{-3} mg/ml, middle: 10^{-2} mg/ml, bottom: 10^{-1} mg/ml) coated with 9AA. All spectra were measured with 40% laser intensity. $m/z = 163$ (analyte)^c, 193 (9AA-H⁺), 194 (9AA)^a, 242 (analyte)^c.

^a see Table S1 and Table S2 for abbreviations; ^b clear assignment not possible; ^c see Figure S96

- S46 -

1. Phosphocholine (PC) sample (positive mode)

Figure S94: First reflectron positive mode MALDI MS spectra of PC (top: 10^{-1} mg/ml, middle: 10^{-2} mg/ml, bottom: 10^{-3} mg/ml) coated with harmonic (left) and PIVHar (right). All spectra were measured with 40% laser intensity. Please mind the differences in Y-axis scale. $m/z = 23$ [Na⁺], 39 [K⁺], 58 (analyte)^c, 72 (analyte)^c, 86 [CCl₃]⁺, 104 [CCl₃]⁺, 122 (analyte)^c, 198 (matrix)^b, 213 [Har+H⁺], 227 (matrix)^b, 298 (matrix)^b, 422/423 [Har+Har]^c, 437 (matrix)^c.

1. Phosphocholine (PC) sample (negative mode)

Figure S95: First reflectron negative mode MALDI MS spectra of PC (top: 10^{-1} mg/ml, middle: 10^{-2} mg/ml, bottom: 10^{-3} mg/ml) coated with harmonic (left) and PIVHar (right). All spectra were measured with 40% laser intensity. Please mind the differences in Y-axis scale. $m/z = 35$ [Cl⁻], 161 (analyte)^c, 163 (analyte)^c, 169 (analyte)^c, 171 (analyte)^c, 173 (analyte)^c, 175 (analyte)^c, 211 [Har+H⁺], 247 [Har+Cl⁻], 251 (analyte)^c, 257 [Har+Har]^c.

^a see Table S1 and Table S2 for abbreviations; ^b clear assignment not possible; ^c see Figure S86

- S45 -

3.6.3.3.1. DHB / PVDHB

Full spectra of Glycerophosphorylcholine (GPC):

Figure S98: Full reflection positive mode MALDI MS spectra of pure DHB (green), DHB coated on GIPC (1 mg/ml) (red), pure PVDHB (blue) and PVDHB coated on GIPC (1 mg/ml) (black). All spectra were measured with 40% laser intensity.

Figure S99: Full reflection negative mode MALDI MS spectra of pure DHB (green), DHB coated on GIPC (2 mg/ml) (red), pure PVDHB (blue) and PVDHB coated on GIPC (2 mg/ml) (black). All spectra were measured with 40% laser intensity. Please mind differences in Y-axis scale.

- S47 -

1. Glycerophosphorylcholine (GPC) sample (positive mode)

Figure S100: First reflection positive mode MALDI MS spectra of GIPC (top: 10^{-1} mg/ml, middle: 10^{-2} mg/ml, bottom: 10^{-3} mg/ml) coated with DHB (left) and PVDHB (right). All spectra were measured with 40% laser intensity. Please mind the differences in Y-axis scale. $m/z = 23$ [Na], 39 [K], 72 (analyle), 86 [CC1⁺], 104 [C], 136 [DHB-OH+H⁺], 139 (analyle), 154 [DHB⁺], 177 [DHB-Na⁺], 193 [DHB-K⁺], 272 (analyle).

1. Glycerophosphorylcholine (GPC) sample (negative mode)

Figure S101: First reflection negative mode MALDI MS spectra of GIPC (top: 2 mg/ml, middle: 10^{-1} mg/ml, bottom: 10^{-2} mg/ml) coated with DHB (left) and PVDHB (right). All spectra were measured with 40% laser intensity. Please mind the differences in Y-axis scale. $m/z = 63$ (analyle), 79 (analyle), 123 (analyle), 153 [DHB-H⁺], 163 (analyle), 171 (analyle), 197 (analyle), 242 (analyle), 307 [DHB-DHB-H⁺], 315 [DHB-H⁺].

* see Table S1 and Table S2 for abbreviations; * clear assignment not possible; * see Figure S96

- S48 -

1. Glycerophosphorylcholine (GPC) sample (positive mode)

Figure S104: First reflection positive mode MALDI MS spectra of GPC (top: 10^{-2} mg/ml, middle: 10^{-1} mg/ml, bottom: 10^0 mg/ml) coated with pure P(VHar) (blue) and pure Har (red). All spectra were measured with 40% laser intensity. Please mind the differences in Y-axis scale. $m/z = 23$ [Na]⁺, 39 [K]⁺, 40 [Ca]⁺, 72 (analyte)⁺, 86 [CCl₄]⁺, 104 [CCl₄]⁺, 139 (analyte)⁺, 213 [Har+H]⁺, 227 (matrix)⁺, 242 (matrix)⁺, 315 (matrix)⁺, 406/407 (matrix)⁺, 422/423 [Har+Har]⁺.

1. Glycerophosphorylcholine (GPC) sample (negative mode)

Figure S105: First reflection negative mode MALDI MS spectra of GPC (top: 10^{-2} mg/ml, middle: 10^{-1} mg/ml, bottom: 10^0 mg/ml) coated with pure P(VHar) (blue) and pure Har (red). All spectra were measured with 40% laser intensity. Please mind the differences in Y-axis scale. $m/z = 35$ [Cl]⁻, 63 (analyte)⁻, 79 (analyte)⁻, 97 (analyte)⁻, 111 (analyte)⁻, 123 (analyte)⁻, 137 (analyte)⁻, 153 (analyte)⁻, 163 (analyte)⁻, 168 (analyte)⁻, 171 (analyte)⁻, 183 (analyte)⁻, 197 (analyte)⁻, 211 [Har+H]⁻, 242 (matrix)⁻, 247 [Har+Cl]⁻, 421 [Har+Har]⁻.

^a see Table S1 and Table S2 for abbreviations; ^b clear assignment not possible; ^c see Figure S96

- S50 -

3.6.3.3.2. Harmine / P(VHar)

Full spectra of Glycerophosphorylcholine (PC):

Figure S102: Full reflection positive mode MALDI MS spectra of pure harmine (green), harmine coated on GPC (1 mg/ml) (red), pure P(VHar) (blue) and P(VHar) coated on GPC (1 mg/ml) (black). All spectra were measured with 40% laser intensity. Please mind differences in Y-axis scale.

Figure S103: Full reflection negative mode MALDI MS spectra of pure harmine (green), harmine coated on GPC (2 mg/ml) (red), pure P(VHar) (blue) and P(VHar) coated on GPC (2 mg/ml) (black). All spectra were measured with 40% laser intensity. Please mind differences in Y-axis scale.

- S49 -

3.6.4. MALDI Imaging

Figure S106: Effect of the solvent composition on ionic intensity and spatial localization. Overall average mass spectra of tissue regions and heatmaps of the ion at $m/z = 846.4$ when using 75% MeOH and 25% H_2O (red), 50% MeOH and 50% H_2O (blue) and 50% MeOH, 30% H_2O and 20% acetone (black) in the preparation of P(VDHB) solutions.

Figure S107: Optical picture of a MALDI MSI sample covered with P(VDHB).

Figure S108: HE staining of patient-derived breast cancer xenograft. The enlargement shows a morphologically representative region of the breast cancer tissue.

3.6.4.1. MSI Comparison P(VDHB) - DHB

$m/z = 72.111 \pm 27$ mDa

Figure S109: Visualization of an ion at $m/z = 72.111$ in a heatmap measured with P(VDHB) (left) and DHB (right).

$m/z = 268.907 \pm 43$ mDa

Figure S111: Visualization of an ion at $m/z = 268.907$ in a heatmap measured with P(VDHB) (left) and DHB (right).

$m/z = 469.077 \pm 92$ mDa

Figure S113: Visualization of an ion at $m/z = 469.077$ in a heatmap measured with P(VDHB) (left) and DHB (right).

$m/z = 155.036 \pm 71$ mDa

Figure S110: Visualization of the protonated molecule ion of DHB at $m/z = 155.036$ in a heatmap measured with P(VDHB) (left) and DHB (right).

$m/z = 402.668 \pm 27$ mDa

Figure S112: Visualization of an ion at $m/z = 402.668$ in a heatmap measured with P(VDHB) (left) and DHB (right).

$m/z = 486.626 \pm 62$ mDa

Figure S114: Visualization of an ion at $m/z = 486.626$ in a heatmap measured with P(VDHB) (left) and DHB (right).

Table S13: Potential origins of m/z signals detected with P(VDHB) in positive mode.

m/z	Potential origin	m/z	Potential origin
72.1	unknown	666.5	glucosylceramide ^{15:17} [M+Na] ⁺
89.0	pyruvic acid ^{18:20} [M+H] ⁺	762.4	PC 32:1 ^{15:17} [M+K] ⁺
114.0	glycine ^{18:20} [M+K] ⁺	794.4	PS 34:4 ^{15:17} [M+K] ⁺
268.0	alpha-ketoglutarate ^{18:20} [M+H] ⁺	798.5	PC 34:1 ^{15:17} [M+Na] ⁺
229.9	unknown	874.4	PS 40:6 ^{15:17} [M+K] ⁺
360.7	unknown	920.5	lipids PS, CDP, PC ^{15:17}
462.6	unknown		

3.6.4.3. Enlargements of ionic maps with P(VDHB) (negative mode)

$m/z = 102.979 \pm 55$ mDa

Figure S123: Enlarged ionic map of a breast cancer xenograft model cryo-section coated with P(VDHB). The map was measured in negative reflection mode and visualizes the ion at $m/z = 102.979 \pm 55$ mDa.

$m/z = 279.22 \pm 25$ mDa

Figure S125: Enlarged ionic map of a breast cancer xenograft model cryo-section coated with P(VDHB). The map was measured in negative reflection mode and visualizes the ion at $m/z = 279.22 \pm 25$ mDa.

$m/z = 283.251 \pm 25$ mDa

Figure S127: Enlarged ionic map of a breast cancer xenograft model cryo-section coated with P(VDHB). The map was measured in negative reflection mode and visualizes the ion at $m/z = 283.251 \pm 25$ mDa.

$m/z = 255.231 \pm 25$ mDa

Figure S124: Enlarged ionic map of a breast cancer xenograft model cryo-section coated with P(VDHB). The map was measured in negative reflection mode and visualizes the ion at $m/z = 255.231 \pm 25$ mDa.

$m/z = 281.248 \pm 39$ mDa

Figure S126: Enlarged ionic map of a breast cancer xenograft model cryo-section coated with P(VDHB). The map was measured in negative reflection mode and visualizes the ion at $m/z = 281.248 \pm 39$ mDa.

$m/z = 702.427 \pm 25$ mDa

Figure S128: Enlarged ionic map of a breast cancer xenograft model cryo-section coated with P(VDHB). The map was measured in negative reflection mode and visualizes the ion at $m/z = 702.427 \pm 25$ mDa.

3.6.4.2. Enlargements of ionic maps with P(VDHB) (positive mode)

$m/z = 113.991 \pm 27$ mDa

Figure S115: Enlarged ionic map of a breast cancer xenograft model cryo-section coated with P(VDHB). The map was measured in positive reflection mode and visualizes the ion at $m/z = 113.991 \pm 27$ mDa.

$m/z = 666.449 \pm 27$ mDa

Figure S117: Enlarged ionic map of a breast cancer xenograft model cryo-section coated with P(VDHB). The map was measured in positive reflection mode and visualizes the ion at $m/z = 666.449 \pm 27$ mDa.

$m/z = 794.391 \pm 96$ mDa

Figure S119: Enlarged ionic map of a breast cancer xenograft model cryo-section coated with P(VDHB). The map was measured in positive reflection mode and visualizes the ion at $m/z = 794.391 \pm 96$ mDa.

$m/z = 874.449 \pm 27$ mDa

Figure S121: Enlarged ionic map of a breast cancer xenograft model cryo-section coated with P(VDHB). The map was measured in positive reflection mode and visualizes the ion at $m/z = 874.449 \pm 27$ mDa.

$m/z = 762.427 \pm 27$ mDa

Figure S118: Enlarged ionic map of a breast cancer xenograft model cryo-section coated with P(VDHB). The map was measured in positive reflection mode and visualizes the ion at $m/z = 762.427 \pm 27$ mDa.

$m/z = 798.475 \pm 27$ mDa

Figure S120: Enlarged ionic map of a breast cancer xenograft model cryo-section coated with P(VDHB). The map was measured in positive reflection mode and visualizes the ion at $m/z = 798.475 \pm 27$ mDa.

$m/z = 920.474 \pm 216$ mDa

Figure S122: Enlarged ionic map of a breast cancer xenograft model cryo-section coated with P(VDHB). The map was measured in positive reflection mode and visualizes the ion at $m/z = 920.474 \pm 216$ mDa.

- S54 -

- S53 -

$m/z = 708.387 \pm 45$ mDa

$m/z = 764.391 \pm 25$ mDa

Figure S129: Enlarged ionic map of a breast cancer xenograft model cryo-section coated with PVDHB). The map was measured in negative reflection mode and visualizes the ion at $m/z = 708.387 \pm 45$ mDa.

Figure S130: Enlarged ionic map of a breast cancer xenograft model cryo-section coated with PVDHB). The map was measured in negative reflection mode and visualizes the ion at $m/z = 764.391 \pm 25$ mDa.

Table S14: Potential origins of m/z signals detected with PVDHB) in negative mode.

m/z	Potential origin	m/z	Potential origin
99.0	unknown	324.7	unknown
103.0	3-hydroxypyruvic acid ¹⁸⁻²⁰ [M-H] ⁺	428.7	unknown
166.9	unknown	465.0	unknown
255.2	FA 16:0 ^{15-17, 21} [M-H] ⁺	518.6	unknown
279.2	FA 18:2 ^{15-17, 21} [M-H] ⁺	702.4	PS 30:2 ¹⁵⁻¹⁷ [M-H] ⁺
281.2	FA 18:1 ^{15-17, 21} [M-H] ⁺	708.4	PC 28:2 ¹⁵⁻¹⁷ [M+Cl] ⁺
283.2	FA 18:0 ^{15-17, 21} [M-H] ⁺	764.4	PS 32:3 ¹⁵⁻¹⁷ [M+Cl] ⁺

3.6.4.4. Enlargements of ionic maps with PVDHB) (positive mode)

$m/z = 86.057 \pm 258$ mDa

$m/z = 104.106 \pm 194$ mDa

Figure S131: Enlarged ionic map of a breast cancer xenograft model cryo-section coated with PVDHB). The map was measured in positive reflection mode and visualizes the ion at $m/z = 86.057 \pm 258$ mDa.

Figure S132: Enlarged ionic map of a breast cancer xenograft model cryo-section coated with PVDHB). The map was measured in positive reflection mode and visualizes the ion at $m/z = 104.106 \pm 194$ mDa.

$m/z = 118.106 \pm 77$ mDa

$m/z = 184.071 \pm 129$ mDa

Figure S133: Enlarged ionic map of a breast cancer xenograft model cryo-section coated with PVDHB). The map was measured in positive reflection mode and visualizes the ion at $m/z = 118.106 \pm 77$ mDa.

Figure S134: Enlarged ionic map of a breast cancer xenograft model cryo-section coated with PVDHB). The map was measured in positive reflection mode and visualizes the ion at $m/z = 184.071 \pm 129$ mDa.

$m/z = 258.074 \pm 77$ mDa

$m/z = 504.243 \pm 77$ mDa

Figure S135: Enlarged ionic map of a breast cancer xenograft model cryo-section coated with PVDHB). The map was measured in positive reflection mode and visualizes the ion at $m/z = 258.074 \pm 77$ mDa.

Figure S136: Enlarged ionic map of a breast cancer xenograft model cryo-section coated with PVDHB). The map was measured in positive reflection mode and visualizes the ion at $m/z = 504.243 \pm 77$ mDa.

$m/z = 738.427 \pm 272$ mDa

Figure S137: Enlarged ionic map of a breast cancer xenograft model cryo-section coated with PVDHB). The map was measured in positive reflection mode and visualizes the ion at $m/z = 738.427 \pm 272$ mDa.

$m/z = 874.338 \pm 272$ mDa

Figure S138: Enlarged ionic map of a breast cancer xenograft model cryo-section coated with PVDHB). The map was measured in positive reflection mode and visualizes the ion at $m/z = 874.338 \pm 272$ mDa.

Table S15: Potential origins of m/z signals detected with PVDHB) in positive mode.

m/z	Potential origin	m/z	Potential origin
86.0	fragment choline ^{15-17, 21}	295.2	FA 16:0 ^{15-17, 21} [M+K] ⁺
104.1	choline ¹⁸⁻²¹	504.3	LysOPC/LysoPE
116.0	proline ¹⁸⁻²⁰ [M+H] ⁺	554.5	unknown
118.1	valine ¹⁸⁻²⁰ [M+H] ⁺	666.5	Glucosylceramide ¹⁵⁻¹⁷
147.0	alpha-ketoglutarate ¹⁸⁻²⁰ [M+H] ⁺	748.7	unknown
162.1	carnitine ²¹ [M+H] ⁺	798.4	Lipids PE, PC, PS ¹⁵⁻¹⁷
166.1	phenylalanine ¹⁸⁻²⁰ [M+H] ⁺	844.5	unknown
184.1	phosphocholine ¹⁸⁻²⁰	874.3	Lipids PS ¹⁵⁻¹⁷
258.1	glycerophosphocholine ¹⁵⁻²⁰ [M+H] ⁺		

3.6.4.5. Enlargements of ionic maps with PVDHB) (negative mode)

$m/z = 87.041 \pm 100$ mDa

$m/z = 89.067 \pm 66$ mDa

Figure S139: Enlarged ionic map of a breast cancer xenograft model cryo-section coated with PVDHB). The map was measured in negative reflection mode and visualizes the ion at $m/z = 87.041 \pm 100$ mDa.

Figure S140: Enlarged ionic map of a breast cancer xenograft model cryo-section coated with PVDHB). The map was measured in negative reflection mode and visualizes the ion at $m/z = 89.067 \pm 66$ mDa.

4. References

- Layton, M. E.; Morales, C. A.; Shair, M. D. Biomimetic Synthesis of (-)-Longithorone A. *J. Am. Chem. Soc.* **2002**, *124* (5), 773-775.
- Adamski-Werner, S. L.; Palaninathan, S. K.; Sacchetti, J. C.; Kelly, J. W. Diflunisal Analogues Stabilize the Native State of Transferrin. Potent Inhibition of Amyloidogenesis. *J. Med. Chem.* **2004**, *47*, 355-374.
- Galan, L. A.; Cordes, D. B.; Slawin, A. M. Z.; Jacquemin, D.; Ogden, M. I.; Massi, M.; Zysman-Colman, E. Analyzing the Relation between Structure and Aggregation Induced Emission (AIE) Properties of Iridium(III) Complexes through Modification of Non-Chromophoric Ancillary Ligands. *Eur. J. Inorg. Chem.* **2019**, 152-163.
- Zhong, B.; Sun, L.; Shi, H.; Li, J.; Chen, C.; Chen, Z. **2017**, WO 2017/176812 A1.
- Kimura, M.; Sakai, R.; Sato, S.; Fukawa, T.; Ikehara, T.; Maeda, R.; Mihara, T. Sensing of Volatile Organic Compounds by TiO₂ Porous Films Covered with Polythiophene Layers. *Adv. Funct. Mater.* **2012**, *22*, 469-476.
- He, Y.; Zhao, H.; Yao, M.; Weiss, R. G. Complex New Materials from Simple Chemistry: Combining an Amino-Substituted Polysiloxane and Carboxylic Acids. *Journal of Polymer Science, Part A: Polymer Chemistry* **2017**, *55*, 3851-3861.
- Gothard, C. M.; Rao, N. A.; Nowick, J. S. Nanometer-Sized Amino Acids for the Synthesis of Nanometer-Scale Water-Soluble Molecular Rods of Precise Length. *J. Am. Chem. Soc.* **2007**, *129*, 7272-7273.
- Ponce, M. A.; Erra-Balsells, R. Synthesis and Isolation of Bromo- β -carbolines obtained by Bromination of β -Carboline Alkaloids. *J. Heterocyclic Chem.* **2001**, *38*, 1087.
- Holle, A.; Ketterlinus, R. *Innovative smartbeam Laser Technology Enhances MALDI-TOF Based Proteomics Applications*. Bruker Daltonics GmbH; 2008; Vol. Technical Note #TN-12.
- Horatz, K.; Giamprà, M.; Karpov, Y.; Sahre, K.; Bednarz, H.; Kiry, A.; Voit, B.; Niehaus, K.; Hadjichristidis, N.; Michels, D. L.; Lissel, F. Conjugated Polymers as a New Class of Dual-Mode Matrices for MALDI Mass Spectrometry and Imaging. *J. Am. Chem. Soc.* **2018**, *140*, 11416-11423.
- Sladkova, K.; Houska, J.; Havel, J. Laser desorption/ionization of red phosphorus clusters and their use for mass calibration in time-of-flight mass spectrometry. *Rapid Commun. Mass Spectrom.* **2009**, *23*, 3114-3118.
- Desmazières, B.; Buchmann, W.; Terrier, P.; Tortajada, J. *Anal. Chem.* **2008**, *80*, 783-792.
- Karas, M.; Hillenkamp, F. Laser Desorption/Ionization of Proteins with Molecular Masses Exceeding 10 000 Daltons. *Anal. Chem.* **1988**, *60*, 2299-2301.
- Meier, M. A. R.; Hoogenboom, R.; Filten, M. W. M.; Schneider, M.; Schubert, U. S. Automated MALDI-TOF-MS Sample Preparation in Combinatorial Polymer Research. *J. Comb. Chem.* **2003**, *5*, 269-374.
- Jiang, L.; Chughtai, K.; Purvine, S.; Bhujwalla, Z. M.; Raman, V.; Paša-Toli, L.; Heeren, R. M. A.; Glunde, K. MALDI-mass spectrometric imaging reveals hypoxia-driven lipids and proteins in a breast tumor model. *Anal. Chem.* **2015**, *87* (12), 5947-5956.
- Angerer, T. B.; Magnusson, Y.; Landberg, G.; Fletcher, J. S. Lipid Heterogeneity Resulting from Fatty Acid Processing in the Human Breast Cancer Microenvironment Identified by GC/MS-TOF-SIMS Imaging. *Anal. Chem.* **2016**, *88* (23), 11946-11954.
- Holzlechner, M.; Eugenin, E.; Pradeaux, B. Mass spectrometry imaging to detect lipid biomarkers and disease signatures in cancer. *Cancer Reports* **2019**, *2*, e1229.
- Mattami, K. R.; Sullivan, M. R.; Heiden, M. G. V. The importance of serine metabolism in cancer. *J. Cell Biol.* **2016**, *214* (3), 249-257.
- Dai, C.; Arceo, J.; Arnold, J.; Sreekumar, A.; Dovichi, N. J.; Littlepage, L. E. Metabolomics of oncogene-specific metabolic reprogramming during breast cancer. *Cancer & Metabolism* **2018**, *6* (5), 1-17.
- Grinde, M. T.; Hilmarsdottir, B.; Tunset, H. M.; Henriksen, I. M.; Kim, J.; Haugen, M. H.; Rye, M. B.; Mælandsmo, G. M.; Moestue, S. A. Glutamine to proline conversion is associated with response to glutamine inhibition in breast cancer. *Breast Cancer Research* **2019**, *21* (61), 1-13.
- Sun, C.; Wang, F.; Zhang, Y.; Yu, J.; Wang, X. Mass spectrometry imaging-based metabolomics to visualize the spatially resolved reprogramming of carnitine metabolism in breast cancer. *Theranostics* **2020**, *10* (16), 7070-7082.

 $m/z = 123.986 \pm 66 \text{ mDa}$

Figure S141: Enlarged ionic map of a breast cancer xenograft model cryo-section coated with P(VHAr). The map was measured in negative reflection mode and visualizes the ion at $m/z = 123.986 \pm 66 \text{ mDa}$.

 $m/z = 283.254 \pm 159 \text{ mDa}$

Figure S143: Enlarged ionic map of a breast cancer xenograft model cryo-section coated with P(VHAr). The map was measured in negative reflection mode and visualizes the ion at $m/z = 283.254 \pm 159 \text{ mDa}$.

 $m/z = 835.443 \pm 66 \text{ mDa}$

Figure S145: Enlarged ionic map of a breast cancer xenograft model cryo-section coated with P(VHAr). The map was measured in negative reflection mode and visualizes the ion at $m/z = 835.443 \pm 66 \text{ mDa}$.

 $m/z = 145.068 \pm 66 \text{ mDa}$

Figure S142: Enlarged ionic map of a breast cancer xenograft model cryo-section coated with P(VHAr). The map was measured in negative reflection mode and visualizes the ion at $m/z = 145.068 \pm 66 \text{ mDa}$.

 $m/z = 303.221 \pm 86 \text{ mDa}$

Figure S144: Enlarged ionic map of a breast cancer xenograft model cryo-section coated with P(VHAr). The map was measured in negative reflection mode and visualizes the ion at $m/z = 303.221 \pm 86 \text{ mDa}$.

 $m/z = 887.483 \pm 66 \text{ Da}$

Figure S146: Enlarged ionic map of a breast cancer xenograft model cryo-section coated with P(VHAr). The map was measured in negative reflection mode and visualizes the ion at $m/z = 887.483 \pm 66 \text{ Da}$.

Table S16: Potential origins of m/z signals detected with P(VHAr) in negative mode.

m/z	Potential origin	m/z	Potential origin
87.0	pyruvic acid ^{18,20} [M-H] ⁺	283.2	FA 18:0 ^{15,17,21} [M-H] ⁺
89.1	lactate ^{18,20} [M-H] ⁺	303.2	FA 20:4 ^{15,17,21} [M-H] ⁺
114.1	proline ^{18,20} [M-H] ⁺	346.0	unknown
124.0	taurine ^{18,20} [M-H] ⁺	835.4	lipids PI ¹⁵⁻¹⁷
145.1	glutamine ^{18,20} [M-H] ⁺	857.5	PI 36:4 ¹⁵⁻¹⁷ [M-H] ⁺
255.2	FA 16:0 ^{15-17,21} [M-H] ⁺	885.5	PI 38:4 ¹⁵⁻¹⁷ [M-H] ⁺
281.2	FA 18:1 ^{15-17,21} [M-H] ⁺	887.5	Lipids PGP ^{15,17}

- S58 -

- S57 -

8.3. Copyright & Permission

Dear Dr. Horatz,

Thank you for contacting ACS Publications Support.

Your permission requested is granted and there is no fee for this reuse. In your planned reuse, you must cite the ACS article as the source, add this direct link <<https://pubs.acs.org/doi/10.1021/jacs.8b06637>>, and include a notice to readers that further permissions related to the material excerpted should be directed to the ACS.

Sincerely,

Budimir Jelic
ACS Publications
Customer Services & Information
Website: <https://help.acs.org>

Incident Information:

Incident #: 3970214
Date Created: 2020-11-25T10:23:36
Priority: 3
Customer: Kilian Horatz

License Details

This Agreement between Mr. Kilian Horatz ("You") and John Wiley and Sons ("John Wiley and Sons") consists of your license details and the terms and conditions provided by John Wiley and Sons and Copyright Clearance Center.

Print

Copy

License Number	4955820411500
License date	Nov 25, 2020
Licensed Content Publisher	John Wiley and Sons
Licensed Content Publication	CHEMPLUSCHEM
Licensed Content Title	Amorphous Conjugated Polymers as Efficient Dual-Mode MALDI Matrices for Low-Molecular-Weight Analytes
Licensed Content Author	Franziska Lissel, Brigitte Voit, Anton Kiri, et al
Licensed Content Date	Jul 10, 2019
Licensed Content Volume	84
Licensed Content Issue	9
Licensed Content Pages	8
Type of Use	Dissertation/Thesis
Requestor type	Author of this Wiley article
Format	Print and electronic
Portion	Full article
Will you be translating?	No
Title	Development of novel polymer matrices for MALDI MS/imaging
Institution name	Leibniz Institute for Polymer Research
Expected presentation date	Apr 2021
Requestor Location	Mr. Kilian Horatz Agnes-Smedley-Straße 8 Dresden, 01187 Germany Attn: Mr. Kilian Horatz EU826007151
Publisher Tax ID	

Appendix

Copyright & Permission

Polymerization as a Strategy to Improve Small Organic Matrices for Low-Molecular-Weight Compound Analytics with MALDI MS and MALDI MS Imaging

Author: Kilian Horatz, Marco Giampà, Zhi Qiao, et al

Publication: ACS Applied Polymer Materials

Publisher: American Chemical Society

Date: Jul 1, 2021

Copyright © 2021, American Chemical Society

PERMISSION/LICENSE IS GRANTED FOR YOUR ORDER AT NO CHARGE

This type of permission/license, instead of the standard Terms and Conditions, is sent to you because no fee is being charged for your order. Please note the following:

- Permission is granted for your request in both print and electronic formats, and translations.
- If figures and/or tables were requested, they may be adapted or used in part.
- Please print this page for your records and send a copy of it to your publisher/graduate school.
- Appropriate credit for the requested material should be given as follows: "Reprinted (adapted) with permission from {COMPLETE REFERENCE CITATION}. Copyright {YEAR} American Chemical Society." Insert appropriate information in place of the capitalized words.
- One-time permission is granted only for the use specified in your RightsLink request. No additional uses are granted (such as derivative works or other editions). For any uses, please submit a new request.

If credit is given to another source for the material you requested from RightsLink, permission must be obtained from that source.

BACK

CLOSE WINDOW

Declaration / Eidesstattliche Erklärung

Hiermit versichere ich, dass ich die vorliegende Arbeit ohne unzulässige Hilfe Dritter und ohne Benutzung anderer als der angegebenen Hilfsmittel angefertigt habe; die aus fremden Quellen direkt oder indirekt übernommenen Gedanken sind als solche kenntlich gemacht. Die Arbeit wurde bisher weder im Inland noch im Ausland in gleicher oder ähnlicher Form einer anderen Prüfungsbehörde vorgelegt.

Die vorliegende Arbeit wurde unter der wissenschaftlichen Betreuung von Frau Dr. Franziska Lissel und Frau Prof. Dr. Brigitte Voit im Zeitraum Oktober 2017 bis März 2021 am Leibniz-Institut für Polymerforschung Dresden e.V. angefertigt.

Dresden, den 25.06.2021

Kilian Horatz